

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED

Precise Radiation Oncology Center
Precise Radiation Oncology Center
Carratera Marginal #2 Esquina 3B, Suite G1 (GPS 18
Urb. Hermanas Davila
Bayamon, PR 00960

REPORT NUMBER(S) 2025001

2. NRC/REGIONAL OFFICE

Region 1
475 Allendale Rd
Suite 102
King of Prussia, PA 19406

3. DOCKET NUMBER(S)

030-38713

4. LICENSE NUMBER(S)

52-35129-01

5. DATE(S) OF INSPECTION

05/01/2025 - 05/20/2025

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

License Condition 13 of Amendment No. 2 to NRC License No. 52-35129-01 required, in part, that except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below.

The section titled "Radiation Safety Training Program" contained in the application dated February 6, 2014, listed in License Condition 13.A of Amendment No. 2 to NRC License No. 52-35129-01 required, in part, that all radiation workers and ancillary personnel whose duties required them to work in the vicinity of radioactive materials or radiation producing machines would receive instruction. The training frequency was listed, in part, as before assuming duties with, or in the vicinity of, radioactive materials and during annual refresher training.

Contrary to the above, for the calendar years 2022 and 2023, the licensee failed to conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed in the license condition. Specifically, all radiation workers and ancillary personnel whose duties required them to work in the vicinity of radioactive materials or radiation producing machines did not receive annual refresher training. The licensee performed treatments in November 2022 and January and February 2023 without associated calendar year 2022 or 2023 annual refresher training for the required staff.

This is a Severity Level IV violation (Enforcement Policy 6.3).

As corrective action, the licensee committed to conducting radiation safety training prior to performing licensed activities in the future, including generating an annual reminder to perform refresher training. Following the onsite inspection, the licensee submitted a request to transition into "Standby – No Operations" mode. Prior to transitioning back to active use status, the licensee has committed to reperforming radiation safety training for all required personnel.

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED

Precise Radiation Oncology Center
Precise Radiation Oncology Center
Carratera Marginal #2 Esquina 3B, Suite G1 (GPS 18
Urb. Hermanas Davila
Bayamon, PR 00960

REPORT NUMBER(s) 2025001

2. NRC/REGIONAL OFFICE

Region 1
475 Allendale Rd
Suite 102
King of Prussia, PA 19406

3. DOCKET NUMBER(S)

030-38713

4. LICENSE NUMBER(S)


52-35129-01

5. DATE(S) OF INSPECTION

05/01/2025 - 05/20/2025

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE	Adriana Cordova Ayuso, RSO		5/23/25
NRC INSPECTOR	Jonathan Pfingsten		
BRANCH CHIEF	Farrah Gaskins		