

Materials Inspection Report

1. Licensee/Location	n Inspected:		2. NRC/Regional Office		
Cardiac Imaging, Inc. 2 TransAm Plaza Drive Oakbrook Terrace, IL 60181 Location Inspected: Job site in Merrillville, IN Report Number(s) 2025-002			Region III U. S. Nuclear Regulatory Commission 2056 Westings Avenue, Suite 400 Naperville, IL 60563-2657		
3. Docket Number(s)		4 License N	umbor(s)		5. Date(s) of Inspection
150-00012)	4. License Number(s) IL-02531-01 (reciprocity)			05/13/25 - 06/30/25
		12 02001			00/10/20 - 00/00/20
Nuclear Regulatory	an examination of the activities co Commission (NRC) rules and reg ocedures and representative recor	julations and t	the conditions of your license. Th	e inspect	ion consisted of selective
✓ 1. Based of	on the inspection findings, no viola	itions were ide	entified.		
2. Previou	s violation(s) closed.				
	this inspection, certain of your acti seessed at Severity Level IV, in ac			ere in viol	ation of NRC requirements, and
identifi were s	plation(s), specifically described to ed, non-repetitive, corrective actio atisfied. hited violation(s) was/were discussion	n was or is be	eing taken, and the remaining crit		
which	lowing violation(s) is/are being cite may be subject to posting in accor ions and Corrective Actions)			y. This fo	rm is a NOTICE OF VIOLATION,
		Statement of	of Corrective Actions		
actions is made in a	thin 30 days, the actions described by ccordance with the requirements of 1 hieved). I understand that no further w	0 CFR 2.201 (corrective steps already taken, corr	rective ste	ps which will be taken, date when ful
TITLE	PRINTED NAME SIGNAT			GNATURE	AND DATE
LICENSEE'S REPRESENTATIVE	(not required)				
NRC INSPECTOR	Ryan Craffey		Poly Coffrey		Digitally signed by RYAN CRAFFEY Date: 2025.07.07 08:27:44 -05'00'
BRANCH CHIEF	Rhex Edwards		12	_ \	Digitally signed by RHEX EDWARDS Date: 2025.07.18 10:01:15 -05'00'
				<i>U</i>	

NRC FORM 592M (10-04-2022) .sar Regu,			U.S. NU	CLEAR REGULATORY COMMISSION	
TO OF ZOZZ	Materials Insp	pection Record			
1. Licensee Name:	2. Docket Num	2. Docket Number(s):		3. License Number(s)	
Cardiac Imaging, Inc.	150-0001	150-00012		IL-02531-01 (reciprocity)	
4. Report Number(s):	!	5. Date(s) of Inspection:	· ·		
2025-002		May 13, 2025; in-office review through June 30, 2025			
6. Inspector(s):		7. Program Code(s):	8. Priority:	9. Inspection Guidance Used:	
Ryan Craffey		02220	3	IP 87130	
10. Licensee Contact Name(s):	11. Licensee E-mail Address:	E-mail Address:		12. Licensee Telephone Number(s):	
Anthony Wallace RT(R)(N)	awallace@cardiacscan.org		772-341-1574		
13. Inspection Type: Initial 14. I	ocations Inspected: Hyb	rid 15. Next Inspection I	Date (MM/DD/YY	YY):	
Routine Announced	Main Office Fiel	d Office TE	3D	Normal Extended	
✓ Non-Routine ✓ Unannounced ✓	Temporary Job Site Ren	note		Reduced 🗸 No change	
16. Location(s) Inspected List:					
Queen of Hearts Cardiovascular,	, 8684 Connecticut Str	eet, Merrillville, IN 464	10		

17. Scope and Observations:

Cardiac Imaging, a mobile medical imaging service provider, was authorized by the State of Illinois to use byproduct material for diagnostic administrations of radiopharmaceuticals at its facility in Mokena, and at temporary job sites in State jurisdiction. The company performed work at temporary job sites in NRC jurisdiction per a reciprocity notification filed in January 2025 for the calendar year. The scope of this inspection was limited to an evaluation of licensed activities (cardiac PET scans using a Rb-82 infusion machine) on a PET/CT trailer parked at a client's facility in Merrillville, Indiana.

The inspector visited the trailer in Merrillville. The trailer was adequately secured, controlled, and posted. Independent surveys on its exterior throughout the day remained well below regulatory limits to members of the public. Independent surveys inside the trailer found no evidence of contamination. The inspector observed the conduct of several cardiac PET scans using weight-based doses of Rb-82 administered via infusion machine and QC testing (including generator breakthrough) prior to administrations. The NMT wore personnel dosimetry as assigned and was knowledgeable of infusion machine use and maintenance as well as safe handling practices for radioactive material. The NMT demonstrated exchanging the Rb-82 generator as well as blocking and bracing the infusion machine for transport using equipment and instructions provided by the manufacturer that appeared adequate for the purpose. All other calibration and reference sources were adequately secured for transport. Calibrated survey instruments were available for end-of-day surveys.

The inspector also reviewed a selection of records, including infusion machine preventative maintenance and training records, dose administration records, the licensee's MOU with the client in Merrillville, personnel dosimetry reports, consultant audits including instrument calibrations, quality control, and sealed source leak tests and inventories, sealed source return documentation, and Type A package test results for the infusion machine. The inspector also requested documentation of instructions from the sole AU on the Illinois license for cardiac PET scans. The licensee located these during the period of in-office review. The "radiopharmaceutical standing order" thus provided confirmed that the AU approved of the weight-based dosage in use.

NRC Form 592M (10-04-2022) Page 1 of 2

NRC FORM 592M (10-04-2022)	U.S. NUCLEAR REGULATORY COMMISSION					
(1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-	Materials Inspection Record (Continued)					
No violations of NRC requirements were identified as a result of this inspection.						