



Materials Inspection Report

1. Licensee/Location Inspected:

Henry Ford Health Systems
d/b/a Henry Ford Health
20200 W Outer Dr.
Dearborn, MI 48124

Report Number(s) 2025-001

2. NRC/Regional Office

Region III
U. S. Nuclear Regulatory Commission
2056 Westings Avenue, Suite 400
Naperville, IL 60563-2657

3. Docket Number(s)

030-38009

4. License Number(s)

21-32740-01

5. Date(s) of Inspection

7/14/2025

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



1. Based on the inspection findings, no violations were identified.



2. Previous violation(s) closed.



3. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements, and were assessed at Severity Level IV, in accordance with the NRC Enforcement Policy.

A. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy were satisfied.

(Non-cited violation(s) was/were discussed involving the following requirement(s))

B. The following violation(s) is/are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE AND DATE
LICENSEE'S REPRESENTATIVE		
NRC INSPECTOR	Daniel Fiedorek	DANIEL FIDOREK Digitally signed by DANIEL FIDOREK Date: 2025.07.24 10:43:02 -05'00'
BRANCH CHIEF	Rhex Edwards	 Digitally signed by RHEX EDWARDS Date: 2025.07.25 11:12:40 -05'00'



Materials Inspection Record

1. Licensee Name: Henry Ford Health System		2. Docket Number(s): 030-38009		3. License Number(s) 21-32740-01	
4. Report Number(s): 2025-001			5. Date(s) of Inspection: 7/14/2025		
6. Inspector(s): Daniel Fiedorek		7. Program Code(s): 02201		8. Priority: 5	9. Inspection Guidance Used: 87130
10. Licensee Contact Name(s): Alan Jackson, RSO		11. Licensee E-mail Address: alanj@rad.hfh.edu		12. Licensee Telephone Number(s): 313-916-2739	
13. Inspection Type: ☒ Routine ☐ Announced ☐ Non-Routine ☒ Unannounced		14. Locations Inspected: ☐ Main Office ☒ Field Office ☐ Temporary Job Site ☐ Remote		15. Next Inspection Date (MM/DD/YYYY): 7/14/2030 ☒ Normal ☐ Extended ☐ Reduced ☐ No change	
16. Location(s) Inspected List: 20200 W Outer Drive, Dearborn, MI 48124					
17. Scope and Observations: This was an unannounced routine inspection of a cardiac clinic in the Henry Ford Health system with locations in Dearborn and Gibraltar, MI. the licensee was authorized to possess and use byproduct material in accordance with 10 CFR 35.200. At the time of the inspection the licensee had recently been bought by Henry Ford Health and had just completed the transition from private practice to joining the Henry Ford system. At the Dearborn location the licensee conducted approximately 4-7 cardiac studies per day Monday-Thursday. The licensee had three nuclear medicine technologists (NMT) on staff at the Dearborn location. The licensee retained the services of a medical physicist for audits and instrument calibrations and maintenance. The inspection consisted of a tour of the licensee Dearborn clinic and records review. The Gibraltar location was not included in this inspection. The inspector toured the licensee clinic and conducted independent surveys that did not show radiation levels exceeding posting requirements in either restricted or unrestricted areas of the clinic. The NMT demonstrated daily instrument checks on the well counter and dose calibrator, demonstrated package acceptance surveys, and verified the locations of all sealed sources listed on the most recent inventory. The inspector discussed licensee emergency procedures, regulatory requirements, and radiation safety principles with the NMT and found the NMT to be knowledgeable of procedures, regulatory requirements, and radiation safety principles. No issues were identified from the tour of the licensee clinic. The inspector also reviewed licensee records of quarterly physicist audits and instrument calibrations, daily instrument checks, daily surveys, weekly wipes, survey meter calibrations, dose administration history, inventories and leak tests, decay in storage logs, training records, and dosimetry. No issues were identified from reviewing the licensee records. No violations of NRC requirements were identified from this inspection.					

Materials Inspection Record (Continued)

Signature and Date - Branch Chief



Digitally signed by RHEX EDWARDS
Date: 2025.07.25 11:12:19 -05'00'

From: [Jackson, Alan \(RSO\)](#)
To: [Daniel Fiedorek](#)
Subject: [External_Sender] RE: NRC Inspection Reports for Henry Ford Wyandotte and Henry Ford Health
Date: Friday, July 25, 2025 11:55:28 AM
Attachments: image002.png

Thanks Daniel!

I hope that I am able to meet you in person for the next inspection.

-Alan

Alan M. Jackson, MS. CHP

Radiation Safety Officer

Henry Ford Hospital
2799 West Grand Boulevard
Detroit, MI 48202
(734) 657-4133 Cell
E-mail: AlanJ@rad.hfh.edu

Pronouns: He, him, his (gender neutral pronouns can also be used)

HENRY FORD HEALTH

From: Daniel Fiedorek <Daniel.Fiedorek@nrc.gov>
Sent: Friday, July 25, 2025 12:46 PM
To: Jackson, Alan (RSO) <alanj@rad.hfh.edu>
Subject: NRC Inspection Reports for Henry Ford Wyandotte and Henry Ford Health

Caution: External Email!

Do not open attachments or click on links if you do not recognize the sender.
If this email is suspicious, please click the *Report Phish* button in your Outlook app.

Hi Alan

Please see the attached NRC 591Ms documenting my inspections of Henry Ford Wyandotte and Henry Ford Health (docket no. 030-38009). As discussed earlier this week, no violations were identified for either license, and no further actions are required beyond keeping a copy of both records with their respective licenses.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter, its enclosure, and your response, if you choose to provide one, will be made available electronically for public inspection in the NRC's Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC's website at <http://www.nrc.gov/reading-rm/adams.html>. To the extent

possible, your response should not include any personal privacy, proprietary, or safeguards information so that it can be made publicly available without redaction.

Thanks!

Daniel

Daniel Fiedorek
U.S. NRC | Region III
Division of Radiological Safety and Security (DRSS)
Materials Inspection Branch (MIB)
daniel.fiedorek@nrc.gov | 630-829-9836



CONFIDENTIALITY NOTICE: This email contains information from the sender that may be CONFIDENTIAL, LEGALLY PRIVILEGED, PROPRIETARY or otherwise protected from disclosure. This email is intended for use only by the person or entity to whom it is addressed. If you are not the intended recipient, any use, disclosure, copying, distribution, printing, or any action taken in reliance on the contents of this email, is strictly prohibited. If you received this email in error, please contact the sending party by reply email, delete the email from your computer system and shred any paper copies.

Note to Patients: There are a number of risks you should consider before using e-mail to communicate with us. See our Privacy & Security page on www.henryford.com for more detailed information as well as information concerning MyChart, our new patient portal. If you do not believe that our policy gives you the privacy and security protection you need, do not send e-mail or Internet communications to us.