

**From:** [Jason VonEhr](#)  
**To:** [Lyons, Erik](#); [ext Christopher Smitherman](#); [radrem@aol.com](mailto:radrem@aol.com)  
**Cc:** [Steger, Ted](#); [Hisel, Gregory](#)  
**Subject:** NRC HDR Amendments - 2 of 4 (St. Vincent's Med Ctr)  
**Date:** Tuesday, August 19, 2025 10:39:00  
**Attachments:** image002.jpg  
image003.jpg

---

Erik & Co. (I've also included Greg and Gerald),

Regarding the St. Vincent's HDR amendment, and consistent with discussions last week, the amendment request letter should be revised to address:

- Your request did not state whether there was a change in the emergency response equipment (NUREG-1556, Vol 9, Rev 3, Section 8.9.5);
- For occupational exposure monitoring (NUREG-1556, Vol 9, Rev 3, Section 8.10.2), your license currently has, more or less, the 'standard' NUREG commitment – please supplement this with a commitment to provide and require the use of extremity monitoring in the case of an HDR emergency response;
- Your request and your license do not clearly prohibit the installation, maintenance, adjustment, repair, or inspection of your HDR. Please provide a commitment that these activities will only be provided by a licensed third-party (e.g., the manufacturer) (NUREG-1556, Vol 9, Rev 3, Section 8.10.7);
- If you'd be so kind – the emergency procedure from the manufacturer were labeled "Attachment 3" on the document itself (but "attachment 2" as the pdf document title). Unless there was a further attachment intended, please either fix this (10 pages... 10 mentions of "Attachment 3") or clearly state that it was mislabeled in the cover letter so no one goes wondering where the 'ghost' attachment 2 went.

Given the accelerated timelines to complete CT-related actions, I would request the revised and mgmt.-signed response no-later-than Friday this week.

**Gerald** (Charlotte Hungerford): I'm catching up from being out-of-office Friday and Monday, but I did not see a communication from you whether Charlotte Hungerford was going to retain the request for "withdrawal" of the submitted amendment. Mr. Hisel and I spoke and he confirmed that his request would remain with the NRC.

All: Hartford's request is coming next.

Regards,

--

**Jason vonEhr**  
**Sr. Health Physicist**  
**Medical and Licensing Assistance Branch,**  
**Division of Radiological Safety and Security,**  
**U.S. Nuclear Regulatory Commission, Region I**  
**475 Allendale Road, Suite 102**  
**King of Prussia, PA 19406-1415**

**Office: (610) 337-5256**  
**Email: [jason.vonehr@nrc.gov](mailto:jason.vonehr@nrc.gov)**

DRSS logo



---

**From:** Lyons, Erik <Erik.Lyons@hhchealth.org>

**Sent:** Thursday, August 14, 2025 9:59 AM

**To:** Jason VonEhr <Jason.VonEhr@nrc.gov>

**Cc:** ext\_Christopher\_Smitherman <christopher.smitherman@hhchealth.org>; Steger, Ted <Ted.Steger@hhchealth.org>

**Subject:** [External\_Sender] Amendment request RAML 06-00843-03

Jason,

Attached, please find the amendment request and supporting documents to add a new HDR unit at St. Vincent's Medical Center. Thanks for your guidance with this.

Erik Lyons, RT (R)(N) ARRT, CNMT  
Assistant Radiation Safety Officer-RAM Use  
HHC Radiation Protection Program  
Hartford HealthCare  
Office: 860-972-2227  
Cell: 959-282-7735  
[erik.lyons@hhchealth.org](mailto:erik.lyons@hhchealth.org)



*This e-mail message, including any attachments, is for the sole use of the intended recipient(s) and may contain confidential and privileged information. Any unauthorized review, use, disclosure, or distribution is prohibited. If you are not the intended recipient, or an employee or agent responsible for delivering the message to the intended recipient, please contact the sender by reply e-mail and destroy all copies of the original message, including any attachments.*