



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2056 WESTINGS AVENUE, SUITE 400
NAPERVILLE, IL 60563-2657

September 15, 2025

Laura M. Luna, BS
Radiation Safety Officer
OSF Healthcare St. Francis Hospital
and Medical Group
3401 Ludington St.
Escabana, MI 49829

**SUBJECT: REQUEST FOR ADDITIONAL INFORMATION – PENDING NRC LICENSE
RENEWAL – OSF HEALTHCARE ST. FRANCIS HOSPITAL AND MEDICAL GROUP**

Dear Laura Luna,

We have reviewed your NRC license renewal request dated June 30, 2025 (ML25195A257), and have determined that additional information is required for continued processing of the licensing action. As such, your response is requested to address the following items:

1. When reviewing active medical licenses in the State of Michigan, I could not locate an active medical license for Dr. Nuha Krad. Please confirm that Dr. Krad should continue to be listed on the license as an Authorized User and provide their medical license number.
2. The facility diagram provided labeled as “OSF Healthcare St. Francis Hospital Nuclear Medicine Department” appears to indicate a limited area of the facility that was remodeled in 2022, as described in the letter dated April 18, 2022. Please clarify if this diagram indicates the full extent of the areas within the nuclear medicine department where radioactive materials will be used.
3. In the second facility diagram labelled as “OSF Healthcare St. Francis Hospital OVERVIEW OF PET AREA,” it is unclear what this diagram is intended to show. In the above referenced letter dated April 22, 2022, it appears that use of PET isotopes is limited to use within the trailer area only. Please clarify if there are any areas within the main facility where PET isotopes are used, such as patient waiting areas. Additionally, the resolution of the diagram provided is too low to read the provided labels and room numbers, particularly in the shaded area.
4. Please provide a copy of the shielding evaluation for each area where PET isotopes are used.
5. For PET radionuclide use, please provide a brief description of additional radiation safety equipment in use, such as remote handling devices and shielding available such as syringe shields, shielded waste containers, etc.

We request that you submit your response to this letter by Friday, October 17, 2025. In your response, please refer to the license, docket, and control number specified below.

If you have questions, require additional time to respond, or require clarification on any of the information stated above, please contact me at (630) 829-9712 or laura.cender@nrc.gov.

Sincerely,

LAURA CENDER  Digitally signed by LAURA CENDER
Date: 2025.09.15 11:11:08 -05'00'

Laura Cender
Sr. Health Physicist (Licensing)
Nuclear Regulatory Commission
Materials Licensing Branch, RIII

License No. 21-16481-01
Docket No. 030-11102
Control No. 647791

From: [Laura Cender](#)
To: [Laurie Luna](#)
Subject: Pending NRC License Renewal for OSF Healthcare St. Francis Hospital and Medical Group - Additional Information Required
Date: Monday, September 15, 2025 11:28:00 AM
Attachments: Request for Additional Information to OSF Healthcare St. Francis Hospital and Medical Group.pdf

Hello Laura,

I hope that you are well. I have reviewed your license renewal request dated June 30, 2025 for OSF Healthcare St. Francis Hospital and Medical Group, and I have attached a letter detailing the additional clarification required to finalize the licensing action.

Your response is requested by no later than Friday, October 17, 2025. Please reach out to me at 630-829-9712 or via email with any questions.

Thank you,

Laura

Laura Cender
Sr. Health Physicist
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712