



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION I
475 ALLENDALE ROAD – SUITE 102
KING OF PRUSSIA, PA 19406-1415

April 24, 2026

Ganesh Asaithambi, M.D., Owner, RSO
Beckley Cardiology, PLLC
1832 Harper Road, 2nd Floor
Beckley, WV 25801

**SUBJECT: BECKLEY CARDIOLOGY, PLLC, REQUEST FOR ADDITIONAL
INFORMATION, MAIL CONTROL NO. 658401**

Dear Ganesh Asaithambi:

This is in reference to your application received April 9, 2026, requesting to amend NRC License No. 47-35564-01. In order to continue our review, we need the following additional information. Please be aware that, unless stated otherwise, all “Item”, “Section”, and “Appendix” references below are referring to NUREG-1556, Volume 9, Revision 3, “Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses” which can be found at <https://www.nrc.gov/docs/ML1925/ML19256C219.pdf>.

1. *Item 3. Addresses Where Licensed Material Will be Used or Possessed* – In your application you submitted the address, “153 Mayor R. T. Wills, Beckley, WV 25801”, as a new location of use to be listed on your license. This address does not appear to be registered on any online maps at this time. Please confirm the address and, if it does not yet exist, provide GPS coordinates for the requested location of use.
2. *Item 9. Facility Diagram* – In your application you submitted a diagram of the facility you are requesting to be added to your license. Please indicate where radioactive materials will be used or stored. Additionally, for facility diagrams, NUREG-1556, Vol. 9, Rev. 3 requests the following:
 - a. Drawings should be to scale, and the scale used should be indicated. The direction of north should be indicated.
 - b. Doors should be indicated, and specify which doors are access controlled (i.e., locked).
 - c. NUREG-1556, Vol. 9, Rev. 3 also requests shielding calculations for areas of use involving PET radionuclides. Please confirm that you do not intend to use PET radionuclides;

OR

Provide shielding calculations for all areas of use involving PET radionuclides.

3. *Item 9. Radiation Monitoring Instruments* – In your application you submitted a brief description of the radiation instruments you intend to use at your facility. The description does not appear to include a dose calibrator. Please include a description of the dose calibrator or other instrument you intend to use to measure the activity of dosages prior to administration;

OR

Describe how you will verify the activity of dosages prior to administration.

We will continue our review upon receipt of this information. Please reply to my attention at:

R1DRSSMail.Resource@nrc.gov

Cc: CFR@nrc.gov

Reference – Christopher Rapson

Mail Control No. 658401

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

An electronic version of the NRC's regulations is available on the NRC Web Site at: www.nrc.gov. Additional information regarding medical uses of radioactive materials may be obtained on the NRC Web Site at: <http://www.nrc.gov/materials/miau/med-use-toolkit.html>. This site also provides the updated Training and Experience NRC Form 313A series of forms and guidance, as well as information on the revised regulations for naturally-occurring and accelerator-produced radioactive materials (NARM).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice and Procedure," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web Site at: <http://www.nrc.gov/reading-rm/adams.html>. Please be aware that you may request that certain portions of your submittal to the NRC be withheld from public disclosure as proprietary information. To do this, you must execute an affidavit as specified in 10 CFR 2.390. You must list all the portions that you wish to be held proprietary, along with your reasoning as to why that is appropriate. While it is allowable, please refrain from submitting proprietary information in support of a license unless necessary. Keep in mind that all NRC licenses are considered to be in the public domain and therefore may be viewed by any member of the public who requests to see them.

If you have any questions regarding this request for additional information, please contact me at 610-337-5198 or via electronic mail at CFR@nrc.gov.

Thank you for your cooperation.

Sincerely,

Christopher Rapson, Health Physicist
Medical and Licensing Assistance Branch
Division of Radiological Safety and Security
Region I

License No. 47-35564-01
Docket No. 030-39191
Mail Control No. 658401