



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION I
475 ALLENDALE ROAD – SUITE 102
KING OF PRUSSIA, PA 19406-1415

April 16, 2026

Darlene Baptiste, CEO
Governor Juan F. Luis Hospital and Medical Center
4007 Estate Diamond Ruby
Christiansted
St. Croix, VI 00820

SUBJECT: GOVERNOR JUAN F. LUIS HOSPITAL AND MEDICAL CENTER, REQUEST FOR ADDITIONAL INFORMATION, MAIL CONTROL NO. 644983

Dear Darlene Baptiste:

This is in reference to your application dated January 27, 2025, requesting to renew NRC License No. 55-25547-01 and your letter dated April 13, 2026, responding to our March 28, 2025, request for additional information (RAI). In order to continue our review, we need the following additional information. Please be aware that, unless stated otherwise, all “Item”, “Section”, and “Appendix” references below are referring to NUREG-1556, Volume 9, Revision 3, “Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses” found at <https://www.nrc.gov/docs/ML1925/ML19256C219.pdf>.

1. Items 5 and 6, Radioactive Material and Use, and Item 7, Authorized Users (AU) – Your RAI response letter included a description of a training course and recurring training needs but did not confirm that each of your Authorized Users had received this training recently (i.e., within the last 7 years) as requested per 10 CFR 35.59.

Please provide evidence of continuing education and experience for each of your AUs, including generator experience per 10 CFR 35.290(c)(1)(ii)(G), in support of your requested transition to active use status.

2. Item 9, Facility Diagram – Please confirm that the prior nuclear medicine areas for storage and use have not yet been released for unrestricted use.
3. Item 9, Radiation Monitoring Instruments – Your RAI response included the statement that you had developed and implemented survey meter calibration procedures; however, you are not currently authorized for materials necessary for the calibration of survey meters. As such, either confirm that radiation monitoring instruments will be calibrated by a vendor who is licensed by the NRC or an Agreement State to perform instrument calibrations or, if you desire to calibrate your own meters, additional licensed calibration materials would need to be authorized and possessed.

We will continue our review upon receipt of this information. Please reply to my attention at Jonathan.Pfingsten@nrc.gov.

In order to continue prompt review of your application, we request that you submit your response to this letter within 15 calendar days from the date of this letter.

An electronic version of the NRC's regulations is available on the NRC Web Site at: www.nrc.gov. Additional information regarding medical uses of radioactive materials may be obtained on the NRC Web Site at: <http://www.nrc.gov/materials/miau/med-use-toolkit.html>. This site also provides the updated Training and Experience NRC Form 313A series of forms and guidance, as well as information on the revised regulations for naturally-occurring and accelerator-produced radioactive materials (NARM).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web Site at: <http://www.nrc.gov/reading-rm/adams.html>. Please be aware that you may request that certain portions of your submittal to NRC be withheld from public disclosure as proprietary information. To do this, you must execute an affidavit as specified in 10 CFR 2.390. You must list all portions that you wish to be held proprietary, along with your reasoning as to why that is appropriate. While it is allowable, please refrain from submitting proprietary information in support of a license unless necessary. Keep in mind that all NRC licenses are considered to be in the public domain, and therefore may be viewed by any member of the public who requests to see them.

If you have any questions regarding this request for additional information, please contact me at (630) 829-9853 or via electronic mail at Jonathan.Pfingsten@nrc.gov.

Thank you for your cooperation.

Sincerely,

Jonathan Pfingsten, Senior Health Physicist
Materials Inspection Branch
Division of Radiological Safety and Security
Region III, on behalf of Region I

License No. 55-25547-01
Docket No. 030-35612
Mail Control No. 644983

cc: Angelo Galiber, Radiation Safety Officer

GOVERNOR JUAN F. LUIS HOSPITAL AND MEDICAL CENTER, REQUEST FOR
ADDITIONAL INFORMATION, MAIL CONTROL NO. 644983 DATED APRIL 16, 2026

SUNSI Review Complete: Jonathan Pfingsten

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/end "E" = Copy w/ attach/end "N" = No copy

OFFICE	RIII:DRSS	N					
NAME	JPfingsten/JBP						
DATE	4/16/2026						

OFFICIAL RECORD COPY