



Materials Inspection Report

1. Licensee/Location Inspected:

Cardiology Physicians, LLC
dba Cardiovascular Physicians of Delaware
One Centurian Drive, Suite 310
Newark, DE 19713

Report Number(s) 2026001

2. NRC/Regional Office

Region 1
475 Allendale Rd
Suite 102
King of Prussia, PA 19406

3. Docket Number(s)

030-35921

4. License Number(s)

07-30713-01

5. Date(s) of Inspection

05/13/2026 - 07/09/2026

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☒ 2. Previous violation(s) closed.
- ☒ 3. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements, and were assessed at Severity Level IV, in accordance with the NRC Enforcement Policy.

A. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy were satisfied.
(Non-cited violation(s) was/were discussed involving the following requirement(s))

B. The following violation(s) is/are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION.
which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

1) 10 CFR 35.67(b)(2) requires, in part, that a licensee in possession of a sealed source shall test the source for leakage at intervals not to exceed 6 months.

Contrary to the above, the licensee in possession of a sealed source did not test the source for leakage at intervals not to exceed 6 months. Specifically, the licensee performed sealed source leak testing on March 31, 2025 at their Middletown, DE location, and did not repeat this testing again until January 21, 2026, which is a period greater than 6 months.

This is a Severity Level IV violation (NRC Enforcement Policy Section 6.7).

Corrective actions included hiring a new health physics group, installing NMSS software at all locations, and retraining staff on sealed source regulatory requirements.

2) 10 CFR 35.67(g) requires, in part, that a licensee in possession of sealed sources shall conduct a semi-annual physical inventory of all such sources in their possession.

Contrary to the above, the licensee in possession of sealed sources did not conduct a semi-annual physical inventory of all such sources in their possession. Specifically, the licensee conducted a sealed source inventory on March 31, 2025 at their Middletown, DE location, and did not repeat the inventory again until January 21, 2026, which is a period greater than semi-annual.

This is a Severity Level IV violation (NRC Enforcement Policy Section 6.7).

Corrective actions included hiring a new health physics group, installing NMSS software at all locations, and retraining staff on sealed source regulatory requirements.



Materials Inspection Report (Continued)

1. Licensee/Location Inspected:

Cardiology Physicians, LLC
dba Cardiovascular Physicians of Delaware
One Centurian Drive, Suite 310
Newark, DE 19713

Report Number(s) 2026001

2. NRC/Regional Office

Region 1
475 Allendale Rd
Suite 102
King of Prussia, PA 19406

3. Docket Number(s)

030-35921

4. License Number(s)

07-30713-01

5. Date(s) of Inspection

05/13/2026 - 07/09/2026

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE AND DATE
LICENSEE'S REPRESENTATIVE	Brian Sarter, M.D.	 Brian Sarter (Jul 10, 2026 09:53:21 EDT)
NRC INSPECTORS	Janice Nguyen, Netra Patel JANICE NGUYEN	 Digitally signed by JANICE NGUYEN Date: 2026.07.09 13:27:12 -04'00' NETRA PATEL Digitally signed by NETRA PATEL Date: 2026.07.09 13:36:22 -04'00'
BRANCH CHIEF	Monica Ford	 MONICA FORD Digitally signed by MONICA FORD Date: 2026.07.14 07:49:51 -04'00'

R07-30713-01.2026-001.NRC Form 591M Part 1

Final Audit Report

2026-07-10

Created:	2026-07-09
By:	Zoila Rodriguez (zoila@ushphealth.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAfsVCsJG3kkFED14Y52fCmm6YXfHY0LVq

"R07-30713-01.2026-001.NRC Form 591M Part 1" History

-  Document digitally presigned by JANICE NGUYEN (Janice.Nguyen@nrc.gov)
2026-07-09 - 5:27:12 PM GMT
-  Document digitally presigned by NETRA PATEL (netra.patel@nrc.gov)
2026-07-09 - 5:36:22 PM GMT
-  Document created by Zoila Rodriguez (zoila@ushphealth.com)
2026-07-09 - 7:38:00 PM GMT
-  Document emailed to brian@cardiocppa.com for signature
2026-07-09 - 7:38:24 PM GMT
-  Email viewed by brian@cardiocppa.com
2026-07-10 - 1:52:33 PM GMT
-  Signer brian@cardiocppa.com entered name at signing as Brian Sarter
2026-07-10 - 1:53:19 PM GMT
-  Document e-signed by Brian Sarter (brian@cardiocppa.com)
Signature Date: 2026-07-10 - 1:53:21 PM GMT - Time Source: server - Signature Appearance Selected: TYPE
-  Agreement completed.
2026-07-10 - 1:53:21 PM GMT