

August 3, 1999

SECY-99-201

FOR: The Commissioners

FROM: William D. Travers /s/
Executive Director for Operations

SUBJECT: DRAFT FINAL RULE - 10 CFR PART 35, "MEDICAL USE OF
BYPRODUCT MATERIAL"

PURPOSE:

This paper provides the Commission with: (1) draft final rule language for the revision of 10 CFR Part 35; (2) a summary of the public comments received on the proposed rule, the staff's draft responses to the comments, and resulting changes made in the proposed rule; and (3) a comparison of the requirements in the current Part 35, as codified in Title 10, and the draft final rule. The paper also requests Commission approval to complete the final Part 35 rulemaking package using the draft final rule language and to notify specialty boards that we will begin accepting requests for recognition of specialty boards before publication of the final rule.

SUMMARY:

In its Staff Requirements Memorandum (SRM) dated April 23, 1999 ("Staff Requirements: Briefing on Part 35 Rulemaking," Attachment 1) the Commission requested that staff provide it with a paper that includes a draft Federal Register notice, containing the draft final rule text and enough information to allow comparison of the changes from the current rule to the proposed rule and the current rule to the draft final rule. This paper provides the requested information and discusses three key issues for Commission consideration.

CONTACT: Catherine Haney, NMSS/IMNS
(301) 415-6825

Diane S. Flack, NMSS/IMNS
(301) 415-5681

BACKGROUND:

In its SRM-COMSECY-96-057, "Materials/Medical Oversight (DSI 7)," dated March 20, 1997 (Attachment 2), the Commission directed the revision and restructuring of Part 35 into a risk-informed, more performance-based regulation. During the rulemaking process, the staff forwarded several Commission Papers that either provided information on the major issues addressed during the rulemaking or requested direction on specific issues. A chronological list of these papers and associated SRMs is provided in Attachment 3. Any outstanding issues from the SRM are highlighted in Attachment 4.

Members of the public, stakeholders, Agreement States, non-Agreement States, and the Advisory Committee on the Medical Uses of Isotopes (ACMUI) were given the opportunity to provide input and to discuss the proposed requirements for medical use licensees on numerous occasions. Since August 1997, the staff has held seven facilitated public workshops with stakeholders, four of which were held during the public comment period on the proposed rule, and has made formal presentations at approximately 20 professional society meetings. We also discussed key rulemaking issues twice with the full ACMUI and 4 times with the ACMUI subcommittees. We held two workshops with the Agreement States. In addition, we have maintained a continuing dialogue with the professional societies on particular issues of interest to them [e.g., training and experience (T&E) requirements].

We worked closely with the Organization of Agreement States (OAS) and the Conference of Radiation Control Program Directors, Inc. (CRCPD) in preparing the proposed and draft final rule. Two individuals from the OAS and CRCPD participated as members of the Part 35 Working Group and a representative of the OAS participated as a member of the Steering Group. In addition, the Chair and a member of the Part 35 Working Group participated in meetings with the CRCPD SR-6 Committee, which is tasked with revising Part G, "Medical Uses of Radionuclides," of the "Suggested State Regulations". As a result of the interactions with the SR-6 Committee, we have identified four areas in which the Committee plans on recommending that the States adopt more restrictive rule language than that proposed in the draft final rule. These areas are identified in Attachment 5.

In preparing the draft final rule, we also carefully reviewed and considered approximately 225 written letters that were submitted during the public comment period. All written and oral comments on the rule were consolidated in a Lotus notes database that is available to the Commission for review. In addition, all comment letters, transcripts and written summaries of the public meetings, and the April 30, 1999, version of the draft final rule were placed on the NRC rulemaking website for review by stakeholders. All working group meetings were open to the public and stakeholder representatives frequently attended.

On March 25, 1999, representatives of the staff and the ACMUI briefed the Commission on the status of the rulemaking and on the staff's proposed resolution of comments on the major issues. Since the briefing, the staff has considered comments made by the Commission and

ACMUI (Attachment 6, ACMUI minutes), as well as comments from various stakeholders that were submitted after the March meeting, in further developing the draft final rule text and the responses to comments. As a result, we have revised our position on some of the issues that were presented at the March 25 briefing. Subsequent to the briefing, the Commission directed us, in an SRM dated April 23, 1999 (“Staff Requirements: Briefing on Part 35 Rulemaking,” Attachment 1) to provide the Commission with a paper that provides a draft Federal Register notice to include the draft final rule text and enough information to allow comparison of the changes from the current rule to the proposed rule and from the current rule to the draft final rule.

DISCUSSION:

Attachment 7 is a draft Federal Register notice for your review. This document summarizes and responds to comments received on the proposed rule (“Supplementary Information Section III and IV”). This document also provides a narrative comparison between the current rule and the draft final rule text (Section V). Attachment 8 provides a comparison between the current, proposed, and draft final rule. Attachment 9 provides detailed information on the compatibility levels associated with the draft final rule.

A risk-informed approach was used in developing the draft final rule and in responding to comments. The staff used risk insights, together with other factors, to establish requirements that better focus licensee and regulatory attention on design and operational issues commensurate with their importance to health and safety. We considered all comments, provided by the stakeholders and the public, about general risk issues as well as comments about specific sections in the rule. We reviewed several extensive assessments, including the external review and report by the National Academy of Sciences-Institute of Medicine, a 1993 Nuclear Regulatory Commission (NRC) internal senior management review and report, and the Commission’s Strategic Assessment and Rebaselining Project. We also reviewed events in the Nuclear Materials Event Database and technical issues previously addressed in technical assistance requests from the regions.

A risk-informed, performance-based approach was also used in revising the licensing requirements. The draft final rule significantly reduces the amount of information needed from an applicant or licensee to possess and use byproduct material for medical purposes. For example, most applicants will no longer be required to submit detailed procedures in support of their license applications or amendment requests. License or amendment applicants will only be required, if applicable, to submit procedures required by Subpart H, “Photon Emitting Remote Afterloader Units, Teletherapy Units, and Gamma Stereotactic Radiosurgery Units.” This change will significantly reduce unnecessary regulatory burden on both diagnostic nuclear medicine licensees and NRC licensing staff. The Commission should note that we received numerous comments both in support of, and in opposition to, the proposed revision of NRC’s licensing policy (see Attachment 7).

Although Attachment 7 provides proposed resolutions for all medical use issues associated with this rulemaking, there are several issues that we believe should be brought to the Commission’s attention. These issues are discussed below.

1. Training and Experience

T&E was one of the most important issues addressed during the rulemaking because adequately trained personnel are key to the safe use of byproduct material in medicine. In addition, the majority of comments submitted on the rule were in this area. The summary of public comments and our proposed response to those comments, as well as a more detailed discussion of the two T&E issues identified below, are provided in Attachment 7, Section III.

There are two significant differences in the T&E requirements between the proposed rule and the draft final rule. First, we have deleted the proposed requirement for an examination to demonstrate that an individual has sufficient knowledge in radiation safety commensurate with the use requested. Second, we increased the requirements for individuals who want authorized user status for medical uses under 10 CFR 35.200, "Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required," and 10 CFR 35.300, "Use of unsealed byproduct material for which a written directive is required." Both of these changes were made because of the comments received on the proposed rule.

The T&E requirements in the draft final rule also differ from the T&E requirements staff presented at the March 25, 1999, Commission briefing. At that time, the staff proposed specific hourly/yearly requirements for authorized users, authorized nuclear pharmacists, authorized medical physicists, and Radiation Safety Officers. We also proposed that the training be obtained in an NRC-approved program. Since the briefing, we added specific sections for use of strontium-90 for ophthalmic use and sodium iodide I-131 in quantities less than or equal to 1.22 Gigabecquerels (33 microcuries) and in quantities greater than 1.22 Gigabecquerels (33 microcuries). In addition, we are no longer recommending that NRC review and approve training programs. We believe that NRC review of training programs is not warranted because of comments received from the ACMUI during its March meeting and additional comments received from stakeholders after the Commission briefing. The draft final rule provides for NRC recognition of boards whose certification processes includes all of the requirements in a given section containing T&E requirements.

As a result of the changes noted above, we believe the T&E requirements in Subpart J of the proposed rule can be deleted. The T&E requirements in Subpart J were maintained in the proposed rule because alternative routes to the revised T&E criteria were needed until radiation safety examinations and specialty boards were approved by NRC. (More detailed information on deletion of Subpart J can be found in Attachment 7, Section III.) To ensure that the specialty boards have adequate time to request NRC recognition and for NRC to recognize the boards, we request Commission approval to begin the recognition process before the publication of the final rule. Attachment 10 is a sample letter that would be mailed to the specialty boards. This letter explains that the Commission will accept requests for recognition and identifies the information that must be provided to the Commission.

2. Notification Following a Medical Event (10 CFR 35.3045)

We received numerous comments supporting the deletion of the requirement for licensees to notify and provide a written report to the patient or responsible relative after a medical event. These comments are summarized in Attachment 7, Section III, 10 CFR 35.3045.

We have retained the current reporting requirements in the draft final rule, except for a modification to the patient notification requirement. However, as requested in the April 1999 SRM, we have evaluated the pros and cons of revising the Part 35 patient notification provisions to require: (1) verbal notification of the patient (or the individual's responsible relative or guardian) of the medical event, while retaining the current caveat that allows referring physicians to exercise discretion on notifying the patient; (2) written documentation of the medical event in the patient's file; and (3) a written certification by the referring physician to NRC stating that the patient was notified, unless discretion was exercised. The pros and cons of this approach and the requirements in the proposed rule are provided in Attachment 3. The Attachment also contains suggested rule text that could be used to implement the approach outlined in the SRM.

3. Reporting Threshold for Unintended Exposure to an Embryo, Fetus, and Nursing Child (10 CFR 35.3047)

The proposed rule contained a requirement for licensees to notify us of any unintended exposure to an embryo, fetus, or nursing child that exceeds 5 milliSievert (500 millirem). This requirement was included in the rule to satisfy one of the provisions, in the "Abnormal Occurrence Policy," that requires us to report to Congress any unintended radiation exposure, to any minor, resulting in an annual total effective dose equivalent of 50 mSv (5 rem) or more, or to an embryo/fetus, resulting in a dose equivalent of 50 milliSievert (500 millirem) or more. [A more extensive discussion on the need for this rulemaking can be found in the Federal Register notice for the proposed rule (63 FR 43516)]. Note that the requirement is a reporting requirement, not a dose limit.

We received numerous comments from the stakeholders on this reporting requirement. Many commenters believe that this proposed requirement should not be in the final rule because it will interfere with the practice of medicine. However, they indicated that if the requirement were included in the final rule, the reporting level should be raised to 5 rem because: (1) a lower threshold would negatively affect the practice of medicine; and (2) no deterministic effects of radiation exposure are noted in the embryo, fetus and nursing child at 5 milliSievert (500 millirem). A summary of the public comments and our proposed resolution of those comments are provided in Attachment 7, Section III, 10 CFR 35.3047.

At the March 25, 1999, Commission briefing, the staff presented the Commission with two options for addressing this issue. Option 1 was to place the requirement in 10 CFR Part 20, "Standards for Protection Against Radiation." This was our preferred option because the requirement would apply to all uses of byproduct, source, and special nuclear material, rather than just medical users. Option 2 was to place the requirement in Part 35, but raise the reporting threshold to 50 milliSievert (5 rem). After further

review, now we believe it is more appropriate to go forward with Option 2, at this time, because the majority of incidents that would warrant reporting occur during medical use. Subsequently, we would evaluate whether rulemaking is needed to add a similar requirement in 10 CFR Part 20 or Parts 30, 40, and 70.

NOTE:

This paper does not contain an enforcement and inspection plan. Office of Nuclear Material Safety and Safeguards (NMSS) staff, in coordination with Office of Enforcement staff, will develop a plan and will provide it to the Commission when the final rulemaking is submitted for Commission approval.

RESOURCES:

The resources expended on this rulemaking have been somewhat greater than those identified in the fiscal years (FYs) 1998 and 1999 budgets. As a result, resources have been reprogrammed from lower-priority rulemaking activities within NMSS. Rulemaking activities will continue into FY 2000 to accommodate the revised rulemaking schedule. Staff estimates that an additional 3 full-time equivalents will be required in FY 2000 to complete the rulemaking, associated guidance, and related documents. These resources are included in the proposed FY 2000 budget.

RECOMMENDATIONS:

That the Commission

1. Inform the staff, within 2 business days from the date of the Commission Paper, if the paper should not be made available to the public;
2. Approve the staff's draft final rule language and draft responses to comments;
3. Direct staff to submit the revised Medical Policy Statement, if possible, and complete the final Part 35 rulemaking package to the Commission approximately 3 months from the date of the SRM for this paper;
4. Direct staff to prepare a rulemaking plan to revise Part 20 or Parts 30, 40, and 70 to require reporting of unintended exposures to an embryo, fetus, or nursing child;
5. Inform staff if a public briefing on the draft final rule is needed; and
6. Direct the staff to begin the process to recognize specialty boards.

COORDINATION:

The Office of the General Counsel has reviewed this paper and has no legal objection. The Office of the Chief Information Officer has information collection and information management technology implications and has no objections. The Office of the Chief Financial Officer has reviewed this paper for resource implications and has no objection.

William D. Travers
Executive Director
for Operations

Attachments:

1. SRM on March 25, 1999, Briefing, dated April 23, 1999
2. SRM: COMSECY-96-057, dated March 20, 1997
3. "Chronological List of Part 35 Commission Papers and SRMs"
4. "Outstanding Issues from Previously Issued SRMs"
5. "CRCPD SR-6 Committee Concerns"
6. ACMUI Meeting Minutes, March 1999
7. Draft Federal Register Notice
8. Comparison - Current, Proposed, and Draft Rule
9. Detailed Information on Compatibility Designations
10. Sample Letter for Specialty Boards

COORDINATION:

The Office of the General Counsel has reviewed this paper and has no legal objection. The Office of the Chief Information Officer has information collection and information management technology implications and has no objections. The Office of the Chief Financial Officer has reviewed this paper for resource implications and has no objection.

William D. Travers
Executive Director
for Operations

Attachments:

1. SRM on March 25, 1999, Briefing, dated April 23, 1999
2. SRM: COMSECY-96-057, dated March 20, 1997
3. "Chronological List of Part 35 Commission Papers and SRMs"
4. "Outstanding Issues from Previously Issued SRMs"
5. "CRCPD SR-6 Committee Concerns"
6. ACMUI Meeting Minutes, March 1999
7. Draft Federal Register Notice
8. Comparison - Current, Proposed, and Draft Rule
9. Detailed Information on Compatibility Designations
10. Sample Letter for Specialty Boards

RECORD NOTE: A draft copy of the proposed/final rule was sent to OIG for information on _____.

Distribution: (a:\haney\cpfinal3.wpd) EDO WITS No. 199700065/NMSS Ticket No. 199900194

To receive a copy of this document, indicate in the box: "C" = Copy without attachment/enclosure "E" = Copy with attachment/enclosure "N" = No copy

*See previous concurrence

** Verbal and also see 8/2/99 memo

OFFICE	RGB:IMNS	RGB:IMNS	D:IMNS:NMSS	TECH ED	OGC
NAME:	DFlack*	CHaney*	DCool*	EKraus*	STreby*
DATE:	7/ 8 /99	7/ 8 /99	7 /14 /99	7/ 6 /99	7/ 8 /99
OFFICE:	ADM	D:OE	OSP	CFO	OCIO
NAME:	DMeyer*	JLieberman**	PLohaus*	CAbbott*	BStMary-Refer to
DATE:	6/ 22 /99	6/ 11 /99	6/ 24 / 99	6/ 22 /99	6/ 16 /99 email
OFFICE:	D:NMSS	DEDR	EDO		
NAME:	CPaperiello*	FMiraglia, Jr.	WDTravers		
DATE:	7 /14 /99	/ /99	/ /99		

OFFICIAL RECORD COPY

ATTACHMENT 1

April 23, 1999

MEMORANDUM TO: William D. Travers
Executive Director for Operations

FROM: Annette Vietti-Cook, Secretary /s/

SUBJECT: STAFF REQUIREMENTS: BRIEFING ON PART 35 RULEMAKING, 1:30 P.M.
THURSDAY, MARCH 25, 1999, COMMISSIONERS CONFERENCE ROOM, ONE
WHITE FLINT NORTH, ROCKVILLE, MARYLAND (OPEN TO PUBLIC
ATTENDANCE)

The Commission was briefed by the NRC staff and members of the Advisory Committee on Medical Uses of Isotopes on the status of proposed amendments to 10 CFR part 35, "Medical Use of Byproduct Material." The proposed rule, as contained in SECY-98-128 and modified by SRM S98-128, dated July 21, 1998, was issued for public comment from August 13, 1998 to November 12, 1998. SRM S98-263 extended the comment period 30 days.

At the meeting, there were no requirements identified for staff action. However, the Commission indicated that it would provide guidance on the schedule for the completion date of the revised final rule. This staff requirements memorandum (SRM) provides such guidance.

The staff should provide the Commission with a brief paper that provides a draft Federal Register notice (FRN) to include the draft final rule text and those portions of the statements of consideration (SOC) that discuss resolution of public comments and provide enough information to allow comparison of the changes from the current rule to the proposed rule and draft final rule. The staff should offer rule language options in either the draft FRN or as an attachment to the paper on any issues where the staff does not have a preferred approach.

(EDO) (Suspense Date: 07/23/99)

The Commission requests that in addition to the current patient notification language the staff consider and provide the pros and cons in the July paper of revising the Part 35 patient notification provisions to require: 1) verbal notification of the patient (or the individual's responsible relative or guardian) of the medical event while retaining the current caveat regarding referring physician discretion on patient notification; 2) written documentation of the medical event in the patient's file; and 3) a written certification by the referring physician to the NRC that the patient was notified, unless discretion was exercised, as part of the event report submitted to NRC.

The staff should plan to make the July paper publicly available two business days from the date of the paper unless directed otherwise by the Commission. The Commission may choose to conduct a public briefing on the draft final rule after receipt of the paper. An SRM would then be issued to require submittal of the complete final rulemaking package to the Commission approximately three months from the date of the SRM. The package should include the final FRN, the Office of Management and Budget package for clearance, related rule guidance and, if possible, the revised medical policy statement.

The staff is encouraged to keep the Commission informed on these important issues through informal briefings to the Commissioner Assistants previously designated for each office. Mr. John Thoma is designated as the contact for Commissioner Merrifield.

cc: Chairman Jackson

Commissioner Dicus
Commissioner Diaz
Commissioner McGaffigan

OGC

CIO

CFO

OIG

OCA

OPA

Office directors, regions, ACRS, ACNW, ASLBP (by E-Mail)

PDR

DCS

ATTACHMENT 2

March 20, 1997

MEMORANDUM TO: L. Joseph Callan
Executive Director for Operations

FROM: John C. Hoyle, Secretary /s/

SUBJECT: STAFF REQUIREMENTS - COMSECY-96-057 MATERIALS/MEDICAL
OVERSIGHT (DSI 7)

With respect to the overall materials program, the Commission continues to support its preliminary views on this issue which were a combination of two options -- Continue the Ongoing Program with Improvements (Option 2) and Decrease Oversight of Low-Risk Activities with Continued Emphasis in High-Risk Activities (Option 3). For the longer term, the Commission also believes that consideration should be given to broadening NRC's regulatory oversight to include one or more of the higher-risk activities identified in Option 1.

With respect to the medical program, the Commission was not persuaded by the National Academy of Sciences, Institute of Medicine (IOM) report that recommends that NRC should not be the Federal agency involved in the regulation of ionizing radiation in medicine. The Commission continues to believe that the conclusions in the report were not substantiated and that the recommendations should not be pursued.

The Commission continues to support the use of ACMUI and professional medical organizations and societies in developing regulatory guides and standards as was proposed in the Commission's preliminary views. In the longer term, the Commission would be willing to consider taking on broader regulatory responsibilities for higher risk activities involving other sources of ionizing radiation but such efforts should not divert resources from the 10 CFR Part 35 rulemaking discussed below.

In lieu of a rulemaking plan in the context of Management Directive 6.3 the staff should submit a program for Commission approval for revising 10 CFR Part 35, and associated guidance documents, and the Commission's 1979 Medical Policy Statement, if necessary. The program should describe how 10 CFR Part 35 can be restructured into a risk-informed, more performance-based regulation by a suspense date of 6/30/99. In developing the program the staff should consider the following:

- (1) Focusing Part 35 on those procedures that pose the highest risk.
- (2) For diagnostic procedures, staff should consider regulatory oversight alternatives consistent with the lower overall risk of these procedures.
- (3) The staff should address how best to capture not only relevant safety-significant events, but also precursor events.
- (4) Changing the nomenclature from "misadministration" to "medical event" or comparable terminology.

- (5) Part 35 should be redesigned so that it can incorporate necessary regulatory requirements for new treatment modalities in a timely manner.
- (6) The Quality Management Program provisions (10 CFR Part 35.32) should be re-evaluated and revised to focus on those requirements that are essential for patient safety, e.g., confirming patient identity, requiring written prescriptions and verifying dose. To the maximum extent possible, the requirements should be revised to be risk-informed. Given this objective, a mixed approach of performance-based rules and otherwise prescriptive regulations should be pursued.
- (7) The staff should consider the viability of using or referencing available industry guidance and standards within Part 35 and related guidance to the extent that they meet NRC needs.
- (8) The staff should consider a rulemaking process that provides more opportunity for input from potentially affected parties than is provided by the normal notice and comment rulemaking process but would be less consumptive of resources and time than the process recently used in the development of NRC's rule on radiological criteria for license termination.

The staff's program to implement the above should be submitted to the Commission for its consideration no later than June 6, 1997. The program should target June 30, 1999 as the date for completing the rulemaking process. This rulemaking and associated guidance development is a very high priority for the Commission. The Commission is prepared to provide additional resources to the extent necessary to complete the rulemaking process on this schedule.

(EDO - Program)	(SECY Suspense: 6/6/97)
(EDO - Complete Rulemaking)	(SECY Suspense: 6/30/99)

cc: Chairman Jackson
Commissioner Rogers
Commissioner Dicus
Commissioner McGaffigan
Commissioner Diaz
K. Cyr
D. Rathbun
H. Bell
A. Galante
R. Scroggins
W. Beecher

ATTACHMENT 3

CHRONOLOGICAL LIST OF PART 35
COMMISSION PAPERS AND SRMS

April 23, 1999 Staff Requirements: Briefing on Part 35 Rulemaking, March 25, 1999

November 13, 1998 Staff Requirements: SECY-98-263- Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material

November 9, 1998 SECY-98-263: Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material

July 21, 1998 Staff Requirements: SECY-98-128- Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material

July 9, 1998 Staff Requirements: SECY-98-127- Draft Proposed Policy Statement on the Medical Use of Byproduct Material

June 26, 1998 Memo from John C. Hoyle to L. Joseph Callan on SECY-98-054: Commission Resolution of Significant Issues Associated with the Revision of 10 CFR Part 35, Medical Uses of Byproduct Material

June 4, 1998 SECY-98-128: Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material

June 4, 1998 SECY-98-127: Draft Proposed Policy Statement on the Medical Use of Byproduct Material

March 20, 1998 SECY-98-054: Commission Resolution of Significant Issues Associated with the Revision of 10 CFR Part 35, Medical Uses of Byproduct Material

June 30, 1997 Staff Requirements: SECY-97-115- Program for Revision of 10 CFR Part 35, Medical Uses of Byproduct Material and Associated Federal Register Notice

June 20, 1997 SECY-97-131: Supplemental Information on SECY-97-115, Program for Revision of 10 CFR Part 35, Medical Uses of Byproduct Material and Associated Federal Register Notice

June 5, 1997 SECY-97-115: Program for Revision of 10 CFR Part 35, Medical Uses of Byproduct Material and Associated Federal Register Notice

March 20, 1997 Staff Requirements: COMSECY-96-057- Materials/Medical Oversight (DSI 7)

ATTACHMENT 4

OUTSTANDING ISSUES FROM PREVIOUSLY ISSUED SRMS

The staff has received direction on the Part 35 rulemaking in five separate SRMs. The purpose of this attachment is to provide the Commission with sufficient information on any outstanding issues in those SRMs and to close them.

SRM-SECY-98-128 Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material

This SRM provided Commission approval to publish the proposed rulemaking in the Federal Register, subject to the comments provided in the SRM. These comments required staff to provide the Commission with additional information when the draft final rule was forwarded to the Commission or revise the rulemaking package. When completing the proposed rulemaking package, we either revised the Statements of Consideration for the rule to request public comments on specific issues or revised rule text prior to publishing the proposed rule. The following discussion provides information on any open items associated with the SRM.

Radiation Safety Committee

The Commission asked that we solicit specific comment on the proposed changes which impact the Radiation Safety Committee (RSC) and the licensee's effectiveness in carrying out radiation protection programs. A summary of the public comments on this topic can be found in Attachment 7, Supplementary Information Section III, Section 35.24, of this Commission Paper.

Training and Experience

The Commission briefing of the proposed rule included a discussion of the proposed increase in the number of hours of training and experience required for authorized user status for treatment of hyperthyroidism and thyroid carcinoma. The SRM directed staff to provide a more thorough explanation of the basis for raising the number of training hours for each category of use where the hours were increased in the FRN. As a result of public comment, the number of hours of training and experience required for authorized user status for treatment of hyperthyroidism and thyroid carcinoma is status quo with the current rule. However, the final rule contains an increase in the number of training hours for one group of users: authorized users of unsealed byproduct material that requires a written directive (§ 35.300). The basis for raising the hours can be found in Attachment 7, Supplementary Information Section III, of this Commission Paper.

The Commission also asked that we provide more detailed information, in the final rule package, on the FTE effort that would be required to implement the proposed third party examination process for the adequacy of licensee training and experience and that we solicit specific public comments on the need for this additional regulatory layer. Based on public comments, we have not included the requirement for the third party examination in the draft final rule. The basis for this decision can be found in Attachment 7, Supplementary Information Section III, of this Commission Paper.

Calibration of Brachytherapy Sources

The Commission asked that the requirement for calibration of all brachytherapy sources by the licensee before initial use should be revised to allow calibration by the manufacturer traceable to NIST or other recognized bodies. Such calibration would be an acceptable means of meeting the requirement in the proposed rule, unless the staff could provide evidence that there have been problems with the manufacturer's calibration. We were also asked to seek

specific public comments on this area of the proposed rule. We revised the requirement in § 35.432 to address manufacturers' calibrations prior to issuing the proposed rule. A summary of the public comments on this requirement can be found in Attachment 7, Supplementary Information Section III, Section § 35.432, of this Commission Paper.

Recordkeeping Requirements

The Commission asked staff to continue to review the proposed recordkeeping requirements to determine if there are any that could be deleted, and, if so, specifically solicit comment on those requirements. As a result of public comments and revisions to some regulatory requirements, some recordkeeping requirements have been deleted, while others have been revised to eliminate some elements of the recordkeeping requirements. A summary of the public comments on recordkeeping requirements can be found in Attachment 7, Supplementary Information Section III, Subpart L, of this Commission Paper.

SRM-SECY-98-263 Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material

In SECY-98-263 Proposed Rule: Revision of 10 CFR Part 35, Medical Use of Byproduct Material, we asked for Commission direction on whether the public comment period for the proposed rule should be extended and whether a risk assessment should be performed in support of this rulemaking.

In the related SRM, the Commission approved Option 2B in the Paper and requested that the staff provide additional information on the need to perform a risk assessment. Specifically, the Commission directed the staff to extend the comment period for the proposed rule and the final rule due date for 30 days, but not to perform a formal risk assessment. The Commission asked that we include a discussion of the risk assessment as an option and provide the pros and cons for this option when the proposed final rule is submitted to the Commission. The pros and cons for performing a risk assessment are listed below.

Pros and Cons of Whether a Risk Assessment is Needed

Pros:

1. Risk assessment would provide additional information on the basis and need for revising regulations.
2. Responsive to comments made by stakeholders.

Some stakeholders believe that Part 35 should be developed using a “risk-based” approach rather than a “risk-informed” approach. Commenters believe that NRC had not adequately assessed the risk of the various modalities for medical use because a formal risk assessment was not done. Three major comments were made on the need for a risk assessment: (1) NRC had not fully accounted for the low risk in diagnostic nuclear medicine; (2) relative risk must be considered because risk can occur from both action and inaction in medicine; and (3) NRC should perform a risk assessment, with stakeholders involved in the selection of the methodology, and publish the results for comment before it proceeds with any revision of its medical use program.

3. Responsive to written request from Organization of Agreement States to “extend the comment period for the ongoing revision of 10 CFR Part 35 to allow for the development of the risk analysis and the rule accordingly” (November 6, 1998).

Cons:

1. If the decision is made to perform a risk assessment, the publication date for the final rule would be significantly delayed. This would not be beneficial to many licensees since the draft final rule provides regulatory relief to licensees using unsealed byproduct material under §§ 35.100 and 35.200. The rule also codifies requirements for remote afterloaders and gamma stereotactic radiosurgery units.
2. NRC resource intensive. We estimate that a formal risk assessment, with significant stakeholder involvement, could require up to 10 FTE over a 5-year period and several million dollars for an outside contractor to perform the risk assessment. These resources are not included in the present budget submission or in the Operating Plan.
3. Not responsive to comments from Agreement State representatives at the 1998 All Agreement States Workshop on Part 35 and stakeholders.

Conclusion:

The staff believes that no additional risk assessment should be performed.

SRM: Briefing on Part 35

The Commission's SRM of April 23, 1999, requested that, in addition to the current patient notification language, the staff consider, and provide the pros and cons of revising the patient notification provisions to require: (1) verbal notification of the patient (or the individual's responsible relative or guardian) of the medical event while retaining the current caveat regarding referring physician discretion on patient notification; (2) written documentation of the medical event in the patient's file; and (3) written certification by the referring physician to the NRC that the patient was notified, unless discretion was exercised, as part of the event report submitted to NRC.

Discussion

The difference between this alternative and the draft final rule text in § 35.3045, "Report and notification of a medical event" would be deletion of paragraph (f). That paragraph would require a licensee to furnish individuals notified of medical events with a written report consisting of either the report submitted to NRC or a brief description of the event and the consequences as they may affect the patient. Thus, our discussion below sets forth the pros and cons of deleting the requirement for the licensee to provide a written report to the patient.

As part of our consideration of this alternative, we have interpreted the Commission's intent in the third part of the alternative as requiring the licensee, rather than the referring physician, to certify that the patient was notified of the medical event. This interpretation avoids the issue whether such a requirement should be imposed on an individual (the referring physician), who in all likelihood is neither an NRC licensee nor an individual otherwise subject to NRC reporting requirements (e.g., Section 206 of the Energy Reorganization Act of 1974).¹ Therefore, we have provided pros and cons of requiring the licensee, rather than the referring physician, to make the certification to the NRC.

Enclosure 1 contains regulatory text that could be used in the final rule to incorporate the approach described above. The regulatory text is identical to that proposed by staff in the draft final rule text with the exception of paragraph (f). Note, we have used the term "medical record" in the rule rather than the term "patient's file" because this is the term more commonly used in medicine and law to describe the collection of documents or other information pertaining to the physical or mental examination of or treatment administered to anyone. These records include, but are not limited to, patient histories, examination results, test results, records of drugs prescribed, dispensed, or administered, records of surgical or other medical procedures, and reports of consultations and hospitalizations.

¹ In addition, in some medical administrations, there may be no referring physician (i.e., the authorized user physician prescribes the administration of byproduct material and there is no referral from another physician).

Pros

- Is more consistent with medical policy goals of (1) avoiding or minimizing NRC intrusion into the practice of medicine, and (2) recognizing that physicians have the primary responsibility for the protection of their patients and that they will act in the best interest of their patients. (Reference 1)
- Places greater reliance on the physician-patient relationship, in which physicians are to provide patients with complete and accurate information (and in which patients have the right to obtain copies of their medical records). (Reference 2)
- Is consistent with SRM direction to revise Part 35 to be risk-informed, i.e., a requirement to provide the patient with copy of written report to NRC (or other written report) may unduly magnify, in the patient's eyes, an event that could be of minimal safety significance. (Reference 3)
- Is consistent with aspects of another Federal patient notification requirement, specifically, in "The Mammography Quality Standards Reauthorization Act" of 1998 (Pub. L. No. 105-248), under which notification (not specified as oral or written)² of a patient and referring physician may be required for certain events (i.e., when a patient has received mammography from a facility whose quality is found to be "so inconsistent with quality standards as to present a risk to individual or public health"). (Reference 4)
- Is responsive to commenters from medical community who objected to the requirement to provide the patient (or responsible relative) with a written report.

Cons

- Does not ensure that patients will be fully informed of a medical event (unless that information would be harmful to them), and therefore, patients may not receive information, viewed necessary by NRC, to make informed medical care decisions. (Reference 5)
- Is not consistent with other NRC requirements, in 10 CFR 20.2205 and 19.13(d), requiring licensees to provide to an individual a copy of the licensee's written report to NRC or other report of radiation exposures to that individual. (Reference 6)
- May not effectuate the specific Commission determination that patients have a right to fully know when they have been involved in a medical event unless that information would be harmful to them. (Reference 7)
- Is unduly prescriptive (i.e., requirement for written documentation of the medical event in the patient's file) as compared to the current rule (10 CFR 35.33(b) and proposed rule (10 CFR 35.3045(h), which do not dictate where the required record of a medical event be made.
- Does not recognize the right of individuals to know information about themselves which is contained in records both inside and outside the Federal sector, unless they have obtained their medical records. (Reference 8)

²However, a written report of the mammography results must be sent directly to the patient if the patient's physician is not available or if there is no such physician. Whether or not such a physician is available or there is no physician, a summary of the written report of mammography results must be sent directly to the patient. 42 U.S.C. § 263b(f)(1)(G)(ii)(III)-(IV).

- May weaken protection of patients, articulated by the Commission as a basis for the misadministration rule (45 FR at 31702; May 14, 1989), due to the deletion of the requirement that a copy of the written report to NRC, or other written report, be provided to the patient.
- Does not account for inconsistent or incomplete adherence by physicians to the AMA Code of Medical Ethics provision for physicians to provide patients with complete and accurate information. (Staff has previously provided copies of professional medical journal articles on this subject to the Commission.)
- AMA threshold for informing patients of physicians' mistakes ("significant complications" to the patient) may not capture NRC "medical events." (Reference 9)

References

1. "Medical Use of Byproduct Material; Draft Policy Statement," 63 FR 43580, 43584-85 (August 13, 1998); "Regulation of the Medical Uses of Radioisotopes; Statement of General Policy," 42 FR 8242 (February 9, 1979).
2. American Medical Association Council on Ethical and Judicial Affairs, "Code of Medical Ethics," Fundamental Element 1.
3. COMSECY-96-057 "Materials/Medical Oversight (DS1 7), March 20, 1997.
4. 42 U.S.C. § 263b(h)(2)(1998).
5. "Quality Management Program and Misadministrations," 56 FR 34104, 34117 (July 25, 1991).
6. "Basic Quality Assurance Program, Records and Reports of Misadministrations or Events," 55 FR 1439, 1444 (January 16, 1990); "Misadministration Reporting Requirements," 45 FR 31701, 31702 (May 2, 1980).
7. Basic Quality Assurance Program, Records and Reports of Misadministrations or Events, 55 FR 1439, 1444 (January 16, 1990), "Misadministration Reporting Requirements," 45 FR 31701, 31702 (May 2, 1980).
8. 55 FR 1444; 45 FR 31701.
9. "AMA Council on Ethical and Judicial Affairs, "Code of Medical Ethics," Current Opinions with Annotations at xxxix, 120, 125 (§ 8.12), 1996-1997.

Enclosure 1

Draft Alternative Rule Language

Differences between the draft final rule are marked to facilitate review.

§ 35.2045 Records of medical events.

- (a) A licensee shall retain a record of medical events reported in accordance with § 35.3045 for 3 years.
- (b) The record must include--
 - (1) The licensee's name;
 - (2) Names of all the individuals involved;
 - (3) The affected or potentially affected individual's social security number or other identification number of one has been assigned;
 - (4) A brief description of the medical event and why it occurred;
 - (5) The effect, if any, on the individual; and
 - (6) The actions, if any, taken or planned to prevent recurrence.

§ 35.3045 Report and notification of a medical event.

- (a) A licensee shall report any event, except for an event that results from patient intervention, in which the administration of byproduct material or radiation from byproduct material results in --
 - (1) A dose that differs from the prescribed dose by more than 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin; and
 - (i) The total dose delivered differs from the prescribed dose by 20 percent or more;
 - (ii) The total dosage delivered differs from the prescribed dosage by 20 percent or more or falls outside the prescribed dosage range; or
 - (iii) The fractionated dose delivered differs from the prescribed dose, for a single fraction, by 50 percent or more.
 - (2) A dose that exceeds 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin from any of the following --
 - (i) An administration of a wrong radioactive drug containing byproduct material;
 - (ii) An administration of a radioactive drug containing byproduct material by the wrong route of administration;
 - (iii) An administration of a dose or dosage to the wrong individual or human research subject;
 - (iv) An administration of a dose or dosage delivered by the wrong mode of treatment; or
 - (v) A leaking sealed source.
 - (3) A dose to the skin or an organ or tissue other than the treatment site that exceeds by 0.5 Sv (50 rem) to an organ or tissue and 50 percent or more of the dose expected from the administration defined in the written directive (excluding, for permanent implants, seeds that were implanted in the correct site but migrated outside the treatment site).
- (b) A licensee shall report any event resulting from intervention of a patient or human research subject in which the administration of byproduct material or radiation from byproduct material results or will result in an unintended permanent functional damage to an organ or a physiological system, as determined by a physician.
- (c) The licensee shall notify by telephone the NRC Operations Center³ no later than the next calendar day after discovery of the medical event.

³ The commercial telephone number of the NRC Operations Center is (301) 951-0550.

(d) The licensee shall submit a written report to the appropriate NRC Regional Office listed in § 30.6 of this chapter within 15 days after discovery of the medical event.

(1) The written report must include --

(i) The licensee's name;

(ii) The name of the prescribing physician;

(iii) A brief description of the event;

(iv) Why the event occurred;

(v) The effect, if any, on the individual(s) who received the administration;

(vi) What actions, if any, have been taken or are planned to prevent recurrence; and

(vii) Certification that whether the licensee notified the individual (or the individual's responsible relative or guardian), and if not, why not; and

~~(viii) If there was notification, what information was provided.~~

(2) The report may not contain the individual's name or any other information that could lead to identification of the individual.

(e) The licensee shall notify the referring physician and also notify the individual affected by the medical event no later than 24 hours after its discovery, unless the referring physician personally informs the licensee either that he or she will inform the individual or that, based on medical judgement, telling the individual would be harmful. The licensee is not required to notify the individual without first consulting the referring physician. If the referring physician or the affected individual cannot be reached within 24 hours, the licensee shall notify the individual as soon as possible thereafter. The licensee may not delay any appropriate medical care for the individual, including any necessary remedial care as a result of the medical event, because of any delay in notification. To meet the requirements of this section, the notification of the individual receiving the medical event may be made instead to that individual's responsible relative or guardian.

~~(f) If the individual was notified under paragraph (e) of this section, the licensee shall also furnish a written report to the individual within 15 days after discovery of the medical event. A licensee may send either --~~

~~—(1) A copy of the report that was submitted to the NRC; or~~

~~—(2) A brief description of both the event and the consequences as they may affect the individual.~~

(f)(g) Aside from the notification requirement, nothing in this section affects any rights or duties of licensees and physicians in relation to each other, to individuals affected by the medical event, or to that individual's responsible relatives or guardians.

(g)(h) A licensee shall retain a record of a medical event in accordance with § 35.2045.

ATTACHMENT 5

CRCPD SR-6 Committee Concerns

The Conference of Radiation Control Program Directors, Inc., SR-6 Committee plans to recommend to the States that they adopt more restrictive requirements than those in Part 35. The revised Part G, "Medical Uses of Radionuclides," of the Suggested State Regulations (SSR) will contain the rule text required to maintain compatibility with Part 35 but will also contain the more restrictive requirements. Any variances will be denoted by the bracketing of text and explanations in the rationale. In addition, the supporting rationale for the revision will recommend that the States adopt the more restrictive rule text.

Specific areas where the SR-6 Committee believes the rule should be more restrictive are listed below in order of the importance the Committee has placed on them:

Area 1 - Training and experience requirements for use of I-131 in quantities greater than 30 microcuries (§§ 35.392 and 35.394).

The SR-6 Committee intends to recommend that the States not adopt either § 35.392 or § 35.394 (80 hours of training plus casework) (currently codified, with minor differences, as §§ 35.932 and 35.934). The Committee plans to recommend that the States require individuals using iodine in quantities requiring a written directive to meet the training and experience requirements in § 35.390 (700 hours plus casework).

The Committee believes that radioiodine is the most radiotoxic radionuclide currently in use in medicine, and poses the greatest radiation risk to ancillary personnel and the general public. They stated that iodine misadministrations are the most common type of therapy misadministration, and that these misadministrations pose a greater biological radiation risk to the patient (based on the larger number of organs which receive unnecessary high exposures) and the public than a sealed source therapy misadministration does. Because of this, the Committee believes that authorized users of iodine therapy doses should not be held to lesser training and experience requirements than diagnostic users of radioiodine or any other diagnostic doses for which a written directive is required. They also believe their approach addresses the stakeholder comments recommending that the regulating agencies assure users are adequately trained in the medical use of byproduct material commensurate with the risk involved.

Note: Comments on this section of the proposed rule are summarized in the draft Federal Register Notice, Supplementary Information Section, Sections III and IV, General Training and Experience.

Area 2 - Need for a reporting requirement for unintended doses to an embryo/fetus or a nursing child (§ 35.3047).

The SR-6 Committee intends to recommend that the States not adopt the requirements in § 35.3047. Committee members believe that the reporting threshold should be 5 mSv

(500 mrem) as is currently required in § 20.2203(a)(2)(iv). In their opinion, the reporting requirement in § 35.3047 establishes a de facto approval on the part of the NRC to allow doses of 50 mSv (5 rem) to the embryo/fetus and nursing child. They do not believe the 5 mSv (500 mrem) limit in Part 20 will negatively impact the practice of diagnostic nuclear medicine because, based on information provided by radiologists and physicists, the majority of diagnostic nuclear medicine procedures will not result in a dose in excess of 5 mSv (500 mrem) to the embryo/fetus or nursing child. Therefore, the SR-6 Committee plans to allow Section D.1203, Aii(4) of the Suggested State Regulations, which requires the licensee to report exposures to an embryo/fetus or nursing child which exceed 5 mSv (500 mrem), to remain and prevail.

Note: Comments on this section of the proposed rule are summarized in the draft Federal Register Notice, Supplementary Information Section, Section III, § 35.3047.

Area 3 - Criteria for releasing individuals containing unsealed byproduct material or implants containing radioactive material (§ 35.75).

The Sr-6 Committee continues to be concerned by the current patient release rule's waiver of regulatory responsibility on the part of the licensee. They are also concerned with the number of radiation alarms at landfills that have been "tripped" by articles contaminated with radioactive materials after the individual is released from the licensee's facility and the State resources which must be expended in responding to these alarms. As a result, the Committee believes the rule should address the licensee's continued responsibility for the contaminated articles.

As a result, the SR-6 Committee plans to recommend that the States incorporate the following rule text into their regulations in this area. (Marked text in paragraphs (c) and (e) identifies where SSR text in this paragraph differs from the draft final rule text in § 35.75.)

Excerpts from draft SSR Part G, Sec.G.27.

(c) *The authorized user shall sign, and the licensee shall maintain, a record of the basis for authorizing the release of an individual, in accordance with § 35.2075(a).*

(e) *Notwithstanding G.27a, the licensee may be held responsible for the proper disposal of any individual's radioactive waste discovered in a solid waste stream that can be traced to the licensee.*

Note: Comments on this section in the proposed rule are summarized in the draft Federal Register Notice, Supplementary Information Section, Sections III and IV, § 35.75.

Area 4 - Safety precautions associated with brachytherapy treatments (§ 35.415).

The SR-6 Committee plans on allowing a patient receiving implant therapy to be in the same room as a nonradiation therapy patient only if the licensee documents that the public dose limits at a meter are met (Part D of the SSR). Currently, they do not intend to allow the housing of more than one unsealed source therapy patient in the same room (as is approved in the draft § 35.315). They plan to keep the requirement to survey both after implantation of brachytherapy sources and after the administration of unsealed therapy doses.

The SR-6 Committee plans to recommend that the States incorporate the following rule text into their regulations in this area.

Excerpts from draft SSR Part G, Sec. G.45 - Safety Precautions.

- a. For each patient receiving implant therapy, a licensee shall:
 - i. Not place the patient in the same room with a patient who is not receiving radiation therapy unless the licensee can demonstrate compliance with the radiation dose limits for individual members of the public as specified in Part D of these regulations at a distance of 1 meter from the implant;
 - iii. Authorize visits by individuals under 18 years of age only on a case-by-case basis with the approval of the authorized user after consultation with the Radiation Safety Officer;
 - iv. Promptly after implanting the sources, survey the dose rates in contiguous restricted and unrestricted areas with a radiation measurement survey instrument to demonstrate compliance with Part D of these regulations and retain for 3 years a record of each survey that includes the time and date of the survey, a sketch of the area or list of points surveyed, the measured dose rate at several points expressed in microsieverts (mrems) per hour, the instrument used to make the survey, and the initials of the individual who made the survey;

- Note:
- 1. Comments on this section of the proposed rule are summarized in the draft Federal Register Notice, Supplementary Information Section, Section III, § 35.415.
 - 2. The draft final Part 35 allows the licensees to quarter patients or human research subjects who are both undergoing a brachtherapy treatment and who cannot be released in accordance with § 35.75 to be quartered in the same room. It does not contain a dose limit for either patient.
 - 3. The draft final Part 35 does not contain a requirement for the licensee to survey the dose rates in contiguous restricted and unrestricted areas. General requirements for surveys are contained in Part 20.

ATTACHMENT 6

SUMMARY MINUTES
ADVISORY COMMITTEE ON THE MEDICAL USES OF ISOTOPES

MARCH 24-25, 1999

The Advisory Committee on the Medical Uses of Isotopes (ACMUI) held a meeting in Rockville, Maryland on March 24-25, 1999. A briefing book, with background information for the issues under discussion, was provided to the ACMUI members in advance of the meeting, and is available through the Public Document Room.

ACMUI members present at the meeting:

Judith Ann Stitt, M.D., Chair

Manuel Cerqueira, M.D.

Nikita Hobson

Ruth McBurney, M.S., CHP

Dennis P. Swanson, M.S., B.C.N.P

John Graham (3/24 only)

Andrew Kang, M.D.

William B. Nelp, M.D. (3/25 only)

Louis K. Wagner, Ph.D.

Invited guests present at the meeting:

Bruce Bower, M.D., representing endocrinology perspectives

Larry Holder, M.D., representing nuclear medicine perspectives

Richard Vetter, Ph.D., representing Radiation Safety Officer perspectives

Jeffrey F. Williamson, Ph.D., representing medical physicists perspectives

Nuclear Regulatory Commission staff present at the meeting:

Cathy Haney, Section Leader, Rulemaking and Guidance Branch (RGB), IMNS, NMSS and Chair of the Part 35 Working Group (Designated Federal Official for the Committee)

Donald Cool, Ph.D., Director, Division of Industrial and Medical Nuclear Safety, NMSS

Robert Ayres, Ph.D., Materials Safety and Inspection Branch, INMS, NMSS

Part 35 Working Group Members present at the meeting:

Marjorie Rothschild, Office of the General Counsel

Penny Lanzisera, Region 1

Diane Flack, RGB

Sam Jones, RGB

Tony Tse, RGB

Tom Young, Region III

Barry Siegel, M.D., medical consultant to the Part 35 Working Group

OPENING REMARKS

Ms. Cathy Haney officially opened the meeting at 9:05 a.m. with general comments on the meeting agenda and the function of the ACMUI. Ms. Haney stated that she had reviewed the Committee members' financial and employment interests, and had not identified any conflict of interest with items to be considered during the meeting. Ms. Haney

stated that any ACMUI member who becomes aware of a potential conflict of interest during the course of the meeting should inform her or Dr. Judith Stitt. Ms. Haney noted that the meeting was announced in the Federal Register on March 3, 1999.

Donald A. Cool, Ph.D., made opening remarks to the Committee. Dr. Cool discussed the importance of the Part 35 rulemaking to the Commission as an effort to move in the direction of more risk-informed and performance-based regulations. He also expressed thanks to three members of the Committee whose terms are ending later in 1999, Dr. Stitt, Mr. Dennis Swanson, and Dr. William Nelp.

PREVIOUS ACMUI SUBCOMMITTEE RECOMMENDATIONS

Ms. Haney updated the Committee on the activities of the Part 35 Working Group since the ACMUI diagnostic and therapeutic subcommittee meetings in February 1999. She asked the Committee members to review the minutes of the subcommittee meetings to ensure that the minutes adequately captured the recommendations of the subcommittees. She indicated that the Working Group had incorporated most of the recommendations of the subcommittees into the revised draft final rule that was provided to the ACMUI in advance of this meeting (hereafter referred to as "draft final rule"). In a few cases, the Working Group wanted further information or dialogue with the Committee about certain recommendations.

TIMELINE FOR REVISION OF 10 CFR PART 35 AND ASSOCIATED ACTIVITIES

Ms. Haney summarized the Part 35 rulemaking activities leading up to the present meeting. She also explained that following this meeting the NRC staff and the ACMUI would brief the Commission on the draft final rule. Following the briefing, the rule would be revised to incorporate any ACMUI recommendations, unless the Commission had previously stated a position that was not consistent with the ACMUI recommendations, e.g., patient notification. This version of the rule would then be provided to the other NRC offices for review and concurrence. Comments from the other NRC offices would be returned by early May. She indicated that the final rulemaking package was due to the Commission by June 1999. Finally, she indicated that during the summer and fall of 1999, the Part 35 Working Group expects to focus on finalizing the guidance document associated with Part 35 (NUREG 1556) and the Medical Policy Statement.

REVIEW OF KEY ISSUES IN PART 35 WORKING GROUP'S DRAFT FINAL RULE

Ms. Haney explained that the Working Group had identified specific issues where they would like ACMUI review and comment. This review would assist the Working Group in finalizing the draft final rule for Commission consideration & approval and in writing the Statements of Consideration.

Section 35.292, Training for imaging and localization. (Note that this section was changed to 35.290 in version being forwarded to the Commission)

Ms. Haney indicated that the training and experience requirements in the proposed rule were focused on radiation safety. The duration of the training and experience program needed for individuals to become authorized users (AUs) for imaging and localization studies (§ 35.200) was significantly reduced from the current rule (1,200 hours to 120 hours). In addition, requirements for an examination in radiation safety and a preceptor affirmation of competence was added to the training requirements for AUs, as well as for the authorized nuclear pharmacist (ANP), authorized medical physicist (AMP), and Radiation Safety Officer (RSO).

Ms. Haney explained that the examination requirement was not included in the draft final rule, but a provision was added for NRC to review and approve training programs. She also explained that the duration of the training program for use of byproduct material for imaging and localization studies (§ 35.200) was increased from 120 hours to 700 hours. (The training and experience requirements in the proposed rule are presented in Attachment 1. The training and experience requirements in the draft final rule are presented in Attachment 2.)

A physician member of the public expressed concern that the training and experience requirements for this category of AUs had been increased since the last public meeting. Ms. Haney explained that the classroom and laboratory training (80 hours) and supervised practical experience (40 hours) would be considered a component within the 700 hours so the actual increase was 580 hours. Dr. Cerqueira stressed that it was important to ensure that the required hours of training could be done concurrently (i.e., 120 within the 700).

Dr. Siegel suggested that eliminating the words "in basic radionuclide handling techniques" would help to clarify that the requirement for 700 hours includes the 120 hours (80 hours of didactic and 40 hours of handling [80/40 split]). Mr. Swanson suggested that the specificity of 80 hours and 40 hours within the 700 hours was too prescriptive, and recommended instead that the requirement be made more flexible by saying "700 hours of training and experience applicable to the medical use of unsealed byproduct material." Dr. Siegel argued that this would provide insufficient information to training organizations on the breakdown between classroom training and practical work experience. Mr. Swanson then modified his recommendation to specify 700 hours of training "that includes 120 hours of training in the following areas: (1) didactic and (2) supervised practical experience."

Ms. McBurney suggested that specifying an 80/40 split in the 120 hours would allow Agreement States to make a determination on the adequacy of the alternate training program. Mr. Graham suggested that without such specification, the entire 120 hours could be provided in a classroom. Dr. Siegel agreed with Mr. Graham and suggested retaining the 80/40 specification. Dr. Williamson also agreed with the split. Dr. Wagner suggested that the requirement say "at least 80 and at least 40." Mr. Swanson moved to leave the proposed final rule language as written, retaining the 80/40 split, that is to leave (c)(1) and (c)(2) of this section with respect to the required hours, unchanged.

Motion 1.1: Leave § 35.292(c)(1) and (c)(2), with respect to the required hours, unchanged.

Vote: 6 in favor, none opposed.

Mr. Swanson expressed concern about the requirements for “supervised practical experience under an authorized user.” He recommended instead that the reference should be to “experience under the preceptor,” since in some cases, such as a centralized nuclear pharmacy, the work would not be done under the supervision of an authorized user. He also questioned whether supervision was the same as preceptorship. Dr. Siegel suggested that the program director would be providing the supervision, while possibly delegating some of the responsibility for training to other persons.

Dr. Vetter supported the requirement for a preceptor. He was concerned that too narrow a restriction would make it impossible for an individual to go elsewhere to obtain specific elements of training. Dr. Wagner suggested that the requirement might be for supervised practical experience approved by a preceptor. Dr. Cerqueira also expressed concern that a cardiologist might obtain clinical training and experience in an Accreditation Council for Graduate Medical Education (ACGME) program, but go outside for didactic training. In that case, the clinical preceptor would have little authority or supervision over the outside program and two preceptor signatures might be needed. Ms. Haney explained that the preceptor will be required to certify that the individual is competent to function independently as an AU and, therefore, two signatures would not be needed.

Dr. Williamson supported revising the definition of preceptor in §35.2. Dr. Siegel thought that changing § 35.292(c)(2) to specify “under the direction of a preceptor” would be sufficient, because the preceptor would be taking the overall responsibility of ensuring that the individual had mastered the necessary material, and “direction” rather than “supervision” is a looser term. Mr. Graham asked if a medical physicist or pharmacist could serve as a preceptor, and was referred to the requirement that the preceptor must meet either §35.292 or § 35.390, which means he or she must be an AU. Dr. Cerqueira recommended using the term “preceptor authorized user” in §35.292(c)(2) and (c)(3) for clarity.

Mr. Swanson suggested that the term “radioactive drug” in the introductory statement to § 35.292 should be retained. Ms. Haney explained that the term radiopharmaceutical did not include biologics, and therefore “radioactive drug” was more appropriate. However, she noted, the FDA definition of radioactive drug could not be used because Part 35 only addresses use of byproduct material. Mr. Swanson noted that in some contexts, even in Part 35, the term should not be limited to byproduct material. For example, the 700 hours of training and experience received by an individual should not be limited to byproduct material because it could include accelerator-produced material, such as thallium. He also noted that in other contexts, use of the broad term “radioactive drug” may not be appropriate unless it is modified by “containing byproduct material.”

Mr. Graham moved to revise § 35.292(c). He moved that the phrase “in basic radionuclide handling techniques” be deleted and the term “radioactive drugs” be substituted for the term “unsealed byproduct material” in the introductory sentence of (c). He also moved that the word “direction” be substituted for the word “supervision” and the term “authorized user” be replaced with the term “preceptor authorized user” in paragraph (c)(2).

Dr. Vetter was concerned that nuclear pharmacists and RSOs may come to an institution having already had their didactic training, and, therefore, the preceptor could not have “directed” that training. Dr. Wagner noted that it is a question not only of direction, but also of approval of the training. Mr. Graham argued that the preceptor does not have to direct all the training, but only must be satisfied that all the training is acceptable. Dr. Cerqueira was concerned that the certifying preceptor could not always verify competence in any training received outside the program. Therefore, more than one preceptor statement might be necessary for individuals who do not obtain board certification, but instead seek AU status via the alternative pathway. Dr. Siegel noted that even medical boards do not require multiple program directors to attest to satisfactory completion of training. Dr. Wagner noted that the rule

language should be revised to state: “under the direction of a preceptor” rather than “under the direction of the preceptor.” This change would clarify that the preceptor who signs the certification does not need to direct all of the training. Secondly, he recommended replacing the phrase “under the direction of” with the phrase “approved by” to add flexibility.

Drs. Cerqueira and Siegel discussed the likelihood that an individual would plan his or her training in advance and use training programs approved by NRC or whether he or she would receive training in different places at different times. Dr. Siegel suggested an individual using a non-traditional approach could seek advance NRC approval. Dr. Cerqueira questioned if a mechanism existed for such approval. Ms. Haney noted that ACMUI assistance would be requested in approving training programs. She also asked whether NRC needed to approve the 120 hours of classroom and laboratory training and the balance of the 700 hours. Dr. Siegel believed that NRC does need to approve the entire 700, just as it would for an ACGME-approved program.

Mr. Graham argued that an individual should have the flexibility of obtaining training under the direction of multiple preceptor AUs. He moved to amend the pending motion to delete the term “authorized user physician” as preceptor and instead use the term “preceptor authorized user.”

Motion 1.2:

1. Section 35.292(c), introductory sentence - The phrase “in basic radionuclide handling techniques” should be deleted and the term “radioactive drugs” should be substituted for the term “unsealed byproduct material.”
2. Section 35.292(c)(2) - The word “direction” should be substituted for the word “supervision” and the term “authorized user” be replaced with the term “preceptor authorized user.”
3. Section 35.292(c)(3) - The phrase “an authorized user” should be replaced with the phrase “preceptor authorized user” and the word “physician” should be deleted.

Vote: 7 in favor, none opposed.

Section 35.392, Training for use of sodium iodide I-131 for which a written directive is required. (Note that this section was divided into 35.392 and 35.394 in the version being forwarded to the Commission.)

Mr. Swanson expressed concern that an individual who is authorized under § 35.392 for the use of I-131 could prepare I-131 capsules with only 80 hours of training. Dr. Siegel pointed out that the individual could not prepare the capsules because they are neither an ANP nor an AU under § 35.292. He indicated that an individual licensed under § 35.392 can only use a drug that is received from an organization licensed under § 32.72 or prepared by an ANP or an AU who meets the requirements in § 35.292. Dr. Siegel indicated that an AU who only meets the requirements of § 35.392 cannot direct anything be done to a drug received from a § 32.72 supplier. Mr. Swanson then moved that § 35.300(b) be amended to say: “who meets the requirements specified in §§ 35.292, 35.390, or an individual under the supervision”

Ms. McBurney noted that the Conference of Radiation Control Program Director’s, Inc., (CRCPD) working group developing State regulations had expressed concern to her about inconsistency between the training and experience requirements in §§ 35.292 and 35.390. They were concerned that the draft final training requirements for individuals that would like to use I-131 for treatment of hyperthyroidism and thyroid cancer were insufficient. Dr. Bower noted that I-131 dosages are administered orally, and have a low risk profile. It was also noted that the training and experience requirements in the proposed rule would have resulted in increased training requirements for an individual who would like to be authorized to administer only I-131 for hyperthyroidism or thyroid cancer. Finally, it was noted that the endocrinology community did not think the increase was warranted in light of their impeccable safety record under the current 80 hour training requirement.

Motion 2: Modify § 35.300(b) to state: "who meets the requirements specified in §§ 35.292, 35.390, or an individual under the supervision . . . "

Vote: 7 in favor, none opposed.

Intravascular Brachytherapy

Dr. Cerqueira suggested that intravascular brachytherapy should be classified under § 35.1000, "Other medical uses of byproduct material or radiation from byproduct material," as an emerging technology. Ms. Haney noted that intravascular brachytherapy is currently addressed under the requirements for brachytherapy, and that the draft final rule does not explicitly classify intravascular brachytherapy as an emerging technology. She explained that the training and experience requirements for intravascular brachytherapy will be addressed after the completion of the Part 35 rulemaking, and in the meantime will be handled on a case-by-case basis.

Dr. Cerqueira made a motion that intravascular brachytherapy, for the prevention of restenosis in the vascular system, be classified as an emerging technology, pending the results of the ongoing Food and Drug Administration (FDA)/broad trials. Mr. Graham suggested that the implications for intravascular brachytherapy would be better understood if the topic was addressed as part of a broader discussion of Subpart K. Ms. McBurney moved to table the motion.

Motion 3: Table the motion to discuss intravascular brachytherapy for prevention of restenosis in the vascular system until discussion of Subpart K.

Vote: 6 in favor, Dr. Cerqueira opposed, believing it appropriate to discuss at this time.

General Discussion - Training and Experience Requirements - Alternative Pathway Chart (Attachment 2)

The Committee reviewed the chart that summarized the training and experience requirements for the alternative pathway to obtain status as an AU, ANP, AMP, or RSO (Attachment 2).

Ms. McBurney expressed concern over the training requirements for an AU that would like to use sodium iodide I-131 for hyperthyroidism and thyroid carcinoma. Dr. Wagner suggested, and Dr. Siegel agreed, that explicitly stating that the requirements pertained only to oral administration of I-131 would be desirable.

Motion 4: Approve the training and experience requirements - Alternative Pathway (Attachment 2).

Vote: 7 in favor, none opposed.

Section 35.24, Authority and responsibility for the radiation protection program.

Radiation Safety Committee (RSC)

Ms. Flack explained under § 35.24(b) that an RSC is required if the institution has two or more different types of uses of byproduct material under Subparts E, F, and H. The ACMUI therapy subcommittee had suggested inserting "or." Ms. Flack asked whether the different types of uses would have to be in different Subparts, or whether an RSC should be required if there are different types of machines that are included in the same subpart, e.g., is an RSC needed if a licensee only has a remote afterloader and a teletherapy unit.

Mr. Graham explained that an RSC should not be required if there was only one type of use. Ms. Haney noted that “type of use” is used to reference use of byproduct material as specified in §§ 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, and 35.1000. As proposed in the draft final rule, if a licensee only used unsealed byproduct material in quantities that required a written directive (§ 35.300), it would not be required to have an RSC. However, a licensee that used unsealed byproduct material in quantities that require a written directive (§ 35.300) and manual brachytherapy (§ 35.400) would be required to have an RSC.

Dr. Williamson noted that two units under § 35.600 would not be covered by the term “types of use.” Dr. Wagner suggested adding a reference to two or more units under § 35.600, so that even two of the same type of units would require a RSC. Dr. Williamson advocated making the criterion two or more uses under § 35.600. Dr. Stitt suggested that the distinction between “units” and “uses” should be clear.

Mr. Graham moved that § 35.24(b) be amended to read that licensees authorized for two or more different types of uses of byproduct material under Subparts E, F, or H or two or more types of units under Subpart H should be required to have an RSC. Dr. Siegel suggested that uses under Subpart K should also be included. Ms. Haney said that the need for an RSC under Subpart K should be addressed on a case-by-case because it could involve a low risk activity. Ms. McBurney moved to modify the motion to state “two or more uses under E, F, and H.”

Motion 5.1: That § 35.24(b) be amended to state “Subparts E, F, and H or two or more types of units under Subpart H.”

Vote: 7 in favor, none opposed.

Dr. Siegel commented that the RSC’s duties should be clarified, to specify that the RSC is not just responsible for the activities that mandate the RSC, but also for all other activities involving the use of byproduct materials in the institution. For example, he explained, if an RSC is required because both Subpart E and Subpart H activities are permitted by the license, the intent is that Subpart D activities at the institution would also be covered by the RSC. Mr. Graham moved to change § 35.24 to read: “establish a radiation safety committee to oversee all uses of byproduct material permitted by the licensee.”

Motion 5.2: That § 35.24 be amended to “establish a radiation safety committee to oversee all uses of byproduct material permitted by the licensee.”

Vote: 7 in favor, none opposed.

Temporary Radiation Safety Officer (RSO)

Ms. Flack noted that the Working Group is revising § 35.24 to allow multiple temporary RSOs. The Committee made no specific comments on this item.

Section 35.27, Supervision.

Ms. Flack stated that the Working Group considered it important to retain the requirements in § 35.27. Mr. Graham noted that ultimately the chief executive officer is responsible for the actions of everyone who is carrying out any activity that is in any way covered by the hospital’s license. He indicated that the draft final rule is correct - licensees are responsible for the acts and omissions of the supervised individual. The Committee had no other comments on this item.

Section 35.40, Written Directives.

Ms. Flack asked the Committee to review § 35.40(b) that lists the information that must be included in written directives. Dr. Williamson recommended that the rule should be revised to state that a written directive for gamma stereotactic radiosurgery include the treatment site, total dose, and number of gamma stereotactic shots for each anatomically distinct treatment site. Other specifications, such as gamma angles and coordinates, should be placed in the treatment plan description. In addition, the written directive should specify number of target coordinate settings, not the target coordinate settings themselves. Dr. Stitt agreed.

Dr. Siegel asked whether the term “anatomically distinct treatment site” was consistent with the definition of treatment site in § 35.2. Dr. Williamson clarified that “distinct” in treatment planning means that the dose contribution from site one is not considered in planning the dose to site two. Dr. Stitt agreed that the language was consistent and would not be burdensome. Dr. Williamson also stated that, to the best of his knowledge, the gamma stereotactic radiosurgery device is only used for single fraction radiosurgery.

Dr. Siegel asked if § 35.40(b) should specify route of administration. Dr. Vetter pointed out that § 35.392 already specifies oral administration, although for sources of I-131 other than sodium iodide, the route of administration should be specified. Mr. Swanson argued that under (b)(1), the written directive could contain a dosage but not the identity of the radiopharmaceutical. Dr. Siegel explained that a form would generally be used to identify the radiopharmaceutical as I-131 sodium iodide. Mr. Graham was unwilling to rely on a form, and argued in favor of adding the radiopharmaceutical name to the minimum requirements for a written directive. The Committee determined that the rule should be modified to address these concerns. Specifically § 35.40(b)(1) should be revised to state: “for an administration of a radiopharmaceutical: the radiopharmaceutical, dosage, and route of administration.” The Committee also agreed that the balance of the list of items to be included in the written directive was acceptable.

Dr. Williamson noted that for remote afterloading brachytherapy, the written directive must include the radionuclide, treatment site, dose per fraction, number of fractions, and total dose. He did not believe the requirements for a low dose-rate (LDR) remote afterloader were appropriate, since the logistics of loading the sources, and doing the treatment planning before the treatment is complete, are identical to the requirements for manual brachytherapy. He believed the requirements should be modified to group the requirements for manual brachytherapy with the requirements for pulsed dose-rate (PDR), medium dose-rate (MDR), and LDR remote afterloaders. Dr. Stitt agreed.

Section 35.63, Determination of dosages of unsealed byproduct material for medical use.

Ms. Haney indicated that the definition for prescribed dosage, as specified in the proposed rule, did not preclude the AU from prescribing a dosage range. In addition, the requirements for written directive in § 35.40 do not preclude an AU from prescribing a dosage range. She also indicated that § 35.63 provides for a 20% deviation between the prescribed and administered dosage.

Dr. Siegel suggested that the rule allow the AU to prescribe a dosage range. He went on to state that if the dosage is specified as a range, the administered dosage must fall within that range. Whereas if the dosage is specified as a single number, the administered dosage should be within 20% of the prescribed dosage. Dr. Vetter questioned whether such a broad latitude would be wise from a radiation safety perspective. Dr. Siegel noted that a physician has latitude to prescribe a dosage range.

Dr. Wagner questioned whether the requirement that administered dosages be within 20% of the prescribed dosage was limited to administrations that require a written directive. He believed a 20% range was acceptable for diagnostic

dosages but might not be acceptable for therapeutic doses or dosages. Dr. Williamson noted that the practice was not to prescribe dose ranges when using sealed therapy sources. Typically, the radiation oncologist specifies a single dose. Dr. Williamson was concerned that allowing AUs to specify a range would cause an increase in treatment delivery errors. Dr. Stitt agreed that a specific dose was desirable for therapeutic uses, although flexibility for diagnostic uses made sense.

Mr. Swanson moved to amend § 35.63(d) to allow a licensee to administer dosages that do not differ from the prescribed dosage by more than 20% or dosages that fall within the prescribed dosage range, unless otherwise directed by the AU.

The Committee discussed at length whether a dosage range should be allowed for unsealed byproduct material that requires a written directive. Dr. Wagner moved to make the distinction according to whether or not a written directive is required.

Mr. Swanson argued that this approach directly interfered with the practice of medicine and implied a level of accuracy and dosage determinations with unsealed byproduct material that does not exist. Dr. Wagner agreed that the issue is defining the boundary between radiation safety and interference with good medical practice. Dr. Williamson was concerned that by allowing a range for certain modalities, an incentive could be created for practitioners to abandon time-honored practices of prescribing a single number in order to avoid potential problems of regulatory enforcement.

Dr. Wagner sought to clarify the motion. He wanted to specify that ranges should not be allowed for the medical use of byproduct material where a written directive is required. Drs. Wagner and Siegel noted examples of situations in which a 20% range could be too narrow. In some cases, administrations could fall outside this range but still be within acceptable medical limits.

Ms. Haney summarized the motion as follows:

Motion 6: NRC regulations should reflect the following:

1. An authorized user may prescribe a dosage range for material administered under §§ 35.100, 35.200, and 35.500.
2. An authorized user may not prescribe a dosage range for material administered under §§ 35.300, 35.400, and 35.600.
3. Administered activities can deviate from a prescribed dosage by 20%.

The Committee decided not to address the issue of a dose range for material used under §§ 35.400 or 35.600.

Vote: Motion withdrawn.

Motion 7: Amend § 35.63(d) to state that unless otherwise directed by the AU, a licensee shall not use a dosage if (a) the dosage differs from the prescribed dosage by more than 20% or (b) the dosage does not fall within the prescribed dosage range.

Vote: 7 in favor, none opposed.

Section 35.3045, Reports of medical events.

Ms. Flack summarized the requirements for medical event reporting, noting that the Working Group revised the requirement to exclude reporting of events that occur as a result of patient intervention, unless a physician determines that the event resulted in permanent function damage to an organ or a physiological system. Ms. Flack also explained that wrong treatment site now involves a dose to the skin or an organ or tissue, other than the treatment site, that exceeds 50 rem to an organ or tissue or 50% of the dose expected from the administration.

Dr. Williamson suggested that the Working Group completely revise the definition of medical event to say it is “an administration of byproduct material in which a technical error on the part of the care giver or device malfunction results in a dose that . . .” Ms. Haney noted that the phrase “technical error” would need to be defined. Drs. Stitt and Wagner agreed that it would be difficult to define “technical error.” Dr. Siegel noted that even with the proposed definition, patient intervention would need to be defined. Dr. Stitt concluded that the proposed final rule was a definite improvement over the current requirements.

Mr. Swanson requested that the Statements of Consideration contain a clear statement that subcutaneous infiltration of a dose that was put into a vein is not considered wrong route of administration.

Ms. Haney requested comments on the phrase “or is expected to result.” She explained that the Working Group used the abnormal occurrence reporting policy, which used similar language, in developing the rule text. Mr. Graham and Dr. Siegel suggested using the phrase “results or will result” rather than the phrase “or is expected to result.” Dr. Stitt agreed with this approach.

Dr. Wagner noted that the phrase “dose to the skin” could mean dose to a point on the skin or dose to all or a large area of the skin as an organ. He moved that all references to “skin, organ, or tissue” be revised to reference “tissue” only.

Motion 8: In § 35.3045, references to “skin, organ, or tissue” should be revised to reference “tissue” only.

Vote: 7 in favor, none opposed.

The Committee agreed, by consensus, to accept the remainder of the editorial rule changes to this section.

Mr. Swanson moved that the Committee go on record as opposing the patient notification requirements in §§ 35.3045 and 35.3047. Dr. Stitt agreed with the statement that the ACMUI does not support any regulation for required notification of physicians and patients, because the requirement is redundant to existing State laws and medical ethics. Ms. Hobson also noted the possible cruelty in notifying patients unnecessarily and frightening them if negative effects of the event were not anticipated. Ms. Haney noted that the Commission believes the requirement to notify patients recognizes the right of individuals to know information about themselves and provides the opportunity for patients to consult with their personal physicians to make timely decisions about their health care. Drs. Stitt and Siegel noted that providing the patient with a copy of the official letter citing Federal regulations harms the physician-patient relationship. Mr. Graham moved to modify the motion to add that notification is redundant to existing State laws and medical ethics.

Motion 9: ACMUI reaffirms that it does not support any regulation requiring notification of physicians and patients as this is redundant to existing State laws and medical ethics.

Vote: 6 in favor, none opposed, 1 abstention. Ms. McBurney abstained, on the grounds that NRC’s reasons for retention of the requirement may be valid, and because she was not an official Committee member during the initial discussion.

Dr. Siegel proposed adding the words "if any" to paragraphs (d)(1)(v), (vi) and (vii). Dr. Stitt agreed, since in some cases there will be no effect on the patient; no improvements are necessary in the program; and no actions need to be taken. The Committee approved the change.

Section 35.3047, Report of a dose to an embryo/fetus or a nursing child.

Ms. Flack indicated that the Working Group had not made any changes to the draft final rule since the subcommittee meetings. However, it has been considering whether the requirement should be removed from the Part 35 rulemaking and instead become the subject of a Part 20 rulemaking that would apply to all NRC licensees. She also indicated that the majority of the Working Group members believed that the reporting threshold for reports should be 50 mSv (5 rem) rather than 5 mSv (500 mrem). Ms. Flack noted that based on information contained in a recent study (Joy R. Russell, et al., "Radiation absorbed dose to the embryo/fetus from radiopharmaceuticals," *Health Physics Journal*, 73:5 (November 1997) pp. 756-769), a 5 mSv (500 mrem) level could result in a large number of reports being submitted to NRC. Dr. Siegel noted that if the Agreement States adopted the rule, the number of reports would be substantially larger because of gallium and thallium use. Ms. Haney noted that representatives from the CRCPD SR-6 Committee indicated to her that they preferred the 5 mSv (500 mrem) level.

Ms. McBurney moved that the requirement be placed in Part 20.

Motion 10: The reporting requirements for unintended exposures to an embryo/fetus or nursing child that are currently in § 35.3047 should be moved to Part 20.

Vote: 7 in favor, none opposed.

Dr. Siegel stated that the 5 mSv (500 mrem) limit could impact use of byproduct material in diagnostic nuclear medicine. He indicated that a number of nuclear medicine procedures could approach the 5 mSv (500 mrem) limit and would impose either a de facto pregnancy testing requirement or a de facto diversion of women of childbearing age away from needed nuclear medicine procedures. Dr. Wagner stressed the importance of the issue, arguing that the 5 mSv (500 mrem) level would lead to a great amount of patient anxiety and stress. He agreed that the rule would require pregnancy testing for a large number of diagnostic nuclear medicine patients. He supported the 50 mSv (5 rem) reporting level. Mr. Graham moved that the ACMUI endorse the subcommittee recommendation that the reporting threshold be 50 mSv (5 rem).

Motion 11: The reporting threshold in § 35.3047 should be 50 mSv (5 rem).

Vote: 7 in favor, none opposed.

Dr. Wagner noted that the only action a physician can do to protect the nursing child is to properly notify and instruct the mother about the precautions to be taken to minimize exposure to the child. He moved that the notification requirements in § 35.3047, that pertain to a nursing child, be restricted to events in which the mother was not properly instructed in accordance with § 35.75 prior to release from the facility.

Motion 12: The reporting requirements in § 35.3047 should be limited to only those events where the mother was not properly instructed in accordance with § 35.75 prior to release from the facility.

Vote: 6 in favor, none opposed, 1 abstention. Ms. McBurney abstained, on the grounds that NRC's reasons for retention of the requirement may be valid, and because she was not an official Committee member during the initial discussion. (Mr. Graham no longer present.)

Dr. Siegel proposed adding the words "if any" to paragraphs (d)(1)(v) and (vi). The Committee agreed.

Mr. Swanson moved that the ACMUI state that it does not support any regulation requiring notification of physicians, mothers, or pregnant women as this is redundant to existing State laws and medical ethics.

Motion 13: ACMUI does not support any regulation requiring notification of physicians, mothers, or pregnant women as this is redundant to existing State laws and medical ethics.

Vote: 5 in favor, none opposed, 1 abstention. Ms. McBurney abstained, because NRC's reasons for the requirement may be valid. (Mr. Graham no longer present.)

Dr. Siegel suggested and Dr. Wagner moved that the language on unintended permanent functional damage be added.

Motion 14: The following phrase should be included with regards to the embryo/fetal reporting requirement in § 35.3047(a): "Has resulted or will result in an unintended permanent functional damage to an organ or a physiological system, as determined by a physician."

Vote: 6 in favor, none opposed (Mr. Graham no longer present).

Dr. Siegel raised the question of how the phrase "specifically approved in advance by the authorized user" will be interpreted, when a dose is given to a patient based on the belief that she is not pregnant, especially during the period when pregnancy is undetectable. Ms. Haney requested advice on what the Statements of Consideration should say on this subject. Dr. Wagner noted that if the threshold is placed at 500 millirem, a large number of reports would be generated concerning doses to individuals within the first two weeks after conception. Dr. Williamson expressed concern that the medical community will be forced to give pregnancy tests to every woman of childbearing age who receives a diagnostic procedure. Dr. Wagner noted that even pregnancy tests would not be effective early in the pregnancy. The Committee concluded that the Statements of Consideration should acknowledge that there is no need to report an exposure to an embryo/fetus if the exposure occurred while the pregnancy was not medically detectable.

The meeting recessed at 5:33 p.m. on March 24, 1999.

The meeting reconvened at 8:08 a.m. on March 25, 1999.

REVIEW OF ISSUES RAISED BY NRC GENERIC ASSESSMENT PANEL

Dr. Bob Ayres explained that the NRC Generic Assessment Panel had two issues requiring ACMUI advice. First, an Agreement State licensee had used an ultrasound device to place iodine-125 (I-125) seeds in a patient's prostate. The ultrasound device did not have the necessary resolution to image the ends of the implant needle. As a result, 31 seeds were placed in the patient's bladder. Dr. Williamson suggested that the problem was a failure of the device quality assurance program, noting that there are standard tests that are done to ensure contrast resolution. Dr. Holder thought the problem was due to inadequate operator training. Specifically, operators should know what they are looking at and be obligated to not proceed if they cannot see properly. Dr. Siegel asked if this event was considered to be a misadministration or an example of malpractice. If a physician performed the actions, he suggested, it might be malpractice and not require NRC involvement. Dr. Ayres said it was a misadministration, since the intended dose was not given. He asked ACMUI to advise whether NRC should describe this event to its licensees as a generic issue. Ms. McBurney did not believe it was a generic issue since this was the only instance

known to have occurred. Dr. Cerqueira thought that it was a malpractice issue and no NRC action was needed. Dr. Stitt indicated that the staff needed to provide additional, more detailed information to the ACMUI prior to any further action.

Second, an Agreement State licensee was giving written instructions to I-125 prostate implant patients that directed the patients to strain their urine for at least the first few days, after source implantation, to capture any I-125 seeds. The instructions directed the patient to return the seeds to the hospital. The licensee indicated that the purpose of the instructions was to ensure that the patient is receiving the "correct dose." Dr. Ayres indicated that the Generic Assessment Panel was concerned that patients would not handle the seeds properly. He requested ACMUI advice on actions, if any, that should be taken by NRC. Dr. Williamson thought that it could be within the medical purview of the physician to ask that the seeds be counted, but he believed the patient should be instructed to then flush them down the toilet rather than collecting and retaining them or transferring them to someone else. Dr. Nelp thought the instruction given to the patient was a matter of medical practice and ACMUI or NRC involvement was not needed. No further recommendations were made.

CONTINUED REVIEW OF KEY ISSUES IN WORKING GROUP'S DRAFT FINAL PART 35 RULE

Section 35.2, Definitions.

Ms. Rothschild questioned whether the duties of the AU, ANP, AMP, and RSO should be placed in the rule or in the guidance document. The Committee agreed that the list of duties should be placed in the guidance rather than in the rule, provided the list did not place any additional requirements on the individuals.

The Committee agreed with the following actions:

1. The definition of diagnostic clinical procedures manual and reference to it should be deleted from the rule.
2. MDR remote afterloaders should be defined.
3. The definition of prescribed dosage should be revised to state: "the quantity or range of radiopharmaceutical activity, as documented in . . ."
4. Ms. Haney and Drs. Stitt and Williamson should develop a definition of manual brachytherapy.
5. The term "radioactive drug" rather than "radiopharmaceutical" should be used generically in the rule. They noted some places where it may be necessary to use the phrase "radioactive drug containing byproduct material."
6. The definition for a unit dosage should be revised to reference dosages prepared "by or under the supervision" of an ANP or AU. The definition should also clearly state that a unit dosage is a dosage intended for medical use, without subsequent manipulation, in a single patient."

Section 35.12, Application for license, amendment, or renewal.

The Committee was asked whether NRC should only review and approve the procedures required in Subpart H as part of the license or amendment process. Mr. Swanson and Dr. Siegel noted that Subpart K should require a more detailed review of procedures for an emerging technology. Dr. Williamson supported the change, but asked why it was necessary for NRC to review the procedures required by Subpart H. Ms. Haney explained that at present only one procedure is required by Subpart H (full calibrations and spot checks).

Drs. Wagner and Williamson suggested that the review of procedures was unnecessarily burdensome and resulted in a situation where the licensee could only change a protocol by a license amendment. Ms. McBurney noted that Agreement States prefer the review to be done by the license reviewer rather than by an inspector.

Dr. Cerqueira moved that NRC should not require license applicants to provide any procedures, including Subpart H procedures, to NRC for review prior to NRC issuance of the license or amendment and, therefore, should not tie licensees to those procedures via license conditions. Dr. Wagner noted that the real issue was whether licensees must follow the procedures unless they apply for and receive a license amendment that allows use of a revised procedure. He suggested the motion should say that the licensee is free to amend the procedures as needed, without prior approval.

Motion 15: ACMUI believes that NRC should not require license applicants to provide any procedures (including Subpart H procedures) to NRC for review prior to NRC issuance of the license (or amendments) and therefore should not tie licensees to these procedures via license conditions.

Vote: 5 in favor; 1 opposed, 1 abstention. Ms. McBurney opposed because States and many licensees want review of procedures. Dr. Nelp abstained because he disliked how the motion was phrased.

Section 35.60, Possession, use and calibration of instruments to measure the activity of unsealed byproduct materials.

The Committee agreed with the proposed revisions to § 35.60. Mr. Swanson recommended that the identity and serial number of any radionuclide standards used in calibrating the instruments be included in the recordkeeping requirement in § 35.2060.

Section 35.63, Determination of dosages of unsealed byproduct material for medical use.

Dr. Siegel noted that § 35.63 needed to be revised to address the issue of tritiated and carbon-14 compounds used in research, which may be obtained from a manufacturer who is not a § 32.72 licensee, radiopharmacist, or AU, but whose activity cannot be directly measured by a licensee without altering the dosage. Dr. Stitt suggested additional work by the Working Group to address the issue.

Section 35.400, Use of sealed sources for manual brachytherapy.

The Committee agreed with the addition of the phrase “in accordance with an effective investigational device exemption application accepted by the FDA” in § 35.400. The ACMUI believed this phrase addresses use of sources not listed in the Sealed Source and Device Registry.

Section 35.410, Safety Instructions.

The Committee agreed with the revision to § 35.410 that requires an AU to be notified if a patient or human research subject has a medical emergency or dies.

Section 35.432, Calibration measurements of brachytherapy sealed sources.

The Committee agreed with allowing a licensee to accept a manufacturer’s calibration of manual brachytherapy sources.

Subpart H, Use of a Sealed Source in Remote Afterloader Units, Teletherapy Units, and Gamma Stereotactic Radiosurgery Units.

The Committee agreed with limiting Subpart H to photon-emitting devices.

Section 35.600, Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit.

The Committee agreed with the addition of the phrase “in accordance with an effective investigational device exemption application accepted by the FDA” in § 35.600. The ACMUI believed this phrase addresses use of sources not listed in the Sealed Source and Device Registry.

Section 35.615, Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Dr. Williamson noted that the requirements for qualified individuals to be present during initiation and ongoing treatments with MDR and HDR afterloaders should not be identical. He believed that requirements for MDR, LDR, and PDR remote afterloaders should be similar.

The Committee reached agreement on the following items:

1. The draft final language in § 35.615 for “an authorized medical physicist and an authorized user or an individual under the supervision of the authorized user who has been trained to remove the applicators in the event of an emergency, to be immediately available during continuation of all patient treatments,” during initiation and ongoing treatments with an MDR afterloader provides needed flexibility.
2. Only a physician designee is needed at the initiation of a treatment with a PDR remote afterloader.
3. Paragraph (b) should be revised to delete the requirement to “immediately” shield the radioactive source for gamma stereotactic radiosurgery, because time is needed to withdraw the patient.

Section 35.632, Full calibration measurements on teletherapy units.

Section 35.633, Full calibration measurements on remote afterloader units.

Section 35.635, Full calibration measurements on gamma stereotactic radiosurgery units.

The Committee agreed with the requirements for calibration in §§ 35.632, 35.633, and 35.635.

Section 35.642, Periodic spot-checks for teletherapy units.

Section 35.643, Periodic spot-checks for remote afterloader units.

Section 35.644, Periodic spot-checks for low dose-rate remote afterloaders.

Section 35.645, Periodic spot-checks for gamma stereotactic radiosurgery units.

The Committee agreed with the removal of the redundancies in §§ 35.642, 35.643, 35.644, and 35.645. Dr. Williamson recommended changing the check on source transfer tubes from quarterly to annually. He also recommended NRC clarify whether LDR remote afterloader licensees need to possess a calibration system if they rely on the manufacturer’s calibrations.

Section 35.657, Therapy-related radiosurgery units.

The Committee suggested that Dr. Williamson review the requirements in § 35.657 directly with the Working Group.

Preparation of Commission Briefing Materials

The Committee members developed a list of issues that they wished to bring to the attention of the Commission and identified the members of the Committee who would present particular parts of the briefing.

10 CFR Part 32, Specific Domestic Licenses to Manufacture or Transfer Certain Items Containing Byproduct Material.

Mr. Swanson noted that Part 32 contains similar requirements to the current Part 35. In particular, § 32.72, "Manufacture, preparation, or transfer for commercial distribution of radioactive drugs containing byproduct material for medical use under part 35," requires that the licensee have instrumentation to measure the radioactivity of drugs and have procedures for use of the instrumentation. Mr. Swanson made a motion that, following completion of the Part 35 rulemaking, the NRC review § 32.72 and make it consistent with the requirements in the final Part 35.

Motion 16: Subsequent to publication of the final Part 35 rule, NRC should review Part 32 for items of consistency.

Vote: 6 in favor, none opposed.

ACGME and NRC Approval of Training Programs

Dr. Cerqueira indicated that only the nuclear medicine program, as described in the ACGME Directory, meets the requirements to be ACGME accredited. The description of the radiology program is general; the cardiology program does not specify details about the hours of training; and the endocrinology program does not specify requirements.

Dr. Cerqueira was concerned that all cardiology, radiology, and endocrinology programs will have to apply to NRC for approval. He noted that there are over 400 cardiology training programs. He was uncertain how quickly programs could make the necessary changes to the ACGME Directory. Dr. Holder suggested that all of the 400 programs would not involve nuclear cardiology and so only a portion of them would seek approval. Dr. Siegel argued that programs, such as radiology, would be able to make the necessary changes in the program description within the two-year rule implementation period.

Mr. Swanson suggested that the cardiology and endocrinology societies should seek information on what would be involved in obtaining ACGME approval of the nuclear cardiology and nuclear endocrinology programs. Dr. Siegel noted that programs have cross-departmental training, so potentially cardiologists or endocrinologists could obtain the necessary training as part of a radiology or nuclear medicine program, as an elective in their cardiology or endocrinology training. The necessary training could be an elective component of an ACGME-approved training program, rather than an essential component of all programs, but the requirements of the program would need to be specified for that elective.

At 12:17 p.m. Ms. Haney adjourned the meeting.

SUMMARY OF MOTIONS

Motion 1.1: Leave § 35.392(c)(1) and (c)(2), with respect to the required hours, unchanged.

Vote: 7 in favor, none opposed.

Motion 1.2:

1. Section 35.292(c), introductory sentence - The phrase "in basic radionuclide handling techniques" should be deleted and the term "radioactive drugs" should be substituted for the term "unsealed byproduct material."
2. Section 35.292(c)(2) - The word "direction" should be substituted for the word "supervision" and the term "authorized user" be replaced with the term "preceptor authorized user."
3. Section 35.292(c)(3) - The phrase "an authorized user" should be replaced with the phrase "preceptor authorized user" and the word "physician" should be deleted.

Vote: 7 in favor, none opposed.

Motion 2: Modify § 35.300(b) to state "who meets the requirements specified in §§ 35.292, 35.390, or an individual under the supervision . . ."

Vote: 7 in favor, none opposed.

Motion 3: Table the motion to discuss intravascular brachytherapy for prevention of restenosis in the vascular system until discussion of Subpart K.

Vote: 6 in favor, Dr. Cerquiera opposed, believing it appropriate to discuss at this time.

Motion 4: Approve the training and experience requirements - alternative pathway (chart)

Vote: 7 in favor, none opposed.

Motion 5.1: That § 35.24(b) be amended to state "Subparts E, F, and H or two or more types of units under Subpart H."

Vote: 7 in favor, none opposed.

Motion 5.2: That § 35.24 be amended to "establish a radiation safety committee to oversee all uses of byproduct material permitted by the licensee."

Vote: 7 in favor, none opposed.

Motion 6: NRC regulations should reflect the following:

1. An authorized user may prescribe a dosage range for material administered under §§ 35.100, 35.200, and 35.500.
2. An authorized user may not prescribe a dosage range for material administered under §§ 35.300, 35.400, and 35.600.
3. Administered activities can deviate from a prescribed dosage by 20%.

Vote: Motion withdrawn.

Motion 7: Amend § 35.63(d) to state that unless otherwise directed by the AU, a licensee shall not use a dosage if (a) the dosage differs from the prescribed dosage by more than 20% or (b) the dosage does not fall within the prescribed dosage range.

Vote: 7 in favor, none opposed.

Motion 8: In §35.3045, references to "skin, organ, or tissue" should be revised to reference "tissue" only.

Vote: 7 in favor, none opposed.

Motion 9: ACMUI reaffirms that it does not support any regulation requiring notification of physicians and patients as this is redundant to existing State laws and medical ethics.

Vote: 6 in favor, None opposed, 1 abstention. Ms. McBurney abstained, on the grounds that NRC's reasons for retention of the requirement may be valid, and because she was not an official Committee member during the initial discussion.

Motion 11: The reporting threshold in § 35.3047 should be 5 rem.

Vote: 7 in favor, none opposed.

Motion 12: The reporting requirements in § 35.3047 should be limited to only those events where the mother was not properly instructed in accordance with § 35.75 prior to release from the facility.

Vote: 6 in favor, none opposed, 1 abstention. Ms. McBurney abstained, on the grounds that NRC's reasons for retention of the requirement may be valid, and because she was not an official Committee member during the initial discussion. (Mr. Graham no longer present.)

Motion 13: ACMUI does not support any regulation requiring notification of physicians, mothers or pregnant women as this is redundant to existing State laws and medical ethics.

Vote: 5 in favor, none opposed, 1 abstention. Ms. McBurney abstained, because NRC's reasons for the requirement may be valid. (Mr. Graham no longer present.)

Motion 14: The following sentence should be included with regards to the embryo/fetal reporting requirement in § 35.3047(a), "Has resulted or will result in an unintended permanent functional damage to an organ or a physiological system, as determined by a physician."

Vote: 6 in favor, none opposed (Mr. Graham no longer present).

Motion 15: ACMUI believes that NRC should not require license applicants to provide any procedures (including Subpart H procedures) to NRC for review prior to NRC issuance of the license (or amendments) and therefore should not tie licensees to these procedures via license conditions.

Vote: 5 in favor; 1 opposed, 1 abstention. Ms. McBurney opposed because States and many licensees want review of procedures. Dr. Nelp abstained because he disliked how the motion was phrased.

Motion 16: Subsequent to publication of the final Part 35 rule, NRC should review Part 32 for items of consistency.

Vote: 7 in favor, none opposed.

Attachment 1

Proposed Rule - August 1998
 Training and Experience Requirements
 Alternative Requirements to Certification by Board Approved by NRC

	Structured Educational Program		Other
	Didactic (hrs)	Practical (hrs)	
35.100, Unsealed - uptake, dilution, excretion	40	20	Physician, preceptor, exam
35.200, Unsealed - imaging and localization	80	40	Physician, preceptor, exam
35.300, Unsealed - written directive required	80	40	Physician, preceptor, exam, 5 cases
35.400, Manual brachytherapy	200	500	Physician, preceptor, exam, 1 yr ACGME program, 2 yrs clinical experience
35.500, Sealed sources for diagnosis	8		Physician, Dentist, Podiatrist
35.600, Therapeutic medical devices	200	500	Physician, preceptor, exam, 1 yr ACGME program, 2 yrs clinical experience
RSO	200		Preceptor, exam, 1 yr or AU
AMP			Preceptor, exam, MS, 2 yrs
ANP	700		Preceptor, exam

Attachment 2
TRAINING AND EXPERIENCE REQUIREMENTS - Alternative Pathway

	Requirements*
§ 35.290 - Training for uptake, dilution, and excretion studies (Written Directive is not required - § 35.100)	<ul style="list-style-type: none"> - 40 hours classroom and laboratory - 20 hours supervised practical
§ 35.292 - Training for imaging and localization studies (Written Directive is not required - § 35.200)	<ul style="list-style-type: none"> - 80 hours classroom and laboratory - 40 hours supervised practical - 580 hours supervised experience in a clinical environment
§ 35.390 - Training for use of unsealed byproduct material (Written directive is required - § 35.300)	<ul style="list-style-type: none"> - 80 hours classroom and laboratory - 40 hours supervised practical - 580 hours supervised experience in a clinical environment - 3 cases each use category requested
§ 35.392 - Training for use of sodium iodide I-131 for which a written directive is required	<ul style="list-style-type: none"> - 80 hours classroom, laboratory and supervised practical - 3 cases each use category requested
§ 35.490 - Training for use of manual brachytherapy sources (§ 35.400)	<ul style="list-style-type: none"> - 200 hours didactic - 500 hours practical - 3 years ACGME program
§ 35.590- Training for use of sealed sources for diagnosis (§ 35.500)	<ul style="list-style-type: none"> - 8 hours classroom and laboratory
§ 35.690 - Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units (§ 35.600)	<ul style="list-style-type: none"> - 200 hours didactic - 500 hours practical - 3 years ACGME program
§35.50 - Radiation Safety Officer	<ul style="list-style-type: none"> - OPTION 1 <ul style="list-style-type: none"> - 200 hours didactic - 1 year supervised experience - OPTION 2 <ul style="list-style-type: none"> - Authorized user for type of use
§35.51 - Authorized Medical Physicist	<ul style="list-style-type: none"> - MS - 2 years experience
§ 35.55 - Authorized Nuclear Pharmacist	<ul style="list-style-type: none"> - 700 hours structured educational program

* Training must be in NEC or A/S approved program. An AU under §§ 35.290, 35.292, 35.390, 35.392, 35.490, 35.690 must be a physician. An AU under § 35.590 may be a physician, dentist, or podiatrist. An AU, AMP, and ANP must also have a preceptor statement.

ATTACHMENT 7

[7590-01-P]

NUCLEAR REGULATORY COMMISSION

10 CFR Parts 20, 32, and 35

RIN 3150-AF74

Medical Use of Byproduct Material; Final Rule

AGENCY: Nuclear Regulatory Commission.

ACTION: Final rule.

SUMMARY: The Nuclear Regulatory Commission (NRC) is amending its regulations regarding the medical use of byproduct material. This final rule is one component of the Commission's overall program for revising its regulatory framework for medical use. The overall goals of this program are to focus NRC's regulations on those medical procedures that pose the highest risk to workers, patients, and the public, and to structure its regulations to be risk-informed and more performance-based, consistent with the NRC's "Strategic Plan for Fiscal Year 1997-Fiscal Year 2002."

EFFECTIVE DATE: This regulation becomes effective on _____ after publication in the Federal Register].

FOR FURTHER INFORMATION CONTACT: Catherine Haney, Office of Nuclear Material Safety and Safeguards, Nuclear Regulatory Commission, Washington, DC 20555-0001, (301) 415-6825, e-mail CXH@nrc.gov or Diane Flack, Office of Nuclear Material Safety and Safeguards, Nuclear Regulatory Commission, Washington, DC 20555-0001, (301) 415-5681, e-mail DSF1@nrc.gov.

SUPPLEMENTARY INFORMATION:

- I. Background.*
- II. Petition for Rulemaking.*
- III. Summary of Public Comments and Responses to Comments.
- IV. Summary of Specific Comments on Agreement State Compatibility.
- V. Summary of Changes Made Between the Current Part 35 and the Revised Part 35.
- VI. Coordination with the Advisory Committee on Medical Uses of Isotopes.*
- VII. Coordination with NRC Agreement States.*
- VIII. Consistency with Medical Policy Statement.*
- IX. Implementation.*
- X. Agreement State Compatibility.

XI. Final Environmental Assessment: Availability.*

XII. Paperwork Reduction Act Statement.*

XIII. Regulatory Analysis.*

XIV. Regulatory Flexibility Certification.*

XV. Backfit Analysis.*

XVI. Small Business Regulatory Enforcement Fairness Act.*

XVII. Criminal Penalties.*

* To be completed after receipt of SRM on draft final rule.

I. Background

To be completed after receipt of SRM on draft final rule.

II. Petition for Rulemaking

To be completed after receipt of SRM on draft final rule.

III. Summary of Public Comments and Responses to Comments

The following summarizes the written and oral comments that we received on the proposed rule. Part I of this section contains a list of the acronyms used in this Section. Part II discusses general issues that affect the rulemaking. Part III discusses specific comments on the proposed rule.

Part I - Acronyms

The following is a list of the acronyms used in this Section. Acronyms are in alphabetical order.

AAPM	American Association of Physicists in Medicine
ABHP	American Board of Health Physics
ABR	American Board of Radiology
ABMS	American Board of Medical Specialities
ABNM	American Board of Nuclear Medicine
ACGME	Accreditation Council for Graduate Medical Education
ACMP	American College of Medical Physics
ACMUI	Advisory Committee on Medical Uses of Isotopes

ACR American College of Radiology

ALARA As low as is reasonably achievable

AMP Authorized medical physicist

ANP Authorized nuclear pharmacist

ANSI American National Standards Institute, Inc.

AO Abnormal Occurrence

AU Authorized user

FDA Food and Drug Administration

Gy/h Gray per hour

GBq Gigabecquerel

HDR High dose-rate remote afterloader

IDE Investigational Device Exemption

IMPEP Integrated Materials Performance Evaluation Program

IND Investigational New Drug Exemption

INPO Institute for Nuclear Power Operations

IRB Institutional Review Board

JCAHO Joint Commission on the Accreditation of Hospitals Organization

LDR Low dose-rate remote afterloader
MBq Megabecquerel
MDR Medium dose-rate remote afterloader
mSv Millisievert
NAS-IOM National Academy of Science-Institute of Medicine
NCRP National Council on Radiation Protection and Measurements
NIST National Institute of Standards and Technology
PDR Pulsed dose-rate remote afterloader
QMP Quality Management Program
SSDR Sealed Source and Device Registry
Sv Sievert
RDRC Radioactive Drug Research Committee
RSC Radiation Safety Committee
RSO Radiation Safety Officer

Part II - General Issues

A. Risk

Issue 1: What is the difference between a risk-informed and a risk-based approach to rulemaking?

Comment. Commenters asked us to explain the difference between a “risk-based” rule and a “risk-informed” rule.

Response. A “risk-based” approach to regulatory decision-making is one in which a safety decision is solely based on the numerical results of a risk assessment. This places a heavier reliance on risk assessment results than currently may be practicable. A “risk-informed” approach to regulatory decision-making represents a philosophy whereby risk insights are considered together with other factors to establish requirements that better focus licensee and regulatory attention on design and operational issues commensurate with their importance to health and safety.

The Commission does not endorse risk-based regulation. In revising Part 35, the Commission used risk insights from available risk information. The Commission considered information in relation to its completeness and reliability and balanced the insights drawn from this information against other factors such as statutory requirements and public and stakeholder interests in formulating policy.

Issue 2: How was risk used in revising Part 35?

Comments. Commenters indicated that the NRC's approach to the Part 35 rulemaking was flawed because a risk analysis had not been performed before initiating the rulemaking. Some commenters did not believe that the NRC has the expertise to perform or manage a rigorous risk analysis that is needed before publishing the final rule. Other commenters believed the proposed rule did not explain NRC's perception of the regulatory problem and how the rulemaking would solve that problem. Commenters asked that the NRC start over by

- (1) Identifying the problem (perform a formal risk-based analysis);
- (2) Revising the Medical Policy Statement;
- (3) Completing the rulemaking; and
- (4) Developing licensing, inspection, and enforcement policies and procedures to support the rule.

Many of these commenters offered possible ways of evaluating risk and asked that stakeholders be allowed to participate in assessing risk. Some commenters indicated that NRC should establish a risk-benefit "filter" to evaluate this and future rulemakings. They believed this approach would be useful in dealing with emerging technologies. They also believed that if NRC had a structured framework for risk analysis, appropriate regulations could be developed to deal with the real risk to the patient, public, and worker.

Other commenters asked that we consider all types of risk before publishing the final rule, e.g. absolute, relative, comparable, perceived, cost, and "pseudo risks." Commenters discussed these types of risks in the following terms and offered the following comments on each type of risk. While most comments were directed at diagnostic nuclear medicine, many of the statements would also apply to therapeutic uses of byproduct material.

Absolute risks are real health effects (deterministic, stochastic) that include harm to the patient, public, and worker. Commenters indicated that diagnostic nuclear medicine procedures do not present measurable health effects to the patients, workers, and the public.

Relative risks are the risks of diagnostic nuclear medicine relative to other diagnostic medical procedures that are currently unregulated for the end-user. The side-effects from many non-radiological medical procedures involve higher risks of harm to the patients than microcurie and millicurie amounts of byproduct material that are used for tracer and localization and imaging studies, where there is no observable radiological or pharmaceutical effect.

Comparable risks are risks of diagnostic nuclear medicine as compared to other industrial risks (radiological and non-radiological) and other human activities that are acceptable to the general public.

Perceived risks involve the public perception of safe and unsafe uses of radiation that eventually influence the licensee to comply with unnecessary NRC requirements in order to compete in the market place. The commenter noted that most cancer patients are willing to accept higher risks for the benefit of cure. The commenter believed that the large number and prescriptiveness of the current regulations add to the misconception that the public has of radiation. By reducing needless requirements on low risk nuclear medicine, the public perception will adjust accordingly, so that NRC regulatory oversight is less burdensome to licensees.

Cost risks result in overspending on the low risk activities. This economic imbalance creates a higher risk for other areas that do not receive the resources that would otherwise be available.

Pseudo risks are unreal risks in which there is no harm associated with the activity or event, e.g., landfill alarms to short-lived, low-activity radioactive waste from diagnostic nuclear medicine.

Response. In March 1997, the Commission directed the revision and restructuring of Part 35 into a risk-informed and, where appropriate, more performance-based regulation. This direction was part of the Commission's overall decision to decrease oversight of lower-risk activities, such as diagnostic nuclear medicine, while retaining oversight of high-risk activities.

Before initiating the rulemaking, the Commission thoroughly reviewed several extensive assessments, including the external review and related report conducted by the National Academy of Sciences-Institute of Medicine (NAS-IOM), "Radiation in Medicine, A Need for Regulatory Reform;" a 1993 NRC internal senior management review and report, and the Commission's Strategic Assessment and Rebaselining initiative. During the development of the overall revision of Part 35, we considered information on risk provided by members of the public and professional societies, professional medical standards of practice, and event databases maintained by NRC to determine where oversight of lower-risk activities could be decreased and where continuation, or even broadening, of the regulations governing higher-risk activities was needed. In addition, throughout the development of the proposed

rule and associated proposed guidance, public workshops were held and early opportunities for comment from potentially affected parties were provided. These interactions included significant discussions on the risk associated with medical uses of byproduct material.

While we did not perform a formal risk assessment, we believe that we have adequately evaluated and considered the risks associated with use of byproduct material in medicine. We have eliminated requirements in the current Part 35 that are contained elsewhere in the Commission's regulations, such as the radiation protection requirements in Part 20. Part 35 licensees will continue to be required to comply with these requirements, such as the ALARA provisions in Part 20, but we do not believe that there is a need to duplicate the requirements in Part 35 unless there are specific, additional radiation protection requirements that are applicable to medical uses licensees. We have maintained some prescriptive requirements in the rule where we believe they are necessary to ensure adequate protection of the workers, patients, and public. The statements of consideration for the proposed and final rule and the Regulatory Analysis explain why we believe changes needed to be made in the rule.

Issue 3: Is the risk of byproduct material in diagnostic nuclear medicine low?

Comment. Many commenters provided information indicating that risks associated with the use of byproduct material in diagnostic nuclear medicine is low. The commenters provided reasons for the de-regulation of low risk nuclear medicine uses altogether. The commenters indicated that the average patient dose from administration of a single unit dose is comparable to the average annual radiation dose from natural background radiation in the United States. They believed that a zero risk tolerance is extremely impractical and the NRC should not attempt to regulate diagnostic nuclear medicine to

account for errors that are harmless. Commenters indicated that the NRC should not substitute theoretical risk values for lack of measurable risk values, that “real risk” is based on real harms that are measurable, and that there are no measurable risks involved with diagnostic nuclear medicine.

Commenters went on to state that diagnostic nuclear medicine has an outstanding performance history, and that there have been zero consequences to the patients, workers, and the public. Another commenter stated that in over 300 million applications of radiation for diagnostic purposes, there has been only one death, and that occurred over 30 years ago. Commenters believed that by requiring compliance with regulations where there is no clear hazard or detrimental radiation dose, the NRC is subtracting licensee resources away from higher risk activities, e.g. non-radiological risks related to medical practice. This brand of economics for safety programs creates an unjustifiable imbalance of resource allocation for the licensee. They went on to say an additional risk burden is placed on the higher, non-radiological risk activities because there is competition for finite resources that support NRC requirements for low risk nuclear medicine. In this sense, the NRC requirements are overly burdensome for most licensees.

Response. We agree that the risks associated with the use of byproduct material in diagnostic nuclear medicine is low. For this reason, the final rule is much different from the current rule. In consideration of the low radiation risks in the diagnostic area, we have reduced the unnecessary regulatory burden for diagnostic nuclear medicine licensees by either eliminating or decreasing the prescriptiveness of the regulations that apply to them. Instead, we are relying on a performance-based approach emphasizing the training and experience of the authorized user (AU), authorized medical physicist (AMP), authorized nuclear pharmacist (ANP), and the Radiation Safety Officer (RSO).

Issue 4: Can regulation of diagnostic nuclear medicine be limited to Part 20 and training and experience requirements?

Comment. Commenters stated that the appropriate regulation of diagnostic nuclear medicine would involve only the radiation protection requirements in Part 20 and board certification requirements as an indication of medical competence. Another commenter identified the sections of the proposed rule asserted to perform no useful purpose and to have no risk-based justification. These identified provisions were: §§ 35.6, 35.11(c), 35.13(d), 35.24, 35.27, 35.60, 35.61, 35.62, 35.63, 35.69, 35.204, 35.2024, 35.2060, 35.2061, 35.2063, and 35.2204.

Response. The final rule includes requirements that are needed to protect occupationally exposed individuals, patients, and the public. Certain radiation protection-related requirements unique to medical use are needed in Part 35 because of their contribution to risk reduction. For example, the rule retains requirements to calibrate instrumentation used to measure the radioactivity of patient dosages before they are administered (§ 35.60). For the reason cited above and because we believe these requirements are essential to the safe handling of byproduct material, we believe the sections cited by the commenter should not be deleted from the rule. (Note, §§ 35.60 and 35.62 were combined in the final rule.)

B. Licensing

Issue 1: Should diagnostic nuclear medicine programs be given a general license rather than a specific license?

Comments. Many commenters recommended that the NRC issue a general rather than specific license for diagnostic nuclear medicine programs. The NRC's role would be to establish training and experience requirements for physicians, pharmacists, and RSOs. They indicated that the applicant would provide the NRC with their name, location, and contact information and pay a licensing fee to NRC. Commenters emphasized that, after satisfying the minimum training and experience criteria for low risk nuclear medicine programs, the physician should be authorized to receive and use byproduct material with minimal or no regulatory oversight.

Commenters related the use of byproduct material in diagnostic nuclear medicine to medical uses of naturally-occurring or accelerator-produced radioactive material (NARM), e.g. thallium-201, gallium-67, indium-111. Commenters indicated that several states have no current regulatory authority for NARM. In those states, any physician could receive and use NARM for nuclear medicine procedures without either a registration or a license. There were no training and experience criteria or other radiation safety regulations for medical use of NARM - the medical use of NARM was controlled by current standards for medical care. Commenters believed that the unregulated medical use of NARM products justifies a similar lack of regulations for medical use of byproduct materials that are currently regulated by NRC.

Some commenters suggested that one of the state radiation control agencies should be allowed to establish a pilot program for general licensing of their nuclear medicine licensees. After a period of several years, the NRC could evaluate the pilot program. If the program were found to be successful, the NRC could revise its regulations to issue general licenses for diagnostic nuclear medicine facilities.

Some commenters indicated that it should not be necessary to identify a physician for the medical use program because the focus of the revised Part 35 will be on radiation safety rather than on the physician's (AU's) clinical competence. These commenters recommended that the licensing process be simplified to identify the name and contact information for the management representative responsible for radiation safety and to describe any byproduct material that is normally used and that could become hazardous to public health and safety during a catastrophic event, e.g., an earthquake or a serious fire/explosion. This commenter believed that the NRC should authorize the applicant for broad scope use of byproduct material and should not review the licensee's standard operating procedures before the authorization.

Some Agreement State commenters stated that they were opposed to the use of a general license in the medical use area. Commenters believed that in the past, regulatory difficulties were created by general licenses for other non-medical uses, e.g., fixed gauges containing sealed sources. The Agreement State representatives believed that if this concept could not be supported for non-medical uses, then it was doubtful that it should be endorsed for medical uses. Many also believed a Radiation Safety Committee (RSC) should be retained to review all aspects of the radiation safety program before submitting an application to the regulatory agency and that the regulatory agency should continue to review procedures before the license or amendment was issued.

Response. We believe that diagnostic nuclear medicine programs should continue to be specifically licensed rather than generally licensed. A specific license is needed because the potential exists for individuals in the diagnostic nuclear medicine setting to be exposed to radiation levels in excess of the Part 20 dose limits, because of the possession of significant quantities of unsealed material, and because the training and experience of the AMP, ANP,

AU, and RSO are necessary for the safe handling of byproduct material. We have however reduced the amount of documentation that must be submitted by an individual or organization that is applying for a specific license to use byproduct material in diagnostic nuclear medicine. When applying for such a license, the applicant will only need to provide us with information on its facility and the training and experience of the AMP, ANP, AU, or RSO. The applicant will no longer need to provide us with detailed operating and emergency procedures, e.g., dose calibrator calibration procedures, survey meter calibration procedures, or safe handling procedures. In many cases, the final rule gives licensees the flexibility to use procedures that have been developed by nationally recognized organizations or by manufacturer's instructions. The final rule also reduces the unnecessary regulatory burden on diagnostic nuclear medicine licensees by eliminating or reducing the prescriptiveness of the regulations that pertain to diagnostic nuclear medicine.

C. Inspection

Issue 1: Could NRC use an outside accrediting organization for inspections in diagnostic nuclear medicine?

Comment. Some commenters expressed a belief that the inspection program in diagnostic nuclear medicine was not necessary. They believed that the NRC could allow professional accreditation boards and organizations to conduct inspections on behalf of NRC. They state that these organization are already involved with nationwide monitoring of the quality of nuclear medicine services, in a peer review manner that encourages comprehensive improvement of quality and safe use of radioactive materials. They compared this approach to NRC's recognition of the Institute of Nuclear Power Operations (INPO) for the reactor industry. These commenters went on to state that the low risks to the patient, worker, and public from the use of

byproduct material for diagnostic nuclear medicine practices do not warrant the current level of NRC regulatory oversight.

These commenters also provided two examples in which a similar approach has been used in the medical community. One example is where the medical community and the Food and Drug Administration (FDA) worked closely in implementing the “Mammography Quality Standards Reauthorization Act of 1998” (Public Law No. 105-248). The FDA partnered with the American College of Radiology (ACR) to establish the ACR accreditation standards as Federally mandated practice standards for personnel, equipment, quality assurance, and other activities involved in mammography. These national standards have led to broad improvements in mammography nationwide. A second example is where the State bureaus for hospital standards recognize the Joint Commission on the Accreditation of Hospitals Organization (JCAHO) accreditation as evidence that State laws have been met by the certified institutions. This approach allows State governments to focus their resources on those facilities that are not certifiable by the JCAHO. This reduces duplication of inspection effort and provides cost savings to the medical institutions.

The commenters thought that the NRC should delegate the inspection program to an accrediting organization by rulemaking or by administrative action after the NRC has reviewed the accreditation organizations. They also indicated that this rulemaking or administrative action should result in a reduction in NRC fees assessed to licensees that voluntarily submit to the accreditation process.

Commenters indicated that the NRC should review the accrediting program to assure that the content of the current monitoring (accrediting) program was adequate and equivalent to the NRC inspection program. Commenters indicated that the site review teams would identify deficiencies, recommend corrective actions, allow time for implementation of improvements, and offer an appeal process to the licensees. They believed that the NRC should then recognize the accreditation organization monitoring programs as adequate to evaluate radiation safety practices of nuclear medicine licensees.

Along with the final rule, commenters recommended that NRC should post a list of approved accreditation boards and organizations. The licensees could voluntarily select the appropriate organization to evaluate their radiation safety programs. Accredited licensees would not be subject to direct inspection by NRC. Licensees that did not voluntarily select an NRC-approved accreditation organization would be subject to direct inspection by the NRC or an Agreement State. Commenters indicated that the NRC could audit the site review teams and randomly accompany them to observe the appropriateness of the evaluation process.

Commenters cautioned that the accreditation organizations should not become the enforcement arm of the NRC and should not be required to report detailed, confidential findings to NRC. Commenters believed a pass/fail list of licensees that voluntarily submitted to the site review team could be made available to NRC. Alternatively, the NRC could condition the nuclear medicine licenses to require the licensee to notify NRC upon certification, re-certification, or change in certification status (e.g. probation, suspension, termination).

Some commenters did not agree with this approach to inspection. Commenters did not believe there would be a cost savings associated with this approach. They cited increased costs to utilities because of the INPO standards and to medical facilities because the cost of mammography operations were increased by the Mammography Quality Standards Act. These commenters believed that any cost savings associated with JCAHO certification were offset by increased fees from other organizations.

Commenters that did not favor this approach indicated that site review team members would not have the authority of the Federal government behind them as NRC inspectors do now. Some indicated that the alternative proposed was self-serving and did not account for independent clinics and institutions. These commenters indicated that NRC's endorsement of accreditation process will set up an unfair advantage and will be used only to increase membership in accrediting organizations.

Representatives from some Agreement State did not think it was likely that Agreement States would relinquish their inspection programs to accrediting organizations.

Response. The NRC's inspection program is separate from this rulemaking and may be changed without changing the regulations. We agree that diagnostic nuclear medicine licensees, as a whole, have operated safely in the past and that the radiation risk to the public, patients, and workers is low. The inspection and enforcement history indicates cooperation and successful implementation of radiation protection programs by most of the licensees.

NRC licensees are encouraged to audit their own activities and discover and correct their own violations. A voluntary program of inspection by an accrediting organization is one method to accomplish this goal. For example, if accrediting organizations were noted to be successful in discovering violations and assuring that those violations are corrected, the frequency of inspections at accredited facilities could be decreased. Under this scenario, some NRC inspections could still be performed to verify the effectiveness of the voluntary program undertaken by the accrediting organization, but the overall number of inspections performed by the NRC would be reduced.

In summary, we believe the proposal for involvement of professional accreditation boards and organizations in the inspection program has merit and should be further developed in an ongoing dialogue. In the interim, the NRC will continue to inspect nuclear medicine licensees but will also continue to make improvements to the program, especially in the area of focusing the inspection program on risk.

Issue 2: What changes should be made in the inspection process as a result of the revised Part 35?

Comment. Commenters expressed a concern that NRC inspections were too detailed and focused on records and use of checklists. Some commenters asked that NRC inspectors focus on radiation safety program management. They indicated that, if the program was managed properly, there would be no need to evaluate program records or the written procedures. Commenters believed the inspectors should be satisfied if the big picture does not indicate a violation because the final rule will be less prescriptive, risk-informed, and performance-based. Other commenters asked that inspectors rely on conversations with licensee staff, and independent measurements to form a basis for inspection findings.

Commenters asked that the NRC provide training on the new rule to inspectors before the final rule is published. They also asked that the period between inspections be increased. Commenters believe that the inspector should be able to recognize the differences between the current and final rule. Agreement State representatives also believe that there will be a critical need to provide training on the final rule to their inspectors. Some commenters also asked that inspectors be encouraged to describe the good practices. They believed this would foster a more positive relationship among NRC, workers, management, and the public.

Response. In recent years, we have changed the focus of our medical inspections from a detail oriented inspection (check-list) to a more performance-oriented inspection. Under this approach, inspectors are directed to focus more on observations, interviews, and measurements than on record reviews to assess program adequacy. We have also revised our process for documenting inspection results. Before 1998, routine inspections were documented using a checklist format. In 1998 and 1999, we revised our procedures to allow findings to be documented in narrative form. This revision was designed to give the inspectors more flexibility and to promote a more performance-based inspection process.

In recent years, we have also revised our inspection policy to focus on risk. The inspection policy now requires inspectors to extend the time between inspections for good performers, those licensees that have relatively few violations for several inspections in succession and no escalated enforcement actions. The time between inspections is also based on the radiation risks associated with the use of the byproduct material. For example, a licensee using byproduct material for imaging and localization studies in a hospital setting is scheduled to be inspected every three years. If this licensee is inspected and demonstrates good performance, the next inspection will be scheduled to be conducted after 5 years, rather than 3 years. A licensee using

a high dose rate remote afterloader (HDR) will be inspected every year. If this licensee is inspected and demonstrates good performance, the next inspection will be scheduled to be conducted after 2 years, rather than 1 year.

The NRC plans to conduct [**text may need to be revised depending on status of Commission direction on the pilot**] a pilot program for licensees authorized to use unsealed byproduct material under §§ 35.100, 35.200, and 35.300. This one-year program is intended to streamline the inspection process and focus inspections on radiation safety performance and risk-based outcomes. The intent of the pilot program is to demonstrate that the streamlined approach can --

- (1) Maintain, and potentially enhance, safety;
- (2) Reduce unnecessary burdens on the licensee;
- (3) Increase NRC efficiency and effectiveness; and
- (4) Increase public confidence, by explicitly addressing risk-informed outcomes. If successful, the program will be extended to other NRC material licensee inspection programs.

Under this pilot program, inspectors will shift primary focus away from detailed examination of the licensee's processes, policies, and procedures to an evaluation of the adequacy of outcomes for six radiation safety based and outcome oriented performance indicators. These outcomes are:

- (1) Adequate program surveillance and corrective actions;
- (2) Knowledgeable staff and management;

(3) Occupational and public doses within regulatory limits;

(4) Adequate security and control of licensed material;

(5) Use of licensed material only as authorized; and

(6) Radiopharmaceutical administrations conforming to the physicians written directives.

The extent and depth of the inspection will be guided by the outcomes for the performance indicators and the potential risk associated with licensed activities. If the desired outcomes are not achieved by the licensee, then a detailed evaluation will follow. It will identify root causes and contributing factors for the licensee's apparent failure to conduct a satisfactory radiation protection program. The detailed evaluation will be similar to the approach that has been used during routine NRC inspections in the past, e.g. review of processes, policies, procedures, and observations, and interviews of additional licensee staff members.

We will continue to qualify inspectors using NRC Inspection Manual Chapter 1246, "Formal Qualification Programs in the Nuclear Material Safety and Safeguards Program Area." During the inspector qualification program, the candidate completes self-study exams for the various parts of 10 CFR Chapter I and obtains classroom and practical laboratory experience for each type of medical use. The candidate accompanies other qualified inspectors and the inspection supervisor during inspections of various types of licenses for medical use programs to develop inspection skills necessary to evaluate radiation safety programs independently and to relate inspection findings to the NRC enforcement policy. Finally, individuals must pass an oral qualification board before they become certified to conduct inspections without direct supervision.

The Agreement States also have formal training programs for their inspectors. Agreement State inspector qualification are reviewed during NRC's periodic review of the Agreement State program.

NRC inspectors also participate in ongoing refresher training. This training includes new innovations in the health physics field as well as training in new initiatives underway at the NRC. Individuals performing medical inspections will receive training in the final Part 35 as well as in any guidance documents associated with the rulemaking. Training will focus on the concepts associated with a risk-informed, more performance-based rule. In addition, inspectors will receive training on the pilot program for streamlining inspections before the pilot program is introduced.

Issue 3: Will the Agreement State inspection program change as a result of changes in the NRC program?

Comment. Several commenters stated that Agreement States may experience problems with their inspection programs if they follow NRC's lead in moving from a prescriptive to a more performance-based approach to inspecting. Other commenters stated that, if the NRC adopted an approach in which inspections would be deferred or eliminated, States may not be able to or choose not to follow NRC's example.

Response. Moving from prescriptive to more performance-based inspections will require a period of adjustment for both the NRC and Agreement States as well as for the licensees. NRC and the Agreement States will address any needed adjustments via their internal training programs. In addition,

Agreement States will be provided with copies of guidance documents currently under development by the NRC. Finally, Agreement States are afforded the flexibility to inspect more frequently based on local concerns.

Issue 4: What changes will be made in the enforcement program as a result of the revised Part 35.

Comment. A commenter agreed with the principal of a performance-based regulation, but questioned whether there would be any changes in the enforcement program.

Response. The NRC's enforcement program is separate from this rulemaking and may be changed without changing the regulations. However, as a result of some changes in the rule, we will need to revise to our enforcement policy. In particular, minor changes will be made to the NRC Enforcement Policy (NUREG 1600, Rev. 1, "General Statement of Policy and Procedures for NRC Enforcement Actions") to incorporate the new terms that are used in the final rule.

In a broader effort, the NRC is revising its enforcement policy to make that program more risk-informed and performance-based. Guidance has been issued on non-escalated enforcement actions (EGM 98-007) to the NRC staff in the materials enforcement area to assure that:

(1) Non-cited violations are used for non-repetitive, non-willful Severity Level IV violations;

(2) The use of enforcement discretion not to issue a citation is considered where warranted for Severity Level IV violations in accordance with Sections VII.B.2 through VII.B.6 of the Enforcement Policy;

(3) Responses are not required for cited Severity Level IV violations if the licensee's corrective actions are already available in a docketed report or other correspondence;

(4) RSC meeting minutes and other licensee program audit records are not used to identify violations that the licensee is already aware of unless the corrective action for the violation is not prompt or comprehensive; and

(5) Multiple examples of the same violation are grouped into a single citation when appropriate.

D. Industry Standards

Issue 1: Can standards of practice be used as an alternative to regulation?

Comment. Some commenters asked whether we would consider replacement of regulations with standards of practice or industry standards that are well understood by medical professionals. For instance, one commenter points out that the American Association of Physicists in Medicine (AAPM) has recently published several excellent reports that relate to radiation safety, including the reports of Task Groups 59, 56, and 40.

Some commenters believed that we could allow a licensee to commit to follow an established standard of practice and thereby limit our regulatory oversight. Commenters also pointed out that many current regulations have become the standard of care and, in instrumentation cases, the manufacturer's guidance. Conversely, some commenters believed that we, as regulators, had the role of defining the minimum level of practice necessary to directly enhance safety. The commenters indicated that there are some limited cases where those practicing are not following "voluntary" standards of practice; therefore regulations were needed. Finally, some commenters questioned our role in regulating an activity that is also regulated by another government agency or by the state.

Response. In developing the final rule in therapeutic uses of sealed sources, we consulted several AAPM reports, including the reports from Task Groups 40, 56, and 59, and Report No. 54. In developing several other sections of the rule, we also consulted various other nationally recognized bodies' reports including American National Standards Institute, Inc. (ANSI), ACR, American College of Medical Physics (ACMP). We understand that these and other standards of practice are often voluntary, and as such, medical professionals are not required to follow them. Therefore, we limited the requirements to the performance standard to be achieved and allowed the licensee to select among the various performance standards to meet the objective of the regulation. We believe that this provides the licensee significant flexibility in designing its radiation protection program.

We agree that, in some cases, the licensed community must comply with several different Federal and state regulations for a single type of use. For instance, in the case of sealed radioactive sources for therapeutic medical uses, the licensed community must comply with FDA regulations for

devices and must also comply with NRC regulations on the use of the radioactivity in or on humans. Whenever possible, we reviewed the various state and Federal regulations, including other NRC regulations, to limit duplication of requirements.

E. General training and experience

Issue 1: Why are there two sets of training and experience requirements in the final Part 35?

Comment: One commenter noted that much of Subpart J is redundant with, but not identical to the training and experience requirements listed in the individual sections of the other subparts. The training and experience requirements should be identical if they are included in two subparts within the same part, or they should only be listed once in the part.

Response: We have deleted Subpart J, so there is only one set of training and experience requirements in the final rule. All medical use licensees and 10 CFR Part 32.72 licensees will have to comply with the new training and experience requirements for AMPs, ANPs, AUs, and RSOs in Subparts B and D through H when the rule becomes effective on [insert date 6 months from publication of the final Rule]. Individuals who have status as an AMP, teletherapy physicist, ANP, AU, and RSO at the time the rule becomes effective will be “grandfathered” under § 35.57, and will not have to satisfy the new training and experience requirements.

The training and experience requirements in Subparts B and D through H of the rule provide alternative pathways for individuals who are not board certified, i.e. the rule specifies the total number of hours of training and experience needed to become an AMP, ANP, AU, or RSO.. This was done because we do not believe individuals must be board certified however, they must have adequate training to handle byproduct material safely. The primary difference between the “board certification route” and the “alternative pathways” concerns the regulatory process used for being approved as an AMP, ANP, or AU. For example, if an individual is certified by a board recognized by NRC, a licensee does not need to amend its license before it allows that individual to work as an AU, ANP, or AMP (reference § 35.14(a) and § 35.24(a)). If however, the individual is not board certified, the licensee must apply for and receive an amendment from NRC before it allows that individual to begin work (§ 35.13(b)). In the case of an RSO, a licensee must always amend its license before it allows an individual to work as an RSO unless the individual would be considered a temporary RSO under 35.24(c)).

Issue 2: Would it be best for regulations to be developed, administered, and monitored by medical speciality organizations?

Comment. A commenter believed that the training and experience requirements would be best developed, administered, and monitored by medical speciality organizations with expertise in clinical applications of radiation-related technologies. The commenter cited the Mammography Quality Standards Reauthorization Act as an example of a cooperative public/private partnership that uses the strengths of both established accreditation/certification programs and Federal government enforcement authority.

Response. We acknowledge and value the expertise of medical speciality boards involved in radiation related technologies. We have met with many of these boards and received valuable information that was used to develop the final rule. However, we believe that the administration of this rule is best performed by the NRC.

Issue 3: Should speciality boards be listed by name in the regulations?

Comment. We received several comments regarding board certification. Some commenters recommended that the regulations list the boards, by name, because the boards rarely change. Another set of commenters stated that the Cardiology board should be listed by name in the rule. Other commenters expressed concern that NRC would recognize boards that were not recognized by the American Board of Medical Specialities (ABMS).

Response. We believe that any reference, by name, to boards should be deleted from the regulation because a rulemaking is needed to add new boards or to delete existing boards. This has been a problem with the current Part 35 because individuals requesting AU status have been certified by a board that is not listed in the regulations on several occasions. In these cases, the NRC evaluated the training of these individuals, in consultation with the Advisory Committee on the Medical Uses of Isotopes (ACMUI), on a case-by-case basis. In the future, without need for a rulemaking, NRC could recognize boards in a more timely manner. (Note: We have provisions in §§ 35.50, 35.51, 35.55, 35.190, 35.290, 35.390, 35.392, 35.394, 35.490, 35.491, 35.590, and 35.690 that allow individuals, who are certified by NRC-recognized boards, to function as an ANP, ANP, AU, or RSO.)

Under the final rule, the boards must be recognized by the NRC or an Agreement State. The NRC will recognize a board if its certification process requires or will require an individual to meet all of the applicable requirements listed in the alternative pathway of the training and experience requirements in Subparts B and D through H. For example, the individual must complete the required number of hours of training and experience that covers specific topics; obtain a signed preceptor certification; and complete specifically identified patient casework, if required.

We do not believe that the NRC's recognition of boards should be limited to those boards that are recognized by the ABMS. Our recognition is contingent on whether the certification process includes all the requirements listed in the alternative pathway. Before we recognize a board, we will review the board's submittal with ACMUI. We will maintain a list of recognized boards on our website.

Boards that are listed in current Part 35, as well as any other boards that are not listed in the current rule such as the cardiology boards will need to apply for recognition. We believe it is necessary to obtain a commitment from the boards that their certification meets the criteria in the alternative pathways because it has been several years since NRC reviewed many of them.

Issue 4: Should the board certification process be "approved" or "recognized" by the NRC?

Comment. Commenters questioned the phrase "whose certification process has been approved by the Commission" because the board will continue to exist regardless of whether the Commission approves the board for Commission purposes.

Response. Based on this comment, we changed all training and experience requirements to state that the medical specialty board's certification process must be "recognized" by the Commission.

Issue 5: What is the preceptor's role?

Comment. A commenter stated the proposed regulations place an inappropriate burden on the preceptor to provide written certification that the applicant has satisfactorily completed the didactic instruction in a structured educational program, obtained the required hours of supervised practical experience, and achieved a level of competency to independently function as an AU. The commenter recommended that all didactic training be certified or approved by an independent organization not associated with any society, board, or medical speciality. The commenter stated that the preceptor should not make any judgement regarding competency and should simply attest that an individual completed the training program.

Response. The regulations in the final rule do place a high degree of responsibility on the preceptor. Because the preceptor must be an AMP, ANP, AU, or RSO, we believe that the preceptor is in the best position to certify that the individual has achieved a level of competency sufficient to function independently as an AMP, ANP, AU, or RSO. We do not believe this places an undue burden on a preceptor, rather it demonstrates a high degree of confidence in the preceptor. Further, we believe that such judgements of competency in training and experience are consistent with the duties of individuals who direct training programs or provide training.

Issue 6: What are the training and experience requirements for physicians who perform research on human subjects?

Comment. A commenter asked what the training and experience requirements are for physicians who perform research on human subjects.

Response. There is no difference between the training and experience requirements for the administration of byproduct material or radiation from byproduct material to a human research subject and the training and experience requirements for an administration to a patient. For example, if the research involves using unsealed byproduct material for imaging and localization studies for which a written directive is required, the physician performing the research must meet the qualifications in § 35.390. If the research involves use of sealed byproduct material in a remote afterloader, the physician must meet the qualifications in § 35.690.

Issue 7: Should the training and experience requirements include an examination?

Comment. We received comments both opposed to and in support of a requirement for individual who would like to become an AMP, ANP, AU, or RSO to pass an examination that would assess whether they had sufficient radiation safety knowledge.

Some commenters supported the exam concept. One thought that it would provide an alternative to a requirement for a long training program. Those commenters who supported the examination believe that an examination is an important tool that should be used to assure that individuals have the necessary skill to handle byproduct material safely. Other commenters believed that the examination would be warranted if an individual had not taken an examination as part of a board certification.

Several commenters stressed the practical problems of implementing the requirements for an examination. They noted that establishing an examination program was extremely time-intensive and expensive. According to several commenters, maintaining the confidentiality of questions was a concern. Some commenters said that the examination requirement was unnecessary and should be deleted unless the NRC had information that significant numbers of AMPs, ANPs, AUs, and RSOs were being inadequately trained.

Other commenters indicated that many training organizations already use testing as part of their educational programs. Therefore, the testing requirement would only increase training costs without adding benefit or value.

Some commenters argued that the NRC should not give the exam itself, nor should the NRC determine the passing score. Other commenters suggested that examining organizations submit questions to the NRC and that the NRC should develop the exam. Some commenters recommended that the NRC collaborate with one or more boards to develop the radiation safety exam. Others suggested that several boards collaborate to develop a

radiation safety examination independent of the NRC. Commenters also recommended that the NRC contract either directly or indirectly with a testing service to administer the exam.

Several commenters stated that the proposed requirement, in Appendix A, for examining organizations to ensure that examinations are not given to individuals who have also been instructed by the examining organization, was too prescriptive. One commenter explained that professional organizations must be trusted to both offer instruction and testing while another commenter encouraged the NRC to keep the two functions separate.

Response. We believe that the training and experience requirements in the final rule for AMPs, ANPs, AUs, and RSOs are sufficient to assure that the radiation safety of the public, patients, and human research subjects is maintained. Therefore, we deleted the requirement for an examination from all the training and experience sections. Instead of an examination, we will rely on the preceptor's certification that an individual has completed the required training and experience and has achieved a level of competency sufficient to function independently as an AMP, ANP, AU, or RSO.

Issue 8: Should Part 35 contain training and experience requirements for technologists?

Comment. Many commenters suggested that minimum training and experience requirements be established for nuclear medicine technologists. In addition, they suggested that technologists be required to pass an exam. Commenters stated that there is a need for training and experience requirements for those individuals who actually handle radioactive materials.

One commenter felt that healthcare agencies should mandate licensure requirements for technologists rather than the NRC. Commenters opposed NRC requiring specific training and experience for nuclear medicine technologists, but supported mandated licensure requirements by healthcare agencies.

Response. We recognize that technologists have an important and substantial role in the medical use of byproduct materials. However, the licensee is responsible for ensuring that the training and experience of individuals working under the supervision of an AU or ANP is adequate. We will continue to rely on the regulations in § 35.27, Supervision, to assure individuals working under the supervision of an ANP or AU are provided adequate training. Therefore, we have not established training and experience requirements for technologists or other individuals using byproduct material under the supervision of an ANP or AU.

Issue 9: Will the training and experience requirements for physicians affect training requirements for technologists?

Comment. Commenters were concerned that the reduction in the duration of some of the physicians' training programs would negatively affect the amount of training that licensees expect technologists to have completed. They were concerned that, if NRC reduced the training requirements for AUs that licensees might reduce their training requirements for technologists. The commenters believed that as the technology becomes more sophisticated, a reduction in training could lead to poor quality studies and result in unnecessary radiation exposure to patients.

Response. We believe that AUs will have sufficient training and experience to assure that byproduct material is handled safely. In addition, an AU is required to be a physician, dentist, or podiatrist. It is the licensees' responsibility to determine the level of training and experience, in addition to the instruction required in § 35.27, needed for individuals working under the supervision of an AU.

1. Unsealed Byproduct Material

For the most part, comments received on the following sections related to more than one section. Therefore, we have chosen to summarize comments received on these sections in one portion of this notice. Comments that pertain only to specific sections are discussed under that particular section heading.

As discussed earlier, the training and experience requirements in proposed § 35.290 were moved to final § 35.190 and the training and experience requirements in proposed § 35.292 were moved to final § 35.290. For purpose of the following discussion, the summary of the comments refers to the section in the proposed rule and the response refers to the sections in the final rule.

Section 35.190, Training for uptake, dilution, and excretion studies.

Section 35.290, Training for imaging and localization studies.

Section 35.390, Training for use of unsealed byproduct material for which a written directive is required.

Section 35.392, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).

Section 35.394, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).

Issue 1: Should NRC requirements focus on radiation safety rather than clinical proficiency?

Comment. Commenters were generally in support of the NRC focusing training and experience requirements on radiation safety rather than on clinical competency. Some commenters believe that the training and experience for physicians who wish to use unsealed byproduct material should be based on demonstrated competence in nuclear science and radiation safety. These commenters did not believe that the NRC should define the criteria for clinical competence, rather it should allow clinical training to be defined by relevant medical specialty organizations such as the Accreditation Council for Graduate Medical Education (ACGME)-approved training programs or the ABMS-sanctioned certifying boards. However, commenters noted that “AU status” was frequently equated with clinical competency. As a result, these commenters encouraged the NRC to clearly state that a license granted under Part 35 only reflects the qualifications of a physician to safely handle radioactive material for medical use and not to practice nuclear medicine.

Response. The current training and experience requirements for AUs under §§ 35.100, 35.200 and 35.300 have been revised to focus on radiation safety. The focus of these training requirements should not be clinical proficiency. Clinical proficiency is best addressed by State Medical

Boards, certifying organizations, and hospital credentialing committees. An individual's status as an AU means that the individual has met the requirements to handle byproduct material safely. It does not reflect an assessment of the individual's clinical or professional competency.

Issue 2: Should training and experience be limited to FDA-approved uses of byproduct material?

Comment. A commenter recommended that training and experience be obtained in those activities that are related to FDA-approved uses of byproduct materials, and that all research, drug testing, and related non-FDA approved procedures be excluded from training and experience activities.

Response. These training and experience requirements focus on radiation safety not on clinical proficiency. Therefore, we believe individuals should have training and experience in the safe handling all types of byproduct material. Thus, limiting training and experience to FDA-approved uses of byproduct materials is not needed.

Issue 3: Where should training be obtained?

Comment. A commenter recommended that the NRC not recognize training and experience that has been obtained at a facility that is supported by commercial manufacturers or suppliers. Other commenters recommended that practical training should be in an ACGME-accredited program in

nuclear medicine or a graduate level course at an accredited university. Another commenter recommended that only those physicians completing an accredited residency program in an ABMS-approved speciality be allowed to become AUs under § 35.390.

Response. We do not believe that the rule should specify where the training should be obtained because this level of prescriptiveness is not warranted by the types and levels of byproduct material that are handled under §§ 35.100, 35.200, and 35.300. We will investigate any allegations regarding inadequate training programs on a case-by-case basis. In addition, we do not believe that the rule should prohibit an individual from obtaining training at locations whose activities are supported by commercial manufacturers, suppliers, or the owners/investors. We will rely on the preceptor's written certification for final assurance that an individual has completed the required training and experience and is competent to function independently as an AU.

Issue 4: Should NRC provide "deemed" status to individuals?

Comment: Commenters questioned whether NRC would provide "deemed" status to diplomates of the American Board of Nuclear Medicine (ABNM) and whether diplomates of the American Board of Radiology (ABR) or the ABNM should be licensed to use diagnostic radionuclides without additional education or examination requirements.

Response. Any individual who is an AMP, teletherapy physicist, ANP, AU, or RSO will continue to be considered such by NRC. Hence, such individuals will have "deemed" status as an AMP, ANP, AU, or RSO. (Individuals currently recognized as "teletherapy physicists" would be considered to be an AMP under the final Part 35.) Once the final rule becomes effective, diplomates of boards such as the ABNM and ABR will be considered to have met the training and experience requirements if the boards have been recognized by NRC.

Issue 5: Why are there different requirements of training for users under §§ 35.100, 35.200, and 35.300?

Comment. Commenters questioned why the training and experience requirements for using byproduct materials under §§ 35.100, 35.200, and 35.300 differed. They indicated that the basic radiation safety practices and knowledge of radiation science should be the same regardless of the quantity of byproduct material and how it is used.

Response. We recognize that there is a certain degree of basic radiation safety knowledge that is common among all the types of use, e.g. use of the decay formula and decontamination techniques. However, we also believe that there are some basic differences between use of byproduct material under §§ 35.100, 35.200, and 35.300 that warrant additional training and experience, e.g. increased potential for exposures in excess of Part 20 limits and the potential for adverse biological effects. For example, AUs handling byproduct material for imaging and localization studies, as compared to uptake, dilution, and excretion studies, are generally handling larger quantities and many different radionuclides. Also, AUs meeting the training and experience

requirements in § 35.190 are not authorized to prepare radioactive drugs using generators and reagent kits but, AUs under § 35.290 are authorized to prepare drugs using generators and reagent kits. Finally, AUs under § 35.390 are handling material in quantities that can cause deterministic effects.

Issue 6: How long should the training program be for individuals who would like to become AUs under §§ 35.190, 35.290, and 35.390?

Comment. We received numerous comments both in support of and in opposition to the duration of the proposed training and experience requirements for individuals who would like to become an AU for unsealed byproduct material.

Some commenters strongly supported the proposed reduction of the training and experience requirements for use of unsealed byproduct material in diagnostic nuclear cardiology because of the minimal risk to patients and public safety.

Some commenters believed that NRC should not establish an “arbitrary” number of training and experience hours. They indicated that it may take some individuals more time to master needed information. They believe that classroom training should focus on radiation safety and that there should be a requirement to show evidence of mastery in comprehensive nuclear and radiation science through an exam. In addition, the rule should clearly identify what knowledge and skills an individual should have.

A commenter suggested that the proposed requirements for an individual who would like to use material under § 35.100 be changed from 20 hours of classroom and laboratory experience to 40 hours of supervised practical experience.

A commenter recommended that the proposed requirement for an individual who would like to use material under § 35.200 should be a minimum of 240 hours of supervised practical experience. For the same type of use, another commenter suggested that an individual complete a six month/1200 hour training program in an ACGME-accredited or equivalent training program. Finally, a commenter recommended that individuals certified by the ABR or ABNM should automatically qualify as AUs. These commenters also indicated that as an alternative pathway to board certification, an individual who would like to use material under § 35.200 should be required to complete a dedicated 4-month nuclear medicine/radiology training program that integrates radiation safety training with clinical training and experience. This integrated experience should be obtained in a ACGME-approved residency program in diagnostic radiology or nuclear medicine.

A commenter stated that the current training and experience requirements for physicians authorized for nuclear medicine therapy (§ 35.390) are minimal to a fault. The commenter cited the 1996 National Academy of Sciences-Institutes of Medicine analysis of NRC's medical program that recommended increasing the requirements for a nuclear medicine therapy AU. Another commenter found it inconsistent that the use of unsealed byproduct material for therapy requires far less training than the use of sealed byproduct material. Another position is that therapeutic nuclear medicine

represents a higher risk for patients. Therefore, the training and experience requirements to become an AU for therapy should be greater than those for diagnostic nuclear medicine.

A commenter recommended that the current requirements for an individual who would like to use unsealed byproduct material under § 35.300 should be revised to be at least equal to or greater than the requirements to use material under § 35.200. Another commenter suggested that an individual have 100 hours, rather than 40 hours, of supervised practical experience under the supervision of an AU. The commenter went on to state that this additional time would be used to cover the requirements that pertain to dosages requiring a written directive.

Another commenter stressed the importance of remembering that under § 35.300, byproduct material is used for therapeutic treatments and that the possibility of injury to the patient and others is very real. This commenter stated that he had personally seen significant bone marrow suppression after using strontium for bone pain as well as life-threatening pulmonary edema after treatment of a patient with Iodine-131 (I-131) for metastatic thyroid cancer of the lungs.

Response. We believe the regulatory text should contain a list of the subject areas to be addressed in a training program. We have not included a requirement in the final rule for an examination that would be used to demonstrate an individual has sufficient knowledge in radiation safety. Instead, we will rely on the duration of the training program and the preceptor's written certification that a physician has completed the required training and experience and is competent to function independently as an AU.

The following discussion summarizes the training and experience requirements for use of unsealed byproduct material under §§ 35.100, 35.200, and 35.300. We believe the specified training periods will provide individuals with sufficient knowledge to handle byproduct material safely. We also believe that it is sufficient to specify the overall period for training. We do not believe that any further breakdown is needed in terms of the hours devoted to classroom/laboratory training and work experience. Note, this same approach is used in the current rule for the training and experience requirements for an ANP and will provide needed flexibility in designing and implementing training programs.

In § 35.190, Training for uptake dilution and excretion studies, the total number of hours (i.e., 60 hours) in the proposed rule is the same as the total number of hours in the current rule and in the final rule. AUs, qualified under §§ 35.290, 35.390, or equivalent Agreement State requirements may use byproduct material under § 35.100. AUs qualified under § 35.190 are not be authorized to prepare unsealed byproduct material using generators and reagent kits.

In § 35.290, Training for imaging and localization studies, we agree with the public comments that the proposed 120 hours is not sufficient. AUs in this category are authorized to prepare unsealed byproduct material for medical use using generators and reagent kits. Therefore, we have increased the period of training from 120 hours to 700 hours (essentially four months) in this section of the final rule. This change was necessary to assure that physicians spend an adequate amount of time in an environment in which radioactive drugs are routinely being prepared and/or administered for medical use.

As stated earlier, we have not specified a breakdown between the number of hours of didactic (i.e., classroom and laboratory) and work experience to allow licensees flexibility in designing and implementing training programs. Therefore, the number of hours of classroom and laboratory training needed to address the required subject areas in § 35.290(c)(1)(i) may vary with individual training programs. The remainder of the required 700 hours would be devoted to supervised work experience to include, but not be limited to, the subject areas in § 35.290(c)(1)(ii).

We recognize that physicians in training will not dedicate all of their time specifically to the subject areas in § 35.290(c)(1)(ii) and will be attending to other clinical matters involving the diagnostic use of the material under the supervision of an AU (e.g., reviewing case histories or interpreting scans). Even though such clinical matters are not specifically required by the NRC, such supervised work experience may be counted toward the supervised work experience to obtain the required 700 hours.

We agree that the training and experience requirements should be increased for individuals who would like to use byproduct material for which a written directive is required. In the final rule, § 35.390, Training for use of unsealed byproduct material or for use of unsealed byproduct material that requires a written directive, the hours have been increased from 80 hours to 700 hours. We believe this increase is needed because these physicians would be authorized to elute generators and prepare radioactive drugs as well as administer a wide variety of radionuclides requiring written directives and thus the associated radiation risks of the use could be greater. In addition, the work experience in the administration of such dosages to patients must specifically include at least three cases in each of the following categories for which the individual is requesting AU status:

1. Oral administration of less than or equal to 33 millicuries of sodium iodide I-131;
2. Oral administration of greater than 33 millicuries of sodium iodide I-131;
3. Parenteral administration of any beta-emitter or a photon-emitting radionuclide with a photon energy of less than 150 keV; and/or
4. Parenteral administration of any other radionuclide.

Physicians who are authorized under § 35.390 for all the above types of administrations also meet the requirements in §§ 35.190, 35.290, 35.392, and 35.394.

Issue 7: What are the appropriate training requirements for an individual who would like to use I-131 for treatment of hyperthyroidism and thyroid cancer?

Comment. Commenters were strongly opposed to the proposed changes to the requirements for the administration of I-131 for treatment of hyperthyroidism and thyroid cancer. Commenters felt that there was no justification for revising the current § 35.932, Training for treatment of hyperthyroidism, and to do so would conflict with NRC's guidelines of "minimizing intrusion into medical judgements affecting patients and into other areas considered to be a part of the practice of medicine." These commenters indicated that the increased training was not warranted in light of endocrinologists' impeccable safety record with the use of I-131 and the fact that there have been no records of therapeutic misadministrations of any byproduct material by endocrinologists. In addition, commenters stated that, in reality, most of the practical aspects of handling I-131 that would be

covered in the proposed 40 hours of additional training is already covered in the 80 hours of didactic training and in the supervised clinical training that is currently required by §§ 35.932, Training for treatment of hyperthyroidism, and 35.934, Training for treatment of thyroid carcinoma.

Commenters stated that the clinical endocrinologist is the physician best qualified to take care of patients with thyroid disease and part of their responsibility is to protect their patients from unnecessary burdens. Commenters stated that the practical effect of increasing the basic radiation physics and safety training from 80 hours to 120 hours would be to exclude endocrinologists from administering I-131 to patients with hyperthyroidism and thyroid cancer. Some commenters went on to state that increasing the requirement for licensure would actually result in fewer endocrinologists being able to take care of their own patients and would ultimately place increased and undue strain on the patients such as:

1. Increased costs to the patient. The cost to patients receiving treatment in a hospital setting are double or triple the cost of an endocrinologist administering I-131 in his/her own office.

2. Increased potential safety hazards for the patient. There is much more personal and focused attention given to the patient in the endocrinologist's office. In other settings, the patient is one of dozens of people waiting to be treated with a variety of doses for a variety of diseases. Thus, the possibility of error in communications and for the misadministration of I-131 is greatly increased.

3. Increased emotional trauma during treatment. Patient anxiety and fear will be increased as a result of patients being required to go to nuclear medicine departments where other patients are being treated for all manner of disease, including cancer. This is an unnecessary exposure of the patient to psychological trauma and can be a deterrent to a patient seeking appropriate care.

4. Increased hassles visiting another specialist. With fewer endocrinologists administering I-131, patients will have to endure another layer of speciality consultation, resulting in delays in treatment, inconvenience and loss of time from work, significant increase in the cost of treatment, and exposure to unfamiliar settings and personnel.

Commenters were also concerned that the proposed rule required that the 40 hours of supervised practical experience be obtained at a medical institution. They thought this is a prescriptive requirement which is not warranted because acceptable training could be provided in other clinical settings. Other commenters noted that this requirement would make it more difficult for endocrinologists to receive supervised practical experience from mentors or preceptors who practice and administer radioiodine in their offices rather than in a medical institution.

A commenter thought it paradoxical that the proposed rule would actually decrease the amount of clinical experience needed for licensure. The commenter indicated currently under § 35.932, a physician is required to have documented, supervised clinical experience with 10 patients with hyperthyroidism and under § 35.934 they are required to have experience with 3 patients with thyroid cancer. The commenter indicated that in the proposed rule, an individual must have experience

with 5 cases. This commenter believed this was a step backward from the current regulations, because the clinical experience and practical aspects of the use of radioiodine that are obtained during clinical experience rather than obtained in a classroom setting. According to another commenter, the blanket requirement for 5 cases for each procedure may not always be appropriate. This commenter thought that it might be better to list the procedures and the number of required cases in the regulations.

Response. In the final rule, §§ 35.392 and 35.394 have been added to specifically address oral administrations of sodium iodide I-131. These sections do not increase the duration of training for an endocrinologist over the current requirements in §§ 35.932 and 35.934.

In the final rule, § 35.392 was added to provide the training and experience requirements for physicians who only seek authorization for the oral administration of sodium iodide I-131 in dosages less than or equal to 1.22 gigabecquerel (GBq) (33 millicurie) and do not seek authorization to prepare radioactive drugs using generators and reagent kits. To qualify as an AU under this limited authorization, an individual must have 80 hours of classroom and laboratory training and supervised work experience that includes 3 cases involving the oral administration of sodium iodide I-131 in dosages less than or equal to 1.22 GBq (33 millicurie). We have not specified a breakdown between the number of hours of didactic (i.e., classroom and laboratory) and supervised work experience to allow licensees flexibility in designing and implementing training programs. Therefore, the number of hours of classroom and laboratory training and supervised work experience needed to adequately address the required subject areas can vary with individual training programs. These individuals may not prepare unsealed byproduct materials using generators and reagent kits.

Also, in the final rule, § 35.394 was added to provide training and experience requirements for physicians who only seek authorization for the oral administration of sodium iodide I-131 in dosages greater than 1.22 GBq (33 millicurie) and do not seek authorization to prepare radioactive drugs using generators and reagent kits. This limited authorization requires 80 hours of classroom and laboratory training and work experience that includes 3 cases involving the oral administration of sodium iodide I-131 in quantities greater than 1.22 GBq (33 millicurie). Physicians authorized under § 35.394 would also meet the training and experience criteria in § 35.392. These individuals may not prepare unsealed byproduct materials using generators and reagent kits.

We agree that it is not necessary for the supervised work experience required by §§ 35.392 and 35.394 to be obtained at a medical institution. The essential element of this requirement is who is supervising the individual rather than where the experience is obtained. The final rule allows an individual to obtain work experience at any type of medical facility (e.g., medical institution, clinic, private practice office) provided the experience is under the supervision of an AU who meets the applicable requirements.

Issue 8: Should there be a difference between the training and experience requirements for use of sodium iodide I-131 liquid and capsules?

Comment. A commenter indicated that an individual who only planned on using iodine in a capsule should not be required to have as much training as someone who planned on using liquid iodine. The commenter recommended that only 40 hours of training was needed to learn how to handle I-131 capsules.

Response. The final training and experience requirements do not differentiate between the different forms of I-131. We believe that AUs should have the flexibility to prescribe whatever form of I-131 they believe appropriate. Although there are differences between handling iodine in capsule form and liquid form, e.g. decontamination procedures and volatility issues, we do not believe that the differences are significant enough to warrant a separate category for training.

Issue 9: Should diagnostic use of I-131 be authorized under §§ 35.200 or 35.300?

Comment. A commenter noted that the proposed rule would move requirements for whole body imaging using I-131 sodium iodine from §§ 35.200 to 35.300. The commenter argued that this would prevent physicians who are imaging specialists from performing the procedure and allow therapy specialists to do the procedure. This commenter suggested that the procedure not be included in either, but instead be listed as a line item authorization and that specified training and experience requirements be adopted for it.

Response. We do not believe that training and experience criteria for the use of sodium iodide I-131 for whole body imaging should be excluded from the regulations. The radiation safety considerations associated with the diagnostic use of millicurie quantities of sodium iodine I-131 more closely resemble the therapeutic use of sodium iodide I-131 than most diagnostic imaging and localization studies using technetium-99m. Therefore, the training and experience requirements for the use of sodium iodide I-131 in quantities greater than 1.11 Megabecquerel (MBq) (30 microcurie), regardless of how it will be used, requires additional experience in the administration of these types of dosages.

The final rule reduces the required number of cases, as stated in the proposed rule, from 5 to 3 for each type of use for which authorization is requested. We believe that a physician's involvement in 3 cases will provide him or her with adequate training and experience. In addition, we do not believe that requiring physicians to obtain administration experience or demonstrate they have such experience for three cases of sodium iodide I-131 represents an unwarranted burden nor would it discourage such physicians from becoming authorized to use I-131.

Issue 10: Should both §§ 35.290 and 35.292 in the final rule refer to reagent kits?

Comment. A commenter stated that the proposed § 35.292 does not refer to "reagent kits," although proposed § 35.290 does and questioned whether this was an error.

Response. The training and experience requirements to become an AU for imaging and localization require a physician to have experience with generators and reagent kits because physicians authorized under final § 35.290 may prepare unsealed byproduct material using generator systems and reagent kits. Under final § 35.190, physicians are not authorized to prepare byproduct material using generator systems and reagent kits. Therefore, it is appropriate that § 35.290 and not § 35.190 require experience with eluting generator systems appropriate for preparing unsealed byproduct material for imaging and localization studies, measuring and testing the eluate for radiochemical purity, and processing the eluate with reagent kits.

Issue 11: Is it necessary to require training in calibrating dose calibrators and in calculating and measuring dosages?

Comment. A commenter stated that there was an inconsistency between the training and experience requirements in the proposed §§ 35.292 and 35.390 and the requirement to calibrate dose calibrators in § 35.60 and the requirement to measure unit dosages in § 35.63. The commenter recommended that we replace the phrase “Calculating, measuring, and safely preparing patient or human research subject dosages,” with the phrase “Determining and safely preparing patient or human research subject dosages.”

Response. We believe physicians who plan to use unsealed byproduct material must have training in calibrating instruments used to measure the activity of unsealed byproduct materials, in calculating and measuring dosages, and in eluting generators even though, in practice, an AU may choose to only use unit dosages. We believe that this training is important because AUs who meet the qualifications in the final §§ 35.290 and 35.390 are not restricted to using unit dosages. The training requirements do not interfere with the practice of medicine or pharmacy because the rule provides sufficient flexibility for procuring and preparing unsealed byproduct material.

We have not replaced the words “calculating and measuring” with the word “determining.” Use of the words “calculating and measuring” clearly states our intent that an individual receive training in calculating (perform radioactive decay calculations) and measuring (use instrumentation to determine activity) the activity of unsealed byproduct material.

Issue 12: Were there any other changes made to these sections between the proposed and final rule?

Response: Yes. We revised the requirement for individuals to have experience administering dosages to patients or human research subjects to state: “Administering dosages of radioactive drugs to patients or human research subjects.” This was done to clearly state that experience administering radioactive drugs need not be limited to radioactive drugs containing byproduct material because there is no difference between the safety precautions that must be exercised when administering byproduct or nonbyproduct material.

We revised the requirement for individuals to have experience using procedures to contain spilled byproduct material safely and using proper decontamination procedures to state: “Using procedures to contain spilled radioactive material safely and using proper decontamination procedures.” This was done to clearly state that experience with containing spilled radioactive material and decontaminating areas need not be limited to byproduct material because there is no difference between the safety precautions that must be exercised when handling byproduct or nonbyproduct material.

We revised paragraph (b)(ii)(G) in 35.290 and 35.390 to state: “. . . measuring and testing the eluate for radionuclidic purity. . .” rather than “. . . measuring and testing the eluate for radiochemical purity.” This change was made because it more accurately reflects the testing that licensees actually perform for quality control testing on generator eluates, e.g., determining the molybdenum-99 concentration in the eluate from a molybdenum-99/technetium-99 m generator.

We added a reference to 35.390 in paragraph (b) of 35.100, 35.200, and 35.300. This was done to recognize that an individual who meets the requirements in 35.390 has sufficient training and experience to handle material safely under 35.100, 35.200, and 35.300.

2. Sealed Byproduct Material

For the most part, comments received on the following two sections related to more than one section. Therefore, we have chosen to summarize comments received on these two sections in one portion of this notice. Comments that pertain only to specific sections are discussed under that particular section heading.

Section 35.490, Training for use of manual brachytherapy sources.

Section 35.690, Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Issue 1: What is the appropriate level of training to require?

Comment. Some commenters felt that the current training requirements should be retained and that lessening of the current training requirements could have a tremendous detrimental effect on patient care. Many of these same commenters believed that the training for coronary artery therapy should be of the same level as for all other sealed source therapy. Conversely, some commenters supported lessening of the training requirements to a level that considers only radiation safety and not clinical competence.

Response. We did not change the training levels required by these sections. We believe that individuals should complete a structured educational program that includes both classroom and laboratory training and work experience. We recognize that radiation safety training and clinical competency may be intertwined, especially for therapeutic uses of sealed sources. Therefore, we agree that significant changes should not be made in the current training requirements for AUs in this area.

Issue 2: Can this section be revised to refer to the appropriate review committee and the appropriate time division reviewed by the committee?

Comment. A commenter suggested that §§ 35.490(b)(2) and 35.690(b)(2) should refer to the Residency Review Committee for Radiation Oncology (since 1993). The commenter also stated that the phrase “that includes one year in a formal training program” should be replaced with “in radiation oncology as part of a formal training program.”

Response. We agree with the suggested changes because the changes reflect the changes in the certification process since 1993. We have incorporated the requested revisions in the rule.

Issue 3: Is concurrent training allowed for clinical and work experience?

Comment. A commenter pointed out that, as written in the proposed rule, 6 years of training is required unless concurrent training is allowed. The commenter felt that the proposed rule would require 500 hours of supervised practical experience plus 3 years of supervised clinical experience. The commenter also felt that the proposed rule would require 3 years of training with, for instance, iridium-192 sources, and an additional 3 years of training in order to use gamma stereotactic radiosurgery sources.

Response. We agree that concurrent training should be allowed for the clinical and work (practical) experience requirements in this section to be consistent with allowances in §§ 35.490 and 35.690. Therefore, we revised the regulatory text in §§ 35.490(b)(2) and 35.690(b)(2) to allow for concurrent work and clinical experience.

Issue 4: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We have deleted the phrase "or equivalent program approved by the NRC" from §§ 35.490(b)(2) and 35.690(b)(2) because a program equivalent to the Accreditation Council for Graduate Medical Education (ACGME) program does not exist.

Global changes in the rule

Issue 1: What is the Sealed Source and Device Registry and how do I access the Registry?

Comment. A commenter noted that the proposed revision would be strengthened if there were an indication as to the nature of the Sealed Source and Device Registry and how to obtain a copy.

Response. The Sealed Source and Device Registry, as defined in § 35.2, is the national registry containing all the registration certificates, generated by both NRC and the Agreement States, that summarize the radiation safety information for sealed sources and devices and describe the licensing and use conditions approved for these products. The information contained in the registry is summarized from information provided during registration of the source or device in accordance with § 32.210, Registration of product information. The Commission or Agreement State evaluates the information submitted to register a source (or device) and, if acceptable, issues a “Safety Evaluation of Sealed Source (or Device).” A compilation of these evaluations can be found electronically at the following address: <http://www.hsrdr.gov/nrc/ssdr/ssdrindx.htm>

Issue 2: Should the requirements in the current rule related to possession of survey instruments be deleted?

Comment. A commenter stated that the requirements in the current Part 35 concerning possession of survey instruments are very useful and should not be deleted from the rule (§§ 35.120, 35.220, 35.320, 35.420, 35.520, and 35.620 in the current Part 35). This commenter believed that the Part 20 requirements are not specific enough on this point.

Response. We do not believe specific requirements relating to possession of survey instruments are needed in Part 35. Section 20.1501 requires that the licensee make, or cause to be made, surveys to demonstrate compliance with 10 CFR Part 20. This provision requires, in part, the licensee to ensure that instruments and equipment used to show compliance with Part 20 are periodically calibrated. In addition, § 30.33(a)(2) of this chapter requires licensees to have adequate instrumentation. Information on the types of instruments is available in NUREG-1556, Vol. 9.

Issue 3: Should the term “dose calibrator” be replaced with the term “radionuclide calibrator” in the training and experience requirements for unsealed byproduct material?

Comment. Commenters suggested that we replace the term “dose calibrator” with the term “radionuclide calibrator” in proposed §§ 35.50, 35.55, 35.290, 35.292, 35.390, 35.920 and 35.930.

Response. The reference to “dose calibrators” in §§ 35.50, 35.55, 35.290, and 35.292 has been deleted and replaced with “instruments used to determine the activity of dosages.” (Proposed §§ 35.920 and 35.930 were deleted by the final rule.) As stated under the discussion on § 35.60, this change recognizes that there are various types of instruments that can be used to measure the activity of unsealed byproduct materials. Therefore, we believe individuals should have experience with the different types of instruments and not limit them to only experience with dose calibrators.

Issue 4: Were there any other changes made to the rule between the proposed and final rule?

Response. Yes. References in the proposed rule to § 35.290 have been changed to § 35.190 and references to § 35.292 have been changed to § 35.290. This was done because the training and experience requirements in proposed §§ 35.290 and 35.292 were moved to §§ 35.190 and 35.290, respectively. This change groups the sections that specify the requirements for an individual who would like to become an AU for a specific type of use with the section that provides information on that specific type of use. For example, § 35.100 provides authorization for use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required and § 35.190 contains the training and experience requirements for someone who would like to use material under § 35.100.

Throughout the rule we have replaced the word “promptly” with the phrase “as-soon-as-possible.” In the proposed rule we used both “promptly” and “as-soon-as-possible.” For the purpose of this rule, both could be used interchangeably. Therefore, we have chosen to use the phrase “as-soon-as-possible” to maintain consistency within the rule. The phrase “as-soon-as-possible” is used to indicate that the required action should be taken immediately considering the circumstances. The term “as soon as possible” adds a degree of reasonableness to “immediate.” For example, a notification might be made the next morning rather than in the middle of the night.

Part III - Specific Comments on the Proposed Rule

10 CFR—STANDARDS FOR PROTECTION AGAINST RADIATION

Section 20.1002, Scope.

Issue 1: Were any other changes made to this section between the proposed and final rule?

Response. Yes. We revised this section to replace the phrase “from exposure to individuals administered radioactive material and released in accordance in accordance with § 35.75” with the phrase “from exposure to individuals administered radioactive material and released, which is governed by § 35.75.” This is being done to further clarify that the dose to individual members of the public from a licensed operation does not include doses received by individuals exposed to patients who were released by the licensee under the provisions of § 35.75.

In 1997, we amended the regulations concerning the criteria for the release of patients administered radioactive material to base the criteria for patient release on the potential dose to other individuals exposed to the patient (62 FR 4120; January 29, 1997). As part of that rulemaking, we also amended the regulatory text in §§ 20.1002, 20.1003 and 21.1301 to reflect the Commission’s policy that patient release is governed by § 35.75, not § 20.1301 (62 FR 4120; January 29, 1997, see page 4122).

Current §§ 20.1002, 20.1003, and 20.1301(a)(1) indicate that the dose to individual members of the public or to an occupationally exposed individual from a licensed operation do not include doses received by individuals exposed to patients who were released in accordance with § 35.75. Upon further review, we believe that changes must be made to the current regulatory text in §§ 20.1002, 20.1003, and 20.1301, to further clarify that the

dose limits do not apply to the maximally exposed individual from a patient or human research subject who has been administered unsealed byproduct material or implant containing byproduct material (reference § 35.75) and has been released from licensee control.

Under § 35.75, a licensee may release an individual from its control if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 millisievert (mSv)(0.5 rem). The licensee is required to comply with all the requirements in § 35.75. Failure to comply with any of the provisions in § 35.75 may result in enforcement action. This change in Part 20 makes it clear that any violations will be cited against § 35.75 and not Part 20.

Section 20.1003, Definitions.

Issue 1: Were any other changes made to this section between the proposed and final rule?

Response. Yes. We have made corresponding changes to the definitions for occupational dose and public dose to state that doses do not include doses received by individuals exposed to patients who were released by the licensee under the provisions of 35.75. Specifically, we revised these definitions to replace the phrase “from exposure to individuals administered radioactive material and released in accordance in accordance with § 35.75” with the phrase “from exposure to individuals administered radioactive material and released, which is governed by § 35.75.” This change is discussed in depth under the discussion on § 20.1002, above.

Section 20.1301, Dose limits for individual members of the public.

Issue 1: Who should approve whether a visitor is allowed to receive a dose up to 0.5 rem?

Comment. A commenter suggested that the RSO, not the AU, should be the appropriate individual to approve the merits of allowing a visitor to receive up to 0.5 rem.

Response. AUs have the primary responsibility for the health and safety of their patients. They are also responsible for determining, depending on the patient's conditions, whether visitors can visit their patients and with what limitations. Therefore, we believe the AU should approve whether a visitor is allowed to receive a dose up to 0.5 rem. However, the AU may consult with the RSO at any time regarding visitor control.

Issue 2: Should visitors be allowed to receive a dose up to 0.5 rem?

Comment. The commenter stated that the proposed rule did not meet any standard for justifying an increased exposure to someone visiting a hospitalized (confined) patient. The commenter indicated that one of the reasons for the increased dose limit in § 35.75 was the economic benefit of allowing the patient or human research subject to be released from control earlier. He went on to state that in the case of the proposed revision to §

20.1301, there was no economic benefit to the licensee and that NRC was basing this change on an emotional benefit to the patient rather than an economic benefit.

Response. The justification for this change was discussed in detail in the Statements of Consideration for the proposed rule and in the associated draft Regulatory Analysis. It is restated in Part II of the Supplementary Information section in this document and in the final Regulatory Analysis. Overall, we believe the emotional benefit to the patient or the visitor outweighs the increase in radiation risk to the visitor. We believe AUs should have the flexibility to make a determination, based on their judgment, as to whether a patient or human research subject would benefit from allowing a visitor to receive a dose up to 5 mSv (500 mrem). The AU must consider the patient's condition when determining whether it is appropriate to allow a visitor to receive a dose up to 0.5 rem.

Issue 3: Were any other changes made to this section between the proposed and final rule?

Response. Yes. We have changed the regulatory text in § 20.1301(a)(1) to indicate that the dose to individual members of the public from a licensed operation does not include doses received by individuals exposed to patients who were released by the licensed operation under the provisions of 35.75. Specifically, we revised this section to replace the phrase "from exposure to individuals administered radioactive material and released in accordance in accordance with § 35.75" with the phrase "from exposure to individuals administered radioactive material and released which is governed by § 35.75." This change is discussed in depth under the discussion on § 20.1002, above.

PART 35 - MEDICAL USE OF BYPRODUCT MATERIAL

SUBPART A - General Information

Section 35.1 Purpose and Scope.

Issue 1: How does this rule provide for the radiation safety of patients?

Comment. Commenters did not believe that Part 35 should address radiation safety of patients because it would necessitate NRC making medical judgments. Commenters noted that physicians are trained to make informed decisions on behalf of patients. They believed that the NRC should insure that those practicing nuclear medicine are adequately trained in nuclear science, thus insuring that the radiation safety of patients is provided for.

Response. We did not make any changes to the regulatory text. We believe that the NRC should provide for the radiation safety of the public, workers, and patients. The Commission's goal in regulating nuclear materials safety, as stated in its September 1997 "Strategic Plan" (NUREG-1614, Vol. 1), is to "prevent radiation-related deaths or illnesses due to civilian use of source, byproduct material, and special nuclear material." Protection of the radiation safety of the public, workers, and the patient is central to the fulfillment of the Commission's statutory mandate to "protect health and minimize danger to life."

The Commission has determined to retain its long-standing medical use regulatory program. However, it is doing so with improvements, including decreased oversight of low-risk activities and continued emphasis on high-risk activities. The Commission has long recognized that physicians have the primary responsibility for the diagnosis and treatment of their patients. NRC regulations are predicated on the assumption that properly trained and adequately informed physicians will make decisions that are in the best interest of their patients. However, NRC has a secondary, but necessary, role with respect to the radiation safety of patients. The NRC will, when justified by the risk to patients, regulate their radiation safety, primarily to assure that the use of radionuclides is in accordance with the physician's directions.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We replaced the word "prescribes" with the phrase "contains the" in the first sentence of the section because Part 35 contains the requirements and provisions for the medical use of byproduct material and for issuance of specific licenses authorizing medical use.

Section 35.2 Definitions.

We received numerous comments on the definitions. Commenters asked us to revise, delete, or add definitions for terms used in the rule. We have also added some new terms in this section because of changes made in other sections of the rule. Public comments and our response to the comments, as well as the reasons for other changes to this Section, are presented below in alphabetical order of the terms.

Area of use

Issue1: Were there any other changes made in this definition between the proposed and final rule?

Response: Yes. We added the word "preparing" to the definition. This was done to recognize that licensees not only receive, use, and store byproduct material but, in the case of a medical licensee, they may also prepare the material for use.

Authorized Medical Physicist

Issue 1: Should the term "medical physicist" be used in the rule?

Comment. Commenters believed that a "medical physicist" would better be defined by a unique term, similar to "Authorized User," which has no meaning outside the regulations. They stated that use of the term "authorized physicist" would be consistent with "authorized user."

Response. We have retained the term "authorized medical physicist" in the final rule. This was done to maintain consistency with other terms used in Part 35 (AU and ANP). We also believe the term "authorized physicist" may be too broad and we would like to make it clear that this individual has experience as a medical physicist.

Issue 2: Can an AMP be an AU?

Comment. Commenters questioned whether a medical physicist could be the AU and if so, whether there would be a need to have a physician listed on a nuclear medicine license?

Response. It is always necessary to name an AU on the Part 35 license because only an AU can prescribe dosages or doses of byproduct material for medical use under Part 35. An AU for medical use under §§ 35.100, 35.200, 35.300, 35.400, and 35.600 must be a physician. An AU for medical use under § 35.500 may be a physician, dentist, or podiatrist. An AMP could only be an AU, named in the license, if the AMP meets the criteria in the definition of AU in § 35.2, including the training and experience criteria cited in that section.

Issue 2: Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. In addition to restructuring the definition, to make it more readable, we substituted the word “individual” for the word “physicist.” This change was made so that the definition of the term would be similar to the definition for an RSO.

We also revised the definition for the AMP to include individuals identified as an AMP on (1) a permit issued by a Commission master material licensee or (2) a permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating AMPs. This change, which was also made to the definitions of "ANP," "RSO," and "AU," accounts for the fact that an AMP may be named on a permit issued by a master material licensee. For example, in the first case identified above, if a master material licensee has issued a permit that recognized a particular individual as an AMP, under this definition, the individual would continue to meet the requirements for an AMP under an NRC license. In the second case, if a master material licensee chooses to issue a broad scope permit to a hospital and that hospital has authorization to issue permits designating AMPs, under this definition, an AMP on the permit would also meet the requirements for an AMP under an NRC license.

Authorized Nuclear Pharmacist

Issue 1: What are the duties of an ANP?

Comment. A commenter stated that the responsibilities and duties of the ANP were not codified.

Response. We did not change the regulatory text. We have used the definitions section to provide an understanding of what we mean by a term. We do not believe it is appropriate to list the responsibilities and duties of the ANP in the definitions section or elsewhere in the rule. In most cases, we have not specified who must perform a particular duty. This was done to give the licensee flexibility in how it implements its radiation protection program.

In a limited number of cases, where justified by risk, we have specified who must perform specific duties. For example, the full calibration measurements on remote afterloader must be performed by an AMP (§ 35.633(h)).

Issue 2: Why do nuclear pharmacies have the authority to approve ANP's?

Comment. A commenter did not believe that nuclear pharmacies should be authorized to approve ANPs.

Response. This commenter objected to one way by which an individual may be qualified to be an ANP (i.e., approval as such by a nuclear pharmacy authorized to approve ANPs). This qualification was added to the rule for two reasons. One, the current definition needed to recognize that 10 CFR 35.72(b)(4) allows nuclear pharmacies to designate a pharmacist as an ANP if the individual meets certain requirements. Specifically, § 32.72(b)(4) contains a “grandfathering” provision permitting certain Part 32 nuclear pharmacy licensees to designate, as a nuclear pharmacist (as defined in § 35.2), an individual identified as of December 2, 1994, as an AU on a nuclear pharmacy license issued by the Commission. (If you would like additional information on § 32.72, refer to the regulatory history of the radiopharmacy rule (58 FR 33396; December 2, 1994, see page 33400)). Second, this change is needed because some nuclear pharmacies have a license amendment that would allow them to approve AMPs if the individual meets the training and experience requirements in Part 35. Without this corresponding change in Part 35, the individual would not be allowed to function as an ANP regardless of the nuclear pharmacy's approval.

Issue 3: Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. The definition was restructured to make it more readable. We also revised the definition for the ANP to include pharmacists identified as ANPs on (1) a permit issued by a Commission master material licensee or (2) a permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating ANPs. This change, which was also made to the definitions of "AMP," "RSO," and AU," accounts for the fact that an ANP may be named on a permit issued by a master material licensee. For example, in the first case identified above, if a master material licensee has issued a permit that recognizes a particular individual as an ANP, under this definition the individual would continue to meet the requirements for an ANP under an NRC license. In the second case, if a master material licensee chooses to issue a broad scope permit to a hospital and that hospital has authorization to issue permits designating ANPs, under this definition an ANP on the permit would also meet the requirements for an ANP under an NRC license.

Authorized User

Issue 1: What does an AU do?

Comment. A commenter recommended the definition of "Authorized user" include the duties of an AU.

Response. We did not include the duties of the AU in the definition. We have used the definitions section to provide an understanding of what we mean by a term. Duties that must be performed by the AU are stated in the rule text where appropriate. The issue whether the duties of a licensed individual belong in the definition section is discussed in more detail under the term “authorized nuclear pharmacist.”

Issue 2: Does the rule distinguish between different types of AU’s?

Comment. A commenter recommended we clarify each type of AU, or distinguish between AU’s involved in diagnostic versus therapeutic medical uses.

Response. We do not believe the definition should be modified in this way. Other requirements in this part address the safety requirements for the different types of medical uses and the AU’s actual duties. For example, the training and experience requirements for AU’s as well as other requirements in the regulations differentiate between diagnostic and therapeutic medical uses of byproduct material. The training and experience requirements for an AU who would like to perform uptake, dilution, and excretion studies (§ 35.290) differ from the training and experience requirements for an AU who would like to use unsealed byproduct material for therapy (§ 35.390). Also, radiation safety requirements are not the same for diagnostic as compared to therapeutic medical uses. Finally, the medical use license indicates what materials can be used by an AU.

Issue 3: Can non-physicians be AU’s?

Comment. A commenter noted that although the definition of “AU” refers to “any prescriber,” (i.e., physician, dentist, or podiatrist),” the proposed rule language (in §§ 35.100, 35.200, and 35.300) refers only to a physician. The commenter indicated that if dentists and podiatrists cannot be AU’s, the regulations should state this.

Response. Section 35.2 contains a general definition of an AU. Specific training and experience requirements for AUs are contained elsewhere within the regulatory text of Part 35. Where appropriate, the rule does specify when an AU must be a physician. An AU of materials authorized in §§ 35.100, 35.200, 35.300, 35.400, and 35.600 must be a physician. An AU using materials authorized under § 35.500 can be a physician, dentist, or podiatrist if that individual meets all other training and experience requirements for this type of use.

Issue 4: Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. We also revised the definition for the AU to include physicians, dentists, or podiatrists identified as AUs on (1) a permit issued by a Commission master material licensee or (2) a permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating AUs. This change, which was also made to the definitions of "ANP," "AMP," and "RSO," accounts for the fact that an AU may be named on a permit issued by a master material licensee. For example, in the first case identified above, if a master material licensee has issued a permit that recognizes a particular individual as an AU, under this definition the individual would continue to meet the requirements for an AU under an NRC license. In the second case, if a master material licensee chooses to issue a broad scope permit to a hospital and

that hospital has authorization to issue permits designating AUs, under this definition an AU on the permit would also meet the requirements for an AU under an NRC license.

We also added a reference to new sections in the final rule that lists the training and experience requirements for individuals using only iodine-131 in quantities that would require a written directive (§§ 35.392 and 35.394) and for using strontium-90 for ophthalmic treatments (§ 35.491).

Brachytherapy

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We added a definition for brachytherapy. This was done because we believe it is important to define such a term as it is used in Part 35 so that the regulated community and regulatory agencies have a clear understanding of what we mean when we use the term in the rule.

Brachytherapy source

Issue 1: Were there any changes made in this definition between the proposed and final rule?

Response: We did not receive any public comment on this definition. We did however delete the word “sealed” in the definition. This was done in order to include sources which do not meet the definition of “sealed source” (i.e., “radioactive plated, embedded, and activated” sources).

Client’s address

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We have defined this term because we now use it in § 35.80, “Provision of mobile medical service.” The term “client's address” encompasses an area of use as well as a temporary jobsite. Use of this term in the rule is explained in greater depth under the discussion of § 35.80.

Diagnostic clinical procedures manual

Issue 1: Is this term needed?

Comment. Commenters recommended this term be deleted because it is too prescriptive and should be replaced with the term “radiopharmaceutical prescription/order”. A radiopharmaceutical prescription/order can either be written for an individual patient (e.g., a written directive for therapeutic radiopharmaceuticals) or in the form of specific standing orders. The commenter was concerned that use of the term “clinical procedures

manual” may limit a licensee’s ability to compound radioactive drugs. As such, according to the commenter, the term raises a clinical medical practice issue under state law regarding the practice of medicine and pharmacy. The commenter believed that it would be more appropriate for the NRC to require a general description of the radiation safety procedures used to protect workers, the public, and other patients from unintentional exposures. The commenter indicated the procedure manuals are written by physicians and should only be considered as informational or guidance documents for technologists.

Response. In response to this comment, we have deleted “diagnostic procedures manual” both as a defined term in § 35.2 and from the definition of “prescribed dosage” in § 35.2. It is not used in the regulatory text. Therefore, we no longer need to define it.

As modified, the rule is less prescriptive and does not limit a licensee’s ability to compound certain radioactive drugs. Sections 35.100, 35.200, and 35.300 permit certain uses of unsealed byproduct material which are prepared by an ANP, a physician who is an AU (meeting certain requirements), or an individual under their supervision.

Health Physicist

Comment. A commenter asked that we add a definition for “health physicist.” This individual would be defined to be “a person qualified in the art, science and professional practice of radiation safety as evidenced by current certification by the American Board of Health Physics (ABHP) or an

equivalent certifying body with substantially the same requirements.” The commenter believed that NRC, when identifying physicists, was defining a specific position too narrowly, with delineated duties and responsibilities representing only a portion of the duties and responsibilities of physicists who are involved in radiation safety.

Response. We have not defined the term in Part 35 because it is not used in Part 35. Physicists meeting the requirements for an “authorized medical physicist” or “Radiation Safety Officer” would be recognized on the license as an AMP or RSO, respectively.

High Dose-Rate Remote Afterloader and Low-Dose Rate Remote Afterloader

Issue 1: Should there be another category of “afterloader,” such as a “non-remote” or “beta-only” afterloader?

Comment. A commenter stated that the proposed afterloader definitions don’t distinguish between the beta device that delivers more than 2 Gray/hour (Gy/h) to a target tissue and less than 0.002 Gy/h to the remainder of the body from the afterloader capable of delivering a lethal whole body dose. The proposed definitions will result in confusion for licensees and inspectors. The commenter recommended that another category such as “non-remote” or “beta-only” afterloaders be developed.

Response. We have not distinguished between beta and photon-emitting remote afterloaders in the definition. The purpose of the definition is to categorize afterloaders into different groups based on the dose rate (i.e. high, medium, or low) of a remote afterloader. Requirements for the devices are found in Subpart H. The final rule only addresses use of photon-emitting afterloaders. Use of beta-emitting afterloaders is being addressed on a case-by-case basis at this time because use of such afterloaders is relatively new and both regulators and licensees continue to identify elements of safe operation.

Issue 2: Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. The definition for an HDR was revised to state that it means a device that remotely delivers a dose rate in excess of 12 gray (1200 rads) per hour at the point or surface where the dose is prescribed rather than in excess of 2 gray (200 rads) per hour. As revised, the definition for a low dose-rate remote afterloader (LDR) states it means a device that remotely delivers a dose rate of less than or equal to 2 gray (200 rads) per hour rather than a dose rate of less than 2 gray (200 rads). This change was needed because we added a definition for medium dose-rate remote afterloaders (MDR).

Licensee

Issue 1: Should this term be defined?

Comment. A commenter asked that this term be defined.

Response. We did not define the term in Part 35 because the term is defined in 10 CFR 20.1003, "Definitions," as the holder of a license.

Wherever possible, we have tried to rely on the definitions in other parts of 10 CFR, Chapter I, that apply to medical licensees, rather than duplicate the definition in Part 35.

Management

Issue 1: Who is "management"?

Comment. A commenter asked that we clarify what we mean when we use the term "management." The commenter wanted to know whether management could be the chief executive officer or the head of one or all departments?

Response. We have clarified the regulatory text to define management as the Chief Executive Officer (CEO) or other individual having the authority to manage, direct, or administer the licensee's activities, or those persons' delegate or delegates. If the head of one or all departments is a delegate(s) of the CEO or if the individual has the authority to manage, direct, or administer the licensee's activities, that person(s) would be considered to be part of "management."

Manual Brachytherapy

Issue 1: Should the term “manual brachytherapy” be defined?

Comment. A commenter asked that we define this term because it is not a common or standard term and it is used as a subpart title.

Response. We added a definition for manual brachytherapy. As used in this part, manual brachytherapy has been defined to be a type of brachytherapy in which the radioactive sources (e.g., seeds, ribbons) are manually inserted either into the body cavities that are in close proximity to a treatment site or directly into the tissue volume.

Medical Use

Issue 1: Should the definition of the term “medical use” include the term “byproduct material”?

Comment. A commenter recommended the term “byproduct material” be deleted from the definition of the term “medical use” because the regulations use the term “byproduct material for medical use,” which is redundant. The commenter did not believe it necessary to include the term “byproduct material” in the definition of “medical use” and then to modify the term "medical use" with the phrase “byproduct material” in the regulations.

The commenter stated that deleting the term “byproduct material” from the definition “requires the least amount of correction and simplifies compatibility by Agreement States.”

Response. We recognize that there is some redundancy in using the phrase "Medical use of byproduct material." However, we believe that this level of redundancy in some requirements is not objectionable if it helps to clarify the specific requirements.

Medium dose-rate remote afteloader

Issue 1: Is there a need for definition of the term “medium dose-rate remote afterloader”?

Comment. Commenters were divided in response to our request for comment on whether the proposed rule should define the term “medium dose-rate remote afterloader.” Some commenters recommended that the term be defined because although the regulatory requirements for “high” and “medium” dose-rate afterloaders are very similar, the radiation safety precautions are different and thus these terms require different definitions. Commenters who did not support a definition for an MDR cited various reasons for their position. Some commenters believed that the regulatory requirements for HDR and MDR should be identical, therefore, there was no need to define MDRs. This position is based on the opinion that the risks to patients from high, medium, pulsed and low dose-remote afterloaders, capable of whole body irradiation, are indistinguishable. Other commenters were concerned the definition for an MDR could lead to confusion because the definition would overlap with the current definition of “high dose-rate afterloader.”

Response. We have included a definition for an MDR in the rule. This was done because the final rule contains requirements that apply to MDRs.

Mobile medical service

Issue 1: Were there any changes made in this definition between the proposed and final rule?

Response. We did not receive any public comment on this definition. However, we did change the title from “mobile service” to “mobile medical service.” This was done because we wanted to state clearly that the provisions apply only to medical use. The final rule defines “mobile medical service” to be the transportation of byproduct material and its medical use at the "client’s address" which includes the "area of use" or a "temporary jobsite." In addition, the definition of this term no longer contains the phrase “by the same licensee” because that phrase unduly limited the transportation and medical use of the byproduct material to one licensee.

Output

Issue 1. Were there any changes made in this definition between the proposed and final rule?

Response. The definition for output was revised to add reference to the exposure rate or dose rate from a brachytherapy source, remote afterloader, or a gamma stereotactic radiosurgery unit. The proposed rule only addressed the output from a teletherapy unit. This was done because various sections in Subpart H reference output from these other units.

Patient Intervention

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response: Yes. We added a definition for patient intervention. We believe this definition is needed to clearly state what we mean when we use the term in § 35.3045. Discussion of patient intervention is found in the section of this document responding to comments on § 35.3045.

Prescribed dosage

Issue 1. Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. We revised the definition for "prescribed dosage" and "prescribed dose" to allow the AU to direct the administration of a range of activity and to delete reference to the "diagnostic clinical procedures manual."

Pulsed dose-rate remote afterloader

Issue 1. Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. We restructured the definition to make it easier to read and we added a statement that the source is capable of delivering dose rates in the “high dose-rate” range, but is approximately one-tenth of the activity of typical HDR sources.

Preceptor

Issue: Should the term “preceptor” be defined?

Comment. Commenters recommended the term be defined, using a definition which distinguishes between low-dose radiopharmaceuticals (diagnostic) and “high-dose radiopharmaceuticals” (therapeutic). The former would include “persons designated as authorized physician users of “low-dose radiopharmaceuticals.” Preceptors of “high-dose radiopharmaceuticals” must be “program directors of structured educational programs in medical teaching institutions that consist of didactic training and practical experience.” Commenters believed that the “preceptor” should not be limited to someone in the medical, dental, or podiatry profession.

Commenters believe the term “preceptor” should be defined as an individual who is listed on a license as an AU, RSO, or appointed by licensee management to act in the capacity of a preceptor for the purpose of documenting that an individual has completed a structured educational program and/or practical experience. The preceptor must have demonstrated training and experience that is at least equal to the training and experience of the individuals being trained.

Response. We agree the term “preceptor” should be defined because the term previously did not appear in Part 35. The new definition defines a preceptor to be someone who provides or directs the training and experience required for an individual to become an AU, AMP, ANP, or a RSO. In addition, we agree that the preceptor must have training and experience that is at least equal to the training and experience required by the AU, AMP, ANP, or RSO, as appropriate. This is reflected in the paragraph requiring the preceptor certification in the training and experience requirements in Subpart B and D through H.

Radiation Safety Officer

Issue 1: Were there any changes made in this definition between the proposed and final rule?

Response. Yes. We restructured the definition to make it more readable. We also revised the definition for the RSO to include individuals identified as an RSO on (1) a permit issued by a Commission master material licensee or (2) a permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating AUs. This change, which was also made to the definitions of "ANP," "AMP," and "AU," accounts for the fact that an RSO may be named on a permit issued by a master material licensee. For example, in the first case identified above, if a master material licensee has issued a permit that recognizes a particular individual as an RSO, under this definition, the individual would continue to meet the requirements for an RSO under an NRC license. In the second case, if a master material licensee chooses to issue a broad scope permit to a hospital and that hospital has authorization to issue permits designating RSOs, under this definition, an RSO on the permit would also meet the requirements for an RSO under an NRC license.

Radionuclide or Radiopharmaceutical

Comment. Commenters opposed the use of terms like "radionuclide," or "radiopharmaceutical" in Part 35 because these terms are not defined as specifically containing byproduct material. They indicated that this was very important because NRC's statutory authority for regulating medical use under

Part 35 is limited to byproduct material. They recommended that the regulation should use terms that have been defined to mean “byproduct material radionuclide” or “byproduct material radiopharmaceutical.”

Response. Section 35.1, Scope, specifies that "this part contains the requirements and provisions for the medical use of byproduct material and for the issuance of specific licenses authorizing the medical use of this material." In addition, medical use is defined in § 35.2, to mean the intentional internal or external administration of byproduct material or the radiation from byproduct material to patients or human research subjects under the supervision of an AU. Therefore, we do not believe that the words "radionuclide" or "radiopharmaceutical" need to be modified by the term "byproduct material" in regulatory requirements.

The word "radiopharmaceutical" is only used in §§ 35.204 and in 35.2063. In both cases, it is clear that the requirement applies to radiopharmaceuticals containing byproduct material. The word "radionuclide" is used in §§ 35.13, 35.40, and 35.2067 and is also used in the training and experience sections in Subparts B and D through H. Again, it is clear that the requirements in §§ 35.13, 35.40, and 35.2067 apply to radionuclides containing byproduct material and for the rule text to restate the phrase "containing byproduct material" would be redundant. In the case of the training and experience sections, we have chosen to allow an individual "to take credit for" experience obtained with handling nonbyproduct and byproduct material in meeting the training and experience requirements because there is very little difference between how byproduct and nonbyproduct materials are handled.

Sealed source

Issue 1: Are epoxy vials used for testing dose calibrators “sealed sources”?

Comment. A commenter asked that we clarify whether the epoxy vials used for testing dose calibrators are “sealed sources.” The commenter stated that epoxy vials are more correctly characterized as monoliths and should not be subject to leak testing.

Response. A “sealed source” is defined in § 35.2 as “any byproduct material that is encased in a capsule designed to prevent leakage or escape of the byproduct material.” Under this definition, epoxy vials used for testing of dose calibrators are typically considered sealed sources. However, it is the licensee’s responsibility to verify that a particular manufacturer’s vial is considered by the relevant regulatory agencies to be a sealed source. This can be done by referencing the Sealed Source and Device Registry.

Teletherapy

Issue 1: Were there any changes made in this section between the proposed and final rule?

Response. We added a definition for teletherapy. This was done because we believed it is important to define such a term as used in Part 35, so that the regulated community and the regulatory agencies have a clear understanding of how we have used a term within the rule.

Therapeutic dosage and Therapeutic dose

Issue 1: Were there any changes made in this section between the proposed and final rule?

Response. We added a definition for the terms “therapeutic dosage” and “therapeutic dose.” Both terms are used in § 35.40, “Written directives.” We believe these definitions are needed to eliminate any confusion as to when a written directive is needed.

Type of use

Issue 1: Were there any changes made in this section between the proposed and final rule.

Response. We added a definition for the term “type of use.” This term replaced the term "clinical procedure" in § 35.13(a). We believe this term makes it clear that we are discussing uses in Part 35, e.g. a use of byproduct material as specified in §§ 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, and 35.1000,” rather than a “clinical procedure,” e.g., a bone scan, liver scan, whole body scan.

Unit dosage

Issue 1: Is manipulation of “unit dosages” permitted under the definition of this term?

Comment. A commenter asked to what extent the “end user” would be allowed to manipulate a “unit dosage.” The commenter indicated that the greater the manipulation of the dosage, the greater the chance of an error being made in calculating the activity.

Response. We have revised the regulatory text to make it clear that unit dosages cannot be manipulated after being initially prepared because any manipulation could change the activity in the dosage.

Issue 2: Were there any other changes made in this definition between the proposed and final rule?

Response. Yes. We revised the definition to stipulate that unit dosages must be prepared for medical use for administration as a single dosage to a patient or human research subject without any further manipulation of the dosage after it is initially prepared. This revision acknowledges that preparation of a unit dosage is not limited to a manufacturer or preparer licensed under § 32.72 or equivalent Agreement State requirement. It also highlights that a unit dosage is intended for administration to patient or human research subject without any further manipulation.

Written directive

Issue 1: Does the definition of “written directive” recognize “computerized directives”?

Comment. A commenter asked that the definition of written directive be revised to recognize that many facilities are using computerized systems and are not relying on written documents.

Response. We did not revise the definition. The intent of the definition of “written directive” and the requirements in § 35.40 are to distinguish between an AU’s written versus oral direction for the administration of byproduct material, rather than between written (hard copy) and electronic direction. As used in Part 35, “written” includes information recorded in a computerized system. If a written directive is generated or stored in a computerized system, the licensee must have a method of authenticating the AU’s signature. Reference § 35.5 for additional information on maintenance of records.

Section 35.5, Maintenance of records.

Issue 1: Can required records, other than originals, be authenticated?

Comment. A commenter asked how a copy or microform is authenticated by authorized personnel. The commenter indicated there is no requirement to authenticate records stored in electronic media. The commenter believed that all records should be required to be authenticated in writing when provided for legal purposes or verbally when being reviewed during an inspection.

Response. Any record required by Part 35 must be maintained in accordance with § 35.5. These records must be authenticated regardless of the storage media. The issue of authenticating records was addressed by NRC under a separate rulemaking, published in the Federal Register on May 27, 1988 (53 FR 19240). The following explanation of "authenticated," as stated in that notice, applies to all records retained under NRC's regulatory authority: "Authenticated' denotes that the data has been verified for completeness and accuracy by an authorized individual and that it is a true representation of the original data." (53 FR at 19243)

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We made an editorial change in the second sentence to replace the phrase "original, or a reproduced copy or a microform," with the phrase "original, reproduced copy, or microform."

Section 35.6, Provisions for research involving human subjects.

Issue 1: Should § 35.6 include a requirement that licensees develop, implement, and maintain procedures for evaluating when a medical procedure would be considered to be a research procedure?

Comment. We received a comment in support of the requirement as well as comments opposed to the requirement. The commenter who wrote in favor of requiring such procedures stated there are occasions when a clear definition of what constitutes research would be useful in deciding which procedures must be approved by the Institutional Review Board (IRB) or RSC.

Commenters opposed to a requirement indicated that FDA regulates research through IRB. They believed that existing regulations and guidelines provided adequate oversight of research and that decisions regarding research should be left to the individual licensee and the licensee's IRB. They noted that the IRB must follow the Federal Policy for the Protection of Human Subjects. As a result, they believed that research approved by an IRB and within the scope of the authorized inventory should be permitted. Commenters also noted that similar procedures were not required in other areas of medicine. Finally, commenters indicated that a requirement for procedures would not increase public health and safety.

Response. We do not believe it is necessary to include a separate definition of the term "research" in Part 35 because Section 102 of the Federal Policy for the Protection of Human Research Subjects defines the term "research." (Further information on this can be found in the Federal Register (56 FR 28003; June 18, 1991, see page 28013.)

We agree with the comment that the NRC should not add a requirement in Part 35 for licensees to develop, implement, and maintain procedures for evaluating when a medical procedure would be considered to be research. We believe that the issue of ensuring that all medical procedures and studies that should be subject to the policy are recognized as “research” and reviewed by an IRB, should be resolved as a matter of common policy, rather than in any separate effort by NRC. However, in reaching this conclusion, we do not believe that we must be guided by whether, for any given Commission requirement, there is a comparable requirement for other areas of medicine. The regulatory history of Part 35 shows that the Commission has operated under the assumption that Congress intended a disproportionate degree of Federal regulatory control be exercised over nuclear materials as opposed to other sources of radiation (e.g., x-rays, accelerator produced isotopes). (44 FR 31701; May 14, 1980, see page 31702).

Issue 2: Do broad scope licensees need a license amendment before conducting research?

Comment. A commenter recommended that broad scope licensees be exempted from the requirement to amend their licenses before conducting research involving human subjects using byproduct material.

Response. We believe that broad scope medical use licensees should be required to comply with § 35.6. This section is designed to protect the rights of human research subjects by requiring all licensees to obtain the informed consent of the subjects and by requiring an IRB to give prior review and approval of the research.

Issue 3: Were there any other changes made between the proposed and final rule?

Response. Yes. We restructured the section to make it easier to read. We also added an introductory paragraph to make it clear that research permitted under § 35.6 may only be performed using byproduct material that is already authorized for medical use by the license. For example, if a licensee is authorized to use byproduct material under §§ 35.100, 35.200, and 35.300, it could not conduct research using a remote afterloader. However, the same licensee could conduct research using materials authorized in §§ 35.100, 35.200, or 35.300.

We also added a new paragraph (e). This paragraph codifies the Commission's intent that § 35.6 does not relieve licensees from complying with other provisions in Part 35. In other words, as stated in the regulatory history of § 35.6, the relevant radiation safety provisions of Part 35 are applicable to research involving human subjects. For further information on this issue, you may want to refer to the December 2, 1994, Federal Register (59 FR 61767).

Section 35.8, Information collection requirements: OMB approval.

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (b) was revised to add reference to §§ 35.190, 35.394, 35.491, and 35.615 and to delete reference to §§ 35.633, 35.635, 35.3046, and to the sections in Subpart J that were deleted. These actions are considered conforming changes needed because of the changes made to the proposed regulatory text between the proposed and final rule.

Section 35.10, Implementation.

Issue 1: Should the time period for implementation of the new rule be extended?

Comment. Commenters asked that the implementation period for the new rule be extended up to 1 year from its publication to allow licensees and applicants sufficient time to adjust their budgets for any increased expenditures needed to implement the rule.

Response. We have maintained a 6-month implementation period for all sections of the rule. We believe that 6 months provides adequate time for licensee to develop and implement any changes in their radiation safety programs.

Issue 2: Should the rule provide for relief from restrictive requirements in the rule or license?

Comment. A commenter recommended that § 35.10(e) be revised because otherwise it will maintain the most restrictive requirements of either the revisions of Part 35 or the licensee's current license conditions. The commenter was concerned if a license condition cites a deleted requirement in Part 35, the license condition remains in effect unless the license is amended in order to remove the needless requirements. The need for a license amendment would diminish the projected cost saving of the rule.

Commenters also raised the issue of whether there is a "duality" of new Part 35 and existing license conditions, thus raising a concern about inspection and enforcement. Licensees will have to make significant amendments comparable to submitting a license renewal. Commenters believed that, if feasible and upon written request, licensees should be permitted to comply with the "new" Part 35 without regard to the restrictive nature of the license and without requiring a license amendment. If NRC believes that a regulation can be relaxed or eliminated without a reduction in radiation safety, the NRC should allow licensees to change their programs accordingly.

Response. We have modified the text of § 35.10 to allow for relief from the current rule and in some cases, license conditions. The following discussion explains and summarizes the changes made in this section.

Paragraph (a) requires licensees to implement the provisions in the rule six months after the final rule is published in the Federal Register.

Paragraph (b) states if a license condition exempted a licensee from a provision in the current Part 35, that license condition continues to exempt the licensee from the requirements in the corresponding provision until there is a license amendment or renewal that modifies or removes the condition. As shown in the following example, a corresponding provision may not always have the same numerical section reference. For example, if a licensee is exempted from the requirements in current § 35.57(c), Authorization for calibration and reference sources, the licensee will be exempted from the requirements in the final § 35.65(c), Authorization for calibration, transmission, and reference sources.

Paragraph (c) states that when a regulatory requirement in Part 35 differs from the requirement in an existing license condition, the requirement in Part 35 shall govern. This paragraph primarily applies to those licensees that committed to follow the procedures in Regulatory Guide 10.8, "Guide for the Preparation of Applications for Medical Use Programs." After the revised rule becomes effective, licensees will follow the requirement in Part 35 if it differs from the requirement that the licensee committed to by referencing the Regulatory Guide. For example, most licensees have committed to calibrate their dose calibrators using the procedures in Regulatory Guide 10.8, Appendix C, "Model Procedure for Calibrating Dose Calibrator." These procedures are very prescriptive. The final Part 35 only requires licensees to calibrate instruments used to measure the activity of unsealed byproduct materials in accordance with nationally recognized standards or the manufacturer's instructions. After the effective date of the final rule, a licensee must calibrate its dose calibrators in accordance with nationally recognized standards or the manufacturer's instructions.

Paragraph (d) states that the licensee shall continue to comply with any license condition that requires it to implement procedures for spot-checks on teletherapy, remote afterloaders, or gamma stereotactic units and to implement emergency procedures for photon-emitting remote afterloaders,

teletherapy units, or gamma stereotactic radiosurgery units until there is a license amendment or renewal that modifies or removes the condition. Specifically, licensees must continue to follow any emergency response and spot-check procedures for teletherapy, remote afterloaders, and gamma stereotactic radiosurgery units that were submitted to NRC in support of a licensing action because of the high radiation risk associated with this type of use of byproduct material.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. The proposed paragraphs (b) and (c) were deleted because we have decided to delete the training and experience requirements in the proposed Subpart J. Therefore, these paragraphs were no longer needed. As a result, the remaining sections were renumbered. Reference the General Training and Experience discussion in the beginning of this Section of the Federal Register Notice for more information.

Section 35.11, License Required.

Issue 1: Should the term “person” be used in lieu of “individual”?

Comment. A commenter noted that the word “person” was used in paragraph (a), while in paragraphs (b) and (c), the word “individual” was used. They recommended that the word “person” in paragraph (a) be changed to “individual.”

Response. We did not revise the regulatory text of § 35.11. The term “person” is used in § 35.11(a) because licenses are issued to “persons” as defined in 10 CFR 30.4. Section 30.4 states that a person includes not only individuals (defined in 10 CFR 20.1003 as “any human being”), but also corporations, government agencies other than the Commission, and States. Paragraphs (b) and (c) of § 35.11 use the term “individual” because the activities authorized by those sections are performed by “individuals” (under the supervision of an "authorized user" or "authorized nuclear pharmacist"), but not necessarily by all of the entities which constitute "persons."

Issue 2: Can there be transfer of sources among licensees?

Comment. A commenter indicated that changes in the health care environment have created affiliations between hospital groups which may or may not be under a single NRC license. The commenter believed that this regulation could prohibit the cost savings created by these affiliations. The commenter believed that if sources are received from a licensed distributor and handled properly, there should be some flexibility in transferring the sources between licensees.

Response. We have not revised the regulatory text in this section. However, we did revise the regulatory text of § 35.49 to address this comment. Section 35.11 references conditions of a specific license issued by the Commission or an Agreement State. This license would require the licensee to comply with all provisions of Part 35. One such provision in § 35.49 has been modified to state that a licensee may use a sealed source for medical use which is initially manufactured, labeled, packaged, and distributed in accordance with a 10 CFR Part 30 and § 32.74 (or equivalent requirements of an

Agreement State) license. For example, if two licensees are authorized to possess sealed sources for medical use, they may transfer the sources from one to the other, as long as the source was initially distributed in accordance with § 32.74.

Issue 3: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. The section was restructured to make it easier to use. Paragraphs (b) and (c) were combined into one paragraph because both paragraphs provided information on when a specific license is not needed.

Section 35.12, Application for license, amendment, or renewal.

Issue 1: Who may apply for a license?

Comment. The commenter believed that the requirements in the current § 35.12(a) are inconsistent. According to the commenter, under the current rule, any person may apply for a license for medical use not sited in a medical institution, while only a medical institution's management may apply for a license for medical use sited in a "medical institution." The commenter recommended that the NRC issue the license to a "responsible person" no matter what the license type. The commenter further recommended that the text of the rule be changed to reflect that the NRC will only accept a license application from a financially and/or legally responsible person.

Response. We did not make any changes between the proposed rule and the final rule in response to this comment. Section 35.12(a) of the final rule requires that the license application be signed by the applicant's or licensee's management, regardless of the types of use applied for or authorized. For a sole practitioner, the "management" could be the same as the AU. This paragraph clarifies that "management," by signing the application, is responsible for the license, regardless of the size of the licensee.

Issue 2: Is there a need for a separate license for medical uses covered by § 35.600?

Comment. Commenters stated that license applicants should be permitted to submit one license application covering several uses of radioactive material, as long as the activity is under both the same management and a qualified RSO. Commenters asked that we justify the inconsistent and separate licensing of a medical device such as a Cobalt-60 machine because neither the administrative nor the technical requirements of the radiation safety program are going to be unique for the cobalt-60 unit. Commenters believed that a licensee should not be assessed a separate annual fee just for a medical device. The additional cost will only place a greater burden on the health care delivery system.

Response. We agree with the commenter. We have revised the regulatory text to require only one license application for a license regardless of what medical use modalities the licensee will be performing. It will not be necessary for a licensee or applicant to file a separate application for each medical use of byproduct material, as described in §§ 35.600 or 35.1000.

Issue 3: Can licenses be combined at facilities?

Comment: Commenters believed that it would be advantageous for larger licensees that employ a full-time RSO and that have several existing licenses to unify all specific licenses into a single license. Commenters believed that the RSO should have the freedom and flexibility to manage resources to control all types of use without describing all the individual radiation safety procedures for the NRC. The RSO could appoint specialty RSO's, if needed, to manage the daily radiation safety program in specialty areas, e.g. nuclear medicine, cardiology, radiation therapy, or individual campuses. For example, universities or large hospitals with several campuses could issue sub-licenses under a unified license. The RSO could authorize individual users who qualified under the training and experience criteria, without notifying NRC. This would be appropriate for authorizing physicians for emerging technologies, as well.

Response. We agree that licensees should have the flexibility of combining licenses into one license. This will help to foster a more unified radiation protection program at the licensee's facility. Section 35.12 has been revised to allow applicants to apply for one license for all types of medical uses. For example, it is no longer necessary to have separate licenses for medical uses such as teletherapy, gamma stereotactic radiosurgery unit, or diagnostic nuclear medicine. Licensees have flexibility in structuring their radiation protection program to include speciality RSO's but the Commission holds the RSO named on the license responsible for the radiation protection program. Licensees do not have authority to issue any type of license. Under § 35.24, only licensee management can approve AUs.

Issue 4: Should licensees be required to submit operating procedures to NRC for review and approval as part of the license application?

Comment. We received comments recommending that we review operating procedures as part of the license application. We also received comments indicating that we did not need to review procedures and that licensees should have flexibility in program management.

Some commenters recommended that we should not abandon our practice of reviewing a licensee's or applicant's procedures before issuing a license. These commenters believed it is important for NRC to review procedures as part of the licensing process. This is important because licensee management, AUs, workers, and NRC staff must have a common understanding of what is in the procedures. They believed that this would avoid enforcement problems during subsequent inspections.

Commenters believed licensees should have the flexibility to change certain procedures even if the procedures had been submitted to the NRC, as long as the spirit of the rule is met. Once the procedure is incorporated into the license, the regulatory agency and the licensee know what to expect. NRC review of procedures during the license application or renewal process is a good way to see if the licensee has established procedures in compliance with NRC requirements. Other commenters asked this section be changed to include the requirement that applicants either (1) commit to adopting the model procedures contained within the current version of NUREG-1556, Volume 9, or (2) submit with the application the procedures they wish to use for review and approval by the Commission. These commenters did not believe inspectors have the time or resources during an inspection to both conduct the inspection and determine the adequacy of the licensee's procedures.

Other commenters suggested that the NRC review procedures only at the time of the initial application or when the license is periodically renewed. Procedures would not need to be submitted for license amendments. They believed that this approach would be helpful for smaller licensees that do not employ a full-time RSO and who usually rely on a consultant to write their standard operating procedures.

We also received comments that did not support NRC review of procedures. These commenters indicated that the NRC must recognize there are many acceptable procedures to accomplish a specified goal. A licensee should be able to use any one of a large number of procedures as long as the performance standard is met. No written procedures of any kind need to be submitted to the NRC for review or required as license conditions. Commenters also indicated that because the level of radioactivity involved in diagnostic medical uses of byproduct material is so low, compliance with the requirement for licensees to develop, maintain, and implement procedures provides no additional safety. Such a requirement would only increase the cost to the patients without any corresponding increase in the safety of the patient, hospital worker, or physician. Finally, commenters stated that this licensing approach should be extended to other uses outside Part 35 such as radiography (Part 34) and irradiator (Part 36) licenses.

Response. We have revised the various provisions in the rule to delete, with one exception, the requirement for licensees to develop, implement, and maintain procedures (e.g. § 35.24). We have also revised § 35.12 to state that only procedures required under §§ 35.610, 35.642, 35.643, and 35.645, as applicable, must be submitted to NRC for review as part of the license or amendment application. We agree that submittal of a licensee's operating procedures for NRC review and approval is necessary for certain higher risk medical uses such as those authorized in Subpart H, but not for low risk uses such as in diagnostic nuclear medicine. The lack of a procedure for the high risk modalities could result in situations where the public,

workers, or patients could be exposed to unnecessary radiation. Overall, the final rule reduces the amount of documentation that an applicant must submit, including operating procedures.

Issue 5: What are the information and licensing requirements for “emerging technology”?

Comment. Commenters were concerned that significant resources may be expended by companies for clinical research for “emerging technologies,” without knowing what the actual regulatory requirements will be. Commenters asked that provisions be made for protection of confidential and proprietary information which licensees are required to submit in accordance with § 35.12(d)(1). Commenters also asked whether NRC would be open to a petition for rulemaking proposing an appropriate way to license “emerging technology” such as brachytherapy?

Response. We clarified the regulatory text in § 35.12(d)(1) to make it clear that the information in paragraph (d)(1) must be submitted in addition to the information required by other parts of this section. This section was proposed because the current rule does not provide for the efficient licensing of "emerging technologies" (i.e., those medical uses that are not specifically included in Subparts D through H). This section provides a generic list of all the information needed by NRC to approve a medical use that is not specifically addressed in those Subparts. The specified information is needed because we must verify that the byproduct material will be handled safely. At this time and because of the evolving nature of “emerging technologies,” it is not possible to be more specific as to the necessary information. Applicants for “emerging technology” licenses are encouraged to consult with the NRC staff

as to the required information during the application process. Of course, licensees for these technologies would also be required to comply with all the applicable sections in Part 35 and 10 CFR Chapter I (e.g., Parts 30 and 71).

Provisions are already in place for the protection of trade secrets or privileged or confidential information. Section 2.790(b)(1) contains procedures under which any person who proposes to withhold from public disclosure trade secrets or privileged or confidential information may file an application for withholding accompanied by an affidavit.

Any “interested person” may file a petition for rulemaking pursuant to 10 CFR 2.802. During the NRC review of the petition, the NRC staff will review the interested person’s request and determine whether a rulemaking is needed to address the issue. In some cases, there are existing regulatory requirements that adequately address the petitioner’s request, in other areas, the petition’s request may result in development of a new rule or revision of an existing rule.

Although any “interested person” may file a petition for rulemaking in accordance with 10 CFR 2.802, such a petition should not be necessary for licensing “brachytherapy.” Licensing medical use involving brachytherapy is covered under the rule in Subparts F, “Manual brachytherapy” and Subpart H, “Therapeutic medical devices.” If an applicant believes that the use is not covered in either Subparts F or H, the applicant may request use under § 35.12(d) and Subpart K, “Other Medical Uses of Byproduct Material or Radiation from Byproduct Material.” Subpart K provides a means for licensing such medical use as an “emerging technology.”

Issue 6: Does a broad scope licensee need to amend its license for an emerging technology?

Comment. A commenter stated that broad scope licensees should not be required to amend their licenses simply for emerging technologies. The commenter asked this section be clarified or added to the list of exemptions for broad scope licenses in § 35.15.

Response. We agree with the commenter's recommendation. We have revised § 35.15 to relieve a broad scope licensee from the requirement to file a request for a license or amendment for medical uses of byproduct material, as described in § 35.1000. This regulatory relief only applies if the broad scope licensee is already authorized to possess the type and form of byproduct material.

Issue 7. Were there any other changes made in this section between the proposed and final rule.

Response. Yes. Section 35.12(a) was revised to delete the phrase "of the facility." The proposed rule required that the application be signed by the management of the facility. The final rule requires that the application be signed by the applicant's or licensee's management. The addition of the words "applicant's or licensee's" is discussed under Issue 1 of this section. We deleted the phrase "of the facility" because the word "management" clearly ties the requirement to activities performed by the licensee. (Reference definition of "management" in § 35.2.)

Paragraph (d) was revised to delete the requirement to submit information on the training and experience of proposed users. This requirement was redundant of paragraph (b). Applicants are required to submit the information under paragraph (b).

Section 35.13, License amendments.

Issue 1: Why would a license amendment be necessary for a type of use not authorized in the license?

Comment. A commenter was concerned that this section implies the NRC will be regulating medical procedures through the licensing process (i.e., NRC will use license conditions to prevent the clinical use of certain isotopes). According to the commenters, physicians should not have to wait for the NRC to grant an amendment in order to practice medicine.

Response. We have not made any changes in the regulatory text as a result of this comment. Requiring a licensee to obtain a license amendment for a type of use permitted under Part 35 but not authorized on the current license is not intended to prevent the medical use of certain radionuclides. A licensee must apply for and receive an amendment for such a type of use because it may change the byproduct material program and might increase the potential for radiation exposure to workers and the general public. For example, a licensee would need to amend its license if it is only authorized to use byproduct material for imaging and localization studies and it would like to use a remote afterloader. These changes in the byproduct material program are potentially significant and require a license amendment because:

(1) The NRC must be assured that the licensee has adequate training and experience and facilities before authorizing a change in the type of medical use or the amount of byproduct material used; and

(2) Such a change might also indicate a need for increased inspection frequency.

Issue 2: Should there be provision for a temporary RSO?

Comment. A commenter asked if we planned to add language to this section to codify the discussion in the Statements of Consideration for the proposed rule on § 35.13(c) (53 FR 43516; August 13, 1998) regarding using an AU to fill the RSO position if the RSO leaves with little or no warning. This commenter recommended that we add the following phrase to §35.13(c): “changes permanent Radiation Safety Officer.” Commenters recommended that we allow an ANP or AMP to function as the RSO since both these individuals would meet the qualifications of an RSO in § 35.50.

Response. We have addressed these comments by adding a provision for a "temporary RSO" in § 35.24(c). As stated in § 35.24(c) and discussed in greater detail under the Statements of Consideration for § 35.24, an AU or an individual qualified to be an RSO may function as the temporary RSO. The broader issue of who can be an RSO is discussed in greater detail in the response to comments on § 35.50. A licensee would not need to amend its license for a temporary RSO.

Issue 3: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We made minor editorial changes to the regulatory text in paragraph (b) to make the rule easier to read. For example, we started each requirement by stating to whom the requirement applies, e.g., we replaced the phrase “An authorized user who meets the requirements in . . .” with “For an authorized user, an individual who meets the requirements in . . .” We also revised paragraph (d) requiring the licensee to apply for and receive a license amendment before it receives byproduct material in excess of the amount, in a different form, or a different radionuclide than is authorized in the license. This change makes the regulatory text clearer.

Paragraph (b) was revised to add reference to §§ 35.190(a), 35.394(a), and 35.491(a) and to delete reference to the sections in Subpart J that were deleted. These actions are considered conforming changes needed for the changes made to the proposed regulatory text between the proposed and final rule. In addition, paragraphs (b)(4) and (5) were combined to make the rule easier to use.

Section 35.14, Notifications.

Issue 1: Is the purpose of notification to initiate a license amendment?

Comment. A commenter recommended the title of this section be changed to "Thirty-day Notifications for Amendments." In addition, the commenter stated that an introductory sentence should be added to the section indicating that the notifications should be made to initiate license amendments. Without this sentence, it is not clear that the purpose of the notification is to initiate an amendment.

Response. We have not changed the regulatory text. The purpose of § 35.14 is to identify when a licensee must notify NRC of changes in its program for which it need not apply for a license amendment. For example, if an AU, AMP, or ANP is certified by a specialty board recognized by NRC, the licensee may allow that individual to begin work immediately (without first seeking and obtaining a license amendment). All the licensee must do is notify the NRC within 30 days that the individual has begun working.

Issue 2: Is there a conflict between the requirements in §§ 35.13 (b)(1) and 35.14(b)(1)?

Comment. A commenter indicated this section was confusing because it was not clear whether the board certifications mentioned in § 35.14(a)(1) meant only those boards “adopted by regulation” or those certifying organizations listed in Appendix A. The commenter also believed the section conflicted with § 35.13(b)(1), which permits persons to act as an AU if they met the training and experience requirements in §§ 35.290(a), 35.292(a), 35.390(a), 35.490(a), 35.590(a), or 35.690(a) and § 35.59 and §§ 35.910, 35.920, 35.930, 35.932, 35.934, 35.940, 35.941, 35.950, 35.960 and § 35.49.

Response. Section 35.13 provides information on when a licensee must apply for a license amendment. Section 35.14 provides information on when a licensee must notify NRC of a change in its program. In order to provide some regulatory relief to licensees and to allow individuals to begin work immediately, we have structured these provisions as two parts, to address two different groups of people - those who are certified by a board recognized by NRC and those who are not certified by a board recognized by NRC. In the case of an AU, a licensee would not need to amend its license before allowing that individual to begin work if the individual is certified by a board whose certification process has been recognized by NRC. The licensee would

however need to notify us within 30 days of having allowed the individual to work as an AU. Conversely, a licensee would need to amend its license if the individual is NOT certified by a board that has been recognized by NRC.

We have deleted any references to boards by name in the final rule. In addition, Appendix A to the proposed rule was not included in the final rule. More detailed information on these changes can be found under the general discussion on training and experience at the beginning of this section of the Federal Register Notice.

Issue 3: Is it necessary to name AMP on a license?

Comment. A commenter recommended NRC need only allow individuals who meet the training and experience requirements for an AMP to function as an AMP.

Response. We believe the requirements for naming an AMP and AU, in the license, should be the same. In order to be considered an AMP, the individual must meet the training and experience qualifications in § 35.51. If the individual is certified by a board whose certification process has been recognized by NRC, the licensee may allow that individual to begin work immediately and notify us within 30 days that the individual has begun work. If the individual is not certified by a board whose certification process has been recognized by NRC, the licensee must apply for and obtain an amendment of its license before it allows that individual to begin work as an AMP.

Issue 4: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (b)(4) was revised to state that the licensee must notify NRC when it deletes or otherwise changes the areas where byproduct material is used in accordance with §§ 35.100 and 35.200. This change was made to clarify the regulatory text.

Section 35.19, Specific exemptions.

Issue: Shouldn't this section provide an exemption for diagnostic nuclear medicine?

Comment. Some commenters believed that essentially all diagnostic nuclear medicine procedures should be exempted from regulation because they would not endanger life or property or the common defense or security and are otherwise in the public interest.

Response. We did not make any changes in this section. Section 35.19 recognizes that an applicant for a license or licensee filing an amendment request may seek to be exempted from a specific requirement in this part (50 FR 30616, July 26, 1985, see page 30624). However, this provision does not provide the basis for a "blanket" exemption of an entire category of medical use such as "diagnostic nuclear medicine procedures" from Part 35. Nevertheless, and consistent with making Part 35 more risk-informed, we have decreased the regulatory burden on licensees administering or preparing byproduct material for most diagnostic uses by decreasing the requirements imposed on them in Part 35.

SUBPART B - General Administrative Requirements

Section 35.20, ALARA program.

Issue1:. Should the current Part 35 requirements related to ALARA programs be deleted?

Comment. A commenter supported the deletion of the current Part 35 requirements related to the ALARA program. However, another commenter believed that the requirements in Part 35 related to the ALARA program should be retained. This commenter stated that keeping this regulation in Part 35 is appropriate because Part 20 regulations are not specific enough.

Response. We have deleted § 35.20, which includes prescriptive requirements related to the ALARA program, in its entirety from the revised Part 35. Medical use licensees will continue to be required to comply with § 20.1101 that includes a requirement to implement an ALARA program designed to keep doses as low as reasonably achievable. We believe that deletion of the prescriptive ALARA requirements that are in the current § 35.20 will provide flexibility to licensees in developing and implementing their ALARA programs.

Section 35.24, Authority and responsibilities for the radiation protection program.

Issue 1: Can licensee management delegate its responsibility to approve individuals before allowing them to work as an AU, ANP, or AMP?

Comment. Several commenters said that mandating that licensee management approve individuals before allowing them to work as AUs, ANPs, or AMPs is excessive. Normally, management does not approve other individuals to work in non-NRC licensed areas. The approval to work generally comes from the department chief or the hospital credentialing committee. Therefore, the commenters suggested inserting “or management designee” after “management” in paragraph (a)(2) of this section to allow management to delegate the responsibility for approving individuals to either a responsible individual in the department or the hospital credentialing committee.

Response. In the current Part 35, the RSC has the responsibility to approve AUs, ANPs, and teletherapy physicists before allowing them to work. In the new § 35.24(a)(2), licensee management is given this responsibility for several reasons. First, licensee management has the ultimate responsibility for the radiation protection program in the revised rule. Second, not all licensees are required to have an RSC. Therefore, giving licensee management the responsibility for approval of individuals makes the requirement uniform for all medical licensees, i.e., the authority for approving individuals is not dependent on whether or not a licensee has an RSC.

As defined in § 35.2, management means the chief executive officer or other individual having the authority to manage, direct, or administer the licensee’s activities, *or those persons’ delegate or delegates*. Thus, licensee management could delegate the responsibility for approving individuals before allowing them to work.

Issue 2: Is there a need for a requirement for the RSO to acknowledge responsibility for implementing the radiation protection program in writing?

Comment. We received comments in response to the Commission's question as to whether a requirement for the RSO to acknowledge in writing responsibility for implementing the radiation protection program would impact the licensee's effectiveness in carrying out its radiation protection program. These comments both agreed and disagreed with the requirement in paragraph (b) of this section that an RSO agree in writing to be responsible for implementing the radiation protection program. One commenter supported this requirement, especially in cases where the RSO position is assigned to a junior medical staff member who has significantly more pressing obligations. Another commenter supported the requirement because it enhances the visibility of the RSO position. Several commenters noted that National Council on Radiation Protection and Measurements (NCRP) Report No. 127, Operational Radiation Safety Program (1998), Section 3 on Organization and Administration, includes recommendations for the RSO's responsibilities for the radiation safety program.

Other commenters questioned why the RSO should be required to sign off on his or her duties when the AU, AMP, and ANP are not required to do so. One commenter said that a written agreement seems more appropriate between management and the AUs, or between the AUs and NRC. Increasing the responsibilities of the AUs would provide more incentive for them to become familiar with the details of the radiation safety aspects of the licensed activities. Another suggestion was that there be a requirement for the licensee and AUs to commit in writing to follow the radiation protection program, instructions, and procedures as formalized/approved by the RSO.

Other commenters questioned why there needs to be a paper trail of the RSO's agreement. They questioned whether there is a concern that management may assign the RSO duties to someone who is unaware of their responsibilities or is there a concern because unqualified, uncommitted RSOs have been named in the past. A commenter believes that if an individual agrees to assume the RSO's duties and his or her name is on the license as the RSO, a written statement from the RSO is redundant and unnecessary. Instead, the Commission should require that the individual appointed to be the RSO sign the license amendment naming him or her as RSO, which would not only provide documentation of their acceptance of the RSO duties, but would also provide the licensing staff with a copy of the RSO's signature for future reference.

Another commenter was concerned that the written agreement seems to be more of a legal, contractual matter than it is a radiation safety matter, and it could be later used by management against the RSO.

Response. After reviewing and evaluating the public comments, we have retained the requirement for the RSO to acknowledge in writing responsibility for implementing the radiation protection program. We believe that future confusion over the responsibilities for the radiation protection program can be prevented by having a clear, written agreement between licensee management and the RSO. The RSO is given explicit responsibility in the rule for implementing the radiation protection program. Therefore, we believe it is more appropriate for that individual, rather than the AU, ANP, or AMP, to agree to that responsibility in writing.

Issue 3: Why does the rule increase management oversight of, and consequently limit the RSO's authority over, the radiation safety program?

Comment. Commenters believe that the proposed rule is very prescriptive about the relationship between the RSO and licensee management. The rule implies that licensee management gives the responsibility for maintaining the radiation safety program to the RSO, but does not allow the RSO the authority needed to manage the program. No other radiation protection program in 10 CFR Chapter I has as much management oversight as the medical use program. The NRC should also stipulate that the RSO report directly to *senior* management.

Response. The requirements in paragraphs (d) and (f) of § 35.24 that are associated with the RSO's authority are in the current § 35.23. The revised rule retains all of the RSO's current authority, plus provides the RSO with additional authority to stop unsafe operations. We did not address whether there is the same level of management oversight of other NRC licensees' radiation protection programs because that issue is beyond the scope of this rulemaking. We believe that the requirements for both the RSO's authority and for management oversight are risk-informed and, therefore, appropriate for the risk associated with the medical use of byproduct material.

Issue 4: Should there be a provision for a temporary RSO?

Comment. As noted in Issue 2 under § 35.13, License amendments, a commenter asked if we planned to add regulatory text to allow a licensee to use an AU to fill the RSO position when the RSO leaves a facility with little or no advance warning. Commenters also recommended that we allow an ANP to function as the RSO if these individuals met the qualifications for an RSO in § 35.50.

Response. We added a new provision to paragraph (c) that allows a licensee to have a temporary RSO for up to 60 days a year if the licensee meets the requirements for RSOs in paragraphs (b), (e), (g), and (h) of this section. This new provision was added so that licensees can appoint someone in a timely manner to fulfill the duties and responsibilities of the RSO, following the sudden departure of the permanent RSO named on the license. We also added a new paragraph (d) that allows a licensee to simultaneously appoint more than one temporary RSO, if needed, to ensure that the licensee has an individual that is qualified to be an RSO for each of the different types and uses of byproduct material permitted by the license. Even though we have added a provision for a temporary RSO, a licensee is expected to fill the position of permanent RSO as soon as possible.

Issue 5: Would the proposed deletion of the requirement for an RSC impact the licensee's effectiveness in carrying out its radiation protection program?

Comment. We received a substantial number of comments on whether the proposed deletion of the RSC would impact the licensee's effectiveness in carrying out its radiation protection program. The majority of the comments supported retaining the current requirement for an RSC at medical institutions because the RSC is a valuable resource in this case. The decision to eliminate the RSC could be detrimental to the institution's radiation safety program, especially with the proposed reduction in the training and experience hours for some AUs. Commenters noted that, in a medical institution, the RSC provides a valuable forum with expertise from all aspects of the licensee's medical use operations. The RSC performs many functions, such as developing and mandating the implementation of radiation protection policies and procedures, peer reviewing the radiation safety aspects of research protocols, and responding to enforcement or infractions of radiation safety practices. In addition, it provides the RSO support,

authority, and access to management. It is incorrect to assume that other hospital committees will encompass the area of radiation safety compliance. An accountable RSC, and documentation of its activities, will assure that decisions are made in the interest of radiation safety and regulatory compliance.

Several commenters noted that NCRP Report No. 127, Operational Radiation Safety Program, clearly supports the RSC, especially in the formulation of policies, review and audit of program effectiveness, and guidance of the RSO.

Other commenters supported retaining the requirement for an RSC, but not specifically tying the requirement to medical institution licensees. One recommendation was to retain the RSC for complex, multiple discipline, multi-department, and multi-use licensees. Another recommendation was for eliminating the requirement for small operations authorized under §§35.100 and 35.500, and possibly under §35.200, but making the requirement mandatory for activities under §§35.300, 35.400, and 35.600 and for larger operations involving imaging. Other recommendations included modifying the definition of medical institution to only include those facilities that perform more than one radioactive material modality; and requiring an RSC for facilities with inpatients. Commenters also said that any requirement for facilities with multiple modalities should be qualified by “within the same speciality” because there is no benefit to having physicians of completely separate modalities communicating regularly.

Some commenters supported deletion of the RSC. According to one commenter, there is no evidence that the absence of an RSC jeopardizes public and occupational health and safety. Another commenter noted that, in some cases, other Federal agencies, such as the FDA, have committee requirements that meet radiation safety objectives. Also, facilities comply with Occupational Health and Safety Administration or Environmental Protection

Agency regulations without a requirement for a committee. Therefore, deletion of the RSC would not reduce the effectiveness of the program, but would allow the licensee flexibility in meeting radiation safety objectives and in organizing its operations in the most efficient manner. However, another commenter said that removing the RSC may increase the burden on licensees, especially in conjunction with not requiring procedures to be submitted for review by licensing staff.

Another commenter suggested that rather than eliminating the entire requirement for an RSC, it might be more appropriate to reduce the more prescriptive requirements, such as the meeting, quorum, recordkeeping, and membership requirements.

Response. Based on public comment, we have retained the current requirement, with modifications, for certain medical licensees to have an RSC to oversee all the uses of byproduct material permitted by the license. In the final rule, only licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F, and H, or two or more types of therapy units under Subpart H are required to establish an RSC. Examples of such licensees are those authorized to use therapeutic quantities of unsealed byproduct material (§ 35.300) and manual brachytherapy (§ 35.400), or manual brachytherapy (§ 35.400) and LDR (§ 35.600), or teletherapy (§ 35.600) and gamma stereotactic radiosurgery (§35.600). An example where an RSC would not be required would be a licensee authorized for use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required (§ 35.100) and for use of unsealed byproduct material for imaging and localization studies for which a written directive is not required (§ 35.200). However, we believe that, based on public comments, many other medical use licensees will also continue to use an

RSC to oversee the use of byproduct material. Licensees should note that the requirement for an RSC is no longer tied to medical institutions, which means that it now also applies to “free-standing clinics.”

We have deleted most of the prescriptive list of administrative requirements and committee tasks that are specified in the current rule. For example, the final rule does not include specific requirements for the frequency of meetings, the content of the meeting minutes, or the tasks that the RSC must perform to oversee the use of licensed material. However, based on public comment, we have specified the membership of the committee, as discussed in Issue 6.

Issue 6: If an RSC is required, who should be members of the committee?

Comment. The Commission asked whether the regulatory text should explicitly require that the RSO be a member of the committee if a requirement for a committee to oversee the radiation safety program was included in the final rule. Several commenters said that the membership of the RSC is best left to the licensee. While most licensees would make their RSO a member, there is no obvious reason to require this action. Some commenters said that the RSO should be allowed to decide the committee membership and then submit the specialties to the NRC.

Most commenters agreed that both the RSO and a representative of the licensee’s upper management should be explicitly named as members. Commenters also recommended that representatives of the different users and the nursing staff be on the committee, if the facility is licensed for inpatient

therapies. While the RSO is responsible for implementing the radiation safety program, a successful committee requires both management backing and resources, and user support.

Response. As discussed in Issue 4, the final rule includes a requirement for certain medical licensees to have an RSC. We essentially agree with the commenters' recommendations for the membership of the RSC. We have included a requirement in the final rule that the membership of the RSC must include an AU for each type of use authorized by the license, the RSO, a representative of the nursing service, a representative of management, and other members the licensee considers appropriate.

Issue 7: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (b) was revised to delete the phrase "in the daily operation of the licensee's radiation protection program." This phrase did not add anything to the requirement and was awkwardly worded.

Section 35.26, Radiation protection program changes.

Issue 1: What is meant by changes in a licensee's radiation protection program that "do not reduce radiation safety?"

Comment. Several commenters said that the provision in the proposed § 35.26(a)(2), that radiation protection program changes can be made if the revisions “do not reduce radiation safety,” was ambiguous and subjective and would invite second-guessing by NRC inspectors. There should be objective measures for acceptable changes, such as changes that do not result in a licensee exceeding the limits in Part 20 or only changes that comply with all applicable regulations and license conditions.

Response. We intended for this provision to provide licensees with as much flexibility as possible in making changes in their radiation protection program, without seeking Commission approval. However, in response to comments that the proposed wording was not clear, we revised the rule to allow licensees to make revisions in their radiation protection program that are “ in compliance with the regulations and the license.”

Issue 2: Why is there a requirement to instruct individuals in changes in the radiation protection program?

Comment. Commenters said that the requirement to instruct individuals on changes in the radiation protection program should be removed. This requirement only adds work for licensees, with no resultant increase in safety, and is not consistent with the Commission’s philosophy of risk-based regulations.

Response. This requirement has been retained in the final rule because we believe that it is important to instruct individuals in radiation protection program changes before they are implemented so that individuals have a clear understanding of those changes in the radiation protection program that

may affect them. This instruction may be provided in writing or orally and may be conducted on either an informal or formal basis. For example, the instruction could be provided at an informal staff meeting.

Section 35.27, Supervision.

Issue 1: Why does this section include requirements for supervising individuals?

Comment. Commenters had a number of concerns about the requirements for supervising individuals in this section. One concern was that there is no requirement for a licensee to notify the NRC that it operates in the manner permitted by this section, i.e. a licensee does not have to inform NRC when it allows supervised individuals to use byproduct material. Therefore, this section is not consistent with other sections in the regulations that only allow licensees to conduct activities that are permitted by their licenses. This section should be deleted or changed to require licensees to apply for a supervised user program within their license applications. In addition, commenters noted that if NRC is not made aware of this type of activity, it is not conducive to inspection activities.

Another concern was that this section permits individuals, including physicians, to use byproduct material without completing the training and experience requirements for AU status. This also allows a physician who does not meet the training and experience requirements for an AU to perform the duties of the AU without the AU being present. If the training and experience required to become an AU is necessary, the supervising AU should be

required to be present, for example, during the administration and reading of films and the supervised physician should be required to attain licensure in a specified period of time.

Another commenter also said that this section should be deleted, but said that if the section is retained it should be revised to meet minimal Accreditation Council for Graduate Medical Education teaching requirements for physicians. Recommended changes relate to whether: the supervising physician and the supervised physician must be within the same city (and preferably in the same building); the number of physicians supervised at one time should be limited; the duration of a physician working under the supervision of an AU should be limited; the NRC should verify the ability of the supervising individual to teach; the supervised program should have a curriculum, goals, objectives, handouts, and testing; and the NRC should be notified that a supervised physician program is in effect.

Some commenters said that there was no need for this section because its provisions are covered in other sections of Part 35. For example, § 35.11 (b) and (c) state that a specific license is not needed for individuals receiving, possessing, transferring, preparing, and using byproduct material under the supervision of an AU or ANP. In addition, commenters said that paragraphs (a) and (b) of this section, that contain requirements for supervised individuals to follow the instructions of the supervising AU or ANP, should be deleted. If there is a failure to properly supervise, the licensee, not the supervisor, will ultimately be responsible because paragraph (d) of this section holds the licensee responsible for the acts and omissions of supervised individuals.

In addition, one commenter said that the ANP should be added to paragraph (a) because, in order to prepare material, the material must first be received, possessed, and used.

Response. Only AUs and ANPs identified on a medical use license are allowed, under Part 35, to use or prepare, respectively, byproduct material in the practice of medicine. It is frequently necessary for an AU or an ANP to delegate specific tasks associated with using or preparing byproduct material to other individuals who do not have the same training in the use or preparation of the byproduct material for medical use. This section allows for that delegation, if the individuals are properly supervised and instructed. The supervised individuals must also be required to follow the instructions of the supervisor for medical uses or for preparation of byproduct material for medical uses, the licensee's written radiation protection program procedures and written directive procedures, the license conditions, and the regulations of this chapter. These rule provisions do not require prior notification of the NRC that a licensee has delegated tasks associated with the medical use of byproduct material, e.g., package receipt, administration, and disposal of the radioactive waste. Such a requirement would be an unnecessary burden and negate the flexibility afforded to licensees in conducting their medical use programs.

The AUs and ANPs are best suited to determine what tasks supervised individuals are capable of performing and the degree of supervision that each needs. Consequently, this section does not include prescriptive requirements for training or list delegatable tasks. We believe that the requirements in this section provide the best balance between NRC's responsibility to assure the public health and safety and the licensee's responsibility for the safe use of byproduct material.

We have not added ANP to paragraph (a) of this section because this requirement is tied to § 35.11(b), which only allows individuals to receive, possess, use, or transfer material under the supervision of an AU. Section 35.11(c) permits the preparation of byproduct material for medical use under the supervision of an AU or ANP, unless prohibited by license condition.

Issue 2: Is there a need for licensees to have a policy for supervised individuals to request clarification from AUs or ANPs about procedures or instructions (proposed § 35.27(c))?

Comment. Commenters said that the requirement for licensees to have a policy for supervised individuals to request clarification if they do not understand procedures or instructions should be deleted. This requirement will not stop a misadministration which may be caused by other factors, such as human error or poor management. One commenter said that there were no data demonstrating that the failure to ask clarifying questions had resulted in a misadministration associated with either nuclear medicine or radiation oncology. If misadministration data are being used to justify the requirement, then it should not apply to diagnostic nuclear medicine because there has probably never been an instance where a diagnostic misadministration was the result of someone not understanding procedures or instructions.

Response. We agree with the recommendations in these comments and have deleted paragraph (c) of this section in the final rule.

Issue 3: What is the purpose and intent of the statement in § 35.27(d) that licensees are responsible for the acts and omissions of supervised individuals?

Comment. Commenters raised a number of concerns about the statement in paragraph (d) that licensees that permit supervised activities are responsible for the acts and omissions of supervised individuals. By explicitly stating that the licensee is responsible for the acts and omissions of supervised individuals, the implication is that the licensee is not responsible for the acts and omissions of AUs, ANPs, AMPs, or the RSO. State laws hold the supervising physicians and pharmacists responsible for the actions of all health professionals working under their supervision. Another concern was that licensees would be held responsible for willful actions and omissions of supervised individuals against established policies and/or procedures. One commenter requested a definition of the term “supervising AU.” This term appears to imply that the “AU” is responsible for supervision, while other statements in Part 35 give the authority for supervision to management. In addition, some commenters suggested that this requirement be deleted because it states the obvious and is unnecessary.

Response. This statement of the licensee’s responsibility for the acts and omissions of supervised individuals is in the current § 35.25(c). According to the Statements of Consideration for this provision, it was added to make it clear that a “licensee can not delegate *responsibility* to supervised individuals. If a supervised individual, through misunderstanding, negligence, or commission, acts contrary to the requirements of the license, the regulations, or an order, the *licensee* remains responsible” (51 FR 36932; October 16, 1986). This is still an accurate statement of the Commission’s intent in retaining a provision for supervision by an AU or ANP.

As used in this section, a “supervising AU” is simply an AU who supervises an individual using byproduct material. Even though an individual may be supervised by an AU, the licensee is ultimately responsible for the acts and omissions of supervised individuals.

Issue 4: Should “telesupervision” be allowed for Part 35 licensees?

Comment. One commenter said that the Part 35 rulemaking should address the issue of “telesupervision.” With present technology, AUs can stay in their offices and supervise medical procedures at facilities that are miles away. Due to all of the upcoming challenges of emerging technologies, the NRC should address this issue to ensure protection of public health and continued radiation safety.

Response. We have not addressed “telesupervision” during the revision of Part 35 because the need for either an AU or an individual under the supervision of an AU or a medical physicist to be present during the medical use of byproduct material is dependent on the risk associated with the particular modality. For example, the use of remote afterloader units, requires onsite supervision by individuals who are knowledgeable of the radiological hazards associated with the use of that material.

Issue 5: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. The phrase “in addition to the requirements in § 35.12” was added to both paragraphs (a)(1) and (a)(2) of this section. This addition to § 35.27 was made as a reminder to licensees that they must also comply with the requirements for supervision in § 35.12.

Paragraph (b)(1) of this section was revised to read “ individual’s involvement with byproduct material,” rather than “use of byproduct material,” because the requirement applies to individuals who prepare byproduct material for medical use under the supervision of an ANP.

Section 35.40, Written directives.

Issue 1: Why does Part 35 need to include requirements for written directives?

Comment. Several commenters agreed that the NRC should require licensees to prepare written directives, especially for those procedures that create the greatest risk to the patient from errors and those procedures that are performed by supervised individuals. However, if the written directive is really meant to be a tool for communication between the AU and other health care staff, the proposed requirements for written directives should be revised to allow licensees more flexibility in defining what information must be included in written directives. For example, an AU should be allowed to determine what information is necessary for a supervised individual to administer the byproduct material. One commenter said that the NRC should only require that a written directive be prepared before a treatment to a patient is delivered and should not define even the essential elements of the directive.

Another group of commenters opposed both the use of the term “written directive” and the need for written directives for administrations of unsealed byproduct material in medicine. Written directives, as described in the proposed rule, are “prescriptions,” which are the standard of practice in medicine and pharmacy. Prescriptions are already controlled by the State Board of Medicine and Pharmacy and the Attorney General of each state. Licensees should be allowed to create records that are consistent with other requirements for medical practice and pharmacy, rather than duplicating a “prescription.” The NRC should cite data demonstrating that the traditional method of prescribing medicine is not adequate. If the requirement for a written directive is retained, “radiopharmaceutical “ in § 35.40(a) should be qualified by adding “containing byproduct material” because no other radiopharmaceuticals fall under NRC’s jurisdiction.

Response. We believe that the requirements for written directives in this section only include what is essential to provide high confidence that the byproduct material will be administered as directed by the AU. Licensees have the flexibility to include additional information that they feel is necessary for a supervised individual to perform a procedure according to the directions of the AU.

During the Quality Management and Misadministrations rulemaking [56 FR 23360; May 21, 1991], several medical societies recommended that NRC use the term “written directive” to avoid confusion with the term “prescription” in medical and pharmacy practices. We have retained the use of the term “written directive” so that there continues to be a clear distinction between NRC’s requirements and other requirements for a “prescription.”

This section neither prevents licensees from keeping or creating other pharmacy or medical records, nor requires licensees to create records that duplicate prescriptions. Written directives are not duplicative of prescriptions. They must include information necessary to ensure that byproduct material is administered as directed by the AU. This may require different or more detailed information than is in a prescription.

Most diagnostic procedures are low risk. Therefore, licensees are not required to prepare written directives for most administrations of unsealed byproduct material. This section only requires written directives for the higher-risk administrations, such as sodium iodide I-131 in quantities greater than 30 microcuries. We also agree that the NRC's jurisdiction only covers radioactive drugs containing byproduct material so we have replaced the word "radiopharmaceutical" with "radioactive drug containing byproduct material" throughout Part 35.

Issue 2: Does a written directive need to be prepared if the AU physician performs or is present during the administration?

Comment. Several commenters questioned the need for a written directive when the AU physician performs or is present during the medical use of the byproduct material. In particular, they questioned the benefit of a physician in such a situation having to prepare a written directive if the primary purpose of written directives is to prevent misadministrations in carrying out the physician's directions. Commenters also questioned whether physicians were expected to prepare or revise written directives while simultaneously performing administrations.

Response. Written directives must be prepared in accordance with § 35.40 whether or not the AU physician performs or is present during the procedure that involves the medical use of the byproduct material. We do not expect physicians to either prepare or revise written directives while performing medical procedures. We agree with the commenter that the main reason for requiring written directives is to provide high confidence that the administration is according to the directions of the AU physician, i.e., that there is no misinterpretation of the physician's directions by another physician, pharmacist, or supervised individual.

Licensees are required to retain copies of written directives for 3 years. These copies provide documentation that the actual administrations were according to the written directives prepared before the administrations. Licensees are required to report medical events, in accordance with § 35.3045, based on the differences between the information in the written directives and the actual administrations. Therefore, if written directives, or copies of them, are not available for all administrations for which they are required (e.g., they were not prepared when physicians were present during the administrations) licensees will not be able to demonstrate compliance with either § 35.40 or § 35.3045.

Issue 3: What are the requirements for the AU's signature on written directives?

Comment. One commenter agreed that the requirement for the AU to sign the written directive should be retained. The AU checks the written directive for "appropriateness of study" before signing the document before treatment. This practice is part of the Quality Assurance Program developed by the Joint Review on Accreditation of Hospital Organizations.

Several commenters requested clarification of the requirements and policies associated with signatures on written directives. One commenter said that the requirement for preparing, signing, and dating written directives has been interpreted differently by regulators in the past. The regulations should explicitly state whether a written directive must be signed by an AU, or whether a physician under the supervision of the AU may sign the written directive. Another commenter questioned whether “electronic signatures” or “signatures on file” would be accepted on written directives.

Response. This section allows an individual under the supervision of an AU to prepare a written directive, but requires an AU to sign and date it. We require the signature of the AU on a written directive so that there is a record that the AU has reviewed and approved the information on the written directive.

Section 35.5 allows records to be maintained electronically. Therefore, AUs may use their own electronic signatures if they are signing an electronic version of a written directive. However, licensees may not use the “signature on file” notation on written directives because another individual may add it to a written directive and, therefore, it may or may not mean that the AU has reviewed and approved the written directive.

Issue 4: How soon should oral directives or oral revisions to written directives be documented in writing?

Comment. One commenter recommended that written documentation of oral directives or oral revisions to written directives should be made the next working day. The current requirement for written documentation within 48 hours is unnecessarily restrictive in some cases (e.g, over a weekend) and too lenient in other cases (e.g., during the week).

Response. In situations where a delay in order to revise a written directive or to prepare a written directive would jeopardize the patient's health, the current requirements in § 35.32(a)(1) allow for revisions of written directives to be signed by the AU within 48 hours of the oral revision and for written directives to be prepared within 24 hours of oral directives. We have made no change to the proposed requirement that documentation of both oral directives and oral revisions to written directives must be made within 48 hours. The 48-hour requirement provides more flexibility for AU physicians and also allows them to prepare any written documentation during the workweek, unless they choose to do otherwise.

Issue 5: Do the requirements for written directives allow for prescribing doses or dosages in a range?

Comment. Several commenters said that the NRC should allow AU physicians to prescribe a range of doses and dosages in a written directive. At the time that written directives are prepared, physicians are not always aware of how much radioactive drug will be taken up or how many seeds will actually be implanted. One commenter suggested that an alternative to a dose range in manual brachytherapy is not to specify a dose. This allows the

physician to make a guess at the number of seeds of a certain strength to implant and when the implant is completed to document the number of seeds actually implanted. If this is acceptable, the dosimetry could be done later.

Response. The regulations allow for AU physicians to prescribe a range of dosages, but not doses, in a written directive. Section 35.2 states that prescribed dosage means the specified activity or range of activity of unsealed byproduct material. The definition of dose in § 35.2 is dependent on the modality.

In addition, paragraph (b)(5)(i) of this section allows the physician to change the written directive after the brachytherapy sources (other than HDR) are implanted, but before completion of the procedure, to reflect what actually took place more accurately (number of sources used, total source strength, exposure time, etc.).

Issue 6: What is the basis for requiring written directives for administrations of greater than 1.11 MBq (30 microcuries) of I-131 sodium iodide?

Comment. One commenter questioned why the threshold for preparing a written directive for administrations of I-131 sodium iodide is set at greater than 1.11 GBq (30microcuriesi) when the patient release criteria in § 35.75 indicates that hundreds of millicuries in a patient do not pose undue harm. Another commenter said that the threshold for I-131 should be increased.

Response. The threshold for preparing a written directive for administrations of I-131 sodium iodide was set at 1.11 GBq (30 microcuries) because it results in a 0.5 sievert (SV) (50 rem) dose to the thyroid. The Commission, with the recommendation of the ACMUI, adopted an organ dose of 50 rem as one threshold for identifying medical events (previously “misadministrations”) during the Quality Management Program and Misadministrations rulemaking (56 FR 34104; July 25, 1991). We cited NCRP Commentary No. 7, Misadministrations of Radioactive Byproduct Material-Scientific Background (July, 1991) as stating that this threshold was considered to be well below the onset of acute, clinically detectable adverse effects that may be caused by ionizing radiation. We believe that the current threshold for preparing a written directive for I-131 sodium iodide is appropriate. Therefore, we have retained it in the final rule.

The criteria for licensees to authorize the release of patients in § 35.75 are based on the dose to the maximally exposed individual, not on the quantity of byproduct material associated with the administration to the patient. Under § 35.75, a licensee may authorize the release of any individual from its control who has been administered radioactive drugs or implants containing byproduct material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (0.5 rem).

Issue 7: Should there be any changes to the proposed list of information that is required to be included in written directives?

Comment. For any administrations of quantities greater than 1.11 GBq (30 microcuries) of sodium iodide I-131, the name of the radiopharmaceutical and the route of administration should be provided so that the requirements for written directives for all unsealed byproduct material are consistent.

Response. The requirements are not consistent because there is no need to specify either the name of radiopharmaceutical or the route of administration when sodium iodide is used. Sodium iodide is the name of the radioactive drug administered and it concentrates in the thyroid regardless of the route of administration.

Comment. For gamma stereotactic radiosurgery, the total treatment volume should be deleted because there is no way of determining it numerically.

Response. We agree with the comment, and have deleted the requirement in paragraph(b)(3) to include the total treatment volume in written directives for gamma stereotactic radiosurgery.

Comment. For teletherapy, the inclusion of the overall treatment period is not necessary. Extending the treatment time for one or two missed fractions has no impact on the overall effectiveness of the treatment.

Response. We agree that it is not necessary to include the overall treatment period in written directives for teletherapy. The requirement for overall treatment period has been deleted from paragraph (b)(4) of this section.

Comment. For HDR brachytherapy, the number of fractions and dose per fraction can be used to calculate the total dose. The requirement for total dose should be deleted so that there is no confusion if two different doses (dose per fraction and total dose) are required on the written directive.

Response. We have retained the requirement for the written directive for HDR brachytherapy to specify the total dose because the treatment time is very short compared to other types of brachytherapy.

Comment. For all other brachytherapy, several commenters suggested revision of the requirements for written directives for brachytherapy. One commenter said there was no need to require the dose to be stated if the number and source strengths were included, while another commenter said the opposite. Another commenter suggested separate requirements for permanent and temporary brachytherapy implants.

Response. Following discussion of the comments with the ACMUI, we deleted the requirement in paragraph (b)(6)(i) of this section to provide the number of sources and source strengths before implantation. We do not believe that there needs to be different requirements for permanent and temporary brachytherapy because the rule allows the AU to document certain information after implantation but before the procedure is completed.

Issue 8: Can the footnote be incorporated into the regulatory text of this section?

Comment. One commenter suggested that the footnote in this section be incorporated into the body of the rule text.

Response. We agreed with the commenter and have incorporated the footnote in its entirety into the body of the text. That footnote contains important information about preparing written directives when a patient's health could be jeopardized by any delay in providing medical care. The requirements for written documentation of an oral directive and documentation of a revision to a written directive now appear in paragraphs (a)(1) and (a)(2) of this section, respectively.

Issue 9: Were any other changes made to this section between the proposed and final rules?

Response. Yes. Paragraph (a) was revised to delete the requirement for an AU to prepare a written directive. The change recognizes the fact that written directives are often prepared by supervised individuals.

Paragraph (b)(2) of this section was revised to make it clear that the requirements in this paragraph apply to an administration of a *therapeutic* dosage of unsealed byproduct material.

The requirements for written directives for gamma stereotactic radiosurgery in paragraph (b)(3) of this section were revised to delete “the target coordinates (including gamma angle), collimator size, plug pattern, total dose for the treatment, and the total treatment volume” and to add “for each anatomically distinct treatment site - the total dose, treatment site, and number of target coordinate settings per treatment.” These changes were made to ensure that written directives for gamma stereotactic radiosurgery include the essential information.

Paragraph (b)(5) of this section was revised to make it clear that the requirements in this paragraph apply only to high-dose rate brachytherapy.

Paragraph (b)(6) of this section was revised to make it clear that the requirements in this paragraph apply to all other brachytherapy, including low-, medium-, and pulsed-dose rate remote afterloaders.

Paragraph (b)(6)(i) of this section was revised to delete the requirement for written directives for brachytherapy, before implantation, to include the number of sources and source strengths. The number of sources used is often not known until the procedure is performed.

Paragraph (b)(6)(ii) of this section was revised to include a requirement for written directives for brachytherapy, after implantation but before completion of the procedure, to document the number of sources. The number of sources used is determined during the procedure.

Section 35.41, Procedures for administrations requiring a written directive.

Issue 1: Is there a need for medical licensees to have a quality management program (QMP)?

Comment. Most of the commenters favored deletion of the QMP, as it appears in the current Part 35. The commenters felt that the provisions of the QMP were redundant with requirements that are already in place because of State pharmacy laws or with regulations codifying the routine “standard of care” in medicine. They also noted that the data collected on misadministrations does not show that QMPs have any impact. In particular, there were no data that showed patient identification is a problem. Therefore, the issue of incorrect patients being administered dosages of byproduct material has been exaggerated. Several commenters noted that regulations cannot prevent misadministrations (medical events) that are due to human error, purposeful misconduct, or failure of a supervised individual to ask questions. In addition, commenters welcomed the paperwork relief provided by deletion of some of the QMP review and reporting requirements.

Several commenters favored retention of the current QMP requirements. One commenter said that the requirement for a QMP reinforces the need for a quality improvement committee (QIC) in his institution. The QIC reviews patient records and plans, investigates, checks, and acts on issues of quality improvement. In addition, the QIC periodically reviews compliance with all aspects of the QMP, prepares a report that summarizes the findings of the review and identifies the corrective actions taken, and then submits it to the RSO. Therefore, the QMP can be important in assisting licensees to maintain good radiation protection programs. Another individual supported retention of the QMP for the following reasons: licensees have already developed QMPs that meet the regulations; the annual reviews of the QMPs evaluate the effectiveness of the therapy programs; QMP program reviews are documented and distributed to management; and they provide a mechanism to identify precursor events.

Several commenters favored a more balanced approach. They would delete some of the prescriptive QMP requirements, such as submittal of the QMP plans to NRC for review, but retain some essential requirements, such as identifying the patient and ensuring that each administration is in accordance with the written directive.

Response. We have not retained the current § 35.32, Quality management program, in the final rule. We have decided that only certain essential requirements are necessary to provide high confidence that byproduct material will be administered as directed by the AU. For any administration that requires a written directive to be prepared in accordance with § 35.40, licensees must develop, implement, and maintain written procedures to assure that, before each administration, the patient's or human research subject's identity is verified and that each administration is in accordance with the written directive. These procedures must address certain items applicable to the licensee's use of byproduct material. Beyond these requirements, the rule allows licensees the flexibility to develop procedures to meet their needs. In addition, there is no requirement for submittal of procedures to NRC for its approval as was previously required by the quality management rule.

Issue 2: What is the Commission's intent in requiring procedures for administrations requiring a written directive in § 35.41(a)?

Comment. One commenter noted that the emphasis in § 35.41 seems to be on development of the procedures, rather than on what the Commission is trying to accomplish with the procedures. Another commenter was in favor of the proposed requirements in paragraph (a) if the intent is to

permit licensees to develop their own policies and procedures to prevent patient misadministration, rather than submitting QMP programs requiring prior approval by the NRC.

Response. Our intent in requiring procedures to provide high confidence that the administration will be as directed by an AU is to avoid burdening licensees with an absolute requirement that this objective be met. We do not intend to imply that all errors in the administration of byproduct material can be prevented. For additional information refer to the regulatory history of Part 35 - 56 FR 34104; July 25, 1991, page 34115. Paragraph (a) provides licensees with some flexibility to develop procedures that are appropriate for their uses of byproduct material. We recognize that there is no “absolute” way to achieve the objectives of these procedures, e.g., verifying the patient’s or human research subject’s identity. However, NRC does require that these procedures be sufficient to provide high confidence that the patient’s or human research subject’s identity is verified. For example, just asking an individual his name may not provide high confidence that the administration was given to the correct individual. Although the procedures do not have to be submitted for NRC review and approval, licensees may be requested to make them available for review during an inspection or, following a medical event, to demonstrate that they provide the requisite high degree of confidence.

Issue 3: Does § 35.41(b) include the appropriate items that should be addressed in procedures for written directives?

Comment. Commenters differed on whether the list of items that must, at a minimum, be addressed in the written procedures was too prescriptive or too vague. Commenters noted that if a licensee has procedures that provide high confidence that the patient’s identification is verified and that the

administration is in accordance with the written directive, the procedures will have to include the appropriate information in paragraph (b). Another commenter said that not all of the items to be addressed in paragraph (b) are applicable to all of the uses of byproduct material that require a written directive.

A commenter said that the requirement in paragraph (b) to have procedures for checking the manual and computer-generated dose calculations and verifying that any computer-generated dose calculations are correctly transferred into the consoles of therapeutic medical units is vague and does not state how these should be done. Another commenter recommended adding an “/or” after the word “and” in paragraph (b)(3) to acknowledge that there could be either manual or computer-generated dose calculations.

Response. Paragraph (b) has been retained in the final rule because the Commission believes that these are the minimum items that should be addressed in procedures to provide high confidence that the patient’s identification is verified and that the administration is in accordance with the written directive. The commenter correctly noted that not all of the items in paragraph (b) are applicable to all of the uses of byproduct material that require a written directive. Therefore, paragraph (b) of this section was revised to read that the procedures “must address the following items that are applicable to the licensee’s use of byproduct material.” Paragraph (b)(2) of this section was revised to read “treatment plan, if applicable.” Both of these changes were made because all of the items listed in paragraph (b) may not be applicable to the licensee’s use of byproduct material. We revised paragraph (b)(3) to more correctly state that “both manual and/or computer-generated dose calculations” should be checked. We have not been more specific in order to provide the licensee flexibility in determining how these items should be addressed in the procedures for his or her modality or unit.

Issue 4: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. Paragraph (b)(2) of this section was revised to read “verifying that the administration is in accordance with the treatment plan.”

The phrase “the specific details” was deleted because they are not provided in the regulations.

Paragraph (b)(4) of this section was revised to read “therapeutic medical units” to correspond to the use of “units” in Subpart H.

Section 35.49, Suppliers for sealed sources or devices for medical use.

Issue 1: Are the sealed sources and devices covered by this section only supposed to be for medical uses?

Comment. As worded, one commenter said that the proposed regulation could be interpreted to mean that the sealed sources or devices manufactured, labeled, packaged, and distributed in accordance with a Part 30 and § 32.74 license may be used only for medical use. If the latter interpretation is used, Cesium-137 brachytherapy sources could not be used for shielding evaluations because this is not a medical use.

Response. The intent of the regulatory text is for licensees to use only the sealed sources and devices listed in paragraphs(a) and (b) for medical use. Other sealed sources and devices may not be used for medical use. Therefore, we have revised the regulatory text to make it clear that licensees shall use only the sealed sources and devices that are listed in paragraphs (a) and (b) of this section for medical use. This paragraph does not address what sources may be used for non-medical uses. For example, Cesium-137 brachytherapy sources may be used for shielding evaluations.

Issue 2: Are Iridium-192 seeds and ribbons considered to be sealed sources under Part 35?

Comment. A commenter indicated that iridium-192 seeds and ribbons are not “sealed” sources. Are they included in the reference to sealed sources in this section?

Response. We consider Iridium-192 seeds and ribbons to be sealed sources, as defined in § 35.2.

Issue 3: Under what circumstances can limited-scope licensees participate in medical device trials conducted under FDA-approved Investigative Device Exemptions (IDE)?

Comment. One commenter said that § 35.49, under both the current and proposed regulations, has the effect of prohibiting medical facilities with specific licenses from participating in certain manufacturer-sponsored trials of medical devices conducted under FDA-approved IDE. The commenter

recommended that § 35.49 be modified to permit the participation of limited-scope licensees in multi-site manufacturer-sponsored medical device trials conducted under FDA-approved IDEs.

Response. A specific licensee may have to amend its license before it participates in a trial with a source with an IDE in the following situations: (1) the sealed source/device design or use is changed from that documented in the Sealed Source and Device Registry (SSDR); or (2) the sealed source or device was not initially distributed by a § 32.74 supplier. There are other situations where a specific licensee may use a sealed source under an IDE and not have to amend its license. For example, when the sealed source is the same as the description in the SSDR and the sealed source was originally distributed by a § 32.74 supplier, but the FDA requires an IDE because the description of the sealed source or device differs from that originally described to the FDA.

There are additional regulatory requirements for broad scope medical licensees beyond the requirements for specific licensees. Because the broad scope licensees must comply with additional requirements to ensure the safe use of byproduct material, they have more flexibility than specific licensees in the activities that may be conducted under their licenses.

Issue 4: Should this section also address distribution by § 32.72 licensees?

Comment. One commenter questioned whether § 35.49(a) should include § 32.72 licensees as distributors of the sources.

Response. Section 32.72 applies to unsealed byproduct material distributors. Therefore, these licensees should not be included in § 35.49(a), which applies to sealed sources.

Issue 5: What are the regulations for the use and distribution of sealed sources and devices from international manufacturers?

Comment. A commenter questioned whether the rules prohibit the use of sources and devices from international manufacturers that may not have an NRC or Agreement State license to manufacture, package, and distribute these sources and devices.

Response. In order for an international manufacturer of sealed sources to distribute these sources in the United States, the manufacturer must have both a distribution license and a manufacturing license. The manufacturing license does not have to be from the US. The distribution license must be from the NRC and the sources to be distributed must go through the Sealed Source and Device Registry process.

Issue 6: What other comments were made on the proposed rule?

Comment. One commenter said that “assembled” needed to be added to § 35.49(a).

Response. As used in § 35.49(a), the word “manufactured” includes “assembly” of the sealed sources or devices.

Issue 7: Were there any other changes made in this section between the proposed and final rule?

We have revised paragraph (a) of § 35.49 to make it clear that it only applies to the *initial* manufacturing, labeling, packaging, and distribution of a sealed source or device. Subsequent distribution of the sealed sources or devices is not subject to the requirements of this paragraph, if the sources or devices are distributed to licensees that have a license to possess the source or device. However, the sources and devices cannot be altered from the description and intended use documented in the SSDR. Currently, licensees must obtain an amendment exempting them from the requirements in this section following the initial distribution of the sealed source or device.

Section 35.50, Training for Radiation Safety Officer.

Issue 1: Due to the large variation in authorized uses of byproduct material under medical licenses, what are appropriate training and experience requirements for RSOs listed on such licenses?

Comment. Commenters expressed concern that due to the large variation in the authorized uses of byproduct material under medical licenses, it is difficult to have one set of requirements for RSOs. Other commenters believe that the qualifications of the RSO should be specified in competencies that are commensurate with the scope and complexity of the radiation safety program that the RSO must implement. For example, the required experience in paragraph (b) should be tied to the specific medical uses that are authorized on the license. It is neither necessary nor practical to require a

certified health physicist to be the RSO at a small clinical program that only involves low risk modalities, such as routine nuclear medicine procedures. Alternatively, it is inappropriate for an AU to function as the RSO at a large complex program or one which may involve a broad scope license. A related comment was that certification by the ABHP does not mean that an individual is qualified to be an RSO for a medical licensee because he or she may have no experience in a medical environment.

One commenter said that the issue of acceptable qualifications for an RSO should be dealt with both through the regulations and the licensing process. A licence reviewer should be able to place additional qualifications on an RSO for a more complex byproduct material program.

Another concern was the perceived inconsistencies in the requirements. For example, board certification in paragraph (a) requires many more hours of training and experience than is listed in paragraph (b). In addition, AUs, AMPs, and ANPs are not required to obtain written certification that they have achieved a level of radiation safety knowledge sufficient to independently function as an RSO.

Response. We agree with the commenters that it is very difficult to have a single set of training and experience requirements for RSOs named on medical licenses because of the wide variation in medical uses of byproduct material. Therefore, we made several changes to the current requirements for RSOs to ensure that the RSO has adequate training for the types of uses for which he or she has RSO responsibilities. The final rule requires that an RSO must have one year of full-time radiation safety *experience involving similar types of uses* of byproduct material and a signed preceptor statement

that the individual can function as an RSO for *a medical use licensee*. If an AU, AMP, or ANP is named RSO, he or she must have the required experience with *similar types of uses* of byproduct material for which the individual has RSO responsibilities.

The training and experience of the RSO is reviewed by NRC, as part of the licensing process, to determine if the individual has the qualifications to be named as RSO for the medical uses authorized on that license. A major focus during the rulemaking has been to incorporate *all* of the requirements for medical licensees in Part 35 so that there is no need for additional requirements (via license conditions) to be placed on licensees during the licensing review.

Issue 2: What will be the status of an RSO who satisfies the current training and experience requirements, but not the new training and experience requirements, when the rule becomes effective?

Comment. One commenter said that the regulations need to accommodate older, valuable professionals with years of experience as health physicists and medical health physicists. The preceptor of such an individual may no longer be available (retired or deceased) to provide the written certification. In addition, it serves no purpose for these individuals to satisfy 200 hours of didactic training when they might well be the instructors for such programs.

Response. An individual who is currently listed on a license as an RSO will be “grandfathered” under § 35.57 when the rulemaking becomes final and will not have to satisfy the requirements in § 35.50. The individual will be able to continue as an RSO, including being named as an RSO on a new license application at a future date.

Issue 3: Can a technologist be the RSO for a medical licensee?

Comment. We received comments that both supported and opposed technologists being RSOs for a medical licensees. Some commenters think that nuclear medicine technologists are often the individuals who are most familiar with radiation safety requirements and are in the best position to carry them out. Other commenters think that technologists are more involved in clinical procedures. Therefore, technologists are not as totally oriented to radiation safety as either medical physicists or health physicists. One commenter said that certified or registered technologists would many times be better choices for RSOs than AUs. Another commenter said that one year of full-time experience as a radiation safety technologist does not provide enough opportunity to address all the issues that confront an RSO.

Response. The current Part 35 allows a radiation safety technologist to be an RSO providing that the requirements in § 35.900, Radiation safety officer, are met. We believe that any technologist can be an RSO if he or she successfully completes all of the training and experience requirements in § 35.50, Training for Radiation Safety Officer.

Issue 4: Is the requirement in § 35.50(b) for an RSO to have one year of full-time supervised radiation safety experience involving similar type(s) of use(s) of byproduct material adequate?

Comment: One commenter said that one year of full-time experience is not adequate for an RSO to cover both nuclear medicine and therapy or to cover all aspects of a broad scope licensee's radiation safety program.

Response. We have retained the requirement for one year of full-time supervised experience because that requirement is in the current § 35.900(b)(2) for radiation safety technologists and we have no evidence that the one year requirement has resulted in inadequate experience using byproduct material. This requirement is important because it must involve similar type(s) or use(s) of byproduct material for which the individual will have RSO responsibilities. It should be noted that in addition to the one year of full-time experience, the individual must also satisfy the other training and experience requirements in § 35.50 in order to be named as an RSO on a license.

Issue 5: Why is there a requirement for an RSO to obtain a preceptor statement?

Comment. Several commenters questioned the need for a preceptor statement for RSOs and noted the difficulty of obtaining such statements. One commenter said that preceptors are not common in the health physics profession. RSOs often obtain their training and experience at multiple institutions. Therefore, no single individual would be able to attest to satisfactory completion of all of the training and experience requirements. Several

commenters said that the requirement for a preceptor statement should allow for submission of such documents as resumes or college transcripts that are comparable to a preceptor statement. Another suggestion was that licensee management be able to sign the preceptor statement.

Response. We have retained the requirement for an RSO to obtain written certification that he or she has completed the training and experience requirements in paragraph (b)(1) of § 35.50. We consider such a statement to be an important component of the overall training requirements. The requirement for a preceptor statement for an ANP is in the current Part 35. We are not aware of any difficulties an ANP may have experienced in getting the required written certification. We recognize that professionals very often get their training and experience at multiple locations and there may not be one individual who can attest to completion of all of the training and experience requirements. In that case, the preceptor would be expected to look at the transcripts or possibly check some references for the individual for whom they are preceptoring in order to certify that the individual has satisfied the requirements in paragraph (b)(1) of this section. We have required that the preceptors be RSOs because they are most qualified to judge whether the individual has achieved a level of radiation safety knowledge sufficient to independently function as an RSO for medical uses of byproduct material. Licensee management may not have the same knowledge. Therefore, the licensee may not be in the best position to judge another individual's level of radiation safety knowledge and experience. We discuss the training and experience requirements in the final rule, including the preceptor, in Section III, Part I, of the Supplementary Information section of this document.

Issue 6: Should AUs, AMPs, and ANPs be RSOs?

Comment. We received a number of comments that did not agree with the provision in paragraph (c) that allows AUs, AMPs, or ANPs to be RSOs. Commenters felt that there was an inconsistency between the requirements for an RSO to complete 200 hours of didactic training, and allowing AUs, with as little as 40 hours of didactic training and 20 hours of supervised training, to be RSOs.

There were no comments that recommended that the hours required for RSOs be reduced. Rather, commenters recommended that if AUs, AMPs, and ANPs are allowed to be RSOs, they should be required to satisfy the same requirements as RSOs, including 200 hours of didactic training and supervised experience in the activities listed in paragraph (b)(1)(ii). Another suggestion was to revise the training requirements for AUs to focus on requirements associated with being an RSO. One commenter said that paragraph (c) should be deleted because training and experience requirements for RSOs should be independent of AU, AMP, and AMP status.

Another concern was that physicians typically have AU status for one type, or similar types, of medical use and may not be qualified to be the RSO for other types of medical uses. For example, a physician with AU status in nuclear medicine may be qualified to be an RSO for a licensee that only provides nuclear medicine services, but he or she should not be named as RSO for a brachytherapy device licensee or a broad scope licensee.

Several commenters said that only AUs for § 35.100 and § 35.200 uses should be allowed to be RSOs, while another commenter suggested that an AU for § 35.600 uses could be an RSO for all other uses. One commenter said that in small practices an AU should be allowed to serve as the RSO for the modality in which they have AU status, while in broad scope institutions a “dedicated” RSO is necessary. One commenter said that the regulations

should allow licensees to have more than one RSO, or the regulations should emphasize that an RSO must have training and experience in all of the types of uses for which he or she has RSO responsibilities.

Response. Following our review and evaluation of the public comments, we retained the provision in paragraph (c) that allows AUs, AMPs, and ANPs to be RSOs. The current rule allows AUs that are identified on the licensee's license to be RSOs. Retention of this provision is important for a licensee that is a sole practitioner and must be both the AU and RSO. Not allowing such a licensee to be an RSO would result in unnecessary regulatory burden on such a licensee.

The final rule also allows for AMPs and ANPs to be RSOs. This provides medical licensees even more flexibility in whom they name as their RSO. We believe that AMPs are well aware of the radiation safety issues associated with therapeutic units. In addition, we believe that the 700 hours of training and experience required for ANPs provides them with extensive knowledge of the radiation safety issues associated with the medical use of unsealed byproduct material.

Note that AUs, AMPs, and ANPs may be named as RSO *only if* they have experience with the radiation safety aspects of similar type(s) of use(s) of byproduct material for which the individual will have RSO responsibilities. For example, an AU of unsealed byproduct material cannot be named an RSO for therapeutic medical units, or vice versa, unless he or she has additional training and experience with these types of units.

Part 35 does not allow licensees to have more than one permanent RSO. The RSO named on the license must have training and experience with the radiation safety aspects of *all* types of uses of byproduct material for which the individual will have RSO responsibilities. However, § 35.24(c) in the final rule does allow licensees to name multiple *temporary* RSOs, if necessary. Refer to the discussion of the provision for temporary RSOs in § 35.24.

Issue 7: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We reworded paragraph (b)(2) of this section to more clearly state that the preceptor must certify in writing that the individual has *both* completed the structured educational program in paragraph (b)(1) *and* achieved a level of radiation safety knowledge sufficient to function independently as an RSO for a medical use licensee.

Section 35.51, Training for an authorized medical physicist.

Issue 1: What is the distinction between a physicist, health physicist, and a medical physicist in Part 35?

Comment. One commenter was concerned about the lack of differentiation between a physicist, a health physicist, and a medical physicist in the proposed rule. Health physics is radiation detection and radiation safety. Medical physics involves radiation detection and health physics, but with additional emphasis on treatment planning, therapy, and dosimetry. Under the new regulations, it appears that a solid state physicist with a masters

degree, who had never had a course in medical physics or dosimetry, could work for two years on the radiation safety aspects of the tasks listed in § 35.51(b)(1), learn to calibrate an HDR, take a test on radiation safety, and be an AMP.

Response. The term “authorized medical physicist,” as used in Part 35, is defined in

§ 35.2. We use the term AMP in the new Part 35, rather than “teletherapy physicist” as in the current Part 35, because the regulations now include requirements for photon-emitting remote afterloader units and gamma stereotactic radiosurgery units, in addition to teletherapy units. The terms “physicist” and “health physicist” are not defined in § 35.2 because they are not used in Part 35. Physicists and health physicists that meet the requirements for an AMP or RSO would be recognized on the license as an AMP or RSO, respectively.

The requirements for an AMP in this section are similar to the requirements for a teletherapy physicist in the current § 35.961, Training for teletherapy physicist. As in the current Part 35, a physicist who wants to be an AMP would have to have a master’s or doctor’s degree in physics, biophysics, radiological physics, or health physics; and complete one year of full-time training in therapeutic radiological physics and an additional year of full-time work experience under the supervision of a medical physicist at a medical institution performing the tasks in the sections listed in § 35.51(b)(1). The only new requirement is for an AMP to obtain a preceptor statement that he or she has obtained a level of competency sufficient to independently function as an AMP. We have deleted the proposed requirement for an AMP to demonstrate sufficient knowledge in radiation safety by passing an examination. We discuss the training and experience requirements in the final rule, including the deletion of the examination, in Section III, Part I, of the Supplementary Information section of this document.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. The phrase “or an equivalent training program approved by the NRC” was deleted from paragraph (b)(1) of this section because we are not going to approve training programs under the revised training and experience requirements. For a more detailed discussion of the new training and experience requirements refer to Section III, Part I, of the Supplementary Information section of this document.

Paragraph (b)(1) was revised to include a reference to the new § 35.433, Decay of strontium-90 sources for ophthalmic use. Section § 35.433 requires that only an AMP shall calculate the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments.

In addition, we reworded paragraph (b)(2) of this section to more clearly state that the preceptor must certify in writing that the individual *both* has completed the requirements in paragraph (b)(1) *and* has achieved a level of competency sufficient to function independently as an AMP.

Section 35.55, Training for an authorized nuclear pharmacist.

Issue 1: Should the current requirement for ANPs to complete 700 hours in a structured educational program be retained?

Comment. Most commenters supported the proposal to maintain the current 700 hours of training and experience because they believe that this training is necessary to assure the quality of nuclear pharmacy practitioners. One commenter recommended that the 700 hours of training and experience should specifically include 200 hours of didactic training.

Response. Throughout this rulemaking, we reviewed and discussed the training and experience requirements in Part 35 at facilitated public meetings held both during the development of the proposed rule and during the public comment period on the proposed rule. Based on these discussions and on a review of the written comments received on the proposed rule, we made no changes to the current requirements for an ANP to complete 700 hours in a structured educational program. The current requirements are considered appropriate for the duties and responsibilities of an ANP, as defined in § 35.2.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We reworded paragraph (b)(2) of this section to more clearly state that the preceptor must certify in writing that the individual *both* has completed the structured educational program in paragraph (b)(1) and has achieved a level of competency sufficient to function independently as an ANP. We also reworded this section to more correctly state that the preceptor is certifying that the individual has achieved a level of competency sufficient to function independently as an ANP, rather than to independently operate a nuclear pharmacy. The revised text is consistent with the text used in the other training and experience sections.

Section 35.57, Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

Issue 1: Why doesn't § 35.57 include a reference to § 35.55, Training for an authorized nuclear pharmacist.

Comment. One commenter noted that § 35.57(a) in the proposed rule referred to experienced RSOs, physicists, and nuclear pharmacists, but only referenced the training requirements for RSOs and physicists.

Response. We corrected § 35.57(a) to include the reference to § 35.55, Training for an authorized nuclear pharmacist.

Comment. One commenter noted that § 35.57(b) in the proposed rule referenced training requirements for AUs in Subparts C-H, but there are no training requirements for AUs in Subpart C.

Response. We corrected § 35.57(b) to delete the reference to Subpart C, which includes training requirements for RSOs, AMPs, and ANPs, but not AUs.

Section 35.59, Recentness of training.

Issue 1: How much related continuing education and experience does an individual need to have if their training and experience has not been obtained within 7 years preceding the date of the application?

Comment. A commenter questioned that if the training and experience have not been obtained within the 7 years preceding the date of application, how much related continuing education and experience would the individual need to have and would this be a case-by-case evaluation with input by the ACMUI.

Response. If the training and experience was not obtained within 7 years preceding the date of the application, the continuing education and experience requirements for an individual would be reviewed on a case-by-case basis, with input from the ACMUI, as necessary.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. The reference to Subpart J was deleted because that subpart was deleted in its entirety from Part 35. For additional information on the training and experience requirements in the final rule, including the deletion of Subpart J, refer to Section III, Part I, of the Supplementary Information section of this document.

SUBPART C - General Technical Requirements

Section 35.60, Possession, use, and calibration of instruments to measure the activity of unsealed byproduct materials.

Issue 1: Can all requirements for calibration of instruments used to measure the activity of unsealed byproduct material be combined? Is it necessary to have prescriptive calibration requirements for these instruments?

Comment. Commenters proposed that §§ 35.60 and 35.62 be combined into one section because both sections address calibration of instruments used to measure the activity of unsealed byproduct materials. They also recommended that the prescriptive calibration requirements be deleted so that licensees have the flexibility to develop a calibration program that meets their needs.

Response. We agree that §§ 35.60 and 35.62 should be combined because both sections address instrument calibration. We also agree that the prescriptive requirements should be deleted from the section. The regulatory text was revised to delete prescriptive calibration requirements. The section now requires that licensees calibrate instrumentation in accordance with nationally recognized standards (e.g., voluntary consensus standards, such as ANSI N42.13-1986 (R 1993), "Calibration and Usage of Dose Calibrator Ionization Chambers for the Assay of Radio nuclides.") or with the manufacturer's instructions. This change makes the requirements for instrument calibration more flexible, more adaptable to new technology, and more performance based.

Issue 2: Does this section apply to licensees that use brachytherapy sources?

Comment. A commenter asked that we revise the section to state that the section does not apply to use of brachytherapy sources.

Response. The title of this section has been revised to clarify that it only pertains to instruments used to measure the activity of unsealed byproduct material. The calibration of brachytherapy sources is addressed in § 35.432.

Issue 3: Should licensees that only use unit dosages be required to possess, use, and calibrate instruments to measure the activity of unsealed byproduct materials?

Comment. Some commenters agreed that the NRC should not require unit dosages to be assayed. As a result, they did not believe that it was necessary to require licensees that only use unit dosages to possess, use or calibrate instruments to measure the activity of unsealed byproduct materials. Other commenters disagreed with the proposed provision that did not require direct measurement of unit dosages prior to administration. They believed that all dosages should be assayed. Therefore, all licensees should be required to comply with this section.

Response. We have revised the regulatory text to state clearly that this section only applies to measurements that are made in accordance with § 35.63, which requires licensees to assay (measurement of radioactivity) nonunit dosages except when volumetric measurements and mathematical calculations are used.

As stated in the Statements of Consideration for the proposed rule (63 FR 43533; August 13, 1998), if a licensee administers only unit dosages from manufacturers (or preparers) and uses decay methods to determine the dosages, the licensee is not required to have a measurement instrument and, thus, is exempt from the calibration requirements of this section. However, if a licensee administers unit dosages but chooses to reassay a unit dosage, the licensee must comply with this section. If an instrument is used to measure dosages, it is extremely important that it is calibrated.

Issue 4: Is it necessary to keep a record of instrument calibrations?

Comment. Some commenters did not believe that it was necessary to keep a record of the instrument calibrations.

Response. We have retained the requirement to maintain calibration records because they are needed to document that the instruments have been calibrated. However, we have simplified the recordkeeping requirements § 35.2060 of the final rule by requiring that the licensee record the model and serial number of the instrument and calibration source, date, results, and the name of the individual who performed the calibration.

Section 35.61, Calibration of survey instruments.

Issue 1: Is this section needed in Part 35?

Comment. A commenter believed that this section should be deleted from Part 35 because survey instrument calibration is addressed in 10 CFR 20.1501.

Response. We have not deleted this section from the rule. Section 20.1501 requires that licensees calibrate survey instruments periodically, but it does not provide specific requirements for calibrations of survey instruments. Specific requirements are needed for Part 35 licensees to assure that their radiation survey instruments are properly calibrated. An accurate survey instrument is important because individuals rely on the instrument output to assess radiation levels in areas in or adjacent to nuclear medicine or radiation therapy departments where patients or the public may have access.

Issue 2: Is it necessary to require that survey instrument operability be determined with a check source?

Comment. A commenter stated that the NRC should retain the requirement in the current rule that requires licensees to check survey instrument operability with a dedicated check source. Another commenter indicated that the word “check” should be deleted in the section title because the regulatory text did not include a requirement for an instrument “check.”

Response. The requirement to check survey instrument operability with a dedicated check source was not included in the proposed or final rule because we believe that licensees should have flexibility in how they determine that instruments are operating properly. We deleted the word “check” from the title because the section does not include a requirement for an instrument “check.”

Issue 3: How often should a survey instrument be calibrated?

Comment. Commenters suggested various frequencies for instrument calibrations. Some commenters suggested that instruments be calibrated every 6 months. Others agreed with the 1-year interval in the proposed rule and still others suggested a 2-year interval.

Response: We believe that survey instruments should be calibrated before first use, annually, and following any repair that affects the calibration of the instrument. A 1-year calibration frequency is consistent with nationally recognized standards, such as ANSI (ANSI-N323A-1997).

Issue 4: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. In paragraph (a), we added the phrase "that affects the calibration." This was done to clarify that the licensee does not need to recalibrate an instrument if the repair did not affect the calibration. For example, if the licensee replaced the batteries in the instrument, it would not need to calibrate it. In paragraph (a)(2), we added the word "decade" to account for instruments with digital readouts.

Paragraph (b) was deleted from the proposed rule. We believe the licensee should have flexibility in how it documents information on the status of survey instrument calibrations. Our primary concern is that the instrument is reading accurately. Proposed paragraph (c) states that a licensee may not use a survey instrument if the difference between the indicated exposure rate and the calculated exposure rate exceeds 20 percent. Therefore, we do not believe the requirement in paragraph (b) for a licensee to attach a correction chart is needed. A statement regarding when a licensee shall consider a point calibrated is unnecessary. Because of the deletion of proposed paragraph (b), proposed paragraphs (c) and (d) have been redesignated as paragraphs (b) and (c) in the final rule.

Section 35.62, Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides.

Issue 1: Can this section be combined with § 35.60?

Comment. Commenters proposed that this section be combined with § 35.60.

Response. We agreed that §§ 35.60 and 35.62 could be combined since the Part 35 requirements for instrument calibrations are the same for all types of instruments. (Reference response to similar comments under § 35.60.)

Section 35.63, Determination of dosages of unsealed byproduct material for medical use.

Issue 1: Can this section be combined with § 35.60?

Comment. A commenter proposed that this section be combined with § 35.60.

Response. We did not combine § 35.60 with § 35.63 because these sections have different purposes. Section 35.60 contains the requirements for calibrating instruments used to determine the activity of a dosage. Section 35.63 contains the requirements for determining the activity of a dosage.

Issue 2: Should unit dosages be reassayed before administration?

Comment. Some commenters supported the proposed rule which did not require the licensee to reassay unit dosages. These commenters believed that the administered activity could be based on the activity reported by the nuclear pharmacy. Other commenters did not support the proposed rule. They believed that all dosages should be assayed by the licensee before administration.

Response. We believe a licensee should determine and record the activity of each dosage before medical use. For unit dosages, this determination must be made by a decay correction based on the activity or activity concentration. This activity or activity concentration must have been

determined by a manufacturer or preparer licensed under § 32.72 or equivalent Agreement State requirement or by a NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee (RDRC)-approved protocol or an Investigational New Drug (IND) protocol accepted by FDA. Because the unit dosages have been assayed by the Part 32 licensee or by a licensee for use in research in accordance with an RDRC-approved protocol or an IND protocol accepted by FDA, we do not believe the Part 35 licensee should be required to reassay the dosage. Licensees should note that, if a unit dosage has been changed or manipulated in any way, it is no longer considered to be a unit dosage and will need to be reassayed before it is administered.

Issue 3: Can volumetric measurements be used to determine the activity of a dosage?

Comment. Commenters asked that we clarify whether the phrase “combination of measurements and calculations” would allow a licensee to base the administered activity on the radioactivity measurement made by a manufacturer (or a preparer), with volume measurement and calculation by a licensee. Commenters also asked that we clarify whether the term “direct measurement” means that the activity of the dosage must be based on a measurement of the radioactivity.

Response. We agree that the terms “direct measurement” and “combination of measurements and calculations” in the proposed rule text needed to be clarified. In the final rule, we made two changes:

1. We replaced the term “direct measurement” by “direct measurement of radioactivity,” and

2. We added an alternate method for determining dosage by using the radioactivity measured by a manufacturer or a preparer, with volume measurement and calculation by a licensee.

Issue 4: Should the administered dosage be allowed to deviate from the prescribed dosage?

Comment. Commenters recommended that we delete the requirement in § 35.63(d) that states: “a licensee shall not use a dosage if the dosage differs from the prescribed dosage by more than 20 percent.” Many commenters believed that this was an overly prescriptive requirement. They stated that it is the AU’s responsibility to determine the proper dosage or dosage range for patients.

Response. We believe that the requirement, with some modification to address prescribed dosage ranges, should be maintained in the final rule. AUs are responsible for prescribing the dosage or dosage range. AUs may prescribe a dosage range greater than 20 percent. This range can be case specific or can be a “blanket” range that would cover all administrations of unsealed byproduct material. For example, the AU could establish a policy where all administered dosages may deviate from the prescribed dosage by plus or minus “xx” percent.

In cases where the AU has not prescribed a dosage range, we believe that the regulatory text should allow for some deviation from the prescribed dosage. Without this 20 percent “default” range, all administered dosage would need to exactly match the prescribed dosage at the time of administration. We believe that a 20 percent deviation is reasonable in consideration of current technology. We have not allowed a deviation outside of the prescribed range because, under this provision, the AU has the flexibility of establishing the acceptable range.

Issue 5: Is it necessary to perform a decay correction for long-lived Radio nuclides?

Comment. Commenters asked that the rule be modified so that licensees are not required to perform a decay correction for long-lived Radio nuclides.

Response. We do not believe that the rule should specify when, based on half life, a decay correction should be performed. We believe the rule addresses this issue by permitting a licensee to administer a dosage if the dosage activity is within 20 percent of the prescribed dosage or is within the prescribed dosage range. This requirement gives the licensee responsibility for determining when it is appropriate to perform a decay correction. In the case of a long-lived radionuclide, the licensee may make a determination that a decay correction is not needed to verify that the dosage is within 20 percent of the prescribed dosage or is within the prescribed range because of the long half life of the byproduct material.

Section 35.65, Authorization for calibration, transmission, and reference sources.

Issue 1: Are medical licensees authorized to receive calibration sources from licensees that are licensed under §§ 32.72 and 32.74?

Comment. A commenter asked that this section be revised to allow licensees to receive calibration and reference sources from licensees that are licensed under § 32.72, Manufacture, preparation, or transfer for commercial distribution of radioactive drugs containing byproduct material for medical use under Part 35 and § 35.74, Manufacture and distribution of sources or devices containing byproduct material for medical use.

Response. The rule has been modified to include the reference to § 32.72. Nuclear pharmacies, licensed under § 32.72, are permitted to acquire calibration or reference sources from manufacturers licensed under § 32.74 and then to redistribute the sources to medical use licensees. Therefore, it is appropriate for a Part 35 licensee to have the flexibility of procuring the source from a § 32.72 licensee.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We inserted the work “transmission” in the section title. This was done to clarify that licensees may receive, possess and use any transmission sources that do not exceed the quantity limits in this section.

We corrected an error in paragraphs (a) and (b). Paragraph (a) should have referred to “1.11 GBq (30 mCi)” rather than “1.11 kBq (30 mCi)” and paragraph (b) should have referred to “0.555 GBq (15 mCi)” rather than “0.555 MBq (15 mCi).” In addition, paragraph (c) was clarified. Our intent is to

allow the licensee to receive, possess, and use byproduct material with a half-life longer than 120 days provided individual amounts do not exceed the smaller of 7.4 MBq (200 microcurie) or 1000 times the quantities in Appendix B of 10 CFR Part 30.

Section 35.67, Requirements for possession of sealed sources and brachytherapy sources.

Issue 1: When are leak tests required?

Comment. Some commenters believed that leak tests should only be required if a radioactive source has been abused, misused, or retrieved after being lost. Other commenters questioned whether the rule requires leak testing of small check sources. In addition, some commenters believed that sources should be leak tested annually. Others supported semiannual annual leak testing. Finally, some commenters believed the rule should not require a licensee to leak test certain sources, such as dry radionuclides embedded in acrylic.

Response. Section 35.67(b) contains the leak test requirements for sealed sources. We believe that sealed sources should be leak tested semiannually or in accordance with the interval approved by the Commission or an Agreement State in the SSDR. A semiannual leak testing requirement is consistent with recommendations in ANSI-N542. If licensees are unsure whether a source meets the definition of a sealed source, they should reference the SSDR. This Registry may be accessed at <http://www.hsrdo.gov/nrc/ssdr/ssdrindx.htm>.

We have not included a requirement for a source to be leak tested if it has been “abused, misused, or retrieved after being lost” because the licensee is responsible for assuring that the dose limits in Part 20 are not exceeded. If the licensee suspects that a source may be leaking or could have been damaged, it should evaluate whether a survey (leak test) should be performed.

Paragraph (f) lists the sources that do not need to be leak tested. In particular § 35.67(f)(3) states sources containing 3.7MBq (100 µCi) or less of beta or gamma-emitting material or 0.37 MBq (10 µCi) or less of alpha-emitting material need not be leak tested. If a source contains less than this quantity of material, a leak test is not needed.

We believe leak tests are needed for sources such as dry radionuclide embedded in acrylic because removable contamination could exist due to:

1. Radioactivity contained at the surface of the acrylic;
2. Interaction between any chemicals or solvents that may accidentally come into contact with the acrylic;
3. Aging of the acrylic; or
4. Radiation damage to the acrylic. (Note: if the radioactivity of the acrylic source is less than the quantities in § 35.67(f)(3), leak testing would not be necessary.)

For example, a common dose calibrator source, which is embedded in cast epoxy resin matrix sometimes referred to as an "E Vial," meets the definition of a sealed source and would have to be leak tested in accordance with the requirements in this section. However, E vials containing no more than 3.7 Mbq (100 μ Ci) of a gamma emitting material are exempt from leak testing under § 35.67(f)(3).

Issue 2: When should an inventory of sealed sources and brachytherapy sources be performed?

Comment. Commenters suggested that inventories of sealed sources should be performed quarterly, others suggested semiannually, as in the proposed rule. Other commenters believed that sealed sources that are exempt from leak testing should not be subject to inventory requirements. Another commenter questioned whether extra brachytherapy seeds should be subject to inventory requirements.

Response. Sealed source inventories should be performed semiannually. A review of events where sources have been lost or stolen in the past 10 years indicated that quarterly inventories would not have had a significant impact on preventing the incidents. The change from a quarterly frequency to a semiannual frequency would reduce unnecessary regulatory burden and radiation exposure for individuals performing the inventories.

We believe sealed sources that are not required to be leak tested should be inventoried because handling sources listed in paragraph (f) would not necessarily be considered low risk. For the same reason, extra brachytherapy sources should be inventoried. If one of these sources were lost and were picked up by an individual, the radiation dose received by the individual may exceed Part 20 limits.

Issue 3: What is the appropriate time period for reporting a leaking source?

Comment. A commenter suggested that the time period for reporting a leaking source should be changed from “within 5 days” to “within 15 days.”

Response. We have not changed the time period for reporting a leaking source. We continue to believe that it is important to inform NRC promptly when a licensee discovers that a source is leaking.

Issue 4: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (a) was revised to delete the requirement to maintain a copy of the radiation safety and handling instructions supplied by the manufacturer for the duration of source use because it was overly prescriptive. We believe that this change makes the regulation more performance based. However, deletion of the requirement does not prohibit the licensee from maintaining the instructions.

Paragraphs (d) through (f) were revised to replace the term “leakage test” with the phrase “leak test.” This change was made to reflect common use of the term “leak test.”

Paragraph (f) was revised to indicate clearly that a stored source is exempt from the leak testing requirements in this section, regardless of the length of time that it has been in storage. The current rule does not contain a requirement to leak test stored sources after 10 years. The provision for leak testing after 10 years was added to the proposed rule because, at that time, we believed that leak testing was appropriate given the time of storage and the potential for contamination. At this time, we do not think this prescriptive requirement is warranted because the licensee must test each stored source for leakage before any use or transfer unless it has been leak tested within 6 months before the date of use or transfer.

Section 35.69, Labeling of vials and syringes.

Issue 1: Can this section be deleted?

Comment. Commenters suggested that this section should be deleted because appropriate labeling is the standard of medical and pharmacy practice and is adequately regulated by the FDA, the State Boards of medicine and pharmacy, and the US Pharmacopeia. Syringe shields can be used to maintain exposures ALARA. Under certain circumstances, syringe shields can be hazardous to patients because they could obscure subtle visualization of the syringe content.

Response. We do not think this section should be deleted in its entirety and we do not believe that this requirement duplicates the requirements of the FDA, State Boards of Medicine and Pharmacy, and the U.S. Pharmacopeia. The labeling requirement is needed to provide information to physicians

or technologists that indicates the contents of the syringe. In addition, the label is needed to warn other workers that the syringe contains byproduct material. We have however deleted the requirement for the licensee to develop, implement, and maintain written procedures for labeling each syringe, syringe shield, or vial shield that contains a radiopharmaceutical and for shielding vials and syringes. We also deleted the requirement to provide individuals with instructions on these procedures. Both requirements have been deleted because we believe the rule should focus on how the vial or syringe should be labeled rather than on procedures.

Syringe or vial shields can be used to maintain exposures ALARA. However, we believe licensees should have flexibility to determine whether syringe or vial shields should be used. Thus, we have deleted the requirements to shield the syringe or vial. However, deletion of the requirement does not prohibit the licensee from using syringe or vial shields. Note: When syringe shields or vial shields are used by a licensee, the final rule requires the licensee to label the shields, if the label on the syringe or vial is not visible.

Section 35.70, Surveys for ambient radiation exposure rate.

Issue 1: Is this section needed?

Comment. Some commenters did not believe this section was needed because it was up to the licensee, through the RSO, to ensure radiation safety. Some commenters agreed that surveys should only be required when byproduct material requiring a written directive is used. Other commenters believed that the rule should require surveys of all areas where byproduct material is used.

Response. This section is needed to ensure that a radiation survey is conducted in areas where unsealed byproduct material that requires a written directive was prepared or administered. We believe that a radiation survey, at the end of each day, should be required in Part 35 because patients and other individuals could be present near a nuclear medicine or radiation therapy department. Without surveying ambient radiation levels, it is possible for patients or other individuals to receive unnecessary or excessive radiation exposures.

In order to make the rule more risk-informed, we do not believe all areas need to be surveyed. However, licensees must be prepared to show compliance with the public and occupational dose limits specified in Part 20 of this chapter.

Issue 2: When should surveys be performed?

Comment. Some commenters believed that surveys should be performed after preparation or administration of byproduct material rather than at the end of the day. Some opposed removing the existing requirements to survey areas where radiopharmaceuticals or waste is stored and to survey for

removable contamination. Finally, one commenter asked that we clarify whether the requirement for surveys in paragraph (b) applies only to patients' rooms or whether it also applies to the area where the patient's dosage was prepared.

Response. The general survey requirements are in Part 20. In addition to these requirements, we believe that medical use licensees should be required to perform radiation surveys in areas where unsealed byproduct material requiring a written directive is used or administered once a day, at the end of the day. A medical use licensee, such as a hospital, prepares and administers byproduct material to multiple patients or human research subjects throughout the day. If a survey were required after each preparation or administration of byproduct material, there would be a significant increase in the licensee's burden to comply with this requirement without an associated safety benefit. We believe that the survey at the end of each day of use is sufficient to detect elevated radiation levels. If elevated levels are detected, corrective action, if warranted, could be taken. However, licensees always have the flexibility of performing more frequent surveys.

We do not believe a requirement for weekly surveys for removable contamination is needed because licensees are required to show compliance with public and occupational dose limits in Part 20 of this chapter. In addition, the licensee will need to be able to show compliance with 10 CFR Part 20, Subpart F, Surveys and Monitoring.

We have clarified paragraph (b) to indicate that the survey requirement only applies to areas where patients or human research subjects are confined when they cannot be released under § 35.75.

35.75, Release of individuals containing radiopharmaceuticals or implants.

Issue 1: Should any changes be made to the release criteria?

Comment. Some commenters supported the dose-based release criteria in the proposed rule, while others asked that the criteria be revised.

Those commenters that supported the 5 mSv (500 mrem) release limit believed that § 35.75 provided regulatory relief to the medical profession without an associated increase in radiation risk to the public. These commenters recognized that one of the major obstacles to allowing the release of individuals in accordance with § 35.75 is a possible increase in radiation alarms at landfills. However, they believed the issue of landfill alarms should be addressed in other ways, such as raising the threshold for the alarms to a “more practical” level, rather than revising the release criteria in § 35.75. Commenters also indicated that several studies had been conducted that indicated that radiation exposures to family members from released patients were less than the 5mSv (500 mem.) limit. As a result, they asked that NRC reevaluate information provided in the guidance associated with this requirement.

Other commenters asked that the release criteria be revised because they believed that the criteria were based solely on economics and not on radiation risk. They were also concerned that household waste from an individual who had been released from the hospital could be contaminated and could trigger radiation alarms at landfills. This situation would affect State radiation protection programs because the States would have to investigate incidents in which the alarms had been activated.

Response. We do not believe that any changes are needed to this section as a result of the public comments. We acknowledge that some States have reported an increase in the number of alarms at landfills. However, we have no documentation indicating that the exposure rates to the maximally exposed individuals have exceeded the dose limit in § 35.75. The NRC does not have regulatory jurisdiction over the landfill operators nor over the alarm set points for radiation detectors at landfills. We do, however, encourage continued communication between regulatory bodies and landfill operators to resolve this issue.

We believe that the release criteria provide licensees with needed flexibility in program management. A dose limit of 5 Mev (500 mrem) to individuals knowingly exposed while voluntarily helping in the care, support, and comfort of patients provides adequate protection of these individuals. The dose limit of 5 Mev (500 mrem) to individuals comforting patients is consistent with the recommendations of the National Council on Radiation Protection and Measurements (NCRP) and the International Commission on Radiological Protection (ICRP). If you would like further information on the background to this section, you should reference 62 FR 4120; January 29, 1997.

Finally, we recognize that the values presented in NUREG 1556, Volume 9, for release are based on some conservative values. The licensee may use case-specific information in place of the values used in the guidance document.

Issue 2: What other changes were made in this section between the proposed and final rule?

Response. Paragraph (b) was revised to replace the term “breast feeding infant” with the term “nursing infant.” This was done to maintain consistency within Part 35. Paragraph (d) was revised to state that records of the instructions provided to breast-feeding females should be made in accordance with § 35.2075(b) rather and § 35.2075(c). This change was needed because of a change in the rule text for § 35.2075 (see § 35.2075 for a more detailed discussion).

Section 35.80, Provision of mobile service.

Issue 1: Should mobile service licensees be allowed to operate under reciprocity in other regulatory jurisdictions?

Comment. Commenters indicated that mobile medical services are currently operating under reciprocity in some States. Some Agreement States indicated they do not allow medical licensees to operate under reciprocity, while other Agreement States said they permit mobile medical services to come to their State under reciprocity.

Response. Agreement States have the flexibility of determining whether they will issue mobile medical licenses and whether they will allow NRC or other State licensees to operate in their State under reciprocity. Under reciprocity, an Agreement State may allow a specific licensee from another Agreement State (or the NRC) to work within the Agreement State without requiring the licensee to obtain a license in that State. Similarly, under reciprocity, a specific licensee from an Agreement State may work in NRC jurisdictions, provided the requirements in 10 CFR § 150.20, Recognition of

Agreement State Licensees, are met. Specifically, NRC allows Agreement State mobile medical service licensees to operate in areas under NRC jurisdiction provided they comply with all the requirements in § 150.20, including the submittal of the information required in that section.

Issue 2: Should NRC allow byproduct material to be delivered to a client's address of use?

Comment. A commenter recommended that the NRC permit byproduct material to be delivered to the client's address.

Response. Byproduct material may only be transferred to an NRC or Agreement State licensee because the licensee is responsible for the safe handling of the material. In almost all cases, the client is neither an NRC nor an Agreement State licensee. Therefore, the material must only be transferred to the licensed mobile medical service. Mobile medical service licensees may have byproduct material delivered to them at the client's address, provided the byproduct material is secured against unauthorized removal (§§ 20.1801 and 20.1802).

Issue 3: What checks should be performed on instruments used to measure the activity of unsealed byproduct material at a client's address?

Comment. A commenter recommended that the check for instrument operation at the client's address be limited to a constancy check.

Response. Licensees must check the operation of instruments used to measure the activity of unsealed byproduct materials to assure that the instrument is functioning properly. The regulatory text was revised to require that licensees check instruments used to measure the activity of unsealed byproduct material for constancy before medical use at each client's address or on each day of use, whichever is more frequent. In the case of a mobile medical service, we believe that a constancy check must be performed to assure that the instrument is functioning properly. The need for additional testing on the instruments is dictated by how the licensee addresses compliance with § 35.60.

Issue 4: Is it necessary to check a survey instrument with a dedicated check source?

Comment. A commenter recommended that the requirement to check the survey instrument with a dedicated check source be deleted because this check was no longer included in § 35.61.

Response. We do not believe that the requirement to check survey instruments with a dedicated check source should be deleted from § 35.80. While we have deleted the requirement from § 35.61, we believe it is needed in § 35.80 because there is a greater likelihood that a survey instrument in a mobile unit may become damaged or uncalibrated as a result of extensive movement.

Issue 5: Do mobile medical service licensees need to collect contaminated waste generated by patients after administration of the byproduct material?

Comment. A commenter asked that NRC clarify whether mobile medical service licensees need to return to the client's address to collect contaminated waste generated by patients after the administration of the byproduct material.

Response. The mobile medical service licensee does not need to return to the client's address to collect contaminated waste generated by the patient after the administration because the waste is no longer considered under the licensee's control since the patient would have been released from licensee control under § 35.75.

Issue 6: What other changes were made between the proposed and final rule?

Response. The section was revised to use the term "mobile medical service" rather than "mobile service." This was done to indicate clearly that the provisions in this section only apply to medical use. In addition, in paragraphs (a)(1) through (a)(4), "client's address of use" was replaced by "client's address," which is defined in § 35.2. This was done to recognize that mobile medical service may be provided at an area of use or a temporary jobsite. (Area of use is defined as a portion of an address of use that has been set aside for the purpose of receiving, using, or storing byproduct material.)

Paragraph (a) was also revised to replace the term "each entity" with the phrase "the licensee and the client." We believe this clearly states our intent that the mobile medical service obtain a letter from each client that delineates the authority and responsibility of the licensee and the client.

In paragraph (b), “the client’s address of use” was replaced by “the client.” This was done to clarify that byproduct material cannot be delivered to the client unless the client has a license allowing possession of the byproduct material. Finally, paragraph (a)(2) was revised to clarify that the instruments referred to in this paragraph refer to those instruments used to measure the activity of unsealed byproduct materials.

Section 35.92, Decay-in-storage.

Issue 1: Should this section be moved to Part 20?

Comment. Commenters believed that decay-in-storage should be addressed in Part 20 rather than in Part 35.

Response. Part 20 provides the general requirements for various waste disposal methods, including the decay-in-storage method. Currently, detailed procedures for decay-in-storage are in license conditions. We believe the specific provisions for decay-in-storage that apply to a medical licensee should be codified in Part 35.

Issue 2: Should the rule continue to require that byproduct material be held for 10 half-lives before disposal as nonradioactive material?

Comment. Commenters were divided in response to the NRC's request for specific comment on whether byproduct material should be held for a minimum of 10 half-lives. Commenters in favor of retaining the requirement believed that it would help ensure that the waste is not prematurely disposed of as nonradioactive material due to human error or instrumentation malfunction. They also believed that licensees may not have adequate survey instruments to survey low-energy beta emitters, such as sulfur-35 (S-35).

Commenters supporting the deletion of the requirement indicated that holding the byproduct material for 10 half-lives was in no way a guarantee that the waste could be disposed of as nonradioactive material. They believed that deletion of the requirement to hold the material for 10 half-lives would improve sanitary conditions and provide for more efficient use of storage space. Finally, they indicated that although S-35 is difficult to detect with a survey instrument, S-35 is not a component in any FDA-approved radiopharmaceutical for routine use.

Response. We have not included a requirement in the final rule to hold byproduct material for 10 half-lives before disposing of the material as nonradioactive material. We do not believe this requirement is needed in light of the requirement in paragraph (a)(1) that precludes disposal of byproduct material without regard to its radioactivity until radiation levels adjacent to the material do not exceed background levels.

Issue 3: Does the requirement to obliterate radiation labels only apply to the outermost container, especially if the material will be handled as biohazardous material?

Comment. A commenter questioned whether the obliteration of radiation labels is only required on the outermost container. Specifically, the commenter asked whether labels needed to be defaced on inner containers if the label on the outer container had been defaced and the inner label was not visible.

Response. Radiation labels must be removed or obliterated from both the inner and outer containers once the material has decayed unless the released material will be handled as biomedical waste after it has been released from the licensee. (In almost all cases, biomedical waste is incinerated.) For example, the radiation labels on used syringes/needles cannot be defaced without opening the waste barrels, which would expose employees to biological and radiological hazards. Because the waste barrels will be incinerated, the licensee may deface or remove the labels visible on the outer waste barrels/containers without defacing labels on the individual syringes/needles.

If the waste material is going to be incinerated, the licensee does not need to obliterate the radiation labels. If the waste will not be incinerated, the labels must be obliterated because the containers might be placed in unrestricted areas. In this latter situation, members of the public could be unnecessarily alarmed by the radiation labels.

Issue 4: What type of byproduct material may be held for decay-in-storage?

Comment. A commenter asked whether radioactive “seeds” can be held for decay-in storage.

Response. The rule allows a licensee to hold byproduct material with a physical half-life of less than 120 days for decay-in-storage before disposal without regard to its radioactivity. If a “seed” contains byproduct material with a half-life of less than 120 days, this provision applies.

Issue 5: Were there any other changes made between the proposed and final rule?

Response. Yes. Paragraph (a) was revised to clearly indicate that the provisions in this section pertain only to disposal of the material without regard to its radioactivity. Licensees must continue to comply with any other regulations that pertain to disposal of the material (e.g., Environmental Protection Agency and State biomedical waste regulations).

SUBPART D - Unsealed Byproduct Material - Written Directive Not Required

General comments.

Issue 1: What are the correct titles for Subparts D and E?

Comments. Commenters recommended renaming Subparts D and E to avoid use of the terms “low dose” and “high dose.” A commenter recommended renaming these sections: Subpart D–Unsealed Byproduct Material-Written Directive Not Required and Subpart E–Unsealed Byproduct Material-Written Directive Required.

Response. We agree that the titles of Subparts D and E should be renamed to avoid use of the terms “low dose” and “high dose.” Subparts D and E in the final rule have been renamed to use the requirement for a "written directive" as the basis for associating the use of the material to radiation risk. The new titles are Subpart D–Unsealed Byproduct Material-Written Directive Not Required and Subpart E–Unsealed Byproduct Material-Written Directive Required.

Issue 2: Are the regulations in Part 35 (except the training and experience requirements) needed?

Comment. Commenters proposed removing regulations for diagnostic nuclear medicine, except for training and experience, from Part 35. The commenters believed that properly trained physicians with the assistance of other associated nuclear medicine health care providers and the standards of radiation protection in Part 20 are all that are necessary to protect the public health and safety adequately.

Response. During the development of the proposed rule, we eliminated requirements in the current Part 35 that are contained elsewhere in the Commission's regulations, such as the radiation protection requirements in Part 20. Part 35 licensees will need to comply with these requirements, such as the ALARA provisions in Part 20, but we believe there is no need to duplicate requirements.

Part 20 contains general radiation protection requirements applicable to all licensees whereas Part 35 contains requirements specific to medical use licensees. While some commenters believe that Part 35 should not contain any requirements associated with low risk procedures, certain radiation protection-related requirements specific to medical use are needed in Part 35 because of their contribution to risk reduction. For example, the final rule retains requirements to perform quality control tests on instrumentation used to measure the radioactivity of patient dosages before administration. These regulations are necessary to provide high confidence that the instrumentation used to measure dosages is operating properly.

In other cases, more specific requirements were kept in Part 35 where justified by risk. The majority of those requirements deal with the therapeutic uses of sealed radioactive material. We believe that the requirements in the final rule are necessary, in addition to the requirements in Part 20, to assure that the dosage administered to a patient is as prescribed by the AU and to assure protection of workers and the public.

Issue 3: Should the requirements for diagnostic and therapeutic uses of unsealed byproduct materials for medical use be combined?

Comment. A commenter believed that the proposed rule intermingled requirements for diagnostic and therapeutic nuclear medicine and failed to provide a regulatory scheme appropriate to each.

Response. Early in the rulemaking process, we considered structuring the rule to have completely "stand-alone" subparts for each type of medical use. However, under this approach, there would have been significant duplication of the requirements which would make the entire rule unnecessarily voluminous. For example, if we took this approach, each subpart would have had a section that addressed when a license was needed, criteria for amending a license, or RSO qualifications.

We have structured the rule so that Subparts A, B, C, J, L, M, and N contain the requirements that apply to all licensees. Subparts D, E, F, G, H, and K contain the requirements that apply to a particular modality, e.g., Subpart D provides specific requirements for the use of unsealed byproduct material which does not require a written directive, and Subpart E contains the requirements for the use of unsealed byproduct material which requires a written directive. The subparts for each type of use also contain the specific training and experience requirements for the AU. (Note, the training and experience requirements in Subpart J will only be effective for two years after the effective date of the final rule to allow a transition period for individuals currently in training.)

Section 35.100, Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

Issue 1: Why doesn't the NRC eliminate or reduce the regulation of certain § 35.100 materials?

Comment. A commenter recommended eliminating or reducing regulation of materials in § 35.100 with extremely low doses (e.g., 35 microcuries of I-125 iothalamate, 10 microcuries of I-125 albumin and 1 microcurie of Co-57 cyanocobalamin) because medical use of these materials involves minimal risk.

Response. We do not believe that the requirements for the medical use of byproduct material described in § 35.100 should be eliminated. If this material is not handled safely, the public or occupationally exposed individuals could receive an exposure in excess of the Part 20 dose limits. We have, however, reduced some regulatory requirements that apply to this type of use, e.g., §§ 35.24, 35.61, 35.92, and 35.290. Explanations for these changes can be found in the discussions of the respective sections.

Issue 2: Should §§ 35.100 and 35.200 be combined because the procedures performed in both modalities do not require a written directive?

Comment. A commenter suggested that the two types of studies listed under Subpart D in the proposed rule in §§ 35.100 and 35.200 should be combined into one category, “unsealed byproduct material for which a written directive is not required.”

Response. Early in the development of the proposed rule, we considered combining these two categories into one section. We did not do so because we believe that the training and experience requirements for individuals using byproduct material for imaging and localization should be more rigorous than such requirements for individuals who only use unsealed byproduct material for uptake, dilution, and excretion studies. This is because AUs using unsealed material under § 35.200 are allowed to compound radiopharmaceuticals and, in general, are handling multiple types of radionuclides at higher activity levels than users performing uptake, dilution, and excretion studies.

Issue 3: Is the reference in § 35.100(b) referring to § 35.292 correct?

Comment. A commenter suggested the cross reference in § 35.100(b) to § 35.292 should be § 35.290.

Response. The cross reference in § 35.100(b) to an individual who meets the criteria to become an AU for use of unsealed byproduct material for imaging and localization is correct. (The requirements in the proposed rule § 35.292 were moved to § 35.290 in the final rule.) We also added a reference to § 35.390. Final §§ 35.292 and 35.390 give physicians authorization to prepare radioactive drugs using generators and reagent kits. AUs qualified under the final

§ 35.190 (proposed § 35.290) do not have this type of authorization.

Issue 4: Why aren't FDA-approved IND pharmacokinetic studies addressed in the proposed rule?

Comment. A commenter stated that the proposed rule did not recognize pharmaceutical companies that do not have a 10 CFR Part 35 license but label compounds with byproduct material and transfer them to specific licensees for use in FDA-approved IND pharmacokinetic studies. This commenter proposed addition of a new § 35.100(c) to address this issue.

Response: The final rule addresses this comment and other omissions in the proposed rule. The proposed rule did not recognize pharmaceutical companies not having a Part 32 license but labeling compounds with byproduct materials and transferring them to a specific licensee for use in FDA-approved IND studies. Also, the proposed rule did not recognize the use of unsealed byproduct material obtained from an NRC or Agreement State licensee in accordance with an RDRC protocol. Finally, § 35.100 in the proposed rule did not allow specific medical use licensees, who do not have individuals qualified under §§ 35.292, 35.55, 35.920, or 35.980, to prepare unsealed byproduct material in accordance with an RDRC or IND protocol. These omissions in the rule unduly restricted labeling and transfer of unsealed byproduct material to Part 35 licensees.

The final rule addresses all of the aforementioned situations. Sections 35.100 and 35.200 have been revised to include materials approved by an RDRC or IND protocol.

Section 35.190, Training for uptake, dilution, and excretion.

Issue 1: Were there any other changes made between the proposed and final rule?

Response: Yes. The training and experience requirements that were in the proposed § 35.290 were moved to § 35.190 in the final rule. This is discussed in greater detail under the general discussion on training and experience located at the beginning of this Section in the Federal Register notice.

Section 35.200, Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraphs (c) and (d) were added to this section. These changes are identical to the change made to § 35.100. The reasons for these additions are in the discussion of § 35.100, Issue 4.

Section 35.204, Permissible molybdenum-99 concentration.

Issue 1: Why is it necessary for NRC regulations to address molybdenum-99 concentrations?

Comments. Commenters argued for eliminating this section because U.S. Pharmacopeia (USP) and FDA standards already address this area. Another commenter believed that the proposed requirements were excessive and unnecessary. Some commenters supported the change in the requirement from evaluating molybdenum-99 concentration for only the first elution instead of evaluating it for every elution.

Response. We believe this requirement is necessary as a means to check generator eluate before medical use to ensure that the generator was not damaged in shipment. However, we did revise paragraph (a) to express the permissible concentration level as 0.15 kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (0.15 microcures of molybdenum-99 per millicurie of technetium-99m). This level is identical to that used in the U.S. Pharmacopeia (USP) 23 U.S. Pharmacopeial Convention, Inc., 1995, page 1486-1487.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We revised paragraph (c) to be more precise. We replaced the phrase “measure molybdenum concentration” with the phrase “measure the molybdenum-99 concentration.”

Section 35.205, Control of aerosols and gases (current rule).

Issue 1: Should the current requirements related to aerosols and gases be deleted?

Comment. We received comments both in support of and in opposition to deletion of this section. A commenter supported the deletion of the requirement because the current requirement is too prescriptive. Another commenter believed that the requirement to control radioactive aerosols and gases should be retained. This commenter stated that the requirement of having a negative pressure environment ensures that there is control over “escaping radioactive gas.”

Response. We do not believe this requirement is needed in Part 35. Part 35 licensees must comply with the occupational and public dose limits of 10 CFR Part 20. Additional prescriptive requirements for limiting airborne concentrations of radioactive material are not warranted in Part 35.

Section 35.190, Training for uptake, dilution, and excretion.

Issue 1: Is it necessary for physicians using byproduct materials under § 35.100 to be board certified in nuclear medicine?

Comment. A commenter believed that there should be an alternative training and experience pathway for individuals who are not full board certified nuclear medicine physicians but would like to become an AU for materials authorized under § 35.100.

Response. The final rule contains three pathways for individuals to become AUs for material under § 35.100. The first pathway, § 35.190(a), requires a physician to be certified by a board recognized by NRC. The second pathway, § 35.190(b), allows AUs, qualified under §§ 35.290, 35.390, or equivalent Agreement State requirements, to use byproduct material under § 35.100. The third pathway, § 35.190(c), requires that the physician complete 60 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material for uptake, dilution, and excretion studies that includes classroom and laboratory training and work experience.

Section 35.290, Training for imaging and localization studies.

Issue 1: Should all individuals be required to have experience with eluting generators?

Comment. A commenter recommended that we revise the training and experience requirements in the proposed § 35.292 to state: “To be authorized for possession and use of technetium from a generator system, the applicant must obtain supervised practical experience eluting technetium-99m from generator systems.” The commenter is drawing a distinction between AUs that plan to limit their use to unit dosages rather than preparing the dosages themselves. The commenter believed the requirement, as proposed, would be consistent with actual practice and good radiation safety practices. In addition, the commenter recommended that the preceptor not be required to certify that an individual has achieved a level of competency with regards to use of generators. Another commenter believed that we should delete requirements for individuals to receive training in eluting

generators, measuring and testing the eluate for radiochemical purity and processing the eluate with reagent kits because unit dosages are obtained from a Part 32 licensee.

Response. We have not modified the regulatory text to establish a separate category for AUs only using unit dosages nor have we deleted the training requirement. We believe that a separate category would unnecessarily complicate licensing practices for both the licensees and NRC. Physicians who meet the qualifications in the final § 35.290 are authorized to use generator systems and reagent kits in the preparation of radioactive drugs and must be trained accordingly, even though they may elect to use only unit dosages. For the same reason, we believe that the preceptor should certify that the individual has achieved a level of competency with regards to use of generators. We would unduly limit where a licensee may obtain unsealed byproduct material if we made any further revisions to the regulatory text.

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response: Yes. The requirements in the proposed § 35.290 were moved to the final § 35.190. The requirements in the proposed § 35.292 were moved to the final § 35.290. This is discussed in greater detail under the general discussion on training and experience located at the beginning of this Section of this Federal Register notice.

SUBPART E - Unsealed Byproduct Material - Written Directive Required

Section 35.300, Use of unsealed byproduct material for which a written directive is required.

Issue 1: Were there any changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (b) was revised to add a reference to § 35.390. The proposed rule would have allowed licensees to use any unsealed byproduct material prepared for medical use by an ANP, a physician who is an AU and who meets the requirements specified in § 35.292, or an individual under the supervision of either as specified in § 35.27. Because a physician who meets the requirements specified in § 35.390 also meets the training requirements in § 35.292, we added a reference to § 35.390 in paragraph (b).

Paragraphs (c) and (d) were added to this section. This was done because the proposed rule did not recognize pharmaceutical companies who do not have a 10 CFR Part 32 license but label compounds with byproduct materials and transfer them to a specific licensee for use in FDA-approved IND studies. Finally, the proposed rule did not allow specific medical use licensees to prepare unsealed byproduct material in accordance with an IND protocol. These omissions in the rule unduly restricted labeling and transfer of unsealed byproduct material to Part 35 licensees. The final rule addresses all of these situations.

Sections 35.100 and 35.200 have been revised to address both the RDRC and IND approved material. Note: § 35.300, in contrast to §§ 35.100 and 35.200, does not include reference to RDRC authorizations because the FDA RDRC regulations restrict RDRC approvals to pharmacokinetic and physiological studies. Further, the dose limits for a study that can be approved by an RDRC under 21 CFR 361.1 are as follows:

(1) For a single administration of radioactive drug - whole body, gonads, blood forming organs, and lens--3 rem; all other organs--5 rem; and

(2) For multiple administrations (or annual dose commitment) - whole body, gonads, blood forming organs, and lens--5 rem; all other organs--15 rem.

Section 35.310, Safety instruction.

Issue 1: Who must participate in annual retraining?

Comments. Many commenters questioned the need for the training required in § 35.310. Some commenters found this requirement to be very burdensome. A commenter suggested that posting radiation safety precautions on a patient's door or in the patient's chart could replace the training requirement. Another commenter believed that annual retraining was not needed for certified radiation therapy technologists and therefore recommended that the section specify annual retraining only for "persons without specialized training in handling radioactive materials." Other commenters thought the requirement was too prescriptive, and that licensees should be given the freedom to decide how to assure compliance with the dose limits in § 35.75 on a case-by-case basis. According to another commenter, annual retraining should be required only for health care personnel who were not directly

supervised by trained radiation safety staff. Some commenters argued against placing the training requirement in Part 35 while other commenters suggested that we make the requirement only applicable to allied health workers who are not nurses. The commenter believed that the need for training should be dependent on whether the licensees needed to provide the individual with dosimetry. These commenters suggested that we revise § 35.310(a) to state: “A licensee shall provide radiation safety instruction, initially and at least annually, to personnel, whose exposure rates may approach the limits in Part 20, caring for patient or human research subjects that have received therapy . . .”

Response. We believe that it is important that personnel caring for patients or human research subjects, who cannot be released in accordance with § 35.75, receive instruction in limiting radiation exposure to the public and workers and in the radiation safety actions to be taken in the case of a death or a medical emergency. We believe this provision is needed because exposure in excess of the public dose limits could result unless proper precautions are taken. We also believe this requirement is consistent with ALARA principles. We do not believe that only posting doors or a chart provides adequate information to the licensee's staff without corresponding instruction.

The rule does not require the licensee to instruct all hospital staff. Instruction must only be provided to personnel caring for patients or human research subjects who cannot be released in accordance with § 35.75. We considered the comments regarding who should receive the training and whether the requirement should be linked to a dose limit. We decided that it is more appropriate to specify that instruction must be provided to personnel caring for patients or human research subjects rather than tie the instruction to the dose limits in 10 CFR Part 20. This was done because it is possible for

a licensee's staff member to receive a dose that is less than the occupational dose limits in Part 20 but take an action that could result in a dose to a member of the public that exceeds the public dose limit.

We have given the licensee flexibility on the level and detail of instruction that must be provided. The instruction need only be commensurate with the duties of the personnel. In other words, the licensee can determine the appropriate level of radiation safety instruction to be provided, depending on the level of care provided by the personnel. For example, a primary care nurse may receive detailed instructions on patient and visitor control, but the ward clerk may only need to be instructed to observe the caution signs on the patient's door.

We recognize that certified radiation therapy technologists or other individuals who have received specialized training in handling radioactive materials would have received training in the areas required by this section as part of a training program. However, we believe that refresher training is warranted because of the potential for unnecessary exposure to workers and the public if needed safety precautions are not observed.

Issue 2: Can the AU have a designee?

Comment: A commenter recommended that paragraph (a)(5) be revised to require that personnel be instructed to notify the RSO (or his or her designee) and the AU (or his or her designee) if the patient or the human research subject dies or has a medical emergency.

Response: The final rule provides the RSO flexibility in designating who should be notified to address radiation protection issues. However, the rule does not provide for the AU to have a designee. The AU is the individual who is responsible for the medical use and supervision of other persons using the byproduct material. Therefore, because of the type or dosages that are administered under § 35.300, we believe it is important that an AU be available to be contacted in case of a medical emergency.

Issue 2: Should the current requirements in § 35.315(a)(4) related to surveys be deleted?

Comment. A commenter indicated that removal of the requirements in § 35.315(a)(4) to perform a radiation survey following a therapeutic administration of I-131 would be ill-advised. This commenter also believed that the requirement to perform a careful contamination room survey should not be removed.

Response. We do not believe these survey requirements should be in Part 35. We believe Part 20 contains adequate information regarding radiation surveys. Part 35 licensees are responsible for assuring occupational and public dose limits are not exceeded. As required in § 20.1501, the licensee must make or cause to be made surveys that are needed to comply with the regulations in Part 20.

Issue 4: Were there any other changes made to this section between the proposed and final rule?

Response. Yes. Paragraph (a) was revised to replace the term “radiopharmaceutical therapy” with the phrase “therapy with unsealed byproduct material.” This change was made to clarify that this section addresses both drugs and biologics containing byproduct material. The term radiopharmaceutical does not cover both radioactive drugs and radiobiologics containing byproduct material.

Paragraph (b) was restructured to clarify our intent that, for the purpose of this section, only the RSO may have a designee.

Section 35.315, Safety precautions.

Issue 1: Does the rule allow the licensee to quarter patients or human research subjects receiving therapy with unsealed byproduct material together?

Comment. Commenters did not believe that the requirement to quarter a patient or human research subject, who cannot be released in accordance with § 35.75, in a private room with a private bathroom is justifiable. They believed that the requirement should be deleted, citing calculations suggesting that two patients undergoing identical radiation treatments (unsealed byproduct material) and occupying the same room would each have their total radiation dose increased by less than 1 percent due to the presence of the other patient. Others believed that allowing two patients undergoing treatment in the same room would be helpful as a means of controlling contamination and would therefore support ALARA principles.

Commenters also argued that allowing a nontherapy patient to share a room with a patient undergoing radiation therapy (unsealed byproduct material) was unacceptable. They said this would result in unnecessary exposure to a member of the public and would not be ALARA.

Other commenters opposed allowing the sharing of a posted restricted room with a patient who was not undergoing radiation therapy. These commenters were concerned about the radiation exposure to hospital housecleaning staff. Other commenters supported the requirement for a private room because they were concerned that medical institution management and health care insurance companies would not allow patients or human research subjects to be quartered in private rooms or in a double room (with single occupancy) because it was too expensive.

Response. We revised the rule text to allow the licensee to quarter a patient or human research subject in either (1) a private room with a private sanitary facility or (2) a room, with a private sanitary facility, with another individual who also has received therapy with unsealed byproduct material and who also cannot be released under § 35.75. This requirement does not preclude the licensee from quartering the patient in a private room. We made this revision because we recognize that the exposure patients could receive from each other is insignificant in light of the exposure the patient is receiving from their administered dosages. Conversely, we do not believe that it is appropriate to allow a therapy and nontherapy patient to share a room because the nontherapy patient would not receive a radiation exposure under normal conditions.

We believe that contamination control is essential and that two patients could share the same room without negatively affecting the licensee's ability to control contamination. We do not agree that sharing rooms will increase the exposure to housecleaning staff. Assuming that two patients

require treatment, the exposure to the housekeeping staff should not be significantly different whether the patients are quartered in the same room or different rooms. In either situation, licensees have the responsibility to maintain the exposures below the Part 20 limits.

Issue 2: Should a patient or human research subject be allowed to take contaminated articles home?

Comment. A commenter asked that this section be revised to permit the licensee to package items contaminated with short lived material so that the items could be released at the same time as the patient or human research subject. The commenter went on to state that the section should also include a requirement for the licensee to instruct the individual not to unpack the package and use anything in the package until a predetermined date. Finally, the commenter recommended that the date be calculated to ensure the activity remaining in the package is small.

Response. We have not changed the rule because of the potential for unnecessary radiation exposure to the public if the material were not handled properly once it is released from licensee control. Any items contaminated as a result of medical use are the responsibility of the licensee.

Issue 3: Should additional requirements be added to § 35.315 to address hospitalization of patients released under § 35.75?

Comment. A commenter questioned how a patient, who had been released under § 35.75 but was still hospitalized for another medical condition, should be managed. The commenter was concerned that the nursing staff could be confused by the instructions provided to the patient under § 35.75,

because § 35.315 does not address the management of this type of patient. The commenter suggested that § 35.315 be revised to require licensees to implement radiation safety precautions, to include posting warning signs, whenever patients receiving therapy quantities of radiopharmaceuticals are hospitalized.

Response. It is the licensee's responsibility, under § 35.75, to control any individual who has been administered unsealed byproduct material or implants containing byproduct material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (500 mrem). The requirements for a patient released in accordance with § 35.75 apply to the case in which a patient goes home as well as the case in which a patient would remain an in-patient in the hospital for reasons other than radiation protection. The licensee must identify who would be the maximally exposed individual prior to releasing the hospitalized individual from licensee control (§ 35.75). If that individual would not be released from the hospital immediately, the maximally exposed individual may be a member of the nursing staff. In this case, the licensee should estimate the exposure to a member of the nursing staff and take this into consideration when preparing the instructions required by § 35.75.

We do not believe that § 35.315 should be revised to specifically address patients who are released in accordance with § 35.75 but remain hospitalized for other reasons, because § 35.75 contains adequate provisions to assure that the maximally exposed individual does not receive a dose in excess of 5 mSv (500 mrem).

Issue 4: Are the release limits in § 35.315 appropriate?

Comment. A commenter was strongly in favor of the revised survey requirements because the previous rules were too prescriptive and not warranted for reasons of health and safety. Another commenter believed that the release limits in § 35.315(a)(3) are unnecessarily low and are not logical when compared to the annual limit of intake for I-131 and I-125.

Response. Under § 35.315, material and items from the patient's or the human research subject's room cannot be removed until the radiation levels adjacent to the items are not distinguishable from natural background, unless the material and items are managed as radioactive waste. Because this requirement is consistent with the release requirements in § 35.92 for radioactive waste, we do not believe additional modification is needed.

Issue 5: Should the bioassay requirements in the current § 35.325(a)(8) be included in the final rule?

Comment. A commenter asked that the current § 35.315(a)(8) be revised and incorporated in the final rule. The commenter recommended that the following provision be added: A licensee shall measure the thyroid burden of each individual who helped prepare or administer a dosage of iodine (I-131) within 3 days after administering the dosage if there is a likelihood that the individual would receive more than 10 percent of the Annual Limit on Intake in Appendix B to 10 CFR Part 20.

Response. We have not added the provision. Licensees are required to comply with Part 20. As such, they must limit occupational exposure to the limits in Part 20. In addition, they must develop, document, and implement a radiation protection program commensurate with the scope and extent of licensed activities (§ 20.1101). This would include assessing whether individuals preparing or administering I-131 need bioassays.

Issue 6: Were there any other changes made to this section between the proposed and final rule?

Response. Yes. Paragraph (b) was restructured to clarify our intent in the proposed rule that, for the purpose of this section, only the RSO may have a designee. This same change was made in § 35.310. The reasons for this change are under the discussion on § 35.310, Issue 2.

Section 35.390, Training for use of unsealed byproduct material for which a written directive is required.

Issue 1: Should the training and experience requirement in § 35.390 include instruction in giving radiation safety directions in the event the patient or human research subject dies?

Comment. A commenter recommended that we add a requirement to § 35.390(b)(1) to require that an individual receive instruction on issuing radiation safety directions in the event the patient or human research subject dies.

Response. We do not believe such a change is necessary because this issue should be addressed as part of the licensee's overall radiation safety program. Licensees should have flexibility in how they address radiation safety issues associated with the death of a patient or human research subject.

Section 35.392, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).

Issue 1: Were there any other changes made between the proposed and final rule?

Response: Yes. We added specific training and experience requirements for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries). This addition is discussed in greater detail under the general discussion on training and experience topic located at the beginning of this Section in the Federal Register notice.

Section 35.394, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).

Issue 1: Were there any other changes made between the proposed and final rule?

Response: Yes. We added specific training and experience requirements for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries). This addition is discussed in greater detail under the general discussion on training and experience topic located at the beginning of this Section in the Federal Register notice.

SUBPART F- Manual Brachytherapy

Section 35.400, Use of sources for manual brachytherapy.

Issue 1: Should all therapy sealed sources be required to have National Institute of Science and Technology (NIST) traceability?

Comment. Some commenters felt that all sources used for therapeutic applications should be required by regulation to have a NIST traceable national standard. Conversely, some commenters felt that it is inconsistent to require licensees to calibrate in the absence of national standards for all clinically used sources.

Response. Section 35.432 requires that sealed source output be measured with a dosimetry system that has been calibrated using a source traceable to NIST. We agree with the AAPM position that all therapy sealed sources should be calibrated in accordance with a traceable standard. In limited cases, a traceable standard identical to the therapy sealed source is not available. In these cases, the requirement allows the licensee the flexibility to use protocols accepted by nationally recognized bodies to meet the calibration requirement. As an example, AAPM Report Number 21 recommends that sources used in radiation therapy have calibrations with direct or secondary traceability to national standards. AAPM defines direct traceability as “when a source or calibrator has been calibrated either at NIST or an AAPM-Accredited Dosimetry Calibration Laboratory.” AAPM defines secondary traceability as “when the source is calibrated in comparison with a source of the same design and comparable strength which has direct

traceability or when the source is calibrated using an instrument with direct traceability.” In addition, AAPM Task Group (TG) 56 recommends that, for “sources that do not have a national standard yet, users should develop a constancy check calibrated against the vendor’s standard and use this constancy check to verify the source strength. Another option is to develop one’s own secondary standard.” Therefore, this allows the licensee flexibility in the event that a direct NIST traceable standard does not exist.

Issue 2. Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We added a new paragraph (b) to this section that allows a licensee to use therapy sources in medical research as long as the research is conducted in accordance with an effective IDE application accepted by the FDA. This was done to clarify how research with sealed sources could be conducted if the medical use of the sources differed from the statements found in the SDR for the sources. With this revision, we allow previously registered sources to be used for uses other than those described in the original registration process, as long as the use is in accordance with an effective IDE application accepted by the FDA.

In addition, the first sentence in § 35.400 was revised to read “brachytherapy sealed sources” to make it clear that this subpart only applies to sealed sources.

Section 35.404, Surveys after source implant and removal.

Issue 1: Is the requirement for radiation surveys after brachytherapy source implant necessary?

Comment. Commenters felt that a survey of the patient after brachytherapy sources have been implanted for the purpose of looking for misplaced sources would be difficult. The commenters stated that with the sources in the patient, the background around the patient is too high to detect an errant source. Additionally, some commenters believed that radiation surveys should be deleted from Part 35 because this is a Part 20 issue.

Response. We agree that Part 20 requires surveys and control of licensed material. However, in order to clarify that surveys must be conducted to locate and account for all sources that have not been implanted, the requirements for surveys have been retained in § 35.404(a). Section 20.1501 requires, in part, that each licensee shall make, or cause to be made, surveys that may be necessary for the licensee to comply with the regulations in this part and are reasonable to evaluate: the magnitude and extent of radiation levels; the concentration or quantities of radioactive material; and the potential radiological hazards that could be present. In addition, Subpart I of Part 20 requires that the licensee secure from unauthorized removal or control and maintain constant surveillance of licensed material. Because surveys under § 35.404(a) are not necessarily radiation surveys, the term “radiation” has been removed from the title and from the rule text in paragraph (a). Depending on the area being surveyed and the ability to distinguish from the radiation background around the patient implanted with brachytherapy sources, these surveys may include radiation

surveys of a facility room (e.g., operating room suite) after the patient with implanted sources has been removed from the room, radiation surveys in and around the patient's room after the implant, and visual surveys of the patient's bed after the implant.

Issue 2: Does adjacent area include contiguous restricted and unrestricted areas?

Comment. A commenter requested that we explicitly indicate that “adjacent area” does not categorically include “contiguous restricted and unrestricted areas.” The commenter stated that the latter wording appears in the current § 35.415(a)(4). The commenter indicated there was little rationale for the current requirement and that it has been deservedly removed in the proposed rule.

Response. We deleted the requirement in the current rule (§ 35.415(a)(4)) that required radiation surveys in contiguous restricted and unrestricted areas to demonstrate compliance with the requirements of Part 20. We agree that this requirement is covered by Part 20. Deleting this requirement and relying on Part 20 to ensure that adequate surveys are performed provides the licensee flexibility in performing adequate surveys. For instance, an adequate survey following a brachytherapy implant may include a radiation survey of restricted and unrestricted areas with a maximally loaded patient in a representative patient room. If the circumstances of subsequent brachytherapy patient treatments are equivalent to the initial survey conditions, we believe that the licensee may rely upon the initial survey to show compliance with Part 20.

Section 35.406, Brachytherapy Source Accountability.

Issue 1: Were there any other changes made in this section between the proposed and final rule?

Response: Yes. We changed the title of the section from “Brachytherapy source inventory,” to “Brachytherapy source accountability.” The revised title more accurately reflects the section content. Inventory requirements are in § 35.67.

Section 35.410, Safety Instruction.

Issue 1: Who must participate in annual retraining?

Comment. Many commenters questioned the need for the training required in § 35.410. Some commenters found this requirement to be very burdensome. Another commenter believed that annual retraining was not needed for certified radiation therapy technologists and therefore recommended that the section specify annual retraining only for “persons without specialized training in handling radioactive materials.” Additionally, one commenter stated that initial and annual training of all nurses and all hospital staff was not cost effective.

Response. We believe that it is important that personnel caring for patients or human research subjects, who have received a brachytherapy implant and cannot be released in accordance with § 35.75, receive instruction. This instruction should include information on how to minimize radiation exposures to the public and workers and the radiation safety actions to be taken in the case of a death or a medical emergency. We believe this provision is needed because exposures, in excess of the public dose limits, could result if proper precautions are not taken. We also believe this requirement is consistent with ALARA principles.

We do not require training of all hospital staff. We allow the licensee flexibility in determining the appropriate level of radiation safety instruction to be provided depending on the level of involvement by various personnel caring for the patient or human research subject. The instruction need only be commensurate with the duties of the personnel. For example, a primary care nurse may receive detailed instructions on patient and visitor control, but the ward clerk may only need to be instructed to observe the caution signs on the patient's door.

We recognize that certified radiation therapy technologists, or other individuals who have received specialized training in handling radioactive materials, may have received training in the areas required by this section as part of their training program. However, we believe that refresher training is warranted because of the potential for unnecessary exposure to workers and the public if needed safety precautions are not observed.

Issue 2: When notifying an AU following a patient emergency, can a physician designee be notified if the AU is not available?

Comment. A commenter recommended that for notifications of patient or human research subject medical emergencies, the AU, like the RSO, may not always be readily available and should also have the option to specify a designee, such as another physician.

Response. Sections 35.11 and 35.27 permit an individual to use byproduct material under the supervision of an AU. Nevertheless, an AU, and not a designee, is responsible for the medical use and supervision of the byproduct material. In the event of a medical emergency with a patient or human research subject implanted with brachytherapy source(s), we believe that because of the doses administered under § 35.400, an AU must be notified, and this notification cannot be delegated to a designee.

Issue 3. Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (a)(5) was restructured to clarify our intent that, for the purpose of this section, only the RSO may have a designee.

Section 35.415, Safety precautions.

Issue 1: Is it necessary to list the type and location of emergency response equipment in the regulations?

Comment. Commenters believed that the requirement to list the contents of an emergency pack was too prescriptive and confusing. Additionally, commenters felt that the emergency equipment did not need to be specifically located in the patient's room but could be somewhere accessible in the hospital. Commenters felt that the licensee should have the freedom to adequately stock and locate an emergency pack. One commenter also felt that the phrase "supplies necessary to surgically remove applicators" kept in the patient's room implied that surgery should be conducted in a nonsterile environment.

Response. We agree with these comments because, in a performance-based rule, the essential objectives are stated in the regulatory text. Therefore, we revised the regulatory text to identify the essential objective of possessing emergency response equipment. The list of specific items that are needed for emergency response has been deleted from this section. The licensee has the flexibility to determine the type of emergency response equipment needed to respond to a source that is either dislodged from the patient or lodged within the patient following removal of the source applicators.

We revised paragraph (b) to delete the word "inadvertently." This was done because the emergency equipment is needed regardless of how the source became unshielded or if the source remains within the patient.

We agree that the emergency equipment does not need to be maintained in the treatment room, but that it should be maintained near each treatment room in order to expeditiously respond to an emergency. The rule allows the licensee some flexibility in locating the emergency response

equipment. The issue of whether to conduct surgical removals of applicators or sources within the treatment room that may not be a sterile environment is left to the licensee's discretion.

Issue 2: Can brachytherapy patients be quartered in the same room with a patient not receiving radiation therapy?

Comment. We solicited specific comment on the current requirement that the licensee not quarter a brachytherapy patient in the same room as an individual who is not receiving radiation therapy. The majority of commenters agreed with the requirement that would allow more than one brachytherapy patient in a room, although a few commenters questioned this requirement. Some commenters felt that the final rule should retain the requirement that the licensee not quarter a patient in the same room as an individual who is not receiving radiation therapy. One commenter pointed out that a posted restricted room should not be shared with a patient not involved in the therapy. Another commenter believed that the requirement to prohibit placing a therapy patient in the same room as a nontherapy patient should apply not only to patients confined under § 35.75 but to any patient where another individual in the room could receive over 1mSv (100 mrem). This commenter believed that limiting the requirement to only patients confined under § 35.75 was not "as low as is reasonably achievable." Conversely, other commenters suggested that the provision for a private room be deleted.

Response. In the current Part 35, we permit the sharing of a brachytherapy patient room with another individual undergoing radiation therapy. In the final rule, we have, however, clarified that the other "individual undergoing radiation therapy" refers to other brachytherapy patients. This is consistent with changes made to § 35.315 to allow therapy patients treated with unsealed material to share a room, if they cannot be released pursuant to § 35.75.

We made no change in the final rule in response to comments on the allowable exposure to the patient sharing the room or to individual members of the public. Section 20.1301 requires the licensee to conduct operations so that, in part, the total effective dose equivalent to individual members of the public from the licensed operation does not exceed 1mSv (100 mrem) in a year, exclusive of the dose contributions, in part, from exposure to individuals administered radioactive material and released in accordance with § 35.75. Section 35.75 allows release of patients administered byproduct material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (500 mrem). Therefore, if the licensee confines a patient receiving brachytherapy and has not authorized the release of the patient under § 35.75, the licensee must limit the total effective dose equivalent to individual members of the public to less than 1mSv (100 mrem) in a year. Alternatively, if the licensee authorizes the release of the patient receiving brachytherapy under § 35.75, the licensee must make the determination that the total effective dose equivalent to any other individual is not likely to exceed 5 mSv (500 mrem). The licensee must also provide the released individual, or the individual's parent or guardian, with instructions on actions recommended to maintain doses to other individuals as low as is reasonably achievable, if the total effective dose equivalent to any other individual is likely to exceed 1 mSv (100 mrem). In all cases, the licensee is required, under § 20.1101, to conduct operations to achieve doses that are as low as is reasonably achievable.

Issue 3: Where should “Radioactive Materials” signs be posted?

Comment. A commenter suggested that having the option to put “Radioactive Materials” signs in the chart instead of on the door was not a good idea. This commenter felt that signs should be posted on the door and in the chart.

Response. Section 35.415(a) in the proposed rule specifically stated that the patient's door had to be posted. We revised this section in the final rule to require, in part, that the licensee visibly post the patient's or human research subject's room with a "Radioactive Materials" sign. We revised this section to allow the licensee flexibility in determining where to place the posting so that it is visible. This requirement would not preclude placing a sign on the chart provided the sign would be visible to someone entering the treatment room. Notations as to where and how long visitors may stay may be placed in the patient's chart or posted on the door.

Issue 4: Why is there a difference in the time periods to notify the AU and the RSO, or his or her designee, if the patient or human research subject dies or has a medical emergency?

Comment. A commenter suggested that the notification time period for medical emergency and death should be the same.

Response. We have maintained the difference in the notification time periods. These differences recognize that, in the event of a medical emergency, the notification should be as soon as possible, rather than immediately, because the licensee's primary responsibility during a patient's medical emergency is the care of the patient.

Issue 5: Following a patient emergency, when should an AU versus an RSO be notified and can a physician designee be notified if the AU is not available?

Comment. A commenter felt that the AU should be notified, and the notification of the RSO should be left to the AU's discretion. Another commenter recommended that for notifications of medical emergencies, the AU, like the RSO, may not always be readily available and should also have the option to specify a designee, such as another physician.

Response. Sections 35.11 and 35.27 permit an individual to use byproduct material under the supervision of an AU. Nevertheless, an AU, and not a designee, is responsible for the medical use and supervision of the byproduct material. Therefore, under § 35.415(c) an AU and not a designee must be notified in the event that a patient or human research subject has a medical emergency or dies. The RSO is responsible, in accordance with § 35.24, for implementing the radiation protection program. Therefore, we believe that notification of the RSO, or his or her designee, provides additional assurance that appropriate corrective actions to respond to the radiation safety hazard associated with the emergency or death are taken.

Issue 6. Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (a) was reworded to make it clear that the requirements in § 35.75 apply to the release of individuals, not to the confinement of individuals. In addition, paragraph (c) was restructured to clarify our intent that, for the purpose of this section, only the RSO may have a designee.

Section 35.432, Calibration of brachytherapy sources.

Issue 1: What does the term “nationally recognized body” mean and what is the policy for taking recommendations from these bodies and making them regulations?

Comment. Commenters questioned what was intended by the term “nationally recognized body” and stated that professional protocols may contain items that are recommended but that were never intended to be adopted as regulation.

Response. Examples of nationally recognized bodies would include ANSI, AAPM, ACR, and ACMP. Documents issued by nationally recognized bodies include multiple peer-reviews of the reports, protocols, or standards. The requirements in this subpart are based on recommendations found in AAPM Task Group Reports 40 and 56 and are consistent with the calibration requirements for sealed sources and devices for therapy, including those found in ANSI documents. However, we did not include all recommendations made in the above reports because we recognize the prescriptiveness of various reports. Instead, the essential objectives for the test being required are listed in the rule.

Issue 2: What is the meaning of the term “intervals consistent with 1 percent physical decay?”

Comment. One commenter requested that we clarify whether the requirement meant 1.0000 percent or allowed rounding down to 1 percent. Some commenters felt that 1 percent was too prescriptive since the calibration requirements are higher. Additionally, a commenter stated that correcting the output/activity at “intervals consistent with 1 percent physical decay” was not feasible for short half-life sources.

Response. This section requires that outputs or activities be corrected for physical decay at intervals consistent with 1 percent physical decay. “Rounding” is a mathematical term. “Consistent with 1 percent” includes from 0.51 percent to 1.49 percent. The 1 percent correction is separate from the calibration. The accuracy of the calibration must be within a given percentage provided by the published protocol used to perform the calibration. This calibration is then used to determine the dose delivered to the patient.

Issue 3: Should the rule contain a requirement to perform calibration measurements of brachytherapy sources and, if so, can the licensee rely on the manufacturer’s or distributor’s calibration?

Comment. Some commenters felt that the vendor’s calibration should be verified by the licensee because use of unverified vendor calibrations poses serious hazards for the patient. Other commenters believed that the calibration of brachytherapy sources should be the manufacturer’s responsibility. They also suggested that we could easily verify procedures at a few manufacturers, rather than at multiple hospitals. Some commenters also requested that we require the manufacturer to guarantee the source activity or output within 3 percent.

Response. We believe that it is good practice to verify the calibration provided by the manufacturer because of the high risk associated with therapy doses to patients. Therefore,

§ 35.432 requires a licensee to perform calibration measurements before the first medical use of a brachytherapy sealed source. The licensee shall determine the source output or activity using a dosimetry system that meets the requirements of § 35.630(a); determine source positioning accuracy within applicators; and use published protocols accepted by nationally recognized bodies to meet the previous two requirements.

However, we also believe that licensees should be able to use calibration measurements provided by the source manufacturer or by a calibration laboratory accredited by the AAPM as long as it was done in accordance with a published protocol accepted by a nationally recognized body using appropriately calibrated equipment. In order to ensure the reliability of the outputs or activities reported by the manufacturer, the manufacturer must perform the calibrations in accordance with the same requirements placed on the licensee. This also addresses the issue that the manufacturer guarantee the activity or output because the manufacturer must use at least the same performance standard as the licensee.

Issue 4: What is the meaning of the term “full” in “full calibration?”

Comment. A commenter suggested that the title be changed to “Verification of calibration measurements of brachytherapy sources.” Another commenter requested clarification of the term “full” in “full calibration.” Another commenter suggested that the term “full calibration” be replaced with “spot check” and the phrase “spot check assay” should be added to be consistent with terminology used in AAPM Task Group Reports 40 and 56.

Response. We agree that the term “full” is confusing in the title because we do not define “full.” Therefore, the title has been changed to “Calibration measurements of brachytherapy sealed sources.” Also, the term “full” has been deleted from this section. The terminology, including “calibration,” was selected to be consistent with terminology used in Subpart H of Part 35 and in AAPM and ANSI reports.

Issue 5: When should the brachytherapy sources be calibrated?

Comment. A commenter requested clarification on whether brachytherapy sources should be calibrated before the first medical use period or before the first medical use at a given facility.

Response. As written, the requirement is that each licensee must calibrate its brachytherapy sources before the first medical use at the licensee’s facility. If the licensee is licensed for medical use at more than one facility in a single license, this calibration must only be performed once, before medical use, at any of the facilities listed in the license.

Issue 6: Does the rule allow calibration of a sampling of sources when a batch of sources is received?

Comment. Some commenters suggested that for short half-life sources and pure beta-emitting sources (e.g., I-125 and Pd-103), a sampling of the sources should be allowed.

Response. We do not preclude a sampling of short half-life sources when received in a large batch. The rule requires that the calibration be performed using published protocols accepted by nationally recognized bodies. As an example, a nationally recognized body would include the AAPM. The AAPM, in the report from Task Group 40, recommends for short half-life sources that “for groupings with a large number of loose seeds, a random sample containing at least 10% of the seeds be calibrated” and “for a large number of seeds in ribbons, a minimum of 10% or 2 ribbons (whichever is larger) should be calibrated.” However, this recommendation is made to the end user and as a verification of the source strength measurement performed by the manufacturer. The licensee must ensure that the published protocol allows for sampling of sources that have not been previously calibrated.

Issue 7: Are sources currently in the possession of the licensee exempt from the calibration requirement?

Comment. A commenter suggested that we include an exemption for sources in inventory before the requirement becomes effective.

Response. Since calibration standards and methods have varied over the years, we believe that to ensure that the correct dose is given to the patient in accordance with § 35.41, the brachytherapy source output or activity must be calibrated in accordance with published protocols currently accepted by nationally recognized bodies. Therefore, we did not revise this section to include the requested exemption for sources in inventory prior to the effective date of the rule. Instead, we revised this section to clarify that all brachytherapy sources must be appropriately calibrated before the first medical use after the effective date of this rule. By including this date, the rule now indicates that sources currently possessed by the licensee must be

calibrated prior to the first medical use after the effective date of this rule and in accordance with a published protocol accepted by a nationally recognized body. If the source was previously calibrated in accordance with a currently accepted published protocol and using a dosimetry system that meets the requirements of § 35.630(a), the calibration would not need to be repeated.

Issue 8: Are the calibration requirements for high-dose versus low-dose sources the same?

Comment. A commenter requested that we make a distinction between high-dose and low-dose brachytherapy sources when requiring calibration.

Response. We do not believe that a distinction is needed. We believe that when delivering a therapeutic dose to a patient or human research subject, the licensee is responsible for ensuring that the correct dose is administered regardless of the source strength.

Issue 9: Do the manufacturer's measurements need to be performed consistent with those required by the licensee?

Comment. A commenter suggested that for the manufacturer's accepted measurements, the phrase "that are made in accordance with the requirements of this section" be deleted.

Response. This phrase has been retained. To ensure the same level of calibration, we believe that unverified calibrations performed by the manufacturer must meet the same standard of calibration as the calibrations required of the licensee.

Issue 10: Is the requirement for source positioning accuracy necessary?

Comment. Some commenters felt that the requirement for source positioning accuracy within applicators was vague and may be irrelevant or impossible to comply with.

Response. We believe that in order for the licensee to further ensure that the correct dose is delivered, the applicators used to help deliver the dose must be appropriately tested. We reviewed several standards currently available for calibration of brachytherapy sources. For example, AAPM Task Group 40 recommends, at a minimum, that initial tests be performed on brachytherapy applicators. Task Group 40 states that “of major concern is that the applicators position the source where they are intended to be localized, and that any part of the structures which are used to attenuate the radiation (e.g., rectal and bladder shields) have not shifted.”

Issue 11: Should the accuracy of source activity or output determination be stated in the rule?

Comment. A commenter suggested that the accuracy for I-125 be changed to 10 percent because a 5 percent accuracy is not possible.

Response. We deleted the reference to +/- 5 percent from § 35.432(c)(1) of the proposed rule. We do not believe that the accuracy of the source activity or output measurement needs to be stated in the rule because the published protocol addresses the accuracy requirement.

Issue 12: Is new equipment required by licensees to perform calibrations?

Comment. Several commenters indicated that the new requirement to calibrate brachytherapy sources would require licensees not currently involved in teletherapy or remote afterloader therapy to procure equipment. Additionally, a commenter requested clarification on whether a well ionization chamber (e.g., dose calibrator) was adequate for calibrating low dose rate brachytherapy sources because former chambers have historically been associated with § 35.630.

Response. As represented in the Regulatory Analysis accompanying this rulemaking, we recognize that licensees may need to procure additional equipment to meet this requirement. We believe that the additional expenditure is warranted for the licensee administering brachytherapy doses to ensure that the correct dose is administered to patients. We agree that a well ionization chamber could meet the requirement if the chamber, or source used to calibrate the chamber, is traceable to NIST or an AAPM-accredited calibration laboratory and a published protocol accepted by a nationally recognized body is used.

Section 35.433, Decay of strontium-90 sources for ophthalmic uses.

Issue 1: Were there any other changes made to this subpart between the proposed and final rule?

Response: Yes. We added this new section that requires an AMP to calculate the activity of a strontium-90 source that will be used in determining the treatment time for ophthalmic uses. It also requires that the activity be calculated using the source activity determined under § 35.432.

This section was added because we are aware of numerous misadministrations involving strontium-90 for ophthalmic use that were caused by individuals improperly decaying the sources. Given the risks associated with use of strontium-90 and the numerous misadministrations in this area, a more prescriptive requirement is warranted.

Section 35.457, Therapy-related computer systems.

Issue: Were there any other changes made to this subpart between the proposed and final rule?

Response. Yes. We added this new section that is consistent with the requirement found in § 35.657 for other therapy-related computer systems. The new section in the final rule requires brachytherapy licensees, who use treatment planning systems, to perform acceptance testing on the system in accordance with published protocols accepted by nationally recognized bodies.

Section 35.490, Training for use of manual brachytherapy sources.

General comments on this section are summarized under the General Training topic found at the beginning of this section of the Federal Register notice.

Issue 1: Should training include ordering and inventory of byproduct material?

Comment. A commenter requested that we delete the following from work experience requirements: “ordering” material safely and “maintaining running inventories of material on hand.” The commenter believed that there was no risk associated with these procedures.

Response. Because the AU is responsible for use of byproduct material under the license, we believe that experience in ordering and maintaining inventories of radioactive materials is an important component of a training program for an AU.

Section 35.491, Training for ophthalmic use of strontium-90.

Issue 1: Were there any other changes made in this subpart between the proposed and final rule?

Response: Yes. We added this new section. The proposed rule had deleted specific training and experience requirements for individuals who wanted to use strontium-90 for ophthalmic use. Under the proposed rule, these individuals would need to meet the training and experience requirements in the proposed §§ 35.490 or 35.940. This change was proposed because, at that time, we believed it was warranted in view of: the similarities between the use of strontium-90 eye applicators and the use of sealed byproduct material in medical devices, and recent misadministrations involving strontium-90 eye applicators. Upon further review of the misadministrations, we believe that the majority of the misadministration events could have been prevented if an AMP had decayed the sources, rather than additional training and experience for AUs. Therefore, we added a requirement for an AMP to calculate the activity of the source (§ 35.433) and have included a specific section that provides the training and experience requirements for an individual who would like to use strontium-90 ophthalmic treatments.

This section is identical to § 35.941, Training for ophthalmic use of strontium-90, in the current rule, with minor exceptions. We have deleted the phrase “who is in the active practice of therapeutic radiology or ophthalmology.” This was done because we believe it is important that the individual is a physician and therefore this additional level of prescriptive regulation is not warranted. We have also added a requirement for a written statement, signed by a preceptor AU, stating that the individual has satisfactorily completed the training requirements and has achieved a level of competency sufficient to

function independently as an AU for use of strontium-90 for ophthalmic treatments. This change is consistent with the other training and experience sections within the revised rule. The preceptor statement is discussed in more detail under the General Training topic found at the beginning of this section of the Federal Register notice.

SUBPART G - Sealed Sources for Diagnosis

We received comments on only three areas in Subpart G. They are: (1) SSSDR; (2) availability of survey instruments; and (3) training and experience requirements. The first two topics are summarized under the “Global Changes” topic in the beginning of this section in the Federal Register notice because the same comments pertain to multiple sections in the rule. Comments on the training and experience requirements are summarized under the “General Training” topic found at the beginning of this section of the Federal Register notice.

**SUBPART H - Photon Emitting Remote Afterloader Units, Teletherapy Units,
and Gamma Stereotactic Units**

General Comments.

Issue 1: Can this subpart be revised to eliminate redundant and overly prescriptive requirements?

Comment. A commenter suggested that Subpart H should be rewritten to eliminate redundancy and overprescriptive procedures that we expect licensees to follow. The commenter felt that the licensees should have the ability to develop their own procedures instead of the NRC dictating each step.

Response. We agree that the rule should not be redundant, and we have combined sections whenever possible. For example, we combined § 35.644, Periodic spot-checks for low dose-rate remote afterloaders, with § 35.643, Periodic spot-checks for remote afterloader units. However, the full calibration requirements for all therapy units have been retained in separate sections for each type of unit to avoid confusion on the applicability of certain tests for a given therapy unit.

Subpart H contains requirements for emergency response and operating procedures, including full calibration and spot-check tests. Where warranted by risk, we have maintained prescriptive requirements in the rule. We identified the performance objectives for full calibrations and spot-checks

in the rule. This decision was based on various AAPM and ANSI reports. However, the exact content of these procedures has not been specified. These procedures are required to be developed by the licensee and the AMP. Where applicable, the procedures must use published protocols accepted by a nationally recognized body. We believe that this provides the licensee more flexibility in developing its procedures.

Issue 2: How have national standards been incorporated into the rule?

Comment. Commenters were concerned that we are transforming recommended "practice standards" into excessively prescriptive and unnecessarily burdensome regulatory requirements.

Response. In many sections, the rule allows licensees to develop their own procedures in accordance with these multiple peer-reviewed reports, protocols, or standards. Examples include following recommendations published by the AAPM, ACR, ANSI, and ACMP. We believe this provides licensees with the flexibility needed to develop a procedure as long as it meets the minimum regulatory requirements in this subpart.

Issue 3: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We have changed the title of this subpart to make it clear that the requirements in this section refer to only photon-emitting remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Section 35.600, Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit.

Issue 1: Should all therapy sealed sources be required to have NIST traceability?

Comment. Some commenters said that all sources used for therapeutic applications should be required by regulation to have a NIST traceable national standard. Conversely, some commenters said that it is inconsistent to require licensees to calibrate such sources in the absence of national standards for all clinically used sources.

Response. Sections 35.632, 35.633, and 35.635 require that sealed source output be measured with a dosimetry system that has been calibrated using a source traceable to NIST. We agree with the AAPM position that all therapy sealed sources should be calibrated in accordance with a traceable standard. In limited cases, a traceable standard identical to the therapy sealed source is not available. In these cases, §§ 35.632, 35.633, and 35.635 allow the licensee the flexibility to use protocols accepted by nationally recognized bodies to meet the calibration requirement. As an example, AAPM Report Number 21 recommends that sources used in radiation therapy have calibrations with direct or secondary traceability to national standards. AAPM defines direct traceability as “when a source or calibrator has been calibrated either at NIST or an AAPM-Accredited Dosimetry Calibration Laboratory.” AAPM defines secondary traceability as “when the source is calibrated in comparison with a source of the same design and comparable strength which has direct traceability or when the source is calibrated using an instrument with direct traceability.” In addition, AAPM Task Group (TG) 56 recommends that for “sources that do not have a national standard yet, users should develop a constancy check calibrated against the vendor’s standard and use this

constancy check to verify the source strength. Another option is to develop one's own secondary standard." This option allows the licensee to develop a calibration source in the event that a direct NIST traceable standard does not exist.

Issue 2: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We added a new paragraph (b) to this section that allows a licensee to use therapy sources in medical research as long as the research is conducted in accordance with an effective Investigational Device Exemption (IDE) application accepted by the FDA. This was done to clarify how research with sealed sources could be conducted if the medical use of the sources differed from the statements found in the SSDR for the sources. This revision allows previously registered sources to be used for other uses than those described in the original registration process if the source is used in accordance with an effective IDE application accepted by the FDA.

Section 35.604, Surveys of patients and human research subjects treated with a remote afterloader unit.

Issue 1: What is the purpose of the survey required by this section?

Comment. A commenter requested clarification of the requirement to survey the patient or human research subject and the remote afterloader with a portable radiation detection survey instrument to confirm that the source(s) have been removed from the patient or human research subject and returned to the safe shielded position.

Response. The radiation surveys are needed to ensure that a source does not remain within the patient or outside of the source shield following completion of each treatment with the unit.

Issue 2: Who may perform the survey?

Comment. A commenter requested that the rule be revised to allow the medical physicist to train an assistant to do radiation surveys, required by § 35.604, when the physicist is not available.

Response. The rule does not specify who must perform the surveys required by § 35.604. We believe that the licensee should have the flexibility to decide who should perform the surveys. However, the record of the survey must include the name of the individual who performed the survey, in accordance with § 35.2404.

Section 35.605, Installation, maintenance, adjustment, and repair.

Issue 1: Who may repair a LDR unit?

Comment. We solicited comments on whether the restrictions, in this section, on who may work on a device containing a sealed source should apply to LDR units. Some commenters said that the restrictions should apply to LDR units. Other commenters believed that the restrictions should only apply to LDR units if the device manufacturer recommends the restriction for the particular device. Conversely, some commenters said that the restrictions should not apply to LDR units because the risk from these low dose-rate units is minimal enough that a trained individual knowledgeable of the unit's operation could install, perform maintenance, adjust, or repair the device. They believed that we should not "over-regulate" these units. Some commenters also believed that users of nonmedical devices who perform these types of services must submit procedures that show they have had appropriate training in performing these services on the specific devices. They stated that persons who perform installation, maintenance, and repair of other NRC-regulated devices (that do not apply radiation to humans) are routinely limited to services on the specific devices for which they have training and experience, e.g., fixed gauges, radiography cameras, etc. In addition, repairs of therapy devices are not just an issue of source or cable replacement, but could also include electronics and software modifications. Consequently, they believed that none of the training and experience requirements identified in the proposed regulations provide for this kind of training. Therefore, the service provider's specific training must be evaluated by the NRC.

Response. Because of the risk associated with therapy devices, the final rule only allows an NRC or Agreement State licensed entity to install, maintain, adjust, or repair a therapy device that involves work on the source(s) shielding, the source(s) driving unit, or other electronic or mechanical

component that could expose the source(s), reduce the shielding around the source(s), or compromise the radiation safety of the therapy unit or the source(s). Additionally, these regulations limit the replacement, removal, or relocation of the source(s) in a teletherapy unit, a gamma stereotactic radiosurgery unit, a HDR, MDR, and PDR, to an entity specifically licensed by the NRC or an Agreement State for these activities. For LDR source(s), we allow an AMP or a specifically licensed entity to perform these functions. This provides relief for licensees possessing LDRs when replacing decayed sources or removing and installing sources to render each individualized treatment plan. However, for work on the LDR source(s) safe, the source(s) driving unit, or other electronic or mechanical components that may expose the source(s) or compromise the radiation safety of the unit, we believe that specialized training, in addition to the training required to meet AMP status, is necessary to perform these activities. Therefore, only personnel specifically licensed by the NRC or an Agreement State may perform these activities.

Issue 2: Does install, maintain, adjust, or repair include assembly?

Comment. A commenter suggested that the word "assembly" be added to the list of activities that require a specifically licensed person to perform.

Response. We believe that "assembly" is included within the meaning of installation and repair. Therefore, we made no change in the regulatory text.

Section 35.610, Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Issue 1: Does the rule allow individuals other than the patient to be present in the treatment room?

Comment. Commenters indicated that therapy administrations in cardiac catheterization suites require the presence of other persons for the safety of the patient during the treatment, and may require that individuals have access to the patient, through the treatment room doors, without interruption of the treatment. In such cases, the commenters believed that the exposures to personnel were already limited by Part 20 requirements. A commenter also questioned the term “contraindicated” in the phrase “ensuring that only the patient . . . is in the treatment room before initiating treatment with the source(s), unless contraindicated . . .”

Response. We agree that, in limited cases, the licensee may need to allow other individuals in the treatment room during treatment. We also agree that the scope of “unless contraindicated” needs to be defined. Therefore, we modified the regulatory text to permit the presence of individuals approved by the AU, AMP, or RSO in the treatment room, during treatment.

Issue 2: Must the console and the console keys be secured?

Comment. A commenter suggested that securing both the console and the console keys was redundant. The commenter went on to state that securing a teletherapy or a gamma stereotactic treatment room is unnecessary if the console or console keys are secured because it would be highly unlikely that unauthorized individuals would remove the devices given their bulk and weight. The commenter felt that, in keeping with a performance-based rule, this section should be revised to read “prevention of unauthorized use or removal of the device when not in use or unattended.”

Response. This section specifies the mechanism for ensuring that the licensed material in therapy treatment devices is controlled when the devices are not attended or are not in use. In keeping with a performance-based rule, we removed the requirement for a written procedure for security. This allows the licensee flexibility in determining the appropriate method for meeting this requirement. General requirements for security of byproduct material are addressed in Part 20, Subpart I. However, because of the high risk posed by these sources, we believe that a more prescriptive requirement is warranted.

Issue 3: Where should emergency procedures and instructions be posted?

Comment. Some commenters said that requiring a copy of instructions and procedures to be posted only at the device console was too prescriptive. They suggested that the language should be revised to read “in the immediate vicinity of the device console.” A commenter also suggested that paragraph (c) of this section was unnecessary because it requires posting the location of the procedures, and paragraph (b) requires the procedures be posted. Another commenter suggested that, in some cases, a console may not exist.

Response. We have not revised either paragraph (b) or (c) of this section. Paragraph (b) of this section requires that a copy of the emergency procedures required by paragraph (a) be physically located at the device console. Paragraph (c) of this section requires posting the location of emergency procedures and the telephone numbers for the emergency contacts. Because the emergency procedures for some devices (e.g., HDR units) may consist of several volumes of error codes and their meaning, we do not require that these procedures be posted. However, the location where these procedures are stored must be posted at the unit console to alert individuals on where to find the detailed emergency procedures in the event of an emergency. We agree that this does not specifically require posting on the console but may allow, for instance, posting on the wall in front of the console. We believe that a console exists for “remotely” delivered sources since the sources must be removed from the source shielding from outside of the treatment room. For cardiac units, this may be an infusion console.

Issue 4: Should device operators be listed in the license?

Comment. A commenter felt that operator knowledge was vital to prevent a medical event, but the requirements do not address operator education, training, or experience. The commenter suggested that the operator be named in the license.

Response. It is the licensee’s responsibility to assure operators are trained. In accordance with § 35.27, operators use licensed material and operate licensed devices under the supervision of the AU depending on the activity being conducted. Therefore, we do not believe that NRC’s review of a specific operator’s training is necessary.

Issue 5 : What is the appropriate frequency and scope of instruction?

Comment. Some commenters suggested that we clarify that persons not receiving annual refresher training are simply prohibited from operating the unit until the training is provided and that the individuals need not be removed from authorization in the institutional license. A commenter also felt that the instruction requirements were too prescriptive for the variety of devices. In addition, while it may be possible to perform a drill simulating the removal of a patient from a teletherapy unit, such a drill is not practical for a HDR unit. The commenter requested that the regulatory text be revised to read "a licensee shall provide instruction and practice drills or demonstrations, initially and at least annually . . ." Conversely, some commenters suggested that retraining was not necessary at all because the AMP and the operator routinely perform the procedures.

Response. We revised the regulatory text to clarify the requirements for instruction. We believe that initial instruction and annual retraining are needed to ensure that the correct dose is administered to the patient or human research subject and to ensure that responsible individuals appropriately respond to emergencies. We also believe that emergency drills are appropriate for all devices. The requirement for training on emergency and operating procedures has been revised to clarify that the training provided is "as appropriate to the individual's assigned duties." We believe that the revised rule allows the licensee flexibility in determining the appropriate level of instruction to be provided depending on the level of involvement of personnel in the operation of and emergency response for the therapy unit.

Section 35.615, Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Issue 1: Is it necessary to list the type and location of emergency response equipment in the regulations?

Comment. Commenters believed that the requirement to list the contents of an emergency pack was too prescriptive and confusing. Additionally, commenters felt that the emergency equipment did not need to be specifically located in the patient's room but could be somewhere accessible in the hospital. Commenters felt that the licensee should have the freedom to adequately stock and locate an emergency pack. One commenter also felt that the phrase "supplies necessary to surgically remove applicators" kept in the patient's room implied that surgery should be conducted in a nonsterile environment.

Response. We agree with these comments because in a performance-based rule, the essential objectives of the rule are stated in the regulatory text. Therefore, we revised the regulatory text to identify the essential objective of possessing emergency response equipment. The list of specific items that are needed for emergency response has been deleted from this section. The licensee has the flexibility to determine the type of emergency response equipment needed to respond to a source that remains in the unshielded position or is lodged within the patient following completion of the treatment.

We revised paragraph (b) to delete the word "inadvertently" because emergency equipment is needed regardless of how the source became unshielded or if the source remains within the patient.

We agree that the emergency equipment does not need to be maintained in the treatment room, but that it should be maintained near each treatment room in order to expeditiously respond to an emergency. The rule allows the licensee some flexibility in locating the emergency response equipment. The issue of whether to conduct surgical removals of applicators or sources within the treatment room that may not be a sterile environment is left to the licensee's discretion.

Issue 2: Is this section applicable to remote afterloader units with beta-emitting sources?

Comment. We solicited specific response on whether the safety precautions in this section should apply to beta-emitting sources. Some commenters felt that the requirements in this section should not apply to remote afterloader beta-emitting sources, since the lower doses from the beta-emitting sources present a very low risk. For example, some commenters felt that paragraphs (b), (c), (d), and (g) could be waived. Other commenters did not believe that we should waive the requirements in this section for remote afterloader beta-emitting sources in keeping with ALARA.

Response. We revised the title of this subpart to make it clear that it only applies to photon-emitting units. We agree that when requirements for beta-emitting remote afterloader units are subsequently added to the regulations, many of the types of requirements described in this section may be appropriate. However, until the use and safety issues of beta-emitting remote afterloader units are fully understood, specific requirements for these units have not been incorporated into this subpart.

Issue 3: Who may generate a treatment plan?

Comment. A commenter suggested adding a requirement that only an AMP may generate an HDR treatment plan. The commenter believed that the level of complexity and the chance for error in this area certainly warranted a requirement in this area.

Response. We have not revised the rule. We believe that licensees should determine who will generate the treatment plan. Additionally, we remind licensees that under § 35.41, Procedures for administrations requiring a written directive, the licensee must develop, implement, and maintain written procedures to provide high confidence that each administration is in accordance with the written directives, including providing the correct dose to the patient.

Issue 4: Is an intercom system necessary?

Comment. A commenter requested that the requirement for an intercom be deleted because voice communication with the patient is not necessary during treatment. The commenter also suggested that the requirement to have an intercom system restricts treatments given by a deaf employee.

Response. Based on ANSI and AAPM recommendations and to help ensure patient and worker safety, we retained the requirement for an intercom system in the final rule.

Issue 5: Should the word “expeditious” be used in the rule?

Comment. A commenter suggested that the term “expeditious” implies that, if the source is difficult to remove, the licensee will be cited. The commenter also felt that this requirement could interfere with what the physician considers to be in the best interest of the patient.

Response. The potential dose to the patient from a decoupled or jammed therapy source remaining within the patient is significant. Therefore, we believe that the requirement is warranted and have retained it in the final rule.

Issue 6: Who needs to be present during LDR treatments?

Comment. A commenter felt that treatments with an LDR unit should allow for trained individuals, working under the supervision of an AU, who have been trained in the operation of the device to be physically present during treatment initiation and an AU and AMP immediately available. Another commenter felt that the AU and the AMP should be physically present during the initiation of patient treatments involving LDR devices. This commenter also asked whether the reference to a radiation oncology physician includes a resident in training. Still another commenter requested that we delete the

requirement for an AU and AMP to be present for continuation of LDR treatments because the treatment may last 48-72 hours and it is not possible to have someone continually available.

Response. In response to public comments, the requirements for the presence of trained personnel during LDR, MDR, and PDR treatments were revised. The final rule does not contain any requirements for the presence of trained personnel for LDR treatments. The risk associated with use of byproduct material in a LDR and manual brachytherapy are similar. Therefore, we do not believe regulatory text is needed in this area. The final rule allows an AU to permit a physician, working under his/her supervision and with training specific to operation and emergency response to the unit, to be physically present in place of the AU, during initiation of patient treatment. The final rule also allows the AU to permit an individual, working under his/her supervision and with training in removing source applicator(s), to be "immediately available" in place of the AU during continuation of patient treatment. Since the treatment times for pulsed dose-rate treatments are significantly longer than those for high dose-rate treatments and the activities of pulsed dose-rate sources are approximately one-tenth of the activities of high dose-rate sources, the change in physician attendance during pulsed dose-rate treatments is warranted. Additionally, for normal resumption of treatment controlled by the pulsed dose-rate device during the normal continuation of the treatment, the presence of a medical professional is not required. This revision allows the licensee to determine the appropriate personnel to have physically present or "immediately available" for medical response to patients treated with these units.

Issue 7: Who needs to be present during HDR treatments?

Comment. Some commenters felt that a physician and a properly trained radiation therapy technologist should be present for HDR treatments. The commenters believed that the responsibility for the device is the AU's, since this is an FDA-approved device. Another commenter believed that the physical presence of an AMP is sufficient if an AU, or a physician trained to respond to an emergency, could be summoned to the HDR unit console within 2 minutes. Some commenters also requested that all remote afterloader requirements be combined because the present requirements are repetitive.

Response. We believe that the requirements for HDR units should differ from the requirements for LDR, MDR, and PDR treatments because the treatment times and the source activities differ significantly. We believe that the requirements appropriately address emergency situations.

The final rule allows an AU to permit a physician, working under his or her supervision, to be physically present in place of the AU during continuation of patient treatment as long as the physician has received operating and emergency response training for the device and as long as the AU is physically present during initiation of the patient treatment. We believe that this revision is appropriate because it allows the licensee to determine who should be physically present during treatments involving HDR units.

Issue 8: Who needs to be present during gamma stereotactic radiosurgery treatments?

Comment. A commenter requested that for gamma stereotactic radiosurgery treatments, an AU or anyone trained in the setting of the coordinates and emergency procedures should be present. Another commenter suggested that emergency response could be limited to requiring the presence of a

physician capable of dealing with the patient's medical needs and two individuals trained in emergency procedures particular to the unit. Still another commenter suggested that we require continuous monitoring by one trained individual and monitoring by an AU during the start and the end of the treatment.

Response. We require the physical presence of the AU and the AMP throughout the treatment to ensure appropriate response to an emergency and to ensure that the correct dose is delivered to the patient.

Issue 9: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We revised paragraph (b)(2) to delete the word "promptly." We did not believe the word was needed in the regulatory text since the text clearly indicates that the interlock system must cause the sources to be shielded when an entrance door is opened.

We also revised § 35.615 (f) to allow an AU to designate another medical professional to cover availability or physical presence requirements for some therapy unit treatments. In those cases in which an AU takes advantage of this provision, we added a requirement that an AU or RSO must be notified in the event the patient or human research subject has a medical emergency or dies. This notification requirement is similar to § 35.415(c) and provides consistency in the requirements for therapy devices and manual brachytherapy. Section

35.615(f)(4) requires a licensee to notify the RSO, or his or her designee, and an AU if the patient or human research subject has a medical emergency or dies. In cases where an AU is physically present during the patient treatment, the notification need only be made to the RSO.

Section 35.630, Dosimetry equipment.

Issue: Is calibrated dosimetry equipment needed for low dose-rate therapy?

Comment. A commenter suggested that licensees routinely do not have or have available, other than through a source provider, calibrated dosimetry equipment that is applicable to the lower dose rates used in standard brachytherapy. Therefore, the commenter requested that dosimetry equipment only be required for higher dose-rate procedures.

Response. As noted in the Regulatory Analysis accompanying this rulemaking, we recognize that licensees may need to procure additional equipment to meet this requirement. We believe that the additional expenditure is warranted for the licensee administering therapeutic doses to ensure that the correct dose is administered to patients. However, we added regulatory text so that this section is consistent with the requirements in Subpart F, Manual Brachytherapy. In the final rule, a licensee using an LDR source(s) may rely on the manufacturer's calibration, and hence the manufacturer's calibration equipment, as long as the equipment and source calibration is performed in accordance with protocols accepted by nationally recognized bodies.

Section 35.632, Full calibration measurements on teletherapy units.

Issue 1: What does the term “nationally recognized body” mean and what is the policy for making recommendations from these bodies into regulations?

Comment. Commenters questioned what was intended by the term “nationally recognized body” and stated that professional protocols may contain items that are recommended but that were never intended to be adopted as regulations.

Response. Examples of nationally recognized bodies include ANSI, AAPM, ACR, and ACMP. Documents issued by nationally recognized bodies include multiple peer-reviews of reports, protocols, or standards. The requirements in this subpart are based on recommendations found in ANSI and AAPM reports and are consistent with the calibration requirements for other sealed sources and devices for therapy. However, we did not include all recommendations made in the ANSI and AAPM reports because we recognize the prescriptiveness of various reports. Instead, the essential objectives for the test being required are listed in the rule.

Issue 2: What is the meaning of the term “intervals consistent with 1 percent physical decay”?

Comment. One commenter requested that we clarify whether the requirement meant 1.0000 percent or allowed rounding down to 1 percent. Some commenters felt that 1 percent was too prescriptive because the calibration requirements are higher. Additionally, a commenter requested that the posted values be within 1 percent of the mathematically corrected values.

Response. This section requires that outputs be corrected for physical decay at intervals not exceeding 1 month for cobalt-60, 6 months for cesium-137 or at intervals consistent with 1 percent decay for all other nuclides. “Rounding” is a mathematical term. “Consistent with 1 percent” includes from 0.51 percent to 1.49 percent. The 1 percent correction is separate from the output full calibration. The accuracy of the output full calibration must be within +/- 3 percent in accordance with paragraph (b)(1) of this section. This calibration is then used to determine the dose delivered to the patient.

Issue 3: What is the meaning of the term “calibrate” when referring to timer accuracy and linearity?

Comment. Commenters requested the meaning of “calibrate” when referencing timer accuracy and linearity. The commenters suggested that, if the purpose is to measure these items to assure they are within some tolerance, this purpose should be stated in the regulation.

Response. Procedures for calibrating the timer are provided in various protocols, which include tolerances. Examples include ANSI N449 and N449-1, and AAPM TG-40. As stated in this regulation, the calibration must be performed in accordance with published protocols accepted by nationally recognized bodies. The term calibrate, as used in this context, means to perform measurements to assure that the timer is operating appropriately within

a given tolerance. The tolerances may be found in reports such as AAPM TG-40. The licensee is therefore given flexibility in developing its calibration methods.

Issue 4: Why are repetitive output measurements necessary?

Comment. A commenter agreed with the requirement for full calibration of sources. However, the commenter suggested that repetitive output checks of long-lived sources, such as cesium, was unnecessary, because the output is not going to change as long as the source is not leaking.

Response. When delivering a therapeutic dose to a patient or human research subject, we believe that the licensee is responsible for ensuring that the correct dose is administered. Additionally, in accordance with § 35.41, the licensee must implement procedures to ensure that the dose is administered in accordance with the written directive. As part of ensuring that the correct dose is administered, we believe that the source output for all sources used to administer a therapeutic dose must be calibrated and verified. Also, we agree with published protocols, such as ANSI and AAPM recommendations, that include periodic recalibration of source activity when delivering therapeutic doses. Therefore, we retained the proposed calibration requirements.

Section 35.633, Full calibration measurements on remote afterloader units.

Issue 1: Why are repetitive output measurements necessary and shouldn't the output test requirements reference the equipment calibration requirements?

Comment. A commenter agreed with the requirement for full calibration of sources. However, the commenter suggested that repetitive output checks of long-lived sources, such as cesium, was unnecessary, because the output is not going to change as long as the source(s) is not leaking. Another commenter suggested that the output calibration requirement should reference the requirement for dosimetry equipment in § 35.630.

Response. When delivering a therapeutic dose to a patient or human research subject, we believe that the licensee is responsible for ensuring that the correct dose is administered. Additionally, in accordance with § 35.41, the licensee must implement procedures to ensure that the dose is administered in accordance with the written directive. As part of ensuring that the correct dose is administered, we believe that the source output for all sources used to administer a therapeutic dose must be calibrated and verified. Also, we agree with published protocols, such as AAPM recommendations, that include periodic recalibration of source activity when delivering therapeutic doses. Therefore, we retained the proposed calibration requirements. However, for consistency with manual brachytherapy, which is traditionally low dose-rate, we included an allowance for LDR sources in the final rule. Paragraph (f) allows licensees using LDRs to accept the manufacturer's calibration of the unit and source as long as the manufacturer conducted the calibration in accordance with this section and with a published protocol accepted by a nationally recognized body and used a dosimetry system as described in § 35.630(a) to measure the output.

Issue 2: What system tests and tolerances should be included in calibration requirements?

Comment. Commenters requested the meaning of “calibrate” when referencing source guide tubes, connectors, and timer accuracy and linearity. The commenters suggested that, if the purpose is to measure these items to assure they are within some tolerance, this purpose should be stated in the regulation. Another commenter suggested that timer accuracy is irrelevant to dosimetry as long as the timer functions the same at the time of treatment as at the time of calibration (i.e., consistency), and responds linearly. Some commenters requested deletion of: (1) timer accuracy and linearity for LDR and PDR units; (2) guide tube calibrations; (3) connector length calibrations; (4) autoradiograph of LDR sources to verify inventory (because sources are difficult to remove from the unit); and (5) battery backup checks (should only be performed at preventative maintenance inspection conducted by the manufacturer). Additionally, a commenter suggested that a reasonable positioning accuracy was 2 millimeters for an HDR stepping source and 5 millimeters for an LDR source (reference AAPM TG-59). A commenter also requested that we clarify that tests for tubes and connectors apply to tubes and connectors in use, and no tests are required if the unit is not in use.

Response. Various professional reports provide suggested protocols for quality assurance tests on remote afterloaders. We based the performance objectives for various tests in these requirements on recommendations made by AAPM TG-56. For instance, AAPM TG-56 suggests 1 millimeter positional accuracy for HDR, LDR, and PDR units; initial, annual, and quarterly battery backup checks; timer accuracy tests for LDR units; and autoradiograph of LDR sources. We agree with recommendations made in AAPM reports and believe that the calibration requirements of this section are warranted to ensure that the correct dose is administered to the patient.

The terminology used in this section was chosen to reflect the current language used in practice. AAPM reports use “timer accuracy and linearity, applicators, transfer tubes, and transfer tube-applicator interfaces.” We noted small discrepancies in the terminology used in the proposed requirements versus in AAPM reports. Therefore, we revised the term “source guide tube” to “source transfer tube” and the term “connector” to “transfer tube-applicator interface” in the final rule. The tests apply only to units and accessories in use.

Issue 3: How frequently should recalibrations be performed?

Comment. A commenter stated that a full calibration is always performed immediately after the source exchange. However, it is probable that the source exchange for an iridium-192 HDR source may take more than 120 days. The commenter suggested that a full calibration on the source after 120 days was not necessary if the source was not yet exchanged for a new source. Another commenter agreed with the proposed requirement that HDR units should be calibrated within 120 days and that LDR units should be calibrated annually, within 1 year. A commenter also requested clarification of the phrase “not exceeding one quarter.”

Response. We believe that for iridium-192 HDR sources, the source calibration frequency can be changed to “at source exchange” to allow for source exchanges that slightly exceed the 120-day period. Therefore, the frequency for full recalibration of HDR, MDR, and PDR units has been revised to quarterly for sources whose half-lives exceed 75 days. The requirement to perform a full calibration at source exchange has been retained. We believe that this revision will facilitate the use of sources with short half-lives. We also believe that this revision will not reduce safe use of sources whose

half-lives are less than 75 days (e.g., iridium-192), since these sources are exchanged at the end of their useful life, which is approximately quarterly for iridium-192. The phrase “not exceeding one quarter” can be equated to a 3-month period.

Issue 4: Who is required to perform the decay corrections for source output?

Comment. A commenter requested that dosimetrists be allowed to perform decay corrections.

Response. The AMP remains responsible for performing decay corrections because of the high consequence associated with errors in these corrections.

Issue 5: Were there any other changes made in this section between the proposed and final rule.

Response. Yes. The requirement to repeat the full calibration of the remote afterloader unit and source, whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration, was deleted from this section. We deleted this requirement because the requirement to perform output spot-checks on remote afterloader units was deleted from § 35.643.

We also revised § 35.633(b) to include patient dose delivery components for LDR units that are detailed in AAPM TG-56. Paragraphs (b)(4), (b)(5), and (b)(6) were added to this section to encompass all remote afterloaders, including LDRs. The items added to the final rule include measurement of the length of the source transfer tubes and applicators; measurement of timer accuracy and linearity; and function tests of the source transfer tubes, applicators, and transfer tube-applicator interfaces. We believe that these changes are necessary to ensure that, during acceptance testing of the unit and after source replacement, these additional tests that increase patient radiation safety are performed.

Section 35.635, Full calibration measurements on gamma stereotactic radiosurgery units.

Issue 1: What is the meaning of the term “calibrate” when referring to timer accuracy and linearity?

Comment. Commenters requested the meaning of “calibrate” when referencing timer accuracy and linearity. The commenters suggested that, if the purpose is to measure these items to assure they are within some tolerance, this purpose should be stated in the regulation.

Response. The terminology used in this section reflects the current language used in practice. AAPM reports use “timer accuracy and linearity.” As stated in this regulation, calibrations must be performed in accordance with published protocols accepted by a nationally recognized body. The term calibrate, as used in this context, means to perform measurements to assure that the timer is operating appropriately within a given tolerance. The tolerances may be found in reports such as AAPM TG-21. The licensee is therefore given flexibility in developing its calibration methods.

Issue 2: Can the licensee adopt the manufacturer's measurements for relative helmet factors?

Comment. A commenter suggested that many users currently adopt the manufacturer's recommended relative helmet factors rather than measure them directly. The commenter stated that this was preferable because: (1) there are inherent difficulties in measuring these factors; (2) requiring users to measure their own factors could result in large errors in some situations; and (3) using the manufacturer's factors aids in sharing information among facilities conducting research protocols.

Response. We believe that measurement of helmet factors is inherent in patient dosimetry. Various professional reports provide suggested protocols for quality assurance tests on gamma stereotactic radiosurgery units. The performance objectives for various tests in this section are based on recommendations in AAPM Report No. 54. For example, AAPM Report No. 54 recommends that helmet factors be measured by the end user.

Issue 3: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. We revised this section to add the components related to the delivery of the dose to the patient that are in § 35.645, Periodic spot-checks for gamma stereotactic radiosurgery. This was done because all patient dose delivery components detailed in the periodic spot-check section, § 35.645, were not included in the proposed full calibration requirements, and therefore, were not required during initial quality assurance testing on the unit or after source replacement. The new paragraphs (b)(7) through (b)(10) include tests of the treatment table retraction mechanism, helmet

microswitches, emergency timing circuits, and stereotactic frames and localizing devices (trunnions). We believe that these changes are necessary to ensure that these additional tests that involve patient radiation safety are performed during acceptance testing of the unit and after source replacement. These additions are consistent with the approach used in the teletherapy unit requirements for full calibration and spot-checks.

Section 35.642, Periodic spot-checks for teletherapy units.

Issue: What is the meaning of the term “calibrate” when referring to timer accuracy and linearity?

Comment. Commenters requested the meaning of “calibrate” when referencing timer accuracy and linearity. The commenters suggested that, if the purpose is to measure these items to assure they are within some tolerance, this purpose should be stated in the regulation.

Response. Procedures for calibrating the timer are provided in various protocols, which include tolerances. Examples include ANSI N449 and N449-1, and AAPM TG-40. The term calibrate, as used in this context, means to perform measurements to assure that the timer is operating appropriately within a given tolerance. The tolerances may be found in reports such as AAPM TG-21. As stated in this regulation, the measurements must be performed in accordance with procedures established by the AMP. The licensee is therefore given flexibility in developing its spot-check methods.

Section 35.643, Periodic spot-checks for remote afterloader units.

Issue 1: Is an output spot-check necessary?

Comment. Commenters requested deletion of the output spot-check, since output is calibrated at installation and by the manufacturer, thereby satisfying all requirements in assuring correct dosimetry and administration. A commenter also suggested that a requirement to determine the output with a dosimetry system described in § 35.630(b) be included.

Response. We agree that the full calibration output measurements are adequate. Therefore, the output spot-check requirement has been deleted. We believe that a quarterly test for HDR, MDR, and PDR source output and an annual test of LDR source output are sufficient to ensure that the correct dose is delivered to the patient. In the place of the output check, we have included a requirement to check the computer decayed source activity against a precalculated decay chart to confirm that the unit has decayed the source activity properly. The output checks done in accordance with § 35.633 continue to require the use of an appropriate dosimetry system, described in § 35.630, when performing the output calibration.

Issue 2: How frequently should spot-checks be performed?

Comment. Some commenters suggested that the spot-checks be done each day of use, thereby insuring patient safety and not duplicating weekly checks. A commenter requested that the term “beginning of each day of use” be revised to “prior to the use of the device on a given day.” Another commenter suggested that the frequencies provided in NUREG/CR-6276 should be used. With regard to timer constancy, a commenter felt that a monthly check was adequate for LDR units.

Response. The rule text has been revised to state “prior to the first use of a HDR, MDR, or PDR unit on a given day.” We developed the frequency of the spot checks from recommendations of AAPM TG 40 and 56, meetings with medical physicists, input from the Therapy Subcommittee of the ACMUI, and NUREG/CR-6276. Therefore, we believe that the frequencies of the spot-checks are appropriate.

Issue 3: What is the meaning of the term “calibrate” when referring to timer constancy/accuracy and linearity?

Comment. A commenter requested that timer constancy be deleted because it is not a credible source of risk to the patient with the current timer technology. The commenter stated that this is verified at installation and needs no further monitoring. Commenters also requested the meaning of “calibrate” when referencing timer accuracy and linearity. The commenters suggested that, if the purpose is to measure these items to assure they are within some tolerance, this purpose should be stated in the regulation.

Response. The terminology used in this section was chosen to reflect the current language used in practice. AAPM reports use the terminology “timer accuracy and linearity.” The term calibrate, as used in this context, means to perform measurements to assure that the timer is operating appropriately within a given tolerance. The tolerances may be found in reports such as AAPM TG-40. As stated in this regulation, the measurements must be performed in accordance with procedures established by the AMP. The licensee is given flexibility in developing its spot-check methods. We have also retained timer checks because they are recommended by the AAPM and are similar to ANSI requirements for teletherapy. Spot-checks of timer linearity are not required by this section, since we believe that timer linearity need only be measured during full calibration measurements.

Issue 4: Why must nonexistent source exposure indicator lights be checked?

Comment. A commenter suggested that checks of source exposure indicator lights be deleted, since these lights do not exist on a remote afterloader unit.

Response. We are unaware of any units that do not have source exposure indicator lights. Source position indicator light checks are recommended by the AAPM and are similar to ANSI requirements for teletherapy. Therefore, these requirements have been retained in the final rule.

Issue 5: Is it necessary to perform a simulated cycle of treatment?

Comment. A commenter suggested that the requirement to conduct a simulated cycle of treatment should be deleted because it is vague and will not necessarily provide any higher level of assurance that the remote afterloader unit is working properly than the daily and monthly checks already performed.

Response. We agree with this comment and have deleted this requirement from the final rule.

Issue 6: Does a treatment system have to be locked-out if the system fails safety tests but a backup system is available?

Comment. A commenter suggested that we change the wording in this section to be more flexible. The commenter stated that, in some instances, a backup device may be available that will allow patient treatments to continue without compromising patient safety.

Response. This section does not prohibit the use of the unit if the licensee replaces the malfunctioning system before using the unit for treatment.

Issue 7: Should door interlocks and audiovisual systems apply to LDR units?

Comment. Some commenters felt that LDR units may not require interlocks or audiovisual systems, depending on the dose rate and whether sources are gamma-emitters only. One commenter suggested that we always require interlocks, but require an audiovisual system only when direct visual contact is not available. Another commenter felt that we should always require interlocks and an audiovisual system for LDR units.

Response. We revised the title of this subpart to clarify that it only applies to photon-emitting units. We have retained the requirements for interlocks for LDR units. This is consistent with recommendations in AAPM reports. We have not included a requirement for an audiovisual system.

Issue 8: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. Paragraph (b) was revised to explicitly indicate that the AMP must establish written procedures for performing the spot-check outputs required in paragraph (a).

Section 35.645, Periodic spot-checks for gamma stereotactic radiosurgery units.

Issue 1: How frequently should spot-checks be performed?

Comment. A commenter suggested that the requirement for monthly checks be deleted if spot-checks are performed daily. A commenter specified that the term “beginning of each day of use” be revised to “prior to the use of the device on a given day.” Another commenter suggested that the frequencies provided in NUREG/CR-6324 should be used. Other commenters said that: (1) a daily output measurement was not necessary as long as the user checks the mechanical integrity of the system through a standard run; and (2) the manufacturer recommends that the battery backup system only be tested on a monthly basis.

Response. We agree with the commenter. The rule text has been revised to state “prior to first use of the unit on a given day.” We developed the frequency of the spot-checks from recommendations of AAPM Report No. 54, meetings with medical physicists, input from the Therapy Subcommittee of the ACMUI, and NUREG/CR-6324. We believe that the final rule text distinguishes between the checks that must be done daily or monthly. Additionally, the rule only requires output checks and battery backup checks monthly. Therefore, we believe that the frequencies of the spot-checks are appropriate.

Issue 2: Define “assure proper operation of stereotactic frames and localizing devices?”

Comment. A commenter requested that we clarify what is meant by “assure proper operation of stereotactic frames and localizing devices.”

Response. Various professional reports provide suggested protocols for quality assurance tests on gamma stereotactic radiosurgery units. For instance, reports from AAPM, ACR, ACMP, and ANSI may be used by the licensee in performance of these tests. The phrase “assure proper operation of

stereotactic frames and localizing devices” means to perform quality assurance tests on these devices to assure that they operate appropriately when used to deliver a dose to a patient. The measurements must be performed in accordance with procedures established by the AMP. The licensee is therefore given flexibility in developing its spot-check methods.

Issue 3: What is the meaning of the term “calibrate” when referring to timer accuracy and linearity?

Comment. Commenters requested the meaning of “calibrate” when referencing timer accuracy and linearity. The commenters suggested that, if the purpose is to measure these items to assure they are within some tolerance, this purpose should be stated in the regulation.

Response. The terminology used in this section reflects the current language used in practice. AAPM reports use “timer accuracy and linearity.” The term calibrate, as used in this context, means to perform measurements to assure that the timer is operating appropriately within a given tolerance. The tolerances may be found in reports such as ANSI N449. The measurements must be performed in accordance with procedures established by the AMP. The licensee is therefore given flexibility in developing its spot-check methods.

Section 35.647, Additional technical requirements for mobile remote afterloader units.

Issue 1: What are the requirements for discontinuing use of a malfunctioning unit?

Comment. A commenter noted that this section did not contain a requirement for discontinuation of use of a malfunctioning unit and questioned whether this was an oversight.

Response. We agree with this comment. We believe that a licensee using a mobile unit must also meet the requirements described in other sections of this subpart applicable to the particular device in use. However, for clarification, we added regulatory text that prohibits the use of the unit if a safety check is failed. Paragraph (d) now reads: “If the results of the checks required in paragraph (b) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.”

Section 35.652, Radiation surveys.

Issue 1: Are these surveys limited to therapy units?

Comment. A commenter questioned whether the surveys required by this section were only for therapy devices or if they included other instruments or devices used at medical facilities.

Response. The requirements of Part 35 apply only to medical uses of byproduct material. The requirements in this section apply to licenses issued for uses in this subpart. Therefore, these requirements do not include sealed sources covered by other subparts (e.g., Subparts F and G). We added the phrase “licensed under this subpart” to this section to clarify this issue.

Issue 2: Why do radiation levels around devices differ?

Comment. Commenters suggested that the maximum radiation levels and average radiation levels around devices could be made a generic number, as with radiography cameras and source changers. They also suggested that it may make sense to put in the average acceptable reading for each type of afterloader unit (i.e., high dose-rate, low dose-rate, and pulsed dose-rate).

Response. The radiation levels referenced in the SADR differ greatly by device manufacturer. Therefore, we retained the requirement in paragraph (a) of this section “to assure that the maximum radiation levels and average radiation levels from the surface of the main source safe with the source(s) in the shielded position do not exceed levels stated in the Registry.”

Section 35.657, Therapy-related computer systems.

Issue 1: What is the purpose of acceptance testing on computer operating systems?

Comment. Commenters felt that acceptance testing of computer operating systems should be deleted because no method could guarantee that software would always operate appropriately. A commenter also said that this requirement should be deleted because it appears to be a year 2000 concern with operating systems.

Response. We agree with these concerns and have deleted the requirement to verify operability of computerized operating systems. This concern is addressed by the FDA's regulations of medical devices, which require reliability testing on computerized operating systems.

Issue 2: Should acceptance testing of treatment planning systems be a requirement?

Comment. Commenters believed that the requirement for treatment planning system acceptance testing was warranted. However, they suggested that the methodology for acceptance testing should be left to the licensee. The commenters also questioned the ability to guarantee that the systems are operating appropriately and questioned our interest in the device operating system that is reviewed by the FDA.

Response. Paragraph (a) of this section in the proposed rule would have required the licensee to verify that the computerized operating system and treatment planning system are operating appropriately. Based on these comments, FDA's review of reliability testing on medical devices, and the device's associated computer operating systems, we deleted these requirements from the final rule.

The requirement to perform acceptance testing on treatment planning systems has been retained. We believe that this requirement is appropriate and still provides the licensee flexibility in designing its acceptance testing program. We revised the rule language to incorporate the components of acceptance testing addressed in AAPM TG-56. The licensee is provided flexibility in performing acceptance testing of treatment planning systems as long as a published protocol accepted by a nationally recognized body is used and as long as the minimum testing requirements are met.

Section 35.690, Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. Paragraph (b)(3) has been revised to read “ an authorized user of *each* type of therapeutic unit for which the individual is requesting authorized user status.” This change was made to clarify that the preceptor authorized user must certify that the individual has achieved a level of competency sufficient to function independently as an authorized user for *each* type of unit for which the individual would like authorized user status. However, this does not mean that the individual has to satisfy paragraphs (b)(1) and (b)(2) in their entirety for each type of unit, e.g., an individual does not need 1400 hours in a structured educational program if he or she wants to be an AU for two types of units under § 35.690.

General comments on this section are summarized under the General Training topic found at the beginning of this section of the Federal Register notice.

SUBPART J - Training and Experience Requirements

Issue 1: Why are there two sets of training and experience requirements in the revised Part 35?

Comment. One commenter noted that much of Subpart J is redundant with, but not identical to, the training and experience requirements listed in the individual sections of the other subparts. The training and experience requirements should be identical if they are included in 2 subparts within the same part, or they should only be listed once in the part.

Response. We have deleted Subpart J, so there is only one set of training and experience requirements in the final rule. All medical use licensees will have to comply with the new training and experience requirements for AMPs, ANPs, AUs, and RSOs in Subparts B and D through H when the rule becomes effective on [insert date 6 months from publication of the Final Rule]. Individuals who have status as AUs, AMPs, ANPs, and RSOs at the time the rule becomes effective will be “grandfathered” under § 35.57, and will not have to satisfy the new training and experience requirements.

Issue 2: Why were the lists of certifying medical boards in Subpart J of the current Part 35 not updated during the rulemaking to include other medical specialty boards and other subspecialties?

Comment. Several commenters noted that there are other medical specialty boards and other subspecialties that should be added to the lists of certifying boards in Subpart J.

Response. The suggested updates were not made in the final rule because Subpart J was deleted and there are no lists of certifying specialty boards in the new training and experience requirements in Subparts B and D through H of Part 35. Under the new regulations, we will continue to review the appropriate training and experience requirements of the boards and recognize the boards that satisfy these requirements, but we will provide the lists of recognized boards in a public document (e.g., on NRC's Internet site), rather than in the regulations. Prior to the effective date of the final rule, we encourage the certifying boards to submit their applications for recognition under the new regulations.

Issue 3: Why have the references to ACGME programs been retained in Subpart J?

Comment. Several commenters said that all references to ACGME programs of less than 2 years should be deleted.

Response. Subpart J, including the references to ACGME programs of less than 2 years, was deleted in the final rule.

Issue 4: Why are there no training requirements for endovascular brachytherapy in Subpart J?

Comment. One commenter noted that Subpart J includes no training requirements for endovascular brachytherapy.

Response. Subpart J has been deleted in its entirety. When the research on endovascular brachytherapy is completed, the standard protocol for this technology will be evaluated to determine if it is similar to the modalities currently described in Part 35, or if it should be licensed as an emerging technology under § 35.1000. Following this determination, the training and experience requirements for this modality will be evaluated to see if new requirements are needed for this use, or if it should continue to be regulated as a sealed source therapy.

Section 35.981, Training for experienced nuclear pharmacists.

Issue 1: What is the impact of deleting this section?

Comment. All of the commenters that responded to this question, which we asked in the proposed rule, said that this section could be deleted because the requirements in § 35.55 for an experienced nuclear pharmacist are adequate.

Response. Since neither the commenters nor the NRC staff and the ACMUI could identify a current scenario in which an individual would need to use this section to become an ANP, this section was deleted, along with the other sections of Subpart J.

**SUBPART K - Other Medical Uses of Byproduct Material or
Radiation from Byproduct Material**

Section 35.1000, Other medical uses of byproduct material or radiation from byproduct material.

Issue 1: What is the purpose and scope of this section?

Comment. There were a number of general comments on this section. Comments ranged from an endorsement of the need for this section to concerns that NRC's regulations for emerging technologies will limit the use of new technologies and radiopharmaceuticals and, consequently, affect the delivery of high quality health care.

Some commenters believed that the purpose of this section is vague, undefined, and confusing, and that there needs to be a clearer definition of an emerging technology. One suggestion was that the definition be tied to whether an IND/IRB approval is required. Another commenter said that this section should specifically exempt radiopharmaceuticals because they are regulated by the FDA under RDRC, new drug applications (NDA), biologic product license applications (PLA), and INDs. Thus, all radiopharmaceuticals should fit under Subpart D or E.

One commenter said that emerging technology uses should be reviewed on a case-by-case basis to determine their proper location in the regulations. The commenter proposed a process to determine how an emerging technology should be regulated: propose performance-based regulations for a 90-day comment period; locate the regulations in a separate subpart; and establish that any technology placed in this subpart would have a 5-7 year sunset period at which time the regulations for this technology would be relocated in another appropriate subpart. This process would provide the opportunity for the technology to establish itself and allow the regulations to be amended, based on observed risk.

Response. We added Subpart K to Part 35 so that there would be codified regulatory requirements and a more clearly defined process to obtain a license, or a license amendment, for a new medical use of byproduct material or radiation from byproduct material, i.e., an emerging technology. By adding requirements for emerging technologies to the regulations in §§ 35.12(d) and 35.1000, an applicant for these uses knows what type of information needs to be submitted to NRC.

The scope of this subpart includes all new medical uses of byproduct material or radiation from byproduct material. We have not attempted to define more clearly what is included in this subpart or what is excluded from this subpart because there is no way to predict what types of medical technologies will be developed in the future. The Commission, with input from the ACMUI, will determine if the emerging technology is truly a new technology and is covered by Subpart K, or if the “new” technology is actually a type of use regulated under Subparts D through H.

Issue 2: What process will be used to establish regulatory requirements and evaluate applications for emerging technologies?

Comment. Commenters stated that it is important to have a reasonable regulatory scheme and time frame for approving applications for new technologies. Some commenters expressed concerns about placing so much regulatory burden (e.g., too many safety constraints) on new technologies that there is an impact on the development of new products.

Emerging technologies have an undefined risk. Once the risk becomes clear, the degree of regulation that is needed to minimize the risks to the public can be defined. The NRC might be interested in the design of trials involving emerging technologies, and what kind of data are collected, in order to define the risks from emerging technologies.

A model was suggested for establishing the requirements for emerging technologies. Under the suggested model, appropriate professional societies would establish task forces to examine the issues (e.g., the training requirements) associated with the emerging technology. This model was successful in defining the standards for gamma stereotactic radiosurgery in the late 1980's when it was considered an emerging technology.

Response. We agree with these comments and will take them into consideration in setting up the process for establishing regulatory requirements and for approving applications for emerging technologies. We intend to evaluate each technology on a case-by-case basis and to work with the ACMUI, the medical community, the public, and the developers of the new technology, as appropriate, to determine the specific risks associated with the technology and any additional regulatory requirements for the medical use of the technology.

Issue 3. Will the NRC coordinate its regulations for emerging technologies with the FDA's regulations?

Comment. One commenter has observed that the FDA process works well in addressing patient safety for investigational new drugs and devices. This commenter suggested that the NRC communicate its concerns to the FDA to assure that any radiation safety issues will be included and documented in the investigational research process.

Response. We do not intend to develop requirements that are redundant with those of the FDA. FDA and NRC have different authorities and responsibilities for protection of public health and safety; FDA has the authority to approve investigational new drugs and devices; and NRC has the authority to protect the public, workers, and patients from the medical use of byproduct material. However, we have a "Memorandum of Understanding" with FDA under which we coordinate certain agency functions and share information (58 FR 47300; September 8, 1993 and 62 FR 15740; April 2, 1997, renewal).

Issue 4: Why does this section not include training and experience requirements for AUs of emerging technologies?

Comment. Several commenters said that this section should provide the minimum criteria and training requirements for AUs of these new medical uses. The qualifications of individuals to use emerging technologies are pretty well established by the developers of the emerging technology, and they are aware of the radiation safety problems associated with the new technology. Whether it is an emerging technology or not, there is a need to

understand the properties and hazards of the radioactive material being used, the radiobiological issues, and the measures to be taken in the event of a spill, and to demonstrate the ability to safely handle the radioactive material.

Response. Section 35.1000 does not include any training and experience requirements for AUs of emerging technologies because there is no way of knowing what training requirements will be necessary for the safe use of byproduct material in new technologies. The necessary training will be evaluated on a case-by-case basis with input from the ACMUI and individuals who have been involved with development of the technology, as needed, and other input, as appropriate.

Issue 5: Will cost issues be considered during the development of requirements for emerging technologies?

Comment. Comments were provided on several different cost issues. One commenter said that it is very difficult to spend millions of dollars on clinical research on new technologies and have no idea what the regulatory requirements are going to be. Another commenter said that cost effectiveness needs to be considered during the development of requirements for new technologies. For example, a requirement to have multiple professionals present during a procedure would not only increase the cost of the procedure, but would also limit its availability to patients.

Response. The Commission's approval of license applications for the medical use of byproduct material is based on radiation safety issues associated with use of the byproduct material. Requirements for emerging technologies will be based on the risk posed by the specific modality.

Issue 6: Will intravascular brachytherapy be considered an emerging technology in the revised Part 35?

Comment. Some commenters believe that intravascular brachytherapy is still experimental and covered by § 35.6 and need not be considered in § 35.1000. Other commenters believe that intravascular brachytherapy should be categorized, or specifically mentioned, as an emerging technology under the provisions described in § 35.1000.

One commenter stated that in the proposed rule the standard use of radioisotopes in patients in the field of cardiology was reclassified as experimental and cardiologists had become radiation oncologists.

Response. Section 35.6 contains some specific provisions for protection of human research subjects and does not permit the use of byproduct material for medical uses that are not authorized on the licensee's medical use license. Intravascular brachytherapy is a very complex field with a number of radionuclides being evaluated for use. Currently, we are regulating intravascular brachytherapy as a sealed source therapy. When a standard protocol for intravascular brachytherapy has been established, the technology will be reviewed in light of that protocol by the Commission, with input from the ACMUI, the medical community, and the public, to determine if new regulatory requirements are needed for this use. Pending development of those regulatory requirements, licensees will be able to submit a license request, under the provisions of §§ 35.12 and 35.1000, to incorporate the new modality into their licensed program.

Issue 7. What are the training and experience and radiation safety requirements for intravascular brachytherapy?

Comment. Some commenters felt that intravascular brachytherapy should have the same training and radiation safety requirements as the rest of radiation oncology. Other commenters felt that the training and radiation safety requirements for nuclear cardiology should be reserved until the technology advances enough to develop standard protocols with the assistance of a group of experts. Still other commenters stated that the NRC should develop the training and safety requirements for intravascular brachytherapy.

Response. As we noted in Issue 6, intravascular brachytherapy is currently an evolving medical treatment composed of diverse technologies. Currently, we are regulating intravascular brachytherapy as a sealed source therapy with the associated training and experience requirements for that therapy. The types of sources used vary widely in terms of the type of radiation emitted, the activity, and the level of encapsulation. In fact, intravascular brachytherapy may not evolve into either a standard protocol or a single modality. Pending receipt of additional information, we believe that it is too early to make changes in the level of training and experience for the use of intravascular brachytherapy.

Issue 8: Were there any other changes made in this section between the proposed and final rules?

Response: Yes.

The wording in paragraph (a) of this section was corrected to state that the information that is required to be submitted by an applicant for use of byproduct material under § 35.1000 is in § 35.12(b) through (d), not only in paragraphs (b) and (c).

The wording in paragraph (b) was revised to reflect a change in § 35.12(d) that allows licensees to submit an application for a license amendment, rather than a separate license, for use of byproduct material under § 35.1000. This change is discussed under § 35.12.

SUBPART L - Records

Issue 1: Should all the recordkeeping requirements be grouped into one subpart or should they be incorporated into the section requiring the record?

Comment. Commenters provided a wide range of responses to the Commission's question on whether all of the recordkeeping requirements should be grouped into one subpart, or whether they should be incorporated into the individual sections requiring the records. Some commenters favored having all of the recordkeeping requirements in one subpart because this format provides for easy reference, simplifies licensing, assists licensees in meeting their obligations for the radiation safety program, and simplifies compliance. Other commenters favored having the recordkeeping requirements in the individual sections because this format would place all of the requirements pertaining to a particular area of interest in one section. Therefore, licensees would know exactly what was expected of them in a particular area. They also find the similar separation in 10 CFR Part 20 to be confusing. Several commenters preferred a "balanced approach" in which the recordkeeping requirements would be in the individual sections and then all of the requirements would be summarized in a separate subpart.

Response. After reviewing all of the responses to this question, we have concluded that having all the recordkeeping requirements in one subpart makes it easier for licensees to reference the requirements. However, the final rule is consistent with the "balanced approach" because each section in the final rule that has a recordkeeping requirement includes a cross-reference to the specific recordkeeping requirements in Subpart L.

Issue 2: Are all of the recordkeeping requirements in Part 35 needed?

Comment. Comments on the need for the recordkeeping requirements in Part 35 ranged from all of the records are needed; to the only records that are needed are those that document overexposures, exceeding environmental limits, and leaking sources; to the only records that should be required are those that have a documented history of improving radiation safety; to none of the records are needed.

Response. During preparation of the final rule, each specific recordkeeping requirement was reviewed in light of these comments and changes were made, where appropriate. These changes are noted in the discussions of the individual recordkeeping sections.

Issue 3: Are the recordkeeping requirements too prescriptive?

Comment. The recordkeeping requirements in the proposed revision maintain the detailed, prescriptive elements that are in the current Part 35.

Response. All of the elements in the recordkeeping requirements in the proposed rule were considered important for documenting radiation safety issues associated with a risk-informed regulation. During preparation of the final rule, we reviewed each recordkeeping requirement in light of this comment and made appropriate changes.

Issue 4: Why are there different retention periods for the records required by this subpart?

Comment. One commenter said that compliance with NRC's recordkeeping requirements would be simplified if all of the record retention periods were the same. Another commenter suggested that because most of the records have a retention period of 3 years, it would make more sense to include a separate section that states that all of the records in this subpart are to be maintained for 3 years, unless otherwise stated, than to restate the retention period in each section.

Response. The record retention periods in Part 35 were set according to either the safety significance of the action being recorded or the inspection frequency. As a result, there are several different retention periods for records in Subpart L. Because record retention periods are tied to safety considerations, we believe that the regulations should specifically state the retention period for each recordkeeping requirement even if it means repeating regulatory text.

Issue 5: How can a patient's privacy and confidentiality be protected in records required by NRC?

Comment. The patient's privacy and confidentiality are ignored with NRC recordkeeping requirements for records of the patient's name, social security number, and other personal information.

Response. Any records that must include the patient's name or personal information relating to the patient are to be retained by the licensee. Reports relating to medical events, which licensees provide to the NRC, explicitly must not contain the individual's name or any other information that could lead to identification of the individual.

Issue 6: Can initials be used on a record to identify the individual who performs an activity or an operation?

Comment. The requirement to record the "name of the individual" that performed a certain activity appears throughout this subpart. Several commenters said that because it is common practice to utilize initials as identifiers of individuals, the words "name of the individual" should be replaced with "identification of the individual."

Response. We require that the full name of an individual appear on a record to better ensure future identification of the individual who performed the activity or operation. It is not uncommon for several individuals to have different names, but the same initials. Also, initials are more likely to be illegibly scribbled.

Issue 7: Why do some records require a signature, rather than the name of the individual?

Comment. Several commenters said that requiring a signature on a record is prescriptive, not performance based, and does not necessarily mean that an individual has actually read or reviewed a record.

Response. We have required signatures only on those records where we feel it is important to the radiation safety program to document who approved the action, reviewed the report, performed the calibration, etc. If an individual signs a record saying, for example, that he or she performed an action, we assume that the individual actually did perform whatever action was required and is in compliance with the recordkeeping requirements in this part. Note that most of the recordkeeping requirements in Subpart L require the name of the individual, rather than a signature.

Issue 8: Do the recordkeeping requirements in Part 35 allow for the use of electronic signatures?

Comment. Some commenters were concerned that the requirements for signatures preclude maintaining records electronically.

Response. Section 35.5, Maintenance of records, allows records to be maintained electronically. Therefore, electronic signatures are permitted.

Section 35.2024, Records of authority and responsibilities for radiation protection programs.

Issue 1: Can the requirements in this section be made less prescriptive and therefore less burdensome on licensees?

Comment. Several commenters felt that the requirements in this section are too prescriptive and burdensome, especially for private practices with one physician who is also the owner/president and RSO.

Response. We have retained the requirements in this section because we believe that records associated with the authority and responsibilities of the radiation protection program are fundamental to the safe use of byproduct material by all medical licensees, regardless of their size. Even single practice physicians, who may also serve as RSOs, need to be well aware of and to document their authority, duties, and responsibilities associated with being the RSO named on either an NRC or Agreement State license.

Issue 2: Why is it necessary for licensees to retain records of the licensee's management's written approval of actions associated with the radiation protection program for 5 years?

Comment. One commenter said that the requirement in paragraph (a) of this section to retain records for 5 years is excessive.

Response. We consider the records required by paragraph (a) to be important in documenting actions taken by the licensee's management that affect its radiation protection program. These records include requests for a license application, renewal, or amendment; approval of AUs, AMPs, and ANPs; and radiation protection program changes that do not require a license amendment. The 5-year retention period will ensure that the records that

are key to a licensee's radiation protection program are available for review during inspection of medical use licensees. During the development of the proposed rule, we evaluated the retention period for this requirement and changed the retention period from the duration of the license to 5 years. Therefore, the recordkeeping burden for licensees to comply with the requirements in this paragraph is less than the burden to comply with the current rule.

Issue 3: Why is it necessary for both licensee management and the RSO to sign the authorities, duties, and responsibilities of the RSO?

Comment. Several commenters said that the requirement in paragraph (b) for both licensee management and the RSO to sign the authorities, duties, and responsibilities of the RSO was too prescriptive. They felt that it was unnecessary to require the signature of both of them because other sections only require one signature or name. One commenter was also concerned that, if a problem occurred, the written agreement could be used by licensee management against the RSO.

Response. We retained the requirement for signatures of both licensee management and the RSO because we believe it is important that there is a signed record of what the licensee management and the RSO agree are the authorities, duties, and responsibilities of the RSO. If both the licensee management and the RSO have a clear understanding of the responsibilities of the RSO for the licensee's radiation protection program, problems such as that referred to in the comment could be avoided.

Section 35.2026, Records of radiation protection program changes.

Issue 1: Why is there a requirement for retaining records of changes to a licensee's radiation protection program that "do not reduce safety," and why must these records be signed by licensee management?

Comment. Commenters said that it is excessive and unnecessary to retain records of radiation protection program changes that do not reduce safety. In addition, the commenters believed that it is unnecessary to have licensee management sign the records of radiation protection program changes that had already been reviewed and signed by the RSO, the licensee's radiation safety expert.

Response. Licensees are required to obtain Commission approval for changes in their radiation protection program, except for the revisions authorized by § 35.26. Because licensees are not required to submit these latter changes to NRC for approval, the records of the changes made in accordance with § 35.26 provide the Commission an opportunity to evaluate these changes during the inspection process. We believe that this approach is warranted in light of the importance of changes in a licensee's radiation protection program.

The reference in proposed § 35.26(a)(2) to changes that "do not reduce radiation safety" resulted in many comments that this phrase was "ambiguous" and "subjective." The proposed wording was intended to provide the licensee with as much flexibility as possible in making changes in its radiation protection program, without seeking Commission approval. However, because commenters felt that the proposed wording was not clear, the

regulatory text in paragraph (a)(2) of this section has been revised to state the more objective parameter of changes that are “in compliance with the regulations and the license.”

We have deleted the requirement in § 35.2026 for the RSO to sign the records of radiation protection program changes because licensee management is ultimately responsible for the radiation protection program. Therefore, the final rule includes a requirement for licensee management to sign these records.

Issue 2: Can the requirements in this section be made less prescriptive and therefore less burdensome on licensees?

Comment. Several commenters noted that the recordkeeping requirements in this section are quite prescriptive and suggested that the sentence with the list of items that must be included in the records be deleted or revised to be less prescriptive.

Response. We believe that the recordkeeping requirements in this section are needed to document what changes have been made in the licensee’s radiation protection program. We considered the burden on licensees during development of the final requirements for this section and believe that the requirements for radiation protection changes, and the associated records, provide the licensee more flexibility to manage its radiation protection program than in the current rule and reduce the recordkeeping burden on licensees. For example, licensees must currently retain a record of each

radiation protection change until the license has been renewed or terminated. Under the final rule, licensees are only required to retain these records for 5 years.

Issue 3: Why are licensees required to retain a copy of the old radiation protection procedures?

Comment. One commenter questioned the need to retain a copy of the old radiation protection procedures because they are immaterial to the current procedures and could be confusing to workers.

Response. We believe that licensees should retain a copy of their old radiation protection procedures for 5 years so that they are available during the licensee's next inspection after the procedures were changed. If a "problem" or "event" is discovered during an inspection, the radiation protection procedures that were in place at the time of the event may be very useful in determining the cause of the event.

We suggest retaining the copy of the old radiation protection procedures in the licensee's filing system so that they are not readily available for workers to refer to by mistake.

Issue 4. Were there any other changes made in this section between the proposed and final rules?

Response. Yes. The word “safety” was removed from the title of this section, and the requirement for the signature of the RSO was deleted. Both of these changes were made to correct inconsistencies between the regulatory text in this recordkeeping section and the corresponding § 35.26, Radiation protection program changes.

Section 35.2040, Records of written directives.

Issue 1: Is there a need for an NRC requirement to retain a copy of written directives for therapeutic administrations of unsealed byproduct material?

Comment. One commenter said that the requirement for retaining a copy of written directives should exempt radiopharmaceuticals because state laws already require retention of prescription records.

Response. Section 35.40, Written directives, contains a list of items that must be included in a written directive and requires that an AU sign and date the written directive prior to administration of I-131 sodium iodide greater than 1.11 MBq (30 microcuries) or any therapeutic dosage of unsealed byproduct material. In other words, this section includes specific requirements for preparing written directives prior to administering higher dosages of unsealed byproduct material. Prescriptions for radiopharmaceuticals may or may not be signed by AUs and may or may not include all of the items that are required by § 35.40 for written directives for administrations of therapeutic dosages of unsealed byproduct material. We believe that retaining copies of written directives will help ensure that administrations of therapeutic dosages of unsealed byproduct material are in accordance with the written directives. In addition, a copy of the written directive may be useful in evaluating whether a medical event was a result of a generic problem that may also affect other licensees.

Section 35.2045, Records of medical events.

Issue 1: Can the requirements in this recordkeeping section be made less prescriptive and therefore less burdensome on licensees?

Comment. One commenter noted that the recordkeeping requirements in this section are quite prescriptive and suggested that the list of items that must be included in the records be deleted.

Response. The information that must be included in the licensee's record of a medical event is similar to, but not identical with, the information that a licensee is required to report to NRC in accordance with § 35.3045. Therefore, this recordkeeping requirement results in the least burden possible on the licensee because it does not require the licensee to generate any additional information, other than adding the information on the individual(s) involved, that is not included in the report to the NRC.

Issue 2: Should there be a requirement for maintaining records of significant precursor events?

Comment. One commenter opposed the recordkeeping requirement for significant precursor events.

Response. There are no recordkeeping requirements for significant precursor events in the final rule because no requirements for reporting precursor events have been added to the current Part 35.

Issue 3. Were there any other changes made in this section between the proposed and final rules?

Response. Yes. This section was reworded to read "reported in accordance with" instead of "reported pursuant to" to make the rule language more consistent with the other sections of Part 35.

The section was also reworded to read "the effect, if any, on the individual; and the actions, if any, taken or planned to prevent recurrence." The words "if any" and "planned" were added because there might not be any effect on the individual or any actions taken at the time the record is made.

Section 35.2047, Record of a dose to an embryo/fetus or a nursing child.

Issue. Were there any other changes made in this subpart between the proposed and final rules?

Response. Yes. This recordkeeping section was added because it was inadvertently omitted in the proposed rule. It is needed because of the associated requirement in

§ 35.3047(i) for a licensee to keep a record of a dose to an embryo/fetus or a nursing child. The record must contain the licensee's name; names of all the individuals involved; the affected or potentially affected individual's social security number or other identification number if one has been assigned; a brief description of the event; why it occurred; the effect, if any, on the embryo/fetus or nursing child; and the actions, if any, taken or planned to prevent recurrence. A summary of the comments and responses on the associated reporting requirement appears in § 35.3047.

Section 35.2060, Records of calibrations of instruments used to measure the activity of unsealed byproduct material.

Issue 1: Does this section address “calibrations” or “performance checks”?

Comment. A commenter recommended that the word “calibrations” be replaced with the term “performance checks” because the commenter believes that the tests required by the section are more accurately defined as performance checks.

Response. We did not adopt this comment because this section addresses calibration of all instruments used to measure the activity of unsealed byproduct material, including dose calibrators. We believe this is the appropriate term because the term “calibration” is commonly used within the radiation protection profession.

Issue 2: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. The title of this section was changed to more accurately state that it addresses the calibration of instruments used to measure the activity of unsealed byproduct material. In addition, we revised § 35.2060 to delete prescriptive requirements. This revision is consistent with the revisions made to § 35.60. The licensee is only required to record the model and serial number of the instrument; the date of the calibration; the results of the calibration; and the name of the individual who performed the calibration. We believe that this information will provide adequate documentation of calibrations of instruments used to measure the activity of unsealed byproduct material.

Section 35.2061, Records of radiation survey instrument calibrations.

Issue 1: Is it necessary to keep instrument calibration records?

Comment. Commenters suggested that the requirement to retain records of radiation survey instruments be deleted. Some commenters stated that because the current calibration status and expiration date must be displayed on the instrument, they did not see a benefit to radiation safety by maintaining certificates of calibration. Other commenters stated that this section is already covered in 10 CFR 20.2103.

Response. We believe records of calibration should be kept because they can be used to document that the instrument has been calibrated. This is particularly important when the calibration sticker is unreadable, missing, or in error or when an instrument that was used in a required survey cannot be

located. Section 20.2103 requires that licensees maintain records of calibrations, but it does not provide specific recordkeeping requirements. Therefore, this section is needed to provide medical use licensees with specific information on what items must be maintained in this record.

Issue 2: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. We revised § 35.2061 to delete the requirements to include the description of the calibration procedure and the source used in calibrating the meter; the certified exposure rates from the source; the rates indicated by the instrument being calibrated; and the correction factors deduced from the calibration data. These revisions are consistent with the revisions made to § 35.61. In the final rule, the licensee is required to record the model and serial number of the instrument; the date of the calibration; the results of the calibration; and the name of the individual who performed the calibration. We believe this information will provide adequate documentation of calibrations of radiation survey instruments.

Section 35.2063, Records of dosages of unsealed byproduct material for medical use.

Issue 1: Are records of administered dosages of unsealed byproduct material needed?

Comment. Commenters did not believe this recordkeeping section was needed because prescribing and dispensing records are required by state medical and pharmacy laws. Other commenters did not believe that the recordkeeping requirements should apply to byproduct material administered under §§ 35.100 and 35.200.

Response. We believe it is important to keep records of the dosages administered. These records are needed to document that the byproduct material was administered to a patient or human research subject in accordance with the written directive and to document the amount of byproduct material that was administered. However, if a licensee keeps the same records under state law, the licensee need not retain duplicate records.

Issue 2: Should the expiration date of a radioactive drug be deleted from the regulations?

Comment. A commenter indicated that the current requirement (in § 35.53) to record the expiration date of a radioactive drug should not be deleted from the regulations. The commenter believed the expiration date is important because it can be used, for example, to establish time limits on sterility, dosage, and effectiveness of tagging. The commenter also believed the paperwork burden for including the expiration date is minimal.

Response. We agree that the expiration date of a radioactive drug is important. However, we believe that licensees have to comply with other regulations governing the use of drugs that include noting the expiration date because it is related to stability and sterility. Therefore, we do not believe that it is necessary to have a requirement in Part 35 for licensees to record the expiration date of a radioactive drug.

Issue 3: Should the terms “prescribed dosage” be removed from the requirement?

Comment. A commenter asked that the term “prescribed dosage” be deleted from § 35.2063 because there is no requirement for the AU to prescribe the dosage and, in the case of therapeutic administrations, only a written directive is needed.

Response. We have not deleted the term “prescribed dosage.” The term is defined in § 35.2. In Part 35, only an AU may direct the administration of sealed or unsealed byproduct material for medical use.

Issue 4: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. We restructured § 35.2063 to match the format used in other recordkeeping sections. We also revised § 35.2063 to delete the requirements for the record to include the radionuclide, generic name, trade name, or abbreviation of the radiopharmaceutical and its lot number, and the activity of the determined dosage at the time of determination. These items were deleted to make the rule less prescriptive. The final rule requires that the licensee record the radiopharmaceutical; patient or human research subject’s name, or identification number if one has been assigned; the prescribed dosage; the determined dosage, or a notation that the total activity is less than 1.1 MBq (30 microcurie); the date and time of the dosage determination; and the name of the individual who determined the dosage. This information will provide adequate documentation of dosage administrations.

Section 35.2067, Records of possession of sealed sources and brachytherapy sources.

Issue 1: Why should licensees maintain records of negative leak tests?

Comment. A commenter agreed with retention of positive leak test records but not with the requirement to maintain records of negative tests.

Response. The rule requires records of all leak tests required by § 35.67(b) to show that leak tests were performed. We have changed the final rule to require records of the test results, but a licensee has flexibility in how it records the test results. For negative leak tests, a licensee may simply document that the measured activity is “negative.”

Issue 2: Should this section make a reference to § 35.2406, Records of brachytherapy source inventory?

Comment. A commenter asked that we add a reference which states that additional brachytherapy records may be required by § 35.2406.

Response. We do not believe this reference is needed. We have tried to eliminate redundancy and cross referencing in the rule unless it is needed to make the rule more understandable.

Issue 3: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. We revised § 35.2067 to delete the requirements to record the measured activity of each test sample and a description of the method used to measure each test sample in the record. These items were deleted to make the rule less prescriptive.

Section 35.2070, Records of surveys for ambient radiation exposure rate.

Issue 1: Are contamination surveys included in this section?

Comment. A commenter indicated that the requirement for records of removable contamination should be deleted because § 35.70 does not require removable contamination surveys.

Response. The commenter is correct. We have deleted the requirement for the licensee to record removable contamination in each area expressed in disintegrations per minute per 100 square centimeters and the instrument used to analyze the samples.

Issue 2: Are the requirements in this section already covered by § 20.2103, Records of surveys?

Comment. Commenters did not believe this section was needed because radiation surveys are addressed in § 20.2103.

Response. 10 CFR Part 20 contains general provisions on records. Section 20.2103 requires that licensees maintain records of surveys, but it does not provide specific recordkeeping requirements. This section is needed to specify what Part 35 licensees must document in the record required by this section.

Issue 3: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. We deleted the requirements in this section to record a plan of each area surveyed; the trigger level established for each area; and the detected dose rate at several points in each area expressed in millirem per hour or the removable contamination in each area expressed in disintegrations per minute per 100 square centimeters. These items were deleted to make the rule less prescriptive. The final rule requires the licensee to record the date of the survey; the results of the survey; the instrument used to make the survey; and the name of the individual who performed the survey.

Section 35.2075, Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material.

Issue 1: Should paragraph (b) of this section that requires that a record be kept that instructions were provided to a breast-feeding woman be deleted?

Comment. A commenter stated that the requirements in paragraph (b) [proposed paragraph (c)] are intrusive into medical practice. The commenter believed that instructions should be left to the physician's judgment.

Response. We did not make any changes in paragraph (b) of this section that requires licensees to keep a record that instructions, including written instructions, were provided to a breast-feeding female if the radiation dose to the infant or child from continued breast-feeding could result in a total effective dose equivalent exceeding 5 mSv (0.5 rem). We believe that providing written instructions to patients or human research subjects is necessary because they may not remember all the oral instructions. In addition, written instructions provide needed information to other family members or individuals who are caring for the patient or human research subject. Consequently, we believe that a licensee should retain a record to demonstrate that instructions were provided to a breast-feeding female.

Issue 2: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. We corrected paragraph (a) of this section in the proposed rule because it inadvertently required that licensees maintain records of all releases. This recordkeeping requirement was more restrictive than the current rule. We have revised the rule to require records of the release of

individuals only when the total effective dose equivalent is calculated by using the retained activity rather than the administered activity; using an occupancy factor less than 0.25 at 1 meter; using the biological or effective half-life; or considering the shielding by tissue. We also revised paragraph (b) to specify that the record required by this paragraph must be maintained for three years.

Section 35.2080, Records of administrative and technical requirements that apply to the provision of mobile medical services.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. We revised § 35.2080 to delete the requirement to record a plan of each area surveyed and the measured dose rate at several points in each area of use expressed in millirem per hour. These items were deleted to make the rule less prescriptive. The final rule requires the licensee to record the date of the survey; the results of the survey; the instrument used to make the survey; and the name of the individual who performed the survey. In addition, we clarified that the letter that permits the use of byproduct material must delineate the authority and responsibility of the licensee and the client.

Section 35.2092, Records of decay-in-storage.

Issue 1: Are the requirements in this section already covered by § 20.2103, Records of surveys?

Comment. Commenters did not believe this section was needed because radiation surveys are addressed in § 20.2103.

Response. 10 CFR Part 20 contains general provisions on records. It does not provide specific recordkeeping requirements for disposal of waste through decay-in-storage. Section 35.2092 is needed to specify what Part 35 licensees must document in the records required by § 35.92.

Issue 2: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. We revised the first sentence to replace the term “made in accordance with” with the phrase “as required by.” We believe this makes the sentence more readable. We also deleted the requirement to document the name of the radionuclide that was disposed. We do not believe it is necessary for the licensee to document what material was disposed of because § 35.92 no longer requires that the material be held for 10 half lives. However, this does not preclude the licensee from including this information in the record.

Section 35.2204, Records of molybdenum-99 concentration.

Issue 1: Can this record be deleted?

Comment. Commenters suggested that this section, as well as § 35.204, be deleted. They did not believe the rule should require licensees to measure molybdenum-99 concentrations. (See comments on § 35.204.)

Response. We did not delete the requirement for licensees to measure molybdenum-99 concentrations, nor have we deleted the requirement for licensees to maintain a record of the molybdenum-99 concentration tests required by § 35.204. We believe the record is needed to document that the test has been performed and that the results of the test do not exceed the levels specified in § 35.204.

Section 35.2310, Records of safety instruction.

Issue 1: Is it necessary to maintain records of safety instruction given to non-film badged workers?

Comment. According to commenters, it is excessive to require the licensees to maintain records of training given to non-film badged allied health care workers, who received instruction in accordance with §§ 35.310, 35.410 or 35.610.

Response. Records of all individuals receiving safety instruction in accordance with §§ 35.310, 35.410 or 35.610 are needed to document that the instruction was provided by the licensee. We believe that it is important that personnel, caring for patients or human research subjects who have received

radiopharmaceutical therapy (and cannot be released in accordance with § 35.75), receive instruction in limiting radiation exposure to the public or occupational workers and what actions should be taken in the case of a medical emergency or death.

Issue 4: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. The title of this section was changed to correspond to the title of § 35.310, Safety instruction. That section includes the requirement for licensees to retain a record of individuals receiving safety instruction.

Section 35.2404, Records of radiation surveys of patients and human research subjects.

Issue 1: Is it necessary to maintain records of negative surveys? Also, can the record retention requirement be changed from 3 years to 1 year?

Comment. Some commenters felt that maintenance of negative surveys for 3 years was excessive and suggested that the survey record include only an indication of the survey being performed and the results of any positive surveys. These same commenters also suggested that the record need only be kept for 1 year.

Response. We simplified the recordkeeping requirements in this section by deleting the requirement to record the location of the survey and the patient identifier. These items were deleted to make the rule less prescriptive. We added a requirement to record “the results of the survey” because we do not believe that a requirement to record the results of the survey is excessive, even if the results are that no sources are unaccountable. We have also retained the 3-year recordkeeping period to be consistent with the 3-year inspection period for most medical use licensees.

Issue 2: Could the recordkeeping requirements of this section be less prescriptive, consistent with providing more flexibility in running a radiation protection program?

Comment. A commenter suggested that the contents of the record for radiation surveys be deleted, consistent with providing the licensee flexibility in developing, maintaining, and implementing its radiation protection program. If this cannot be done, the commenter suggested that the “name of the individual” be changed to “the identity of the individual.”

Response. We simplified the recordkeeping requirements in this section by deleting the requirement to record the location of the survey and the patient identifier. As discussed in Issue 6 of the general comments on this subpart, we believe that the full name of an individual must appear on a record to better ensure future identification of the individual who performed the survey.

Issue 3: Were there any changes made in this section between the proposed and final rules?

Response. Yes. Changes were made in both the title and regulatory text of this section to accommodate changes made in § 35.404, Surveys after source implants and removal. For example, the term “radiation” was struck from the section, recognizing that the survey may not necessarily be a radiation survey. The licensee may also perform a visual survey to locate and account for all sources. Other changes are discussed in the comments on § 35.404.

Section 35.2406, Records of brachytherapy source accountability.

Issue 1: Is it necessary to retain a record of permanent implant sources returned to storage if all sources were used during the implant?

Comment. A commenter suggested that, in some permanent implant cases, all of the sources will be utilized. The commenter proposed that the word “unused” be added to item (c)(2) immediately before “sources.”

Response. We changed the regulatory text in this section to require that the record include “the number and activity of sources not implanted.” Therefore, if all of the sources were used, the licensee would have to note that all of the sources were implanted and, consequently, none were returned to storage.

Issue 2: Were there any changes made in this section between the proposed and final rules?

Response. Yes. The title of this section was changed to correspond to the revised title of § 35.406, Brachytherapy source accountability. That section requires licensees to maintain accountability at all times for all brachytherapy sources in storage or use.

Section 35.2432, Records of calibration measurements of brachytherapy sealed sources.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. The title of this section was changed to correspond to the title of § 35.432, Calibration measurements of brachytherapy sealed sources. That section requires licensees to retain records of calibrations performed before the first medical use of brachytherapy sealed sources. Several changes were also made in this section to accommodate changes made in § 35.432. For example, the proposed rule said that the full calibration measurements must include determination of the output or activity within +/- 5 percent, and the final rule says that a licensee must determine the source output or activity using a dosimetry system that meets the requirements in § 35.630(a). Other changes are discussed in the comments on § 35.432.

Section 35.2433, Records of decay of strontium-90 sources for ophthalmic treatments.

Issue 1: Were there any other changes made in this subpart between the proposed and final rules?

Response. Yes. This section was added to the final rule to correspond with the new § 35.433, Decay of strontium-90 sources for ophthalmic treatments. That section includes a requirement that a record be made of the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments. For additional information, see the discussion for § 35.433.

Section 35.2605, Records of installation, maintenance, adjustment, and repair.

Issue 1: Were there any other changes made in this subpart between the proposed and final rules?

Response. Yes. The word “adjustment” was added to the title and text of this section to conform them with the regulatory text. In addition, the phrase “remote afterloader unit, teletherapy unit, or gamma stereotactic unit” was added. This list of units licensed under Subpart H was added to clarify that other types of devices are not included in Subpart H at this time.

Section 35.2630, Records of dosimetry equipment.

Issue 1: Can the record retention period be changed from “for the duration of the license” to 3 years?

Comment. A commenter suggested that the record retention period could be changed to “three years after the last calibration.”

Response. We have not changed the record retention period in this section. The dosimetry equipment calibrations, intercomparisons, and comparisons performed to show compliance with § 35.630 are necessary to document that the correct radiation dose is delivered to the patient or human research subject. If there is a future question about whether the correct radiation dose was delivered to a patient or human research subject, we believe that these records should be available to document that calibration of the therapy unit was made with properly calibrated instruments.

Issue 2: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. Paragraph (b)(2) was revised to require licensees to include the manufacturer's name for the instruments that are calibrated, intercompared, or compared in accordance with § 35.630. This change is consistent with requirements in other sections to include the manufacturer's name of other types of equipment.

Section 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. Changes were made in this section to incorporate the requirements that were in the proposed §§ 35.2633 and 35.2636, which were deleted. Section 35.2632 in the final rule includes the recordkeeping requirements for full calibrations of teletherapy, remote afterloader, and gamma

stereotactic radiosurgery units. Licensees can refer to this section for all of the recordkeeping requirements for full calibrations of the therapy units covered by Subpart H.

Section 35.2633, Records of remote afterloader full calibrations.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. This section was deleted in the final rule because the requirements were moved to § 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations. This change was made so that all of the recordkeeping requirements for full calibrations of therapy units in Subpart H would be in one place for easier reference for licensees.

Section 35.2635, Records of gamma stereotactic radiosurgery unit full calibrations.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. This section was deleted in the final rule because the requirements were moved to § 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations. This change was made so that all of the recordkeeping requirements for full calibrations of the therapy units covered by Subpart H would be in one place for easier reference for licensees.

Section 35.2643, Records of periodic spot-checks for remote afterloader units.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. Several changes were made in this section to accommodate changes made in § 35.643. For example, the spot-check must assure proper operation of the “timer constancy” in the proposed rule and of the “timer accuracy” in the final rule. Other changes are discussed in the comments on § 35.643.

Section 35.2645, Records of periodic spot-checks for gamma stereotactic radiosurgery units.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. Several changes were made in this section to accommodate changes made in § 35.645. These changes are discussed in the comments on § 35.645.

Section 35.2647, Records of additional technical requirements for mobile remote afterloader units.

Issue 1: Were there any changes made in this section between the proposed and final rules?

Response. Yes. Several changes were made in this section to accommodate changes made in § 35.647. For example, the proposed rule said that a licensee shall arrange for prompt repair of any system that is not operating properly, and the final rule states that if the results of the check indicate a malfunction of any system a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system. Other changes are discussed in the comments on § 35.647.

Section 35.2652, Records of surveys of therapeutic treatment units.

Issue: Can the record retention period be changed to 3 years, instead of “for the duration of use of the unit?”

Comment. A commenter suggested that the record retention period could be changed to 3 years.

Response. We have not changed the record retention period in this section. The surveys performed to show compliance with §35.652 are necessary to ensure that the source/device radiation level limits stated in the Sealed Source and Device Registry are not exceeded. We believe that these surveys should be retained for the duration of use of the device.

SUBPART M - Reports

Issue 1: Should all the reporting requirements be grouped into one subpart or should they be incorporated into the section requiring the report?

Comment. Commenters provided diverse responses to the Commission's question on whether all of the reporting requirements should be grouped into one subpart, or whether they should be incorporated into the individual sections requiring the reports. Commenters favored having all of the reporting requirements in one subpart because this format provides for easy reference, simplifies licensing, and assists licensees in determining their reporting requirements, which makes it easier to maintain compliance. Other commenters favored having the reporting requirements in the individual sections because this format is more orderly and informative. They find the similar separation of the actual reporting requirements and the requirements for what needs to be in the reports in Part 20 to be confusing. A number of individuals have misinterpreted sections of Part 20 simply because of the separation. Several commenters preferred a balanced approach where the reporting requirements would be in the individual sections and all of the requirements summarized in a separate subpart.

Response. After reviewing all of the comments responding to this question, we have concluded that having all of the reporting requirements in one subpart makes it easier for licensees to reference those requirements. The final rule is consistent with the balanced approach because each section in the final rule that is associated with a reporting requirement includes a cross reference to the specific reporting requirements in Subpart M.

Section 35.3045, Report and notification of a medical event.

Issue 1: Do stakeholders think that the term “medical event” is an improvement over the use of the term “misadministration” in the current Part 35?

Comment. Commenters supported the use of the term “medical event.” One commenter agreed with the change, but could see no reason for “candy coating” the term “misadministration.”

Response. We have used the term “medical event” in the final rule because some believe the term “misadministration” has a negative connotation that implies negligence on the part of the physician or other hospital workers. The term “medical event” more correctly and simply conveys that the byproduct material or radiation from byproduct material was not administered as directed by the AU.

Issue 2: Are the reporting requirements for medical events necessary?

Comment. Several commenters said that there was no need for the requirements in this section. Events that result from poor radiation protection practices are covered in the primary regulations for the use of radioactive material, e.g., inadequate survey of a patient following an HDR treatment. If such problem areas in licensees’ programs are brought to their attention, licensees can correct the problems before they result in medical events.

Other commenters expressed concern that the overall wording in this section is subject to a great deal of interpretation and debate over whether specific actions are appropriate for a particular patient and whether an event is a reportable medical event. Therefore, the NRC should develop more specific language describing a medical event in order to avoid intrusion into medical judgments. It should be made clear that medical events are major deviations from a planned treatment that have or could have significant effects on the patient. These effects include either a reduction in the possibility of tumor control or an increase in the possibility of complications. In addition, licensees should be able to appeal to medical experts if NRC staff determines that an incident is a reportable medical event.

Response. We believe that the reporting and notification requirements in this section are necessary so that the NRC is aware of events that trigger the thresholds for medical events to determine what actions, if any, need to be taken to prevent recurrence; so that other licensees can be made aware of generic problems that result in medical events; and so that patients can make timely decisions regarding remedial and prospective health care. The requirements throughout Part 35 are more specific for medical use than the general requirements for the use of radioactive material in the other parts, e.g., Part 20 requirements.

During the development of the final rule, we revisited the proposed wording of all sections, including § 35.3045, to see if we could clarify the regulatory text to avoid future misinterpretations and debates about the meaning of rule language. Such a clarifying change was made to exclude reporting medical events that are due to “patient intervention.”

Issue 3: Are the threshold dose levels for reporting medical events set at appropriate levels [§ 35.3045(a)(1)]?

Comment. Some commenters said that the proposed reporting levels for medical events cannot be justified on the basis of any real risk to either patients or the public. Reporting at these levels implies that these events result in harm to the patient, when they often result in no effect on the patient. Therefore, this is an example of a low risk requirement that the 1997 NAS-IOM Report (Radiation in Medicine: A Need for Regulatory Reform, Institute of Medicine, National Academy Press, Washington, DC, 1997) recommended be deleted. In addition, inherent risks do not justify intrusion by NRC into professional activities and the doctor-patient relationship.

Commenters said that the action level criteria for the total dose delivered from brachytherapy procedures or gamma stereotactic radiosurgery procedures should be revised from the prescribed dose to a level at which harm to patients has been demonstrated. Another commenter questioned why the threshold was not similar to FDA's requirements for reporting morbidity and mortality.

One commenter said that the reporting thresholds of 0.05 Sv (5 rem) effective dose equivalent or 0.5 Sv (50 rem) to an organ or tissue were reasonable levels because they are "reasonably significant radiation exposures." Five rem is the annual limit for a radiation worker, and 50 rem to an organ is the level when one might start seeing organ effects. For example, 50 rem to the testicles will result in a decreased sperm count.

Response. We made no change in the proposed threshold reporting levels for medical events. These reporting levels correspond to the annual dose limits in Part 20 for occupational workers and the level for reporting overexposures of occupational workers to NRC. We believe that applying these same thresholds to reporting exposures to patients is reasonable.

NRC uses the information from the reports of medical events that exceed the dose thresholds to reduce the likelihood of other medical events. For example, information from a report may indicate a breakdown in the licensee's program for ensuring that byproduct material or radiation from byproduct material is administered as directed by the AU or may indicate a generic issue that should be reported to other licensees.

Issue 4: Should licensees be required to report events in which the administration of byproduct material or radiation from byproduct material results in a total dose that differs from the prescribed dose by 20 percent or more?

Comment. Commenters said that the 20 percent difference is arbitrary, and that exceeding this limit presents little or no risk to the patient. The limit should be examined and justified. Recommendations ranged from the limit should be 100 percent, to maybe there should not be a limit and the physician can decide when to report harm to a patient, to it is inappropriate to have a single criterion for all procedures.

Commenters believe that the 20 percent limit is reasonable for external beam therapy and unsealed therapeutic radiopharmaceuticals, but that it is too restrictive for brachytherapy, gamma stereotactic radiosurgery, and unsealed diagnostic dosages. Commenters said that they were aware of clinical

data that supported the 20 percent level for external beam therapy. However, they were unaware of any brachytherapy or gamma stereotactic radiosurgery data demonstrating that a 20 percent difference between the prescribed dose and delivered dose would result in harm to the patient. In addition, a few millimeters in brachytherapy can make a tremendous difference in the dose. Some provision should be made to exempt brachytherapy, or to change the 20 percent limit up to 100-120 percent.

Several commenters questioned the applicability of the 20 percent limit to uses of unsealed byproduct material. Exceeding a radiotherapy dosage by 20 percent may be significant, but reporting an administration of a diagnostic dosage that exceeds the prescribed dosage by 20 percent is overregulation.

Response. We have retained the 20 percent difference that is in the current rule. According to the Statements of Consideration for the Quality Management Program and Misadministrations (56 FR 34104; July 25, 1991), the 20 percent differences are required to be reported because they could possibly indicate a deficiency in the licensee's program, not because they necessarily indicate a significant risk to the patient. We agree with this rationale and see no reason to change the threshold.

Licensees should note that they do not have to report an event in which the total dose or dosage delivered differs from the prescribed dose or dosage by 20 percent or more, unless the dose also differs from the prescribed dose by more than 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin.

NRC uses the information from the reports of medical events where the administration of byproduct material or radiation from byproduct material results in a total dose that differs from the prescribed dose by 20 percent or more to reduce the likelihood of other medical events. For example, the difference between the prescribed and administered doses may indicate a breakdown in the licensee's program for ensuring that byproduct material or radiation from byproduct material is administered as directed by the AU.

Issue 5: Does the proposed rule adequately address wrong treatment site
[§ 35.3045(a)(3)]?

Comment. Commenters both agreed and disagreed on whether the proposed rule adequately addressed wrong treatment site. Two commenters said that it was unclear how wrong treatment site will be handled for therapy, especially for brachytherapy where a medical event can occur if the patient moves even a small distance. In addition, commenters questioned how the wrong treatment site criteria will be applied to permanent seed implants that migrate from the prescribed site.

Another comment was that the criteria for a medical event involving the wrong treatment site must be justified. The criteria of a 0.5 Sv (50 rem) tissue/organ dose and difference of 20 percent from the expected dose defined in the written directive are excessively restrictive. Justification can be provided that the percentage deviation could be 100 percent. At a minimum, radiobiological justification can be made for 1 Sv (100 rem) as a significant threshold. The FDA uses this threshold criteria for evaluating lengthy fluoroscopy studies that could result in skin injury.

Response. In the proposed rule, we attempted to more clearly define when exposure of a wrong treatment site is considered a medical event by including both a 0.5 Sv (50 rem) tissue/organ dose limit and a 20 percent deviation from the expected dose defined in the written directive. We believe that the proposed 0.5 Sv (50 rem) tissue/organ dose limit should be retained, but the allowable deviation from the dose in the written directive should be increased to 50 percent. Therefore, we revised paragraph (a)(3) of this section in the final rule to read “50 percent of the dose expected ...” We believe that this change allows for some variation in doses to the wrong treatment site during administrations of radiation from byproduct material, and requires licensees to only report significant doses to the wrong treatment site due to the movement of the patient or source, e.g., during brachytherapy treatments. In addition, we added a statement that is in the current rule, which was inadvertently not included in the proposed rule, that excludes permanent implants of seeds that were implanted in the correct site, but migrated outside the treatment site.

Issue 6: Does the proposed rule adequately address patient intervention
[§ 35.3045(b)]?

Comments. We received a range of responses to the Commission’s question on whether the proposed rule adequately addressed patient intervention, i.e., actions by the patient such as dislodging or removing treatment devices or prematurely terminating treatment. Several commenters said that this issue was adequately addressed in the rule. Other commenters said that any patient intervention should not result in a medical event. One commenter said that an exemption should be provided to the licensee when the cause of a medical event is patient intervention.

A number of commenters said that the phrase in the proposed rule “that could have been prevented by the licensee” was ambiguous and subjective, and should be deleted because it would result in varying interpretations between NRC and licensees. In addition, decisions on what are considered “reasonable medical practices” for patient control infringe on the practice of medicine and should be left to the physician’s professional judgment. Therefore, this requirement is in violation of Statement 2 of the proposed revision of the Medical Policy Statement: NRC will not intrude into medical judgments affecting patients, except as necessary to provide for the radiation safety of workers and the general public.

Response. As part of the medical use rulemaking, the Commission is codifying a common-sense approach to the reporting requirements for medical events that excludes incidents involving patient intervention. In the proposed rule, the phrase “that could not have been reasonably prevented by the licensee” was added to § 35.3045(a) in an attempt to avoid further expenditure of resources by licensees and NRC in trying to determine what constitutes patient intervention, which is not specifically addressed in the current rule. The issue has involved whether or not a licensee did everything it should to prevent patient intervention during a treatment that resulted in a medical event. Following our evaluation of the comments on patient intervention, we have deleted the proposed phrase from § 35.3045(a) because it did not seem to clarify when an event caused by patient intervention must be reported to NRC as a medical event.

In the final rule, we have addressed when an event caused by patient intervention must be reported to NRC as a medical event. Specifically, we have added a definition of that term to § 35.2. As defined, patient intervention means “actions by the patient or human research subject, whether intentional or unintentional, such as dislodging or removing treatment devices or prematurely terminating the administration.” We believe licensees should

only be required to report serious medical events due to patient intervention. Paragraph (b) of this section in the final rule requires licensees to report any event resulting from intervention of a patient or human research subject in which the administration of byproduct material or radiation from byproduct material results or will result in unintended permanent functional damage to an organ or a physiological system, as determined by a physician. As a result of the significantly higher threshold, the NRC will only receive reports involving patient intervention for events with serious consequences, e.g., unintentional permanent functional damage.

This reporting requirement should result in decreased regulatory burden on licensees because in most situations where patients intervene, either voluntarily or involuntarily, in their treatment, there is no permanent functional damage. Therefore, the revised reporting requirement should significantly reduce the resources expended by the NRC and licensees in debating what are considered reasonable medical practices for patient control because the NRC will no longer require most of the reports it currently receives involving patient intervention. In addition, it should avoid intrusion into medical judgments by the NRC because the decision on whether the administration resulted in permanent functional damage to an organ or a physiological system is to be determined by a physician.

Issue 7: Why do licensees need to notify the NRC by telephone no later than the next calendar day after discovery of a medical event [§ 35.3045(c)]?

Comment. Two commenters questioned the need for licensees to notify the NRC no later than the next calendar day after discovery of a medical event because this requirement implies that these events are harmful or hazardous. There are some medical events with serious consequences that should be reported right away, but there is no benefit in reporting events with no medical significance so promptly.

Response. According to the Statements of Consideration for the Quality Management Program and Misadministration final rule [56 FR 34104; July 25, 1991], misadministrations (medical events) warrant telephone notification of the NRC no later than the next calendar day because these events require that a threshold of either 0.05 Sv (5 rem) effective dose equivalent or 0.5 Sv (50 rem) dose equivalent be exceeded. The early telephone notification allows the NRC to promptly take any necessary actions based on the circumstances, e.g., dispatch an inspector or medical consultant or notify other licensees of potential generic problems. We continue to believe that licensees should promptly notify the NRC of medical events that trigger these thresholds because the circumstances of the medical events need to be evaluated as soon as possible to determine if any immediate follow-up or corrective actions are necessary.

All medical events may not be associated with serious consequences. However, we believe that a requirement that allows for different reporting periods, depending on the initial assessment of the event, would lead to differing interpretations and confusion as to whether the magnitude of the event requires notification of the NRC no later than the next calendar day. In addition, there may be a medical event where the seriousness of the consequences would not be immediately apparent and which, therefore, would not be reported.

Issue 8: Should licensees be required to notify the individual (affected by the medical event) about a medical event [§ 35.3045(e)]?

Comment. We received a range of comments on the requirement in § 35.3045(e) to notify the individual affected by the medical event. These ranged from the licensee should always notify the patient or guardian to this requirement should be deleted.

Some commenters suggested modification of the requirement. For example, a licensee should be allowed not to notify an individual if the rationale for withholding the information is noted in the written report to the NRC. Other suggestions were that notification of the patient should not be required unless the medical event results in a detrimental effect to the patient, or it is necessary to ensure patient safety.

Other commenters said that the requirement should depend on the risk of the procedure. In cases of diagnostic and low-risk therapeutic procedures, notification should not be mandatory. For high-risk therapeutic applications, a patient should only be notified if an adverse outcome is probable and only if the patient's mental state would not be adversely affected.

Commenters provided a number of reasons why they felt that this requirement should be deleted: it overlaps with existing medical practice standards; it intrudes into the practice of medicine; it interferes with the physician-patient relationship; there are no data that patients are not being notified; it presents the appearance of much greater harm than there may actually be; there is no precedent in other areas of medicine; and it is in contradiction to NRC's Medical Policy Statement.

Response. We retained the proposed requirements for notifying individuals following a medical event in the final rule. As stated in the proposed rule (63 FR 43516; August 13, 1998), this position reaffirms statements made by the Commission during the misadministration rulemaking, that patient notification “. . . recognizes the right of individuals to know information about themselves which is contained in records both inside and outside the Federal sector” [“Human Uses of Byproduct Material, Misadministration Reporting Requirements,” (43 FR 2927; May 7, 1978)]. We continue to believe that patient notification enables patients, in consultation with their personal physicians, to make timely decisions regarding any remedial and prospective medical care. This approach also codifies existing medical ethical standards obligating physicians to provide complete and accurate information to their patients.

This approach is consistent with aspects of another Federal patient notification requirement specifically in “The Mammography Quality Standards Reauthorization Act of 1998,” Pub. L. No. 105-248, under which notification of a patient may be required for certain events (e.g., when a patient has received mammography from a facility whose quality is found to be “so inconsistent with quality standards as to present a risk to individual or public health”). [42 U.S.C. § 263b(h)(2)(1999)]. By statute, as well as FDA regulations, a summary of the written report of the patient’s mammography results must be sent directly to the patient if the patient’s physician is not available or if there is no such physician. [42 U.S.C. § 263b(f)(1)(G)(ii)(III); 21 CFR § 900.12(e)(1)(2)(ii)(a) and (iii) (1999).]

Issue 9: Should licensees be required to notify the referring physician about a medical event [§ 35.3045(e)]?

Comment. Several commenters disagreed with the need for a regulation for licensees to notify referring physicians about a medical event.

Nuclear medicine physicians and referring physicians have a professional relationship that would be negatively impacted if the nuclear medicine physician provided inaccurate information or withheld information from the referring physician. Therefore, the NRC does not need to mandate notification of the referring physician.

Response. It is important that a referring physician is aware of medical events involving individuals. The referring physician knows the individual and his or her medical history, and is likely to be in the best position to make a decision about whether informing the individual about the medical event would be harmful. That physician may also need to evaluate any follow-up actions relative to the individual's overall health history. Although notification of referring physicians may represent the "standard of care," that practice may not be uniformly followed. Therefore, we retained the current requirement for a licensee to notify the referring physician about a medical event.

The issue of notifying the referring physician was addressed in the Statements of Consideration for the 1995 rulemaking that amended the medical misadministration requirements ("Medical Misadministration of Radiation and Radioactive Material," 60 FR 48623; September 20, 1995). The Commission noted that "If a misadministration occurs because the material was administered to the wrong individual, there may be no referring physician. If there is no referring physician, the licensee is relieved of the responsibility of notifying the referring physician, but must comply with all other requirements of § 35.33."

Issue 10: Why is there a requirement for a licensee to provide a written report to the individual affected by a medical event [§ 35.3045(f)]?

Comment. We received several comments on the need for a licensee to provide a written report to the individual affected by a medical event. Commenters were concerned that providing a written report to the individual may lead to a misunderstanding of the consequences for the patient and jeopardize the individual's confidence in the ability of the physician providing medical care. Another commenter noted that there is no precedent for providing a written report to a patient about a misadministration of other diagnostic agents.

Response. We have retained the current requirement to furnish an individual affected by a medical event with a written report because it provides the individual a permanent record to refer to for information. Individuals may neither retain nor understand the information that is only orally provided to them. Licensees have the option of providing the individual with a copy of the report that was submitted to NRC, or with a brief description of both the event and the consequences as they may affect the individual. We believe that a written report would be especially useful to an individual who needs to make decisions about any follow-up medical care.

Issue 11: What other changes were made as a result of comments?

Comment. It is not clear whether the thresholds in paragraph (a)(1) and either (a)(1)(i) or (ii) need to occur simultaneously for the event to be reported.

Response. We made editorial changes in the text of paragraph (a) to make it clearer that an event is only classified as a reportable medical event if *both* the threshold in paragraph (a)(1) *and* the threshold for the difference between the total dose and prescribed dose in paragraph (a)(1)(i) *or* the difference between the total dosage and prescribed dosage in paragraph (a)(1)(ii) *or* the difference between the fractionated dose delivered and the prescribed dose in paragraph (a)(1)(iii) have been exceeded.

Comment. The word “of” is missing between “20 percent” (50 percent in the final rule) and “the dose expected” in paragraph (a)(3) of this section that addresses the threshold for determining when a dose to a “wrong treatment site” is a reportable medical event.

Response. The text of paragraph (a)(3) of this section has been corrected to read “50 percent of the dose expected from the administration defined in a written directive.”

Comment. Paragraphs (c)(1)(vi) and (vii) could be combined into one paragraph because they both address actions or improvements that have been taken, or are planned, to prevent recurrence of a medical event.

Response. We combined the requirements in the proposed paragraphs into paragraph (d)(1)(vi) in the final rule.

Issue 12: Were there any other changes made in this section between the proposed and final rule?

Response. Yes. The title of this section was revised to more correctly state that this section includes *both* reporting *and* notification requirements for medical events.

The phrase “results from intervention by a patient or human research subject” in paragraph (a) of the proposed rule was deleted and replaced by “an event that results from patient intervention” in the final rule. This change was made because the definition of patient intervention in § 35.2 includes actions by either a patient or human research subject, so paragraph (a) of the proposed rule contained duplicative language.

The phrase “radiation therefrom” in paragraph (a) of the proposed rule was deleted and replaced by “radiation from byproduct material” in the final rule. This change was made so that there is consistency in the regulatory text in Part 35.

Paragraph (a)(1)(i) of the proposed rule that contained the threshold for the difference between the delivered dose or dosage and the prescribed dose or dosage was split into paragraphs (a)(1)(i) and (ii) in the final rule. This change was made to reflect the fact that physicians can prescribe a range of dosages, but not doses, in written directives.

The word “pharmaceutical” in paragraph (a)(2)(i) was replaced by “radioactive drug containing byproduct material” because the requirements in Part 35 are limited to the medical use of byproduct material.

Paragraph (a)(3) was revised to read “50 percent or more” to make it clearer that the dose to a wrong treatment site has to exceed 50 percent or more of the dose expected from the administration defined in the written directive before a licensee is required to report the event to NRC as a medical event.

Paragraphs (d)(1)(v) and (vi) require that information on the effects of the medical event on the individual who received the administration and on the actions to prevent recurrence be included in the written report to the NRC. These paragraphs have been reworded in the final rule to read “the effect, if any, on the individual;” and “what actions, if any, have been taken, or are planned, to prevent recurrence.” The words “if any” and “are planned” were added because there might not be any effect or any actions taken.

Paragraph (e) [paragraph (d) of the proposed rule] was revised in the final rule. The words “as appropriate” were deleted from the last sentence in paragraph (e) because the intent was covered by the phrase “may be made” in the same sentence.

Paragraph (h) was added to the final rule because the reference to the associated recordkeeping requirements in § 35.2045 was inadvertently omitted in the proposed rule. These records are needed to document these events for licensee and Commission review.

Section 35.3047, Report and notification of a dose to an embryo/fetus or a nursing child.

Issue 1: Should the Abnormal Occurrence Policy Statement criteria for reporting of unintended exposures to an embryo/fetus or nursing child be modified?

Comment. Numerous commenters recommended that § 35.3047 be deleted, and the Abnormal Occurrence (AO) Criteria be revised to reflect the deletion of this section.

Response. Information required by this section is needed so that NRC can comply with Section 208 of the Energy Reorganization Act of 1974 (Public Law 93-438, 5848, 42 U.S.C.), as amended, to submit an annual report to Congress of unscheduled incidents or events which the Commission considers significant from the standpoint of public health and safety, e.g., abnormal occurrences. (The “Reports Elimination Act,” Public Law 104-66, changed the Abnormal Occurrence (AO) report to a yearly publication.)

NRC identifies an abnormal occurrence using the revised abnormal occurrence criteria that were published in the Federal Register (62 FR 18820; April 17, 1997). Section II of that policy statement defines unintended radiation exposure as “any occupational exposure, exposure to the general public, or exposure as a result of a medical misadministration (as defined in § 35.2) involving the wrong individual that exceeds the reporting values established in the regulations.” This section also states that “All other reported medical misadministrations will be considered for reporting as an Abnormal Occurrence under the criteria for medical licensees. In addition, unintended radiation exposures include any exposure to a nursing child, fetus, or embryo as a result of an exposure (other than an occupational exposure to an undeclared pregnant woman) to a nursing mother or pregnant woman above

specified values.” Appendix A, Section I.A.2., “Abnormal Occurrence Criteria,” of the policy statement, states that NRC will provide information on “any unintended radiation exposure to any minor (an individual less than 18 years of age) resulting in an annual total effective dose equivalent of 50 mSv (5 rem) or more, or to an embryo/fetus resulting in a dose equivalent of 50 mSv (5 rem) or more.”

At the present time, NRC has no regulatory requirements that require licensees to report those types of events. Therefore, the Commission considered two alternatives: revise the current Abnormal Occurrence Criteria to delete the requirement to report this type of event to Congress; or develop a reporting requirement for licensees that would provide the information needed by the Commission to comply with Section 208.

After extensive discussion and consideration of the public comments, we have decided to pursue the second option. We are not convinced that it is inappropriate for NRC to report this type of event to Congress and that the reporting requirement in § 35.3047 will be overly burdensome or unwarranted. We are also not inclined to further revise the AO criteria because they have recently been revised and limited comments were received on the proposed criteria.

The thresholds for reporting an unintended dose to an embryo/fetus or a nursing child have been raised in the final rule to the reporting levels in Appendix A, Section I.A.2, of the AO policy statement. Licensees are now required to report any unintended dose to an embryo/fetus that is greater than 50 mSv (5 rem) dose equivalent and any dose to a nursing child that is either greater than 50 mSv (5 rem) effective dose equivalent or results in unintended permanent functional damage to an organ or a physiological system, as determined by a physician. We believe that § 35.3047, as revised in the final rule, provides a balanced resolution of this issue. The regulatory burden on licensees will be substantially less than it would have been under the

proposed § 35.3047 because of the higher reporting thresholds in the final rule; and the NRC will receive the information it needs to report to Congress. In addition, because of the more serious consequences associated with these higher thresholds, we believe that the NRC should receive reports of these unintended doses to an embryo/fetus or nursing child.

Issue 2: What is the impact of the proposed reporting requirement on licensee procedures, activities, or medical practices?

Comment. According to the comments, the biggest impact of the proposed reporting requirement on licensees is associated with the need to determine the pregnancy status of individuals. Commenters had many concerns about NRC's expectations of pregnancy testing, such as delays in emergency scans pending the completion of pregnancy tests; the sensitivity of pregnancy tests; false negative tests in early pregnancy; the age range for pregnancy testing; privacy of minors; patients refusing to pay for pregnancy tests; and the method for calculating conception dates.

Commenters were also concerned about the licensees' responsibilities when they find out later that there was an unintended exposure to a pregnant individual. This can happen if, for example, the patient may not be aware of, or opts to conceal, the fact that she is pregnant. Licensees should not be held responsible for what patients do against medical advice, and reporting such incidents will not prevent a recurrence. Unintended exposures may also occur in cases where the AU is not required to examine the patient, consult with the referring physician, or see the patient's chart, e.g., non-iodine diagnostic studies.

Commenters said that the overwhelming majority of nuclear medicine procedures are safe to perform on pregnant women. In fact, they are often the tests of choice for pregnant women because other radiologic procedures frequently involve higher radiation doses. For the few cases in which administration of a pharmaceutical is not recommended (e.g., sodium iodide I-131), pregnancy information is ascertained. They believe that, by default, the proposed requirement will require pregnancy testing on every female of childbearing age. The inaccuracy, costs, etc. of the tests will lead patients to seek alternate, and often less effective, treatments.

Response. The Commission recognizes that the standard of practice for AUs is to assess the pregnancy or nursing status of their patients (reference ACR “Standard for the Performance of Therapy with Unsealed Radionuclide Sources,” 1996, and “Society of Nuclear Medicine General Procedure Guidelines for Imaging with Radionuclides,” 1997). As a result, we do not believe that it is necessary for the NRC to require a licensee to assess the pregnancy or nursing status of patients before a medical treatment involving byproduct material.

We do believe that it is appropriate to require the licensee to inform the NRC when the licensee learns of an unintended dose to an embryo/fetus or a nursing child that exceeds the thresholds in § 35.3047. For example, a licensee must report an unintended dose resulting from an individual not disclosing her pregnancy or nursing status at the time of administration of the byproduct material or radiation from byproduct material. In this situation, the unintended dose could have been prevented if the AU had followed the standard of practice, noted above, to assess the pregnancy status of the patient. The occurrence of such an incident does not necessarily mean that the licensee is in violation of the requirements in Part 35 as long as the licensee

reports it and it is not otherwise in violation of NRC regulatory requirements. For example, a reportable dose to a nursing child under § 35.3047 is not necessarily subject to enforcement action if the licensee has complied with § 35.75.

However, NRC acknowledges that, in some cases, the licensee might not be able to prevent the dose to an embryo/fetus or nursing child. Such cases are not reportable under § 35.3047. For example, there is no way for an AU to prevent administration of an unintended dose to an embryo/fetus if the pregnancy test was negative because it was given very early in the pregnancy.

Issue 3: What should be the reporting threshold for a dose to an embryo/fetus or a nursing child?

Comment. Commenters said that the proposed reporting level of 5 mSv (500 millirem) to an embryo/fetus or a nursing child is not consistent with the Commission's intent of making Part 35 more risk informed and performance based because it cannot be justified on the basis of risk. This reporting level is also not consistent with the NRC's need to submit an annual report to Congress on unscheduled incidents or events which the Commission considers significant from the standpoint of public health and safety, i.e., abnormal occurrences. One commenter noted that significant biological effects would not be observable at this reporting level in either an embryo/fetus or a nursing child, as demonstrated by the healthy births of children who were exposed to radiopharmaceuticals in utero for the purpose of diagnosing the mothers of these children. The only radiation doses that truly present a significant health and safety issue are those which result in actual non-stochastic effects. Therefore, another commenter suggested that the NRC consider only those medical events which result in actual non-stochastic effects as abnormal occurrences. In addition, one commenter said that there is

no similar requirement by agencies regulating diagnostic x-ray machines. Furthermore, the proposed reporting level is going to result in NRC receiving a number of reports of questionable accuracy and utility.

Commenters suggested a range of reporting levels from 1-25 rem dose equivalent. One commenter suggested that the reporting level should be the same as for medical events: 5 rem total effective dose equivalent or 50 rem to an organ or tissue. Another commenter noted that at his institution, genetic counselors do not consider radiation to be a risk until about 15-20 rem to the embryo/fetus. One commenter suggested that licensees report only radiation-induced injuries and deaths from radiopharmaceuticals and radiologic devices that were due to accidents and that were not reportable to the FDA.

A commenter noted that NCRP Report No. 54, "Medical Radiation Exposure of Pregnant and Potentially Pregnant Women" (1977), states that the risk to the embryo/fetus is negligible below 5 rad and is only significant when compared to other risks of pregnancy above 15 rad. This is consistent with the recommendations in AAPM Radiation Therapy Task Group Report No. 36, "Fetal Dose from Radiotherapy with Photon Beams" (1995).

Commenters also noted that the lack of adequate data makes it virtually impossible to accurately calculate radiation doses to an embryo/fetus at various gestational periods from radiopharmaceuticals. They also questioned how the NRC suggests that patients be monitored to ensure that they are complying with instructions about breast feeding if the nursing child could receive a dose in excess of 100 millirem.

Response. Following an evaluation of the comments and further review of published recommendations and literature, we revised the reporting thresholds in § 35.3047 in the final rule. Paragraph (a) requires that a licensee report to the NRC any administration of byproduct material or radiation from byproduct material to a pregnant woman that results in a dose to an embryo/fetus that is greater than 50 mSv (5 rem) dose equivalent unless the administration was specifically approved, in advance, by the AU. It should be emphasized that only unintended exposures must be reported to the NRC. If a licensee knows that an individual is pregnant and makes the decision that it is necessary to proceed with a test involving the administration of byproduct material or radiation from byproduct material, the licensee would not have to report the dose to the pregnant individual as a medical event. Paragraph (b) requires that a licensee report to NRC any administration of byproduct material to a breast-feeding woman that results in a dose to the nursing child that is greater than 50 mSv (5 rem) total effective dose equivalent or a dose that has resulted in unintended permanent functional damage to an organ or a physiological system, as determined by a physician. These reporting levels are consistent with the recommendations in NCRP Commentary No. 9, “Considerations Regarding the Unintended Radiation Exposure of the Embryo, Fetus or Nursing Child” (1994). At a reporting threshold of 50 mSv (5 rem), there are no detectable deterministic effects, and the risk of stochastic effects (e.g., cancer) is less than 1 percent. This report concluded that “setting requirements for action after radiation exposure of the embryo, fetus, or nursing child at some level below an effective dose of 100 mSv (10 rem) to allow for a margin of safety should enable all such incidents with the potential for harm to be dealt with appropriately.”

We believe that the reporting threshold on the final rule is not overly burdensome on licensees. Unintended doses to an embryo/fetus or nursing child exceeding 50 mSv (5 rem) are rarely encountered in the practice of nuclear medicine (refer, for example, to Russell, J.R., et. al., Radiation Absorbed Dose to the Embryo/Fetus from Radiopharmaceuticals, Health Physics 73:756-769;1997).

Issue 4: Should § 35.3047 include a requirement for a licensee to notify a pregnant individual or mother about an event that must be reported to the NRC in accordance with this section?

Comment. The physician should be able to determine whom to notify. The method and extent of notifying a pregnant individual or mother are solely a matter of the physician's judgment, within the context of the physician-patient relationship. In some cases, the best individual to notify may be the pediatrician (or future pediatrician), which is not an option in the rule. The pediatrician, not the mother's referring physician, will be caring for the infant. The notification requirements in this section are an intrusion into the practice of medicine.

Response. We have retained the requirement for notification of the pregnant individual or mother in the final rule. Although notification of the pregnant individual or mother may represent the "standard of care," that practice may not be uniformly followed. We believe that the pregnant individual or mother should be notified so that she can participate in any decisions on follow-up medical care, if necessary.

Issue 5: Is there a better term than "responsible relative or guardian" that could be applied to those situations where the mother is not notified, e.g., in the referring physician's medical judgment telling the mother would be harmful; the mother is a minor; or the mother is not competent to make decisions regarding medical care?

Comment. Several comments were received in response to the above question, which was published in the proposed rule. Some commenters said that the term “responsible relative or guardian” itself was sufficient, and recommended no alternative wording. The term “guardian” appears to be very clear because the only comment on guardian said that it does not need to be fixed.

We also received several comments on the interpretation of “responsible relative.” Several commenters said that hopefully “responsible” is not used as a substitute for “legal.” The term “responsible” should allow for notification of someone who cares for the minor, but who is neither a blood relative nor a legal guardian. Not telling the mother only because she is a minor is not a responsible rule and is inappropriate. The medical community and the laws of each state determine if a mother is allowed information that may affect her child if she is a minor. The other two situations, it would be harmful to the mother or the mother is not competent, should cover when notification of the responsible relative or guardian is necessary. Another commenter said that for an adult, what is really meant by notifying the “responsible relative” is notifying the relative or individual who has medical power of attorney.

Response. The final rule retains the current phrase “responsible relative or guardian” because we received no suggested terms that better capture the intent of this requirement, which is that *someone* be told in those situations where the mother is not notified. We believe this terminology could include an individual who has medical power of attorney. However, it would be unduly restrictive to limit the individual to be notified, in lieu of the patient, to an individual with medical power of attorney. A physician’s decision on whom to notify is based on many factors, including the Code of Medical Ethics of the American Medical Association and state laws that govern the release of a patient’s medical information to another individual.

To assist with the interpretation of the current notification requirements in the misadministration rule, the Commission had previously provided the examples used in the question of when it expects that a “responsible relative or guardian,” rather than the patient, would be notified about a misadministration. These were provided only as examples, and are not part of the actual regulatory text, e.g., we did not intend by the examples that a mother should not necessarily be notified if she is a minor. We believe that the referring physician should have the discretion to either inform the mother or to determine that, based on medical judgment, telling her would be harmful, in which case the mother’s or child’s responsible relative should be notified.

Issue 6: Why do licensees need to notify the NRC, by telephone, within 5 days and in writing no later than 15 days after discovery of a dose to an embryo/fetus or nursing child that requires a report under this section?

Comment. Commenters questioned the need to notify NRC by telephone within 5 days and in writing no later than 15 days after discovery of a dose to an embryo/fetus or nursing child that requires a report under this section. These reporting requirements give the perception that there is much greater harm than there actually is. One commenter said that the licensee should only have to report in writing to the Regional Office within 30 days after discovery of the dose. The other commenter said that notification of the NRC should be changed from 5 days to 15 days after discovery of the event, or at least changed to 5 working days so there is ample time over a holiday period. The additional time is needed for the licensee to assure the validity of the information in the report.

Response. The final rule contains a significantly higher reporting threshold than the proposed rule for reporting an unintended dose to a nursing child or an embryo/fetus as a result of the unintentional administration of byproduct material or radiation from byproduct material. Licensees are now required to report any dose to an embryo/fetus that is greater than 50 mSv (5 rem) dose equivalent and any dose to a nursing child that is either greater than 50 mSv (5 rem) effective dose equivalent or results in unintended permanent functional damage to an organ or a physiological system, as determined by a physician. More serious consequences are associated with these higher thresholds. Therefore, the reporting requirement in the proposed rule to notify the NRC within 5 days after discovery of the unintended dose has been revised to require notification of the NRC no later than the next calendar day. The reporting requirement in the proposed rule to submit a written report to the NRC Regional Office no later than 15 days after discovery of the dose has also been retained in the final rule. Early telephone notification will allow the NRC to promptly take any necessary actions based on the circumstances, e.g., dispatch a medical consultant. Prompt notification of events that trigger these thresholds is important because the circumstances of the medical event may need to be reviewed as soon as possible to determine if any follow-up actions are necessary.

Issue 7: Were there any other changes made in this section between the proposed and final rules?

Response. Yes. The title of this section was revised to more correctly state that this section includes *both* reporting *and* notification requirements following a dose to an embryo/fetus or nursing child that exceeds the thresholds in § 35.3047.

Paragraph (b)(2) was revised to read "... permanent functional damage to an organ or a physiological system of the child..." to make it clear that this reporting criterion applies to the nursing child.

Paragraphs (d)(1)(v) and (vi) in the final rule have been reworded to read "the effect, if any, on the embryo/fetus or the nursing child;" and "what actions, if any, have been taken, or are planned, to prevent recurrence." The words "if any" and "are planned" were added because there might not be any effect or any actions taken at the time the event is reported. Paragraph (d)(1)(vi) in the proposed rule was deleted because it was duplicative of paragraph (d)(1)(vii).

Paragraph (f), which states that the notification may be made to the mother's or child's responsible relative or guardian instead of the mother, was rewritten to read more clearly.

Paragraph (i) was added to the final rule because the reference to the associated recordkeeping requirements in § 35.2047 was inadvertently omitted in the proposed rule. These records are needed to document these events for licensee and Commission review.

Section 35.3067, Report of a leaking source.

Issue: Where there any changes made in this section between the proposed and final rules?

We changed the title of this section so that it refers to a single report. This change makes the title of this section consistent with the titles of the other sections in Subpart M.

We made this section more performance based by using “the results of the test” instead of the more detailed requirements of “the measured activity of each test sample expressed in microcuries” and “a description of the method used to measure each test sample.” These changes are consistent with changes made in response to comments on § 35.2067, Records of leaking sources.

IV. Summary of Specific Comments on Agreement States Compatibility

Part 1: General Questions

Issue 1: How does NRC determine if a requirement should given a health and safety (H&S) classification?

Comment. Several commenters expressed a concern regarding the compatibility categories, especially those designated as D (H&S). Commenters stated that the (H&S) classification has nothing to do with compatibility but does apply to adequacy of a State’s radiation control program. They further stated that, if the NRC finds it necessary to use this classification, then it should define the “significant safety issues” that led to the (H&S) designation. Other commenters stated that H&S designations for Agreement State requirements is a “back door” to compatibility requirements and may

be unevenly and/or inappropriately enforced. Commenters recommended that if a requirement must be adopted by an Agreement State in order for that State's program to be found "adequate," the requirement should be assigned a "compatibility" designation. H&S designations should be assigned only when a requirement has a direct Part 20 connection.

Response. The assignment of compatibility and health and safety designations to each requirement in the final rule follows a procedure which was developed in an open environment with active involvement of the Agreement States. This procedure has been distributed to all Agreement States and is a publicly available document. Findings of Agreement State program compatibility and adequacy are made by the Management Review Board as part of the Integrated Materials Performance Evaluation Program (IMPEP) review of Agreement State programs. These findings are made based on a number of factors, not just regulations.

The NRC has reviewed and revised, where appropriate, the chart detailing the compatibility categories for each requirement in the final rule. Each requirement has a rationale explaining its safety significance.

Issue 2: What flexibility should be given to Agreement States?

Comment. A commenter stated that Part 35 should not be a matter of compatibility for the Agreement States beyond requiring that states have a system for authorizing the medical use of byproduct material. Another commenter stated that the Agreement States should be allowed to regulate

medical users as appropriate and as needed. They believed that the rule should be a low compatibility issue. Another commenter stated that the proposed Part 35 will deal a death blow to the Agreement State Program by demanding that every Agreement State adopt the essential portions of NRC's new Part 35 under threat of being incompatible and inadequate. The commenter stated that the Agreement States want flexibility. A commenter also expressed that this may cause Agreement States to give back their programs, forcing licensees to pay the NRC instead of their states.

On this same topic, a commenter stated that nearly all of NRC's policy on Agreement State adequacy and compatibility should be rejected. The practices of medicine and pharmacy have no "transboundary implications" and should be changed from compatibility Category "B" to "D" because they are State functions. All compatibility category "C" items should be changed to "D" because they are too restrictive. All "Health and Safety" (H and S) requirements for adequacy should be removed because they are not necessary for "Health and Safety." The commenter further stated that, "Health and Safety" is accomplished by starting with qualified professionals who follow professional standards.

In contrast, commenters stated that a uniform or relatively uniform approach nationwide between Agreement State regulations and NRC regulations can be worked out and can be adopted. In particular, the American Association for Nuclear Cardiology requested that the NRC require the new Part 35 requirements to be at least a level C compatibility for the Agreement States.

Response. The assignment of compatibility categories to each requirement in the rule was made in accordance with Management Directive 5.9, “Adequacy and Compatibility of Agreement State Programs” (Adequacy and Compatibility Policy). The assignment of compatibility categories is needed to assure that byproduct material is used with a minimum level of safety nationwide.

Under the Adequacy and Compatibility Policy, and the program for review of the Agreement States, IMPEP, the States are provided with an ample amount of flexibility in administering their programs. Regulations and other program elements identified as having safety significance may be addressed through the promulgation of compatible regulations or other legally binding documents.

Issue 3: Was the comment period on the proposed rule and on compatibility assignments extended?

Comment. Agreement State representatives commented that the comment period was too brief to allow a comprehensive review of the rule, the licensing guide, and the compatibility listing. They also asked that we provide a listing of essential objectives for each section and why particular designations were assigned. In addition, Agreement State representatives asked that the comment period for the rationale for compatibility assignments should be extended up to 90-days after publication of the listing. They further stated that the degree of flexibility allowed the Agreement States is an important issue and should not be omitted from the discussion because information was not available in a timely manner.

Response. Supplement III of this document contains more detailed discussion of the comments that we received on the length of the comment period. As a result of public comment, we extended the comment period on the proposed rule from November 12, 1999 to December 16, 1999.

The proposed rule contained a brief explanation of the compatibility assignments that were made for the proposed rule. Subsequent to that publication, we received requests from Agreement State representatives to provide supporting documentation for how the assignments were made and to provide the essential objectives for each section. This information was made available to the Agreement States in an All Agreement States Letter. We asked that the States provide comment on the assignments by February 12, 1999.

We considered all comments received on the compatibility assignments and, where appropriate, made changes to either the assignment or to the rationale for the assignment. Supplement IV of this document contains a summary of the assignments. A more detailed compatibility chart which provides the essential objectives for each section and why particular designations were assigned is posted on the NRC Website at <http://www.HSRD.ORNL.GOV\NRC\HOME.HTML>.

Issue 4: How has NRC incorporated comments from the Agreement States on Agreement State issues?

Comment. A commenter questioned how the Agreement States comments were considered during the rulemaking.

Response. In the early stages of the rulemaking process, the NRC established a working group and a steering committee comprised of State personnel and NRC staff. One member of the NRC working group was also a member of the Conference of Radiation Control Program Director's, Inc., SR-6 Committee. This Committee is responsible for revising Part G, "Medical Use of Radionuclides," of the Suggested State Regulations. As such, there was a considerable amount of information exchanged between the States and the NRC staff during the development of the proposed and final rule. In addition, we held two facilitated public workshops during the rulemaking process and received numerous comment letters from the States. All comments received were considered in developing the final rule.

Technical comments and our response to the comments are discussed under the specific section heading. More general comments or comments that pertain exclusively to the compatibility level assigned to the requirement are discussed in this section.

Part 2 - Comments on compatibility designations

We received numerous comments on the compatibility designations assigned to specific sections. The following part provides the comments and our response to the comments. In many cases, but not all, we made changes to the compatibility designation based on the comment.

PART 20 - Standards for Protection Against Radiation

Section 20.1301, Dose limits for individual members of the public.

Comment. A commenter stated that this requirement should not be a compatibility category A. The compatibility category for this requirement should be D.

Response. This section meets the criteria for compatibility category A because it is an NRC program element which is generally applicable and is a dose limit. No rule change is required.

PART 35 - Medical Use of Byproduct Material

Section 35.6, Provision for research involving human subjects.

Comment. A commenter stated that compelling Agreement States to adopt this requirement does not reflect that there may be other criteria affecting human research subjects.

Response. A further review of this section indicates that Agreement States should adopt this requirement in order to avoid a gap in the consistent nationwide application of this Federal policy. The compatibility category was changed from “D” to “C.” [We also added a requirement to the section indicating that nothing in this section relieved licensees from complying with the other requirements in Part 35.]

Section 35.24, Authority and responsibilities for the radiation protection program.

Comment. A commenter stated that this requirement should be classified compatibility category D, not D Health and Safety (H&S). The commenter indicated that, while management should be responsible for the areas identified here, there may be other ways to ensure radiation safety. Further, in the opinion of the commenter, the intent of this requirement will be defeated for small facilities where the AU/RSO is management’s designee.

Response. The D (H&S) compatibility category provides the Agreement States with the flexibility needed to use other methods such as legally binding requirements to achieve the essential objective of this rule. In addition, § 35.24 meets the two failure test criteria for the assignment of compatibility category H&S. This designation provides a minimum level of safety in the implementation of a radiation protection program.

Section 35.40, Written directives.

Comment. A commenter stated that the requirement for a written directive may not be contained in the State's radiation regulations. Another commenter stated that written directives do not meet the definition for a compatibility category C in Subpart A, because it does not create a gap or a duplication. It was also noted that written directives are a compatibility category D (H&S) in Subpart B. Another commenter stated that written directives should not be designated compatibility category H&S and that there are other methods to ensure the right dose is delivered to the right patient (e.g., requiring the physician to be present during a therapy treatment).

Response. We believe that it may be possible to ensure the right dose is delivered to the right patient if a legally binding requirement is in effect and there is some documentation by the physician in the routine radionuclide use log. In accordance with the Policy on Adequacy and Compatibility for Agreement State Programs, legally binding requirements may be acceptable in lieu of a specific regulation on written directives if the essential objectives of this rule are achieved. Section 35.40 meets the two failure test criteria for the assignment of compatibility category H&S. This designation provides a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event.

Section 35.61, Calibration of survey instruments.

Comment. A commenter stated that the requirement in § 35.61 to note the date of the calibration on an instrument should not be a compatibility category H&S. The length of time for record retention is not a compatibility category H&S and should be designated a compatibility category C in all areas of the regulations.

Response. We agree with the commenter. The requirement to note the calibration date on a survey instrument and the record retention requirement have been revised from H&S to a compatibility category D in each section of this rule.

Section 35.63, Determination of dosages of unsealed byproduct material for medical use.

Comment. A commenter stated that there may be some confusion regarding the compatibility category assigned to the requirement covering radiopharmaceutical dosages prepared by the medical use licensee under 10 CFR 35.63 versus those prepared by a commercial pharmacy/manufacturer under 10 CFR 32.72.

Response. Both medical licensees and the commercial preparer of radiopharmaceuticals must determine and record the activity of each dosage intended for medical use. Therefore, this requirement is a compatibility category D (H&S).

Section 35.67, Requirements for possession of sealed sources and brachytherapy sources.

Comment. A commenter stated that paragraph (a) should be a category C. The commenter believed that licensees can develop better procedures and should have the opportunity to submit them for review and approval by the licensing agency.

Response. Section 35.67(a) meets the two failure test criteria for the assignment of compatibility category H&S. This designation assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure.

Comment. A commenter stated that paragraph (f) rather than (e) should be a compatibility category D and paragraph (e) should be a compatibility category D (H&S). Another commenter stated that paragraph (f) which provides a waiver of leak test requirements does not meet the criteria for compatibility category H&S.

Response. Paragraph (e) is a compatibility category D because the technical requirements are already addressed in Part 20 and Part 30 and the actual reporting requirement for leaking sources is contained in § 35.3067 which is a compatibility category C. We agree with the commenter. The compatibility category for this requirement was revised from H&S to D.

Section 35.70, Surveys of ambient radiation exposure rate.

Comment. A commenter questioned the need for a compatibility category H & S for paragraph (b).

Response. We agree with the commenters and have revised this section to indicate that § 35.70(b) is assigned a compatibility category D.

Section 35.75, Release of individuals containing radioactive drugs or implants containing byproduct material.

Comment. A commenter stated that 10 CFR 35.75, which has been assigned a compatibility category C, should be changed to category B due to significant transboundary implications.

Response. The assignment of a compatibility category C to this requirement is appropriate because the term transboundary applies to the use of byproduct material by licensees which operate in multiple locations. The category C designation provides a minimum level of safety, while providing some flexibility to Agreement States to be more restrictive.

Section 35.80, Provisions of mobile medical service.

Comment. A commenter did not agree with the our original basis for designating this section as D compatibility. They disagreed with the following statement: “since there is no potential for medical use of byproduct material in other regulatory jurisdictions under reciprocity” the section is designated a D compatibility.”

Other commenters commented on specific paragraph designations. A commenter stated that paragraph (a)(1) should not be a compatibility category H&S issue. Another commenter stated that paragraph (a)(4) should be compatibility category H&S issue but that the designation was is

inconsistent with the requirements for fixed facilities. (Note: Fixed facilities have to conduct surveys only for procedures requiring a written directive (§ 35.70)).

Response. The Agreement State representatives informed the NRC staff that not all Agreement States authorize mobile services and that there are a number of additional State professional and technical licensing issues which complicate this activity. The medical use of byproduct material (diagnostic or therapeutic) as a mobile service was been designated a compatibility category D for all Agreement States (not required for compatibility) and category D (H&S) for those Agreement States which authorize mobile services. This designation (D (H&S)) assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure.

c.

We agree with the specific comments on paragraphs (a)(1) and (a)(4). The compatibility categories were revised from D (H&S) to D in these sections.

Section 35.92, Decay-in-storage.

Comment. A commenter stated that this section should not be a compatibility category H&S issue. The failure scenario is in error in that it assumes waste would be placed in ordinary trash if storage of isotopes with longer or shorter half-lives were permitted. Permitting decay-in-storage does not mean material that has not decayed would be placed in ordinary trash.

Response. This section is a compatibility category D for those States that choose not to allow the decay-in-storage option. For States allowing this option, the compatibility category is D (H&S). The two or fewer failure test scenario was reworded to better reflect the importance of the D (H&S) assignment for this requirement.

Sections 35.100, Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required and 35.200, Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

Comment. A commenter questioned the assignment of a compatibility category H & S to §§ 35.100 and 35.200 because they are very low risk procedures.

Response. Both requirements meet the two or fewer failure test scenario detailed in Management Directive 5.9. These provisions assist in establishing a minimum level of safety in the medical use of agreement materials by reducing the likelihood of a medical event.

Section 35.390, Training for use of unsealed byproduct material for which a written directive is required.

Comment. A commenter believed that Agreement States should have the option of adopting a higher standards for training even if it means the state would become “incompatible.”

Response. A compatibility category C was assigned to this requirement. This provides an appropriate level of safety while providing some flexibility to Agreement States to be more restrictive.

Section 35.432, Calibration measurements of brachytherapy sealed sources.

Comment. A commenter stated that this requirement should not be a compatibility category C.

Response. This requirement was assigned a compatibility category D (H&S) which provides a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event.

Section 35.604, Surveys of patients and human research subjects treated with a remote afterloader unit.

Comment. A commenter stated that the requirement for after implant surveys is not appropriate for a compatibility category C, since it is a Part 20 requirement.

Response. We agree with this comment and have changed the requirement to a compatibility category D (H&S).

Sections 35.610, Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Comment. A commenter stated that § 35.610 should be compatibility category C, as there can be other ways of meeting the essential objectives.

Response. Section 35.610 meets the two failure test criteria for the assignment of compatibility category H&S. This designation assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure.

Section 35.615, Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Comment. A commenter stated that § 35.615 should be compatibility category C, as there can be other ways of meeting the essential objectives.

Response. Section 35.615 meets the two failure test criteria for the assignment of compatibility category H&S. This designation assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure.

General comments on training.

Comment. A commenter stated that when the Part 35 rulemaking becomes effective, Agreement States that have more strict training and experience requirements for non-board certified physicians will not be able to accept individuals who have met the less restrictive requirements needed to become AUs on NRC licenses as authorized.

Response. When the final Part 35 becomes effective, the Agreement States will have up to 3-years to adopt compatible regulations. The training and experience criteria for physicians is a compatibility category C. The training and experience criteria was given this designation because the criteria relieves undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. Agreement States should adopt these requirements in order to maintain a program compatible with that of the NRC. Non-board certified physicians will continue to be afforded the opportunity to present alternate credentials on a case-by-case basis.

**V. Summary of Changes Made Between the Current Part 35
and the Revised Part 35**

Subpart A, General Information, contains general information regarding medical use of byproduct material.

Section 35.1, Purpose and scope, was revised to specify that the requirements and provisions in Part 35 provide for the radiation safety of workers, the general public, patients, and human research subjects. Inclusion of the phrase “patients, and human research subjects” makes it clear that the provisions of this rule apply to the radiation safety of those individuals. This addition is consistent with the proposed revision of the Medical Use Policy Statement that was published in the Federal Register on August 13, 1998 (63 FR 43580), and will be finalized in a separate Federal Register notice. The section was also revised to add a reference to Part 171, “Annual Fees for Reactor Operating Licenses, and Fuel Cycle Licenses and Materials Licenses, Including Holders of Certificates of Compliance, Registrations, and Quality Assurance Program Approvals and Government Agencies Licensed By NRC.” This revision makes it clear that the provisions in Part 171 apply to medical licensees.

Section 35.2, Definitions, was revised. Definitions were either deleted, revised, or added based on the use of the terms within Part 35. Each category of action will be discussed separately.

DELETED DEFINITIONS:

The following terms were deleted from the rule because they do not appear in Part 35 as presented in the final rule: as low as is reasonably achievable (ALARA), dental use, diagnostic clinical procedures manual, ministerial change, misadministration, podiatric use, recordable event, and teletherapy physicist.

REVISED DEFINITIONS:

The definitions for authorized nuclear pharmacist (ANP) and authorized user (AU) were revised to eliminate the specific board certifications by name and to refer to the specific section(s) in Part 35 containing the requirements the individual must meet to be considered an ANP or an AU. Reference to the specific board certifications was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements. In place of listing the boards, the final rule provides for NRC recognition of the boards. In addition, the definitions of ANP and AU were revised to add a new phrase to the definition to include individuals identified as ANPs or AUs in an equivalent permit or license recognized by NRC. Finally, the definition of “authorized nuclear pharmacist” was revised to recognize nuclear pharmacists who have been approved by a nuclear pharmacy that has been authorized by the Commission to approve ANPs.

The definition for a brachytherapy source was revised. This was done to acknowledge current practices within the radiation oncology field.

The definition of management was revised to recognize an individual having the authority to manage, direct, or administer the licensee's activities who may not have the title of Chief Executive Officer.

The definition of medical use was revised to replace the word "therefrom" with "from."

The definition of mobile nuclear medicine was replaced with a definition for *mobile medical service* because it is a broader term that encompasses all modalities that could be performed by a mobile medical service.

The definition of output was revised to add reference to the exposure rate or dose rate coming from a brachytherapy source, remote afterloader, or a gamma stereotactic radiosurgery unit. The current rule only addresses the output from a teletherapy unit.

The definitions of prescribed dosage and prescribed dose were revised. As revised, the definition of prescribed dosage allows the AU to prescribe a range of activity, without reference to the diagnostic clinical procedures manual. The term unsealed byproduct material in this definition replaces the term radiopharmaceutical. The definition of prescribed dose was revised to add a reference to remote afterloaders.

The definition of Radiation Safety Officer (RSO) was revised to include a reference to the specific requirements an individual must meet in order to be authorized as an RSO. This change makes the definition of RSO consistent with the definitions of authorized nuclear pharmacist (ANP), AU (AU), and authorized medical physicist (AMP). All of these definitions were revised to state that these individuals could be identified on a specific license issued by the Commission or Agreement State license; a permit issued by a Commission master material licensee; a permit issued by a Commission or Agreement State broad scope licensee; or a permit issued by a Commission master material license broad scope permittee that has been given authorization to designate RSOs, AUs, ANPs, and AMPs.

The definition of written directive was revised to delete the provisions for the date the directive was signed and for the signature of the AU before administration of any byproduct material or radiation from byproduct material to a specific patient or human research subject. These were considered to be substantive requirements and were therefore moved to § 35.40, Written directives.

NEW DEFINITIONS

The following definitions were added to the rule either because they are used in Part 35 as presented in the final rule or the public asked that definitions of the terms be added to help clarify the regulatory text. Definitions were added for the following terms: authorized medical physicist, brachytherapy, client's address, gamma stereotactic radiosurgery unit, high dose-rate remote afterloader, low dose-rate remote afterloader, manual brachytherapy, medical event, medium dose-rate remote afterloader, patient intervention, preceptor, pulsed dose-rate remote afterloader, radioactive

drug, Sealed Source Device Registry, stereotactic radiosurgery, structured educational program, teletherapy, temporary jobsite, therapeutic dosage, therapeutic dose, treatment site, type of use, and unit dosage.

Section 35.5, Maintenance of records, was revised to insert “and” in the current phrase “drawings and specifications.”

Section 35.6, Provisions for research involving human subjects, was retitled as Provisions for the protection of human research subjects. In addition, this section was restructured to make it easier to read. An introductory paragraph was added to make it clear that research permitted under § 35.6 may only be performed using byproduct material that is already authorized for medical use by the license. For example, if a licensee is authorized to use byproduct material under §§ 35.100, 35.200, and 35.300, it cannot conduct research using a remote afterloader. However, the same licensee can conduct research using materials authorized under §§ 35.100, 35.200, or 35.300. and to add paragraphs (b) and (c).

Paragraph (d) was added to codify the Commission's intent that § 35.6 does not relieve licensees from complying with other provisions in Part 35 and that all relevant radiation safety provisions of Part 35 are applicable to research involving human subjects. This position is further discussed in the regulatory history of § 35.6. For further information on this issue, refer to the Federal Register of December 2, 1994 (59 FR 61767).

Section 35.7, FDA, other Federal, and State requirements, was reworded but there is no change in the requirement.

Section 35.8, Information collection requirements; OMB approval, was revised to reflect the renumbering of some sections within the rule and the additional recordkeeping and reporting sections in the new rule.

Section 35.10, Implementation, is a new section that discusses the provisions for implementing the final rule. A detailed discussion of the implementation provisions can be found in Section IX of the Supplementary Information section of this document. This section replaces the current § 35.999, Resolution of conflicting requirements during transition period.

Section 35.11, License required, was revised. Paragraph (a) was revised to more clearly state that a person may manufacture, produce, acquire, receive, possess, use, or transfer byproduct material for medical use only in accordance with a specific license issued by the Commission or an Agreement State or as allowed in paragraphs (b)(1) and (b)(2) of this section. Paragraphs (b)(1) and (b)(2) were revised to reflect that the requirements for supervision in the current § 35.25 were replaced by the requirements in the final § 35.27.

Section 35.12, Application for license, amendment, or renewal, was revised.

Paragraph (a) was revised to state that any application for a license, amendment, or renewal must be signed by the applicant's or licensee's management. The current rule indicates that any person may apply if the application is for medical use not sited in a medical institution and that only

management may apply for a license if the application is for use in a medical institution. We believe it is important that management apply for a license, regardless of where the byproduct material is used, because NRC holds the licensee responsible for any actions of its employees.

Paragraph (b) was revised to address license applications for uses authorized under §§ 35.600 and 35.1000. Therefore, the current paragraph (c) was no longer needed and was deleted. We no longer require licensees to have separate licenses for teletherapy or gamma stereotactic radiosurgery units. In addition, paragraph (b) was revised to list the items that must be submitted to NRC in support of a license application. The new paragraph (c) provides a list of the items that must be submitted to NRC in support of a license amendment. The lists in paragraphs (b) and (c) codify existing licensing practices. Finally, paragraphs (b) and (c) were revised to delete the reference to the regulatory guides. Guidance for completing an application is in NUREG-1556, Vol. 9, “Consolidated Guidance About Materials Licenses, Program-Specific Guidance about Medical Use Licenses.” NUREG-1556, Vol 9, is available for inspection at the NRC Public Document Room, 2120 L Street NW. (Lower Level), Washington, DC.

The statement in the current paragraph (d) that referenced where to find copies of regulatory guides, application forms, or where to submit an application or an amendment request was deleted. This information is not needed in the regulation. The new paragraph (d) addresses applications for medical use of byproduct material that are not specifically included in Subparts D through H of the proposed rule, referred to as “emerging technologies.” The current rule does not address emerging technologies and therefore does not provide for efficient licensing of emerging technologies. Paragraph (d) provides a list of the information needed by NRC to approve a license or license amendment for a use not specifically addressed in Subparts D through H of the new rule.

Section 35.13, License amendments, was revised . Paragraph (a) was revised to reference “type of use” rather than “clinical procedure.” The term “type of use” is defined in Part 35 and is more appropriate for use in this requirement. Paragraph (b) was revised to delete the reference to an AMP. A medical use licensee is no longer required to amend its license before allowing anyone to work as an ANP if that individual meets the training and experience requirements in § 35.51(a), and the training and experience requirements were met within the 7 years preceding the date of the application. In addition, paragraphs (a) and (b) were reworded to clearly indicate the subject of each paragraph.

Paragraph (c) was revised to delete the requirement for a licensee to apply for a license amendment if the teletherapy physicist changes, provided the individual meets the requirements in §§ 35.51(a) and 35.59. This change is consistent with licensing requirements for AUs and AMPs.

The Commission recognizes that unusual conditions may arise when the RSO leaves a licensee with little to no advance warning. In this event, the licensee may want to consider using an AU or other individual qualified to be an RSO to fill the position, pending appointment of a new RSO. Under these conditions, the licensee must move expeditiously to permanently fill the position of RSO and should contact the appropriate NRC regional office and explain the situation.

Paragraph (d) was revised to require the licensee to apply for and receive a license amendment before it receives byproduct material in excess of the amount or in a different form or it receives a different radionuclide than is authorized on the license. This change was made to clarify that the requirement is tied to a licensee’s authorization to possess, not order, byproduct material and to clarify when an amendment is needed. For example, if a

license authorizes possession of any byproduct material identified in §§ 35.100, 35.200, and 35.300, in any chemical and/or physical form, a licensee would be required to obtain a license amendment if it wanted to possess sealed sources for manual brachytherapy (§ 35.400). This same licensee would not need to amend its license if it wanted to use I-131 sodium iodide for thyroid carcinoma, since that use is authorized by § 35.300. Further, no amendment would be required if the licensee wanted to use Tc-99m labeled methylene diphosphonate (MDP) rather than Tc-99m labeled sestamib because such uses are authorized by § 35.200.

To reduce regulatory burden, paragraph (e) was revised to delete the requirement for a licensee to apply for a license amendment if there is a change in the areas where byproduct material is used under §§ 35.100 and 35.200. In addition, the requirement in the current paragraph (e) for a licensee to apply for an amendment before it changes the address(es) of use identified in the application or on the license was moved to the final paragraph (f).

Section 35.14, Notifications, was revised. Paragraph (a) was revised to include a requirement for the licensee to notify NRC no later than 30 days after the date the licensee permits an individual to work as an AMP under § 35.13(b). This change was needed because we would like to be notified when an AMP who has been approved by the licensee begins work. (Reference change made to § 35.13(b)). Paragraph (b) was revised to require that the licensee notify NRC when an AMP permanently discontinues performance of duties under the license. Paragraph (b) was also revised to require that a licensee notify NRC when the licensee changes its name. This provision applies only if there is no change in ownership, as described in § 30.34 of this chapter. If there is a change in ownership, the licensee must take appropriate action to have its license amended. A requirement was also added to

paragraph (b) for a licensee to notify NRC of any changes in areas where byproduct material is used in accordance with §§ 35.100 and 35.200. These revisions to the requirements for notifications were warranted because of the associated revisions to the requirements for license amendments in § 35.13.

Section 35.15, Exemptions regarding Type A specific licenses of broad scope, was revised to add the term “authorized medical physicist” to paragraph (e). This revision is needed because under the revised requirements in § 35.13 broad scope licensees have the authority to appoint AUs, ANPs, or AMPs without applying for a license amendment, if the individuals meet the approved criteria in Subparts B and D through H.

A new paragraph (f) was added to also exempt these broad scope licensees from § 35.49(a). This change codifies in the regulations an exemption currently provided to these licensees through a standard license condition. NRC’s medical use licensees with a Type A specific license of broad scope currently receive a standard license condition that exempts the licensee from only receiving sealed sources or devices manufactured from licensees with medical distribution licenses issued in accordance with § 32.74. This change would replace the license condition.

Section 35.18, License issuance, was revised. Paragraph (a) lists the conditions that must be met in order for the Commission to issue a license. Requirements for a mobile medical service license were added as paragraph (b). The NRC will issue a license for mobile service if the applicant meets the requirements specified in paragraph (a) of the section and if the individual or human research subject to whom the applicant administers byproduct

material, or radiation from byproduct material, may be released following treatment in accordance with § 35.75. The later provision is necessary because mobile service licensees will not have the capability of controlling individuals who cannot be released under § 35.75.

Section 35.19, Specific exemptions, was revised to delete the statement that the Commission will review requests for exemptions from training and experience requirements with the assistance of its Advisory Committee on the Medical Uses of Isotopes (ACMUI). This statement is a matter of Commission policy rather than a regulatory requirement.

Subpart B, General Administrative Requirements, contains the general administrative requirements regarding medical use of byproduct material.

Section 35.20, ALARA program, was deleted in its entirety. ALARA is discussed in § 20.1101, Radiation protection programs, and medical licensees must comply with the requirements of that section. That section requires, in part, that a licensee develop, document, and implement a radiation protection program and use, to the extent practicable, procedures and engineering controls to achieve occupational doses and doses to members of the public ALARA. Therefore, we do not believe that § 35.20 is needed in light of the requirements in § 20.1101. A medical use licensee should have flexibility in developing, maintaining, and implementing a radiation protection program that meets the requirements of Part 20.

Section 35.21, Radiation Safety Officer, was deleted in its entirety. The requirements in paragraph (a) were moved to § 35.24. The list of the RSO's duties in paragraph (b) was deleted because it is overly prescriptive and in some cases overlaps with the requirements in § 20.1101. We believe

that the licensee should have flexibility in developing, maintaining, and implementing its radiation protection program, including establishing the RSO's duties.

Section 35.22, Radiation Safety Committee, was deleted in its entirety. The issue of whether the NRC should require a Radiation Safety Committee (RSC) was identified as a cross-cutting issue and, therefore, was discussed at public meetings throughout the rulemaking process. Comments received on this topic are discussed in Section III of the Supplementary Information section of this document. The basic requirement for certain medical licensees to have an RSC to oversee all uses of byproduct material permitted by the license was moved to § 35.24. However, the requirement was modified, so that only licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F and H, or two or more types of units under Subpart H, are required to establish an RSC. Several other requirements that are currently in § 35.22 were also moved to § 35.24 and are discussed under that section. However, most of the requirements that are currently in § 35.22 have been deleted to provide licensees with more flexibility in how they use the Committee to oversee the radiation safety aspects of the medical use of byproduct material.

Section 35.23, Statements of authority and responsibilities, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.24.

Section 35.24, Authority and responsibilities for the radiation protection program, is a new section. A number of the current, prescriptive requirements associated with the radiation protection program have been deleted to provide licensees more flexibility in achieving the objective of radiation safety.

Paragraph (a) requires licensee management to approve, in writing, licensing actions; individuals before allowing them to work as an AU, ANP, or AMP; and radiation protection program changes that do not require a license amendment and are permitted under § 35.26. We believe that licensee management should be responsible for these approvals, as part of their overall responsibility for the radiation protection program. This is a change from the current § 35.22, which gives the RSC the responsibility for two of these approvals: approval of individuals before allowing them to work as an RSO, AU, ANP, or AMP; and approval of radiation protection program changes that do not require a license amendment.

The requirement in paragraph (b) to appoint an RSO is currently required by § 35.21. Paragraph (b) also includes a new requirement that the RSO agree, in writing, to be responsible for implementing the radiation protection program. The requirements in paragraphs (e) and (g), associated with the authorities, duties, and responsibilities of the RSO, are similar to the requirements in the current § 35.23.

Paragraph (c) includes a new provision that allows a licensee to have a temporary RSO for up to 60 days a year if the licensee meets the requirements for RSOs in paragraphs (b), (e), (g), and (h) of this section. This new provision was added so that licensees can appoint someone to fulfill the duties and responsibilities of the RSO in a timely manner, following the sudden departure of the permanent RSO named on the license. Paragraph (d)

allows a licensee to simultaneously appoint more than one temporary RSO, if needed, to ensure that the licensee has an individual that is qualified to be an RSO for each of the different types and uses of byproduct material permitted by the license.

Paragraph (f) contains a requirement for certain medical licensees to have an RSC to oversee all the uses of byproduct material permitted by the license. The current requirement in § 35.22 was modified so that only licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F, and H, or two or more types of units under Subpart H, are required to establish an RSC. For example, licensees that are permitted on their license to use therapeutic quantities of unsealed byproduct material (§ 35.300) and manual brachytherapy (§ 35.400), or manual brachytherapy (§ 35.400) and low dose-rate remote afterloaders (§ 35.600), or teletherapy (§ 34.600) and gamma stereotactic radiosurgery (§ 35.600) would be required to have an RSC. However, we believe that many other medical licensees will also continue to use an RSC to oversee the use of byproduct material. Licensees should note that the requirement for an RSC is no longer tied to medical institutions, which means that it now also applies to free-standing clinics.

The new requirement is much less prescriptive than the requirements in the current § 35.22. For example, paragraph (f) does not include the list of administrative requirements and committee tasks that are specified in the current rule. However, based on public comment, we have specified that the membership of the committee should include an AU of each type of use permitted by the license, the RSO, a representative of the nursing service, a representative of management who is neither an AU nor an RSO, and other members the licensee considers appropriate.

Paragraph (h) requires that the licensee retain a record of management's approval of actions in paragraph (a); written acceptance of RSO duties as specified in paragraph (b); and the duties, responsibilities, and authority of the RSO specified in paragraph (e) in accordance with § 35.2024, Records of authority and responsibilities for radiation protection programs.

Section 35.25, Supervision, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.27. The requirements in paragraphs (a)(3) and (b)(3) for periodic reviews of the work of supervised individuals were deleted and not moved to § 35.27. We believe that these requirements are too prescriptive. Licensees should have flexibility in how they evaluate supervised individuals because they are held responsible for their acts and omissions.

Section 35.26, Radiation protection program changes, is a new section. The requirements in this section are similar to the requirements in the current § 35.31, which was deleted. This section allows licensees to revise their radiation protection programs without Commission approval if the revision does not require an amendment in accordance with § 35.13; if the revision is in compliance with the regulations and license; if the change has been reviewed and approved by the RSO, and reviewed and approved in writing by licensee management; and if the affected individuals have been instructed on the revised program before the changes are implemented. This requirement provides licensees with flexibility to manage their radiation protection programs and clearly defines the situations that will not require Commission approval of an amendment to their license. We believe that many licensees were reluctant to make changes to their current program because the term "ministerial changes," as defined in the current § 35.2 and as used in

the current § 35.31, was not clearly understood. This change is intended to provide clear guidance to licensees on when they can revise their radiation protection programs without obtaining Commission approval.

We believe that it is important to instruct individuals in program changes, including those permitted under § 35.26, before they are implemented. This instruction may be provided in writing or orally and may be conducted on an informal or formal basis. It is not necessary to document that this instruction has been provided to affected parties, because these changes should not reduce radiation safety. At the time of inspection, NRC inspectors may question whether this instruction was provided.

Section 35.27, Supervision, is a new section. The requirements in this section are similar to the requirements in the current § 35.25, which were deleted. Paragraphs (a)(1) and (b)(1) were revised to delete the requirement to instruct individuals in the principles of radiation safety. This type of instruction is adequately addressed by § 19.12, Instructions to workers, of this chapter. Paragraph (a)(1) was also revised to require that, in addition to the requirements in § 19.12, the licensee shall instruct supervised individuals in the written radiation protection procedures, written directives procedures, regulations of this chapter, and license conditions. Paragraph (a)(2) was revised to require supervised individuals to follow the instructions of the supervising AU for medical uses of byproduct material, written radiation protection procedures, regulations, and license conditions with respect to the medical use of byproduct material. Paragraphs (a)(3) and (b)(3) of the current § 35.25 were deleted because the licensee should have flexibility in evaluating employee performance. Paragraph (b)(2) was revised to require supervised individuals to follow the instructions of the supervising AU or ANP regarding the preparation of byproduct material for medical use, the written radiation protection procedures, and the regulations of this chapter and license

conditions. The statement in paragraph (c) that licensees are responsible for the acts and omissions of supervised individuals is similar to the statement in the current § 35.25(c).

Section 35.29, Administrative requirements that apply to the provision of mobile service, was deleted in its entirety. The conditions for the Commission to issue a mobile medical service license were moved to § 35.18. The requirements in paragraphs (b) and (d) were moved to § 35.80. Paragraph (c) was deleted because this requirement, which addressed the client's responsibilities, was viewed as being overly prescriptive. Mobile medical service licensees are required to comply with all the provisions of the license that authorize the use, possession, and transfer of material.

Section 35.31, Radiation safety program changes, was deleted in its entirety. The requirements, with some modifications, were moved to § 35.26 so that all the requirements pertaining to management of the licensee's radiation protection program appear in one area of Subpart B.

Section 35.32, Quality management program, was deleted in its entirety. The issue of whether the Commission should continue to require that a licensee develop, implement, and maintain a quality management program was identified as a cross-cutting issue and was discussed at public meetings throughout the rulemaking. Comments received on this topic are discussed in Section III of the Supplementary Information section of this document. Based on these comments, the Commission deleted the requirements for a quality management program. However, the Commission believes there are three elements of the current quality management program that should continue to be addressed in the rule: confirming patient identity, requiring written directives, and verifying dose. The requirements for these three elements are in §§ 35.40 and 35.41. However, we believe that licensees will continue to

implement other elements of the current quality management program, as part of the “standard of care” in medicine. In this regard, the Commission acknowledges that other factors, such as accreditation, have resulted in medical institutions adopting programs similar to those specified in the current rule.

Section 35.33, Notifications, reports, and records of misadministrations, was deleted. The recordkeeping and reporting requirements were moved to Subparts L and M, respectively.

Section 35.40, Written directives, is a new section. This section contains requirements for preparation of written directives that are similar to the requirements in the current §§ 35.2 and 35.32. Written directives are no longer required for administrations of I-125 sodium iodide because I-131 sodium iodide is primarily used now. Based on public comments and discussions with the ACMUI, changes were made in the information that must be included in written directives. For gamma stereotactic radiosurgery, the requirements for target coordinates, collimator size, plug pattern, and total dose have been deleted, and requirements for total dose, treatment site, and number of target coordinate settings per treatment for each anatomically distinct treatment site have been added. For teletherapy, the requirement for overall treatment period has been deleted and a requirement for number of fractions has been added. For high dose-rate remote afterloading brachytherapy, requirements have been added for the dose per fraction and the number of fractions. For all other brachytherapy, before implantation, the requirements for number of sources and source strengths have been deleted and requirements for treatment site and dose have been added; and after implantation, but before completion of the procedure, a requirement for the number of sources has been added. Licensees should refer to § 35.41 for the requirements for procedures for administrations requiring written directives.

Section 35.41, Procedures for administrations requiring a written directive, is a new section. Paragraph (a) of this section requires licensees to develop, implement, and maintain written procedures to assure that, before each administration, the patient's or human research subject's identity is verified and that each administration is in accordance with the written directive. The specific details to be included in the written directives are in § 35.40. Paragraph (b) of this section specifies the items that must, at a minimum, be addressed in the procedures. The items identified in § 35.41 are viewed by the Commission as key elements of a program that will provide high confidence that byproduct material will be administered as directed by the AU. However, the regulations are not prescriptive about how these objectives are met, allowing licensees the flexibility to develop procedures to meet their needs. This section includes no requirement for submittal or approval of the procedures, as was previously required by the quality management rule.

Section 35.49, Suppliers for sealed sources or devices for medical use, has been retained, with one modification. We have revised paragraph (a) to make it clear that it only applies to the initial manufacturing, labeling, packaging, and distribution of a sealed source or device. Subsequent distribution of the sealed sources or devices is not subject to the requirements in this paragraph if they are distributed to licensees that have a license to possess the source or device. Currently, licensees must obtain an amendment exempting them from the requirements following initial distribution of the sealed source or device.

Section 35.50, Training for Radiation Safety Officer, is a new section. The training and experience requirements for an RSO were moved, with some modifications, from the current § 35.900, Radiation Safety Officer. Two changes made in the new section should be noted. First, the listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats

automatically fulfill the training and experience requirements for RSOs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an RSO. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

The current § 35.50, Possession, use, calibration and check of dose calibrators, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.60.

Section 35.51, Training for an authorized medical physicist, is a new section. The training and experience requirements for an AMP were moved, with some modifications, from the current § 35.961, Training for an authorized medical physicist. Three changes made in the new section should be noted. First, the title of this section was revised because the training and experience requirements in this section now apply to AMPs, rather than just teletherapy physicists, because requirements for gamma stereotactic radiosurgery units and remote afterloader units have been codified in the revised Part 35. Second, the listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AMPs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AMP. Section

III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

The current § 35.51, Calibration and check of survey instruments, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.61.

Section 35.52, Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.60.

Section 35.53, Measurements of dosages of unsealed byproduct material for medical use, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.63.

Section 35.55, Training for an authorized nuclear pharmacist, is a new section. The training and experience requirements for an ANP were moved, with some modifications, from the current § 35.980, Training for an authorized nuclear pharmacist. One change made in the new section should be noted. The listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for ANPs. In place of listing the boards, the final rule provides for NRC

recognition of the boards. Section III of the Supplementary Information section of this document contains a detailed discussion of the new training and experience requirements in Part 35.

Section 35.57, Training for an experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist, is a new section that replaces the current requirements in §§ 35.901, 35.970, and 35.981, which were deleted. All individuals who are identified as RSOs, teletherapy or medical physicists, AUs, and ANPs, on an NRC or Agreement State license issued before the effective date of the final rule, do not need to comply with the new training and experience requirements.

The current § 35.57, Authorization for calibration and reference sources, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.65.

Section 35.59, Recentness of training, is a new section that replaces the current requirements in § 35.972. Although this is not a new requirement, questions have recently been raised regarding whether all elements of the requirements must have been obtained in the last 7 years. It is expected that (1) either the individual has been board certified, or has completed the training specified in the alternative pathway within the 7 years preceding the date of the application; or that (2) the individual has had related continuing education and experience since completing the required training and experience requirements. Continuing education and experience requirements are reviewed on a case-by-case basis, with input from the ACMUI, as necessary. The

text in the current § 35.972 was revised to reference Subparts B, D, E, F, G, and H because the revised training and experience requirements appear in the subparts with their associated modality.

The current § 35.59, Requirements for possession of sealed sources and brachytherapy sources, was deleted in its entirety. The requirements in this section, with some modifications, were moved to § 35.67.

Subpart C, General Technical Requirements, contains general technical requirements regarding medical use of byproduct material.

Section 35.60, Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material, is a new section that replaces the current §§ 35.50 and 35.52. This section addresses calibration of all instruments used to measure the activity of all unsealed byproduct materials, rather than only dose calibrators used to measure the activity of dosages of photon-emitting radionuclides (§ 35.50) or instruments used to measure dosages of alpha- or beta-emitting radionuclides (§ 35.52). The change recognizes that there are various types of instruments that can be used to measure the activity of unsealed byproduct materials. This change also gives licensees flexibility in developing a calibration program which meets their program needs.

The prescriptive calibration requirements in the current §§ 35.50 and 35.52 were deleted. Paragraph (b) in the final rule requires that licensees calibrate the instrumentation in accordance with nationally recognized standards (e.g., voluntary consensus standards, such as ANSI N42.13-1986 (R

1993), “Calibration and Usage of Dose Calibrator Ionization Chambers for the Assay of Radionuclides”) or with manufacturer’s instructions. This change makes the regulation more flexible, more adaptable to new technology, and more performance-based.

Licensees should note that they are required by § 35.63 to determine the activity of each dosage before medical use. If they use only unit dosages of radioactive drugs that meet the definition in § 35.2, then § 35.63 allows the licensee to determine the dosage by a decay correction based on the activity or activity concentration determined by either a manufacturer or preparer licensed pursuant to § 32.72 or equivalent Agreement State requirements; or an NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an Investigational New Drug (IND) protocol accepted by the FDA. If a licensee chooses to determine the dosage using this method, a licensee would not be required to possess instrumentation to measure the activity of the dosage, i.e., the licensee would not be required to comply with § 35.60. However, if a licensee chooses to reassay a unit dosage for the purpose of adjusting the activity, it would no longer be considered a unit dosage once it was altered, and the licensee must comply with this section. This requirement is appropriate because confirmation of a dosage, or adjustment of dosages, must be based on properly-calibrated equipment.

The recordkeeping requirements for this section are in § 35.2060, Records of calibrations of instruments used to measure the activity of unsealed byproduct material.

Requirements in the current § 35.60, with minor modifications, were moved to § 35.69.

Section 35.61, Calibration of survey instruments, is a new section that replaces the current § 35.51. The requirements in the current § 35.51 to note the apparent exposure rate from a dedicated check source, as determined at the time of calibration; to attach a correction chart or graph to the instrument; and to check each survey instrument for proper operation with a dedicated check source each day of use were deleted. These changes give the licensee greater flexibility in calibrating instruments.

Paragraph (a) in the new § 35.61 now requires the licensee to calibrate survey instruments used to show compliance with this part and with Part 20 before first use, annually, and following repair that would effect the calibration. Paragraph (b) requires that survey instruments be removed from use if the indicated exposure rate differs from the calculated exposure rate by more than 20 percent. Previously, there was no threshold for removing instruments from use. The requirements in this section are generally consistent with ANSI N323-1978 (R 1993), "Radiation Protection Instrumentation Test and Calibration."

The recordkeeping requirements for this section are in § 35.2061, Records of radiation survey instrument calibrations.

The requirements in the current § 35.61, with minor modifications, were moved to the final § 35.69.

Section 35.63, Determination of dosages of unsealed byproduct material for medical use, is a new section that replaces the current § 35.53. This section requires licensees to determine and record the activity of each dosage before medical use. For unit dosages as defined in § 35.2, paragraph (b)

allows the licensee to determine the dosage by a decay correction based on the activity or activity concentration determined by either a manufacturer or preparer licensed pursuant to § 32.72 or equivalent Agreement State requirements; or an NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an Investigational New Drug (IND) protocol accepted by the FDA.

Because the unit dosages have been assayed by the Part 32 licensee or by a licensee for use in research in accordance with an RDRC-approved protocol or an IND protocol accepted by FDA, we do not believe the Part 35 licensee should be required to reassay the dosage. Licensees should note that if a unit dosage is changed or manipulated in any way it is no longer considered to be a unit dosage and will need to be reassayed before it is administered.

For other than unit doses, paragraph (c) allows the licensee to determine the dosage by direct measurement of radioactivity; combination of direct measurement of radioactivity and mathematical calculations; or by combination of volumetric measurements and mathematical calculations based on the measurement made by a manufacturer or preparer licensed pursuant to § 32.72 or an equivalent Agreement State requirement. The current rule limited the licensee to using direct measurement for determining the activity of a photon-emitting radionuclide, but allowed alpha- or beta-emitting radionuclides to be measured either by direct measurement or by combination of measurements and calculations. This change allows licensees flexibility in determining dosages and does not distinguish between the type of the radiation (e.g., alpha, beta, or photon) and the way the determination is made.

Paragraph (d) permits a licensee to use a dosage if the activity does not differ from the prescribed dosage by more than 20 percent or if the activity falls within the prescribed dosage range. We believe that the rule should allow for some deviation from the prescribed dosage, if the licensee chooses to prescribe a dosage rather than a dosage range. Without this allowed deviation, the administered dosage would need to match the prescribed dosage.

We have not allowed a deviation outside of the prescribed range because we believe that allowing the AU to establish a dosage range provides the AU with the needed flexibility. The final paragraph (d) codifies requirements that are currently imposed on licensees by license conditions and provides guidance regarding allowed deviations for a dosage range. This does not prevent an AU from revising the prescribed dosage at any time prior to the administration.

The recordkeeping requirements for this section would appear in § 35.2063, Records of dosages of unsealed byproduct material for medical use.

Section 35.65, Authorization for calibration, transmission, and reference sources, is a new section that replaces the current § 35.57. In paragraphs (b) and (c) of the final rule, the references in the current § 35.57 to §§ 35.100 and 35.200 were deleted because specific radionuclides were not listed in these sections. Paragraph (a) has also been revised to reference § 32.72, Manufacture and distribution of sources or devices containing byproduct material for medical use. This was done because Part 35 licensees should be allowed to receive calibration and reference sources distributed by licensees under § 32.72. Paragraph (b) was revised to allow possession of calibration and reference sources with half-lives of less than 120 days. The current section only allows possession of sources with half-lives less than 100 days. This change was done so that the section would be consistent with the financial assurance regulations in Part 30. Paragraph (c) was revised to allow possession of any byproduct material with a half-life longer than 120 days in individual amounts that do not exceed the smaller of the following two values: 7.4 Megabecquerels (MBq) (200 uCi) or 1000 times the quantities in Appendix B of Part 30. This change was made to limit the possession activity below the level where financial assurance is required. In paragraph (d),

the possession limit for Tc-99m was deleted. The Commission believes that it is not necessary to limit the possession of Tc-99m for calibration and reference sources because there are no possession limits for Tc-99m associated with the use of Tc-99m under § 35.100 or § 35.200.

Section 35.67, Requirements for possession of sealed sources and brachytherapy sources, is a new section that replaces the current § 35.59. Paragraph (a) continues to require that the licensee follow the radiation safety and handling instructions supplied by the manufacturer, but the requirement to maintain the instructions for the duration of source use has been deleted. Paragraph (b) requires that a source be tested for leakage before its first use, unless the licensee has a certificate from the supplier indicating that the source was tested within 6 months, and the source is tested for leakage at intervals not to exceed 6 months or at other intervals approved in the Sealed Source and Device Registry.⁴ The Sealed Source and Device Registry certificates, in most cases, will include a requirement for leak-testing. Approved intervals for testing are based on information regarding source design construction that is provided by the manufacturer.

Paragraph (c) retains the detection level for leakage at 185 Becquerels (Bq) (0.005 microcuries). The prescriptive requirements on how to satisfy the leak test requirements in the current § 35.59(c) were deleted to reflect the risk-informed, performance-based nature of this final rule. Paragraph (d) requires that leak test records be maintained in accordance with § 35.2067, Records of possession of sealed sources and brachytherapy sources. Paragraph (e) was revised to give the licensee two additional alternatives for action after a leaking source has been identified. The final rule gives the

⁴ A national registry that contains all the registration certificates generated by both NRC and the Agreement States. Registration certificates summarize the radiation safety information submitted by the applicant, and describe the licensing and use conditions approved for the product.

licensee the added flexibility of repairing or disposing of the source in accordance with Parts 20 and 30 if the leakage test reveals the presence of 185 Bq (0.005 microcuries) or more of removable contamination. The current rule only allows the licensee to withdraw the sealed source from use and store it in accordance with the requirements in Parts 20 and 30. The licensee is still required to report to the NRC if a leakage test reveals the presence of 185 Bq (0.005 microcuries) or more of removable contamination. Reporting requirements for this section are in § 35.3067, Report of a leaking source.

Paragraph (g) was revised to change the frequency for source inventories from quarterly to semi-annually to reduce the regulatory burden on licensees. It was also revised to exempt gamma stereotactic radiosurgery sources from the requirement for physical inventories. However, the final rule does not preclude the licensee from conducting an inventory on a more frequent basis. The recordkeeping requirements for this section were moved to § 35.2067, Records of possession of sealed sources and brachytherapy sources.

Paragraphs (h) and (i) in the current § 35.59 were deleted because radiation surveys are addressed under Part 20.

Section 35.69, Labeling of vials and syringes, is a new section that replaces the current §§ 35.60 and 35.61. It requires that syringes and vials containing radioactive drugs be labeled to identify the radioactive drug. It also requires that syringe shields and vial shields be labeled unless the label on the syringe or vial is visible when shielded. These requirements are needed because the Commission does not believe that the labeling requirements in Part 20 are sufficient to ensure that syringes, vials, syringe shields, or vial shields are properly labeled to identify the radioactive drug. In addition, the Commission believes that labeling helps to reduce administration errors.

Shielding of vials and syringes was not addressed in this section. Licensees are required to show compliance with the public and occupational dose limits specified in Part 20 of this chapter. We believe that the licensee should have flexibility in complying with these limits.

Section 35.70, Surveys of ambient radiation exposure rate, was revised and retitled. The term “contamination” was deleted from the title because this section no longer addresses contamination surveys. The final rule requires that licensees survey, at the end of each day of use, all areas where radiopharmaceuticals requiring a written directive were prepared for use or administered except in an area(s) where patients or human research subjects are confined when they cannot be released under § 35.75. Maintaining the requirement for surveys in areas where radioactive drugs requiring a written directive are used is consistent with the Commission's direction for a risk-informed rule.

Licensees are required to show compliance with the public and occupational dose limits specified in Part 20 of this chapter and specifically to develop, document, and implement a radiation protection program commensurate with the scope and extent of licensed activities (§ 20.1101). In situations where radioactive material is used at levels that would not require a survey pursuant to this section, the licensee should be aware that a survey may be required by § 20.1501. The Commission believes that licensees will continue to perform radiation surveys as dictated by “good health physics” practices.

The recordkeeping requirements for this section are in § 35.2070, Records of surveys for ambient radiation exposure rate. All other requirements in the current § 35.70 were deleted.

Section 35.75, Release of individuals containing unsealed byproduct material or implants containing byproduct material, was retitled and revised. The title of the section and paragraph (a) were revised to delete the term "permanent." This was done to clarify that this section applies to all individuals released from licensee control. Paragraph (b) was revised to specify that licensees may provide instructions to either the released individual or to the individual's parent or guardian and to replace the term "dose" with the term "total effective dose equivalent." The first change acknowledges that, in some cases, it is not appropriate to provide the individual being released with instructions (e.g., the individual is a minor or incapable of understanding the instructions). The later change was made to clarify what is meant by "dose" in this section.

Paragraph (b)(2) was modified to state "potential consequences, if any," of failure to follow the guidance. The Commission recognizes that, at low doses, there may be no consequences to continued breast-feeding. A patient may be unnecessarily alarmed if he/she is provided with information on consequences. Therefore, if consequences are not anticipated, the licensee would not be required to provide information to the individual.

The footnote was revised to reference NUREG 1556, Volume 9, "Consolidated Guidance About Materials Licenses, Program-Specific Guidance About Medical Licenses," that superseded Regulatory Guide 8.39.

Paragraphs (c) and (d) were revised to indicate that the recordkeeping requirements for this section are in § 35.2075, Records of the release of individuals containing radioactive drugs or implants containing byproduct material.

Section 35.80, Provision of mobile medical service, was retitled and revised. The title was changed to make it clear that the provisions in this part apply to all mobile medical services and not just to mobile nuclear medicine services. The current paragraphs (a), (b), and (c) were deleted because radiopharmaceutical usage is limited by the requirements in §§ 35.100 and 35.200, and control and security of material are addressed in Part 20. The remainder of the current requirements were incorporated into final paragraphs (a) or (c).

Paragraph (a) requires the mobile medical service provider to obtain a letter from its client that permits the use of byproduct material at the client's address. This letter should clearly delineate the authority and responsibility of the licensee and the client. This paragraph also requires that the mobile medical service provider check instruments used to measure the activity of unsealed byproduct materials for constancy before use at each address of use or on each day of use, whichever is more frequent. For example, if a mobile medical service licensee provides service to more than one client in a day, the instruments would need to be checked at each client's address. The Commission recognizes that the standard of practice is to check other types of equipment, such as gamma cameras, for proper operation at each place of use. Therefore, the Commission has not included any requirements to check this type of equipment in the final rule.

Paragraph (a) also requires that the licensee check survey instruments for proper operation with a dedicated check source, before use, at each client's address. We believe this is appropriate because extensive movement in a transport vehicle may cause the instruments to become damaged or uncalibrated. Finally, paragraph (a) requires the licensee to survey all areas of use to comply with the dose limits in Part 20 before leaving each client's address. This is necessary to assure that all radioactive material is removed from a client's facility.

Paragraph (b) addresses the delivery of byproduct material. It does not allow byproduct material to be delivered from the manufacturer or the distributor to the client's address, unless the client has a license allowing possession of the byproduct material. This requirement is similar to the requirement in the current § 35.29 (which was deleted by this rulemaking).

The recordkeeping requirements for this section are in § 35.2080, Records of administrative and technical requirements that apply to the provision of mobile medical services.

The current § 35.90, Storage of volatiles and gases, was deleted in its entirety. Licensees are required to comply with the public and occupational public dose limits in Part 20 and to maintain exposures ALARA. We believe that licensees should have flexibility in complying with Part 20, and, therefore, a prescriptive requirement in Part 35 is not needed.

Section 35.92, Decay-in-storage, was revised to allow decay-in-storage for byproduct material with a physical half-life of less than 120 days. Under the current rule, decay-in-storage was only authorized for material with a half-life of less than 65 days. Licensees that would like to decay material with a physical half life greater than 120 days would have to apply for and receive an amendment that would permit the decay-in-storage. This change provides licensees with greater flexibility in handling radioactive waste. This revision to § 35.92 codifies current licensing practice.

The requirement in the current paragraph (a)(1) to hold byproduct material for 10 half-lives was deleted. This requirement was not needed in light of the requirement in paragraph (a) of the final rule that precludes disposal of radioactive material until radiation levels adjacent to the material do not exceed background levels. Paragraph (a) still requires the licensee to remove or obliterate all radiation labels, except for material that would be handled as biomedical waste after it has been released. We recognize that material being handled as biomedical waste is typically being incinerated.

The requirement in the current paragraph (a)(4) to separate and monitor each generator column was deleted. This change recognized that the current level of prescriptiveness is not needed because of the requirements in paragraph (a)(1).

The recordkeeping requirements for this section are in § 35.2092, Records of decay-in-storage.

Subpart D was retitled Unsealed Byproduct Material - Written Directive Not Required. This subpart combines the requirements in the current subpart D, Uptake, dilution, and excretion and subpart E, Imaging and localization. This change was made to consolidate specific requirements for the use of unsealed byproduct material where a written directive is not required into one subpart. These changes are consistent with the Commission's intent to make Part 35 modality specific where appropriate. We believe that administrations of unsealed byproduct material not requiring a written directive are in a lower risk category than those administrations requiring a written directive. Therefore, we are using the requirement for a written directive as the threshold to distinguish between the two level of risks.

Section 35.100, Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required, was retitled and revised. The title and introductory paragraph were changed to clearly state that the provisions in this subpart do not apply to the medical use of byproduct material that would require a written directive.

Paragraph (a) was revised to change the format for citing Title 10 of the Code of Federal Regulations (CFR). The reference to Title 10 is now stated as “of this chapter” instead of using the format “10 CFR”.

Paragraph (b) was revised to reflect changes to the numbering in the final rule (i.e., requirements in §§ 35.25 and 35.920 were moved, with some modification, to §§ 35.27 and 35.290, respectively). Also, a reference to § 35.390 was added because physicians meeting these training and experience criteria can now elute generators and prepare radioactive drugs. This paragraph permits medical use licensees to prepare radioactive drugs from any unsealed byproduct materials (e.g., radiochemicals) provided the drug is prepared by an ANP or AU.

Paragraph (c) was added to allow specific licensees to obtain unsealed byproduct material prepared by other NRC or Agreement State licensees for use in research in accordance with a Radioactive Drug Research Committee (RDRC)-approved protocol or an Investigational New Drug Application (IND) protocol accepted by the Food and Drug Administration (FDA). This change was made because the current rule did not allow a licensee to use material from a supplier, who was not a § 32.72 licensee, unless the supplier had obtained a license exemption from the NRC. The final rule allows a

medical use licensee to receive radioactive drugs, for use in RDRC-approved or IND research protocols, prepared and distributed by NRC or Agreement State licensees who are not § 32.72 licensees.

Paragraph (d) was added to allow any individual to prepare a radioactive drug from any unsealed byproduct material (e.g., radiochemicals) for use in research in accordance with either an RDRC-approved protocol or an IND protocol. This change was made because an AU meeting the qualifications in § 35.910 of the current rule could not prepare radioactive drugs under an RDRC-approved protocol or an IND protocol. Therefore, if a licensee was only authorized to use byproduct material under § 35.100, it could not prepare byproduct material for use under an RDRC-approved protocol or an IND protocol unless the material had been prepared by an ANP or AU who was qualified to prepare radioactive drugs. The final rule resolves the issue by allowing any individual to prepare a radioactive drug in accordance with either an RDRC-approved protocol or an IND protocol.

Section 35.120, Possession of survey instruments, was deleted because these specific requirements are not needed in Part 35. Section 20.1501 of this chapter requires that the licensee make, or cause to be made, surveys to demonstrate compliance with Part 20, and requires the licensee to ensure that instruments and equipment used to show compliance with Part 20 are periodically calibrated. In addition, § 30.33(a)(2) of this chapter requires the licensee to have adequate instrumentation. Guidance on the types of instruments medical licensees could consider using is available in draft NUREG-1556, Vol. 9.

Section 35.190, Training for uptake, dilution, and excretion studies, is a new section. The training and experience requirements for an AU for unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required were moved, with some modifications, from the current § 35.910, Training for uptake, dilution, and excretion studies. Three changes made in the new section should be noted. First, the listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AUs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Second, the new requirements require a total of 60 hours of training and experience that must include classroom, laboratory, and supervised work experience. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Section 35.200, Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required, was retitled and revised. The title and introductory paragraph were changed to clearly state that the provisions in this part do not apply to the medical use of byproduct material that would require a written directive.

Paragraph (a) was revised to change the format for citing Title 10 of the Code of Federal Regulations (CFR). The reference to Title 10 is now stated as "of this chapter" instead of using the format "10 CFR".

Paragraph (b) was revised to reflect changes to the numbering in the final rule (i.e., requirements in §§ 35.25 and 35.920 were moved, with some modification, to §§ 35.27 and 35.290, respectively). Also, a reference to § 35.390 was added because physicians meeting these training and experience criteria can now elute generators and prepare radioactive drugs. This paragraph permits medical use licensees to prepare radioactive drugs from any unsealed byproduct materials (e.g., radiochemicals) provided the drug is prepared by an ANP or AU.

Paragraph (c) was added to allow specific licensees to obtain unsealed byproduct material prepared by other NRC or Agreement State licensees for use in research in accordance with an RDRC-approved protocol or an IND protocol accepted by the FDA. This change was made because the current rule did not allow a licensee to use material from a supplier, who was not a § 32.72 licensee, unless the supplier had obtained a license exemption from the NRC. The final rule allows a medical use licensee to receive radioactive drugs, for use in RDRC-approved or IND research protocols, prepared and distributed by NRC or Agreement State licensees who are not § 32.72 licensees.

Section 35.204, Permissible molybdenum-99 concentration, was revised. Paragraph (a) was revised to express the permissible concentration level as 0.15 kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (0.15 microcurie of molybdenum-99 per microcuries of technetium-99m). This level is identical to that used in the U.S. Pharmacopeia (USP) 23 U.S. Pharmacopial Convention, Inc., 1994, page 486-487. Paragraph (b) was revised to require that a licensee measure the molybdenum-99 concentration of the first eluate from a generator. We believe that the licensee should measure the molybdenum-99 concentration in the first elution of a generator after the generator is received at the licensee's facility. Although the frequency of molybdenum breakthrough is exceedingly rare, an initial check may detect generators that have been damaged in transport. The term

“extract” was deleted because the term is no longer needed. NRC is not aware of any licensees that prepare technetium-99m by the solvent extraction method.

The recordkeeping requirements for this section were moved to § 35.2204, Records of molybdenum-99 concentration.

Section 35.205, Control of aerosols and gases, was deleted in its entirety. Part 35 licensees must comply with the occupational and public dose limits of Part 20. Additional prescriptive requirements for limiting airborne concentrations of radioactive material are not needed in Part 35.

Section 35.220, Possession of survey instruments, was deleted in its entirety because specific requirements are not needed in Part 35. Section 20.1501 of this chapter requires that the licensee make, or cause to be made, surveys to demonstrate compliance with Part 20, and requires the licensee to ensure that instruments and equipment used to show compliance with Part 20 are periodically calibrated. In addition, § 30.33(a)(2) of this chapter requires licensees to have adequate instrumentation. Guidance on the types of instruments medical licensees could consider using is available in draft NUREG-1556, Vol. 9.

Section 35.290, Training for imaging and localization studies, is a new section. The training and experience requirements for an AU for unsealed byproduct material for imaging and localization studies for which a written directive is not required were moved, with some modifications, from the current § 35.920, Training for imaging and localization studies. Three changes made in the new section should be noted. First, the listing of specialty boards by

name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AUs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Second, the new requirements require a total of 700 hours of training and experience that must include classroom, laboratory, and supervised work experience. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Subpart E was retitled, Unsealed byproduct material - written directive required. The subpart contains the requirements for any medical use of unsealed byproduct material for which a written directive is required. This subpart would replace the requirements in the current subpart F, Radiopharmaceuticals for therapy.

Section 35.300, Use of unsealed byproduct material for which a written directive is required, was retitled and revised. The title and introductory paragraph were changed to clearly state that the provisions in this subpart apply to the medical use of unsealed byproduct material that would require a written directive. The first paragraph in this section was revised to clearly state that medical uses under this section require a written direction. Also, the phrase "therapeutic administration", used in the current rule, was deleted because some medical uses in this modality will require a written directive but are not "therapeutic administrations" (e.g., diagnostic whole body imaging with I-131 sodium iodide).

Paragraph (a) was revised to change the format for citing Title 10 of the Code of Federal Regulations (CFR). The reference to Title 10 is now stated as “of this chapter” instead of using the format “10 CFR”.

Paragraph (b) was revised to reflect changes to the numbering in the final rule (i.e., requirements in §§ 35.25 and 35.920 were moved, with some modification, to §§ 35.27 and 35.290, respectively). Also, a reference to § 35.390 was added because physicians meeting these training and experience criteria can now elute generators and prepare radioactive drugs. This paragraph permits medical use licensees to prepare radioactive drugs from any unsealed byproduct materials (e.g., radiochemicals) provided the drug is prepared by an ANP or AU.

Paragraph (c) was added to allow specific licensees to obtain unsealed byproduct material prepared by other NRC or Agreement State licensees for use in medical research in accordance with an IND protocol accepted by the FDA. This change was made because the current rule did not allow a licensee to use material from a supplier, who was not a § 32.72 licensee, unless the supplier had obtained a license exemption from the NRC. The final rule allows a medical use licensee to receive radioactive drugs, for use in IND research protocols, prepared and distributed by NRC or Agreement State licensees who are not § 32.72 licensees. This paragraph is similar to the regulatory text added to §§ 35.100 and 35.200. However, we have not included a reference to RDRC-approved protocols because RDRCs are authorized to approve radioactive drugs for certain types of research uses intended to obtain basic information regarding the metabolism of a radioactive drug, or regarding human physiology, pathophysiology, or biochemistry, but not intended for immediate diagnostic, therapeutic, or similar purposes. Additionally, the maximum radiation dose from a single administration of a radioactive drug in an RDRC-approved protocol must be less than 3 rem to the whole body, active blood forming organs, lens of the eye, and gonads, and less than 5

rem to other organs. It is expected that doses from materials requiring a written directive would exceed these limits, and thus research with such materials could not be conducted under the aegis of RDRC approval.

Paragraph (d) was added to allow any individual to prepare a radioactive drug from any unsealed byproduct material (e.g., radiochemicals) in accordance with an IND protocol. This change was made because an AU meeting the qualifications in §§ 35.930, 35.932, or 35.934 of the current rule could not prepare radioactive drugs under an IND protocol. Therefore, if a licensee was only authorized to use byproduct material under § 35.300, it could not prepare byproduct material for use under an IND protocol unless the material had been prepared by an ANP or AU who was qualified to prepare radioactive drugs. The final rule resolves the issue by allowing any individual to prepare a radioactive drug in accordance with an IND protocol.

Section 35.310, Safety instruction, was revised to explicitly state that the instruction requirements of this section are in addition to, and not in lieu of, the training requirements in § 19.12. We believe it is important that personnel caring for patients or human research subjects that have received a dosage requiring a written directive, and cannot be released in accordance with § 35.75, receive instruction in limiting radiation exposure to the public or occupational workers and the actions to be taken in the case of a death or medical emergency.

Paragraph (a) in the final rule requires that safety instruction be provided initially and at least annually. The current rule does not specify when instructions must be given. Typically the frequency of training has been handled during the licensing process. It is not expected that the same level of training be provided to all individuals caring for the patient. The level of training should be commensurate with the potential radiation exposure the

caregiver may receive based on the level of contact the individual is expected to have with the patient or human research subject. For example, the instruction provided to the registered nurse will not necessarily be the same as the instruction provided to a nursing assistant. We have deleted the reference to “licensee’s procedures” in paragraph (a) because we have chosen to focus this section on instruction rather than on procedures. We believe the licensee should have flexibility in program management and recognize that licensees may develop alternative ways of addressing the issues in paragraphs (a)(1) through (a)(5). Paragraph (a)(2) was also revised to require that instruction on visitor control include instruction on routine visitation authorized under the provisions in § 20.1301(a)(1), as well as visitation that is authorized under the final provisions of § 20.1301(c). Paragraph (a)(5) was revised to state that personnel should notify the RSO, or his or her designee, and the AU if the patient or the human research subject dies or has a medical emergency. This change was made to allow the RSO to designate an individual to act in his or her behalf, in such cases, to address radiation protection issues and to ensure that the AU is notified. The recordkeeping requirements for this section are in § 35.2310, Records of instruction and training.

Section 35.315, Safety precautions, was revised. Paragraph (a) was revised to clarify that the requirements in this section only apply if a patient or research subject cannot be released in accordance with § 35.75. Paragraph (a)(1) was revised to give the licensee flexibility in quartering patients. Option 1 is identical to the current rule, i.e. it allows the licensee to quarter the patient or human research subject in a private room with a private sanitary facility. Option 2 allows the licensee to quarter the individual in a room, with a private sanitary facility, with another individual who also has received therapy with a radioactive drug containing byproduct material and who also cannot be released pursuant to § 35.75. Option 2 was included in the final

rule because we believe that the dose patients receive from each other would be inconsequential in light of the dose that they receive from the medical treatment that they have undergone.

Paragraph (a)(2) was revised to require that the patient's room, rather than the door, be visibly posted to give the licensee some flexibility in determining where to place the posting. These requirements are in addition to the posting requirements in Part 20. We believe that posting requirements in Part 20 are not adequate to ensure that individuals entering the room would be aware of the presence of radioactive materials in the room. The current requirements in paragraphs (a)(3), (4), (6), (7), and (8) were deleted because they are radiation protection requirements that are covered under Part 20. Paragraph (b) was revised to state that the licensee shall notify the RSO, or his/her designee, and the AU, as soon as possible, if the patient or human research subject has a medical emergency or if the patient dies. This change was made to allow the RSO to designate an individual to act in his or her behalf, in such cases, to address radiation protection issues and to ensure that the AU is notified.

Section 35.320, Possession of survey instruments, was deleted in its entirety because these specific requirements are not needed in Part 35. Section 20.1501 of this chapter requires that the licensee make or cause to be made surveys to demonstrate compliance with Part 20 and requires the licensee to ensure that instruments and equipment used to show compliance with Part 20 are periodically calibrated. In addition, § 30.33(a)(2) requires a licensee to have adequate equipment. Guidance on the types of instruments medical licensees could consider using is available in draft NUREG-1556, Vol. 9.

Section 35.390, Training for use of unsealed byproduct material for which a written directive is required, is a new section. The training and experience requirements for an AU for unsealed byproduct material for which a written directive is required were moved, with some modifications, from the current § 35.930, Training for therapeutic use of unsealed byproduct material. Three changes made in the new section should be noted. First, the listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AUs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Second, the new requirements require a total of 700 hours of training and experience that must include classroom, laboratory, and supervised work experience. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Section 35.392, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (GBq) (33 millicuries), is a new section. The training and experience requirements for an AU for iodine-131 for treatment of hyperthyroidism were moved, with some modifications, from the current 35.932, Training for treatment of hyperthyroidism. Three changes made in the new section should be noted. First, the section is no longer limited to use of iodine-131 for treatment of hyperthyroidism. Second, the final rule provides for NRC recognition of the boards. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirement in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the

Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Section 35.394, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 GBq (33 millicuries), is a new section. The training and experience requirements for an AU for iodine-131 for treatment of thyroid carcinoma were moved, with some modifications, from the current 35.934, Training for treatment of thyroid carcinoma. Three changes made in the new section should be noted. First, the section is no longer limited to use of iodine-131 for treatment of thyroid carcinoma. Second, the rule provides for NRC recognition of the boards. Third, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirement in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Subpart F was retitled Manual Brachytherapy. This subpart contains the requirements for medical use of sealed sources for manual brachytherapy and replaces the requirements in the current Subpart G, Sources for Brachytherapy.

Section 35.400, Use of sealed sources for manual brachytherapy, was retitled and revised to delete the specific sources and uses listed in the current paragraphs (a) through (g). This conforms with the risk-informed, performance-based nature of this final rule. The licensee has the flexibility to use sealed sources for therapeutic medical uses as approved in the Sealed Source and Device Registry. In addition, a new paragraph (b) was added to

allow the use of therapy sealed sources in medical research as long as the research is conducted in accordance with an effective Investigational Device Exemption (IDE) application accepted by the FDA. With this revision, we allow previously registered sources to be used for uses other than those described in the original sealed source registration process if the research is conducted under an effective IDE application accepted by the FDA.

Section 35.404, Surveys after source implant and removal, was retitled and revised. The current paragraph (a) was redesignated paragraph (b) and was revised to delete the requirement that a licensee may not release a patient or a human research subject treated by temporary implant until all sources have been removed. The release of patients or human research subjects is addressed in § 35.75. The reference to radiation when referring to the survey was also removed because this was repetitive of the requirement to perform the survey with a radiation detection survey instrument. The new paragraph (a) contains the requirements, with minor modifications, that were previously required by § 35.406(c). The survey required by paragraph (a) is performed to locate and account for all sources that have not been implanted, but it does not necessarily have to be a radiation survey. Depending on the area being surveyed and the ability to distinguish from the radiation background around the patient implanted with brachytherapy sources, the survey may be a visual or a radiation survey. Therefore, this section includes all of the survey requirements for this subpart. The recordkeeping requirements for this section are in § 35.2404, Records of surveys after source implant and removal.

Section 35.406, Brachytherapy sources accountability, was retitled and revised. Paragraph (a) requires that the licensee maintain accountability for all brachytherapy sources in storage or use. The majority of the prescriptive requirements and associated recordkeeping requirements in the current section were deleted in this final section to give the licensee flexibility in program management. The requirements in the current paragraph (c) were

moved to § 35.404. We believe that the requirements that were retained in this section are essential to the radiation safety program. The recordkeeping requirements for this section are in § 35.2406, Records of brachytherapy source accountability.

Section 35.410, Safety instruction, was revised to explicitly state that the instruction requirements in this section are in addition to, and not in lieu of, the training requirements of § 19.12. We believe that it is important that personnel caring for patients or human research subjects that have received implant therapy (and cannot be released in accordance with § 35.75), receive instruction in limiting radiation exposure to the public and workers and the actions to be taken in the case of a death or medical emergency.

Paragraph (a) in the final rule requires that safety instruction be provided initially and at least annually. The current rule does not specify when instructions must be given. The frequency of training is typically handled during the licensing process. It is not expected that the same level of training be provided to all individuals caring for the patient. The level of training should be commensurate with the type of care that the personnel may render to the patient or human research subject. We have deleted the reference to “procedures” in paragraph (a) because we have chosen to focus this section on instruction rather than on procedures. We believe the licensee should have flexibility in program management and recognize that licensees may develop alternative ways of addressing the issues in paragraphs (a)(1) through (a)(5). Paragraph (a) was also revised to require that instruction on visitor control include instruction on routine visitation authorized under the provisions in § 20.1301(a)(1), as well as visitation that is authorized under the final provisions of § 20.1301(c). Paragraph (a)(5) was revised to state that personnel should notify the RSO, or his or her designee, and an AU, if the patient or human research subject dies or has a medical emergency. This change was made to provide the RSO flexibility in designating who should be notified to address

radiation protection issues and to ensure that an AU is notified. The recordkeeping requirements for this section are in § 35.2310, Records of safety instruction.

Section 35.415, Safety precautions, was revised. Paragraph (a) was revised to clarify that the requirements in this section only apply if a patient or human research subject is receiving brachytherapy and cannot be released in accordance with § 35.75. Paragraph (a)(1) was revised to clarify that a patient or human research subject who is receiving brachytherapy can only share a room with another brachytherapy patient. Paragraph (a)(2) was revised to allow posting of the patient's room, rather than specifying the door. This provides the licensee flexibility in determining where to place the posting so it is visible. These posting requirements are in addition to the posting requirements in Part 20. We believe that the posting requirements in Part 20 are not adequate to ensure that individuals entering the room would be aware of the presence of radioactive materials in the room. The requirement to put a note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or human research subject's room was moved from the current paragraph (a)(2) to the new paragraph (a)(3). The current requirements in paragraphs (a)(3) and (4) were deleted because they are radiation protection requirements that are covered under Part 20. A new requirement (paragraph b) was added that requires the licensee to have emergency response equipment available near each treatment room. This addition codifies requirements that are currently imposed on licensees by license conditions. The current paragraph (b) was redesignated as paragraph (c) and was revised to state that the licensee shall notify the RSO, or his/her designee, and an AU as soon as possible, if the patient or human research subject has a medical emergency or dies. This change was made: (1) to recognize that in a medical emergency, the licensee's primary responsibility is the care of the patient; (2) to provide the RSO flexibility in whom should be notified to address radiation protection issues; and(3) to ensure that the AU is notified.

Section 35.420, Possession of survey instruments, was deleted in its entirety because these specific requirements are not needed in Part 35. Section 20.1501 of this chapter requires that the licensee make, or cause to be made, surveys to demonstrate compliance with Part 20 and requires the licensee to ensure that instruments and equipment used to show compliance with Part 20 are periodically calibrated. In addition, § 30.33(a)(2) of this chapter requires licensees to have adequate equipment. Guidance on the types of instruments medical licensees could consider using is available in draft NUREG-1556, Vol. 9.

Section 35.432, Calibration measurements of brachytherapy sealed sources, is a new section that requires a licensee authorized to use brachytherapy sources for medical use to perform calibration measurements on brachytherapy sources before the first medical use of the source(s) after the effective date of this rule. The requirements in this section are based on recommendations found in American Association of Physicists in Medicine (AAPM) Task Group 40 - Comprehensive QA for Radiation Oncology (1994) and 56 - Code of Practice for Brachytherapy Physics (1997) and are consistent with the calibration requirements for sealed sources and devices for therapy. The final rule allows the licensee to rely on the output measurement provided by the source manufacturer or by a calibration laboratory accredited by the American Association of Physicists in Medicine, as long as the calibration was conducted in accordance with a published protocol accepted by a nationally recognized body and appropriately calibrated equipment was used. As discussed in the Regulatory Impact Statement, we recognize that licensees may need to procure additional equipment to meet this requirement. We believe that the additional expenditure is warranted in order for the licensee administering brachytherapy doses to ensure that the correct dose is delivered to patients. The recordkeeping requirements for this section are in § 35.2432, Records of calibration measurements of brachytherapy sealed sources.

Section 35.433, Decay of strontium-90 sources for ophthalmic treatment, is a new section. This section requires that only an AMP may calculate the activity of a strontium-90 source that is used to determine the treatment times for ophthalmic treatments. It also requires that the decay must be based on the activity determined under § 35.432. This section was added because we are aware of numerous misadministrations involving strontium-90 for ophthalmic use that were caused by individuals improperly decaying the sources. Given the risks associated with the use of strontium-90 and the numerous misadministrations in this area, more prescriptive requirements are warranted to ensure that the activity of strontium-90 sources are correctly determined. The recordkeeping requirements for this section are in § 35.2433, Records of decay of strontium-90 sources for ophthalmic treatments.

Section 35.457, Therapy-related computer systems, is a new section that requires acceptance testing on the licensee's treatment planning system in accordance with published protocols accepted by nationally recognized bodies. The requirements in this section are based on recommendations found in AAPM Task Group 56 - Code of Practice for Brachytherapy Physics (1997). The components of the testing are provided in this section. However, the licensee retains the flexibility in developing the acceptance testing program. We believe that these new requirements are warranted in order for the licensee administering brachytherapy doses to ensure that the correct dose is delivered to patients.

Section 35.490, Training for use of manual brachytherapy sources, is a new section. The training and experience requirements for an AU of manual brachytherapy sources were moved, with some modifications, from the current § 35.940, Training for use of brachytherapy sources. Two changes made in the new section should be noted. First, the listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer

incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AUs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Second, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Section 35.491, Training for ophthalmic use of strontium-90, is a new section. The training and experience requirements for an AU of strontium-90 sources for ophthalmic treatment were moved, with some modifications, from the current § 35.941, Training for ophthalmic use of strontium-90. One change made in the new section should be noted. An individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Subpart G was retitled Sealed Sources for Diagnosis. This subpart contains the requirements for diagnostic medical use of sealed sources and replaces the requirements in the current Subpart H, Sealed Sources for Diagnosis.

Section 35.500, Use of sealed sources for diagnosis, was revised to delete the specific sources and uses listed in paragraphs (a) and (b). This conforms with the risk-informed, performance-based nature of this final rule. The licensee has the flexibility to use sealed sources for diagnostic medical uses as approved in the Sealed Source and Device Registry.

Section 35.520, Availability of survey instrument, was deleted in its entirety because these specific requirements are not needed in Part 35. Section 20.1501 of this chapter requires that the licensee make or cause to be made surveys to demonstrate compliance with 10 CFR Part 20 and requires the licensee to ensure that instruments and equipment used to show compliance with 10 CFR Part 20 are periodically calibrated. In addition, § 30.33(a)(2) of this chapter requires the licensee to have adequate instrumentation. Guidance on the types of instruments medical licensees could consider using is available in draft NUREG-1556, Vol. 9.

Section 35.590, Training for use of sealed sources for diagnosis, is a new section. The training and experience requirements for an AU of a diagnostic sealed source in a device were moved, with some modifications, from the current § 35.950, Training for use of sealed sources for diagnosis. One change made in the new section should be noted. The listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AUs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Subpart H, Photon Emitting Remote Afterloader Units, Teletherapy Units, and Gamma Stereotactic Radiosurgery Units, was retitled and revised to address all medical uses of photon emitting sealed sources in devices for therapy. Devices such as teletherapy, remote afterloaders, and gamma stereotactic radiosurgery units are addressed in this subpart. This section does not contain requirements for manual brachytherapy, which are in subpart F, nor does it include requirements for beta emitting devices, such as beta emitting intravascular brachytherapy devices. This subpart replaces the requirements in the current subpart I, Teletherapy.

Section 35.600, Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit, was retitled and revised to delete any references to specific radionuclides and devices. The licensee has the flexibility to use sealed sources in photon emitting devices for therapeutic medical uses as approved in the Sealed Source and Device Registry. In addition, paragraph (b) was added to allow the use of therapy sealed sources in medical research as long as the research is conducted in accordance with an effective IDE application accepted by the FDA. With this revision, we allow previously registered sources to be used for uses other than those described in the original sealed source registration process, if the research is conducted under an effective IDE application accepted by the FDA.

Section 35.604, Surveys of patients and human research subjects treated with a remote afterloader unit, is a new section. This section requires that a licensee make a radiation survey of a patient or human research subject to confirm that the sources have been removed from the individual and returned to a shielded position before releasing the individual from licensee control. For fractionated low dose-rate or pulsed dose-rate treatments where the patient is not releasable under § 35.75, surveys need only be performed after the last time the source is returned to the shielded position. For

example, a survey of the patient is not required every time that the source is retracted into the shielded safe when nursing personnel enter the patient treatment room to provide care to patients undergoing fractionated treatments using a low or pulsed dose-rate remote afterloader unit. This new requirement was previously imposed on remote afterloader licensees by license condition. The recordkeeping requirements for this section are in § 35.2404, Records of radiation surveys of patients and human research subjects.

Section 35.605, Installation, maintenance, adjustment, and repair, was retitled and revised to clarify that only a person specifically licensed by the Commission or an Agreement State can install, maintain, adjust, or repair a unit that involves work on the source shielding, source driving unit, or other electronic or mechanical mechanism that could expose the source, reduce the shielding around the source, or compromise the radiation safety of the unit or the sources. The types of units referred to in this section were revised to include remote afterloader units and gamma stereotactic radiosurgery units, rather than just teletherapy units.

Paragraph (b) also specifies that, except for low dose-rate remote afterloader units, only a person specifically licensed by the Commission or an Agreement State shall install, replace, relocate, or remove a sealed source or source contained in a device. For low dose-rate remote afterloader units, installation, replacement, relocation, or removal of a sealed source must be done by a person specifically licensed by the Commission or an Agreement State or by an AMP. The exception to allow an AMP to perform these activities for low dose-rate remote afterloader units was included in the final rule because we believe that the radiation hazards associated with installation, replacement, relocation, or removal of a sealed source in these devices are

similar to that of manipulation of manual brachytherapy sources. The recordkeeping requirements for this section are in § 35.2605, Records of installation, maintenance, adjustment, and repair.

Section 35.606, License amendments, was deleted in its entirety. The requirements in the current paragraphs (a), (b), and (d) are addressed in the final § 35.13(e). Paragraph (c) was deleted because the licensees must comply with the dose limit requirements in Part 20, and no further limitations are warranted. Paragraph (e) was deleted because the requirement to file an amendment before allowing an individual to perform the duties of the AMP is addressed in the final § 35.13(b). Paragraph (e) was deleted because the requirements in subpart H require that the AMP perform specific duties. Any deviations from these requirements would necessitate an exemption from Part 35.

Section 35.610, Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units, was retitled and revised to include remote afterloader units and gamma stereotactic radiosurgery units.

Paragraph (a) requires that a licensee secure the unit, console, console keys, and treatment room when not in use or unattended; permit only approved individuals into the treatment room during treatment; prevent dual operation of radiation producing devices; and develop, implement, and maintain written emergency response procedures.

Paragraphs (a) (1) and (a)(3) codify requirements that are currently imposed on licensees by license conditions related to use of remote afterloaders. Because of the applicability of the requirements to all therapy units, they were added to the rule with the intent of having the requirements apply to all such units. Paragraph (a)(2) was expanded to recognize that there are certain design conditions that will necessitate an individual, other than the patient, being in the treatment room during the treatment. An example of this condition is use of a low energy gamma source in a therapeutic medical device where the AU may need to be in the room with the patient. This exception does not relieve the licensees from complying with the dose limits for occupationally-exposed individuals or the general public in Part 20. Paragraph (a)(4) was expanded to codify requirements that are currently imposed on licensees by license conditions related to emergency procedures.

Paragraph (b) was revised to require that a copy of the licensee's procedures be physically located at the unit console. Paragraph (c) was revised to require that the location of the procedures and emergency response telephone numbers be posted. Previously, all of these procedures were required to be posted. This was impractical with the addition of remote afterloaders because error conditions and responses are often several pages in length.

Paragraphs (d) and (e), previously paragraph (b), were revised to require that the licensee provide initial and annual instruction in specifically identified procedures to all individuals who operate the device, and initial and annual practice drills in emergency procedures to unit operators, AMPs, and AUs. The level of instruction should be commensurate with the individual's assigned duties. For example, an individual need not be instructed in equipment inspection, unless it is expected that during the normal course of the day, the individual will be required to inspect the unit. We believe that due

to the complexity of therapeutic treatment units, refresher training and practice drills on emergency response are warranted. The recordkeeping requirements for this section are in § 35.2310, Records of instruction and training.

Section 35.615, Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units, was retitled and revised to include remote afterloader units and gamma stereotactic radiosurgery units. The current requirements in paragraphs (a) and (b) remain essentially the same, with minor changes to the language to support requirements for remote afterloader units and gamma stereotactic radiosurgery units. Many of the prescriptive requirements [e.g., beam condition indicator light [current paragraph (c)] and radiation monitor [current paragraph (d)]] were deleted from this section because they are addressed in Part 20.

The new requirement in paragraph (d) for intercom systems, and the requirements in paragraphs (e), (f), and (g) were added to codify requirements that are currently imposed on licensees by license conditions. Current license conditions were modified when they were incorporated into the final rule. For example, the presence of an AU and an AMP during patient treatments was clarified for each type of unit. As used in this provision, physically present means to be within hearing distance of normal voice. Immediately available means that the individual is available on an on-call basis to respond to an emergency. At a minimum, this person must be available by telephone.

We believe that the inherent risk of these procedures justifies the prescriptiveness of this regulation and that it is important for a properly trained physician to be available at all times to respond to an emergency requiring source removal.

Section 35.620, Possession of survey instruments, was deleted in its entirety because these specific requirements are not needed in Part 35. Section 20.1501 of this chapter requires that the licensee make, or cause to be made, surveys to demonstrate compliance with Part 20, and that the licensee ensure that instruments and equipment used to show compliance with Part 20 are periodically calibrated. In addition, § 30.33(a)(2) of this chapter requires licensees to have adequate equipment. Guidance on the types of instruments medical licensees could consider using is available in draft NUREG-1556, Vol. 9.

Section 35.630, Dosimetry equipment, was revised to provide calibration requirements for instruments used in this subpart and subpart F. Paragraph (a)(1) requires that dosimetry systems be calibrated using a source or system traceable to National Institute of Science and Technology (NIST) and in accordance with published protocols accepted by a nationally recognized body; or by a calibration laboratory accredited by AAPM. This change gives licensees two alternatives for direct traceability of dosimetry equipment calibration, i.e., either a source or the measurement instrument (e.g., well chamber) can be calibrated against a national standard. We acknowledge that the industry standards for instrument calibration provide adequate assurance that equipment is properly calibrated. Paragraph (a)(2) was revised to delete the reference to intercomparison meetings sanctioned by a calibration laboratory or radiologic physics centers accredited by the AAPM. This provision is no longer necessary because the AAPM does not sanction intercomparison meetings. References to cobalt-60 and cesium-137 contained within teletherapy units were deleted from the rule text to make the section applicable to dosimetry equipment for all radionuclides and therapy units. In addition, licensees using low dose-rate remote afterloader units only are not required to possess dosimetry equipment if they rely on the source output or activity determined by the manufacturer, as long as the manufacturer uses appropriately calibrated equipment and performs the calibration in accordance with published protocols accepted by a nationally recognized body. This

allowance was made to be consistent with the requirements for manual brachytherapy sources. The recordkeeping requirements for this section are in § 35.2630, Records of dosimetry equipment.

Section 35.632, Full calibration measurements on teletherapy units, was revised and retitled to clarify that the requirements in this section apply to teletherapy units. Paragraph (d) was revised to delete the reference to the AAPM Task Group Reports and replace it with a requirement that full calibration measurements be done in accordance with published protocols accepted by nationally recognized bodies. This allows the licensee more flexibility in choosing appropriate protocols. We acknowledge that the industry standards for teletherapy unit calibration provide adequate assurance that equipment is properly calibrated. Paragraph (e) was revised to include mathematical correction of output for sources other than cobalt-60 and cesium-137. Paragraph (f) was revised to replace the term "teletherapy physicist" with the term "authorized medical physicist." The recordkeeping requirements for this section are in § 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.

Section 35.633, Full calibration measurements on remote afterloader units, is a new section that contains the requirements for the calibration of remote afterloader units. This section is similar in content to § 35.632. Requirements in this section were based on recommendations found in AAPM Task Group Report No. 56 - Code of Practice for Brachytherapy Physics (1997) and AAPM Task Group Report No. 59. The recordkeeping requirements for this section are in § 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.

Section 35.634, Periodic spot-checks, was deleted in its entirety, and the requirements of this section, with minor modifications, were moved to § 35.642.

Section 35.635, Full calibration measurements on gamma stereotactic radiosurgery units, is a new section that contains the requirements for the calibration of gamma stereotactic radiosurgery units. This section is similar in content to § 35.632. Requirements in this section are based on recommendations found in AAPM Report No. 54 - Stereotactic Radiosurgery (Task Group 42, 1995). The recordkeeping requirements for this section are in § 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.

The current § 35.636, Safety checks for teletherapy facilities, was deleted in its entirety, and the requirements in this section were extended to all therapy units and incorporated into the final §§ 35.642, 35.643, 35.645, and 35.647.

The current § 35.641, Radiation surveys for teletherapy facilities, was deleted in its entirety. Radiation surveys at the surface of the main source safe of therapy units were addressed in the final § 35.652. The remaining requirements in the current § 35.641 were deleted to allow the licensee more flexibility in managing its radiation protection program.

Section 35.642, Periodic spot-checks for teletherapy units, is a new section that contains the requirements that were previously found in § 35.634, Periodic spot-checks. The phrase "teletherapy physicist" was replaced with the term "authorized medical physicist" throughout the section. The

requirement in paragraph (c) to maintain a copy of the physicist's notification of the results of spot-checks to the licensee was deleted to reduce the recordkeeping requirements for licensees. Paragraph (d) was modified to require that the safety spot-checks be performed monthly and after each source installation. This revision replaces the safety check requirements after each source replacement in the current § 35.636, which is deleted in the final rule. Paragraph (d)(3) was modified to replace the term "beam condition indicator" with "source exposure indicator" to clarify that indicators were needed to note whether the source was exposed and note to what degree the source was exposed. Paragraph (d)(4) was revised to include a requirement for an intercom system that was previously imposed on licensees by license condition. An intercom is needed to assure that the licensee's staff and the patients have the ability to communicate verbally in addition to the ability to communicate visually. Paragraph (e) was revised to require that the licensee lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system, if a malfunction is identified during a safety spot-check. This revision is intended to make § 35.642 consistent with the requirement in the current § 35.636 regarding immediate actions to be taken when a malfunctioning system is identified. The recordkeeping requirements for this section are in § 35.2642, Records of periodic spot-checks for teletherapy units.

Section 35.643, Periodic spot-checks for remote afterloader units, is a new section that replaces the current § 35.643, Modification of teletherapy unit or room before beginning a treatment program. The requirements in the current § 35.643 were deleted because they were considered overly prescriptive. This allows the licensee more flexibility in designing a radiation protection program that is specific to its facility and which assures that the dose limits in Part 20 are not exceeded.

The new section contains the requirements for periodic spot-checks of remote afterloader units, and is similar in content to § 35.642. Requirements in this section are based on recommendations in AAPM Task Group Report Nos. 40 - Comprehensive QA for Radiation Oncology (1994) and 56 - Code of Practice for Brachytherapy Physics (1997). The recordkeeping requirements for this section are in § 35.2643, Records of periodic spot-checks for remote afterloader units.

Section 35.645, Periodic spot-checks for gamma stereotactic radiosurgery units, is a new section that replaces the current § 35.645, Reports of teletherapy surveys, checks, tests, and measurements. The requirements in the current § 35.645 were deleted to reduce the reporting burden on medical use licensees. We believe that there is no need to submit survey results to the appropriate Regional Office because the survey results are maintained by a licensee to show compliance with Part 20 and, therefore, are available for review.

The new section contains requirements for periodic spot-checks of gamma stereotactic radiosurgery units, and is similar in content to § 35.642. Requirements in this section are based on recommendations found in AAPM Report No. 54, Stereotactic Radiosurgery (1995). Additional requirements were added to make the safety checks, and associated corrective actions, consistent with the requirements in § 35.642. The recordkeeping requirements for this section are in § 35.2645, Records of periodic spot-checks for gamma stereotactic radiosurgery units.

Section 35.647, Additional technical requirements for mobile remote afterloader units, replaces the current § 35.647, Five-year inspection. Requirements in the current § 35.647 were moved to § 35.655. This section contains the requirements for mobile remote afterloader units which were

previously listed in an internal NRC document entitled, "Supplement 1 to Policy and Guidance Directive FC 86-4; Revision 1, Mobile Remote Afterloading Brachytherapy Licensing Module." The recordkeeping requirements for this section are in § 35.2647, Records of additional technical requirements for mobile remote afterloader units.

Section 35.652, Radiation surveys, is a new section. This section replaces the current requirements in § 35.641. This section requires that, in addition to the surveys required by § 20.1501, the licensee make surveys to assure that the maximum radiation levels and average radiation levels from the surface of the main source safe do not exceed the levels stated in the Sealed Source and Device Registry. These surveys provide added assurance that a device has been manufactured and that source(s) have been installed properly. The recordkeeping requirements for this section are in § 35.2652, Records of surveys of therapeutic treatment units.

Section 35.655, Five-year inspection for teletherapy and gamma stereotactic radiosurgery units, is a new section and contains the requirements for inspections that were in the current § 35.647. Section 35.655 requires that teletherapy units and gamma stereotactic radiosurgery units be inspected and serviced during source replacement, or at intervals not to exceed 5 years, to assure proper functioning of the source exposure mechanism. Most gamma stereotactic radiosurgery licensees are required, by license condition, to inspect the units every 7 years; however, professionals in the medical community have indicated that the units are inspected on a more frequent basis. We believe that the risk associated with using gamma stereotactic radiosurgery units justifies a change in the inspection frequency to a frequency consistent with teletherapy units, i.e., 5 years. The recordkeeping requirements for this section are in § 35.2655, Records of 5-year inspection for teletherapy and gamma stereotactic surgery units.

Section 35.657, Therapy-related computer systems, is a new section that requires licensees to perform acceptance testing on the treatment planning systems in accordance with published protocols accepted by nationally recognized bodies. These changes are consistent with recommendations found in AAPM Task Group Report No. 56 - Code of Practice for Brachytherapy Physics (1997). The components of the testing are provided in this section. However, the licensee retains flexibility in developing the acceptance testing program. We believe that these new requirements are warranted for the licensee administering therapy doses to ensure that the correct dose is delivered to patients.

Section 35.690, Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units, is a new section. This section contains the training and experience requirements for an AU of teletherapy, remote afterloader, and gamma stereotactic radiosurgery units. The current section, § 35.960, Training for teletherapy, was expanded to include the training for AUs of remote afterloaders and gamma stereotactic radiosurgery units because requirements for gamma stereotactic radiosurgery units and remote afterloader units have been codified in the revised Part 35. Two changes made in the new section should be noted. First, the listing of specialty boards by name was deleted because the regulatory text in Part 35 will no longer incorporate a listing of specialty boards whose diplomats automatically fulfill the training and experience requirements for AUs. In place of listing the boards, the final rule provides for NRC recognition of the boards. Second, an individual must obtain written certification from a preceptor indicating that the individual has satisfactorily completed the requirements in the section and has achieved a level of competency sufficient to function independently as an AU. Section III of the Supplementary Information section of this document contains a detailed discussion of the Commission's changes to the training and experience requirements in Part 35.

Subpart J, Training and Experience Requirements, was deleted in its entirety. The revised training and experience requirements are in Subparts B and D through H. A detailed discussion of the changes to the training and experience requirements is in Section III of the Supplementary Information section of this document.

Section 35.900, Radiation Safety Officer, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.50, Training for Radiation Safety Officer.

Section 35.901, Training for experienced Radiation Safety Officer, was deleted in its entirety. The requirements, with some modifications, were moved to the new § 35.57, Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

Section 35.910, Training for uptake, dilution, and excretion studies, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.190, Training for uptake, dilution, and excretion studies.

Section 35.920, Training for imaging and localization studies, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.290, Training for imaging and localization studies.

Section 35.930, Training for therapeutic use of unsealed byproduct material, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.390, Training for use of unsealed byproduct material for which a written directive is required.

Section 35.932, Training for treatment of hyperthyroidism, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.392, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (GBq) (33 millicuries)

Section 35.934, Training for treatment of thyroid carcinoma, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.394, Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (GBq) (33 millicuries)

Section 35.940, Training for use of brachytherapy sources, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.490, Training for use of manual brachytherapy sources.

Section 35.941, Training for ophthalmic use of strontium-90, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.491, Training for ophthalmic use of strontium-90.

Section 35.950, Training for use of sealed sources for diagnosis, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.590, Training for use of sealed sources for diagnosis.

Section 35.960, Training for teletherapy, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.690, Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Section 35.961, Training for teletherapy physicist, was deleted in its entirety. The requirements in this section, with some modifications, were moved to the new § 35.51, Training for an authorized medical physicist.

Section 35.970, Training for experienced authorized users, was deleted in its entirety. The requirements of this section, with minor modifications, were moved to the new § 35.57, Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

Section 35.971, Physicians training in a three-month program, was deleted in its entirety because three-month nuclear medicine programs are no longer available. The criteria for AUs are now specified in other areas of the rule.

Section 35.972, Recentness of training, was deleted in its entirety. The requirements of this section, with minor modifications, were moved to the new § 35.59, Recentness of training.

Section 35.980, Training for an authorized nuclear pharmacist, was deleted in its entirety. The requirements of this section, with minor modifications, were moved to the new § 35.55, Training for an authorized nuclear pharmacist.

Section 35.981, Training for experienced nuclear pharmacists, was deleted in its entirety. The Federal Register notice for the proposed rule solicited comments on the impact of deleting this section. All of the commenters that responded to this question said that there would be no impact if this section were deleted because the requirements for experienced nuclear pharmacists are covered in § 35.57, Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

Section 35.990, Violations, was deleted in its entirety, and the requirements of this section, with minor modifications, were moved to the new § 35.4001.

Section 35.991, Criminal penalties, was deleted in its entirety, and the requirements of this section, with minor modifications, were moved to the new § 35.4002.

Section 35.999, Resolution of conflicting requirements during transition period, was deleted in its entirety, and the requirements of this section, with modifications, were moved to the new § 35.10.

Subpart K, Other Medical Uses of Byproduct Material or Radiation from Byproduct Material, is a new subpart. This subpart includes all new medical uses of byproduct material or radiation from byproduct material, i.e., types of uses that are not regulated under Subparts D through H.

Section 35.1000, Other medical uses of byproduct material or radiation from byproduct material, is a new section. It has been added so that there are codified regulatory requirements and a more clearly defined process to obtain a license, or an amendment to a license, for a new medical use of byproduct material or radiation from byproduct material, i.e., an emerging technology. The specific information that must be provided to the Commission in support of an application for use under § 35.1000 is provided in § 35.12(d). The Commission intends to evaluate each application on a case-by-case basis and to work with the ACMUI, the medical community, and the developers of the new technology, as appropriate, to determine the risks associated with the technology and the appropriate regulatory requirements, including the training and experience requirements, for use of the technology.

Subpart L, Records, is a new subpart. This subpart contains all the specific recordkeeping requirements necessary to implement the requirements in Part 35. The general requirements for record maintenance, such as electronic storage, are provided in § 35.5. The records are grouped in one subpart to facilitate use by the licensees. A licensee may reference this subpart to determine whether something must be recorded, instead of having to review the entire regulation to find out if there is a particular recordkeeping requirement. Many of the

recordkeeping requirements remain unchanged from the current Part 35. However, some new sections have been added as a result of new requirements, especially in subpart H that codifies requirements for remote afterloaders and gamma stereotactic radiosurgery units that are currently imposed by license conditions.

Section 35.2024, Records of authority and responsibilities for radiation protection programs, requires the licensee to retain a record of actions taken by the licensee's management in accordance with § 35.24(a) for 5 years. The Commission believes that it is important to document the licensee's management review and approval of licensing actions and changes to the radiation protection program. The record of licensing actions and radiation protection program changes must include a summary of the actions taken and a signature of licensee management. The 5-year retention period is a reduction from the current requirements to maintain records of the approval of licensing actions, individuals, and radiation protection program changes. Similar records in the current §§ 35.23 and 35.31 are required to be maintained for the duration of the license. The 5-year retention period will decrease the recordkeeping burden on licensees and will also allow sufficient time for NRC to review records of licensee actions.

Paragraph (b) of this section requires the licensee to retain a current copy of the authorities, duties, and responsibilities of the RSO in accordance with § 35.24(e) and to retain a signed copy of the RSO's agreement to be responsible for implementing the radiation safety program, in accordance with § 35.24(b). These records must include the signatures of both the RSO and licensee management. The current Part 35 requires that the signed copy of the authorities, duties, and responsibilities of the RSO be retained until the Commission terminates the license. The final rule requires that only a current copy of this document be retained. This change decreases the recordkeeping burden on the licensees.

Section 35.2026, Records of radiation protection program changes, requires the licensee to retain a record of each radiation protection program change made in accordance with § 35.26 for 5 years. The record must include a copy of the old and new procedures, the effective date of the change, and the signature of the licensee management that reviewed and approved the change. The requirements in the current § 35.31 to include the reasons for the change, and a summary of radiation safety matters that were considered before making the change, have been deleted. The Commission recognizes that the requirement for management's signature is redundant with the requirement in § 35.2024. However, it believes this approach is warranted in light of the importance of these actions. This record is needed to document what radiation changes were made in the program to facilitate the Commission's evaluation of minor radiation safety program changes, and provides licensees with a record of the changes. Currently, licensees must retain a record of each "radiation safety program" change until the license has been renewed or terminated. Therefore, the 5-year retention period in the final rule represents a reduction in the licensee's recordkeeping burden.

Section 35.2040, Records of written directives, requires the licensee to retain a copy of written directives required by § 35.40 for 3 years. The final rule includes only minor changes to the specific items that must currently be recorded in written directives in accordance with § 35.32. These records will help to ensure that administrations are in accordance with the written directives. The 3-year recordkeeping retention period corresponds with the current retention period for written directives in § 35.32(d). These changes are discussed under § 35.40.

Section 35.2045, Records of medical events, requires that the licensee maintain a record of medical events reported in accordance with § 35.3045 for 3 years. This section is intended, in part, to replace the current recordkeeping requirements in § 35.33. The records made under § 35.3045 must

contain the licensee's name; the names of all the individuals involved; the affected or potentially affected individual's social security number or other identification number if one has been assigned; a brief description of the medical event and why it occurred; the effect, if any, on the individual; and the actions, if any, taken or planned to prevent recurrence. This record is needed to document medical events for licensee and Commission review. The requirement to maintain records of medical events is similar to the current requirement for maintaining records of misadministrations. However, the new requirement provides for a reduction in licensee burden because medical event records are required to be maintained for 3 years, rather than the 5-year requirement for records of misadministrations under the current § 35.33.

Section 35.2047, Record of a dose to an embryo/fetus or a nursing child, requires that the licensee maintain a record of events reported in accordance with § 35.3047 for 3 years. This is a new recordkeeping requirement in the final rule that has been added to correspond to the new reporting requirements in § 35.3047, Report and notification of a dose to an embryo/fetus or nursing child. The records made under § 35.3047 must contain the licensee's name; the names of all individuals involved; the affected or potentially affected individual's social security number or other identification number if one has been assigned; a brief description of the event and why it occurred; the effect, if any, on the embryo/fetus or nursing child; and the actions, if any, taken or planned to prevent recurrence. This record is needed to document these events for licensee and Commission review.

Section 35.2060, Records of calibrations of instruments used to measure the activity of unsealed byproduct material, requires the licensee to maintain a record of instrument calibrations performed in accordance with § 35.60 for 3 years. These records are required to document that the instruments are calibrated properly. This section replaces the requirements in the current § 35.50 (e) and adds recordkeeping requirements for

instruments used to measure the activity of dosages of nonphoton-emitting radionuclides. The prescriptive requirements for the record were deleted because licensees should have flexibility in determining how the results of the calibration are recorded. The final rule requires that the name of the individual who performed the calibration be documented in the record, rather than the initials of the individual who performed the constancy check and the identity of the individual for all other required tests. We believe this change is needed because recording the name of the individual who performed the calibration will better ensure future identification of the individual who performed the calibration. The change is also needed because it gives the licensee the flexibility of using paper records or computer-generated records. This requirement does not prohibit licensees from continuing to have the individual who performed the calibration sign the record. The 3-year recordkeeping retention period is consistent with the current retention period for instrument calibrations.

The final rule requires that the record contain the model and serial number of the instruments; the date of the calibration, the results of the calibration; and the name of the individual who performed the calibration.

Section 35.2061, Records of radiation survey instrument calibrations, requires the licensee to maintain a record of radiation survey instrument calibrations required by § 35.61 for 3 years. This record is needed to provide adequate documentation of instrument calibration. This section replaces the requirements in the current § 35.51(d). We deleted the requirement to include the descriptions of the calibration procedure and the source used; the certified exposure rates from the source and the rates indicated by the instrument being calibrated; and the correction factors deduced from the calibration

data. This revision is consistent with the revisions made to § 35.61. The 3-year recordkeeping retention period is consistent with the current retention period for instrument calibrations.

The final rule requires that the licensee record the model and serial number of the instrument; the date of the calibration; the results of the calibration; and the name of the individual who performed the calibration.

Section 35.2063, Records of dosage of unsealed byproduct material for medical use, requires the licensee to maintain a record of dosage determinations required by § 35.63 for 3 years. This record is needed to show that material has been administered to a patient or human research subject. This section replaces the requirements in the current § 35.53(c). Changes have been made from the current recordkeeping requirements for dosage measurement. We deleted the requirement to include the generic name, trade name, or abbreviation of the radiopharmaceutical; its lot number and expiration date; and the activity of the dosage at the time of measurement. With the exception of the expiration date, the requirements were deleted to make the rule less prescriptive. We deleted the expiration date because it is primarily related to drug stability and sterility. The term “dosage measurement” was replaced by the term “dosage determination” to be consistent with the changes made in § 35.63. Finally, a change was made to require that the name of the individual who determined the dosage be documented rather than the initial of the individual who made the record. We believe that this change is needed because recording the name of the individual will better ensure future identification of the individual who determined the dosage. The 3-year recordkeeping retention period corresponds with the current retention period for dosage records.

The final rule requires that licensees record the radiopharmaceutical; the patient's or human research subject's name, or identification number if one has been assigned; the prescribed dosage, the determined dosage, or a notation that the total activity is less than 1.1 MBq (30 uCi); the date and time of the dosage determination; and the name of the individual who determined the dosage.

Section 35.2067, Records of possession of sealed sources and brachytherapy sources, requires the licensee to retain records of the leak tests and inventory required by § 35.67(b) and (g), respectively, for 3 years. Leak test records are required to show that the leak test was done at the appropriate time interval and that sealed sources are not leaking. Inventory records are necessary to show that the possession of sealed sources did not exceed the amount authorized by the license. This section replaces the requirements in the current § 35.59(d) and (g). We have deleted the requirement to record the measured activity of each leak test sample and a description of the method used to measure each test sample. These changes were done to make the rule less prescriptive. We also revised the rule to require that the name of the individual performing the leak test and inventory be recorded rather than the signature of the RSO. We believe this change is needed because recording the name of the individual will ensure future identification of the individual who performed the leak test or inventory. The record retention period was reduced from 5 years to 3 years to reduce regulatory burden. The Commission does not believe the longer record retention period is warranted.

The final rule requires that leak test records must contain the model number, and serial number if one has been assigned, of each source tested; the identity of each source radionuclide and its estimated activity; the results of the test; the date of the test; and the name of the individual who performed

the test. Inventory records must contain the model number of each source, and serial number if one has been assigned; the identity of each source radionuclide and its nominal activity; the location of each source; and the name of the individual who performed the inventory.

Section 35.2070, Records of surveys for ambient radiation exposure rate, requires the licensee to maintain records of radiation surveys for 3 years. These records are needed to document that surveys were performed. This section replaces the requirements in the current § 35.70(h). We revised the current requirements to delete the need to record a plan of each area surveyed; the trigger level established for each area; and the detected dose rate at several points in each area expressed in millirem per hour or the removable contamination in each area expressed in disintegrations per minute per 100 square centimeters. These deletions were done to make the rule less prescriptive and to delete reference to surveys for removable contamination. The final rule requires that the name of the individual performing the survey be recorded rather than the initials of the individual. We believe this change is needed because recording the name of the individual will ensure easier identification of the individual who performed the survey. The 3-year recordkeeping retention period is consistent with the current retention period for radiation surveys.

The final rule requires that the record include the date of the survey; the results of the survey; the instrument used to make the survey; and the name of the individual who performed the survey.

Section 35.2075, Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material, requires the licensee to maintain records of patient release required by § 35.75 for 3 years. This record is needed to show compliance with the requirements in § 35.75. No changes have been made from the recordkeeping requirements in the current § 35.75 (c) and (d).

Section 35.2080, Records of administrative and technical requirements that apply to the provision of mobile medical services, requires the licensees to maintain a copy of the letter that permits the use of byproduct material at a client's address of use for 3 years after the last provision of service; and to retain the records of the surveys for 3 years. The records are needed to show compliance with the requirements in § 35.80. We deleted the requirements to record a plan of each area that was surveyed and the measured dose rate at several points in each area of use expressed in millirem per hour. This change was done to make the rule less prescriptive. The final rule requires that the name of the individual performing the survey rather than the initials of the individual be recorded. We believe this change is needed because recording the name of the individual will ensure easier identification of the individual who performed the survey.

Paragraph (a) of the final rule requires that the record include a copy of the letter(s) that permits the use of byproduct material at a client's address. Paragraph (b) requires that the record of each survey include the date of survey, the result of the survey, the instrument used to make the survey, and the name of the individual who performed the survey.

Section 35.2092, Records of decay-in-storage, requires the licensee to maintain records of the disposal of licensed materials made in accordance with § 35.92 for 3 years. This record is needed to document that radioactive material is not disposed of as ordinary waste. This section replaces the requirements in the current § 35.92 (b). We deleted the requirement to record the date that the material was placed in storage and the radionuclides because the requirement to store material for 10 half-lives was deleted. The 3-year recordkeeping retention period is consistent with the current retention period for waste disposal records.

The final rule requires that the record include the date of the disposal; the survey instrument used; the background dose rate; the dose rate measured at the surface of each waste container; and the name of the individual who performed the disposal.

Section 35.2204, Records of molybdenum-99 concentrations, requires the licensee to maintain a record of the molybdenum-99 concentration tests required by § 35.204(b) for 3 years. This record is needed to document that the concentration measurement was made and that the maximum molybdenum-99 concentration level was not exceeded. This section replaces the requirements in the current § 35.204 (c). We deleted the requirements to record the measured activity of the technetium expressed in millicuries and the measured activity of the molybdenum expressed in microcuries. The 3-year recordkeeping retention period is consistent with the current retention period for records of molybdenum-99 concentration.

The final rule requires that the record include, for each measured elution of technetium-99m, the ratio for the measures expressed as kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (microcuries of molybdenum per millicurie of technetium); the time and date of the measure; and the name of the individual who made the measurement.

Section 35.2310, Records of safety instruction, requires the licensee to maintain a record of radiation safety instructions required by §§ 35.310, 35.410, and 35.610 for 3 years. This record is needed to document that the instruction was given. This section replaces the requirements in §§ 35.310, 35.410, and 35.610. The rule has been revised to require that the licensee record the topics covered rather than a description of the instruction. We believe the term “description of the instruction” was too vague and could have been interpreted too broadly. For example, the licensee could question whether the rule required a listing of the topics or a general description, e.g., such as laboratory or classroom training. The change makes it clear that the record should contain the topics, e.g., patient, visitor, waste, or contamination control. The 3-year recordkeeping retention period is consistent with the current retention period for training records.

The final rule requires that the record include a list of the topics covered, the date of the instruction, the name(s) of the attendee(s), and the name(s) of the individual(s) who provided the instruction.

Section 35.2404, Records of surveys after source implant or removal, requires the licensee to maintain a record of the surveys required by §§ 35.404 and 35.604 for 3 years. The licensee is no longer specifically required to record the dose rate from the patient or the human research subject expressed as millirem per hour and measured at 1 meter from the patient or human research subject. Each record must include the date and results of

the survey, the survey instrument used, and the name of the individual who made the survey. These records are used to show that sources have not been misplaced and that all sources have been removed from the patient. The 3-year recordkeeping retention period is consistent with the current retention period for surveys found in Part 20.

Section 35.2406, Records of brachytherapy source accountability, requires the licensee to maintain a record of brachytherapy source accountability required by § 35.406 for 3 years. Changes have been made in the recordkeeping requirements found in the current rule. The licensee is no longer required to record the following items because they were deleted from § 35.406: the names of the individuals permitted to handle the sources; name and room number of the patient or the human research subject receiving the implant; number and activity of the sources in storage after the removal; and the number and activity of sources in storage after the return.

The final rule requires that, for temporary implants, the record must include the number and activity of sources removed from and returned to storage; the time and date they were removed from and returned to storage; the name(s) of the individual(s) who removed them from and returned them to storage; the location of use; the time and date they were returned to storage. For permanent implants, the record must include the number and activity of sources removed from storage; the number and activity of sources permanently implanted in the patient or human research subject; the number and activity of sources not implanted; the date they were removed from and returned to storage; and the name(s) of the individual(s) who removed and returned the sources to storage. This record is required so that if a brachytherapy source is misplaced or missing the licensee is immediately alerted and can take appropriate action,. The 3-year recordkeeping retention period is consistent with the current retention period for inventory records.

Section 35.2432, Records of calibration measurements of brachytherapy sealed sources, requires the licensee to retain a record of the results of brachytherapy source calibrations required by § 35.432 for 3 years after the last use of the source. This is a new recordkeeping section. The record must contain the date of the calibration; the manufacturer's name, model number, and serial number for the source and instruments used to calibrate the source; the source output or activity; the source positioning accuracy within the applicators; and the signature of the AMP. These records are needed to document that the brachytherapy sources have been calibrated.

Section 35.2433, Records of decay of strontium-90 sources for ophthalmic treatments, requires the licensee to maintain a record of the activity of a strontium-90 source, as required by § 35.433, for the life of the source. This is a new recordkeeping section. The records for each strontium-90 source must include the initial activity of the source and date; and, for each decay calculation, the date and the source activity as determined under § 35.432. These records are needed to document that the treatment times for ophthalmic uses of strontium-90 are based on properly decayed sources.

Section 35.2605, Records of installation, maintenance, adjustment, and repair, requires the licensee to retain a record of the installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units, as required by § 35.605, for 3 years. This is a new recordkeeping section. Previously, licensees were not required to keep records of installation, maintenance, adjustment, and repair. For each installation, maintenance, adjustment, and repair, the record must include the date, description of the service, and name(s) of the individual(s) who performed the work. This record is necessary to document that the units are properly installed, maintained, adjusted, and repaired; to establish trends in unit performance; and to establish a service history that may be used in evaluation of generic equipment problems.

Section 35.2630, Records of dosimetry equipment, requires the licensee to retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with § 35.630 for the duration of the license. Some changes have been made in the recordkeeping requirements from the current rule. For example, a requirement, similar to requirements for other instruments, has been added to record the manufacturer's name of the instruments that were calibrated. These records are needed to show that calibrations of medical units were made with properly calibrated instruments.

Section 35.2632, Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations, requires the licensee to maintain a record of the full calibrations required by §§ 35.632, 35.633, and 35.635 for 3 years. The record retention period was decreased from the duration of the use of the unit's source to 3 years to reduce regulatory burden. The term "teletherapy physicist" was replaced with the term "authorized medical physicist." In addition, the current recordkeeping requirements for this section were reduced to recording the date of the calibration; manufacturer's name, model number, and serial number for the unit, source and instruments used to calibrate the unit; the results and assessment of the calibration; the results of the autoradiograph required for low dose-rate remote afterloader units; and the signature of the AMP who performed the full calibration. These records are needed to document that calibrations were performed in accordance with §§ 35.632, 35.633, and 35.635.

Section 35.2642, Records of periodic spot-checks for teletherapy units, requires the licensee to retain a record of each periodic spot-check for teletherapy units required by § 35.642 for 3 years. Minor changes have been made in the recordkeeping requirements from the current rule. For instance, the licensee is no longer required to record the operability of the beam condition indicator light, but is required to record the operability of the

source exposure indicator light. This change reflects corresponding changes made in § 35.642. These records are needed to document that spot-checks were performed in accordance with § 35.642. The 3-year recordkeeping retention period is consistent with the current retention period for periodic spot-checks.

Section 35.2643, Records of periodic spot-checks for remote afterloader units, requires the licensee to retain a record of each spot-check for remote afterloader units required by § 35.643 for 3 years. This is a new recordkeeping section. The record must include the date of the spot-check; the manufacturer's name, model number, and serial number for both the remote afterloader unit and source; an assessment of timer accuracy; notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom systems, clock and decayed source activity in the unit's computer; the name of the individual who performed the periodic spot-check; and the signature of the AMP who reviewed the record of the spot-check. These records are needed to document that spot-checks were performed in accordance with § 35.643

Section 35.2645, Records of periodic spot-checks for gamma stereotactic radiosurgery units, requires the licensee to retain a record of each spot-check for gamma stereotactic radiosurgery units required by § 35.645 for 3 years. This is a new recordkeeping section. The record must include the date of the spot-check; the manufacturer's name, model number, and serial number for the gamma stereotactic radiosurgery unit and the instrument used to measure the output of the unit; an assessment of timer linearity and accuracy; the calculated on-off error; a determination of trunnion centricity; the difference between the anticipated output and the measured output; an assessment of source output against computer calculations; notations indicating the operability of radiation monitors, helmet microswitches, emergency timing circuits, emergency off buttons, electrical interlocks, source exposure

indicator lights, viewing and intercom systems, timer termination, treatment table retraction mechanism, stereotactic frames and localizing devices (trunnions); the name of the individual who performed the periodic spot-check; and the signature of the AMP who reviewed the periodic spot-check. This record is needed to show that spot-checks were performed in accordance with § 35.645.

Section 35.2647, Records of additional technical requirements for mobile remote afterloader units, requires the licensee to retain a record of each check for mobile remote afterloader units required by § 35.647 for 3 years. This is a new recordkeeping section. The record must include the date of the check; the manufacturer's name, model number, and serial number for the remote afterloader unit; notations accounting for all sources before departing from a facility; notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom system, applicators, source transfer tubes, and transfer tube applicator interfaces, and source positioning accuracy; and the signature of the individual who performed the check. This record is needed to show that required spot-checks were performed in accordance with § 35.647 and that the unit is operable.

Section 35.2652, Records of surveys of therapeutic treatment units, requires the licensee to maintain a record of radiation surveys made in accordance with § 35.652 for the duration of use of the unit. This recordkeeping requirement has been changed to require that the records of radiation surveys of the treatment unit must be maintained for the duration of use of the unit, rather than for the duration of the license, to reduce regulatory burden. In addition, the licensee is no longer required by this section to maintain a plan of the areas surrounding the treatment room that were surveyed, the measured dose rate at several points in each area expressed in millirem per hour, and the calculated maximum quantity of radiation over a period of 1

week for each restricted and unrestricted area. This change reflects corresponding changes made in § 35.652. The record must include the date of the measurements; the manufacturer's name, model number and serial number of the treatment unit, source, and instrument used to measure radiation levels; each dose rate measured around the source while the unit is in the off position and the average of all measurements; and the signature of the individual who performed the surveys. This record is needed to document radiation levels in areas surrounding therapeutic devices in accordance with § 35.652.

Section 35.2655, Records of 5-year inspection for teletherapy and gamma stereotactic radiosurgery units, requires the licensee to maintain a record of the 5-year inspection for teletherapy and gamma stereotactic radiosurgery units required by § 35.655 for the duration of the unit. This recordkeeping requirement has been changed to require that the records of inspections of the treatment units must be maintained for the duration of use of the unit, rather than for the duration of the license, to reduce the regulatory burden. A minor change was made to delete the requirement to maintain a record of the components replaced to also reduce the regulatory burden. The record must contain the inspector's radioactive materials license number; the date of inspection; the manufacturer's name, model number and serial number for both the treatment unit and source; a list of components inspected and serviced; the type of service; and the signature of the inspector. This record is needed to document the type of service that was performed in accordance with § 35.655.

Subpart M, Reports, is a new subpart in Part 35. This subpart contains all the reporting requirements necessary to implement the requirements in Part 35. Grouping of reporting requirements into one subpart was done to facilitate use by licensees. A licensee may reference this section when

determining whether something must be reported, rather than having to review the entire regulation to find out if there is a particular reporting requirement. Two of the reporting requirements appear in the current §§ 35.33 and 35.59. A third reporting requirement was added so that the NRC can comply with the requirement to submit an annual report to Congress of unscheduled incidents or events which the Commission considers significant from the standpoint of public health and safety.

Section 35.3045, Report and notification of a medical event, provides criteria for reporting and notifying individuals about a medical event. The requirements in the final rule are based on the current requirements in § 35.33, Notifications, reports, and records of misadministrations. Changes were made to make the reporting threshold dose-based where possible; to add a dose threshold of 0.5 Sv (50 rem) shallow dose equivalent to the skin; and to address two areas that have caused problems in implementing the current requirements for reporting misadministrations -- patient intervention and wrong treatment site.

Patient intervention is not specifically addressed in the current rule. However, a licensee is expected to act reasonably, in accordance with prevailing standards of care, to prevent patient intervention from causing a misadministration. This situation has resulted in numerous debates over whether or not a licensee had done everything it should to prevent patient intervention during treatment. In order to correct the current situation, we have defined patient intervention to mean intentional or unintentional actions taken by a patient or human research subject such as dislodging or removing treatment devices or prematurely terminating the administration. We have also added a specific requirement for reporting medical events that occur as a result of patient intervention. Licensees are required to report any event resulting from intervention of a patient or human research subject in which the

administration of byproduct material or radiation from byproduct material results or will result in unintended permanent functional damage to an organ or a physiological system, as determined by a physician. This reporting requirement should result in minimal regulatory burden on licensees because in most situations where patients or human research subjects intervene, either voluntarily or involuntarily, in their treatment there is no resultant permanent medical damage. Even though there is a high threshold for reporting in the final rule, licensees are expected to continue to act reasonably, as required under the current rule, to prevent medical events caused by patient intervention.

The final rule includes specific criteria for determining when a dose to a wrong treatment site is a reportable medical event: a dose to the skin or an organ or tissue other than the treatment site that exceeds by 0.5 Sv (50 rem) to an organ or tissue and 50 percent or more of the dose expected from the administration defined in the written directive (excluding, for permanent implants, seeds that were implanted in the correct site but migrated outside the treatment site).

The final rule retains the current requirement in § 35.33 that licensees notify the NRC Operations Center, by telephone, no later than the next calendar day after discovery of the medical event. The final rule also retains the current requirement for licensees to submit a written report to the appropriate NRC Regional Office listed in 10 CFR 30.6 within 15 days after discovery of the medical event. In addition, the licensee is required to notify the referring physician and the individual affected by the medical event, or the responsible relative or guardian, no later than 24 hours after its discovery, unless the referring physician personally informs the licensee either that he will inform the individual or that, based on medical judgment, telling the

individual would be harmful. This reporting requirement is needed to ensure that NRC is aware of medical events. Refer to Section III of the Supplementary Information of this document for additional information on the notification requirements in § 35.3045.

One change was made in the current requirement for a written report to be provided to the affected individual within 15 days of discovery of the medical event. In the current rule, licensees can provide the individual with a brief description of both the event and the consequences as they may affect the individual, if they include a statement that the individual can also obtain from the licensee a copy of the report that was submitted to the NRC. In the final rule, the licensee is not required to include this statement because knowledge that a report had to be submitted to the NRC might unduly alarm an individual involved in a medical event, with no added benefit. However, licensees are expected to provide the affected individual with a written report that is comprehensive enough to provide the individual with enough information about the effects, if any, of the medical event to enable him or her to make any decisions about remedial or prospective health care.

Section 35.3047, Report and notification of a dose to an embryo/fetus or a nursing child, is a new section. Paragraph (a) requires that a licensee report to NRC any administration of byproduct material, or radiation from byproduct material, to a pregnant female that results in a dose to an embryo/fetus that is greater than 50 mSv (5 rem) dose equivalent unless the administration was specifically approved, in advance, by the AU. It should be emphasized that only unintended exposures are required to be reported to NRC.

Paragraph (b) requires that a licensee report to NRC any administration of byproduct material to a breast feeding woman that results in a dose to a nursing child that is greater than 50 mSv (5 rem) total effective dose equivalent or a dose that has resulted in unintended permanent functional damage to an organ or a physiological system of the child, as determined by a physician, unless the administration was specifically approved, in advance, by the AU.

The reporting requirements in this section are similar to the reporting requirements for medical events. The final rule requires that licensees notify the NRC Operations Center, by telephone, no later than the next calendar day after discovery of a dose to an embryo/fetus or a nursing child that requires a report. The licensee is required to submit a written report to the appropriate NRC Regional Office listed in 10 CFR 30.6 within 15 days after discovery of a dose to an embryo/fetus or a nursing child. In addition, the licensee is required to notify the referring physician and the pregnant individual or mother within 5 days after its discovery, unless the referring physician personally informs the licensee either that he/she will inform the mother or that, based on medical judgment, telling the mother would be harmful. Refer to Section III of the Supplementary Information of this document for additional information on the notification requirements in § 35.3047.

Information required by this section is needed so that the NRC can comply with Section 208 of the Energy Reorganization Act of 1974 (Public Law 93-438), as amended, to submit an annual report to Congress of unscheduled incidents or events which the Commission considers significant from the standpoint of public health and safety, e.g., abnormal occurrences.

NRC identifies an abnormal occurrence using the revised abnormal occurrence criteria that were published in the Federal Register on April 17, 1997 (62 FR 18820). Section II of the policy statement defines unintended radiation exposure as “any occupational exposure, exposure to the general public or exposure as a result of a medical misadministration (as defined in § 35.2) involving the wrong individual that exceeds the reporting values established in the regulations.” This section also states that “All other reported medical misadministrations will be considered for reporting as an Abnormal Occurrence under the criteria for medical licensees. In addition, unintended radiation exposures include any exposure to a nursing child, fetus, or embryo as a result of an exposure (other than an occupational exposure to an undeclared pregnant woman) to a nursing mother or pregnant woman above specified values.” Appendix A, Section I. A, of the policy statement, states that NRC will provide information on “any unintended radiation exposure to any minor (an individual less than 18 years of age) resulting in an annual total effective dose equivalent of 50 mSv (5 rem) or more, or to an embryo/fetus resulting in a dose equivalent of 50 mSv (5 rem) or more.”

At the present time, the NRC has no regulatory requirements that require reporting of those types of events. The Commission considered two alternatives that could be pursued: revise the current Abnormal Occurrence Criteria to delete the requirement to inform Congress of this type of event; or develop a reporting requirement that would provide the information needed by the Commission to comply with Section 208. The Commission did not pursue the first option because the Abnormal Occurrence reporting criteria were recently reviewed and revised.

The Commission recognizes that the standard of practice for AUs is to assess the pregnancy or nursing status of their patients (reference American College of Radiology “Standard for the Performance of Therapy with Unsealed Radionuclide Sources,” 1996, and “Society of Nuclear Medicine

General Procedure Guidelines for Imaging with Radionuclides,” 1997). As a result, the NRC does not believe that it is appropriate to have a rule that requires a licensee to assess the pregnancy or nursing status of patients prior to a medical treatment involving byproduct material. However, it does believe that it is appropriate to require the licensee to inform the NRC when the licensee learns of an unintended dose to an embryo/fetus or a nursing child that exceeds the thresholds in § 35.3047. For example, a licensee must report an unintended dose resulting from an individual not disclosing her pregnancy or nursing status at the time of administration of the byproduct material or radiation from byproduct material. In this situation, the unintended dose could have been prevented if the AU had followed the standard of practice, noted above, to assess the pregnancy status of the patient. The occurrence of such an incident does not necessarily mean that the licensee is in violation of the requirements in Part 35, as long as the licensee reports it and it is not otherwise in violation of NRC regulatory requirements. For example, a reportable dose to a nursing child under § 35.3047 is not necessarily subject to enforcement action if the licensee has complied with § 35.75.

However, NRC acknowledges that, in some cases, the licensee might not be able to prevent the dose to an embryo/fetus or nursing child. Such cases are not reportable under § 35.3047. For example, there is no way for an AU to prevent administration of an unintended dose to an embryo/fetus if the pregnancy test was negative because it was given very early in the pregnancy.

Section 35.3067, Report of a leaking source, requires the licensee to file a report with the appropriate NRC Office listed in § 30.6 of this chapter, with a copy to the Director, Office of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, within

5 days if a leak test required by § 35.67 reveals the presence of 185 Bq (0.005 microcurie) or more of removable contamination. This reporting requirement is similar to the reporting requirements for leaking sources in the current § 35.59, but the final rule does not require that as much prescriptive information be included in the record. The report must contain the model number and serial number, if assigned, of the leaking source; the radionuclide and its estimated activity; the results of the test; the date of the test; and the action taken.

Subpart N, Enforcement, contains statements regarding enforcement. This subpart contains the statements in the current Subpart K, Enforcement.

Section 35.4001, Violations, is a new section that replaces the current § 35.990 which was deleted. Other than changing the number of this section to reflect the new numbering system, no changes were made in the current statements regarding violations.

Section 35.4002, Criminal penalties, is a new section that replaces the current § 35.991 which was deleted. Other than changing the numbers of this section and the sections referenced under paragraph (b) to reflect the new numbering system, no changes were made in the current statements regarding criminal penalties.

VI. Coordination with the Advisory Committee on Medical Uses of Isotopes.

To be completed after receipt of SRM on draft final rule.

VII. Coordination with NRC Agreement States.

To be completed after receipt of SRM on draft final rule.

VIII. Consistency with Medical Policy Statement.

To be completed after receipt of SRM on draft final rule.

IX. Implementation.

To be completed after receipt of SRM on draft final rule.

X. ISSUES OF COMPATIBILITY FOR AGREEMENT STATES

Under the “Policy Statement on Adequacy and Compatibility of Agreement State Programs” approved by the Commission on June 30, 1997 (62 FR 46517), specific requirements within this rule should be adopted by Agreement States for purposes of compatibility or because of their health and safety significance. Implementing procedures for the Policy Statement establish specific categories which have been applied to categorize the requirements in Parts 20, 32, and 35. A Compatibility Category “A” designation means the requirement is a basic radiation protection standard or deals with related definitions, signs, labels or terms necessary for a common understanding of radiation protection principles. Compatibility Category “A” designated Agreement State requirements should be essentially identical to those of the NRC. Compatibility Category “B” designation means the requirement has significant direct transboundary implications. A Compatibility Category “B” designated Agreement State requirements should be essentially identical to those of the NRC. A Compatibility Category “C” designation means the essential objectives of the requirement should be adopted by the State to avoid conflicts, duplications, or gaps. The manner in which the essential objectives are addressed in the Agreement State requirement need not be the same as NRC provided the essential objectives are met. A Compatibility Category “D” designation means the requirement does not need to be adopted by an Agreement State for purposes of compatibility. The Compatibility Category Health and Safety (H&S) identifies requirements that are not required for compatibility, but which have particular health and safety significance. Agreement States should adopt the essential objectives of such requirements in order to maintain an adequate program.

The following discussion identifies the compatibility designations for each part.

PART 20 - STANDARDS FOR RADIATION PROTECTION

Section 20.1002, "Scope" is assigned to Compatibility Category "D."

Section 20.1003, "Definitions," of occupational dose and public dose is assigned to Compatibility Category "A."

Section 20.1301(a) and 20.1301(a)(3), "Dose limits for individual members of the public" is assigned to Compatibility Category "A".

Part 32 - SPECIFIC DOMESTIC LICENSES TO MANUFACTURE OR TRANSFER CERTAIN ITEMS CONTAINING BYPRODUCT MATERIAL

Section 32.72 (b)(1) and (b)(2)(ii), "Manufacture, preparation, or transfer for commercial distribution of radioactive drugs containing byproduct material for medical use under Part 35"; and Section 32.74 (a) and (a)(3), "Manufacture and distribution of sources or devices containing byproduct material for medical use" are assigned to Compatibility Category "B".

10 CFR PART 35 - MEDICAL USES OF BYPRODUCT MATERIAL

Subpart A, "General Information" § 35.2, "Definitions" are assigned to Compatibility Category "D" with the exception of the terms "Agreement State, authorized medical physicist, authorized nuclear pharmacist, authorized user, medical use, prescribed dosage, prescribed dose, radiation safety officer,

sealed source and treatment site.” The terms “Agreement State and sealed source” are assigned to Compatibility Category “B”, because they have significant direct transboundary implications. The terms “authorized medical physicist, authorized nuclear pharmacist, authorized user, medical use, prescribed dosage, prescribed dose, radiation safety officer and treatment site” have been assigned to Compatibility Category “C.” Section 35.6, “Provisions for protection of human research subjects”; and Section 35.11, “License required” are assigned to Compatibility Category “C.”

Subpart B, “General Administrative Requirements,” is assigned to Compatibility Category “D,” with the exception of ten sections. Section 35.24 (b) and (f), “Authority and responsibilities for the radiation protection program”; § 35.27, “Supervision”; § 35.40 (a) and (b) “Written directive”; and § 35.41(a), “Procedures for administrations requiring a written directive”; which are all assigned to Compatibility Category D/H&S. Section 35.49, “Suppliers for sealed sources or devices for medical use”; § 35.50, “Training for radiation safety officer”; § 35.51, “Training for an authorized medical physicist”; § 35.55, “Training for an authorized nuclear pharmacist”; § 35.57, “Training for experienced radiation safety officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist ”; and § 35.59, “Recentness of training” are assigned to Compatibility Category “C”.

Subpart C, “General Technical Requirements,” is assigned to Compatibility Category “D”, with the exception of nine sections. Section 35.60(a) and (b), “Possession, use, and calibration of instruments to measure the activity of unsealed byproduct material”; § 35.61(a)(1) and (2), (b) and (c), “Calibration of survey instruments”; § 35.63(a),(b),(c) and (d), “Determination of dosages of unsealed byproduct material for medical use”; § 35.67(a), (b), (c), (d), (e) and (g), “Requirements for possession of sealed sources and brachytherapy sources”; § 35.69, “Labeling of vials and syringes”; § 35.70(a), “Surveys of ambient radiation exposure rate”; § 35.80(a)(2), (a)(3) and (a)(4), “Provision of mobile service”

(for those States that authorize this activity); and § 35.92, “Decay-in-storage” (for those States that authorize this activity) are assigned to Compatibility Category D/H&S. Section 35.75(a) and (b), “Release of individuals containing unsealed byproduct material or implants containing byproduct material” is assigned to Compatibility Category “C”.

Subpart D, “Unsealed Byproduct Material - Written Directive Not Required”; and Subpart E, “Unsealed Byproduct Material - Written Directive Required” are assigned to Compatibility Category “D”, except for Section 35.100, “Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required”; § 35.200, “Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required”; § 35.204(a) and (b), “Permissible molybdenum-99 concentration”; and § 35.300, “Use of unsealed byproduct material for which a written directive is required”; § 35.310(a), “Safety instruction”; and § 35.315, “Safety precautions” which are assigned to Compatibility Category D/H&S. Sections 35.190, “Training for uptake, dilution, and excretion studies”; § 35.290, “Training for imaging and localization studies”; § 35.390, “Training for use of unsealed byproduct material for which a written directive is required”; § 35.392, “Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries)” is assigned to Compatibility Category “C”; and § 35.394, “Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries)” are assigned to Compatibility Category “C”.

Subpart F, "Manual Brachytherapy" is assigned to Compatibility Category "D", with the exception of ten sections. Section 35.404 (a) and (b) "Surveys after source implant and removal"; § 35.406(a) and (b), "Brachytherapy sources accountability"; § 35.410(a), "Safety instruction"; § 35.415, "Safety precautions"; § 35.432 (a-e), "Calibration measurements of brachytherapy sealed sources"; § 35.433(a), "Decay of strontium-90 sources for ophthalmic treatments"; and § 35.457, "Therapy-related computer systems" are assigned to Compatibility Category D/H&S. Section 35.400, "Use of sealed sources for manual brachytherapy"; § 35.490, "Training for use of manual brachytherapy sources"; and § 35.491, "Training for ophthalmic use of strontium-90" are assigned to Compatibility Category "C".

Subpart G, "Sealed Sources for Diagnosis" Section 35.500, "Use of sealed sources for diagnosis" and § 35.590, "Training for use of sealed sources for diagnosis" are assigned to Compatibility Category "C".

Subpart H, "Therapeutic Medical Devices" is assigned to Compatibility Category "D" with the exception of seventeen sections. The following sections are assigned to the Health and Safety Category: §§ 35.604(a); 35.605(a-c); 35.610(a-e); 35.615; 35.630(a) and (b); 35.632(a-f); 35.633(a-g); 35.635(a-f); 35.642(a-e); 35.643(a-d); 35.645(a-f); 35.647(a-d) for States that authorize this activity; 35.652(a) and (b); 35.655(a) and (b); and 35.657. Section 35.600, "Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit"; and § 35.690, "Training for use of remote afterloader units, teletherapy units, or gamma stereotactic radiosurgery units" are assigned to Compatibility Category "C".

Subpart K, "Other Medical Uses of Byproduct Material or Radiation from Byproduct Material," and Subpart L, "Records" are assigned to Compatibility Category "D". Subpart M, "Reports", Section 35.3045(a) "Report and notification of a medical event"; § 35.3047(a), (b), (d), (e), (f), (g) and (h), "Report and notification of a dose to an embryo/fetus or a nursing child" and § 35.3067, "Report of a leaking source" are assigned to Compatibility Category "C". Sections 35.3045(b) and (f); and § 35.3047(c) are assigned to Compatibility Category "D".

Subpart N, "Enforcement," is assigned to Compatibility Category "D".

XI. Final Environmental Assessment: Availability.

To be completed after receipt of SRM on draft final rule.

XII. Paperwork Reduction Act Statement.

To be completed after receipt of SRM on draft final rule.

XIII. Regulatory Analysis.

To be completed after receipt of SRM on draft final rule.

XIV. Regulatory Flexibility Certification.

To be completed after receipt of SRM on draft final rule.

XV. Backfit Analysis.

To be completed after receipt of SRM on draft final rule.

XVI. Small Business Regulatory Enforcement Fairness Act.

To be completed after receipt of SRM on draft final rule.

XVII. Criminal Penalties.

To be completed after receipt of SRM on draft final rule.

List of Subjects

10 CFR Part 20

Byproduct material, Criminal penalties, Licensed material, Nuclear materials, Nuclear power plants and reactors, Occupational safety and health, Packaging and containers, Radiation protection, Reporting and recordkeeping requirements, Special nuclear material, Source material, Waste treatment and disposal.

10 CFR Part 32

Byproduct material, Criminal penalties, Labeling, Nuclear materials, Radiation protection, Reporting and recordkeeping requirements.

10 CFR Part 35

Byproduct material, Criminal penalties, Drugs, Health facilities, Health professions, Medical devices, Nuclear materials, Occupational safety and health, Radiation protection, Reporting and recordkeeping requirements.

For the reasons set out in the preamble and under the authority of the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974, as amended, and 5 U.S.C. 552 and 553, the NRC is adopting the following amendments to 10 CFR parts 20, 32 and 35.

PART 20--STANDARDS FOR PROTECTION AGAINST RADIATION

1. The authority citation for Part 20 continues to read as follows:

AUTHORITY: Secs. 53, 63, 65, 81, 103, 104, 161, 182, 186, 68 Stat. 930, 933, 935, 936, 937, 948, 953, 955, as amended, sec. 1701, 106 Stat. 2951, 2952, 2953 (42 U.S.C. 2073, 2093, 2095, 2111, 2133, 2134, 2201, 2232, 2236, 2297f), secs. 201, as amended, 202, 206, 88 Stat. 1242, as amended, 1244, 1246 (42 U.S.C. 5841, 5842, 5846).

2. Section 20.1002 is revised to read as follows:

§ 20.1002 Scope.

The regulations in this part apply to persons licensed by the Commission to receive, possess, use, transfer, or dispose of byproduct, source, or special nuclear material or to operate a production or utilization facility under parts 30 through 35, 39, 40, 50, 60, 61, 70, or 72 of this chapter. The limits in this part do not apply to doses due to background radiation, to exposure of patients to radiation for the purpose of medical diagnosis or therapy, to

exposure from individuals administered radioactive material and released, which is governed by § 35.75, or to exposure from voluntary participation in medical research programs.

3. In § 20.1003, the definitions for occupational dose and public dose are revised to read as follows:

§ 20.1003 Definitions

* * * * *

Occupational dose means the dose received by an individual in the course of employment in which the individual's assigned duties involve exposure to radiation or to radioactive material from licensed and unlicensed sources of radiation, whether in the possession of the licensee or other person. Occupational dose does not include doses received from background radiation, from any medical administration the individual has received, from exposure to individuals administered radioactive material and released, which is governed by § 35.75, from voluntary participation in medical research programs, or as a member of the public.

* * * * *

Public dose means the dose received by a member of the public from exposure to radiation or to radioactive material released by a licensee, or to any other source of radiation under the control of a licensee. Public dose does not include occupational dose or doses received from background

radiation, from any medical administration the individual has received, from exposure to individuals administered radioactive material and released, which is governed by § 35.75, or from voluntary participation in medical research programs.

* * * * *

4. In § 20.1301, paragraph (a) is revised, paragraphs (c), (d), and (e) are redesignated as paragraphs (d), (e), and (f), and a new paragraph (c) is added to read as follows:

§ 20.1301 Dose limits for individual members of the public.

(a) Each licensee shall conduct operations so that --

(1) The total effective dose equivalent to individual members of the public from the licensed operation does not exceed 1 millisievert (0.1 rem) in a year, exclusive of the dose contributions from background radiation, from any medical administration the individual has received, from exposure to individuals administered radioactive material and released, which is governed by § 35.75, from voluntary participation in medical research programs, and from the licensee's disposal of radioactive material into sanitary sewerage in accordance with § 20.2003, and

* * * * *

(c) Notwithstanding paragraph (a)(1) of this section, a licensee may permit visitors to individuals who cannot be released in accordance with § 35.75 to receive a radiation dose greater than (1 mSv) 0.1 rem if--

(1) The radiation dose received does not exceed (5 mSv) 0.5 rem; and

(2) The authorized user, as defined in 10 CFR Part 35, determines that it is appropriate.

* * * * *

**PART 32--SPECIFIC DOMESTIC LICENSES TO MANUFACTURE OR TRANSFER
CERTAIN ITEMS CONTAINING BYPRODUCT MATERIAL**

5. The authority citation for Part 32 continues to read as follows:

AUTHORITY: Secs. 81, 82, 161, 182, 183, 68 Stat. 935, 948, 953, 954, as amended (42 U.S.C. 2111, 2201, 2232, 2233); sec. 201, 88 Stat. 1242, as amended (42 U.S.C. 5841).

§ 32.72 [Amended]

6. In § 32.72, in paragraph (b)(1), the reference to "10 CFR 35.25" is revised to read "10 CFR 35.27" and in paragraph (b)(2)(ii), the reference to "10 CFR 35.980(b) and 35.972" is revised to read "10 CFR 35.55(b) and 35.59."

§ 32.74 [Amended]

7. In § 32.74, in the introductory text of paragraph (a), the reference to "§§ 35.400 and 35.500" is revised to read "§§ 35.400, 35.500, and 35.600" and in paragraph (a)(3), the reference to "§§ 35.57, 35.400, or 35.500" is revised to read "§§ 35.65, 35.400, 35.500, and 35.600."

8. 10 CFR Part 35 is revised to read as follows:

PART 35--MEDICAL USE OF BYPRODUCT MATERIAL

Subpart A-- General Information

- 35.1 Purpose and scope.
- 35.2 Definitions.
- 35.5 Maintenance of records.
- 35.6 Provisions for the protection of human research subjects.
- 35.7 FDA, other Federal, and State requirements.
- 35.8 Information collection requirements: OMB approval.
- 35.10 Implementation.
- 35.11 License required.
- 35.12 Application for license, amendment, or renewal.

- 35.13 License amendments.
- 35.14 Notifications.
- 35.15 Exemptions regarding Type A specific licenses of broad scope.
- 35.18 License issuance.
- 35.19 Specific exemptions.

Subpart B-- General Administrative Requirements

- 35.24 Authority and responsibilities for the radiation protection program.
- 35.26 Radiation protection program changes.
- 35.27 Supervision.
- 35.40 Written directives.
- 35.41 Procedures for administrations requiring a written directive.
- 35.49 Suppliers for sealed sources or devices for medical use.
- 35.50 Training for Radiation Safety Officer.
- 35.51 Training for an authorized medical physicist.
- 35.55 Training for an authorized nuclear pharmacist.
- 35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

35.59 Recentness of training.

Subpart C-- General Technical Requirements

35.60 Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material.

35.61 Calibration of survey instruments.

35.63 Determination of dosages of unsealed byproduct material for medical use.

35.65 Authorization for calibration, transmission, and reference sources.

35.67 Requirements for possession of sealed sources and brachytherapy sources.

35.69 Labeling of vials and syringes.

35.70 Surveys of ambient radiation exposure rate.

35.75 Release of individuals containing unsealed byproduct material or implants containing byproduct material.

35.80 Provision of mobile medical service.

35.92 Decay-in-storage.

Subpart D-- Unsealed Byproduct Material - Written Directive Not Required

35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

35.190 Training for uptake, dilution, and excretion studies.

35.200 Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

35.204 Permissible molybdenum-99 concentration.

35.290 Training for imaging and localization studies.

Subpart E-- Unsealed Byproduct Material -Written Directive Required

35.300 Use of unsealed byproduct material for which a written directive is required.

35.310 Safety instruction.

35.315 Safety precautions.

35.390 Training for use of unsealed byproduct material for which a written directive is required.

35.392 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).

35.394 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).

Subpart F-- Manual Brachytherapy

35.400 Use of sealed sources for manual brachytherapy.

35.404 Surveys after source implant and removal.

35.406 Brachytherapy sources accountability.

35.410 Safety instruction.

35.415 Safety precautions.

35.432 Calibration measurements of brachytherapy sealed sources.

35.433 Decay of strontium-90 sources for ophthalmic treatments.

35.457 Therapy-related computer systems.

35.490 Training for use of manual brachytherapy sources.

35.491 Training for ophthalmic use of strontium-90.

Subpart G-- Sealed Sources for Diagnosis

35.500 Use of sealed sources for diagnosis.

35.590 Training for use of sealed sources for diagnosis.

Subpart H-- Photon Emitting Remote Afterloader Units, Teletherapy Units, and Gamma Stereotactic Radiosurgery Units

35.600 Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit.

35.604 Surveys of patients and human research subjects treated with a remote afterloader unit.

35.605 Installation, maintenance, adjustment, and repair.

35.610 Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

35.615 Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

35.630 Dosimetry equipment.

35.632 Full calibration measurements on teletherapy units.

35.633 Full calibration measurements on remote afterloader units.

35.635 Full calibration measurements on gamma stereotactic radiosurgery units.

35.642 Periodic spot-checks for teletherapy units.

35.643 Periodic spot-checks for remote afterloader units.

35.645 Periodic spot-checks for gamma stereotactic radiosurgery units.

35.647 Additional technical requirements for mobile remote afterloader units.

35.652 Radiation surveys.

35.655 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units.

35.657 Therapy-related computer systems.

35.690 Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Subpart I-- Reserved

Subpart J-- Reserved

Subpart K-- Other Medical Uses of Byproduct Material or Radiation from Byproduct Material

35.1000 Other medical uses of byproduct material or radiation from byproduct material.

Subpart L-- Records

35.2024 Records of authority and responsibilities for radiation protection programs.

35.2026 Records of radiation protection program changes.

35.2040 Records of written directives.

35.2045 Records of medical events.

35.2047 Record of a dose to an embryo/fetus or a nursing child.

35.2060 Records of calibrations of instruments used to measure the activity of unsealed byproduct materials.

35.2061 Records of radiation survey instrument calibrations.

35.2063 Records of dosages of unsealed byproduct material for medical use.

35.2067 Records of possession of sealed sources and brachytherapy sources.

35.2070 Records of surveys for ambient radiation exposure rate.

35.2075 Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material.

- 35.2080 Records of administrative and technical requirements that apply to the provision of mobile medical services.
- 35.2092 Records of decay-in-storage.
- 35.2204 Records of molybdenum-99 concentrations.
- 35.2310 Records of safety instruction.
- 35.2404 Records of surveys after source implant and removal.
- 35.2406 Records of brachytherapy source accountability.
- 35.2432 Records of calibration measurements of brachytherapy sealed sources.
- 35.2433 Records of decay of strontium-90 sources for ophthalmic treatments.
- 35.2605 Records of installation, maintenance, adjustment, and repair.
- 35.2630 Records of dosimetry equipment.
- 35.2632 Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.
- 35.2642 Records of periodic spot-checks for teletherapy units.
- 35.2643 Records of periodic spot-checks for remote afterloader units.
- 35.2645 Records of periodic spot-checks for gamma stereotactic radiosurgery units.
- 35.2647 Records of additional technical requirements for mobile remote afterloader units.
- 35.2652 Records of surveys of therapeutic treatment units.
- 35.2655 Records of 5-year inspection for teletherapy and gamma stereotactic radiosurgery units.

Subpart M-- Reports

- 35.3045 Report and notification of a medical event.
- 35.3047 Report and notification of a dose to an embryo/fetus or a nursing child.
- 35.3067 Report of a leaking source.

Subpart N-- Enforcement

- 35.4001 Violations.
- 35.4002 Criminal penalties.

AUTHORITY: Secs. 81, 161, 182, 183, 68 Stat. 935, 948, 953, 954, as amended (42 U.S.C. 2111, 2201, 2232, 2233); sec. 201, 88 Stat. 1242, as amended (42 U.S.C. 5841).

Subpart A--General Information

§ 35.1 Purpose and scope.

This part contains the requirements and provisions for the medical use of byproduct material and for issuance of specific licenses authorizing the medical use of this material. These requirements and provisions provide for the radiation safety of workers, the general public, patients, and human research subjects. The requirements and provisions of this part are in addition to, and not in substitution for, others in this chapter. The requirements and provisions of parts 19, 20, 21, 30, 71, 170, and 171 of this chapter apply to applicants and licensees subject to this part unless specifically exempted.

§ 35.2 Definitions.

Address of use means the building or buildings that are identified on the license and where byproduct material may be received, used, or stored.

Agreement State means any State with which the Commission or the Atomic Energy Commission has entered into an effective agreement under subsection 274b of the Atomic Energy Act of 1954, as amended.

Area of use means a portion of an address of use that has been set aside for the purpose of receiving, preparing, using, or storing byproduct material.

Authorized medical physicist means an individual who --

- (1) Meets the requirements in §§ 35.51(a) and 35.59; or
- (2) Is identified as an authorized medical physicist or teletherapy physicist on --
 - (i) A specific license issued by the Commission or Agreement State;
 - (ii) A permit issued by a Commission master material licensee;

(iii) A permit issued by a Commission or Agreement State broad scope licensee; or

(iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue

permits designating authorized medical physicists.

Authorized nuclear pharmacist means a pharmacist who --

(1) Meets the requirements in §§ 35.55(a) and 35.59; or

(2) Is identified as an authorized nuclear pharmacist on --

(i) A specific license issued by the Commission or Agreement State;

(ii) A permit issued by a Commission master material licensee;

(iii) A permit issued by a Commission or Agreement State broad scope licensee; or

(iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue

permits designating authorized nuclear pharmacists; or

(3) Is designated as an authorized nuclear pharmacist in accordance with § 32.72(b)(4).

Authorized user means a physician, dentist, or podiatrist who --

(1) Meets the requirements in §§ 35.59 and 35.190(a), 35.290(a), 35.390(a), 35.392(a), 35.394(a), 35.490(a), 35.491, 35.590(a), or 35.690(a); or

(2) Is identified as an authorized user on --

(i) A specific license issued by the Commission or Agreement State;

(ii) A permit issued by a Commission master material licensee;

(iii) A permit issued by a Commission or Agreement State broad scope licensee; or

(iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue

permits designating authorized users.

Brachytherapy means a method of radiation therapy in which sealed sources are used to deliver a radiation dose at a distance of up to a few centimeters by surface, intracavitary, intraluminal, or interstitial application.

Brachytherapy source means a radioactive source or a manufacturer-assembled source train or a combination of these sources that is designed to deliver a therapeutic dose within a distance of a few centimeters.

Client's address means the area of use or a temporary jobsite for the purpose of providing mobile medical service in accordance with § 35.80.

Dedicated check source means a radioactive source that is used to assure the constant operation of a radiation detection or measurement device over several months or years.

Dentist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice dentistry.

High dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate in excess of 12 gray (1200 rads) per hour at the point or surface where the dose is prescribed.

Low dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate of less than or equal to 2 gray (200 rads) per hour at the point or surface where the dose is prescribed.

Management means the chief executive officer or other individual having the authority to manage, direct, or administer the licensee's activities, or those persons' delegate or delegates.

Manual brachytherapy, as used in this part, means a type of brachytherapy in which the radioactive sources (e.g., seeds, ribbons) are manually inserted either into the body cavities that are in close proximity to a treatment site or directly into the tissue volume.

Medical event means an event that meets the criteria in § 35.3045(a).

Medical institution means an organization in which several medical disciplines are practiced.

Medical use means the intentional internal or external administration of byproduct material or the radiation from byproduct material to patients or human research subjects under the supervision of an authorized user.

Medium dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate of greater than 2 gray (200 rads), but less than 12 gray (1200 rads) per hour at the point or surface where the dose is prescribed.

Mobile medical service means the transportation of byproduct material to and its medical use at the client's address.

Output means the exposure rate, dose rate, or a quantity related in a known manner to these rates from a brachytherapy source or a teletherapy, remote afterloader, or gamma stereotactic radiosurgery unit for a specified set of exposure conditions.

Patient intervention means actions by the patient or human research subject, whether intentional or unintentional, such as dislodging or removing treatment devices or prematurely terminating the administration.

Pharmacist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice pharmacy.

Physician means a medical doctor or doctor of osteopathy licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to prescribe drugs in the practice of medicine.

Podiatrist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice podiatry.

Preceptor means an individual who provides or directs the training and experience required for an individual to become an authorized user, an authorized medical physicist, an authorized nuclear pharmacist, or a Radiation Safety Officer.

Prescribed dosage means the specified activity or range of activity of unsealed byproduct material as documented --

- (1) In a written directive; or
- (2) In accordance with the directions of the authorized user for procedures performed pursuant to §§ 35.100 and 35.200.

Prescribed dose means --

- (1) For gamma stereotactic radiosurgery, the total dose as documented in the written directive;
- (2) For teletherapy, the total dose and dose per fraction as documented in the written directive;
- (3) For brachytherapy, either the total source strength and exposure time or the total dose, as documented in the written directive; or
- (4) For remote afterloaders, the total dose and dose per fraction as documented in the written directive.

Pulsed dose-rate remote afterloader, as used in this part, means a special type of remote afterloading device that uses a single source capable of delivering dose rates in the “high dose-rate” range, but --

- (1) Is approximately one-tenth of the activity of typical high dose-rate remote afterloader sources; and

(2) Is used to simulate the radiobiology of a low dose rate treatment by inserting the source for a given fraction of each hour.

Radiation Safety Officer means an individual who --

(1) Meets the requirements in §§ 35.50(a) and 35.59; or

(2) Is identified as a Radiation Safety Officer on --

(i) A specific license issued by the Commission or Agreement State;

(ii) A permit issued by a Commission master material licensee;

(iii) A permit issued by a Commission or Agreement State broad scope licensee; or

(iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue

permits designating Radiation Safety Officers.

Sealed source means any byproduct material that is encased in a capsule designed to prevent leakage or escape of the byproduct material.

Sealed Source and Device Registry means the national registry that contains all the registration certificates, generated by both NRC and the Agreement States, that summarize the radiation safety information for the sealed sources and devices and describe the licensing and use conditions approved for the product.

Stereotactic radiosurgery means the use of external radiation in conjunction with a stereotactic guidance device to very precisely deliver a dose to a tissue volume.

Structured educational program means an educational program designed to impart particular knowledge and practical education through interrelated studies and supervised training.

Teletherapy, as used in this part, means a method of radiation therapy in which collimated gamma rays are delivered at a distance from the patient or human research subject.

Temporary jobsite means a location where mobile medical services are conducted other than those location(s) of use authorized on the license.

Therapeutic dosage means a dosage of an unsealed byproduct material that is intended to deliver a radiation dose to a patient or human research subject for palliative or curative treatment.

Therapeutic dose means a radiation dose delivered from a source containing byproduct material to a patient or human research subject for palliative or curative treatment.

Treatment site means the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive.

Type of use means use of byproduct material under §§ 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, or 35.1000.

Unit dosage means a dosage prepared for medical use for administration as a single dosage to a patient or human research subject without any further manipulation of the dosage after it is initially prepared.

Written directive means an authorized user's written order for the administration of byproduct material or radiation from byproduct material to a specific patient or human research subject, as specified in § 35.40.

§ 35.5 Maintenance of records.

Each record required by this part must be legible throughout the retention period specified by each Commission regulation. The record may be the original, a reproduced copy, or a microform provided that the copy or microform is authenticated by authorized personnel and the microform is capable of producing a clear copy throughout the required retention period. The record may also be stored in electronic media with the capability for producing legible, accurate, and complete records during the required retention period. Records such as letters, drawings, and specifications must include all pertinent information such as stamps, initials, and signatures. The licensee shall maintain adequate safeguards against tampering with and loss of records.

§ 35.6 Provisions for the protection of human research subjects.

(a) A licensee may conduct research involving human research subjects only if using the byproduct materials specified on its license for the uses authorized on its license.

(b) If the research is conducted, funded, supported, or regulated by another Federal agency that has implemented the Federal Policy for the Protection of Human Subjects (Federal Policy), the licensee shall, prior to conducting research --

- (1) Obtain review and approval of the research from an "Institutional Review Board," as defined and described in the Federal Policy; and
- (2) Obtain "informed consent," as defined and described in the Federal Policy, from the human research subject.

(c) If the research will not be conducted, funded, supported, or regulated by another Federal agency that has implemented the Federal Policy, the license shall, prior to conducting research, apply for and receive a specific amendment to its NRC medical use license. The amendment request must include a written commitment that the licensee will, prior to conducting research --

- (1) Obtain review and approval of the research from an "Institutional Review Board," as defined and described in the Federal Policy; and
 - (2) Obtain "informed consent", as defined and described in the Federal Policy, from the human research subject.
- (d) Nothing in this section relieves licensees from complying with the other requirements in this part.

§ 35.7 FDA, other Federal, and State requirements.

Licensees are required to comply with applicable FDA, other Federal, and State requirements governing radioactive drugs or devices.

§ 35.8 Information collection requirements: OMB approval.

(a) The Commission has submitted the information collection requirements contained in this part to the Office of Management and Budget (OMB) for approval as required by the Paperwork Reduction Act (44 U.S.C. 3501 et seq.). The NRC may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. OMB has approved the information collection requirements in this part under control number 3150-0010.

(b) The approved information collection requirements contained in this part appear in §§ 35.6, 35.12, 35.13, 35.14, 35.19, 35.24, 35.26, 35.27, 35.40, 35.41, 35.50, 35.51, 35.55, 35.57, 35.60, 35.61, 35.63, 35.67, 35.69, 35.70, 35.75, 35.80, 35.92, 35.190, 35.204, 35.290, 35.310, 35.315, 35.390, 35.392, 35.394, 35.404, 35.406, 35.410, 35.415, 35.432, 35.490, 35.491, 35.590, 35.604, 35.605, 35.610, 35.615, 35.630, 35.632, 35.633, 35.635, 35.642, 35.643, 35.645, 35.647, 35.652, 35.655, 35.690, 35.2024, 35.2026, 35.2040, 35.2045, 35.2047, 35.2060, 35.2061, 35.2063, 35.2067, 35.2070, 35.2075, 35.2080, 35.2092, 35.2204, 35.2310, 35.2404, 35.2406, 35.2432, 35.2433, 35.2605, 35.2630, 35.2632, 35.2642, 35.2643, 35.2645, 35.2647, 35.2652, 35.2655, 35.3045, 35.3047, and 35.3067.

(c) This part contains information collection requirements in addition to those approved under the control number specified in paragraph (a) of this section. These information collection requirements and the control numbers under which they are approved are as follows:

(1) In § 35.12, NRC Form 313, including NRC Forms 313A and 313B, which licensees may use to provide supplemental information, is approved under control number 3150-0120.

(2) [Reserved]

§ 35.10 Implementation.

(a) A licensee shall implement the provisions in this part on or before [insert date 6 months from publication of the Final Rule].

(b) If a license condition exempted a licensee from a provision of Part 35 on [insert date 6 months from publication of the Final Rule], then the license condition continues to exempt the licensee from the requirements in the corresponding provision until there is a license amendment or license renewal that modifies or removes the license condition.

(c) When a requirement in this part differs from the requirement in an existing license condition, the requirement in this part shall govern.

(d) Licensees shall continue to comply with any license condition that requires it to implement procedures required by §§ 35.610, 35.642, 35.643, and 35.645 until there is a license amendment or renewal that modifies the license condition.

§ 35.11 License required.

(a) A person may manufacture, produce, acquire, receive, possess, use, or transfer byproduct material for medical use only in accordance with a specific license issued by the Commission or an Agreement State, or as allowed in paragraphs (b)(1) or (b)(2) of this section.

(b) A specific license is not needed for an individual who--

(1) Receives, possesses, uses, or transfers byproduct material in accordance with the regulations in this chapter under the supervision of an authorized user as provided in § 35.27, unless prohibited by license condition; or

(2) Prepares unsealed byproduct material for medical use in accordance with the regulations in this chapter under the supervision of an authorized nuclear pharmacist or authorized user as provided in § 35.27, unless prohibited by license condition.

§ 35.12 Application for license, amendment, or renewal.

(a) An application must be signed by the applicant's or licensee's management.

(b) An application for a license for medical use of byproduct material as described in §§ 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, and 35.1000 must be made by --

(1) Filing an original and one copy of NRC Form 313, "Application for Material License," that includes the facility diagram, equipment, and training and experience qualifications of the Radiation Safety Officer, authorized user(s), authorized medical physicist(s), and authorized nuclear pharmacist(s); and

(2) Submitting procedures required by §§ 35.610, 35.642, 35.643, and 35.645, as applicable.

(c) A request for a license amendment or renewal must be made by --

(1) Submitting an original and one copy in letter format; and

(2) Submitting procedures required by §§ 35.610, 35.642, 35.643, and 35.645, as applicable.

(d) In addition to the requirements in paragraphs (b) and (c) of this section, an application for a license or amendment for medical use of byproduct material as described in § 35.1000 must also include information regarding any radiation safety aspects of the medical use of the material that is not addressed in Subparts A through C of this part.

(1) The applicant shall also provide specific information on --

(i) Radiation safety precautions and instructions;

- (ii) Methodology for measurement of dosages or doses to be administered to patients or human research subjects; and
 - (iii) Calibration, maintenance, and repair of instruments and equipment necessary for radiation safety.
- (2) The applicant or licensee shall also provide any other information requested by the Commission in its review of the application.
- (e) An applicant that satisfies the requirements specified in § 33.13 of this chapter may apply for a Type A specific license of broad scope.

§ 35.13 License amendments.

A licensee shall apply for and must receive a license amendment --

(a) Before it receives or uses byproduct material for a type of use that is permitted under this part, but that is not authorized on the licensee's current license issued under this part;

(b) Before it permits anyone to work as an authorized user, authorized nuclear pharmacist, or authorized medical physicist under the license, except --

(1) For an authorized user, an individual who meets the requirements in §§ 35.59 and 35.190(a), 35.290(a), 35.390(a), 35.392(a), 35.394(a), 35.490(a), 35.491, 35.590(a), or 35.690(a);

(2) For an authorized nuclear pharmacist, an individual who meets the requirements in §§ 35.55(a) and 35.59;

(3) For an authorized medical physicist, an individual who meets the requirements in §§ 35.51(a) and 35.59;

(4) Identified as an authorized user, an authorized nuclear pharmacist, or authorized medical physicist on--

(i) A Commission or Agreement State license or other equivalent permit or license recognized by NRC that authorizes the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively; or

(ii) A permit issued by a Commission or Agreement State specific license of broad scope that is authorized to permit the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively.

(c) Before it changes Radiation Safety Officers, except as provided in § 35.24(c);

(d) Before it receives byproduct material in excess of the amount or in a different form, or receives a different radionuclide than is authorized on the license;

(e) Before it adds to or changes the areas of use identified in the application or on the license, except for areas of use where byproduct material is used only in accordance with

§§ 35.100 and 35.200; and

(f) Before it changes the address(es) of use identified in the application or on the license.

§ 35.14 Notifications.

(a) A licensee shall provide the Commission a copy of the board certification, the Commission or Agreement State license, the permit issued by an NRC master materials licensee, or the permit issued by a licensee of broad scope for each individual no later than 30 days after the date that the licensee

permits the individual to work as an authorized user, an authorized nuclear pharmacist, or an authorized medical physicist, under § 35.13 (b)(1) through (b)(4).

(b) A licensee shall notify the Commission by letter no later than 30 days after:

(1) An authorized user, an authorized nuclear pharmacist, a Radiation Safety Officer, or an authorized medical physicist permanently discontinues performance of duties under the license or has a name change;

(2) The licensee's mailing address changes;

(3) The licensee's name changes, but the name change does not constitute a transfer of control of the license as described in § 30.34(b) of this chapter; or

(4) The licensee has deleted or otherwise changed the areas where byproduct material is used in accordance with §§ 35.100 and 35.200.

(c) The licensee shall mail the documents required in this section to the appropriate address identified in § 30.6 of this chapter.

§ 35.15 Exemptions regarding Type A specific licenses of broad scope.

A licensee possessing a Type A specific license of broad scope for medical use, issued under Part 33, is exempt from --

(a) The provisions of § 35.12(d) regarding the need to file an amendment to the license for medical uses of byproduct material, as described in § 35.1000;

(b) The provisions of § 35.13(b);

- (c) The provisions of § 35.13(e) regarding additions to or changes in the areas of use only at the addresses specified in the license;
- (d) The provisions of § 35.14(a);
- (e) The provisions of § 35.14(b)(1) for an authorized user, an authorized nuclear pharmacist, or an authorized medical physicist;
- (f) The provisions of § 35.49(a).

§ 35.18 License issuance.

(a) The Commission shall issue a license for the medical use of byproduct material if --

- (1) The applicant has filed Form NRC-313 "Application for Material License" in accordance with the instructions in § 35.12;
- (2) The applicant has paid any applicable fee as provided in Part 170 of this chapter;

(3) The Commission finds the applicant equipped and committed to observe the safety standards established by the Commission in this Chapter

for the protection of the public health and safety; and

(4) The applicant meets the requirements of Part 30 of this chapter.

(b) The Commission shall issue a license for mobile medical services if the applicant:

(1) Meets the requirements in paragraph (a) of this section; and

(2) Assures that individuals or human research subjects to whom unsealed byproduct material or radiation from implants containing byproduct material will be administered may be released following treatment in accordance with § 35.75.

§ 35.19 Specific exemptions.

The Commission may, upon application of any interested person or upon its own initiative, grant exemptions from the regulations in this part that it determines are authorized by law and will not endanger life or property or the common defense and security and are otherwise in the public interest.

Subpart B--General Administrative Requirements

§ 35.24 Authority and responsibilities for the radiation protection program.

(a) In addition to the radiation protection program requirements of § 20.1101 of this chapter, a licensee's management shall approve in writing --

(1) Requests for a license application, renewal, or amendment before submittal to the Commission;

(2) Any individual before allowing that individual to work as an authorized user, authorized nuclear pharmacist, authorized medical physicist; and

(3) Radiation protection program changes that do not require a license amendment and are permitted under § 35.26;

(b) A licensee's management shall appoint a Radiation Safety Officer, who agrees, in writing, to be responsible for implementing the radiation protection program. The licensee, through the Radiation Safety Officer, shall ensure that radiation safety activities are being performed in accordance with licensee-approved procedures and regulatory requirements.

(c) For up to 60 days each year, a licensee may permit an authorized user or an individual qualified to be a Radiation Safety Officer to function as a temporary Radiation Safety Officer and to perform the functions of a Radiation Safety Officer, as provided in paragraph (g) of this section, if the licensee takes the actions required in paragraphs (b), (e), (g), and (h) of this section.

(d) A licensee may simultaneously appoint more than one temporary Radiation Safety Officer in accordance with paragraph (c) of this section, if needed to ensure that the licensee has a temporary Radiation Safety Officer that satisfies the requirements to be a Radiation Safety Officer for each of the different uses of byproduct material permitted by the license.

(e) A licensee shall establish the authority, duties, and responsibilities of the Radiation Safety Officer in writing.

(f) Licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F, and H, or two or more types of units under Subpart H shall establish a Radiation Safety Committee to oversee all uses of byproduct material permitted by the license. The Committee must include an authorized user of each type of use permitted by the license, the Radiation Safety Officer, a representative of the nursing service, and a representative of management who is neither an authorized user nor a Radiation Safety Officer. The Committee may include other members the licensee considers appropriate.

(g) A licensee shall provide the Radiation Safety Officer sufficient authority, organizational freedom, time, resources, and management prerogative, to --

- (1) Identify radiation safety problems;
- (2) Initiate, recommend, or provide corrective actions;
- (3) Stop unsafe operations; and,

(4) Verify implementation of corrective actions.

(h) A licensee shall retain a record of actions taken under paragraphs (a), (b), and (e) of this section in accordance with § 35.2024.

§ 35.26 Radiation protection program changes.

(a) A licensee may revise its radiation protection program without Commission approval

if --

(1) The revision does not require an amendment under § 35.13;

(2) The revision is in compliance with the regulations and the license ;

(3) The revision has been reviewed and approved by the Radiation Safety Officer and licensee management; and

(4) The affected individuals are instructed on the revised program before the changes are implemented.

(b) A licensee shall retain a record of each change in accordance with § 35.2026.

§ 35.27 Supervision.

(a) A licensee that permits the receipt, possession, use, or transfer of byproduct material by an individual under the supervision of an authorized user or as allowed by § 35.11(b)(1) shall --

(1) In addition to the requirements in § 19.12, instruct the supervised individual in the licensee's written radiation protection procedures, written directive procedures, regulations of this chapter, and license conditions with respect to the use of byproduct material; and

(2) Require the supervised individual to follow the instructions of the supervising authorized user for medical uses of byproduct material, written radiation protection procedures established by the licensee, regulations of this chapter, and license conditions with respect to the medical use of byproduct material.

(b) A licensee that permits the preparation of byproduct material for medical use by an individual under the supervision of an authorized nuclear pharmacist or physician who is an authorized user, as allowed by § 35.11(b)(2), shall --

(1) In addition to the requirements in § 19.12, instruct the supervised individual in the preparation of byproduct material for medical use, as appropriate to that individual's involvement with byproduct material; and

(2) Require the supervised individual to follow the instructions of the supervising authorized user or authorized nuclear pharmacist regarding the preparation of byproduct material for medical use, the written radiation protection procedures established by the licensee and the regulations of this chapter, and license conditions.

(c) A licensee that permits supervised activities under paragraphs (a) and (b) of this section is responsible for the acts and omissions of the supervised individual.

§ 35.40 Written directives.

(a) A written directive must be dated and signed by an authorized user before the administration of I-131 sodium iodide greater than 1.11 Megabequerels (MBq) (30 microcuries (μCi)), any therapeutic dosage of unsealed byproduct material or any therapeutic dose of radiation from byproduct material.

(1) If, because of the emergent nature of the patient's condition, a delay in order to provide a written directive would jeopardize the patient's health, an oral directive is acceptable. The information contained in the oral directive must be documented as soon as possible in writing in the patient's record. A written directive must be prepared within 48 hours of the oral directive.

(b) The written directive must contain the patient or human research subject's name and the following information--

(1) For any administration of quantities greater than 1.11 MBq (30 μCi) of sodium iodide I-131: the dosage;

(2) For an administration of a therapeutic dosage of unsealed byproduct material other than sodium iodide I-131: the radioactive drug, dosage, and route of administration;

(3) For gamma stereotactic radiosurgery: the total dose, treatment site, and number of target coordinate settings per treatment for each anatomically distinct treatment site;

(4) For teletherapy: the total dose, dose per fraction, number of fractions, and treatment site;

(5) For high dose-rate remote afterloading brachytherapy: the radionuclide, treatment site, dose per fraction, number of fractions, and total dose;

or

(6) For all other brachytherapy, including low, medium, and pulsed dose rate remote afterloaders:

(i) Before implantation: treatment site, the radionuclide, and dose; and

(ii) After implantation but before completion of the procedure: the radionuclide, treatment site, number of sources, and total source strength and exposure time (or, equivalently, the total dose).

(c) A written revision to an existing written directive may be made if the revision is dated and signed by an authorized user before the administration of the dosage of unsealed byproduct material, the brachytherapy dose, the gamma stereotactic radiosurgery dose, the teletherapy dose, or the next fractional dose.

(1) If, because of the patient's condition, a delay in order to provide a written revision to an existing written directive would jeopardize the patient's health, an oral revision to an existing written directive is acceptable. The oral revision must be documented as soon as possible in the patient's record. A revised written directive must be signed by the authorized user within 48 hours of the oral revision.

(d) The licensee shall retain the written directive in accordance with § 35.2040.

§ 35.41 Procedures for administrations requiring a written directive.

(a) For any administration requiring a written directive, the licensee shall develop, implement, and maintain written procedures to provide high confidence that:

- (1) The patient's or human research subject's identity is verified before each administration; and
- (2) Each administration is in accordance with the written directive.

(b) At a minimum, the procedures required by paragraph (a) of this section must

address the following items that are applicable to the licensee's use of byproduct material--

- (1) Verifying the identity of the patient or human research subject;
- (2) Verifying that the administration is in accordance with the treatment plan, if applicable, and the written directive;
- (3) Checking both manual and computer-generated dose calculations; and
- (4) Verifying that any computer-generated dose calculations are correctly transferred into the consoles of therapeutic medical units authorized by

§ 35.600.

§ 35.49 Suppliers for sealed sources or devices for medical use.

For medical use, a licensee may only use --

- (a) Sealed sources or devices initially manufactured, labeled, packaged, and distributed in accordance with a license issued pursuant to 10 CFR Part 30 and 10 CFR 32.74 of this chapter or the equivalent requirements of an Agreement State; or
- (b) Teletherapy sources manufactured and distributed in accordance with a license issued pursuant to 10 CFR Part 30 or the equivalent requirements of an Agreement State.

§ 35.50 Training for Radiation Safety Officer.

Except as provided in § 35.57, the licensee shall require an individual fulfilling the responsibilities of the Radiation Safety Officer as provided in § 35.24 to be an individual who --

- (a) Is certified by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or
- (b)(1) Has completed a structured educational program consisting of both:
 - (i) 200 hours of didactic training in the following areas--
 - (A) Radiation physics and instrumentation;
 - (B) Radiation protection;
 - (C) Mathematics pertaining to the use and measurement of radioactivity;
 - (D) Radiation biology; and
 - (E) Radiation dosimetry; and
 - (ii) One year of full-time radiation safety experience under the supervision of the individual identified as the Radiation Safety Officer on a Commission or Agreement State license that authorizes similar type(s) of use(s) of byproduct material involving the following--
 - (A) Shipping, receiving, and performing related radiation surveys;
 - (B) Using and performing checks for proper operation of instruments used to determine the activity of dosages, survey meters, and instruments used to measure radionuclides;
 - (C) Securing and controlling byproduct material;

- (D) Using administrative controls to avoid mistakes in the administration of byproduct material;
- (E) Using procedures to prevent or minimize radioactive contamination and using proper decontamination procedures; and
- (F) Disposing of byproduct material; and

(2) Has obtained written certification, signed by a preceptor Radiation Safety Officer, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of radiation safety knowledge sufficient to function independently as a Radiation Safety Officer for a medical use licensee; or

(c) Is an authorized user, authorized medical physicist, or authorized nuclear pharmacist identified on the licensee's license and has experience with the radiation safety aspects of similar types of use of byproduct material for which the individual has Radiation Safety Officer responsibilities.

§ 35.51 Training for an authorized medical physicist.

The licensee shall require the authorized medical physicist to be an individual who --

(a) Is certified by a specialty board whose certification process includes all of the training and experience requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Holds a master's or doctor's degree in physics, biophysics, radiological physics, medical physics, or health physics and has completed one year of full-time training in therapeutic radiological physics and an additional year of full-time work experience under the supervision of a medical physicist at a medical institution that includes the tasks listed in §§ 35.67, 35.433, 35.632, 35.633, 35.635, 35.642, 35.643, 35.645, and 35.652, as applicable; and

(2) Has obtained written certification, signed by a preceptor authorized medical physicist, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized medical physicist.

§ 35.55 Training for an authorized nuclear pharmacist.

The licensee shall require the authorized nuclear pharmacist to be a pharmacist who --

(a) Is certified as a nuclear pharmacist by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed 700 hours in a structured educational program consisting of both:

(i) Didactic training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(ii) Supervised practical experience in a nuclear pharmacy involving --

(A) Shipping, receiving, and performing related radiation surveys;

(B) Using and performing checks for proper operation of instruments used to determine the activity of dosages, survey meters, and, if appropriate, instruments used to measure alpha- or beta-emitting radionuclides;

(C) Calculating, assaying, and safely preparing dosages for patients or human research subjects;

(D) Using administrative controls to avoid medical events in the administration of byproduct material; and

(E) Using procedures to prevent or minimize radioactive contamination and using proper decontamination procedures; and

(2) Has obtained written certification, signed by a preceptor authorized nuclear pharmacist, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized nuclear pharmacist.

§ 35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

(a) An individual identified as a Radiation Safety Officer, a teletherapy or medical physicist, or a nuclear pharmacist on a Commission or Agreement State license before [insert date 6 months from publication of the Final Rule] need not comply with the training requirements of §§ 35.50, 35.51, or 35.55, respectively.

(b) Physicians, dentists, or podiatrists identified as authorized users for the medical use of byproduct material on a Commission or Agreement State license issued before [insert date 6 months from publication of the Final Rule] who perform only those medical uses for which they were authorized on that date need not comply with the training requirements of Subparts D-H.

§ 35.59 Recentness of training.

The training and experience specified in Subparts B, D, E, F, G, and H must have been obtained within the 7 years preceding the date of application or the individual must have had related continuing education and experience since the required training and experience was completed.

Subpart C--General Technical Requirements

§ 35.60 Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material.

(a) For direct measurements performed in accordance with § 35.63, a licensee shall possess and use instrumentation to measure the activity of unsealed byproduct material before it is administered to each patient or human research subject.

(b) A licensee shall calibrate the instrumentation required in paragraph (a) of this section in accordance with nationally recognized standards or the manufacturer's instructions.

(c) A licensee shall retain a record of each instrument calibration required by this section in accordance with § 35.2060.

§ 35.61 Calibration of survey instruments.

(a) A licensee shall calibrate the survey instruments used to show compliance with this part and 10 CFR Part 20 before first use, annually, and following a repair that affects the calibration. A licensee shall --

- (1) Calibrate all scales with readings up to 10 mSv (1000 mrem) per hour with a radiation source;
- (2) Calibrate two separated readings on each scale or decade that will be used to show compliance; and
- (3) Conspicuously note on the instrument the date of calibration.

(b) A licensee may not use survey instruments if the difference between the indicated exposure rate and the calculated exposure rate is more than 20 percent.

(c) A licensee shall retain a record of each survey instrument calibration in accordance with § 35.2061.

§ 35.63 Determination of dosages of unsealed byproduct material for medical use.

(a) A licensee shall determine and record the activity of each dosage before medical use.

(b) For a unit dosage, this determination must be made by--

(1) A decay correction, based on the activity or activity concentration determined by --

(i) A manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(ii) An NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an

Investigational New Drug (IND) protocol accepted by FDA.

(c) For other than unit dosages, this determination must be made by--

(1) Direct measurement of radioactivity;

(2) Combination of direct measurement of radioactivity and mathematical calculations; or

(3) Combination of volumetric measurements and mathematical calculations, based on the measurement made by a manufacturer or preparer

licensed under § 32.72 of this chapter or equivalent Agreement State requirements.

(d) Unless otherwise directed by the authorized user, a licensee shall not use a dosage if the dosage does not fall within the prescribed dosage range or if the dosage differs from the prescribed dosage by more than 20 percent.

(e) A licensee shall retain a record of the dosage determination required by this section in accordance with § 35.2063.

§ 35.65 Authorization for calibration, transmission, and reference sources.

Any person authorized by § 35.11 for medical use of byproduct material may receive, possess, and use any of the following byproduct material for check, calibration, transmission, and reference use.

(a) Sealed sources manufactured and distributed by a person licensed under §§ 32.72 and 32.74 of this chapter or equivalent Agreement State regulations and that do not exceed 1.11GBq (30 mCi) each.

(b) Any byproduct material with a half-life not longer than 120 days in individual amounts not to exceed 0.555 GBq (15 mCi).

(c) Any byproduct material with a half-life longer than 120 days in individual amounts not to exceed the smaller of 7.4 MBq (200 µCi) or 1000 times the quantities in Appendix B of Part 30 of this chapter.

(d) Technetium-99m in amounts as needed.

§ 35.67 Requirements for possession of sealed sources and brachytherapy sources.

(a) A licensee in possession of any sealed source or brachytherapy source shall follow the radiation safety and handling instructions supplied by the manufacturer.

(b) A licensee in possession of a sealed source shall --

(1) Test the source for leakage before its first use unless the licensee has a certificate from the supplier indicating that the source was tested within 6 months before transfer to the licensee; and

(2) Test the source for leakage at intervals not to exceed 6 months or at other intervals approved by the Commission or an Agreement State in the Sealed Source and Device Registry.

(c) To satisfy the leak test requirements of this section, the licensee shall measure the sample so that the leak test can detect the presence of 185 Bq (0.005 μ Ci) of radioactive material in the sample.

(d) A licensee shall retain leak test records in accordance with § 35.2067.

(e) If the leak test reveals the presence of 185 Bq (0.005 μ Ci) or more of removable contamination, the licensee shall --

(1) Immediately withdraw the sealed source from use and store, dispose, or cause it to be repaired in accordance with the requirements in parts 20 and 30 of this chapter; and

(2) File a report within 5 days of the leak test in accordance with § 35.3067.

(f) A licensee need not perform a leak test on the following sources:

(1) Sources containing only byproduct material with a half-life of less than 30 days;

(2) Sources containing only byproduct material as a gas;

- (3) Sources containing 3.7 MBq (100 μ Ci) or less of beta or gamma-emitting material or 0.37 MBq (10 μ Ci) or less of alpha-emitting material;
- (4) Seeds of iridium-192 encased in nylon ribbon; and
- (5) Sources stored and not being used. However, the licensee shall test each such source for leakage before any use or transfer unless it has been leak tested within 6 months before the date of use or transfer.

(g) A licensee in possession of sealed sources or brachytherapy sources, except for gamma stereotactic radiosurgery sources, shall conduct a semi-annual physical inventory of all such sources in its possession. The licensee shall retain each inventory record in accordance with § 35.2067.

§ 35.69 Labeling of vials and syringes.

Each syringe and vial that contains unsealed byproduct material must be labeled to identify the radioactive drug. Each syringe shield and vial shield must also be labeled unless the label on the syringe or vial is visible when shielded.

§ 35.70 Surveys of ambient radiation exposure rate.

(a) In addition to the surveys required by Part 20 of this chapter, a licensee shall survey with a radiation detection survey instrument at the end of each day of use. A licensee shall survey all areas where unsealed byproduct material requiring a written directive was prepared for use or administered.

(b) A licensee does not need to perform the surveys required by paragraph (a) of this section in an area(s) where patients or human research subjects are confined when they cannot be released under § 35.75.

(c) A licensee shall retain a record of each survey in accordance with § 35.2070.

§ 35.75 Release of individuals containing unsealed byproduct material or implants containing byproduct material.

(a) A licensee may authorize the release from its control of any individual who has been administered unsealed byproduct material or implants containing byproduct material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (0.5 rem).⁵

(b) A licensee shall provide the released individual, or the individual's parent or guardian, with instructions, including written instructions, on actions recommended to maintain doses to other individuals as low as is reasonably achievable if the total effective dose equivalent to any other individual is likely to exceed 1 mSv (0.1 rem). If the total effective dose equivalent to a nursing infant or child could exceed 1 mSv (0.1 rem) assuming there were no interruption of breast-feeding, the instructions must also include --

- (1) Guidance on the interruption or discontinuation of breast-feeding; and
- (2) Information on the potential consequences, if any, of failure to follow the guidance.

(c) A licensee shall maintain a record of the basis for authorizing the release of an individual in accordance with § 35.2075(a).

(d) The licensee shall maintain a record of instructions provided to a breast-feeding female in accordance with § 35.2075(b).

⁵ NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Licenses," describes methods for calculating doses to other individuals and contains tables of activities not likely to cause doses exceeding 0.5 rem (5 mSv).

§ 35.80 Provision of mobile medical service.

(a) A licensee providing mobile medical service shall --

(1) Obtain a letter signed by the management of each client for which services are rendered that permits the use of byproduct material at the client's address and clearly delineates the authority and responsibility of the licensee and the client;

(2) Check instruments used to measure the activity of unsealed byproduct materials for proper function before medical use at each client's address or on each day of use, whichever is more frequent. At a minimum, the check for proper function required by this paragraph must include a constancy check;

(3) Check survey instruments for proper operation with a dedicated check source before use at each client's address; and

(4) Before leaving a client's address, survey all areas of use to ensure compliance with the requirements in Part 20 of this chapter.

(b) A mobile medical service may not have byproduct material delivered from the manufacturer or the distributor to the client unless the client has a license allowing possession of the byproduct material. Byproduct material delivered to the client must be received and handled in conformance with the client's license.

(c) A licensee providing mobile medical services shall retain the letter required in paragraph (a)(1) and the record of each survey required in paragraph (a)(4) of this section in accordance with § 35.2080.

§ 35.92 Decay-in-storage.

(a) A licensee may hold byproduct material with a physical half-life of less than 120 days for decay-in-storage before disposal without regard to its radioactivity if it --

(1) Monitors byproduct material at the surface before disposal and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding; and

(2) Removes or obliterates all radiation labels except for material that will be handled as biomedical waste after it has been released from the licensee.

(b) A licensee shall retain a record of each disposal permitted under paragraph (a) of this section in accordance with § 35.2092.

Subpart D--Unsealed Byproduct Material - Written Directive Not Required

§ 35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

Except for quantities that require a written directive pursuant to § 35.40(b)(2), a licensee may use any unsealed byproduct material prepared for medical use for uptake, dilution, or excretion studies that is --

(a) Obtained from a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in §§ 35.290 or 35.390, or an individual under the supervision of either as specified in § 35.27; or

(c) Obtained from an NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an Investigational New Drug (IND) protocol accepted by FDA; or

(d) Prepared by the licensee for use in research in accordance with a Radioactive Drug Research Committee-approved application or an Investigational New Drug (IND) protocol accepted by FDA.

§ 35.190 Training for uptake, dilution, and excretion studies.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.100 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under §§ 35.290 or 35.390 or equivalent Agreement State requirements; or

(c)(1) Has completed 60 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material for uptake, dilution, and excretion studies; the training and experience must include --

(i) Classroom and laboratory training in the following areas --

- (A) Radiation physics and instrumentation;
 - (B) Radiation protection;
 - (C) Mathematics pertaining to the use and measurement of radioactivity;
 - (D) Chemistry of byproduct material for medical use; and
 - (E) Radiation biology; and
- (ii) Work experience, under the supervision of an authorized user who meets the requirements in § 35.190, § 35.290, or § 35.390 or equivalent

Agreement State requirements, involving --

- (A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (B) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation of survey meters;
- (C) Calculating, measuring, and safely preparing patient or human research subject dosages;
- (D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;
- (E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and
- (F) Administering dosages of radioactive drugs to patients or human research subjects; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.190, 35.290, or 35.390 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (c)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under § 35.100.

§ 35.200 Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

Except for quantities that require a written directive under § 35.40(b)(2), a licensee may use any unsealed byproduct material prepared for medical use for imaging and localization studies that is --

(a) Obtained from a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in §§ 35.290 or 35.390, or an individual under the supervision of either as specified in § 35.27; or

(c) Obtained from an NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an Investigational New Drug (IND) protocol accepted by FDA; or

§ 35.204 Permissible molybdenum-99 concentration.

(a) A licensee may not administer to humans a radiopharmaceutical that contains more than 0.15 kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (0.15 microcurie of molybdenum-99 per millicurie of technetium-99m).

(b) A licensee that uses molybdenum-99/technetium-99m generators for preparing a technetium-99m radiopharmaceutical shall measure the molybdenum-99 concentration of the first eluate after receipt of a generator to demonstrate compliance with paragraph (a) of this section.

(c) If a licensee is required to measure the molybdenum-99 concentration, the licensee shall retain a record of each measurement in accordance with § 35.2204.

§ 35.290 Training for imaging and localization studies.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.200 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under § 35.390 or equivalent Agreement State requirements; or

(c)(1) Has completed 700 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material for imaging and localization studies; the training and experience must include, at a minimum, --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use;

(E) Radiation biology; and

(ii) Work experience, under the supervision of an authorized user, who meets the requirements in §§ 35.290 or 35.390 or equivalent Agreement

State requirements, involving --

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;

(E) Using procedures to safely contain spilled radioactive material and using proper decontamination procedures;

(F) Administering dosages of radioactive drugs to patients or human research subjects; and

(G) Eluting generator systems appropriate for preparation of radioactive drugs for imaging and localization studies, measuring and testing the

eluate for radionuclidic purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.290 or 35.390 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (c)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under §§ 35.100 and 35.200.

Subpart E--Unsealed Byproduct Material - Written Directive Required

§ 35.300 Use of unsealed byproduct material for which a written directive is required.

A licensee may use any unsealed byproduct material prepared for medical use and for which a written directive is required that is --

(a) Obtained from a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in §§ 35.290

or 35.390, or an individual under the supervision of either as specified in § 35.27; or

(c) Obtained from an NRC or Agreement State licensee for use in research in accordance with an Investigational New Drug (IND) application accepted by FDA; or

(d) Prepared by the licensee for use in accordance with an Investigational New Drug (IND) protocol accepted by FDA.

§ 35.310 Safety instruction.

In addition to the requirements of § 19.12 of this chapter,

(a) A licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that have received therapy with unsealed byproduct material and cannot be released in accordance with § 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include --

(1) Patient or human research subject control;

(2) Visitor control, including --

(i) Routine visitation to hospitalized individuals in accordance with § 20.1301(a)(1) of this chapter; and

(ii) Visitation authorized in accordance with § 20.1301(c) of this chapter;

(3) Contamination control;

(4) Waste control; and

(5) Notification of the Radiation Safety Officer, or his or her designee, and the authorized user if the patient or the human research subject has a medical emergency or dies.

(b) A licensee shall retain a record of individuals receiving instruction in accordance with § 35.2310.

§ 35.315 Safety precautions.

(a) For each patient or human research subject that cannot be released in accordance with § 35.75, a licensee shall --

(1) Quarter the patient or the human research subject either in --

(i) A private room with a private sanitary facility; or

(ii) A room, with a private sanitary facility, with another individual who also has received therapy with unsealed byproduct material and who also cannot be released pursuant to

§ 35.75;

(2) Visibly post the patient's or the human research subject's room with a "Radioactive Materials" sign.

(3) Note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or the human research subject's room; and

(4) Either monitor material and items removed from the patient's or the human research subject's room to determine that their radioactivity cannot be distinguished from the natural background radiation level with a radiation detection survey instrument set on its most sensitive scale and with no interposed shielding, or handle such material and items as radioactive waste.

(b) A licensee shall notify the Radiation Safety Officer, or his or her designee, and the authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

§ 35.390 Training for use of unsealed byproduct material for which a written directive is required.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.300 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed 700 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material requiring a written directive; the training and experience must include --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(ii) Work experience, under the supervision of an authorized user who meets the requirements in § 35.390(a) or § 35.390(b)(1)(G)(1), (2), (3), or

(4), as appropriate, or equivalent Agreement State requirements, involving --

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating instruments used to determine the activity of dosages, and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;

(E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures;

(F) Eluting generator systems, measuring and testing the eluate for radiochemical purity, and processing the eluate with reagent kits to prepare

labeled radioactive drugs; and

(G) Administering dosages of radioactive drugs to patients or human research subjects involving a minimum of three cases in each of the

following categories for which the individual is requesting authorized user status--

- (1) Oral administration of less than or equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131;
- (2) Oral administration of greater than 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131⁶;
- (3) Parenteral administration of any beta emitter or a photon-emitting radionuclide with a photon energy less than 150 keV; and/or
- (4) Parenteral administration of any other radionuclide; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.390(a) or § 35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under § 35.300(a) or § 35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate.

§ 35.392 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).

Except as provided in § 35.57, the licensee shall require an authorized user for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries), to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

⁶Experience with at least 3 cases in Category (2) also satisfies the requirement in Category (1).

(b) Is an authorized user under §§ 35.390(a), 35.390(b)(1)(ii)(G)(1) or (2), or 35.394 or equivalent Agreement State requirements; or

(c)(1) Has successfully completed 80 hours of classroom and laboratory training, applicable to the medical use of sodium iodide I-131 for

procedures requiring a written directive; the training must include --

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics pertaining to the use and measurement of radioactivity;
- (iv) Chemistry of byproduct material for medical use; and
- (v) Radiation biology; and

(2) Has work experience, under the supervision of an authorized user who meets the requirements in §§ 35.390(b)(1)(ii)(G)(1) or (2) or 35.392 or

equivalent Agreement State requirements, involving --

- (i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (ii) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation for survey meters;
- (iii) Calculating, measuring, and safely preparing patient or human research subject dosages;
- (iv) Using administrative controls to prevent a medical event involving the use of byproduct material;
- (v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and
- (vi) Administering dosages to patients or human research subjects, that includes at least 3 cases involving the oral administration of less than or

equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.390(b)(1)(ii)(G)(1), 35.390(b)(1)(ii)(G)(2), 35.392, or 35.394, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (c)(1) and (c)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user for medical uses of unsealed byproduct material using sodium iodide I-131.

§ 35.394 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).

Except as provided in § 35.57, the licensee shall require an authorized user for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries), to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under §§ 35.390(a), 35.390(b)(1)(ii)(G)(2) or equivalent Agreement State requirements; or

(c)(1) Has successfully completed 80 hours of classroom and laboratory training, applicable to the medical use of sodium iodide I-131 for procedures requiring a written directive; the training must include --

(i) Radiation physics and instrumentation;

(ii) Radiation protection;

(iii) Mathematics pertaining to the use and measurement of radioactivity;

(iv) Chemistry of byproduct material for medical use; and

(v) Radiation biology; and

(2) Has work experience, under the supervision of an authorized user who meets the requirements in §§ 35.390(a), 35.390(b)(1)(ii)(G)(2) or

35.394 or equivalent Agreement State requirements, involving --

(i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(ii) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation for survey meters;

(iii) Calculating, measuring, and safely preparing patient or human research subject dosages;

(iv) Using administrative controls to prevent a medical event involving the use of byproduct material;

(v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and

(vi) Administering dosages to patients or human research subjects, that includes at least 3 cases involving the oral administration of greater than

1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.390(a), 35.390(b)(1)(ii)(G)(2) or

35.394 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (c)(1) and (c)(2) of

this section and has achieved a level of competency sufficient to function independently as an authorized user for medical uses of unsealed byproduct

material using sodium iodide I-131.

Subpart F-- Manual Brachytherapy

§ 35.400 Use of sealed sources for manual brachytherapy.

A licensee shall use only brachytherapy sealed sources for therapeutic medical uses:

(a) As approved in the Sealed Source and Device Registry; or

(b) In research in accordance with an effective Investigational Device Exemption (IDE) application accepted by the FDA provided the requirements of § 35.49(a) are met.

§ 35.404 Surveys after source implant and removal.

(a) Immediately after implanting sources in a patient or a human research subject, the licensee shall make a survey to locate and account for all sources that have not been implanted.

(b) Immediately after removing the last temporary implant source from a patient or a human research subject, the licensee shall make a survey of the patient or the human research subject with a radiation detection survey instrument to confirm that all sources have been removed.

(c) A licensee shall retain a record of the surveys required by paragraphs (a) and (b) of this section in accordance with § 35.2404.

§ 35.406 Brachytherapy sources accountability.

(a) A licensee shall maintain accountability at all times for all brachytherapy sources in storage or use.

(b) As soon as possible after removing sources from a patient or a human research subject, a licensee shall return brachytherapy sources to a secure storage area.

(c) A licensee shall maintain a record of the brachytherapy source accountability in accordance with § 35.2406.

§ 35.410 Safety instruction.

In addition to the requirements of § 19.12 of this chapter,

(a) The licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that are undergoing implant therapy and cannot be released in accordance with § 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include the --

- (1) Size and appearance of the brachytherapy sources;
- (2) Safe handling and shielding instructions;
- (3) Patient or human research subject control;
- (4) Visitor control, including both:

- (i) Routine visitation of hospitalized individuals in accordance with § 20.1301(a)(1) of this chapter; and
 - (ii) Visitation authorized in accordance with § 20.1301(c) of this chapter; and
 - (5) Notification of the Radiation Safety Officer, or his or her designee, and an authorized user if the patient or the human research subject has a medical emergency or dies.
- (b) A licensee shall retain a record of individuals receiving instruction in accordance with § 35.2310.

§ 35.415 Safety precautions.

- (a) For each patient or human research subject that is receiving brachytherapy and cannot be released in accordance with § 35.75, a licensee shall --
- (1) Not quarter the patient or the human research subject in the same room as an individual who is not receiving brachytherapy;
 - (2) Visibly post the patient's or human research subject's room with a "Radioactive Materials" sign; and
 - (3) Note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or human research subject's room.
- (b) A licensee shall have applicable emergency response equipment available near each treatment room to respond to a source --
- (1) Dislodged from the patient; and
 - (2) Lodged within the patient following removal of the source applicators.

(c) A licensee shall notify the Radiation Safety Officer, or his or her designee, and an authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

§ 35.432 Calibration measurements of brachytherapy sealed sources.

(a) Before the first medical use of a brachytherapy sealed source on or after [insert date 6 months from publication of the Final Rule], a licensee shall –

- (1) Determine the source output or activity using a dosimetry system that meets the requirements of § 35.630(a);
- (2) Determine source positioning accuracy within applicators; and
- (3) Use published protocols accepted by nationally recognized bodies to meet the requirements of paragraphs (a)(1) and (a)(2) of this section.

(b) A licensee may use measurements provided by the source manufacturer or by a calibration laboratory accredited by the American Association of Physicists in Medicine that are made in accordance with paragraph (a) of this section.

(c) A licensee shall mathematically correct the outputs or activities determined in paragraph (a) of this section for physical decay at intervals consistent with 1 percent physical decay.

(d) A licensee shall retain a record of each calibration in accordance with § 35.2432.

§ 35.433 Decay of strontium-90 sources for ophthalmic treatments.

(a) Only an authorized medical physicist shall calculate the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments. The decay must be based on the activity determined under § 35.432.

(b) A licensee shall retain a record of the activity of each strontium-90 source in accordance with § 35.2433.

§ 35.457 Therapy-related computer systems.

The licensee shall perform acceptance testing on the treatment planning system in accordance with published protocols accepted by nationally recognized bodies. At a minimum, the acceptance testing must include, as applicable, verification of:

- (a) The source-specific input parameters required by the dose calculation algorithm;
- (b) The accuracy of dose, dwell time, and treatment time calculations at representative points;
- (c) The accuracy of isodose plots and graphic displays; and
- (d) The accuracy of the software used to determine radioactive source positions from radiographic images.

§ 35.490 Training for use of manual brachytherapy sources.

Except as provided in § 35.57, the licensee shall require an authorized user of a manual brachytherapy source for the uses authorized under § 35.400 to be a physician who --

- (a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed a structured educational program in basic radionuclide handling techniques applicable to the use of manual brachytherapy

sources that includes --

(i) 200 hours of classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity; and

(D) Radiation biology; and

(ii) 500 hours of work experience, under the supervision of an authorized user who meets the requirements in § 35.490 or equivalent Agreement

State requirements at a medical institution, involving --

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Checking survey meters for proper operation;

(C) Preparing, implanting, and removing sealed sources;

(D) Maintaining running inventories of material on hand;

(E) Using administrative controls to prevent a medical event involving the use of byproduct material;

(F) Using emergency procedures to control byproduct material; and

(2) Has obtained three years of supervised clinical experience in radiation oncology, under an authorized user who meets the requirements in §

35.490 or equivalent Agreement State requirements, as part of a formal training program approved by the Residency Review Committee for Radiation

Oncology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association. This experience may be obtained concurrently with the supervised work experience required by paragraph (b)(1)(ii) of this section; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.490 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (b)(1) and (b)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user of manual brachytherapy sources for the medical uses authorized under § 35.400.

§ 35.491 Training for ophthalmic use of strontium-90.

Except as provided in § 35.57, the licensee shall require the authorized user of only strontium-90 for ophthalmic radiotherapy to be a physician who has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of strontium-90 for ophthalmic radiotherapy, and a period of supervised clinical training in ophthalmic radiotherapy as follows --

- (a) 24 hours of classroom and laboratory training that includes --
 - (1) Radiation physics and instrumentation;
 - (2) Radiation protection;
 - (3) Mathematics pertaining to the use and measurement of radioactivity; and
 - (4) Radiation biology;

(b) Supervised clinical training in ophthalmic radiotherapy under the supervision of an authorized user at a medical institution that includes the use of strontium-90 for the ophthalmic treatment of five individuals that includes --

- (1) Examination of each individual to be treated;
- (2) Calculation of the dose to be administered;
- (3) Administration of the dose; and
- (4) Follow up and review of each individual's case history.

(c) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in 35.490, 35.491, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (a) and (b) of this section and has achieved a level of competency sufficient to function independently as an authorized user of strontium-90 for ophthalmic use.

Subpart G--Sealed Sources for Diagnosis

§ 35.500 Use of sealed sources for diagnosis.

A licensee shall use only sealed sources for diagnostic medical uses as approved in the Sealed Source and Device Registry.

§ 35.590 Training for use of sealed sources for diagnosis.

Except as provided in § 35.57, the licensee shall require the authorized user of a diagnostic sealed source for use in a device authorized under § 35.500 to be a physician, dentist, or podiatrist who --

(a) Is certified by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Has had 8 hours of classroom and laboratory training in basic radionuclide handling techniques specifically applicable to the use of the device; the training must include --

- (1) Radiation physics and instrumentation;
- (2) Radiation protection;
- (3) Mathematics pertaining to the use and measurement of radioactivity;
- (4) Radiation biology; and
- (5) Training in the use of the device for the uses requested.

**Subpart H-- Photon Emitting Remote Afterloader Units, Teletherapy Units, and
Gamma Stereotactic Radiosurgery Units**

§ 35.600 Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit.

A licensee shall use sealed sources in photon emitting remote afterloader units, teletherapy units, or gamma stereotactic radiosurgery units for therapeutic medical uses:

(a) As approved in the Sealed Source and Device Registry; or

(b) In research in accordance with an effective Investigational Device Exemption (IDE) application accepted by the FDA provided the requirements of § 35.49(a) are met.

§ 35.604 Surveys of patients and human research subjects treated with a remote afterloader unit.

(a) Before releasing a patient or a human research subject from licensee control, a licensee shall survey the patient or the human research subject and the remote afterloader unit with a portable radiation detection survey instrument to confirm that the source(s) has been removed from the patient or human research subject and returned to the safe shielded position.

(b) A licensee shall retain a record of these surveys in accordance with § 35.2404.

§ 35.605 Installation, maintenance, adjustment, and repair.

(a) Only a person specifically licensed by the Commission or an Agreement State shall install, maintain, adjust, or repair a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit that involves work on the source(s) shielding, the source(s) driving unit, or other electronic or mechanical component that could expose the source(s), reduce the shielding around the source(s), or compromise the radiation safety of the unit or the source(s).

(b) Except for low dose-rate remote afterloader units, only a person specifically licensed by the Commission or an Agreement State shall install, replace, relocate, or remove a sealed source or source contained in other remote afterloader units, teletherapy units, or gamma stereotactic units.

(c) For a low dose-rate remote afterloader unit, only a person specifically licensed by the Commission or an Agreement State or an authorized medical physicist shall install, replace, relocate, or remove a sealed source(s) contained in the unit.

(d) A licensee shall retain a record of the installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units in accordance with § 35.2605.

§ 35.610 Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

(a) A licensee shall --

(1) Secure the unit, the console, the console keys, and the treatment room when not in use or unattended;

(2) Permit only individuals approved by the authorized user, Radiation Safety Officer, or authorized medical physicist to be present in the treatment room during treatment with the source(s) ;

(3) Prevent dual operation of more than one radiation producing device in a treatment room if applicable; and

(4) Develop, implement, and maintain written procedures for responding to an abnormal situation when the operator is unable to place the source(s) in the shielded position, or remove the patient or human research subject from the radiation field with controls from outside the treatment room.

These procedures must include --

(i) Instructions for responding to equipment failures and the names of the individuals responsible for implementing corrective actions;

(ii) The process for restricting access to and posting of the treatment area to minimize the risk of inadvertent exposure; and

(iii) The names and telephone numbers of the authorized users, the authorized medical physicist, and the Radiation Safety Officer to be contacted

if the unit or console operates abnormally.

(b) A copy of the procedures required by paragraph (a)(4) of this section must be physically located at the unit console.

(c) A licensee shall post instructions at the unit console to inform the operator of --

(1) The location of the procedures required by paragraph (a)(4) of this section; and

(2) The names and telephone numbers of the authorized users, the authorized medical physicist, and the Radiation Safety Officer to be contacted

if the unit or console operates abnormally.

(d) A licensee shall provide instruction, initially and at least annually, to all individuals who operate the unit, as appropriate to the individual's assigned duties, in --

(1) The procedures identified in paragraph (a)(4) of this section; and

(2) The operating procedures for the unit.

(e) A licensee shall ensure that operators, authorized medical physicists, and authorized users participate in drills of the emergency procedures, initially and at least annually.

(f) A licensee shall retain a record of individuals receiving instruction required by paragraph (d) of this section, in accordance with § 35.2310.

§ 35.615 Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

(a) A licensee shall control access to the treatment room by a door at each entrance.

(b) A licensee shall equip each entrance to the treatment room with an electrical interlock system that will --

(1) Prevent the operator from initiating the treatment cycle unless each treatment room entrance door is closed;

(2) Cause the source(s) to be shielded when an entrance door is opened; and

(3) Prevent the source(s) from being exposed following an interlock interruption until all treatment room entrance doors are closed and the source(s) on-off control is reset at the console.

(c) A licensee shall require any individual entering the treatment room to assure, through the use of appropriate radiation monitors, that radiation levels have returned to ambient levels.

(d) Except for low-dose remote afterloader units, a licensee shall construct or equip each treatment room with viewing and intercom systems to permit continuous observation of the patient or the human research subject from the treatment console during irradiation.

(e) For licensed activities where sources are placed within the patient's or human research subject's body, a licensee shall only conduct treatments which allow for expeditious removal of a decoupled or jammed source.

(f) In addition to the requirements specified in paragraphs (a) through (e) of this section, a licensee shall --

(1) For medium dose-rate and pulsed dose-rate remote afterloader units, require --

(i) An authorized medical physicist and either an authorized user or a physician, under the supervision of an authorized user, who has been trained in the operation and emergency response for the unit to be physically present during the initiation of all patient treatments involving the unit; and

(ii) An authorized medical physicist and either an authorized user or an individual, under the supervision of an authorized user, who has been trained to remove the source applicator(s) in the event of an emergency involving the unit, to be immediately available during continuation of all patient treatments involving the unit.

(2) For high dose-rate remote afterloader units, require --

(i) An authorized user and an authorized medical physicist to be physically present during the initiation of all patient treatments involving the unit;
and

(ii) An authorized medical physicist and either an authorized user or a physician, under the supervision of an authorized user, who has been trained in the operation and emergency response for the unit, to be physically present during continuation of all patient treatments involving the unit.

(3) For gamma stereotactic radiosurgery units, require an authorized user and an authorized medical physicist to be physically present throughout all patient treatments involving the unit.

(4) Notify the Radiation Safety Officer, or his/her designee, and an authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

(g) A licensee shall have applicable emergency response equipment available near each treatment room to respond to a source --

(1) Remaining in the unshielded position; and

(2) Lodged within the patient following completion of the treatment.

§ 35.630 Dosimetry equipment.

(a) Except for low dose-rate remote afterloader sources where the source output or activity is determined by the manufacturer, a licensee shall have a calibrated dosimetry system available for use. To satisfy this requirement, one of the following two conditions must be met.

(1) The system must have been calibrated using a system or source traceable to the National Institute of Standards and Technology (NIST) and published protocols accepted by nationally recognized bodies; or by a calibration laboratory accredited by the American Association of Physicists in Medicine (AAPM). The calibration must have been performed within the previous 2 years and after any servicing that may have affected system calibration; or

(2) The system must have been calibrated within the previous 4 years. Eighteen to thirty months after that calibration, the system must have been intercompared with another dosimetry system that was calibrated within the past 24 months by NIST or by a calibration laboratory accredited by the AAPM. The results of the intercomparison must indicate that the calibration factor of the licensee's system had not changed by more than 2 percent. The licensee may not use the intercomparison result to change the calibration factor. When intercomparing dosimetry systems to be used for calibrating sealed sources for therapeutic units, the licensee shall use a comparable unit with beam attenuators or collimators, as applicable, and sources of the same radionuclide as the source used at the licensee's facility.

(b) The licensee shall have a dosimetry system available for use for spot-check output measurements, if applicable. To satisfy this requirement, the system may be compared with a system that has been calibrated in accordance with paragraph (a) of this section. This comparison must have been performed within the previous year and after each servicing that may have affected system calibration. The spot-check system may be the same system used to meet the requirement in paragraph (a) of this section.

(c) The licensee shall retain a record of each calibration, intercomparison, and comparison in accordance with § 35.2630.

§ 35.632 Full calibration measurements on teletherapy units.

(a) A licensee authorized to use a teletherapy unit for medical use shall perform full calibration measurements on each teletherapy unit --

(1) Before the first medical use of the unit; and

(2) Before medical use under the following conditions:

- (i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;
 - (ii) Following replacement of the source or following reinstallation of the teletherapy unit in a new location;
 - (iii) Following any repair of the teletherapy unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and
- (3) At intervals not exceeding 1 year.
- (b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of --
 - (1) The output within +/-3 percent for the range of field sizes and for the distance or range of distances used for medical use;
 - (2) The coincidence of the radiation field and the field indicated by the light beam localizing device;
 - (3) The uniformity of the radiation field and its dependence on the orientation of the useful beam;
 - (4) Timer constancy and linearity over the range of use;
 - (5) On-off error; and
 - (6) The accuracy of all distance measuring and localization devices in medical use.
 - (c) A licensee shall use the dosimetry system described in § 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.
 - (d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.

(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section for physical decay for intervals not exceeding 1 month for cobalt-60, 6 months for cesium-137, or at intervals consistent with 1 percent decay for all other nuclides.

(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.

(g) A licensee shall retain a record of each calibration in accordance with § 35.2632.

§ 35.633 Full calibration measurements on remote afterloader units.

(a) A licensee authorized to use a remote afterloader unit for medical use shall perform full calibration measurements on each unit --

(1) Before the first medical use of the unit;

(2) Before medical use under the following conditions:

(i) Following replacement of the source or following reinstallation of the unit in a new location outside the facility; and

(ii) Following any repair of the unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and

(3) At intervals not exceeding 1 quarter for high dose-rate, medium dose-rate, and pulsed dose-rate remote afterloader units with sources whose half-life exceeds 75 days; and

(4) At intervals not exceeding 1 year for low dose-rate remote afterloader units.

(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include, as applicable, determination of:

(1) The output within +/- 5 percent;

(2) Source positioning accuracy to within +/- 1 millimeter;

(3) Source retraction with backup battery upon power failure;

(4) Length of the source transfer tubes;

(5) Timer accuracy and linearity over the typical range of use;

(6) Length of the applicators; and

(7) Function of the source transfer tubes, applicators, and transfer tube-applicator interfaces.

(c) A licensee shall use the dosimetry system described in § 35.630(a) to measure the output.

(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.

(e) In addition to the requirements for full calibrations for low dose-rate remote afterloader units in paragraph (b) of this section, a licensee shall perform an autoradiograph of the source(s) to verify inventory and source(s) arrangement at intervals not exceeding 1 quarter.

(f) For low dose-rate remote afterloader units, a licensee may use measurements provided by the source manufacturer that are made in accordance with paragraphs (a) through (e) of this section.

(g) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section for physical decay at intervals consistent with 1 percent physical decay.

(h) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (g) of this section must be performed by the authorized medical physicist.

(i) A licensee shall retain a record of each calibration in accordance with § 35.2632.

§ 35.635 Full calibration measurements on gamma stereotactic radiosurgery units.

(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform full calibration measurements on each unit -

(1) Before the first medical use of the unit;

(2) Before medical use under the following conditions --

(i) Whenever spot-check measurements indicate that the output differs by more than

5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;

(ii) Following replacement of the sources or following reinstallation of the gamma stereotactic radiosurgery unit in a new location; and

(iii) Following any repair of the gamma stereotactic radiosurgery unit that includes removal of the sources or major repair of the components

associated with the source assembly; and

(3) At intervals not exceeding 1 year.

(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of --

(1) The output within +/-3 percent;

(2) Relative helmet factors;

(3) Isocenter coincidence;

(4) Timer accuracy and linearity over the range of use;

- (5) On-off error;
- (6) Trunnion centricity;
- (7) Treatment table retraction mechanism, using backup battery power or hydraulic backups with the unit off;
- (8) Helmet microswitches;
- (9) Emergency timing circuits; and
- (10) Stereotactic frames and localizing devices (trunnions).

(c) A licensee shall use the dosimetry system described in § 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.

(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.

(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section at intervals not exceeding 1 month for cobalt-60 and at intervals consistent with 1 percent physical decay for all other radionuclides.

(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.

(g) A licensee shall retain a record of each calibration in accordance with § 35.2632.

§ 35.642 Periodic spot-checks for teletherapy units.

(a) A licensee authorized to use teletherapy units for medical use shall perform output spot-checks on each teletherapy unit once in each calendar month that include determination

of --

- (1) Timer constancy, and timer linearity over the range of use;
- (2) On-off error;
- (3) The coincidence of the radiation field and the field indicated by the light beam localizing device;
- (4) The accuracy of all distance measuring and localization devices used for medical use;
- (5) The output for one typical set of operating conditions measured with the dosimetry system described in § 35.630(b); and
- (6) The difference between the measurement made in paragraph (b)(5) of this section and the anticipated output, expressed as a percentage of the anticipated output (i.e., the value obtained at last full calibration corrected mathematically for physical decay).

(b) A licensee shall perform measurements required by paragraph (a) of this section in accordance with written procedures established by the authorized medical physicist. That individual need not actually perform the spot-check measurements.

(c) A licensee shall have the authorized medical physicist review the results of each spot-check within 15 days. The authorized medical physicist shall notify the licensee as soon as possible in writing of the results of each spot-check.

(d) A licensee authorized to use a teletherapy unit for medical use shall perform safety spot-checks of each teletherapy facility once in each calendar month and after each source installation to assure proper operation of --

- (1) Electrical interlocks at each teletherapy room entrance;

(2) Electrical or mechanical stops installed for the purpose of limiting use of the primary beam of radiation (restriction of source housing angulation or elevation, carriage or stand travel and operation of the beam on-off mechanism);

(3) Source exposure indicator lights on the teletherapy unit, on the control console, and in the facility;

(4) Viewing and intercom systems;

(5) Treatment room doors from inside and outside the treatment room; and

(6) Electrically assisted treatment room doors with the teletherapy unit electrical power turned off.

(e) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(f) A licensee shall retain a record of each spot-check required by paragraphs (a) and (d) of this section, in accordance with § 35.2642.

§ 35.643 Periodic spot-checks for remote afterloader units.

(a) A licensee authorized to use a remote afterloader unit for medical use shall perform spot-checks of each remote afterloader facility and on each unit:

(1) Prior to the first use of a high dose-rate, medium dose-rate, or pulsed dose-rate remote afterloader unit on a given day;

(2) Before each patient treatment with a low dose-rate remote afterloader unit; and

(3) After each source installation.

(b) A licensee shall perform the measurements required by paragraph (a) of this section in accordance with written procedures established by the authorized medical physicist. That individual need not actually perform the spot check measurements.

(c) A licensee shall have the authorized medical physicist review the results of each spot-check within 15 days. The authorized medical physicist shall notify the licensee as soon as possible in writing of the results of each spot-check.

(d) To satisfy the requirements of paragraph (a) of this section, spot-checks must, at a minimum, assure proper operation of --

- (1) Electrical interlocks at each remote afterloader unit room entrance;
- (2) Source exposure indicator lights on the remote afterloader unit, on the control console, and in the facility;
- (3) Viewing and intercom systems in each high dose-rate, medium dose-rate, and pulsed dose-rate remote afterloader facility;
- (4) Emergency response equipment;
- (5) Radiation monitors used to indicate the source position;
- (6) Timer accuracy;
- (7) Clock (date and time) in the unit's computer; and
- (8) Decayed source(s) activity in the unit's computer.

(e) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(f) A licensee shall retain a record of each check required by paragraph (d) of this section in accordance with § 35.2643.

§ 35.645 Periodic spot-checks for gamma stereotactic radiosurgery units.

(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform spot-checks of each gamma stereotactic radiosurgery facility and on each unit --

- (1) Monthly;
- (2) Prior to the first use of the unit on a given day; and
- (3) After each source installation.

(b) The licensee shall have the authorized medical physicist --

- (1) Establish written procedures for performing the spot-checks required in paragraph (a) of this section; and
- (2) Review the results of each spot-check required by paragraph (a)(1) of this section within 15 days of the check. The authorized medical

physicist need not actually perform the spot-check measurements.

(c) To satisfy the requirements of paragraph (a)(1) of this section, spot-checks must, at a minimum --

- (1) Assure proper operation of --
 - (i) Treatment table retraction mechanism, using backup battery power or hydraulic backups with the unit off;
 - (ii) Helmet microswitches;
 - (iii) Emergency timing circuits; and
 - (iv) Stereotactic frames and localizing devices (trunnions).

(2) Determine --

(i) The output for one typical set of operating conditions measured with the dosimetry system described in § 35.630(b);

(ii) The difference between the measurement made in paragraph (c)(2)(i) of this section and the anticipated output, expressed as a percentage of the anticipated output (i.e., the value obtained at last full calibration corrected mathematically for physical decay);

(iii) Source output against computer calculation;

(iv) Timer accuracy and linearity over the range of use;

(v) On-off error; and

(vi) Trunnion centricity.

(d) To satisfy the requirements of paragraphs (a)(2) and (a)(3) of this section, spot-checks must assure proper operation of --

(1) Electrical interlocks at each gamma stereotactic radiosurgery room entrance;

(2) Source exposure indicator lights on the gamma stereotactic radiosurgery unit, on the control console, and in the facility;

(3) Viewing and intercom systems;

(4) Timer termination;

(5) Radiation monitors used to indicate room exposures; and

(6) Emergency off buttons.

(e) A licensee shall arrange, as soon as possible, for repair of any system identified in paragraph (c) of this section that is not operating properly.

(f) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(g) A licensee shall retain a record of each check required by paragraphs (c) and (d) of this section in accordance with § 35.2645.

§ 35.647 Additional technical requirements for mobile remote afterloader units.

(a) A licensee providing mobile remote afterloader service shall --

(1) Check survey instruments before medical use at each address of use or on each day of use, whichever is more frequent; and

(2) Account for all sources before departure from a client's address of use.

(b) In addition to the periodic spot-checks required by § 35.643, a licensee authorized to use mobile afterloaders for medical use shall perform checks on each remote afterloader unit before use at each address of use. At a minimum, checks must be made to verify the operation of --

(1) Electrical interlocks on treatment area access points;

(2) Source exposure indicator lights on the remote afterloader unit, on the control console, and in the facility;

(3) Viewing and intercom systems;

(4) Applicators, source transfer tubes, and transfer tube-applicator interfaces;

(5) Radiation monitors used to indicate room exposures;

(6) Source positioning (accuracy); and

(7) Radiation monitors used to indicate whether the source has returned to a safe shielded position.

(c) In addition to the requirements for checks in paragraph (b), a licensee shall ensure overall proper operation of the remote afterloader unit by conducting a simulated cycle of treatment before use at each address of use.

(d) If the results of the checks required in paragraph (b) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(e) A licensee shall retain a record of each check required by paragraph (b) of this section in accordance with § 35.2647.

§ 35.652 Radiation surveys.

(a) In addition to the survey requirement in § 20.1501 of this chapter, a person licensed under this subpart shall make such surveys as defined in the Sealed Source and Device Registry to assure that the maximum radiation levels and average radiation levels from the surface of the main source safe with the source(s) in the shielded position do not exceed the levels stated in the Registry.

(b) The licensee shall make the survey required by paragraph (a) of this section at installation of a new source and following repairs to the source(s) shielding, the source(s) driving unit, or other electronic or mechanical component that could expose the source, reduce the shielding around the source(s), or compromise the radiation safety of the unit or the source(s).

(c) A licensee shall retain a record of the radiation surveys required by paragraph (a) of this section in accordance with § 35.2652.

§ 35.655 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units.

(a) A licensee shall have each teletherapy unit and gamma stereotactic radiosurgery unit fully inspected and serviced during source replacement or at intervals not to exceed 5 years, whichever comes first, to assure proper functioning of the source exposure mechanism.

(b) This inspection and servicing may only be performed by persons specifically licensed to do so by the Commission or an Agreement State.

(c) A licensee shall keep a record of the inspection and servicing in accordance with § 35.2655.

§ 35.657 Therapy-related computer systems.

The licensee shall perform acceptance testing on the treatment planning system in accordance with published protocols accepted by nationally recognized bodies. At a minimum, the acceptance testing must include, as applicable, verification of:

(a) The source-specific input parameters required by the dose calculation algorithm;

(b) The accuracy of dose, dwell time, and treatment time calculations at representative points;

(c) The accuracy of isodose plots and graphic displays;

(d) The accuracy of the software used to determine radioactive source positions from radiographic images; and

(e) The accuracy of electronic transfer of the treatment delivery parameters to the treatment delivery unit from the treatment planning system.

§ 35.690 Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Except as provided in § 35.57, the licensee shall require an authorized user of a sealed source for a use authorized under § 35.600 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed a structured educational program in basic radionuclide techniques applicable to the use of a sealed source in a therapeutic medical unit that includes --

(i) 200 hours of classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity; and

(D) Radiation biology; and

(ii) 500 hours of work experience, under the supervision of an authorized user who meets the requirements in § 35.690 or equivalent Agreement State requirements at a medical institution, involving --

(A) Reviewing full calibration measurements and periodic spot-checks;

(B) Preparing treatment plans and calculating treatment doses and times;

(C) Using administrative controls to prevent a medical event involving the use of byproduct material;

(D) Implementing emergency procedures to be followed in the event of the abnormal operation of the medical unit or console;

(E) Checking and using survey meters; and

(F) Selecting the proper dose and how it is to be administered; and

(2) Has completed three years of supervised clinical experience in radiation oncology, under an authorized user who meets the requirements in § 35.690 or equivalent Agreement State requirements, as part of a formal training program approved by the Residency Review Committee for Radiation Oncology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association. This experience may be obtained concurrently with the supervised work experience required by paragraph (b)(1)(ii) of this section; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.690 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (b)(1) and (b)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user of each type of therapeutic medical unit for which the individual is requesting authorized user status.

Subpart I -- [Reserved]

Subpart J -- [Reserved]

**Subpart K--Other Medical Uses of Byproduct Material or
Radiation from Byproduct Material**

§ 35.1000 Other medical uses of byproduct material or radiation from byproduct material.

A licensee may use byproduct material or a radiation source approved for medical use which is not specifically addressed in subparts D through H of this part if --

(a) The applicant or licensee has submitted the information required by § 35.12(b) through (d); and

(b) The applicant or licensee has received written approval from the Commission in a license or license amendment and uses the material in accordance with the regulations and specific conditions the Commission considers necessary for the medical use of the material.

Subpart L--Records

§ 35.2024 Records of authority and responsibilities for radiation protection programs.

(a) A licensee shall retain a record of actions taken by the licensee's management in accordance with § 35.24(a) for 5 years. The record must include a summary of the actions taken and a signature of licensee management.

(b) The licensee shall retain a current copy of the authority, duties, and responsibilities of the Radiation Safety Officer as required by § 35.24(e), and a signed copy of the Radiation Safety Officer's agreement to be responsible for implementing the radiation safety program, as required by § 35.24(b). The records must include the signature of the Radiation Safety Officer and licensee management.

§ 35.2026 Records of radiation protection program changes.

A licensee shall retain a record of each radiation protection program change made in accordance with § 35.26(a) for 5 years. The record must include a copy of the old and new procedures; the effective date of the change; and the signature of the licensee management that reviewed and approved the change.

§ 35.2040 Records of written directives.

A licensee shall retain a copy of each written directive as required by § 35.40 for 3 years.

§ 35.2045 Records of medical events.

(a) A licensee shall retain a record of medical events reported in accordance with

§ 35.3045 for 3 years.

(b) The record must include--

- (1) The licensee's name;
- (2) Names of all the individuals involved;
- (3) The affected or potentially affected individual's social security number or other identification number if one has been assigned;
- (4) A brief description of the medical event and why it occurred;
- (5) The effect, if any, on the individual; and
- (6) The actions, if any, taken or planned to prevent recurrence.

§ 35.2047 Record of a dose to an embryo/fetus or a nursing child.

(a) A licensee shall retain a record of a dose to an embryo/fetus or a nursing child reported in accordance with § 35.3047 for 3 years.

(b) The record must include--

- (1) The licensee's name;
- (2) Names of all the individuals involved;
- (3) The affected or potentially affected individual's social security number or other identification number if one has been assigned;
- (4) A brief description of the event and why it occurred;

- (5) The effect, if any, on the embryo/fetus or nursing child; and
- (6) The actions, if any, taken or planned to prevent recurrence.

§ 35.2060 Records of calibrations of instruments used to measure the activity of unsealed byproduct materials.

A licensee shall maintain a record of instrument calibrations required by § 35.60 for 3 years. The records must include the model and serial number of the instrument, the date of the calibration, the results of the calibration, and the name of the individual who performed the calibration.

§ 35.2061 Records of radiation survey instrument calibrations.

A licensee shall maintain a record of radiation survey instrument calibrations required by § 35.61 for 3 years. The record must include the model and serial number of the instrument, the date of the calibration, the results of the calibration, and the name of the individual who performed the calibration.

§ 35.2063 Records of dosages of unsealed byproduct material for medical use.

- (a) A licensee shall maintain a record of dosage determinations required by § 35.63 for 3 years.
- (b) The record must contain--

- (1) The radiopharmaceutical;
- (2) The patient's or human research subject's name, or identification number if one has been assigned;
- (3) The prescribed dosage, the determined dosage, or a notation that the total activity is less than 1.1 MBq (30 μ Ci);
- (4) The date and time of the dosage determination; and
- (5) The name of the individual who determined the dosage.

§ 35.2067 Records of possession of sealed sources and brachytherapy sources.

(a) A licensee shall retain records of leak tests required by § 35.67(b) for 3 years. The records must include the model number, and serial number if one has been assigned, of each source tested, the identity of each source radionuclide and its estimated activity, the results of the test, the date of the test, and the name of the individual who performed the test.

(b) A licensee shall retain records of the semi-annual physical inventory of sealed sources and brachytherapy sources required by § 35.67(g) for 3 years. The inventory records must contain the model number of each source, and serial number if one has been assigned, the identity of each source radionuclide and its nominal activity, the location of each source, and the name of the individual who performed the inventory.

§ 35.2070 Records of surveys for ambient radiation exposure rate.

A licensee shall retain a record of each survey required by § 35.70 for 3 years. The record must include the date of the survey, the results of the survey, the instrument used to make the survey, and the name of the individual who performed the survey.

§ 35.2075 Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material.

(a) A licensee shall retain a record of the basis for authorizing the release of an individual in accordance with § 35.75, if the total effective dose equivalent is calculated by --

- (1) Using the retained activity rather than the activity administered;
- (2) Using an occupancy factor less than 0.25 at 1 meter;
- (3) Using the biological or effective half-life; or
- (4) Considering the shielding by tissue.

(b) A licensee shall retain a record that the instructions required by § 35.75(b) were provided to a breast-feeding female if the radiation dose to the infant or child from continued breast-feeding could result in a total effective dose equivalent exceeding 5 mSv (0.5 rem).

(c) The records required by paragraphs (a) and (b) of this section must be retained for 3 years after the date of release of the individual.

§ 35.2080 Records of administrative and technical requirements that apply to the provision of mobile medical services.

(a) A licensee shall retain a copy of the letter(s) that permits the use of byproduct material at a client's address, as required by § 35.80(a)(1). This letter must clearly delineate the authority and responsibility of the licensee and the client and must be retained for 3 years after the last provision of service.

(b) A licensee shall retain the record of each survey required by § 35.80(a)(4) for 3 years. The record must include the date of the survey, the results of the survey, the instrument used to make the survey, and the name of the individual who performed the survey.

§ 35.2092 Records of decay-in-storage.

A licensee shall maintain records of the disposal of licensed materials, as required by § 35.92, for 3 years. The record must include the date of the disposal, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.

§ 35.2204 Records of molybdenum-99 concentrations.

A licensee shall maintain a record of the molybdenum-99 concentration tests required by § 35.204(b) for 3 years. The record must include, for each measured elution of technetium-99m, the ratio of the measures expressed as kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (microcuries of molybdenum per millicurie of technetium), the time and date of the measurement, and the name of the individual who made the measurement.

§ 35.2310 Records of safety instruction.

A licensee shall maintain a record of safety instructions required by §§ 35.310, 35.410, and 35.610 for 3 years. The record must include a list of the topics covered, the date of the instruction, the name(s) of the attendee(s), and the name(s) of the individual(s) who provided the instruction.

§ 35.2404 Records of surveys after source implant and removal.

A licensee shall maintain a record of the surveys required by §§ 35.404 and 35.604 for 3 years. Each record must include the date and results of the survey, the survey instrument used, and the name of the individual who made the survey.

§ 35.2406 Records of brachytherapy source accountability.

(a) A licensee shall maintain a record of brachytherapy source accountability required by § 35.406 for 3 years.

(b) For temporary implants, the record must include --

(1) The number and activity of sources removed from storage, the time and date they were removed from storage, the name of the individual who removed them from storage, and the location of use; and

(2) The number and activity of sources returned to storage, the time and date they were returned to storage, and the name of the individual who returned them to storage.

(c) For permanent implants, the record must include --

(1) The number and activity of sources removed from storage, the date they were removed from storage, and the name of the individual who removed them from storage;

(2) The number and activity of sources not implanted, the date they were returned to storage, and the name of the individual who returned them to storage; and

(3) The number and activity of sources permanently implanted in the patient or human research subject.

§ 35.2432 Records of calibration measurements of brachytherapy sealed sources.

(a) A licensee shall maintain a record of the calibrations of brachytherapy sources required by § 35.432 for 3 years after the last use of the source.

(b) The record must include--

- (1) The date of the calibration;
- (2) The manufacturer's name, model number, and serial number for the source and the instruments used to calibrate the source;
- (3) The source output or activity;
- (4) The source positioning accuracy within the applicators; and
- (5) The signature of the authorized medical physicist.

§ 35.2433 Records of decay of strontium-90 sources for ophthalmic treatments.

(a) A licensee shall maintain a record of the activity of a strontium-90 source required by § 35.433 for the life of the source.

(b) The record must include--

(1) The initial activity of the source and date; and

(2) For each decay calculation, the date and the source activity as determined under

§ 35.432.

§ 35.2605 Records of installation, maintenance, adjustment, and repair.

A licensee shall retain a record of the installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units as required by § 35.605 for 3 years. For each installation, maintenance, adjustment and repair, the record must include the date, description of the service, and name(s) of the individual(s) who performed the work.

§ 35.2630 Records of dosimetry equipment.

(a) A licensee shall retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with § 35.630 for the duration of the license. (b) For each calibration, intercomparison, or comparison, the record must include --

(1) The date;

(2) The manufacturer's name, model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by paragraphs (a) and (b) of

§ 35.630;

(3) The correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and

(4) The names of the individuals who performed the calibration, intercomparison, or comparison.

§ 35.2632 Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.

(a) A licensee shall maintain a record of the teletherapy unit, remote afterloader unit, and gamma stereotactic radiosurgery unit full calibrations required by §§ 35.632, 35.633, and 35.635 for 3 years.

(b) The record must include --

(1) The date of the calibration;

- (2) The manufacturer's name, model number, and serial number of the teletherapy, remote afterloader, and gamma stereotactic radiosurgery unit(s), the source, and the instruments used to calibrate the unit(s);
- (3) The results and an assessment of the full calibrations;
- (4) The results of the autoradiograph required for low dose-rate remote afterloader units; and
- (5) The signature of the authorized medical physicist who performed the full calibration.

§ 35.2642 Records of periodic spot-checks for teletherapy units.

- (a) A licensee shall retain a record of each periodic spot-check for teletherapy units required by § 35.642 for 3 years.
- (b) The record must include --
 - (1) The date of the spot-check;
 - (2) The manufacturer's name, model number, and serial number of the teletherapy unit, source and instrument used to measure the output of the teletherapy unit;
 - (3) An assessment of timer linearity and constancy;
 - (4) The calculated on-off error;
 - (5) A determination of the coincidence of the radiation field and the field indicated by the light beam localizing device;

(6) The determined accuracy of each distance measuring and localization device;

(7) The difference between the anticipated output and the measured output;

(8) Notations indicating the operability of each entrance door electrical interlock, each electrical or mechanical stop, each source exposure indicator light, and the viewing and intercom system and doors; and

(9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

§ 35.2643 Records of periodic spot-checks for remote afterloader units.

(a) A licensee shall retain a record of each spot-check for remote afterloader units required by § 35.643 for 3 years.

(b) The record must include, as applicable --

(1) The date of the spot-check;

(2) The manufacturer's name, model number, and serial number for the remote afterloader unit and source;

(3) An assessment of timer accuracy;

(4) Notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom systems, and clock and decayed source activity in the unit's computer; and

(5) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

§ 35.2645 Records of periodic spot-checks for gamma stereotactic radiosurgery units.

(a) A licensee shall retain a record of each spot-check for gamma stereotactic radiosurgery units required by § 35.645 for 3 years.

(b) The record must include --

(1) The date of the spot-check;

(2) The manufacturer's name, model number, and serial number for the gamma stereotactic radiosurgery unit and the instrument used to measure the output of the unit;

(3) An assessment of timer linearity and accuracy;

(4) The calculated on-off error;

(5) A determination of trunnion centricity;

(6) The difference between the anticipated output and the measured output;

(7) An assessment of source output against computer calculations;

(8) Notations indicating the operability of radiation monitors, helmet microswitches, emergency timing circuits, emergency off buttons, electrical interlocks, source exposure indicator lights, viewing and intercom systems, timer termination, treatment table retraction mechanism, and stereotactic frames and localizing devices (trunnions); and

(9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

§ 35.2647 Records of additional technical requirements for mobile remote afterloader units.

(a) A licensee shall retain a record of each check for mobile remote afterloader units required by § 35.647 for 3 years.

(b) The record must include --

(1) The date of the check;

(2) The manufacturer's name, model number, and serial number of the remote afterloader unit;

(3) Notations accounting for all sources before the licensee departs from a facility;

(4) Notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom system, applicators, source transfer tubes, and transfer tube applicator interfaces, and source positioning accuracy; and

(5) The signature of the individual who performed the check.

§ 35.2652 Records of surveys of therapeutic treatment units.

(a) A licensee shall maintain a record of radiation surveys of treatment units made in accordance with § 35.652 for the duration of use of the unit.

(b) The record must include --

(1) The date of the measurements;

(2) The manufacturer's name, model number and serial number of the treatment unit, source, and instrument used to measure radiation levels;

(3) Each dose rate measured around the source while the unit is in the off position and the average of all measurements; and

(4) The signature of the individual who performed the test.

§ 35.2655 Records of 5-year inspection for teletherapy and gamma stereotactic radiosurgery units.

(a) A licensee shall maintain a record of the 5-year inspections for teletherapy and gamma stereotactic radiosurgery units required by § 35.655 for the duration of use of the unit.

(b) The record must contain --

(1) The inspector's radioactive materials license number;

- (2) The date of inspection;
- (3) The manufacturer's name and model number and serial number of both the treatment unit and source;
- (4) A list of components inspected and serviced, and the type of service; and
- (5) The signature of the inspector.

Subpart M--Reports

§ 35.3045 Report and notification of a medical event.

(a) A licensee shall report any event, except for an event that results from patient intervention, in which the administration of byproduct material or radiation from byproduct material results in --

- (1) A dose that differs from the prescribed dose by more than 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin; and
 - (i) The total dose delivered differs from the prescribed dose by 20 percent or more;
 - (ii) The total dosage delivered differs from the prescribed dosage by 20 percent or more or falls outside the prescribed dosage range; or
 - (iii) The fractionated dose delivered differs from the prescribed dose, for a single fraction, by 50 percent or more.

(2) A dose that exceeds 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin from any of the following –

- (i) An administration of a wrong radioactive drug containing byproduct material;
- (ii) An administration of a radioactive drug containing byproduct material by the wrong route of administration;
- (iii) An administration of a dose or dosage to the wrong individual or human research subject;
- (iv) An administration of a dose or dosage delivered by the wrong mode of treatment; or
- (v) A leaking sealed source.

(3) A dose to the skin or an organ or tissue other than the treatment site that exceeds by 0.5 Sv (50 rem) to an organ or tissue and 50 percent or more of the dose expected from the administration defined in the written directive (excluding, for permanent implants, seeds that were implanted in the correct site but migrated outside the treatment site).

(b) A licensee shall report any event resulting from intervention of a patient or human research subject in which the administration of byproduct material or radiation from byproduct material results or will result in an unintended permanent functional damage to an organ or a physiological system, as determined by a physician.

(c) The licensee shall notify by telephone the NRC Operations Center² no later than the next calendar day after discovery of the medical event.

² The commercial telephone number of the NRC Operations Center is (301) 951-0550.

(d) The licensee shall submit a written report to the appropriate NRC Regional Office listed in § 30.6 of this chapter within 15 days after discovery of the medical event.

(1) The written report must include --

(i) The licensee's name;

(ii) The name of the prescribing physician;

(iii) A brief description of the event;

(iv) Why the event occurred;

(v) The effect, if any, on the individual(s) who received the administration;

(vi) What actions, if any, have been taken or are planned to prevent recurrence;

(vii) Whether the licensee notified the individual (or the individual's responsible relative or guardian), and if not, why not; and

(viii) If there was notification, what information was provided.

(2) The report may not contain the individual's name or any other information that could lead to identification of the individual.

(e) The licensee shall notify the referring physician and also notify the individual affected by the medical event no later than 24 hours after its discovery, unless the referring physician personally informs the licensee either that he or she will inform the individual or that, based on medical judgement, telling the individual would be harmful. The licensee is not required to notify the individual without first consulting the referring physician. If the referring physician or the affected individual cannot be reached within 24 hours, the licensee shall notify the individual as soon as possible thereafter. The licensee may not delay any appropriate medical care for the individual, including any necessary remedial

care as a result of the medical event, because of any delay in notification. To meet the requirements of this section, the notification of the individual receiving the medical event may be made instead to that individual's responsible relative or guardian.

(f) If the individual was notified under paragraph (e) of this section, the licensee shall also furnish a written report to the individual within 15 days after discovery of the medical event. A licensee may send either --

(1) A copy of the report that was submitted to the NRC; or

(2) A brief description of both the event and the consequences as they may affect the individual.

(g) Aside from the notification requirement, nothing in this section affects any rights or duties of licensees and physicians in relation to each other, to individuals affected by the medical event, or to that individual's responsible relatives or guardians.

(h) A licensee shall retain a record of a medical event in accordance with § 35.2045.

§ 35.3047 Report and notification of a dose to an embryo/fetus or a nursing child.

(a) A licensee shall report any dose to an embryo/fetus that is greater than 50 mSv (5 rem) dose equivalent that is a result of an administration of byproduct material or radiation from byproduct material to a pregnant individual unless the dose to the embryo/fetus was specifically approved, in advance, by the authorized user.

(b) A licensee shall report any dose to a nursing child that is a result of an administration of byproduct material to a breast-feeding individual that --

(1) Is greater than 50 mSv (5 rem) total effective dose equivalent; or

(2) Has resulted in unintended permanent functional damage to an organ or a physiological system of the child, as determined by a physician.

(c) The licensee shall notify by telephone the NRC Operations Center no later than the next calendar day after discovery of a dose to the embryo/fetus or nursing child that requires a report in paragraphs (a) or (b) in this section.

(d) The licensee shall submit a written report to the appropriate NRC Regional Office listed in § 30.6 of this chapter no later than 15 days after discovery of a dose to the embryo/fetus or nursing child that requires a report in paragraphs (a) or (b) in this section.

(1) The written report must include --

(i) The licensee's name;

(ii) The name of the prescribing physician;

(iii) A brief description of the event;

(iv) Why the event occurred;

(v) The effect, if any, on the embryo/fetus or the nursing child; and

(vi) What actions, if any, have been taken or are planned to prevent recurrence;

(2) The report must not contain the individual's or child's name or any other information that could lead to identification of the individual or child.

(e) The licensee shall notify the referring physician and also notify the pregnant individual or mother, both hereafter referred to as the mother, no later than 24 hours after discovery of an event that would require reporting under paragraph (a) or (b) of this section, unless the referring physician personally informs the licensee either that he or she will inform the mother or that, based on medical judgement, telling the mother would be harmful.

(f) To meet the requirements of this section, the notification may be made to the mother's or child's responsible relative or guardian instead of the mother, when appropriate.

(g) The licensee is not required to notify the mother without first consulting with the referring physician. If the referring physician or mother cannot be reached within 24 hours, the licensee shall make the appropriate notifications as soon as possible thereafter. The licensee may not delay any appropriate medical care for the embryo/fetus or for the nursing child, including any necessary remedial care as a result of the event, because of any delay in notification.

(h) If notification was made under paragraphs (e) and (f) of this section, the licensee shall also furnish a written report to the mother or responsible relative or guardian within 15 days after discovery of the event. A licensee may send either --

(1) A copy of the report that was submitted to the NRC; or

(2) A brief description of both the event and the consequences as they may affect the embryo/fetus or nursing child.

(i) A licensee shall retain a record of a dose to an embryo/fetus or a nursing child in accordance with § 35.2047.

§ 35.3067 Report of a leaking source.

A licensee shall file a report within 5 days if a leak test required by § 35.67 reveals the presence of 185 Bq (0.005 µCi) or more of removable contamination. The report must be filed with the appropriate NRC Regional Office listed in § 30.6 of this chapter, with a copy to the Director, Office of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001. The written report must include the model number and serial number if assigned, of the leaking source; the radionuclide and its estimated activity; the results of the test; the date of the test; and the action taken.

Subpart N--Enforcement

§ 35.4001 Violations.

(a) The Commission may obtain an injunction or other court order to prevent a violation of the provisions of --

(1) The Atomic Energy Act of 1954, as amended;

(2) Title II of the Energy Reorganization Act of 1974, as amended; or

(3) A regulation or order issued under those Acts.

(b) The Commission may obtain a court order for the payment of a civil penalty imposed under Section 234 of the Atomic Energy Act:

(1) For violations of --

- (i) Sections 53, 57, 62, 63, 81, 82, 101, 103, 104, 107, or 109 of the Atomic Energy Act of 1954, as amended;
 - (ii) Section 206 of the Energy Reorganization Act;
 - (iii) Any rule, regulation, or order issued under the sections specified in paragraph (b)(1)(i) of this section;
 - (iv) Any term, condition, or limitation of any license issued under the sections specified in paragraph (b)(1)(i) of this section.
- (2) For any violation for which a license may be revoked under Section 186 of the Atomic Energy Act of 1954, as amended.

§ 35.4002 Criminal penalties.

(a) Section 223 of the Atomic Energy Act of 1954, as amended, provides for criminal sanctions for willful violation of, attempted violation of, or conspiracy to violate, any regulation issued under sections 161b, 161i, or 161o of the Act. For purposes of Section 223, all the regulations in 10 CFR part 35 are issued under one or more of sections 161b, 161i, or 161o, except for the sections listed in paragraph (b) of this section.

(b) The regulations in 10 CFR part 35 that are not issued under subsections 161b, 161i, or 161o for the purposes of Section 223 are as follows: §§ 35.1, 35.2, 35.7, 35.8, 35.12, 35.15, 35.18, 35.19, 35.65, 35.100, 35.200, 35.300, 35.600, 35.4001, and 35.4002 .

Dated at Rockville, Maryland, this ____ day of _____, 1999.

For the Nuclear Regulatory Commission.

Annette L. Vietti-Cook,

Secretary of the Commission.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

PART 35--MEDICAL USE OF BYPRODUCT MATERIAL

PART 35--MEDICAL USE OF BYPRODUCT MATERIAL

PART 35--MEDICAL USE OF BYPRODUCT MATERIAL

Subpart A--General Information

Subpart A--General Information

Subpart A--General Information

Sec.

Sec.

- 35.1 Purpose and scope.
- 35.2 Definitions.
- 35.5 Maintenance of records.
- 35.6 Provisions for research involving human subjects.

- 35.7 FDA, other Federal, and State requirements.
- 35.8 Information collection requirements: OMB approval.

- 35.11 License required.
- 35.12 Application for license, amendment, or renewal.
- 35.13 License amendments.
- 35.14 Notifications.
- 35.15 Exemptions regarding Type A specific licenses of broad scope.
- 35.18 License issuance.
- 35.19 Specific exemptions.

- 35.1 Purpose and scope.
- 35.2 Definitions.
- 35.5 Maintenance of records.
- 35.6 Provisions for research involving human subjects.

- 35.7 FDA, other Federal, and State requirements.
- 35.8 Information collection requirements: OMB approval.
- 35.10 Implementation.
- 35.11 License required.
- 35.12 Application for license, amendment, or renewal.
- 35.13 License amendments.
- 35.14 Notifications.
- 35.15 Exemptions regarding Type A specific licenses of broad scope.
- 35.18 License issuance.
- 35.19 Specific exemptions.

- 35.1 Purpose and scope.
- 35.2 Definitions.
- 35.5 Maintenance of records.
- 35.6 Provisions for the protection of human research subjects.

- 35.7 FDA, other Federal, and State requirements.
- 35.8 Information collection requirements: OMB approval.
- 35.10 Implementation.
- 35.11 License required.
- 35.12 Application for license, amendment, or renewal.
- 35.13 License amendments.
- 35.14 Notifications.
- 35.15 Exemptions regarding Type A specific licenses of broad scope.
- 35.18 License issuance.
- 35.19 Specific exemptions.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart B--General
Administrative Requirements

Subpart B--General
Administrative Requirements

Subpart B--General
Administrative Requirements

35.20 ALARA
program.[20.1101]
35.21 Radiation Safety Officer.
35.22 Radiation Safety
Committee.[35.24]
35.23 Statements of authority
and responsibilities.[35.24]

35.24 Authority and
responsibilities for the
radiation protection
program.

35.24 Authority and
responsibilities for the radiation
protection program.

35.25 Supervision.
35.29 Administrative
requirements that apply to the
provision of mobile nuclear
medicine service.
35.31 Radiation safety
program changes.
35.32 Quality management
program.
35.33 Notifications, reports,
and records of
misadministrations.

35.26 Radiation protection
program changes.

35.26 Radiation protection
program changes.

35.27 Supervision.

35.27 Supervision.

35.40 Written directives.

35.40 Written directives.

35.41 Procedures for
administrations requiring a
written directive.

35.41 Procedures for
administrations requiring a
written directive.

35.49 Suppliers for sealed
sources or devices for medical
use.

35.49 Suppliers for sealed
sources or devices for medical
use.

35.49 Suppliers for sealed
sources or devices for medical
use.

35.50 Training for Radiation
Safety Officer.

35.50 Training for Radiation
Safety Officer.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
35.51 Training for an authorized medical physicist.	35.51 Training for an authorized medical physicist.	35.51 Training for an authorized medical physicist.
35.55 Training for an authorized nuclear pharmacist.	35.55 Training for an authorized nuclear pharmacist.	35.55 Training for an authorized nuclear pharmacist.
35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.	35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.	35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.
35.59 Recentness of training.	35.59 Recentness of training.	35.59 Recentness of training.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart C--General Technical Requirements

Subpart C--General Technical Requirements

Subpart C--General Technical Requirements

35.50 Possession, use, calibration, and check of dose calibrators.

35.60 Possession, use, calibration, and check of instruments to measure the activity of photon-emitting radionuclides.

35.60 Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material.

35.51 Calibration and check of survey instruments.

35.61 Calibration and check of survey instruments.

35.61 Calibration of survey instruments.

35.52 Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides.

35.62 Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides.

35.61 Calibration of survey instruments.

35.53 Measurement of dosages of unsealed byproduct material for medical use.

35.63 Determination of dosages of unsealed byproduct material for medical use.

35.63 Determination of dosages of unsealed byproduct material for medical use.

35.57 Authorization for calibration and reference sources.

35.65 Authorization for calibration and reference sources.

35.65 Authorization for calibration, transmission, and reference sources.

35.59 Requirements for possession of sealed sources and brachytherapy sources.

35.67 Requirements for possession of sealed sources and brachytherapy sources.

35.67 Requirements for possession of sealed sources and brachytherapy sources.

35.60 Syringe shields and labels.

35.69 Labeling and shielding of vials and syringes.

35.69 Labeling of vials and syringes.

35.61 Vial shields and labels.

35.70 Surveys for contamination and ambient radiation exposure rate.

35.70 Surveys for ambient radiation exposure rate.

35.70 Surveys of ambient radiation exposure rate.

35.75 Release of patients or human research subjects containing

35.75 Release of individuals containing radiopharmaceuticals or implants.

35.75 Release of individuals containing unsealed byproduct material or implants containing byproduct material.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

radiopharmaceuticals or permanent implants.

35.80 Technical requirements that apply to the provision of mobile nuclear medicine service.

35.90 Storage of volatiles and gases.

35.92 Decay-in-storage.

35.80 Provision of mobile service.

35.92 Decay-in-storage.

35.80 Provision of mobile medical service.

35.92 Decay-in-storage.

Subpart D--Uptake, Dilution, and Excretion

35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies.

35.120 Possession of survey instrument.

Subpart E--Imaging and Localization

35.200 Use of unsealed byproduct material for imaging and localization studies.

Subpart D--Unsealed Byproduct Material--Low Dose

35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

35.200 Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

Subpart D--Unsealed Byproduct Material - Written Directive Not Required

35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

35.190 Training for uptake, dilution, and excretion studies.

35.200 Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

35.204 Permissible molybdenum-99 concentration.
35.205 Control of aerosols and gases.
35.220 Possession of survey instruments.

35.204 Permissible molybdenum-99 concentration.

35.290 Training for uptake, dilution, and excretion studies.
35.292 Training for imaging and localization studies.

35.204 Permissible molybdenum-99 concentration.

35.290 Training for imaging and localization studies.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart F--
Radiopharmaceuticals for
Therapy

Subpart E--Unsealed
Byproduct Material--High
Dose

Subpart E--Unsealed Byproduct
Material -Written Directive
Required

35.300 Use of unsealed
byproduct material for
therapeutic administration.
35.310 Safety instruction.
35.315 Safety precautions.
35.320 Possession of survey
instruments.

35.300 Use of unsealed
byproduct material for which a
written directive is required.
35.310 Safety instruction.
35.315 Safety precautions.

35.300 Use of unsealed
byproduct material for which a
written directive is required.
35.310 Safety instruction.
35.315 Safety precautions.

35.390 Training for use of
unsealed byproduct material
for therapy or for use of
unsealed byproduct material
that requires a written
directive.

35.390 Training for use of
unsealed byproduct material for
which a written directive is
required.

35.392 Training for the oral
administration of sodium iodide I-
131 requiring a written directive
in quantities less than or equal to
1.22 Gigabecquerels (33
millicuries).

35.394 Training for the oral
administration of sodium iodide I-
131 requiring a written directive
in quantities greater than 1.22
Gigabecquerels (33 millicuries).

Subpart G--Sources for
Brachytherapy

Subpart F--Manual
Brachytherapy

Subpart F-- Manual
Brachytherapy

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

35.400 Use of sources for brachytherapy.	35.400 Use of sources for manual brachytherapy.	35.400 Use of sealed sources for manual brachytherapy.
35.404 Release of patients or human research subjects treated with temporary implants.	35.404 Radiation surveys of patients or human research subjects treated with implants.	35.404 Surveys after source implant and removal.
35.406 Brachytherapy sources inventory.	35.406 Brachytherapy sources inventory.	35.406 Brachytherapy sources accountability.
35.410 Safety instruction.	35.410 Safety instruction.	35.410 Safety instruction.
35.415 Safety precautions.	35.415 Safety precautions.	35.415 Safety precautions.
35.420 Possession of survey instrument.		
	35.432 Full calibration measurements of brachytherapy sources.	35.432 Calibration measurements of brachytherapy sealed sources.
		35.433 Decay of strontium-90 sources for ophthalmic treatments.
		35.457 Therapy-related computer systems.
	35.490 Training for use of manual brachytherapy sources.	35.490 Training for use of manual brachytherapy sources.
		35.491 Training for ophthalmic use of strontium-90.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart H--Sealed Sources for Diagnosis

Subpart G--Sealed Sources for Diagnosis

Subpart G--Sealed Sources for Diagnosis

35.500 Use of sealed sources for diagnosis.

35.500 Use of sealed sources for diagnosis.

35.500 Use of sealed sources for diagnosis.

35.520 Availability of survey instrument.

35.590 Training for use of sealed sources for diagnosis.

35.590 Training for use of sealed sources for diagnosis.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

CURRENT RULE	PROPOSED RULE	FINAL RULE
Subpart I--Teletherapy	Subpart H--Therapeutic Medical Devices	Subpart H--Photon Emitting Remote Afterloader Units, Teletherapy Units, and Gamma Stereotactic Radiosurgery Units
35.600 Use of a sealed source in a teletherapy unit.	35.600 Use of a sealed source in a device for therapeutic medical uses.	35.600 Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit.
	35.604 Radiation surveys of patients and human research subjects treated with remote afterloaders.	35.604 Surveys of patients and human research subjects treated with a remote afterloader unit.
35.605 Maintenance and repair restrictions.	35.605 Installation, maintenance, and repair.	35.605 Installation, maintenance, adjustment, and repair.
35.606 License amendments.		
35.610 Safety instruction.	35.610 Safety procedures and instructions for remote afterloaders, teletherapy units, and gamma stereotactic radiosurgery units.	35.610 Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.
35.615 Safety precautions.	35.615 Safety precautions for remote afterloaders, teletherapy units, and gamma stereotactic radiosurgery units.	35.615 Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.
35.620 Possession of survey instrument.		
35.630 Dosimetry equipment.	35.630 Dosimetry equipment.	35.630 Dosimetry equipment.
35.632 Full calibration measurements.	35.632 Full calibration measurements on teletherapy units.	35.632 Full calibration measurements on teletherapy units.
	35.633 Full calibration measurements on remote afterloaders.	35.633 Full calibration measurements on remote afterloader units.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	35.635 Full calibration measurements on gamma stereotactic radiosurgery units.	35.635 Full calibration measurements on gamma stereotactic radiosurgery units.
35.634 Periodic spot-checks. 35.636 Safety checks for teletherapy facilities. 35.641 Radiation surveys for teletherapy facilities.	35.642 Periodic spot-checks for teletherapy units.	35.642 Periodic spot-checks for teletherapy units.
35.643 Modification of teletherapy unit or room before beginning a treatment program.	35.643 Periodic spot-checks for high dose-rate and pulsed dose-rate remote afterloaders.	35.643 Periodic spot-checks for remote afterloader units.
35.645 Reports of teletherapy surveys, checks, tests, and measurements. 35.647 Five-year inspection.	35.644 Periodic spot-checks for low dose-rate remote afterloaders. 35.645 Periodic spot-checks for gamma stereotactic radiosurgery units. 35.647 Additional technical requirements for mobile remote afterloaders. 35.652 Radiation surveys. 35.655 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units. 35.657 Therapy-related computer systems.	35.645 Periodic spot-checks for gamma stereotactic radiosurgery units. 35.647 Additional technical requirements for mobile remote afterloader units. 35.652 Radiation surveys. 35.655 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units. 35.657 Therapy-related computer systems.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
35.690 Training for use of therapeutic medical devices.	35.690 Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.	
Subpart I--Reserved	Subpart I--Reserved	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart J--Training and Experience Requirements

Subpart J--Training and Experience Requirements

Subpart J--Reserved

35.900 Radiation Safety Officer.[35.50, 900]
 35.901 Training for experienced Radiation Safety Officer.
 35.910 Training for uptake, dilution, and excretion studies.
 35.920 Training for imaging and localization studies.
 35.930 Training for therapeutic use of unsealed byproduct material.
 35.932 Training for treatment of hyperthyroidism.
 35.934 Training for treatment of thyroid carcinoma.
 35.940 Training for use of brachytherapy sources.
 35.941 Training for ophthalmic use of strontium-90.

 35.950 Training for use of sealed sources for diagnosis.
 35.960 Training for teletherapy.
 35.961 Training for teletherapy physicist.
 35.970 Training for experienced authorized users.
 35.971 Physician training in a three month program.

35.900 Radiation Safety Officer.

 35.910 Training for uptake, dilution, and excretion studies.
 35.920 Training for imaging and localization studies.
 35.930 Training for therapeutic use of unsealed byproduct material.
 35.932 Training for treatment of hyperthyroidism.
 35.934 Training for treatment of thyroid carcinoma.
 35.940 Training for use of brachytherapy sources.
 35.941 Training for ophthalmic use of strontium-90.
 35.950 Training for use of sealed sources for diagnosis.
 35.960 Training for use of therapeutic medical devices.
 35.961 Training for an authorized medical physicist.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

35.972 Recentness of training.[35.57]

35.980 Training for an authorized nuclear pharmacist.

35.981 Training for experienced nuclear pharmacists.

35.980 Training for an authorized nuclear pharmacist.

35.981 Training for experienced nuclear pharmacists.

Subpart K--Other Medical Uses of Byproduct Material or Radiation from Byproduct Material

35.1000 Other medical uses of byproduct material or radiation from byproduct material.

Subpart K--Other Medical Uses of Byproduct Material or Radiation from Byproduct Material

35.1000 Other medical uses of byproduct material or radiation from byproduct material.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
Subpart L--Records	Subpart L--Records	
<p>35.2024 Records of authority and responsibilities for radiation protection programs.</p> <p>35.2026 Records of radiation program safety changes.</p> <p>35.2040 Records of written directives.</p> <p>35.2045 Records of medical events.</p>	<p>35.2024 Records of authority and responsibilities for radiation protection programs.</p> <p>35.2026 Records of radiation protection program changes.</p> <p>35.2040 Records of written directives.</p> <p>35.2045 Records of medical events.</p>	
35.2060 Records of instrument calibrations.	<p>35.2047 Record of a dose to an embryo/fetus or a nursing child.</p> <p>35.2060 Records of calibrations of instruments used to measure the activity of unsealed byproduct materials.</p>	
<p>35.2061 Records of radiation survey instrument calibrations.</p> <p>35.2063 Records of dosages of unsealed byproduct material for medical use.</p>	<p>35.2061 Records of radiation survey instrument calibrations.</p> <p>35.2063 Records of dosages of unsealed byproduct material for medical use.</p>	
35.2067 Records for possession of sealed sources and brachytherapy sources.	35.2067 Records of possession of sealed sources and brachytherapy sources.	
35.2070 Records of surveys for ambient radiation exposure rate.	35.2070 Records of surveys for ambient radiation exposure rate.	
35.2075 Records of the release of individuals containing radiopharmaceuticals or implants.	35.2075 Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material.	
35.2080 Records of administrative and technical	35.2080 Records of administrative and technical	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
requirements that apply to the provision of mobile services.	requirements that apply to the provision of mobile medical services.	requirements that apply to the provision of mobile medical services.
35.2092 Records of waste disposal.	35.2092 Records of decay-in-storage.	35.2092 Records of decay-in-storage.
35.2204 Records of molybdenum-99 concentration.	35.2204 Records of molybdenum-99 concentrations.	35.2204 Records of molybdenum-99 concentrations.
35.2310 Records of instruction and training.	35.2310 Records of safety instruction.	35.2310 Records of safety instruction.
35.2404 Records of radiation surveys of patients and human research subjects.	35.2404 Records of surveys after source implant and removal.	35.2404 Records of surveys after source implant and removal.
35.2406 Records of brachytherapy source inventory.	35.2406 Records of brachytherapy source accountability.	35.2406 Records of brachytherapy source accountability.
35.2432 Records of full calibrations on brachytherapy sources.	35.2432 Records of calibration measurements of brachytherapy sealed sources.	35.2432 Records of calibration measurements of brachytherapy sealed sources.
35.2605 Records of installation, maintenance, and repair.	35.2433 Records of decay of strontium-90 sources for ophthalmic treatments.	35.2605 Records of installation, maintenance, adjustment, and repair.
35.2630 Records of dosimetry equipment.	35.2630 Records of dosimetry equipment.	35.2630 Records of dosimetry equipment.
35.2632 Records of teletherapy full calibrations.	35.2632 Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.	35.2632 Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.
35.2633 Records of remote afterloader full calibrations.		

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
35.2635 Records of gamma stereotactic radiosurgery unit full calibrations.		
35.2642 Records of periodic spot-checks for teletherapy units.	35.2642 Records of periodic spot-checks for teletherapy units.	
35.2643 Records of periodic spot-checks for remote afterloaders.	35.2643 Records of periodic spot-checks for remote afterloader units.	
35.2645 Records of periodic spot-checks for gamma stereotactic radiosurgery units.	35.2645 Records of periodic spot-checks for gamma stereotactic radiosurgery units.	
35.2647 Records of additional technical requirements for mobile remote afterloaders.	35.2647 Records of additional technical requirements for mobile remote afterloader units.	
35.2652 Records of surveys of therapeutic treatment units.	35.2652 Records of surveys of therapeutic treatment units.	
35.2655 Records of 5-year inspection for teletherapy and gamma stereotactic radiosurgery units.	35.2655 Records of 5-year inspection for teletherapy and gamma stereotactic radiosurgery units.	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	Subpart M--Reports	Subpart M--Reports
	35.3045 Reports of medical events.	35.3045 Report and notification of a medical event.
	35.3047 Report of a dose to an embryo/fetus or a nursing child.	35.3047 Report and notification of a dose to an embryo/fetus or a nursing child.
	35.3067 Reports of leaking sources.	35.3067 Report of a leaking source.

Subpart K--Enforcement	Subpart N--Enforcement	Subpart N--Enforcement
35.990 Violations.	35.4001 Violations.	35.4001 Violations.
35.991 Criminal penalties.	35.4002 Criminal penalties.	35.4002 Criminal penalties.
35.999 Resolution of conflicting requirements during transition period.		

Appendix A to 10 CFR Part 35--Examining Organization or Entity

Authority: Secs. 81, 161, 182, 183, 68 Stat. 935, 948, 953, 954, as amended (42 U.S.C. 2111, 2201, 2232, 2233); sec. 201, 88 Stat. 1242, as amended (42 U.S.C. 5841).	Authority: Secs. 81, 161, 182, 183, 68 Stat. 935, 948, 953, 954, as amended (42 U.S.C. 2111, 2201, 2232, 2233); sec. 201, 88 Stat. 1242, as amended (42 U.S.C. 5841).
---	---

Source: 51 FR 36951, Oct. 16, 1986, unless otherwise noted.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart A--General Information

Subpart A--General Information

Subpart A--General Information

Section 35.1 Purpose and scope.

Sec. 35.1 Purpose and scope.

§ 35.1 Purpose and scope.

This part prescribes requirements and provisions for the medical use of byproduct material and for issuance of specific licenses authorizing the medical use of this material. These requirements and provisions provide for the protection of the public health and safety. The requirements and provisions of this part are in addition to, and not in substitution for, others in this chapter. The requirements and provisions of parts 19, 20, 21, 30, 71, and 170 of this chapter apply to applicants and licensees subject to this part unless specifically exempted.

This part prescribes requirements and provisions for the medical use of byproduct material and for issuance of specific licenses authorizing the medical use of this material. These requirements and provisions provide for the radiation safety of workers, the general public, patients, and human research subjects. The requirements and provisions of this part are in addition to, and not in substitution for, others in this chapter. The requirements and provisions of parts 19, 20, 21, 30, 71, 170, and 171 of this chapter apply to applicants and licensees subject to this part unless specifically exempted.

This part contains the requirements and provisions for the medical use of byproduct material and for issuance of specific licenses authorizing the medical use of this material. These requirements and provisions provide for the radiation safety of workers, the general public, patients, and human research subjects. The requirements and provisions of this part are in addition to, and not in substitution for, others in this chapter. The requirements and provisions of parts 19, 20, 21, 30, 71, 170, and 171 of this chapter apply to applicants and licensees subject to this part unless specifically exempted.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.2 Definitions.

Sec. 35.2 Definitions.

§ 35.2 Definitions.

Address of use means the building or buildings that are identified on the license and where byproduct material may be received, used, or stored.

Address of use means the building or buildings that are identified on the license and where byproduct material may be received, used, or stored.

Address of use means the building or buildings that are identified on the license and where byproduct material may be received, used, or stored.

Agreement State means any State with which the Commission or the Atomic Energy Commission has entered into an effective agreement under subsection 274b of the Atomic Energy Act of 1954, as amended.

Agreement State means any State with which the Commission or the Atomic Energy Commission has entered into an effective agreement under subsection 274b of the Atomic Energy Act of 1954, as amended.

Agreement State means any State with which the Commission or the Atomic Energy Commission has entered into an effective agreement under subsection 274b of the Atomic Energy Act of 1954, as amended.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

ALARA (as low as reasonably achievable) means making every reasonable effort to maintain exposures to radiation as far below the dose limits as is practical:

(1) Consistent with the purpose for which the licensed activity is undertaken,

(2) Taking into account the state of technology, the economics of improvements in relation to benefits to the public health and safety, and other societal and socioeconomic considerations, and

(3) In relation to utilization of nuclear energy in the public interest.

Area of use means a portion of an address of use that has been set aside for the purpose of receiving, using, or storing byproduct material.

Area of use means a portion of an address of use that has been set aside for the purpose of receiving, using, or storing byproduct material.

Area of use means a portion of an address of use that has been set aside for the purpose of receiving, preparing, using, or storing byproduct material.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Authorized medical physicist means a physicist who--

- (1) Meets the requirements in Secs. 35.51(a) and 35.59 or Secs. 35.961 and 35.59; or
- (2) Is identified as a medical physicist on a Commission or Agreement State license; or
- (3) Is identified as a medical physicist on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the use of byproduct material.

Authorized medical physicist means an individual who --

- (1) Meets the requirements in §§ 35.51(a) and 35.59; or
- (2) Is identified as an authorized medical physicist on --
 - (i) A specific license issued by the Commission or Agreement State license;
 - (ii) A permit issued by a Commission master material licensee;
 - (iii) A permit issued by a Commission or Agreement State broad scope licensee; or
 - (iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating authorized medical physicists.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Authorized nuclear pharmacist means a pharmacist who is:

(1) Board certified as a nuclear pharmacist by the Board of Pharmaceutical Specialties;

(2) Identified as an authorized nuclear pharmacist on a Commission or Agreement State license that authorizes the use of byproduct material in the practice of nuclear pharmacy; or

(3) Identified as an authorized nuclear pharmacist on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the use of byproduct

Authorized nuclear pharmacist means a pharmacist who--

(1) Meets the requirements in Secs. 35.55(a) and 35.59 or Secs. 35.980(a) and 35.59; or

(2) Is identified as an authorized nuclear pharmacist on a Commission or Agreement State license that authorizes the use of byproduct material in the practice of nuclear pharmacy; or

(3) Is identified as an authorized nuclear pharmacist on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the use of byproduct

Authorized nuclear pharmacist means a pharmacist who --

(1) Meets the requirements in §§ 35.55(a) and 35.59; or

(2) Is identified as an authorized nuclear pharmacist on --

(i) A specific license issued by the Commission or Agreement State license;

(ii) A permit issued by a Commission master material licensee;

(iii) A permit issued by a Commission or Agreement State broad scope licensee; or

(iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating authorized nuclear pharmacists; or

(3) Is designated as an authorized nuclear pharmacist in accordance with § 32.72(b)(4).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

material in the practice of nuclear pharmacy.

material in the practice of nuclear pharmacy; or
(4) Is approved as an authorized nuclear pharmacist by a nuclear pharmacy authorized by the Commission to approve authorized nuclear pharmacists.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Authorized user means a physician, dentist, or podiatrist who is:

(1) Board certified by at least one of the boards listed in Paragraph (a) of Sections 35.910, 35.920, 35.930, 35.940, 35.950, or 35.960;

(2) Identified as an authorized user on a Commission or Agreement State license that authorizes the medical use of byproduct material; or

(3) Identified as an authorized user on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the medical use of byproduct material.

Authorized user means a physician, dentist, or podiatrist who--

(1) Meets the requirements in Secs. 35.290(a), 35.292(a), 35.390(a), 35.490(a), 35.590(a), or 35.690(a) and Sec. 35.59, or Secs. 35.910, 35.920, 35.930, 35.932, 35.934, 35.940, 35.941, 35.950, 35.960 and Sec. 35.59; or

(2) Is identified as an authorized user on a Commission or Agreement State license that authorizes the medical use of byproduct material; or

(3) Is identified as an authorized user on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the medical use of byproduct material.

Authorized user means a physician, dentist, or podiatrist who --

(1) Meets the requirements in §§ 35.59 and 35.190(a), 35.290(a), 35.390(a), 35.392(a), 35.394(a), 35.490(a), 35.491, 35.590(a), or 35.690(a); or

(2) Is identified as an authorized user on --

(i) A specific license issued by the Commission or Agreement State license;

(ii) A permit issued by a Commission master material licensee;

(iii) A permit issued by a Commission or Agreement State broad scope licensee; or

(iv) A permit issued by a Commission master material licensee broad scope permittee where that permittee has been given authorization to issue permits designating authorized users.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Brachytherapy means a method of radiation therapy in which sealed sources are used to deliver a radiation dose at a distance of up to a few centimeters by surface, intracavitary, intraluminal, or interstitial application.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Brachytherapy source means an individual sealed source or a manufacturer-assembled source train that is not designed to be disassembled by the user.

Brachytherapy source means a radioactive sealed source or a manufacturer-assembled source train or a combination of these sources that is designed to deliver a therapeutic dose within a distance of a few centimeters.

Brachytherapy source means a radioactive source or a manufacturer-assembled source train or a combination of these sources that is designed to deliver a therapeutic dose within a distance of a few centimeters.

Client's address means the area of use or a temporary jobsite for the purpose of providing mobile medical service in accordance with § 35.80.

Dedicated check source means a radioactive source that is used to assure the constant operation of a radiation detection or measurement device over several months or years.

Dedicated check source means a radioactive source that is used to assure the constant operation of a radiation detection or measurement device over several months or years.

Dedicated check source means a radioactive source that is used to assure the constant operation of a radiation detection or measurement device over several months or years.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Dental use means the intentional external administration of the radiation from byproduct material to human beings in the practice of dentistry in accordance with a license issued by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico.

Dentist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice dentistry.

Dentist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice dentistry.

Dentist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice dentistry.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Diagnostic clinical procedures manual means a collection of written procedures that describes each method (and other instructions and precautions) by which the licensee performs diagnostic clinical procedures; where each diagnostic clinical procedure has been approved by the authorized user and includes the radiopharmaceutical, dosage, and route of administration.

Diagnostic clinical procedures manual means a collection of written procedures that describes each method (and other instructions and precautions) by which the licensee performs diagnostic clinical procedures; where each diagnostic clinical procedure has been approved by the authorized user and includes the radiopharmaceutical, dosage, and route of administration.

High dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate in excess of 2 gray (200 rads) per hour at the point or surface where the dose is prescribed.

High dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate in excess of 12 gray (1200 rads) per hour at the point or surface where the dose is prescribed.

Low dose-rate remote afterloader as used in this part, means a device that remotely delivers a dose rate of less than 2 gray (200 rads) per hour at the point or surface where the dose is prescribed.

Low dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate of less than or equal to 2 gray (200 rads) per hour at the point or surface where the dose is prescribed.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Management means the chief executive officer or that person's delegate or delegates.

Management means the chief executive officer or that person's delegate or delegates.

Management means the chief executive officer or other individual having the authority to manage, direct, or administer the licensee's activities, or those persons' delegate or delegates.

Manual brachytherapy, as used in this part, means a type of brachytherapy in which the radioactive sources (e.g., seeds, ribbons) are manually inserted either into the body cavities that are in close proximity to a treatment site or directly into the tissue volume.

Medical event means an event that meets the criteria in Sec. 35.3045(a).

Medical event means an event that meets the criteria in § 35.3045(a).

Medical Institution means an organization in which several medical disciplines are practiced.

Medical institution means an organization in which several medical disciplines are practiced.

Medical institution means an organization in which several medical disciplines are practiced.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

Medical use means the intentional internal or external administration of byproduct material or the radiation therefrom to patients or human research subjects under the supervision of an authorized user.

PROPOSED RULE

Medical use means the intentional internal or external administration of byproduct material or the radiation from byproduct material to patients or human research subjects under the supervision of an authorized user.

FINAL RULE

Medical use means the intentional internal or external administration of byproduct material or the radiation from byproduct material to patients or human research subjects under the supervision of an authorized user.

Ministerial change means a change that is made, after ascertaining the applicable requirements, by persons in authority in conformance with the requirements and without making a discretionary judgment about whether those requirements should apply in the case at hand.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Misadministration
means the administration of:

(1) A

radiopharmaceutical dosage
greater than 30 microcuries of
either sodium iodide I-125 or
I-131:

(i) Involving the wrong
individual, or wrong
radiopharmaceutical; or

(ii) When both the
administered dosage differs
from the prescribed dosage by
more than 20 percent of the
prescribed dosage and the
difference between the
administered dosage and
prescribed dosage exceeds 30
microcuries.

(2) A therapeutic
radiopharmaceutical dosage,
other than sodium iodide I-125
or I-131:

(i) Involving the wrong
individual, wrong
radiopharmaceutical, or wrong
route of administration; or

(ii) When the
administered dosage differs
from the prescribed dosage by
more than 20 percent of the
prescribed dosage.

(3) A gamma
stereotactic radiosurgery
radiation dose:

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(i) Involving the wrong individual, or wrong treatment site; or

(ii) When the calculated total administered dose differs from the total prescribed dose by more than 10 percent of the total prescribed dose.

(4) A teletherapy radiation dose:

(i) Involving the wrong individual, wrong mode of treatment, or wrong treatment site;

(ii) When the treatment consists of three or fewer fractions and the calculated total administered dose differs from the total prescribed dose by more than 10 percent of the total prescribed dose;

(iii) When the calculated weekly administered dose exceeds the weekly prescribed dose by 30 percent or more of the weekly prescribed dose; or

(iv) When the calculated total administered dose differs from the total prescribed dose by more than 20 percent of the total prescribed dose.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(5) A brachytherapy radiation dose:

(i) Involving the wrong individual, wrong radioisotope, or wrong treatment site (excluding, for permanent implants, seeds that were implanted in the correct site but migrated outside the treatment site);

(ii) Involving a sealed source that is leaking;

(iii) When, for a temporary implant, one or more sealed sources are not removed upon completion of the procedure; or

(iv) When the calculated administered dose differs from the prescribed dose by more than 20 percent of the prescribed dose.

(6) A diagnostic radiopharmaceutical dosage, other than quantities greater than 30 microcuries of either sodium iodide I - 125 or I - 131, both:

(i) Involving the wrong individual, wrong radiopharmaceutical, wrong route of administration, or when the administered dosage differs from the prescribed dosage; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(ii) When the dose to the individual exceeds 5 rems effective dose equivalent or 50 rems dose equivalent to any individual organ.

Medium dose-rate remote afterloader, as used in this part, means a device that remotely delivers a dose rate of greater than 2 gray (200 rads), but less than 12 gray (1200 rads) per hour at the point or surface where the dose is prescribed.

Mobile nuclear medicine service means the transportation and medical use of byproduct material.

Mobile service means the transportation and medical use of byproduct material by the same licensee at temporary jobsites.

Mobile medical service means the transportation of byproduct material to and its medical use at the client's address.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

Output means the exposure rate, dose rate, or a quantity related in a known manner to these rates from a teletherapy unit for a specified set of exposure conditions.

PROPOSED RULE

Output means the exposure rate, dose rate, or a quantity related in a known manner to these rates from a teletherapy unit for a specified set of exposure conditions.

FINAL RULE

Output means the exposure rate, dose rate, or a quantity related in a known manner to these rates from a brachytherapy source or a teletherapy, remote afterloader, or gamma stereotactic radiosurgery unit for a specified set of exposure conditions.

Patient intervention means actions by the patient or human research subject, whether intentional or unintentional, such as dislodging or removing treatment devices or prematurely terminating the administration.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Pharmacist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice pharmacy.

Pharmacist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice pharmacy.

Pharmacist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice pharmacy.

Physician means a medical doctor or doctor of osteopathy licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to prescribe drugs in the practice of medicine.

Physician means a medical doctor or doctor of osteopathy licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to prescribe drugs in the practice of medicine.

Physician means a medical doctor or doctor of osteopathy licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to prescribe drugs in the practice of medicine.

Podiatric use means the intentional external administration of the radiation from byproduct material to human beings in the practice of podiatry in accordance with a license issued by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

Podiatrist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice podiatry.

PROPOSED RULE

Podiatrist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice podiatry.

FINAL RULE

Podiatrist means an individual licensed by a State or Territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to practice podiatry.

Preceptor means an individual who provides or directs the training and experience required for an individual to become an authorized user, an authorized medical physicist, an authorized nuclear pharmacist, or a Radiation Safety Officer.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Prescribed dosage means the quantity of radiopharmaceutical activity as documented:

- (1) In a written directive; or
- (2) Either in the diagnostic clinical procedures manual or in any appropriate record in accordance with the directions of the authorized user for diagnostic procedures.

Prescribed dosage means the quantity of radiopharmaceutical activity as documented—

- (1) In a written directive; or
- (2) Either in the diagnostic clinical procedures manual or in any appropriate record in accordance with the directions of the authorized user for diagnostic procedures.

Prescribed dosage means the specified activity or range of activity of unsealed byproduct material as documented --

- (1) In a written directive; or
- (2) In accordance with the directions of the authorized user for procedures performed pursuant to §§ 35.100 and 35.200.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Prescribed dose means:
 (1) For gamma stereotactic radiosurgery, the total dose as documented in the written directive;
 (2) For teletherapy, the total dose and dose per fraction as documented in the written directive; or
 (3) For brachytherapy, either the total source strength and exposure time or the total dose, as documented in the written directive.

Prescribed dose means--
 (1) For gamma stereotactic radiosurgery, the total dose as documented in the written directive;
 (2) For teletherapy, the total dose and dose per fraction as documented in the written directive;
 (3) For brachytherapy, either the total source strength and exposure time or the total dose, as documented in the written directive; or
 (4) For remote afterloaders, the total dose as documented in the written directive.

Pulsed dose-rate remote afterloader means a special type of remote afterloading device that uses a single source capable of delivering dose rates in the “high dose rate” range, but is used to simulate the radiobiology of a low dose rate treatment by inserting the source for a given fraction of each hour.

Prescribed dose means --
 (1) For gamma stereotactic radiosurgery, the total dose as documented in the written directive;
 (2) For teletherapy, the total dose and dose per fraction as documented in the written directive;
 (3) For brachytherapy, either the total source strength and exposure time or the total dose, as documented in the written directive; or
 (4) For remote afterloaders, the total dose and dose per fraction as documented in the written directive.

Pulsed dose-rate remote afterloader, as used in this part, means a special type of remote afterloading device that uses a single source capable of delivering dose rates in the “high dose-rate” range, but --
 (1) Is approximately one-tenth of the activity of typical high dose-rate remote afterloader sources; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(2) Is used to simulate the radiobiology of a low dose rate treatment by inserting the source for a given fraction of each hour.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Radiation Safety Officer means the individual identified as the Radiation Safety Officer on a Commission license.

Radiation Safety Officer means the individual identified as the Radiation Safety Officer on a Commission license who—

- (1) Meets the requirements in Secs. 35.50 and 35.59 or Secs. 35.900 and 35.59; or
- (2) Is identified as a Radiation Safety Officer on a Commission or Agreement State license.

Radiation Safety Officer means an individual who --

- (1) Meets the requirements in §§ 35.50(a) and 35.59; or
- (2) Is identified as a Radiation Safety Officer on --
 - (i) A specific license issued by the Commission or Agreement State license;
 - (ii) A permit issued by a Commission master material licensee;
 - (iii) A permit issued by a Commission or Agreement State broad scope licensee; or
 - (iv) A permit issued by a Commission master material license broad scope permittee where that permittee has been given authorization to issue permits designating Radiation Safety Officers.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Recordable event means the administration of:

(1) A

radiopharmaceutical or radiation without a written directive where a written directive is required;

(2) A

radiopharmaceutical or radiation where a written directive is required without daily recording of each administered radiopharmaceutical dosage or radiation dose in the appropriate record;

(3) A

radiopharmaceutical dosage greater than 30 microcuries of either sodium iodide I-125 or I-131 when both:

(i) The administered dosage differs from the prescribed dosage by more than 10 percent of the prescribed dosage, and

(ii) The difference between the administered dosage and prescribed dosage exceeds 15 microcuries;

(4) A therapeutic radiopharmaceutical dosage, other than sodium iodide I-125 or I-131, when the

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

administered dosage differs from the prescribed dosage by more than 10 percent of the prescribed dosage;

(5) A teletherapy radiation dose when the calculated weekly administered dose exceeds the weekly prescribed dose by 15 percent or more of the weekly prescribed dose; or

(6) A brachytherapy radiation dose when the calculated administered dose differs from the prescribed dose by more than 10 percent of the prescribed dose.

Sealed source means any byproduct material that is encased in a capsule designed to prevent leakage or escape of the byproduct material.

Sealed source means any byproduct material that is encased in a capsule designed to prevent leakage or escape of the byproduct material.

Sealed source means any byproduct material that is encased in a capsule designed to prevent leakage or escape of the byproduct material.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sealed Source and Device Registry means the national registry that contains all the registration certificates, generated by both NRC and the Agreement States, that summarize the radiation safety information for the sealed sources and devices and describe the licensing and use conditions approved for the product.

Sealed Source and Device Registry means the national registry that contains all the registration certificates, generated by both NRC and the Agreement States, that summarize the radiation safety information for the sealed sources and devices and describe the licensing and use conditions approved for the product.

Stereotactic radiosurgery means the use of external radiation in conjunction with a stereotactic guidance device to very precisely deliver a dose to a tissue volume.

Stereotactic radiosurgery means the use of external radiation in conjunction with a stereotactic guidance device to very precisely deliver a dose to a tissue volume.

Structured educational program means an educational program designed to impart particular knowledge and practical education through interrelated studies and supervised training.

Structured educational program means an educational program designed to impart particular knowledge and practical education through interrelated studies and supervised training.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Teletherapy physicist means the individual identified as the teletherapy physicist on a Commission license.

Teletherapy, as used in this part, means a method of radiation therapy in which collimated gamma rays are delivered at a distance from the patient or human research subject.

Temporary jobsite means a location where mobile services are conducted other than those location(s) of use authorized on the license.

Temporary jobsite means a location where mobile medical services are conducted other than those location(s) of use authorized on the license.

Therapeutic dosage means a dosage of an unsealed byproduct material that is intended to deliver a radiation dose to a patient or human research subject for palliative or curative treatment.

Therapeutic dose means a radiation dose delivered from a source containing byproduct material to a patient or human research subject for palliative or curative treatment.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Treatment site means the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive.

Treatment site means the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive.

Type of use means use of byproduct material under §§ 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, or 35.1000.

Unit dosage means a dosage intended for medical use in a single patient or human research subject that has been obtained from a manufacturer or preparer licensed pursuant to Sec. 32.72 of this chapter or equivalent Agreement State requirements.

Unit dosage means a dosage prepared for medical use for administration as a single dosage to a patient or human research subject without any further manipulation of the dosage after it is initially prepared.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

Written directive means an order in writing for a specific patient or human research subject, dated and signed by an authorized user prior to the administration of a radiopharmaceutical or radiation, except as specified in paragraph (6) of this definition, containing the following information:

(1) For any administration of quantities greater than 30 microcuries of either sodium iodide I-125 or I-131: the dosage;

(2) For a therapeutic administration of a radiopharmaceutical other than sodium iodide I-125 or I-131: the radiopharmaceutical, dosage, and route of administration;

(3) For gamma stereotactic radiosurgery: target coordinates, collimator size, plug pattern, and total dose;

(4) For teletherapy: the total dose, dose per fraction, treatment site, and overall treatment period;

(5) For high-dose-rate remote afterloading brachytherapy: the

PROPOSED RULE

Written directive means an authorized user's written order for the administration of byproduct material or radiation from byproduct material to a specific patient or human research subject, as specified in Sec. 35.40.

FINAL RULE

Written directive means an authorized user's written order for the administration of byproduct material or radiation from byproduct material to a specific patient or human research subject, as specified in § 35.40.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

radioisotope, treatment site,
and total dose; or
 (6) For all other
brachytherapy:
 (i) Prior to implantation:
the radioisotope, number of
sources, and source strengths;
and
 (ii) After implantation
but prior to completion of the
procedure: the radioisotope,
treatment site, and total
source strength and exposure
time (or, equivalently, the total
dose).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.5 Maintenance of records.

Each record required by this part must be legible throughout the retention period specified by each Commission regulation. The record may be the original or a reproduced copy or a microform provided that the copy or microform is authenticated by authorized personnel and that the microform is capable of producing a clear copy throughout the required retention period. The record may also be stored in electronic media with the capability for producing legible, accurate, and complete records during the required retention period. Records such as letters, drawings, specifications, must include all pertinent information such as stamps, initials, and signatures. The licensee shall maintain adequate safeguards against tampering with and loss of records.

Sec. 35.5 Maintenance of records.

Each record required by this part must be legible throughout the retention period specified by each Commission regulation. The record may be the original or a reproduced copy or a microform provided that the copy or microform is authenticated by authorized personnel and that the microform is capable of producing a clear copy throughout the required retention period. The record may also be stored in electronic media with the capability for producing legible, accurate, and complete records during the required retention period. Records such as letters, drawings, and specifications, must include all pertinent information such as stamps, initials, and signatures. The licensee shall maintain adequate safeguards against tampering with and loss of records.

§ 35.5 Maintenance of records.

Each record required by this part must be legible throughout the retention period specified by each Commission regulation. The record may be the original, a reproduced copy, or a microform provided that the copy or microform is authenticated by authorized personnel and the microform is capable of producing a clear copy throughout the required retention period. The record may also be stored in electronic media with the capability for producing legible, accurate, and complete records during the required retention period. Records such as letters, drawings, and specifications must include all pertinent information such as stamps, initials, and signatures. The licensee shall maintain adequate safeguards against tampering with and loss of records.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.6 Provisions for research involving human subjects.

A licensee may conduct research involving human subjects using byproduct material provided that the research is conducted, funded, supported, or regulated by another Federal Agency which has implemented the Federal Policy for the Protection of Human Subjects. Otherwise, a licensee shall apply for and receive approval of a specific amendment to its NRC license before conducting such research. Both types of licensees shall, at a minimum, obtain informed consent from the human subjects and obtain prior review and approval of the research activities by an "Institutional Review Board" in accordance with the meaning of these terms as defined and described in the Federal Policy for the Protection of Human Subjects,

Sec. 35.6 Provisions for research involving human subjects.

A licensee may conduct research involving human subjects using byproduct material provided that the research is conducted, funded, supported, or regulated by another Federal Agency which has implemented the Federal Policy for the Protection of Human Subjects. Otherwise, a licensee shall apply for and receive approval of a specific amendment to its NRC license before conducting such research. Both types of licensees shall, at a minimum, obtain informed consent from the human subjects and obtain prior review and approval of the research activities by an "Institutional Review Board" in accordance with the meaning of these terms as defined and described in the Federal Policy for the Protection of Human Subjects.

§ 35.6 Provisions for the protection of human research subjects.

(a) A licensee may conduct research involving human research subjects only if using the byproduct materials specified on its license for the uses authorized on its license.

(b) If the research is conducted, funded, supported, or regulated by another Federal agency that has implemented the Federal Policy for the Protection of Human Subjects (Federal Policy), the licensee shall, prior to conducting research --

(1) Obtain review and approval of the research from an "Institutional Review Board," as defined and described in the Federal Policy; and

(2) Obtain "informed consent," as defined and described in the Federal Policy, from the human research subject.

(c) If the research will not be conducted, funded, supported, or regulated by another Federal agency that

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

has implemented the Federal Policy, the license shall, prior to conducting research, apply for and receive a specific amendment to its NRC medical use license. The amendment request must include a written commitment that the licensee will, prior to conducting research --

(1) Obtain review and approval of the research from an "Institutional Review Board," as defined and described in the Federal Policy; and

(2) Obtain "informed consent", as defined and described in the Federal Policy, from the human research subject.

(d) Nothing in this section relieves licensees from complying with the other requirements in this part.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.7 FDA, other Federal, and State requirements.

Sec. 35.7, FDA, other Federal, and State requirements.

§ 35.7 FDA, other Federal, and State requirements.

Nothing in this part relieves the licensee from complying with applicable FDA, other Federal, and State requirements governing radioactive drugs or devices.

Nothing in this part relieves the licensee from complying with applicable FDA, other Federal, and State requirements governing radioactive drugs or devices.

Licensees are required to comply with applicable FDA, other Federal, and State requirements governing radioactive drugs or devices.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.8 Information collection requirements: OMB approval.

(a) The Nuclear Regulatory Commission has submitted the information collection requirements contained in this part to the Office of Management and Budget (OMB) for approval as required by the Paperwork Reduction Act (44 U.S.C. 3501 et seq.). The NRC may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. OMB has approved the information collection requirements contained in this part under control number 3150-0010.

(b) The approved information collection requirements contained in this part appear in Sections 35.6, 35.12, 35.13, 35.14, 35.20, 35.21, 35.22, 35.23, 35.29, 35.31, 35.50, 35.51, 35.52, 35.53, 35.59, 35.60, 35.61, 35.70, 35.75, 35.80, 35.92, 35.204, 35.205, 35.310, 35.315, 35.404, 35.406,

Sec. 35.8 Information collection requirements: OMB approval.

(a) The Commission has submitted the information collection requirements contained in this part to the Office of Management and Budget (OMB) for approval as required by the Paperwork Reduction Act (44 U.S.C. 3501 et seq.). The NRC may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. OMB has approved the information collection requirements in this part under control number 3150-0010.

(b) The approved information collection requirements contained in this part appear in Secs. 35.6, 35.12, 35.13, 35.14, 35.24, 35.26, 35.27, 35.40, 35.41, 35.50, 35.51, 35.55, 35.60, 35.61, 35.62, 35.63, 35.67, 35.69, 35.70, 35.75, 35.80, 35.92, 35.204, 35.290, 35.292, 35.310, 35.315, 35.390,

§ 35.8 Information collection requirements: OMB approval.

(a) The Commission has submitted the information collection requirements contained in this part to the Office of Management and Budget (OMB) for approval as required by the Paperwork Reduction Act (44 U.S.C. 3501 et seq.). The NRC may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. OMB has approved the information collection requirements in this part under control number 3150-0010.

(b) The approved information collection requirements contained in this part appear in §§ 35.6, 35.12, 35.13, 35.14, 35.19, 35.24, 35.26, 35.27, 35.40, 35.41, 35.50, 35.51, 35.55, 35.57, 35.60, 35.61, 35.63, 35.67, 35.69, 35.70, 35.75, 35.80, 35.92, 35.190, 35.204, 35.290, 35.310, 35.315,

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

35.410, 35.415, 35.606,
35.610, 35.615, 35.630,
35.632, 35.634, 35.636,
35.641, 35.643, 35.645,
35.647, 35.980, 35.981.

35.404, 35.406, 35.410,
35.415, 35.432, 35.490,
35.590, 35.604, 35.605,
35.610, 35.630, 35.632,
35.633, 35.635, 35.642,
35.643, 35.644, 35.645,
35.647, 35.652, 35.655,
35.690, 35.900, 35.910,
35.920, 35.930, 35.940,
35.950, 35.960, 35.961,
35.980, 35.981, 35.2024,
35.2026, 35.2040, 35.2045,
35.2060, 35.2061, 35.2063,
35.2067, 35.2070, 35.2075,
35.2080, 35.2092, 35.2204,
35.2310, 35.2404, 35.2406,
35.2432, 35.2605, 35.2630,
35.2632, 35.2633, 35.2635,
35.2642, 35.2643, 35.2645,
35.2647, 35.2652, 35.2655,
35.3045, 35.3047, 35,3067,
and Appendix A.

35.390, 35.392, 35.394,
35.404, 35.406, 35.410,
35.415, 35.432, 35.490,
35.491, 35.590, 35.604,
35.605, 35.610, 35.615,
35.630, 35.632, 35.633,
35.635, 35.642, 35.643,
35.645, 35.647, 35.652,
35.655, 35.690, 35.2024,
35.2026, 35.2040, 35.2045,
35.2047, 35.2060, 35.2061,
35.2063, 35.2067, 35.2070,
35.2075, 35.2080, 35.2092,
35.2204, 35.2310, 35.2404,
35.2406, 35.2432, 35.2433,
35.2605, 35.2630, 35.2632,
35.2642, 35.2643, 35.2645,
35.2647, 35.2652, 35.2655,
35.3045, 35.3047, and
35.3067.

(c) This part contains information collection requirements in addition to those approved under the control number specified in paragraph (a) of this section. These information collection requirements and the control numbers under which they are approved as follows:

(1) In Section 35.12, Form NRC-313 is approved

(c) This part contains information collection requirements in addition to those approved under the control number specified in paragraph (a) of this section. These information collection requirements and the control numbers under which they are approved as follows:

(1) In Sec. 35.12, NRC Form 313, including NRC Forms 313A, and 313B which

(c) This part contains information collection requirements in addition to those approved under the control number specified in paragraph (a) of this section. These information collection requirements and the control numbers under which they are approved are as follows:

(1) In § 35.12, NRC Form 313, including NRC Forms 313A and 313B, which

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

under control number 3150-0120.

licensees may use to provide supplemental information, is approved under control number 3150-0120.

licensees may use to provide supplemental information, is approved under control number 3150-0120.

(2) [Reserved]

(2) [Reserved]

(2) [Reserved]

(d) OMB has assigned control number 3150-0171 for the information collection requirements contained in Sections 35.32 and 35.33.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
Sec. 35.10 Implementation.	§ 35.10 Implementation.	
(a) A licensee shall implement the provisions in this part on or before [date 6 months from publication of the Final Rule], with the exception of the requirements listed in paragraph (b) of this section.	(a) A licensee shall implement the provisions in this part on or before [insert date 6 months from publication of the Final Rule].	
(b) A licensee shall implement the training requirements in Secs. 35.50(a), 35.51(a), 35.55(a), 35.59, 35.290(a), 35.292(a), 35.390(a), 35.490(a), 35.590(a), and 35.690(a) on or before [date--2 years from publication of the Final Rule].		
(c) Prior to [date--2 years from publication of the Final Rule], a licensee shall satisfy the training requirements of this part for a Radiation Safety Officer, an authorized medical physicist, an authorized nuclear pharmacist, or an authorized user by complying with either:		
(1) The appropriate training requirements in subpart J; or		
(2) The appropriate training requirements in		

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

subpart B or subparts D through H.

(d) If the requirements of this part are more restrictive than the existing license condition, the licensee shall comply with this part unless exempted by paragraph (f) of this section.

(e) Any existing license condition that is more restrictive than a requirement in this part remains in effect until there is a license amendment or license renewal.

(f) If a license condition exempted a licensee from a provision of part 35 on [date--6 months from publication of the Final Rule], it will continue to exempt a licensee from the corresponding provision in this part.

(g) If a license condition cites provisions in part 35 that will be deleted on [date--6 months from publication of the Final Rule], then the license condition remains in effect until there is a license amendment or license renewal that modifies or removes this condition.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) If a license condition exempted a licensee from a provision of Part 35 on [insert date--6 months from publication of the Final Rule], then the license condition continues to exempt the licensee from the requirements in the corresponding provision until there is a license amendment or license renewal that modifies or removes the license condition.

(c) When a requirement in this part differs from the requirement in an existing license condition, the requirement in this part shall govern.

(d) Licensees shall continue to comply with any license conditions that require it to implement procedures required by §§ 35.610, 35.642, 35.643, and 35.645 until there is a license amendment or renewal that modifies the license condition.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.11 License required.

Sec. 35.11 License required.

§ 35.11 License required.

(a) A person shall not manufacture, produce, acquire, receive, possess, use, or transfer byproduct material for medical use except in accordance with a specific license issued by the Commission or an Agreement State, or as allowed in paragraph (b) or (c) of this section.

(a) A person may not manufacture, produce, acquire, receive, possess, use, or transfer byproduct material for medical use except in accordance with a specific license issued by the Commission or an Agreement State, or as allowed in paragraph (b) or (c) of this section.

(a) A person may manufacture, produce, acquire, receive, possess, use, or transfer byproduct material for medical use only in accordance with a specific license issued by the Commission or an Agreement State, or as allowed in paragraph (b) or (c) of this section.

(b) An individual may receive, possess, use, or transfer byproduct material in accordance with the regulations in this chapter under the supervision of an authorized user as provided in Section 35.25, unless prohibited by license condition.

(b) An individual may receive, possess, use, or transfer byproduct material in accordance with the regulations in this chapter under the supervision of an authorized user as provided in Sec. 35.27, unless prohibited by license condition.

(b) A specific license is not needed for an individual who--

(1) Receives, possesses, uses, or transfers byproduct material in accordance with the regulations in this chapter under the supervision of an authorized user as provided in § 35.27, unless prohibited by license condition; and

(c) An individual may prepare unsealed byproduct material for medical use in

(c) An individual may prepare unsealed byproduct material for medical use in

(2) Prepares unsealed byproduct material for medical use in accordance with the

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

accordance with the regulations in this chapter under the supervision of an authorized nuclear pharmacist or authorized user as provided in Section 35.25, unless prohibited by license condition.

accordance with the regulations in this chapter under the supervision of an authorized nuclear pharmacist or authorized user as provided in Sec. 35.27, unless prohibited by license condition.

regulations in this chapter under the supervision of an authorized nuclear pharmacist or authorized user as provided in § 35.27, unless prohibited by license condition.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.12 Application for license, amendment, or renewal.

Sec. 35.12 Application for license, amendment, or renewal.

§ 35.12 Application for license, amendment, or renewal.

(a) If the application is for medical use sited in a medical institution, only the institution's management may apply. If the application is for medical use not sited in a medical institution, any person may apply.

(a) An application must be signed by the management of the facility.

(a) An application must be signed by the applicant's or licensee's management.

(b) An application for a license for medical use of byproduct material as described in Sections 35.100, 35.200, 35.300, 35.400, and 35.500 of this part must be made by filing an original and one copy of Form NRC-313, "Application for Materials License." For guidance in completing the form, refer to the instructions in the most current versions of the appropriate Regulatory Guides. A request for a license amendment or renewal may be submitted as an original and one copy in letter format.

(b) An application for a license for medical use of byproduct material as described in Secs. 35.100, 35.200, 35.300, 35.400, and 35.500, and for medical use of remote afterloaders in Sec. 35.600, must be made by filing an original and one copy of NRC Form 313, "Application for Material License." A request for a license amendment or renewal may be submitted as an original and one copy in letter format.

(b) An application for a license for medical use of byproduct material as described in §§ 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, and 35.1000 must be made by --

(1) Filing an original and one copy of NRC Form 313, "Application for Material License," that includes the facility diagram, equipment, and training and experience qualifications of the Radiation Safety Officer, authorized user(s), authorized medical physicist(s), and authorized nuclear pharmacist(s); and

(2) Submitting procedures required by §§ 35.610, 35.642, 35.643, and 35.645, as applicable.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(c) An application for a license for medical use of byproduct material as described in Section 35.600 of this part must be made by filing an original and one copy of Form NRC-313. For guidance in completing the form, refer to the instructions in the most current version of the appropriate Regulatory Guide. A request for a license amendment or renewal may be submitted as an original and one copy in letter format.

(d) For copies of regulatory guides, application forms, or to submit an application or an amendment request, refer to Section 30.6 of this chapter.

(c) Except for medical use of remote afterloaders, a separate license application must be filed for each medical use of byproduct material as described in Sec. 35.600 by filing an original and one copy of NRC Form 313. A request for a license amendment or renewal may be submitted as an original and one copy in letter format.

(d) An application for a license for medical use of byproduct material as described in Sec. 35.1000 must be made by filing an original and one copy of NRC Form 313.

(1) In addition to the information required in NRC Form 313, the application must also include information regarding any radiation safety aspects of the medical use of the material that is not addressed in subparts A through C of this part, as well as any specific information necessary for--

- (i) Radiation safety precautions and instructions;
- (ii) Training and experience of proposed users;

(c) A request for a license amendment or renewal must be made by --

- (1) Submitting an original and one copy in letter format; and
- (2) Submitting procedures required by §§ 35.610, 35.642, 35.643, and 35.645, as applicable.

(d) In addition to the requirements in paragraphs (b) and (c) of this section, an application for a license or amendment for medical use of byproduct material as described in § 35.1000 must also include information regarding any radiation safety aspects of the medical use of the material that is not addressed in Subparts A through C of this part.

(1) The applicant shall also provide specific information on --

- (i) Radiation safety precautions and instructions;
- (ii) Methodology for measurement of dosages or doses to be administered to patients or human research subjects; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(iii) Methodology for measurement of dosages or doses to be administered to patients or human research subjects; and

(iv) Calibration, maintenance, and repair of instruments and equipment necessary for radiation safety.

(2) The applicant or licensee shall also provide any other information requested by the Commission in its review of the application.

(e) An applicant that satisfies the requirements specified in Sec. 33.13 may apply for a Type A specific license of broad scope.

(iii) Calibration, maintenance, and repair of instruments and equipment necessary for radiation safety.

(2) The applicant or licensee shall also provide any other information requested by the Commission in its review of the application.

(e) An applicant that satisfies the requirements specified in § 33.13 of this chapter may apply for a Type A specific license of broad scope.

(e) An applicant that satisfies the requirements specified in 10 CFR 33.13 may apply for a Type A specific license of broad scope.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.13 License amendments.

Sec. 35.13 License amendments.

§ 35.13 License amendments.

A licensee shall apply for and must receive a license amendment:

A licensee shall apply for and must receive a license amendment--

A licensee shall apply for and must receive a license amendment --

(a) Before it receives or uses byproduct material for a clinical procedure permitted under this part but not permitted by the license issued pursuant to this part;

(a) Before it receives or uses byproduct material for a type of use that is permitted under this part, but that is not authorized on the licensee's current license issued pursuant to this part;

(a) Before it receives or uses byproduct material for a type of use that is permitted under this part, but that is not authorized on the licensee's current license issued under this part;

(b) Before it permits anyone to work as an authorized user or authorized nuclear pharmacist under the license, except an individual who is:

(b) Before it permits anyone to work as an authorized user, authorized nuclear pharmacist, or authorized medical physicist under the license, except an individual who is--

(b) Before it permits anyone to work as an authorized user, authorized nuclear pharmacist, or authorized medical physicist under the license, except --

(1) An authorized user certified by the organizations specified in paragraph (a) of Section 35.910, Section 35.920, Section 35.930, Section 35.940, Section 35.950, or Section 35.960;

(1) An authorized user who meets the requirements Secs. 35.290(a), 35.292(a), 35.390(a), 35.490(a), 35.590(a), or 35.690(a) and Sec. 35.59, or Secs. 35.910, 35.920, 35.930, 35.932, 35.934, 35.940, 35.941, 35.950, 35.960 and Sec. 35.59;

(1) For an authorized user, an individual who meets the requirements in §§ 35.59 and 35.190(a), 35.290(a), 35.390(a), 35.392(a), 35.394(a), 35.490(a), 35.491, 35.590(a), or 35.690(a);

(2) An authorized nuclear pharmacist certified by the organization specified in paragraph (a) of Section 35.980;

(2) An authorized nuclear pharmacist who meets the requirements in Sec. 35.55(a) and Sec. 35.59; or Secs. 35.980 and 35.59;

(2) For an authorized nuclear pharmacist, an individual who meets the requirements in §§ 35.55(a) and 35.59;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(3) Identified as an authorized user or an authorized nuclear pharmacist on a Commission or Agreement State license that authorizes the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively; or

(4) Identified as an authorized user or an authorized nuclear pharmacist on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively.

(c) Before it changes Radiation Safety Officers or Teletherapy Physicists;

(3) An authorized medical physicist who meets the requirements in Sec. 35.51(a) and Sec. 35.59; or Secs. 35.961 and 35.59;

(4) Identified as an authorized user, an authorized nuclear pharmacist, or authorized medical physicist on a Commission or Agreement State license that authorizes the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively; or

(5) Identified as an authorized user, an authorized nuclear pharmacist, or authorized medical physicist on a permit issued by a Commission or Agreement State specific licensee of broad scope that is authorized to permit the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively.

(c) Before it changes Radiation Safety Officers;

(3) For an authorized medical physicist, an individual who meets the requirements in §§ 35.51(a) and 35.59;

(4) Identified as an authorized user, an authorized nuclear pharmacist, or authorized medical physicist on--

(i) A Commission or Agreement State license or other equivalent permit or license recognized by NRC that authorizes the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively; or

(ii) A permit issued by a Commission or Agreement State specific license of broad scope that is authorized to permit the use of byproduct material in medical use or in the practice of nuclear pharmacy, respectively.

(c) Before it changes Radiation Safety Officers, except as provided in § 35.24(c);

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(d) Before it orders byproduct material in excess of the amount, or radionuclide or form different than authorized on the license; and

(e) Before it adds to or changes the areas of use or address or addresses of use identified in the application or on the license.

(d) Before it orders byproduct material in excess of the amount, or radionuclide or form that is different than the radionuclide or form authorized on the license;

(e) Before it adds to or changes the areas identified in the application or on the license, except for areas where byproduct material is used in accordance with Secs. 35.100 and 35.200; and

(f) Before it changes the address(es) of use identified in the application or on the license.

(d) Before it receives byproduct material in excess of the amount or in a different form, or receives a different radionuclide than is authorized on the license;

(e) Before it adds to or changes the areas identified in the application or on the license, except for areas where byproduct material is used only in accordance with §§ 35.100 and 35.200; and

(f) Before it changes the address(es) of use identified in the application or on the license.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.14 Notifications.

Sec. 35.14 Notifications.

§ 35.14 Notifications.

(a) A licensee shall provide to the Commission a copy of the board certification, the Commission or Agreement State license, or the permit issued by a licensee of broad scope for each individual no later than 30 days after the date that the licensee permits the individual to work as an authorized user or an authorized nuclear pharmacist pursuant to Section 35.13 (b)(1) through (b)(4).

(a) A licensee shall provide to the Commission a copy of the board certification, the Commission or Agreement State license, or the permit issued by a licensee of broad scope for each individual no later than 30 days after the date that the licensee permits the individual to work as an authorized user, an authorized nuclear pharmacist, an authorized medical physicist, pursuant to Sec. 35.13 (b)(1) through (b)(5).

(a) A licensee shall provide the Commission a copy of the board certification, the Commission or Agreement State license, or the permit issued by a licensee of broad scope for each individual no later than 30 days after the date that the licensee permits the individual to work as an authorized user, an authorized nuclear pharmacist, or an authorized medical physicist, under § 35.13 (b)(1) through (b)(4).

(b) A licensee shall notify the Commission by letter no later than 30 days after:

(b) A licensee shall notify the Commission by letter no later than 30 days after:

(b) A licensee shall notify the Commission by letter no later than 30 days after:

(1) An authorized user, an authorized nuclear pharmacist, Radiation Safety Officer, or teletherapy physicist permanently discontinues performance of duties under the license or has a name change; or

(1) An authorized user, an authorized nuclear pharmacist, a Radiation Safety Officer, or an authorized medical physicist permanently discontinues performance of duties under the license or has a name change;

(1) An authorized user, an authorized nuclear pharmacist, a Radiation Safety Officer, or an authorized medical physicist permanently discontinues performance of duties under the license or has a name change;

(2) The licensee's mailing address changes.

(2) The licensee's mailing address changes;

(2) The licensee's mailing address changes;

(3) The licensee's name changes, but the name change does not constitute a

(3) The licensee's name changes, but the name change does not constitute a

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

transfer of control of the license as described in Sec. 30.34(b) of this chapter; or

(4) The licensee has added to or changed the areas where byproduct material is used in accordance with Secs. 35.100 and 35.200.

(c) The licensee shall mail the documents required in this section to the appropriate address identified in Section 30.6 of this chapter.

(c) The licensee shall mail the documents required in this section to the appropriate address identified in Sec. 30.6 of this chapter.

transfer of control of the license as described in § 30.34(b) of this chapter; or

(4) The licensee has deleted or otherwise changed the areas where byproduct material is used in accordance with §§ 35.100 and 35.200.

(c) The licensee shall mail the documents required in this section to the appropriate address identified in § 30.6 of this chapter.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.15 Exemptions regarding Type A specific licenses of broad scope.

Sec. 35.15 Exemptions regarding Type A specific licenses of broad scope.

§ 35.15 Exemptions regarding Type A specific licenses of broad scope.

A licensee possessing a Type A specific license of broad scope for medical use is exempt from the following:

A licensee possessing a Type A specific license of broad scope for medical use is exempt from--

A licensee possessing a Type A specific license of broad scope for medical use, issued under Part 33, is exempt from --

(a) The provisions of Section 35.13(b);

(a) The provisions of Sec. 35.13(b);

(a) The provisions of § 35.12(d) regarding the need to file an amendment to the license for medical uses of byproduct material, as described in § 35.1000;

(b) The provisions of Section 35.13(e) regarding additions to or changes in the areas of use only at the addresses specified in the license;

(b) The provisions of Sec. 35.13(e) regarding additions to or changes in the areas of use only at the addresses specified in the license;

(b) The provisions of § 35.13(b);

(c) The provisions of Section 35.14(a); and

(c) The provisions of Sec. 35.14(a);

(c) The provisions of § 35.13(e) regarding additions to or changes in the areas of use only at the addresses specified in the license;

(d) The provisions of Section 35.14(b)(1) for an authorized user or an authorized nuclear pharmacist.

(d) The provisions of Sec. 35.14(b)(1) for an authorized user, an authorized nuclear pharmacist, or an authorized medical physicist; and

(d) The provisions of § 35.14(a);

(e) The provisions of Sec. 35.49(a).

(e) The provisions of § 35.14(b)(1) for an authorized

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

user, an authorized nuclear
pharmacist, or an authorized
medical physicist;
(f) The provisions of §
35.49(a).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.18 License issuance.

Sec. 35.18 License issuance.

§ 35.18 License issuance.

The Commission shall issue a license for the medical use of byproduct material if:

(a) The Commission shall issue a license for the medical use of byproduct material if--

(a) The Commission shall issue a license for the medical use of byproduct material if --

(a) The applicant has filed Form NRC-313 "Application for Materials License" in accordance with the instructions in Section 35.12;

(1) The applicant has filed Form NRC-313 "Application for Materials License" in accordance with the instructions in Sec. 35.12;

(1) The applicant has filed Form NRC-313 "Application for Material License" in accordance with the instructions in § 35.12;

(b) The applicant has paid any applicable fee as provided in Part 170 of this chapter;

(2) The applicant has paid any applicable fee as provided in part 170 of this chapter;

(2) The applicant has paid any applicable fee as provided in Part 170 of this chapter;

(c) The Commission finds the applicant equipped and committed to observe the safety standards established by the Commission in this Chapter for the protection of the public health and safety; and

(3) The Commission finds the applicant equipped and committed to observe the safety standards established by the Commission in this chapter for the protection of the public health and safety; and

(3) The Commission finds the applicant equipped and committed to observe the safety standards established by the Commission in this Chapter for the protection of the public health and safety; and

(d) The applicant meets the requirements of part 30 of this chapter.

(4) The applicant meets the requirements of part 30 of this chapter.

(4) The applicant meets the requirements of Part 30 of this chapter.

(b) The Commission shall issue a license for mobile services if the applicant:

(b) The Commission shall issue a license for mobile medical services if the applicant:

(1) Meets the requirements in paragraph (a) of this section; and

(1) Meets the requirements in paragraph (a) of this section; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(2) Assures that individuals or human research subjects to whom radiopharmaceuticals or radiation from implants will be administered may be released following treatment in accordance with Sec. 35.75.

(2) Assures that individuals or human research subjects to whom unsealed byproduct material or radiation from implants containing byproduct material will be administered may be released following treatment in accordance with § 35.75.

Section 35.19 Specific exemptions.

Sec. 35.19 Specific exemptions.

§ 35.19 Specific exemptions.

The Commission may, upon application of any interested person or upon its own initiative, grant such exemptions from the regulations in this part as it determines are authorized by law and will not endanger life or property or the common defense and security and are otherwise in the public interest. The Commission will review requests for exemptions from training and experience requirements with the assistance of its Advisory Committee on the Medical Uses of Isotopes.

The Commission may, upon application of any interested person or upon its own initiative, grant such exemptions from the regulations in this part as it determines are authorized by law and will not endanger life or property or the common defense and security and are otherwise in the public interest.

The Commission may, upon application of any interested person or upon its own initiative, grant exemptions from the regulations in this part that it determines are authorized by law and will not endanger life or property or the common defense and security and are otherwise in the public interest.

Subpart B--General
Administrative Requirements

Subpart B--General
Administrative Requirements

Subpart B--General Administrative
Requirements

Section 35.20 ALARA
program.

(a) Each licensee shall develop and implement a written radiation protection program that includes provisions for keeping doses ALARA.

(b) To satisfy the requirement of paragraph (a) of this section:

(1) At a medical institution, management, the Radiation Safety Officer, and all authorized users must participate in the program as requested by the Radiation Safety Committee.

(2) For licensees that are not medical institutions, management and all authorized users must participate in the program as requested by the Radiation Safety Officer.

(c) The program must include notice to workers of the program's existence and workers' responsibility to help keep dose equivalents ALARA, a review of summaries of the types and amounts of byproduct material

used, occupational doses, changes in radiation safety procedures and safety measures, and continuing education and training for all personnel who work with or in the vicinity of byproduct material. The purpose of the review is to ensure that licensees make a reasonable effort to maintain individual and collective occupational doses ALARA.

Sec. 35.24 Authority and responsibilities for the radiation protection program.

(a) In addition to the radiation protection program requirements of Sec. 20.1101 of this chapter, a licensee's management must approve in writing--

(1) Requests for license application, renewal, or amendments before submittal to the Commission;

(2) Any individual before allowing that individual to work as an authorized user, authorized nuclear pharmacist, authorized medical physicist; and

(3) Radiation protection program changes that do not require a license amendment and are permitted under Sec. 35.26;

(b) A licensee with multiple modalities or multiple users shall also develop, implement, and maintain written administrative procedures for interdepartmental/interdisciplinary coordination of the licensee's radiation protection program.

§ 35.24 Authority and responsibilities for the radiation protection program.

(a) In addition to the radiation protection program requirements of § 20.1101 of this chapter, a licensee's management shall approve in writing --

(1) Requests for a license application, renewal, or amendment before submittal to the Commission;

(2) Any individual before allowing that individual to work as an authorized user, authorized nuclear pharmacist, authorized medical physicist; and

(3) Radiation protection program changes that do not require a license amendment and are permitted under § 35.26;

Section 35.21 Radiation Safety Officer.

(a) A licensee shall appoint a Radiation Safety Officer responsible for implementing the radiation safety program. The licensee, through the Radiation Safety Officer, shall ensure that radiation safety activities are being performed in accordance with approved procedures and regulatory requirements in the daily operation of the licensee's byproduct material program.

§ 35.24 Authority and responsibilities for the radiation protection program.

(c) A licensee's management shall appoint a Radiation Safety Officer, who agrees in writing to be responsible for implementing the radiation protection program. The licensee, through the Radiation Safety Officer, shall ensure that radiation safety activities are being performed in accordance with licensee-approved procedures and regulatory requirements in the daily operation of the licensee's radiation protection program.

§ 35.24 Authority and responsibilities for the radiation protection program.

(b) A licensee's management shall appoint a Radiation Safety Officer, who agrees, in writing, to be responsible for implementing the radiation protection program. The licensee, through the Radiation Safety Officer, shall ensure that radiation safety activities are being performed in accordance with licensee-approved procedures and regulatory requirements.

(c) For up to 60 days each year, a licensee may permit an authorized user or an individual qualified to be a Radiation Safety Officer to function as a temporary Radiation Safety Officer and to perform the functions of a Radiation Safety Officer, as provided in paragraph (g) of this section, if the licensee takes the actions required in paragraphs (b), (e), (g), and (h) of this section.

(d) A licensee may simultaneously appoint more than one temporary Radiation Safety Officer in accordance with paragraph (c) of this section, if needed to ensure that the

licensee has a temporary Radiation Safety Officer that satisfies the requirements to be a Radiation Safety Officer for each of the different uses of byproduct material permitted by the license.

(b) The Radiation Safety Officer shall:

(1) Investigate overexposures, accidents, spills, losses, thefts, unauthorized receipts, uses, transfers, disposals, misadministrations, and other deviations from approved radiation safety practice and implement corrective actions as necessary;

(2) Establish, collect in one binder or file, and implement written policy and procedures for:

(i) Authorizing the purchase of byproduct material;

(ii) Receiving and opening packages of byproduct material;

(iii) Storing byproduct material;

(iv) Keeping an inventory record of byproduct material;

(v) Using byproduct material safely;

- (vi) Taking emergency action if control of byproduct material is lost;
 - (vii) Performing periodic radiation surveys;
 - (viii) Performing checks of survey instruments and other safety equipment;
 - (ix) Disposing of byproduct material;
 - (x) Training personnel who work in or frequent areas where byproduct material is used or stored;
 - (xi) Keeping a copy of all records and reports required by the Commission regulations, a copy of these regulations, a copy of each licensing request and license and amendments, and the written policy and procedures required by the regulations.
- (3) Brief management once each year on the byproduct material program;
- (4) Establish personnel exposure investigational levels that, when exceeded, will initiate an investigation by the Radiation Safety Officer of the cause of the exposure;
- (5) Establish personnel exposure investigational levels that, when exceeded, will initiate a prompt investigation by the Radiation Safety Officer

of the cause of the exposure and a consideration of actions that might be taken to reduce the probability of recurrence;

(6) For medical use not at a medical institution, approve or disapprove minor changes in radiation safety procedures that are not potentially important to safety with the advice and consent of management; and

(7) For medical use at a medical institution, assist the Radiation Safety Committee in the performance of its duties.

Section 35.22 Radiation
Safety Committee.

Each medical institution licensee shall establish a Radiation Safety Committee to oversee the use of byproduct material.

(a) Each Committee must meet the following administrative requirements:

(1) Membership must consist of at least three individuals and must include an authorized user of each type of use permitted by the license, the Radiation Safety Officer, a representative of the nursing service, and a representative of management who is neither an authorized user nor a Radiation Safety Officer. Other members may be included as the licensee deems appropriate.

(2) The Committee must meet at least quarterly.

(3) To establish a quorum and to conduct business, at least one-half of the Committee's membership must be present, including the Radiation Safety Officer and the management's representative.

(4) The minutes of each Radiation Safety

(f) Licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F, and H, or two or more types of units under Subpart H shall establish a Radiation Safety Committee to oversee all uses of byproduct material permitted by the license. The Committee must include an authorized user of each type of use permitted by the license, the Radiation Safety Officer, a representative of the nursing service, and a representative of management who is neither an authorized user nor a Radiation Safety Officer. The Committee may include other members the licensee considers appropriate.

Committee meeting must include:

- (i) The date of the meeting;
- (ii) Members present;
- (iii) Members absent;
- (iv) Summary of deliberations and discussions;
- (v) Recommended actions and the numerical results of all ballots; and
- (vi) ALARA program reviews described in Section 35.20(c).

(5) The Committee must promptly provide each member with a copy of the meeting minutes, and retain one copy for the duration of the license.

(b) To oversee the use of licensed material, the Committee must:

(1) Review recommendations on ways to maintain individual and collective doses ALARA;

(2)(i) Review, on the basis of safety and with regard to the training and experience standards in subpart J of this part, and approve or disapprove any individual who is to be listed as an authorized user, an authorized nuclear pharmacist, the Radiation Safety Officer, or a teletherapy

physicist before submitting a license application or request for amendment or renewal; or

(ii) Review, pursuant to Section 35.13 (b)(1) through (b)(4), on the basis of the board certification, the license, or the permit identifying an individual, and approve or disapprove any individual prior to allowing that individual to work as an authorized user or authorized nuclear pharmacist;

(3) Review on the basis of safety, and approve with the advice and consent of the Radiation Safety Officer and the management representative, or disapprove minor changes in radiation safety procedures that are not potentially important to safety and are permitted under Section 35.31 of this part;

(4) Review quarterly, with the assistance of the Radiation Safety Officer, a summary of the occupational radiation dose records of all personnel working with byproduct material;

(5) Review quarterly, with the assistance of the Radiation Safety Officer, all incidents involving byproduct material with respect to cause

and subsequent actions taken;
and

(6) Review annually,
with the assistance of the
Radiation Safety Officer, the
radiation safety program.

Section 35.23 Statements of authority and responsibilities.

(a) A licensee shall provide the Radiation Safety Officer, and at a medical institution the Radiation Safety Committee, sufficient authority, organizational freedom, and management prerogative, to:

- (1) Identify radiation safety problems;
- (2) Initiate, recommend, or provide corrective actions; and
- (3) Verify implementation of corrective actions.

(b) A licensee shall establish and state in writing the authorities, duties, responsibilities, and radiation safety activities of the Radiation Safety Officer, and at a medical institution the Radiation Safety Committee, and retain the current edition of these statements as a record until the Commission terminates the license.

(e) A licensee shall provide the Radiation Safety Officer sufficient authority, organizational freedom, time, resources, and management prerogative, to--

- (1) Identify radiation safety problems;
- (2) Initiate, recommend, or provide corrective actions;
- (3) Stop unsafe operations; and,
- (4) Verify implementation of corrective actions.

(d) A licensee shall establish in writing the authority, duties, and responsibilities of the Radiation Safety Officer.

(g) A licensee shall provide the Radiation Safety Officer sufficient authority, organizational freedom, time, resources, and management prerogative, to --

- (1) Identify radiation safety problems;
- (2) Initiate, recommend, or provide corrective actions;
- (3) Stop unsafe operations; and,
- (4) Verify implementation of corrective actions.

(e) A licensee shall establish the authority, duties, and responsibilities of the Radiation Safety Officer in writing.

(f) A licensee shall retain a record of actions taken pursuant to paragraphs (a), (c), and (d) of this section in accordance with Sec. 35.2024.

(h) A licensee shall retain a record of actions taken under paragraphs (a), (b), and (e) of this section in accordance with § 35.2024.

Sec. 35.26 Radiation protection program changes.

(a) A licensee may revise its radiation protection program without Commission approval if--

(1) The revisions do not require an amendment under Sec. 35.13;

(2) The revisions do not reduce radiation safety;

(3) The revisions have been reviewed and approved by the Radiation Safety Officer and licensee management; and

(4) The affected individuals are instructed on the revised program before the changes are implemented.

(b) A licensee shall retain a record of each change in accordance with Sec. 35.2026.

§ 35.26 Radiation protection program changes.

(a) A licensee may revise its radiation protection program without Commission approval if --

(1) The revision does not require an amendment under § 35.13;

(2) The revision is in compliance with the regulations and the license ;

(3) The revision has been reviewed and approved by the Radiation Safety Officer and licensee management; and

(4) The affected individuals are instructed on the revised program before the changes are implemented.

(b) A licensee shall retain a record of each change in accordance with § 35.2026.

Section 35.25 Supervision.

(a) A licensee that permits the receipt, possession, use, or transfer of byproduct material by an individual under the supervision of an authorized user as allowed by Section 35.11(b) of this part shall:

(1) Instruct the supervised individual in the principles of radiation safety appropriate to that individual's use of byproduct material and in the licensee's written quality management program;

(2) Require the supervised individual to follow the instructions of the supervising authorized user, follow the written radiation safety and quality management procedures established by the licensee, and comply with the regulations of this chapter and the license conditions with respect to the use of byproduct material; and

(3) Periodically review the supervised individual's use of byproduct material and the records kept to reflect this use.

Sec. 35.27 Supervision.

(a) A licensee that permits the receipt, possession, use, or transfer of byproduct material by an individual under the supervision of an authorized user or as allowed by Sec. 35.11(b) shall--

(1) Instruct the supervised individual in the licensee's written radiation protection procedures, written directive procedures, regulations of this chapter, and license conditions with respect to the use of byproduct material; and

(2) Require the supervised individual to follow the instructions of the supervising authorized user for medical uses of radioactive material, written radiation protection procedures established by the licensee, regulations of this chapter; and license conditions with respect to the medical use of byproduct material.

§ 35.27 Supervision.

(a) A licensee that permits the receipt, possession, use, or transfer of byproduct material by an individual under the supervision of an authorized user or as allowed by § 35.11(b)(1) shall --

(1) In addition to the requirements in § 35.12, instruct the supervised individual in the licensee's written radiation protection procedures, written directive procedures, regulations of this chapter, and license conditions with respect to the use of byproduct material; and

(2) Require the supervised individual to follow the instructions of the supervising authorized user for medical uses of byproduct material, written radiation protection procedures established by the licensee, regulations of this chapter, and license conditions with respect to the medical use of byproduct material.

(b) A licensee that permits the preparation of byproduct material for medical use by an individual under the supervision of an authorized nuclear pharmacist or physician who is an authorized user, as allowed by Section 35.11(c), shall:

(1) Instruct the supervised individual in the preparation of byproduct material for medical use and the principles of and procedures for radiation safety and in the licensee's written quality management program, as appropriate to that individual's use of byproduct material;

(2) Require the supervised individual to follow the instructions given pursuant to paragraph (b)(1) of this section and to comply with the regulations of this chapter and license conditions; and

(3) Require the supervising authorized nuclear pharmacist or physician who is

(b) A licensee that permits the preparation of byproduct material for medical use by an individual under the supervision of an authorized nuclear pharmacist or physician who is an authorized user, as allowed by Sec. 35.11(c), shall--

(1) Instruct the supervised individual in the preparation of byproduct material for medical use, as appropriate to that individual's use of byproduct material; and

(2) Require the supervised individual to follow the instructions of the supervising authorized user or authorized nuclear pharmacist regarding the preparation of byproduct material for medical use, the written radiation protection procedures established by the licensee and the regulations of this chapter, and license conditions.

(b) A licensee that permits the preparation of byproduct material for medical use by an individual under the supervision of an authorized nuclear pharmacist or physician who is an authorized user, as allowed by § 35.11(b)(2), shall --

(1) In addition to the requirements in § 35.12, instruct the supervised individual in the preparation of byproduct material for medical use, as appropriate to that individual's involvement with byproduct material; and

(2) Require the supervised individual to follow the instructions of the supervising authorized user or authorized nuclear pharmacist regarding the preparation of byproduct material for medical use, the written radiation protection procedures established by the licensee and the regulations of this chapter, and license conditions.

an authorized user to periodically review the work of the supervised individual as it pertains to preparing byproduct material for medical use and the records kept to reflect that work.

(c) A licensee shall establish, implement, and maintain a policy for all supervised individuals to request clarification, as needed, from--

(1) The authorized user, before initiating or continuing any procedure that requires a written directive, if the supervised individual has any question about what should be done or how it should be done; and

(2) The authorized user or authorized nuclear pharmacist about the instructions and requirements provided to the supervised individual in accordance with paragraphs (a) and (b) of this section.

(c) A licensee that supervises an individual is responsible for the acts and omissions of the supervised individual.

(d) A licensee that permits supervised activities under paragraph (a) and (b) of this section is responsible for the acts and omissions of the supervised individual.

(c) A licensee that permits supervised activities under paragraphs (a) and (b) of this section is responsible for the acts and omissions of the supervised individual.

Section 35.29 Administrative requirements that apply to the provision of mobile nuclear medicine service.

Cross Reference to proposed and final § 35.80

(a) The Commission will license mobile nuclear medicine service only in accordance with subparts D, E and H of this part and Section 31.11 of this chapter.

(b) Mobile nuclear medicine service licensees shall obtain a letter signed by the management of each client for which services are rendered that authorizes use of byproduct material at the client's address of use. The mobile nuclear medicine service licensee shall retain the letter for three years after the last provision of service.

(c) If a mobile nuclear medicine service provides services that the client is also authorized to provide, the client is responsible for assuring that services are conducted in accordance with the regulations in this chapter while the mobile nuclear medicine service is under the client's direction.

(d) A mobile nuclear medicine service may not order byproduct material to be

delivered directly from the manufacturer or distributor to the client's address of use.

Section 35.31 Radiation safety program changes.

Cross Reference proposed and final § 35.26

(a) A licensee may make minor changes in radiation safety procedures that are not potentially important to safety, i.e., ministerial changes, that were described in the application for license, renewal, or amendment except for those changes in Sections 35.13 and 35.606 of this part. Examples of such ministerial changes include: editing of procedures for clarity or conformance with local drafting policy or updating names, telephone numbers, and addresses; adoption of model radiation safety procedures published in NRC Regulatory Guides; replacement of equipment; reassignment of tasks among employees; or assignment of service contracts for services such as personnel dosimetry, radiation safety equipment repair or calibration, waste disposal, and safety surveys. A licensee is responsible for assuring that any change made is in compliance with the requirements of the regulations and the license.

(b) A licensee shall retain a record of each change until the license has been renewed or terminated. The record must include the effective date of the change, a copy of the old and new radiation safety procedures, the reason for the change, a summary of radiation safety matters that were considered before making the change, the signature of the Radiation Safety Officer, and the signatures of the affected authorized users and of management or, in a medical institution, the Radiation Safety Committee's chairman and the management representative.

Section 35.32 Quality management program.

(a) Each applicant or licensee under this part, as applicable, shall establish and maintain a written quality management program to provide high confidence that byproduct material or radiation from byproduct material will be administered as directed by the authorized user. The quality management program must include written policies and procedures to meet the following specific objectives:

(1) That, prior to administration, a written directive 1_/ is prepared for:

1_/ If, because of the patient's condition, a delay in order to provide a written revision to an existing written directive would jeopardize the patient's health, an oral revision to an existing written directive will be acceptable, provided that the oral revision is documented immediately in the patient's record and a revised written directive is signed by the authorized user within 48 hours of the oral revision.

Also, a written revision to an existing written directive may be made for any diagnostic or therapeutic procedure provided that the revision is dated and signed by an authorized user prior to the administration of the radiopharmaceutical dosage, the brachytherapy dose, the gamma stereotactic radiosurgery dose, the teletherapy dose, or the next teletherapy fractional dose.

If, because of the emergent nature of the patient's condition, a delay in order to provide a written directive would jeopardize the patient's health, an oral directive will be acceptable, provided that the information contained in the oral directive is documented immediately in the patient's record and a written directive is prepared within 24 hours of the oral directive.

(i) Any teletherapy radiation dose;

(ii) Any gamma stereotactic radiosurgery radiation dose;

(iii) Any brachytherapy radiation dose;

(iv) Any administration of quantities greater than 30 microcuries of either sodium iodide I-125 or I-131; or

(v) Any therapeutic administration of a radiopharmaceutical, other than sodium iodide I-125 or I-131;

(2) That, prior to each administration, the patient's or human research subject's identity is verified by more than one method as the individual named in the written directive;

(3) That final plans of treatment and related calculations for brachytherapy, teletherapy, and gamma stereotactic radiosurgery are in accordance with the respective written directives;

(4) That each administration is in accordance with the written directive; and

(5) That any unintended deviation from the written directive is identified and evaluated, and appropriate action is taken.

(b) The licensee shall:

(1) Develop

procedures for and conduct a review of the quality management program including, since the last review, an evaluation of:

(i) A representative sample of patient and human research subject administrations,

(ii) All recordable events, and

(iii) All misadministrations to verify compliance with all aspects of the quality management program; these reviews shall be conducted at intervals no greater than 12 months;

(2) Evaluate each of these reviews to determine the effectiveness of the quality management program and, if required, make modifications to meet the objectives of paragraph (a) of this section; and

(3) Retain records of each review, including the evaluations and findings of the review, in an auditable form for three years.

(c) The licensee shall evaluate and respond, within 30 days after discovery of the

recordable event, to each recordable event by:

(1) Assembling the relevant facts including the cause;

(2) Identifying what, if any, corrective action is required to prevent recurrence; and

(3) Retaining a record, in an auditable form, for three years, of the relevant facts and what corrective action, if any, was taken.

(d) The licensee shall retain:

(1) Each written directive; and

(2) A record of each administered radiation dose or radiopharmaceutical dosage where a written directive is required in paragraph (a)(1) above, in an auditable form, for three years after the date of administration.

(e) The licensee may make modifications to the quality management program to increase the program's efficiency provided the program's effectiveness is not decreased. The licensee shall furnish the modification to the appropriate NRC Regional Office within 30 days after the modification has been made.

(f)(1) Each applicant for a new license, as applicable, shall submit to the appropriate NRC Regional Office in accordance with 10 CFR 30.6 a quality management program as part of the application for a license and implement the program upon issuance of the license by the NRC.

(2) Each existing licensee, as applicable, shall submit to the appropriate NRC Regional Office in accordance with 10 CFR 30.6 by January 27, 1992 a written certification that the quality management program has been implemented along with a copy of the program.

Sec. 35.40 Written directives.

(a) A written directive must be prepared, dated, and signed by an authorized user prior to administration of I-131 sodium iodide greater than 1.11 Megabequerels (Mbq) (30 microcuries (μCi)), any therapeutic dosage of a radiopharmaceutical, or any therapeutic dose of radiation from byproduct material.

§ 35.40 Written directives.

(a) A written directive must be dated and signed by an authorized user before the administration of I-131 sodium iodide greater than 1.11 Megabequerels (MBq) (30 microcuries (μCi)), any therapeutic dosage of unsealed byproduct material or any therapeutic dose of radiation from byproduct material.

If, because of the emergent nature of the patient's condition, a delay in order to provide a written directive would jeopardize the patient's health, an oral directive will be acceptable, provided that the information contained in the oral directive is documented immediately in writing in the patient's record a written directive is prepared within 48 hours of the oral directive.

If, because of the patient's condition, a delay in order to

(1) If, because of the emergent nature of the patient's condition, a delay in order to provide a written directive would jeopardize the patient's health, an oral directive is acceptable. The information contained in the oral directive must be documented as soon as possible in writing in the patient's record. A written directive must be prepared within 48 hours of the oral directive.

provide a written revision to an existing written directive would jeopardize the patient's health, an oral revision to an existing written directive will be acceptable, provided that the oral revision is documented immediately in the patient's record and a revised written directive is signed by the authorized user within 48 hours of the oral revision.

Also, a written revision to an existing written directive may be made by any diagnostic or therapeutic procedure provided that the revision is dated and signed by an authorized user prior to the administration of the radiopharmaceutical dosage, the brachytherapy dose, the gamma stereotactic radiosurgery dose, the teletherapy dose, or the next teletherapy fractional dose.

(b) The written directive must contain the patient or human research subject's name and the following:
(1) For any administration of quantities greater than 1.11 MBq (30 μ Ci) of sodium iodide I-131: the dosage;

(b) The written directive must contain the patient or human research subject's name and the following information--
(1) For any administration of quantities greater than 1.11 MBq (30 μ Ci) of sodium iodide I-131: the dosage;

(2) For a therapeutic administration of a radiopharmaceutical other than sodium iodide I-131: the radiopharmaceutical, dosage, and route of administration;

(3) For gamma stereotactic radiosurgery: target coordinates (including gamma angle), collimator size, plug pattern, total dose for the treatment, and the total treatment volume;

(4) For teletherapy: the total dose, dose per fraction, number of fractions, treatment site, and overall treatment period;

(5) For remote afterloading brachytherapy: the radionuclide, treatment site, dose per fraction, number of fractions, and total dose; or

(6) For all other brachytherapy:

(i) Prior to implantation: treatment site, the radionuclide, number of sources and source strengths or dose; and

(ii) After implantation but prior to completion of the procedure: the radionuclide,

(2) For an administration of a therapeutic dosage of unsealed byproduct material other than sodium iodide I-131: the radioactive drug, dosage, and route of administration;

(3) For gamma stereotactic radiosurgery: the total dose, treatment site, and number of target coordinate settings per treatment for each anatomically distinct treatment site;

(4) For teletherapy: the total dose, dose per fraction, number of fractions, and treatment site;

(5) For high dose-rate remote afterloading brachytherapy: the radionuclide, treatment site, dose per fraction, number of fractions, and total dose; or

(6) For all other brachytherapy, including low, medium, and pulsed dose rate remote afterloaders:

(i) Before implantation: treatment site, the radionuclide, and dose; and

(ii) After implantation but before completion of the procedure: the radionuclide, treatment site, number of sources, and total source strength and exposure time (or, equivalently, the total dose).

treatment site, and total source strength and exposure time (or, equivalently, the total dose).

(c) The licensee shall retain the written directive in accordance with Sec. 35.2040.

(c) A written revision to an existing written directive may be made if the revision is dated and signed by an authorized user before the administration of the dosage of unsealed byproduct material, the brachytherapy dose, the gamma stereotactic radiosurgery dose, the teletherapy dose, or the next fractional dose.

(1) If, because of the patient's condition, a delay in order to provide a written revision to an existing written directive would jeopardize the patient's health, an oral revision to an existing written directive is acceptable. The oral revision must be documented as soon as possible in the patient's record. A revised written directive must be signed by the authorized user within 48 hours of the oral revision.

(d) The licensee shall retain the written directive in accordance with § 35.2040.

Sec. 35.41 Procedures for administrations requiring a written directive.

(a) For any administration requiring a written directive, the licensee shall develop, implement, and maintain written procedures to provide high confidence that:

(1) The patient's or human research subject's identity is verified before each administration; and

(2) Each administration is in accordance with the written directive.

(b) The procedures required by paragraph (a) of this section must, at a minimum, address--

(1) Verifying the identity of the patient or human research subject;

(2) Verifying that the specific details of the administration are in accordance with the written directive and treatment plan;

(3) Checking both manual and computer-generated dose calculations; and

(4) Verifying that any computer-generated dose

§ 35.41 Procedures for administrations requiring a written directive.

(a) For any administration requiring a written directive, the licensee shall develop, implement, and maintain written procedures to provide high confidence that:

(1) The patient's or human research subject's identity is verified before each administration; and

(2) Each administration is in accordance with the written directive.

(b) At a minimum, the procedures required by paragraph (a) of this section must address the following items that are applicable to the licensee's use of byproduct material--

(1) Verifying the identity of the patient or human research subject;

(2) Verifying that the administration is in accordance with the treatment plan, if applicable, and the written directive;

(3) Checking both manual and computer-generated dose calculations; and

calculations are correctly transferred into the consoles of therapeutic medical devices authorized by Sec. 35.600.

(4) Verifying that any computer-generated dose calculations are correctly transferred into the consoles of therapeutic medical units authorized by § 35.600.

Section 35.33 Notifications, reports, and records of misadministrations.

(a) For a misadministration:

(1) The licensee shall notify by telephone the NRC Operations Center 2_/ no later than the next calendar day after discovery of the misadministration.

2_/ The commercial telephone number of the NRC Operations Center is (301) 816-5100.

(2) The licensee shall submit a written report to the appropriate NRC Regional Office listed in 10 CFR 30.6 within 15 days after discovery of the misadministration. The written report must include the licensee's name; the prescribing physician's name; a brief description of the event; why the event occurred; the effect on the individual who received the misadministration; what improvements are needed to prevent recurrence; actions taken to prevent recurrence; whether the licensee notified

the individual (or the individual's responsible relative or guardian), and if not, why not; and if there was notification, what information was provided. The report must not contain the individual's name or any other information that could lead to identification of the individual. To meet the requirements of this section, the notification of the individual receiving the misadministration may be made instead to that individual's responsible relative or guardian, when appropriate.

(3) The licensee shall notify the referring physician and also notify the individual receiving the misadministration of the misadministration no later than 24 hours after its discovery, unless the referring physician personally informs the licensee either that he will inform the individual or that, based on medical judgement, telling the individual would be harmful. The licensee is not required to notify the individual without first consulting the referring physician. If the referring physician or the individual receiving the

misadministration cannot be reached within 24 hours, the licensee shall notify the individual as soon as possible thereafter. The licensee may not delay any appropriate medical care for the individual, including any necessary remedial care as a result of the misadministration, because of any delay in notification.

(4) If the individual was notified, the licensee shall also furnish, within 15 days after discovery of the misadministration, a written report to the individual by sending either:

(i) A copy of the report that was submitted to the NRC; or

(ii) A brief description of both the event and the consequences as they may affect the individual, provided a statement is included that the report submitted to the NRC can be obtained from the licensee.

(b) Each licensee shall retain a record of each misadministration for 5 years. The record must contain the names of all individuals involved (including the prescribing physician, allied

health personnel, the individual who received the misadministration, and that individual's referring physician, if applicable), the individual's social security number or other identification number if one has been assigned, a brief description of the misadministration, why it occurred, the effect on the individual, improvements needed to prevent recurrence, and the actions taken to prevent recurrence.

(c) Aside from the notification requirement, nothing in this section affects any rights or duties of licensees and physicians in relation to each other, to individuals receiving misadministrations, or to that individual's responsible relatives or guardians.

Section 35.49 Suppliers for sealed sources or devices for medical use.

A licensee may use for medical use only:

(a) Sealed sources or devices manufactured, labeled, packaged, and distributed in accordance with a license issued pursuant to 10 CFR Part 30 and 10 CFR 32.74 or the equivalent requirements of an Agreement State; or

(b) Teletherapy sources manufactured and distributed in accordance with a license issued pursuant to 10 CFR Part 30 or the equivalent requirements of an Agreement State.

Sec. 35.49 Suppliers for sealed sources or devices for medical use.

A licensee may use for medical use only--

(a) Sealed sources or devices manufactured, labeled, packaged, and distributed in accordance with a license issued pursuant to 10 CFR part 30 and Sec. 32.74 of this chapter or the equivalent requirements of an Agreement State; or

(b) Teletherapy sources manufactured and distributed in accordance with a license issued pursuant to 10 CFR part 30 or the equivalent requirements of an Agreement State.

§ 35.49 Suppliers for sealed sources or devices for medical use.

For medical use, a licensee may only use --

(a) Sealed sources or devices initially manufactured, labeled, packaged, and distributed in accordance with a license issued pursuant to 10 CFR Part 30 and 10 CFR 32.74 of this chapter or the equivalent requirements of an Agreement State; or

(b) Teletherapy sources manufactured and distributed in accordance with a license issued pursuant to 10 CFR Part 30 or the equivalent requirements of an Agreement State.

Sec. 35.50 Training for
Radiation Safety Officer

Except as provided in Sec. 35.57, the licensee shall require an individual fulfilling the responsibilities of the Radiation Safety Officer (RSO) as provided in Sec. 35.24 to be an individual who--

(a) Is certified by a speciality board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission or;

(b)(1) Has completed a structured educational program consisting of both:

(i) 200 hours of didactic training in the following areas--

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Radiation biology; and

(E) Radiation dosimetry; and

(ii) One year of full-time radiation safety experience under the supervision of the

§ 35.50 Training for Radiation
Safety Officer.

Except as provided in § 35.57, the licensee shall require an individual fulfilling the responsibilities of the Radiation Safety Officer as provided in § 35.24 to be an individual who --

(a) Is certified by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed a structured educational program consisting of both:

(i) 200 hours of didactic training in the following areas--

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Radiation biology; and

(E) Radiation dosimetry;

and

individual identified as the RSO on a Commission or Agreement State license that authorizes similar type(s) of use(s) of byproduct material involving the following;

(A) Shipping, receiving, and performing related radiation surveys;

(B) Using and performing checks for proper operation of dose calibrators, survey meters, and instruments used to measure radionuclides;

(C) Securing and controlling byproduct material;

(D) Using administrative controls to avoid mistakes in the administration of byproduct material;

(E) Using procedures to prevent or minimize radioactive

contamination and using proper decontamination procedures; and

(F) Disposing of byproduct material; and

(2) Has obtained written certification, signed by a preceptor RSO, that the requirements in paragraph (b)(1) of this section have been satisfactorily completed and that the individual has

(ii) One year of full-time radiation safety experience under the supervision of the individual identified as the Radiation Safety Officer on a Commission or Agreement State license that authorizes similar type(s) of use(s) of byproduct material involving the following--

(A) Shipping, receiving, and performing related radiation surveys;

(B) Using and performing checks for proper operation of instruments used to determine the activity of dosages, survey meters, and instruments used to measure radionuclides;

(C) Securing and controlling byproduct material;

(D) Using administrative controls to avoid mistakes in the administration of byproduct material;

(E) Using procedures to prevent or minimize radioactive contamination and using proper decontamination procedures; and

(F) Disposing of byproduct material; and

(2) Has obtained written certification, signed by a preceptor Radiation Safety Officer, that the individual has satisfactorily

achieved a level of competency sufficient to independently function as an RSO for medical uses of byproduct material; and

(3) Following completion of the requirements in paragraph (b) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part; or

(c) Is an authorized user, authorized medical physicist, or authorized nuclear pharmacist identified on the licensee's license and has experience with the radiation safety aspects of similar types of use of byproduct material for which the individual has RSO responsibilities.

completed the requirements in paragraph (b)(1) of this section and has achieved a level of radiation safety knowledge sufficient to function independently as a Radiation Safety Officer for a medical use licensee; or

(c) Is an authorized user, authorized medical physicist, or authorized nuclear pharmacist identified on the licensee's license and has experience with the radiation safety aspects of similar types of use of byproduct material for which the individual has Radiation Safety Officer responsibilities.

Sec. 35.51 Training for authorized medical physicist.

The licensee shall require the authorized medical physicist to be an individual who--

(a) Is certified by a speciality board whose certification process includes all of the training and experience requirements in paragraph (b) of this section and whose certification has been approved by the Commission; or

(b)(1) Holds a master's or doctor's degree in physics, biophysics, radiological physics, medical physics, or health physics, or an equivalent training program approved by the NRC, and has completed one year of full-time training in therapeutic radiological physics and an additional year of full-time practical experience under the supervision of a medical physicist at a medical institution that includes the tasks listed in Secs. 35.67, 35.632, 35.633, 35.635, 35.642, 35.643, 35.644, 35.645 and 35.652, as applicable; and

§ 35.51 Training for an authorized medical physicist.

The licensee shall require the authorized medical physicist to be an individual who --

(a) Is certified by a specialty board whose certification process includes all of the training and experience requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Holds a master's or doctor's degree in physics, biophysics, radiological physics, medical physics, or health physics and has completed one year of full-time training in therapeutic radiological physics and an additional year of full-time work experience under the supervision of a medical physicist at a medical institution that includes the tasks listed in §§ 35.67, 35.433, 35.632, 35.633, 35.635, 35.642, 35.643, 35.645, and 35.652, as applicable; and

(2) Has obtained written certification, signed by a preceptor authorized medical physicist, that the requirements in paragraph (b)(1) in this section have been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently function as an authorized medical physicist; and,

(3) Following completion of the requirements in paragraph (b)(1) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part.

(2) Has obtained written certification, signed by a preceptor authorized medical physicist, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized medical physicist.

Sec. 35.55 Training for an authorized nuclear pharmacist.

The licensee shall require the authorized nuclear pharmacist to be a pharmacist who--

(a) Is certified as a nuclear pharmacist by a speciality board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission, or

(b)(1) Has completed 700 hours in a structured educational program consisting of both:

(i) Didactic training in the following areas--

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(ii) Supervised practical experience in a nuclear pharmacy involving--

§ 35.55 Training for an authorized nuclear pharmacist.

The licensee shall require the authorized nuclear pharmacist to be a pharmacist who --

(a) Is certified as a nuclear pharmacist by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed 700 hours in a structured educational program consisting of both:

(i) Didactic training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(A) Shipping, receiving, and performing related radiation surveys;

(B) Using and performing checks for proper operation of dose calibrators, survey meters, and, if appropriate, instruments used to measure alpha- or beta-emitting radionuclides;

(C) Calculating, assaying, and safely preparing dosages for patients or human research subjects;

(D) Using administrative controls to avoid medical events in the administration of byproduct material; and

(E) Using procedures to prevent or minimize radioactive contamination and using proper decontamination procedures; and

(2) Has obtained written certification, signed by a preceptor authorized nuclear pharmacist, that the requirements in paragraph (b)(1) have been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently operate a nuclear pharmacy; and

(ii) Supervised practical experience in a nuclear pharmacy involving --

(A) Shipping, receiving, and performing related radiation surveys;

(B) Using and performing checks for proper operation of instruments used to determine the activity of dosages, survey meters, and, if appropriate, instruments used to measure alpha- or beta-emitting radionuclides;

(C) Calculating, assaying, and safely preparing dosages for patients or human research subjects;

(D) Using administrative controls to avoid medical events in the administration of byproduct material; and

(E) Using procedures to prevent or minimize radioactive contamination and using proper decontamination procedures; and

(2) Has obtained written certification, signed by a preceptor authorized nuclear pharmacist, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to

(3) Following completion of the requirements in paragraph (b)(1) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part.

function independently as an authorized nuclear pharmacist.

Sec. 35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

(a) An individual identified as a Radiation Safety Officer, a teletherapy or medical physicist, or a nuclear pharmacist on a Commission or Agreement State license before [date--6 months from publication of the Final Rule] need not comply with the training requirements of Secs. 35.50 and 35.51, respectively.

(b) Physicians, dentists, or podiatrists identified as authorized users for the medical, dental, or podiatric use of byproduct material on a Commission or Agreement State license issued before [date--6 months from publication of the Final Rule] who perform only those medical uses for which they were authorized on that date need not comply with the training requirements of subparts C through H.

Sec. 35.59 Recentness of training.

§ 35.57 Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist.

(a) An individual identified as a Radiation Safety Officer, a teletherapy or medical physicist, or a nuclear pharmacist on a Commission or Agreement State license before [insert date 6 months from publication of the Final Rule] need not comply with the training requirements of §§ 35.50, 35.51, or 35.55, respectively.

(b) Physicians, dentists, or podiatrists identified as authorized users for the medical use of byproduct material on a Commission or Agreement State license issued before [insert date 6 months from publication of the Final Rule] who perform only those medical uses for which they were authorized on that date need not comply with the training requirements of Subparts D -H.

§ 35.59 Recentness of training.

The training and experience specified in subparts B, D, E, F, G, H, and J must have been obtained within the 7 years preceding the date of application or the individual must have had related continuing education and experience since the required training and experience was completed.

The training and experience specified in Subparts B, D, E, F, G, and H must have been obtained within the 7 years preceding the date of application or the individual must have had related continuing education and experience since the required training and experience was completed.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart C--General Technical Requirements

Subpart C--General Technical Requirements

Subpart C--General Technical Requirements

Section 35.50 Possession, use, calibration, and check of dose calibrators.

Sec. 35.60 Possession, use, calibration, and check of instruments to measure the activity of photon-emitting radionuclides.

§ 35.60 Possession, use, and calibration of instruments to measure the activity of unsealed byproduct material.

(a) A licensee shall possess and use a dose calibrator to measure the activity of dosages of photon-emitting radionuclides prior to administration to each patient or human research subject.

(a) For other than unit dosages, a licensee shall possess and use instrumentation to measure the activity of photon-emitting radionuclides prior to administration to each patient or human research subject.

(a) For direct measurements performed in accordance with § 35.63, a licensee shall possess and use instrumentation to measure the activity of unsealed byproduct material before it is administered to each patient or human research subject.

(b) A licensee shall:

(b) If a licensee uses instrumentation to measure the activity of dosages of photon-emitting radionuclides, including unit dosages, it shall develop, implement, and maintain written procedures for proper operation of the instrumentation. At a minimum, a licensee shall--

(b) A licensee shall calibrate the instrumentation required in paragraph (a) of this section in accordance with nationally recognized standards or the manufacturer's instructions.

(1) Check each dose calibrator for constancy with a dedicated check source at the beginning of each day of use. To satisfy the requirement of this paragraph, the check must be done on a frequently used setting

(1) Perform tests, before initial use and following repair, on each instrument for accuracy, linearity, and geometry dependence;

(2) Perform an accuracy test annually;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

with a sealed source of not less than 10 microcuries of radium-226 or 50 microcuries of any other photon-emitting radionuclide;

(2) Test each dose calibrator for accuracy upon installation and at least annually thereafter by assaying at least two sealed sources containing different radionuclides whose activity the manufacturer has determined within 5 percent of its stated activity, whose activity is at least 10 microcuries for radium-226 and 50 microcuries for any other photon-emitting radionuclide, and at least one of which has a principal photon energy between 100 keV and 500 keV;

(3) Test each dose calibrator for linearity upon installation and at least quarterly thereafter over a range from the highest dosage that will be administered to a patient or human research subject to 1.1 megabecquerels (30 microcuries); and

(4) Test each dose calibrator for geometry dependence upon installation over the range of volumes and volume configurations for which it

PROPOSED RULE

(3) Perform a linearity test annually over the range of medical use; and

(4) Check each instrument for constancy and proper operation at the beginning of each day of use.

(c) Accuracy tests must be performed with source(s) with a principal photon energy of between 100 and 500 keV whose activity is traceable to the National Institute of Standards and Technology (NIST) or by a supplier who has compared the source to a source that was calibrated by NIST.

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

will be used. The licensee shall keep a record of this test for the duration of the use of the dose calibrator.

(c) A licensee shall also perform appropriate checks and tests required by this section following adjustment or repair of the dose calibrator.

(d) A licensee shall mathematically correct dosage readings for any geometry or linearity error that exceeds 10 percent if the dosage is greater than 10 microcuries and shall repair or replace the dose calibrator if the accuracy or constancy error exceeds 10 percent.

(e) A licensee shall retain a record of each check and test required by this section for three years unless directed otherwise. The records required in paragraphs (b)(1) through (b)(4) of this section must include:

(1) For paragraph (b)(1) of this section, the model and serial number of the dose calibrator, the identity of the radionuclide contained in the check source, the date of the check, the activity measured, and the initials of the individual who performed the check;

(d) A licensee shall mathematically correct dosage readings for any geometry or linearity error that exceeds 10 percent if the dosage is greater than 1.11 MBq (30 μ Ci) and shall repair or replace the instrumentation if the accuracy or constancy error exceeds 10 percent.

(e) A licensee shall retain a record of each check and test required by this section in accordance with Sec. 35.2060.

(c) A licensee shall retain a record of each instrument calibration required by this section in accordance with § 35.2060.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(2) For paragraph (b)(2) of this section, the model and serial number of the dose calibrator, the model and serial number of each source used, the identity of the radionuclide contained in the source and its activity, the date of the test, the results of the test, and the identity of the individual performing the test.

(3) For paragraph (b)(3) of this section, the model and serial number of the dose calibrator, the calculated activities, the measured activities, the date of the test, and the identity of the individual performing the test.

(4) For paragraph (b)(4) of this section, the model and serial number of the dose calibrator, the configuration of the source measured, the activity measured for each volume measured, the date of the test, and the identity of the individual performing the test.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.51 Calibration and check of survey instruments.

Sec. 35.61 Calibration and check of survey instruments.

§ 35.61 Calibration of survey instruments.

(a) A licensee shall calibrate the survey instruments used to show compliance with this part before first use, annually, and following repair. The licensee shall:

(a) A licensee shall calibrate the survey instruments used to show compliance with this part and 10 CFR part 20 before first use, annually, and following repair. A licensee shall--

(a) A licensee shall calibrate the survey instruments used to show compliance with this part and 10 CFR Part 20 before first use, annually, and following a repair that affects the calibration. A licensee shall --

(1) Calibrate all scales with readings up to 1000 millirem per hour with a radiation source;

(1) Calibrate all scales with readings up to 10 mSv (1000 mrem) per hour with a radiation source;

(1) Calibrate all scales with readings up to 10 mSv (1000 mrem) per hour with a radiation source;

(2) Calibrate two separated readings on each scale that must be calibrated; and

(2) Calibrate two separated readings on each scale that will be used to show compliance with this part; and

(2) Calibrate two separated readings on each scale or decade that will be used to show compliance; and

(3) Conspicuously note on the instrument the apparent exposure rate from a dedicated check source as determined at the time of calibration, and the date of calibration.

(3) Conspicuously note on the instrument the date of calibration.

(3) Conspicuously note on the instrument the date of calibration.

(b) When calibrating a survey instrument, the licensee shall consider a point as calibrated if the indicated exposure rate differs from the calculated exposure rate by not more than 20 percent, and shall conspicuously attach a correction chart or graph to the instrument.

(b) A licensee shall consider a point as calibrated if the indicated exposure rate differs from the calculated exposure rate by not more than 20 percent, and conspicuously attach a correction chart or graph to the instrument if the indicated exposure rate differs from the calculated exposure rate by more than 10 percent.

(b) A licensee may not use survey instruments if the difference between the indicated exposure rate and the calculated exposure rate is more than 20 percent.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(c) A licensee shall check each survey instrument for proper operation with the dedicated check source each day of use. A licensee is not required to keep records of these checks.

(d) A licensee shall retain a record of each survey instrument calibration for three years. The record must include:

- (1) A description of the calibration procedure; and
- (2) The date of the calibration, a description of the source used and the certified exposure rates from the source, and the rates indicated by the instrument being calibrated, the correction factors deduced from the calibration data, and the signature of the individual who performed the calibration.

PROPOSED RULE

(c) Survey instruments must be removed from use if the indicated exposure rate differs from the calculated exposure rate by more than 20 percent.

(d) A licensee shall retain a record of each survey instrument calibration in accordance with Sec. 35.2061.

FINAL RULE

(c) A licensee shall retain a record of each survey instrument calibration in accordance with § 35.2061.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.52 Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides.

(a) This section does not apply to unit dosages of alpha- or beta- emitting radionuclides that are obtained from a manufacturer or preparer licensed pursuant to 10 CFR 32.72 or equivalent Agreement State requirements.

(b) For other than unit dosages obtained pursuant to paragraph (a) of this section, a licensee shall possess and use instrumentation to measure the

Sec. 35.62 Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides.

(a) For other than unit dosages, a licensee shall possess and use instrumentation to measure the radioactivity of alpha- or beta-emitting radionuclides. A licensee shall measure, by direct measurement or by combination of measurements and calculations, the amount of radioactivity in dosages of alpha-or beta-emitting radionuclides prior to administration to each patient or human research subject.

(b) A licensee shall develop, implement, and maintain written procedures for use of the instrumentation. At a minimum, a licensee shall--

(1) Perform tests before initial use, and following repair, on each instrument for accuracy, linearity, and geometry dependence, unless it is not appropriate for the use of the instrument; and make adjustments when necessary;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

radioactivity of alpha- or beta-emitting radionuclides. The licensee shall have procedures for use of the instrumentation. The licensee shall measure, by direct measurement or by combination of measurements and calculations, the amount of radioactivity in dosages of alpha- or beta-emitting radionuclides prior to administration to each patient or human research subject. In addition, the licensee shall:

(1) Perform tests before initial use, periodically, and following repair, on each instrument for accuracy, linearity, and geometry dependence, as appropriate for the use of the instrument; and make adjustments when necessary; and

(2) Check each instrument for constancy and proper operation at the beginning of each day of use.

(2) Perform accuracy annually;
(3) Perform linearity tests annually over the range of medical use; and
(4) Check each instrument for constancy and proper operation at the beginning of each day of use.
(c) Accuracy tests must be performed with source(s) that are traceable to NIST or by a supplier who has compared the source to a source that was calibrated by NIST.
(d) A licensee shall retain a record of each check and test required by this section in accordance with Sec. 35.2060.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.53 Measurement of dosages of unsealed byproduct material for medical use.

Sec. 35.63 Determination of dosages of unsealed byproduct material for medical use.

§ 35.63 Determination of dosages of unsealed byproduct material for medical use.

A licensee shall:

(a) A licensee shall

(a) A licensee shall

(a) Measure the activity of each dosage of a photon-emitting radionuclide prior to medical use.

determine and record the activity of each dosage prior to medical use.

determine and record the activity of each dosage before medical use.

(b) Measure, by direct measurement or by combination of measurements and calculations, the activity of each dosage of an alpha- or a beta-emitting radionuclide prior to medical use, except for unit dosages obtained from a manufacturer or preparer licensed pursuant to 10 CFR 32.72 or equivalent Agreement State requirements;

(b) For a unit dosage of an alpha-, beta-, or photon-emitting radionuclide, this determination must be made either by direct measurement or by a decay correction, based on the measurement made by a manufacturer or preparer licensed pursuant to Sec. 32.72 of this chapter or equivalent Agreement State requirements.

(b) For a unit dosage, this determination must be made by--

(1) A decay correction, based on the activity or activity concentration determined by --

(c) Retain a record of the measurements required by this section for three years. To satisfy this requirement, the record must contain the:

(c) For a dosage of a alpha-, beta-, or photon-emitting radionuclide prepared by the licensee, this determination must be made

(i) A manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or
(ii) An NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an Investigational New Drug (IND) protocol accepted by FDA.

(1) Generic name, trade name, or abbreviation of the radiopharmaceutical, its lot number, and expiration dates and the radionuclide;

(c) For other than unit dosages, this determination must be made by--

(2) Patient's or human research subject's name, and

(1) Direct measurement of radioactivity;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

identification number if one has been assigned;

(3) Prescribed dosage and activity of the dosage at the time of measurement, or a notation that the total activity is less than 1.1 megabecquerels (30 microcuries);

(4) Date and time of the measurement; and

(5) Initials of the individual who made the record.

by direct measurement or by combination of measurements and calculations.

(d) Unless otherwise directed by the authorized user, a licensee shall not use a dosage if the dosage does not fall within the prescribed dosage range or if the dosage differs from the prescribed dosage by more than 20 percent.

(e) A licensee shall retain a record of the dosage determination required by this section in accordance with Sec. 35.2063.

(2) Combination of direct measurement of radioactivity and mathematical calculations; or

(3) Combination of volumetric measurements and mathematical calculations, based on the measurement made by a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements.

(e) A licensee shall retain a record of the dosage determination required by this section in accordance with § 35.2063.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.57 Authorization for calibration and reference sources.

Any person authorized by Section 35.11 of this part for medical use of byproduct material may receive, possess, and use the following byproduct material for check, calibration, and reference use:

(a) Sealed sources manufactured and distributed by a person licensed pursuant to Section 32.74 of this chapter or equivalent Agreement State regulations and that do not exceed 15 millicuries each;

(b) Any byproduct material listed in Sections 35.100 or 35.200 with a half-life not longer than 100 days in individual amounts not to exceed 15 millicuries;

(c) Any byproduct material listed in Sections 35.100 or 35.200 with a half-life longer than 100 days in individual amounts not to exceed 200 microcuries each; and

Sec. 35.65 Authorization for calibration and reference sources.

Any person authorized by Sec. 35.11 for medical use of byproduct material may receive, possess, and use the following byproduct material for check, calibration, and reference use:

(a) Sealed sources manufactured and distributed by a person licensed pursuant to Sec. 32.74 of this chapter or equivalent Agreement State regulations and that do not exceed 1.11 kBq (30 mCi) each;

(b) Any byproduct material with a half-life not longer than 120 days in individual amounts not to exceed 0.555 MBq (15 mCi);

(c) Any byproduct material with a half-life longer than 120 days in individual amounts not to exceed 7.4 MBq (200 <greek-l>Ci) each and not to exceed 1000 times the quantities in appendix B of Part 30 of this chapter

§ 35.65 Authorization for calibration, transmission, and reference sources.

Any person authorized by § 35.11 for medical use of byproduct material may receive, possess, and use any of the following byproduct material for check, calibration, transmission, and reference use.

(a) Sealed sources manufactured and distributed by a person licensed under §§ 32.72 and 32.74 of this chapter or equivalent Agreement State regulations and that do not exceed 1.11GBq (30 mCi) each.

(b) Any byproduct material with a half-life not longer than 120 days in individual amounts not to exceed 0.555 GBq (15 mCi).

(c) Any byproduct material with a half-life longer than 120 days in individual amounts not to exceed the smaller of 7.4 MBq (200 μCi) or 1000 times the quantities in Appendix B of Part 30 of this chapter.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(d) Technetium-99m in individual amounts not to exceed 50 millicuries.

whichever is more limiting;
and
(d) Technetium-99m in amounts as needed.

(d) Technetium-99m in amounts as needed.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.59 Requirements for possession of sealed sources and brachytherapy sources.

(a) A licensee in possession of any sealed source or brachytherapy source shall follow the radiation safety and handling instructions supplied by the manufacturer, and shall maintain the instructions for the duration of source use in a legible form convenient to users.

(b) A licensee in possession of a sealed source shall:

(1) Test the source for leakage before its first use unless the licensee has a certificate from the supplier indicating that the source was tested within six months before transfer to the licensee; and

(2) Test the source for leakage at intervals not to exceed six months or at other intervals approved by the Commission or an Agreement State and described in the label or brochure that accompanies the source.

Sec. 35.67 Requirements for possession of sealed sources and brachytherapy sources.

(a) A licensee in possession of any sealed source or brachytherapy source shall follow the radiation safety and handling instructions supplied by the manufacturer, and shall maintain the instructions for the duration of source use in a legible form convenient to users.

(b) A licensee in possession of a sealed source shall--

(1) Test the source for leakage before its first use unless the licensee has a certificate from the supplier indicating that the source was tested within 6 months before transfer to the licensee; and

(2) Test the source for leakage at intervals not to exceed 6 months or at other intervals approved by the Commission or an Agreement State in the Sealed Source and Device Registry.

§ 35.67 Requirements for possession of sealed sources and brachytherapy sources.

(a) A licensee in possession of any sealed source or brachytherapy source shall follow the radiation safety and handling instructions supplied by the manufacturer.

(b) A licensee in possession of a sealed source shall --

(1) Test the source for leakage before its first use unless the licensee has a certificate from the supplier indicating that the source was tested within 6 months before transfer to the licensee; and

(2) Test the source for leakage at intervals not to exceed 6 months or at other intervals approved by the Commission or an Agreement State in the Sealed Source and Device Registry.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(c) To satisfy the leak test requirements of this section, the licensee must:

(1) Take a wipe sample from the sealed source or from the surfaces of the device in which the sealed source is mounted or stored on which radioactive contamination might be expected to accumulate or wash the source in a small volume of detergent solution and treat the entire volume as the sample;

(2) Take teletherapy and other device source test samples when the source is in the "off" position; and

(3) Measure the sample so that the leakage test can detect the presence of 0.005 microcuries of radioactive material on the sample.

(d) A licensee shall retain leakage test records for five years. The records must contain the model number, and serial number if assigned, of each source tested, the identity of each source radionuclide and its estimated activity, the measured activity of each test sample expressed in microcuries, a description of the method used to measure

(c) To satisfy the leak test requirements of this section, the licensee shall measure the sample so that the leakage test can detect the presence of 185 Bq (0.005 μ Ci) of radioactive material on the sample.

(d) A licensee shall retain leakage test records in accordance with Sec. 35.2067.

(c) To satisfy the leak test requirements of this section, the licensee shall measure the sample so that the leak test can detect the presence of 185 Bq (0.005 μ Ci) of radioactive material in the sample.

(d) A licensee shall retain leak test records in accordance with § 35.2067.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

each test sample, the date of the test, and the signature of the Radiation Safety Officer.

(e) If the leakage test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall:

(1) Immediately withdraw the sealed source from use and store it in accordance with the requirements in parts 20 and 30 of this chapter; and

(2) File a report within five days of the leakage test with the appropriate NRC Office listed in Section 30.6 of this chapter, with a copy to Director, Office of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555, describing the equipment involved, the test results, and the action taken.

(f) A licensee need not perform a leakage test on the following sources:

(1) Sources containing only byproduct material with a half-life of less than 30 days;

(e) If the leakage test reveals the presence of 185 Bq (0.005 μ Ci) or more of removable contamination, the licensee shall--

(1) Immediately withdraw the sealed source from use and store, dispose, or cause it to be repaired in accordance with the requirements in parts 20 and 30 of this chapter; and

(2) File a report within 5 days of the leakage test in accordance with Sec. 35.3067.

(f) A licensee need not perform a leakage test on the following sources:

(1) Sources containing only byproduct material with a half-life of less than 30 days;

(e) If the leak test reveals the presence of 185 Bq (0.005 μ Ci) or more of removable contamination, the licensee shall --

(1) Immediately withdraw the sealed source from use and store, dispose, or cause it to be repaired in accordance with the requirements in parts 20 and 30 of this chapter; and

(2) File a report within 5 days of the leak test in accordance with § 35.3067.

(f) A licensee need not perform a leak test on the following sources:

(1) Sources containing only byproduct material with a half-life of less than 30 days;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(2) Sources containing only byproduct material as a gas;

(3) Sources containing 100 microcuries or less of beta or gamma-emitting material or 10 microcuries or less of alpha-emitting material;

(4) Sources stored and not being used. The licensee shall, however, test each such source for leakage before any use or transfer unless it has been leakage-tested within six months before the date of use or transfer; and

(5) Seeds of iridium-192 encased in nylon ribbon.

(g) A licensee in possession of a sealed source or brachytherapy source shall conduct a quarterly physical inventory of all such sources in its possession. The licensee

(2) Sources containing only byproduct material as a gas;

(3) Sources containing 3.7 MBq (100 μ Ci) or less of beta or gamma-emitting material or 0.37 MBq (10 μ Ci) or less of alpha-emitting material;

(4) Sources stored for less than a 10-year period and not being used. The licensee shall, however, test each such source for leakage before any use or transfer unless it has been leakage-tested within 6 months before the date of use or transfer; and

(5) Seeds of iridium-192 encased in nylon ribbon.

(g) A licensee in possession of sealed sources or brachytherapy sources, except for gamma stereotactic radiosurgery sources, shall conduct a semi-annual

(2) Sources containing only byproduct material as a gas;

(3) Sources containing 3.7 MBq (100 μ Ci) or less of beta or gamma-emitting material or 0.37 MBq (10 μ Ci) or less of alpha-emitting material;

(4) Seeds of iridium-192 encased in nylon ribbon; and

(5) Sources stored and not being used. However, the licensee shall test each such source for leakage before any use or transfer unless it has been leak tested within 6 months before the date of use or transfer.

(g) A licensee in possession of sealed sources or brachytherapy sources, except for gamma stereotactic radiosurgery sources, shall conduct a semi-annual

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

shall retain each inventory record for five years. The inventory records must contain the model number of each source, and serial number if one has been assigned, the identity of each source radionuclide and its nominal activity, the location of each source, and the signature of the Radiation Safety Officer.

(h) A licensee in possession of a sealed source or brachytherapy source shall measure the ambient dose rates quarterly in all areas where such sources are stored. This does not apply to teletherapy sources in teletherapy units or sealed sources in diagnostic devices.

(i) A licensee shall retain a record of each survey required in paragraph (h) of this section for three years. The record must include the date of the survey, a plan of each area that was surveyed, the measured dose rate at several points in each area expressed in millirem per hour, the survey instrument used, and the signature of the Radiation Safety Officer.

physical inventory of all such sources in its possession. The licensee shall retain each inventory record in accordance with Sec. 35.2067.

physical inventory of all such sources in its possession. The licensee shall retain each inventory record in accordance with § 35.2067.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.60 Syringe shields and labels.

(a) A licensee shall keep syringes that contain byproduct material to be administered in a radiation shield.

(b) To identify its contents, a licensee shall conspicuously label each syringe or syringe radiation shield that contains a syringe with a radiopharmaceutical. The label must show the radiopharmaceutical name or its abbreviation, the clinical procedure to be performed, or the patient's or the human research subject's name.

(c) A licensee shall require each individual who prepares a radiopharmaceutical kit to use a syringe radiation shield when preparing the kit and shall require each individual to use a syringe radiation shield when administering a radiopharmaceutical by injection unless the use of the shield is contraindicated for that patient or human research subject.

Sec. 35.69 Labeling and shielding of vials and syringes.

(a) A licensee shall develop, implement, and maintain written procedures for--

(1) Labeling each syringe, syringe shield, or vial shield that contains a radiopharmaceutical to identify the radiopharmaceutical name, or its abbreviation, and to ensure that the contents are conspicuously identified as containing radioactive material; and

(2) Shielding vials and syringes containing radiopharmaceuticals.

(b) A licensee shall instruct individuals, commensurate with the individual's assigned duties, in the procedures required by paragraph (a) of this section.

§ 35.69 Labeling of vials and syringes.

Each syringe and vial that contains unsealed byproduct material must be labeled to identify the radioactive drug. Each syringe shield and vial shield must also be labeled unless the label on the syringe or vial is visible when shielded.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.61 Vial shields and labels.

(a) A licensee shall require each individual preparing or handling a vial that contains a radiopharmaceutical to keep the vial in a vial radiation shield.

(b) To identify its contents, a licensee shall conspicuously label each vial radiation shield that contains a vial of a radiopharmaceutical. The label must show the radiopharmaceutical name or its abbreviation.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.70 Surveys for contamination and ambient radiation exposure rate.

(a) A licensee shall survey with a radiation detection survey instrument at the end of each day of use all areas where radiopharmaceuticals are routinely prepared for use or administered.

(b) A licensee shall survey with a radiation detection survey instrument at least once each week all areas where radiopharmaceuticals or radiopharmaceutical waste is stored.

(c) A licensee shall conduct the surveys required by paragraphs (a) and (b) of this section so as to be able to detect dose rates as low as 0.1 millirem per hour.

(d) A licensee shall establish radiation dose rate trigger levels for the surveys required by paragraphs (a) and (b) of this section. A licensee shall require that the individual performing the survey immediately notify the Radiation Safety Officer if a

Sec. 35.70 Surveys for ambient radiation exposure rate.

(a) Except as provided in paragraph (b) of this section, a licensee shall survey with a radiation detection survey instrument at the end of each day of use all areas where radiopharmaceuticals requiring a written directive were prepared for use or administered.

(b) A licensee does not need to perform the surveys required by paragraph (a) of this section in an area(s) where patients or human research subjects can not be released pursuant to Sec. 35.75.

(c) A licensee shall retain a record of each survey in accordance with Sec. 35.2070.

§ 35.70 Surveys of ambient radiation exposure rate.

(a) In addition to the surveys required by Part 20 of this chapter, a licensee shall survey with a radiation detection survey instrument at the end of each day of use. A licensee shall survey all areas where unsealed byproduct material requiring a written directive was prepared for use or administered.

(b) A licensee does not need to perform the surveys required by paragraph (a) of this section in an area(s) where patients or human research subjects are confined when they cannot be released under § 35.75.

(c) A licensee shall retain a record of each survey in accordance with § 35.2070.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

dose rate exceeds a trigger level.

(e) A licensee shall survey for removable contamination once each week all areas where radiopharmaceuticals are routinely prepared for use, administered, or stored.

(f) A licensee shall conduct the surveys required by paragraph (e) of this section so as to be able to detect contamination on each wipe sample of 2000 disintegrations per minute.

(g) A licensee shall establish removable contamination trigger levels for the surveys required by paragraph (e) of this section. A licensee shall require that the individual performing the survey immediately notify the Radiation Safety Officer if contamination exceeds the trigger level.

(h) A licensee shall retain a record of each survey for three years. The record must include the date of the survey, a plan of each area surveyed, the trigger level established for each area, the detected dose rate at several

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

points in each area expressed in millirem per hour or the removable contamination in each area expressed in disintegrations per minute per 100 square centimeters, the instrument used to make the survey or analyze the samples, and the initials of the individual who performed the survey.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.75 Release of individuals containing radiopharmaceuticals or Implants.

(a) The licensee may authorize the release from its control of any individual who has been administered radiopharmaceuticals or permanent implants containing radioactive material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 millisieverts (0.5 rem).^{1/}

^{1/} Regulatory Guide 8.39, "Release of Patients Administered Radioactive Materials," describes methods for calculating doses to other individuals and contains tables of activities not likely to cause doses exceeding 5 millisieverts (0.5 rem).

(b) The licensee shall provide the released individual with instructions, including written instructions, on actions recommended to maintain doses to other individuals as

Sec. 35.75 Release of individuals containing radiopharmaceuticals or implants.

(a) A licensee may authorize the release from its control of any individual who has been administered radiopharmaceuticals or implants containing radioactive material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (0.5 rem).^{2\}

^{2\} Regulatory Guide 8.39, "Release of Patients Administered Radioactive Materials," describes methods for calculating doses to other individuals and contains tables of activities not likely to cause doses exceeding 0.5 rem (5 mSv).

(b) A licensee shall provide the released individual, or the individual's parent or guardian, with instructions, including written instructions, on actions

§ 35.75 Release of individuals containing unsealed byproduct material or implants containing byproduct material.

(a) A licensee may authorize the release from its control of any individual who has been administered unsealed byproduct material or implants containing byproduct material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (0.5 rem).¹

(b) A licensee shall provide the released individual, or the individual's parent or guardian, with instructions, including written instructions, on actions

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

low as is reasonably achievable if the total effective dose equivalent to any other individual is likely to exceed 1 millisievert (0.1 rem). If the dose to a breast-feeding infant or child could exceed 1 millisievert (0.1 rem) assuming there were no interruption of breast-feeding, the instructions shall also include:

(1) Guidance on the interruption or discontinuation of breast-feeding and

(2) Information on the consequences of failure to follow the guidance.

(c) The licensee shall maintain a record of the basis for authorizing the release of an individual, for 3 years after the date of release, if the total effective dose equivalent is calculated by:

(1) Using the retained activity rather than the activity administered,

(2) Using an occupancy factor less than 0.25 at 1 meter,

(3) Using the biological or effective half-life, or

recommended to maintain doses to other individuals as low as is reasonably achievable if the total effective dose equivalent to any other individual is likely to exceed 1 mSv (0.1 rem). If the total effective dose equivalent to a breast-feeding infant or child could exceed 1 mSv (0.1 rem) assuming there were no interruption of breast-feeding, the instructions shall also include--

(1) Guidance on the interruption or discontinuation of breast-feeding; and

(2) Information on the potential consequences, if any, of failure to follow the guidance.

(c) A licensee shall maintain a record of the basis for authorizing the release of an individual, in accordance with Sec. 35.2075(a).

(d) The licensee shall maintain a record of instructions provided to breast-feeding women in accordance with Sec. 35.2075(c).

recommended to maintain doses to other individuals as low as is reasonably achievable if the total effective dose equivalent to any other individual is likely to exceed 1 mSv (0.1 rem). If the total effective dose equivalent to a nursing infant or child could exceed 1 mSv (0.1 rem) assuming there were no interruption of breast-feeding, the instructions must also include --

(1) Guidance on the interruption or discontinuation of breast-feeding; and

(2) Information on the potential consequences, if any, of failure to follow the guidance.

(c) A licensee shall maintain a record of the basis for authorizing the release of an individual in accordance with § 35.2075(a).

(d) The licensee shall maintain a record of instructions provided to a breast-feeding female in accordance with § 35.2075(b).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(4) Considering the shielding by tissue.

(d) The licensee shall maintain a record, for 3 years after the date of release, that instructions were provided to a breast-feeding woman if the radiation dose to the infant or child from continued breast-feeding could result in a total effective dose equivalent exceeding 5 millisieverts (0.5 rem).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.80 Technical requirements that apply to the provision of mobile nuclear medicine service.

Sec. 35.80 Provision of mobile service.

§ 35.80 Provision of mobile medical service.

A licensee providing mobile nuclear medicine service shall:

(a) A licensee providing mobile service shall--

(a) A licensee providing mobile medical service shall --

(a) Transport to each address of use only syringes or vials containing prepared radiopharmaceuticals or radiopharmaceutical kits;

(1) Obtain a letter signed by the management of each client for which services are rendered that permits the use of byproduct material at the client's address of use and clearly delineates the authority and responsibility of each entity;

(1) Obtain a letter signed by the management of each client for which services are rendered that permits the use of byproduct material at the client's address and clearly delineates the authority and responsibility of the licensee and the client;

(b) Bring into each address of use all byproduct material to be used and, before leaving, remove all unused byproduct material and all associated waste;

(2) Check instruments as described in Secs. 35.60 and 35.62 for proper function before medical use at each address of use or on each day of use, whichever is more frequent;

(2) Check instruments used to measure the activity of unsealed byproduct materials for proper function before medical use at each client's address or on each day of use, whichever is more frequent. At a minimum, the check for proper function required by this paragraph must include a constancy check;

(c) Secure or keep under constant surveillance and immediate control all byproduct material when in transit or at an address of use;

(d) Check survey instruments and dose calibrators as described in Sections 35.50 and 35.51, and check all other transported equipment for proper function before medical use at each address of use;

(3) Check survey instruments for proper operation with a dedicated check source before use at each address of use;

(3) Check survey instruments for proper operation with a dedicated check source before use at each client's address; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(e) Carry a radiation detection survey meter in each vehicle that is being used to transport byproduct material, and, before leaving a client address of use, survey all radiopharmaceutical areas of use with a radiation detection survey meter to ensure that all radiopharmaceuticals and all associated waste have been removed;

(f) Retain a record of each survey required in paragraph (e) of this section for three years. The record must include the date of the survey, a plan of each area that was surveyed, the measured dose rate at several points in each area of use expressed in millirem per hour, the instrument used to make the survey, and the initials of the individual who performed the survey.

(4) Before leaving a client's address of use, survey all areas of use to ensure compliance with the requirements in part 20 of this chapter; and

(b) A mobile nuclear medicine service may not have byproduct material delivered from the manufacturer or the distributor to the client's address of use, unless the client has a license allowing possession of the byproduct material. Radioactive material delivered to the client's address of use must be received and handled in conformance with the client's license.

(c) A licensee providing mobile nuclear services shall retain the letter required in paragraph (a)(1) and the record of each survey required in paragraph (a)(4) of this section in accordance with Sec. 35.2080.

(4) Before leaving a client's address, survey all areas of use to ensure compliance with the requirements in Part 20 of this chapter.

(b) A mobile medical service may not have byproduct material delivered from the manufacturer or the distributor to the client unless the client has a license allowing possession of the byproduct material. Byproduct material delivered to the client must be received and handled in conformance with the client's license.

(c) A licensee providing mobile medical services shall retain the letter required in paragraph (a)(1) and the record of each survey required in paragraph (a)(4) of this section in accordance with § 35.2080.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.90 Storage of
volatiles and gases.

A licensee shall store
volatile radiopharmaceuticals
and radioactive gases in the
shipper's radiation shield and
container. A licensee shall
store a multi-dose container in
a fume hood after drawing the
first dosage from it.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.92
Decay-in-storage.

Sec. 35.92 Decay-in-storage.

§ 35.92 Decay-in-storage.

(a) A licensee may hold byproduct material with a physical half-life of less than 65 days for decay-in-storage before disposal in ordinary trash and is exempt from the requirements of Section 20.2001 of this chapter if it:

(a) A licensee may hold byproduct material with a physical half-life of less than 120 days for decay-in-storage before disposal in ordinary trash if it--

(a) A licensee may hold byproduct material with a physical half-life of less than 120 days for decay-in-storage before disposal without regard to its radioactivity if it --

(1) Holds byproduct material for decay a minimum of ten half-lives;

(1) Monitors byproduct material at the surface before disposal as ordinary trash and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding;

(1) Monitors byproduct material at the surface before disposal and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding; and

(3) Removes or obliterates all radiation labels; and

(2) Removes or obliterates all radiation labels;

(2) Removes or obliterates all radiation labels except for material that will be handled as biomedical waste after it has been released from the licensee.

(4) Separates and monitors each generator column individually with all radiation shielding removed to ensure that it has decayed to

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

background radiation level before disposal.

(b) A licensee shall retain a record of each disposal permitted under paragraph (a) of this section for three years. The record must include the date of the disposal, the date on which the byproduct material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.

(b) A licensee shall retain a record of each disposal permitted under paragraph (a) of this section in accordance with Sec. 35.2092.

(b) A licensee shall retain a record of each disposal permitted under paragraph (a) of this section in accordance with § 35.2092.

1. NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Licenses," describes methods for calculating doses to other individuals and contains tables of activities not likely to cause doses exceeding 0.5 rem (5 mSv).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart D--Uptake, Dilution, and Excretion

Subpart D--Unsealed Byproduct Material--Low Dose

Subpart D--Unsealed Byproduct Material - Written Directive Not Required

Section 35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies.

Sec. 35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

§ 35.100 Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required.

A licensee may use for uptake, dilution, or excretion studies any unsealed byproduct material prepared for medical use that is either:

A licensee may use for uptake, dilution, or excretion studies any unsealed byproduct material, except in quantities that require a written directive pursuant to Sec. 35.40, prepared for medical use that is either--

Except for quantities that require a written directive pursuant to § 35.40(b)(2), a licensee may use any unsealed byproduct material prepared for medical use for uptake, dilution, or excretion studies that is --

(a) Obtained from a manufacturer or preparer licensed pursuant to 10 CFR 32.72 or equivalent Agreement State requirements; or

(a) Obtained from a manufacturer or preparer licensed pursuant to Sec. 32.72 of this chapter or equivalent Agreement State requirements; or

(a) Obtained from a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in Section 35.920, or an individual under the supervision of either as specified in Section 35.25.

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in Sec. 35.292, or an individual under the supervision of either as specified in Sec. 35.27.

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in §§ 35.290 or 35.390, or an individual under the supervision of either as specified in § 35.27; or

(c) Obtained from an NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

approved protocol or an Investigational New Drug (IND) protocol accepted by FDA; or
(d) Prepared by the licensee for use in research in accordance with a Radioactive Drug Research Committee-approved application or an Investigational New Drug (IND) protocol accepted by FDA.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.120 Possession of survey instrument.

A licensee authorized to use byproduct material for uptake, dilution, and excretion studies shall have in its possession a portable radiation detection survey instrument capable of detecting dose rates over the range 0.1 millirem per hour to 100 millirem per hour.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.190 Training for uptake, dilution, and excretion studies.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.100 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under §§ 35.290 or 35.390 or equivalent Agreement State requirements; or

(c)(1) Has completed 60 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material for uptake, dilution, and excretion studies; the training and experience must include --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(B) Radiation protection;
(C) Mathematics
pertaining to the use and
measurement of radioactivity;
(D) Chemistry of
byproduct material for medical
use; and
(E) Radiation biology;
and
(ii) Work experience,
under the supervision of an
authorized user who meets the
requirements in § 35.190, §
35.290, or § 35.390 or equivalent
Agreement State requirements,
involving --
(A) Ordering, receiving,
and unpacking radioactive
materials safely and performing
the related radiation surveys;
(B) Calibrating
instruments used to determine
the activity of dosages and
performing checks for proper
operation of survey meters;
(C) Calculating,
measuring, and safely preparing
patient or human research
subject dosages;
(D) Using administrative
controls to prevent a medical
event involving the use of
unsealed byproduct material;
(E) Using procedures to
contain spilled byproduct material

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

safely and using proper decontamination procedures; and

(F) Administering dosages of radioactive drugs to patients or human research subjects; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.190, 35.290, or 35.390 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (c)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under § 35.100.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.200 Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

A licensee may use for imaging and localization studies any unsealed byproduct material, except in quantities that require a written directive pursuant to Sec. 35.40, prepared for medical use that is either--

(a) Obtained from a manufacturer or preparer licensed pursuant to Sec. 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in Sec. 35.292, or an individual under the supervision of either as specified in Sec. 35.27.

§ 35.200 Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required.

Except for quantities that require a written directive under § 35.40(b)(2), a licensee may use any unsealed byproduct material prepared for medical use for imaging and localization studies that is --

(a) Obtained from a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in §§ 35.290 or 35.390, or an individual under the supervision of either as specified in § 35.27; or

(c) Obtained from an NRC or Agreement State licensee for use in research in accordance with a Radioactive Drug Research Committee-approved protocol or an Investigational New Drug (IND) protocol accepted by FDA.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.204 Permissible molybdenum-99 concentration.

(a) A licensee may not administer to humans a radiopharmaceutical that contains more than 0.15 kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (0.15 microcurie of molybdenum-99 per millicurie of technetium-99m).

(b) A licensee that uses molybdenum-99/technetium-99m generators for preparing a technetium-99m radiopharmaceutical shall measure the molybdenum-99 concentration of the first eluate after receipt of a generator to demonstrate compliance with paragraph (a) of this section.

(c) If a licensee is required to measure the molybdenum-99 concentration, the licensee shall retain a record of each measurement in accordance with § 35.2204.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.290 Training for imaging and localization studies.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.200 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under § 35.390 or equivalent Agreement State requirements; or

(c)(1) Has completed 700 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material for imaging and localization studies; the training and experience must include, at a minimum, --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(B) Radiation protection;
(C) Mathematics
pertaining to the use and
measurement of radioactivity;
(D) Chemistry of
byproduct material for medical
use;
(E) Radiation biology;
and
(ii) Work experience,
under the supervision of an
authorized user, who meets the
requirements in §§ 35.290 or
35.390 or equivalent Agreement
State requirements, involving --
(A) Ordering, receiving,
and unpacking radioactive
materials safely and performing
the related radiation surveys;
(B) Calibrating
instruments used to determine
the activity of dosages and
performing checks for proper
operation of survey meters;
(C) Calculating,
measuring, and safely preparing
patient or human research
subject dosages;
(D) Using administrative
controls to prevent a medical
event involving the use of
unsealed byproduct material;
(E) Using procedures to
safely contain spilled radioactive

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

material and using proper decontamination procedures;

(F) Administering dosages of radioactive drugs to patients or human research subjects; and

(G) Eluting generator systems appropriate for preparation of radioactive drugs for imaging and localization studies, measuring and testing the eluate for radionuclidic purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.290 or § 35.390 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (c)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under §§ 35.100 and 35.200.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart E--Unsealed Byproduct Material - Written Directive Required

§ 35.300 Use of unsealed byproduct material for which a written directive is required.

A licensee may use any unsealed byproduct material prepared for medical use and for which a written directive is required that is --

(a) Obtained from a manufacturer or preparer licensed under § 32.72 of this chapter or equivalent Agreement State requirements; or

(b) Prepared by an authorized nuclear pharmacist, a physician who is an authorized user and who meets the requirements specified in §§ 35.290 or 35.390, or an individual under the supervision of either as specified in § 35.27; or

(c) Obtained from an NRC or Agreement State licensee for use in research in accordance with an Investigational New Drug (IND) application accepted by FDA; or

(d) Prepared by the licensee for use in accordance

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart E--Imaging and
Localization

with an Investigational New Drug
(IND) protocol accepted by FDA.

Section 35.200 Use of unsealed
byproduct material for imaging
and localization studies.

A licensee may use for
imaging and localization studies
any unsealed byproduct material
prepared for medical use that is
either:

(a) Obtained from a
manufacturer or preparer
licensed pursuant to 10 CFR
32.72 or equivalent Agreement
State requirements; or

(b) Prepared by an
authorized nuclear pharmacist, a
physician who is an authorized
user and who meets the
requirements specified in Section
35.920, or an individual under the
supervision of either as specified
in Section 35.25.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.204 Permissible molybdenum-99 concentration.

(a) A licensee may not administer to humans a radiopharmaceutical containing more than 0.15 microcurie of molybdenum-99 per millicurie of technetium-99m.

(b) A licensee that uses molybdenum-99/technetium-99m generators for preparing a technetium-99m radiopharmaceutical shall measure the molybdenum-99 concentration in each eluate or extract.

(c) A licensee that must measure molybdenum concentration shall retain a record of each measurement for three years. The record must include, for each elution or extraction of technetium-99m, the measured activity of the technetium expressed in millicuries, the measured activity of the molybdenum expressed in microcuries, the ratio of the

Sec. 35.204 Permissible molybdenum-99 concentration.

(a) A licensee may not administer to humans a radiopharmaceutical containing more than 5.55 kBq (0.15 <greek-m>Ci) of molybdenum-99 per millicurie of technetium-99m.

(b) A licensee that uses molybdenum-99/technetium-99m generators for preparing a technetium-99m radiopharmaceutical shall measure the molybdenum-99 concentration of the first eluate after receipt of a generator to demonstrate compliance with paragraph (a) of this section.

(c) A licensee that must measure molybdenum concentration shall retain a record of each measurement in accordance with Sec. 35.2204.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

measures expressed as microcuries of molybdenum per millicurie of technetium, the time and date of the measurement, and the initials of the individual who made the measurement.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.205 Control of aerosols and gases.

(a) A licensee that administers radioactive aerosols or gases shall do so in a room with a system that will keep airborne concentrations low enough so as not to exceed the limits prescribed by Sections 20.1201 and 20.1301 of this chapter. The system must either be directly vented to the atmosphere through an air exhaust or provide for collection and decay or disposal of the aerosol or gas in a shielded container.

(b) A licensee shall administer radioactive gases only in rooms that are at negative pressure compared to surrounding rooms.

(c) Before receiving, using, or storing a radioactive gas, the licensee shall calculate the amount of time needed after a spill to reduce the concentration in the room low enough so as not to exceed the limits prescribed by Section 20.1201 of this chapter. The calculation must be based on the highest activity of gas handled in a single container, the air volume of the room, and the

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

measured available air exhaust rate.

(d) A licensee shall make a record of the calculations required in paragraph (c) of this section that includes the assumptions, measurements, and calculations made and shall retain the record for the duration of use of the area. A licensee shall also post the calculated time and safety measures to be instituted in case of a spill at the area of use.

(e) A licensee shall check the operation of reusable collection systems each month, and measure the ventilation rates available in areas of radioactive gas use each six months.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.220 Possession of survey instruments.

A licensee authorized to use byproduct material for imaging and localization studies shall have in its possession a portable radiation detection survey instrument capable of detecting dose rates over the range of 0.1 millirem per hour to 100 millirem per hour, and a portable radiation measurement survey instrument capable of measuring dose rates over the range 1 millirem per hour to 1000 millirem per hour.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.290 Training for uptake, dilution, and excretion studies.

Except as provided in Secs. 35.57, the licensee shall require the authorized user of a radiopharmaceutical for uses authorized under Sec. 35.100 to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission or--

(b)(1) Has completed a structured educational program in basic radionuclide handling techniques applicable to the use of diagnostic radiopharmaceuticals, consisting of both--

(i) 40 hours of didactic training in the following areas--
(A) Radiation physics and instrumentation;
(B) Radiation protection;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(C) Mathematics pertaining to the use and measurement of radioactivity;
(D) Chemistry of byproduct material for medical use; and
(E) Radiation biology; and
(ii) 20 hours of supervised practical experience under the supervision of an authorized user involving—

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating dose calibrators and diagnostic instruments and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of byproduct material;

(E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(F) Administering dosages to patients or human research subjects; and

(2) Has obtained written certification, signed by a preceptor authorized user, that the requirements in paragraph (b)(1) of this section have been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently function as an authorized user of a diagnostic radiopharmaceutical for the uses listed in Sec. 35.100; and

(3) Following completion of the requirements in paragraph (b)(1) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.292 Training for imaging and localization studies.

Except as provided in Secs. 35.57, the licensee shall require the authorized user of radiopharmaceuticals and generators for the uses authorized under Sec. 35.200 to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission; or

(b)(1) Has completed a structured educational program in basic radionuclide handling techniques applicable to the use of diagnostic radiopharmaceuticals and generators, consisting of both--

(i) 80 hours of didactic training in the following areas--

(A) Radiation physics and instrumentation;

(B) Radiation protection;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(C) Mathematics pertaining to the use and measurement of radioactivity;
(D) Chemistry of byproduct material for medical use; and
(E) Radiation biology; and
(ii) 40 hours of supervised practical experience under the supervision of an authorized user involving--

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating dose calibrators and diagnostic instruments and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of byproduct material;

(E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(F) Eluting technetium-99m from generator systems, measuring and testing the eluate for molybdenum-99, and processing the eluate with reagent kits to prepare technetium-99m labeled radiopharmaceuticals; and

(G) Administering dosages to patients or human research subjects; and

(2) Has obtained written certification, signed by a preceptor authorized user, that the requirements in paragraph (b)(1) of this section have been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently function as an authorized user of diagnostic radiopharmaceuticals and generators for the uses listed in Sec. 35.200; and

(3) Following completion of the requirements in paragraph (b)(1) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

by the Commission in accordance with appendix A of this part.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.310 Safety instruction.

In addition to the requirements of § 19.12 of this chapter,

(a) A licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that have received therapy with unsealed byproduct material and cannot be released in accordance with § 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include --

(1) Patient or human research subject control;

(2) Visitor control, including --

(i) Routine visitation to hospitalized individuals in accordance with § 20.1301(a)(1) of this chapter; and

(ii) Visitation authorized in accordance with § 20.1301(c) of this chapter;

(3) Contamination control;

(4) Waste control; and

(5) Notification of the Radiation Safety Officer, or his or her designee, and the authorized

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

user if the patient or the human research subject has a medical emergency or dies.

(b) A licensee shall retain a record of individuals receiving instruction in accordance with § 35.2310.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.315 Safety precautions.

(a) For each patient or human research subject that cannot be released in accordance with § 35.75, a licensee shall --

(1) Quarantine the patient or the human research subject either in --

(i) A private room with a private sanitary facility; or

(ii) A room, with a private sanitary facility, with another individual who also has received therapy with unsealed byproduct material and who also cannot be released pursuant to § 35.75;

(2) Visibly post the patient's or the human research subject's room with a "Radioactive Materials" sign.

(3) Note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or the human research subject's room; and

(4) Either monitor material and items removed from the patient's or the human research subject's room to determine that their radioactivity cannot be distinguished from the natural background radiation level

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

with a radiation detection survey instrument set on its most sensitive scale and with no interposed shielding, or handle such material and items as radioactive waste.

(b) A licensee shall notify the Radiation Safety Officer, or his or her designee, and the authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.390 Training for use of unsealed byproduct material for which a written directive is required.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.300 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b)(1) Has completed 700 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material requiring a written directive; the training and experience must include --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology;

and

(ii) Work experience, under the supervision of an authorized user who meets the requirements in § 35.390(a) or § 35.390(b)(1)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, involving --

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating instruments used to determine the activity of dosages, and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;

(E) Using procedures to contain spilled byproduct material

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

safely and using proper decontamination procedures;

(F) Eluting generator systems, measuring and testing the eluate for radiochemical purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs; and

(G) Administering dosages of radioactive drugs to patients or human research subjects involving a minimum of three cases in each of the following categories for which the individual is requesting authorized user status--

(1) Oral administration of less than or equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131;

(2) Oral administration of greater than 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131¹;

(3) Parenteral administration of any beta emitter or a photon-emitting radionuclide with a photon energy less than 150 keV; and/or

(4) Parenteral administration of any other radionuclide; and

(2) Has obtained written certification, signed by a preceptor authorized user who

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

meets the requirements in § 35.390(a) or § 35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under § 35.300(a) or § 35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.392 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).

Except as provided in § 35.57, the licensee shall require an authorized user for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries), to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under §§ 35.390(a), 35.390(b)(1)(ii)(G)(1) or (2), or 35.394 or equivalent Agreement State requirements; or

(c)(1) Has successfully completed 80 hours of classroom and laboratory training, applicable to the medical use of sodium iodide I-131 for procedures

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

requiring a written directive; the training must include --

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics pertaining to the use and measurement of radioactivity;
- (iv) Chemistry of byproduct material for medical use; and
- (v) Radiation biology; and

(2) Has work experience, under the supervision of an authorized user who meets the requirements in §§ 35.390(b)(1)(ii)(G)(1) or (2) or 35.392 or equivalent Agreement State requirements, involving --

- (i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (ii) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation for survey meters;
- (iii) Calculating, measuring, and safely preparing patient or human research subject dosages;
- (iv) Using administrative controls to prevent a medical

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

event involving the use of byproduct material;

- (v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and
- (vi) Administering dosages to patients or human research subjects, that includes at least 3 cases involving the oral administration of less than or equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.390(b)(1)(ii)(G)(1), 35.390(b)(1)(ii)(G)(2), 35.392, or 35.394, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (c)(1) and (c)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user for medical uses of unsealed byproduct material using sodium iodide I-131.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.394 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).

Except as provided in § 35.57, the licensee shall require an authorized user for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries), to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Is an authorized user under § 35.390(a), 35.390(b)(1)(ii)(G)(2) or equivalent Agreement State requirements; or

(c)(1) Has successfully completed 80 hours of classroom and laboratory training, applicable to the medical use of sodium iodide I-131 for procedures requiring a written directive; the training must include --

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(i) Radiation physics and instrumentation;

(ii) Radiation protection;

(iii) Mathematics pertaining to the use and measurement of radioactivity;

(iv) Chemistry of byproduct material for medical use; and

(v) Radiation biology; and

(2) Has work experience, under the supervision of an authorized user who meets the requirements in §§ 35.390(a), 35.390(b)(1)(ii)(G)(2) or 35.394 or equivalent Agreement State requirements, involving --

(i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(ii) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation for survey meters;

(iii) Calculating, measuring, and safely preparing patient or human research subject dosages;

(iv) Using administrative controls to prevent a medical

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

event involving the use of byproduct material;

(v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and

(vi) Administering dosages to patients or human research subjects, that includes at least 3 cases involving the oral administration of greater than 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.390(a), 35.390(b)(1)(ii)(G)(2) or 35.394 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (c)(1) and (c)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user for medical uses of unsealed byproduct material using sodium iodide I-131.

1.Experience with at least 3 cases in Category (2) also satisfies the requirement in Category (1).

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart F--
Radiopharmaceuticals for
Therapy

Subpart E--Unsealed Byproduct
Material--High Dose

Subpart E--Unsealed Byproduct
Material - Written Directive
Required

Section 35.300 Use of unsealed
byproduct material for therapeutic
administration.

Sec. 35.300 Use of unsealed
byproduct material for which a
written directive is required.

§ 35.300 Use of unsealed
byproduct material for which a
written directive is required.

A licensee may use for
therapeutic administration any
unsealed byproduct material
prepared for medical use that is
either:

A licensee may use any
unsealed byproduct material
prepared for medical use and for
which a written directive is
required that is either--

A licensee may use any
unsealed byproduct material
prepared for medical use and for
which a written directive is
required that is --

(a) Obtained from a
manufacturer or preparer
licensed pursuant to 10 CFR
32.72 or equivalent Agreement
State requirements; or

(a) Obtained from a
manufacturer or preparer
licensed pursuant to Sec. 32.72
of this chapter or equivalent
Agreement State requirements;
or

(a) Obtained from a
manufacturer or preparer
licensed under § 32.72 of this
chapter or equivalent Agreement
State requirements; or

(b) Prepared by an
authorized nuclear pharmacist, a
physician who is an authorized
user and who meets the
requirements specified in Section
35.920, or an individual under the
supervision of either as specified
in Section 35.25.

(b) Prepared by an
authorized nuclear pharmacist, a
physician who is an authorized
user and who meets the
requirements specified in Sec.
35.292, or an individual under the
supervision of either as specified
in Sec. 35.27.

(b) Prepared by an
authorized nuclear pharmacist, a
physician who is an authorized
user and who meets the
requirements specified in §§
35.290 or 35.390, or an individual
under the supervision of either as
specified in § 35.27; or

(c) Obtained from an
NRC or Agreement State
licensee for use in research in
accordance with an
Investigational New Drug (IND)
application accepted by FDA; or

(d) Prepared by the
licensee for use in accordance

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

with an Investigational New Drug
(IND) protocol accepted by FDA.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.310 Safety instruction.

Sec. 35.310 Safety instruction.

§ 35.310 Safety instruction.

In addition to the requirements of Sec. 19.12 of this chapter,

In addition to the requirements of § 19.12 of this chapter,

(a) A licensee shall provide radiation safety instruction for all personnel caring for the patient or the human research subject receiving radiopharmaceutical therapy and hospitalized for compliance with Section 35.75 of this chapter. To satisfy this requirement, the instruction must describe the licensee's procedures for:

(a) A licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that have received radiopharmaceutical therapy and can not be released in accordance with Sec. 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include--

(a) A licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that have received therapy with unsealed byproduct material and cannot be released in accordance with § 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include --

- (1) Patient or human research subject control;
- (2) Visitor control;

- (1) Patient or human research subject control;
- (2) Visitor control, including--

- (1) Patient or human research subject control;
- (2) Visitor control, including --

- (i) Routine visitation to hospitalized individuals in accordance with Sec. 20.1301(a)(1) of this chapter; and
- (ii) Visitation authorized in accordance with Sec. 20.1301(a)(3);

- (i) Routine visitation to hospitalized individuals in accordance with § 20.1301(a)(1) of this chapter; and
- (ii) Visitation authorized in accordance with § 20.1301(c) of this chapter;

- (3) Contamination control;
- (4) Waste control; and
- (5) Notification of the Radiation Safety Officer in case of the patient's or the human

- (3) Contamination control;
- (4) Waste control; and
- (5) Notification of the authorized user and the Radiation Safety Officer, or his designee, if

- (3) Contamination control;
- (4) Waste control; and
- (5) Notification of the Radiation Safety Officer, or his or

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

research subject's death or medical emergency.

(b) A licensee shall keep for three years a list of individuals receiving instruction required by paragraph (a) of this section, a description of the instruction, the date of instruction, and the name of the individual who gave the instruction.

the patient or the human research subject dies or has a medical emergency.

(b) A licensee shall retain a record of individuals receiving instruction in accordance with Sec. 35.2310.

her designee, and the authorized user if the patient or the human research subject has a medical emergency or dies.

(b) A licensee shall retain a record of individuals receiving instruction in accordance with § 35.2310.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.315 Safety precautions.

(a) For each patient or human research subject receiving radiopharmaceutical therapy and hospitalized for compliance with Section 35.75 of this chapter, a licensee shall:

(1) Provide a private room with a private sanitary facility;

(2) Post the patient's or the human research subject's door with a "Radioactive Materials" sign and note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or the human research subject's room;

(3) Authorize visits by individuals under age 18 only on a case-by-case basis with the approval of the authorized user

Sec. 35.315 Safety precautions.

(a) For each patient or human research subject that cannot be released in accordance with Sec. 35.75, a licensee shall--

(1) Provide a private room with a private sanitary facility;

(2) Visibly post the patient's or the human research subject's room with a "Radioactive Materials" sign and note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or the human research subject's room; and

§ 35.315 Safety precautions.

(a) For each patient or human research subject that cannot be released in accordance with § 35.75, a licensee shall --

(1) Quarter the patient or the human research subject either in --

(i) A private room with a private sanitary facility; or

(ii) A room, with a private sanitary facility, with another individual who also has received therapy with unsealed byproduct material and who also cannot be released pursuant to § 35.75;

(2) Visibly post the patient's or the human research subject's room with a "Radioactive Materials" sign.

(3) Note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or the human research subject's room; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

after consultation with the Radiation Safety Officer;
(4) Promptly after administration of the dosage, measure the dose rates in contiguous restricted and unrestricted areas with a radiation measurement survey instrument to demonstrate compliance with the requirements of part 20 of this chapter, and retain for three years a record of each survey that includes the time and date of the survey, a plan of the area or list of points surveyed, the measured dose rate at several points expressed in millirem per hour, the instrument used to make the survey, and the initials of the individual who made the survey.

(5) Either monitor material and items removed from the patient's or the human research subject's room to determine that their radioactivity cannot be distinguished from the natural background radiation level with a radiation detection survey instrument set on its most sensitive scale and with no interposed shielding, or handle them as radioactive waste.

(6)[Reserved 62 FR 4120.].

(3) Either monitor material and items removed from the patient's or the human research subject's room to determine that their radioactivity cannot be distinguished from the natural background radiation level with a radiation detection survey instrument set on its most sensitive

(4) Either monitor material and items removed from the patient's or the human research subject's room to determine that their radioactivity cannot be distinguished from the natural background radiation level with a radiation detection survey instrument set on its most sensitive scale and with no

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(7) Survey the patient's or the human research subject's room and private sanitary facility for removable contamination with a radiation detection survey instrument before assigning another patient or human research subject to the room. The room must not be reassigned until removable contamination is less than 200 disintegrations per minute per 100 square centimeters; and

(8) Measure the thyroid burden of each individual who helped prepare or administer a dosage of iodine-131 within three days after administering the dosage, and retain for the period required by Section 20.1206(a) of this chapter a record of each thyroid burden measurement, its date, the name of the individual whose thyroid burden was measured, and the initials of the individual who made the measurements.

(b) A licensee shall notify the Radiation Safety Officer immediately if the patient or the human research subject dies or has a medical emergency.

scale and with no interposed shielding, or handle such material and items as radioactive waste.

interposed shielding, or handle such material and items as radioactive waste.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) A licensee shall notify the authorized user and the Radiation Safety Officer, or his or her designee, as soon as possible, if the patient or human research subject has a medical emergency and, immediately, if the patient dies.

(b) A licensee shall notify the Radiation Safety Officer, or his or her designee, and the authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

Section 35.320 Possession of survey instruments.

A licensee authorized to use byproduct material for radiopharmaceutical therapy shall have in its possession a portable radiation detection survey instrument capable of detecting dose rates over the range 0.1 millirem per hour to 100 millirem per hour, and a portable radiation measurement survey instrument capable of measuring dose rates over the range 1 millirem per hour to 1000 millirem per hour.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.390 Training for use of unsealed byproduct material for therapy or for use of unsealed byproduct material that requires a written directive.

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a radiopharmaceutical for the uses authorized under Sec. 35.300 to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission; or

(b)(1) Has completed a structured educational program in basic radionuclide handling techniques applicable to the use of unsealed byproduct material consisting of both--

(i) 80 hours of didactic training in the following areas--

(A) Radiation physics and instrumentation;

(B) Radiation protection;

§ 35.390 Training for use of unsealed byproduct material for which a written directive is required.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.300 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission; or

(b)(1) Has completed 700 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material requiring a written directive; the training and experience must include --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics

pertaining to the use and measurement of radioactivity;

§ 35.390 Training for use of unsealed byproduct material for which a written directive is required.

Except as provided in § 35.57, the licensee shall require an authorized user of unsealed byproduct material for the uses authorized under § 35.300 to be a physician who --

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission; or

(b)(1) Has completed 700 hours of training and experience in basic radionuclide handling techniques applicable to the medical use of unsealed byproduct material requiring a written directive; the training and experience must include --

(i) Classroom and laboratory training in the following areas --

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics

pertaining to the use and measurement of radioactivity;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(C) Mathematics pertaining to the use and measurement of radioactivity; and

(D) Radiation biology; and

(ii) 40 hours of supervised practical experience under the supervision of an authorized user at a medical institution involving--

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating dose calibrators, as appropriate, and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of byproduct material;

(E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(ii) Work experience, under the supervision of an authorized user who meets the requirements in § 35.390(a) or § 35.390(b)(1)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, involving --

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating instruments used to determine the activity of dosages, and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;

(E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(ii) Work experience, under the supervision of an authorized user who meets the requirements in § 35.390(a) or § 35.390(b)(1)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, involving --

(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;

(B) Calibrating instruments used to determine the activity of dosages, and performing checks for proper operation of survey meters;

(C) Calculating, measuring, and safely preparing patient or human research subject dosages;

(D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;

(E) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(2) Has had experience, obtained under the direct supervision of an authorized user, involving at least five cases for each procedure with radiation safety hazards similar to that use for which the individual is requesting authorized user status. This experience may be obtained concurrently with the supervised practical experience required by paragraph (b)(1)(ii) of this section;

(3) Has obtained written certification, signed by a preceptor authorized user, that

PROPOSED RULE

(F) Eluting generator systems, measuring and testing the eluate for radiochemical purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs; and

(G) Administering dosages of radioactive drugs to patients or human research subjects involving a minimum of three cases in each of the following categories for which the individual is requesting authorized user status--

(1) Oral administration of less than or equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131;

(2) Oral administration of greater than 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131¹;

(3) Parenteral administration of any beta emitter or a photon-emitting radionuclide with a photon energy less than 150 keV; and/or

(4) Parenteral administration of any other radionuclide; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.390(a) or §

FINAL RULE

(F) Eluting generator systems, measuring and testing the eluate for radiochemical purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs; and

(G) Administering dosages of radioactive drugs to patients or human research subjects involving a minimum of three cases in each of the following categories for which the individual is requesting authorized user status--

(1) Oral administration of less than or equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131;

(2) Oral administration of greater than 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131²;

(3) Parenteral administration of any beta emitter or a photon-emitting radionuclide with a photon energy less than 150 keV; and/or

(4) Parenteral administration of any other radionuclide; and

(2) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.390(a) or §

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

the requirements in paragraphs (b) (1) and (2) of this section have been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently function as an authorized user of unsealed byproduct material for the uses listed in Sec. 35.300; and
 (4) Following completion of the requirements in paragraph (b) (1) and (2) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part.

35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under § 35.300(a) or § 35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate.

35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraph (b)(1) of this section and has achieved a level of competency sufficient to function independently as an authorized user for the medical uses authorized under § 35.300(a) or § 35.390(b)(1)(ii)(G)(1), (2), (3), or (4), as appropriate.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.392 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).

Except as provided in § 35.57, the licensee shall require an authorized user for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries), to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission; or

(b) Is an authorized user under §§ 35.390(a), 35.390(b)(1)(ii)(G)(1) or (2), or 35.394 or equivalent Agreement State requirements; or

(c)(1) Has successfully completed 80 hours of classroom and laboratory training, applicable to the medical use of sodium iodide I-131 for procedures requiring a written directive; the training must include --

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(i) Radiation physics and instrumentation;
(ii) Radiation protection;
(iii) Mathematics pertaining to the use and measurement of radioactivity;
(iv) Chemistry of byproduct material for medical use; and
(v) Radiation biology; and
(2) Has work experience, under the supervision of an authorized user who meets the requirements in §§ 35.390(b)(1)(ii)(G)(1) or (2) or 35.392 or equivalent Agreement State requirements, involving --
(i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
(ii) Calibrating instruments used to determine the activity of dosages and performing checks for proper operation for survey meters;
(iii) Calculating, measuring, and safely preparing patient or human research subject dosages;
(iv) Using administrative controls to prevent a medical event involving the use of byproduct material;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and

(vi) Administering dosages to patients or human research subjects, that includes at least 3 cases involving the oral administration of less than or equal to 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.390(b)(1)(ii)(G)(1), 35.390(b)(1)(ii)(G)(2), 35.392, or 35.394, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (c)(1) and (c)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user for medical uses of unsealed byproduct material using sodium iodide I-131.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.394 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).

Except as provided in § 35.57, the licensee shall require an authorized user for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries), to be a physician who--

(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (c) of this section and whose certification has been recognized by the Commission; or

(b) Is an authorized user under § 35.390(a), 35.390(b)(1)(ii)(G)(2) or equivalent Agreement State requirements; or

(c)(1) Has successfully completed 80 hours of classroom and laboratory training, applicable to the medical use of sodium iodide I-131 for procedures requiring a written directive; the training must include --

(i) Radiation physics and instrumentation;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(ii) Radiation protection;
(iii) Mathematics
pertaining to the use and
measurement of
radioactivity;
(iv) Chemistry of
byproduct material for
medical use; and
(v) Radiation biology; and
(2) Has work experience,
under the supervision of an
authorized user who meets the
requirements in §§ 35.390(a),
35.390(b)(1)(ii)(G)(2) or 35.394 or
equivalent Agreement State
requirements, involving --
(i) Ordering, receiving,
and unpacking radioactive
materials safely and performing
the related radiation surveys;
(ii) Calibrating
instruments used to determine
the activity of dosages and
performing checks for proper
operation for survey meters;
(iii) Calculating,
measuring, and safely preparing
patient or human research
subject dosages;
(iv) Using administrative
controls to prevent a medical
event involving the use of
byproduct material;
(v) Using procedures to
contain spilled byproduct material

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

safely and using proper decontamination procedures; and
(vi) Administering dosages to patients or human research subjects, that includes at least 3 cases involving the oral administration of greater than 1.22 Gigabecquerels (33 millicuries) of sodium iodide I-131; and
(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in §§ 35.390(a), 35.390(b)(1)(ii)(G)(2) or 35.394 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (c)(1) and (c)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user for medical uses of unsealed byproduct material using sodium iodide I-131.

- 1.Experience with at least 3 cases in Category (2) also satisfies the requirement in Category (1).
- 2.Experience with at least 3 cases in Category (2) also satisfies the requirement in Category (1).

Subpart G--Sources for
Brachytherapy

Section 35.400 Use of sources
for brachytherapy.

A licensee shall use the
following sources in accordance
with the manufacturer's radiation
safety and handling instructions:

- (a) Cesium-137 as a
sealed source in needles and
applicator cells for topical,
interstitial, and intracavitary
treatment of cancer;
- (b) Cobalt-60 as a sealed
source in needles and applicator
cells for topical, interstitial, and
intracavitary treatment of cancer;
- (c) Gold-198 as a sealed
source in seeds for interstitial
treatment of cancer;
- (d) Iridium-192 as seeds
encased in nylon ribbon for
interstitial treatment of cancer;
- (e) Strontium-90 as a
sealed source in an applicator for
treatment of superficial eye
conditions; and
- (f) Iodine-125 as a sealed
source in seeds for interstitial
treatment of cancer.
- (g) Palladium-103 as a
sealed source in seeds for
interstitial treatment of cancer.

Subpart F-- Manual
Brachytherapy

Sec. 35.400 Use of sources for
manual brachytherapy.

A licensee shall use only
brachytherapy sources for
therapeutic medical uses as
approved in the Sealed Source
and Device Registry.

Subpart F-- Manual
Brachytherapy

§ 35.400 Use of sealed sources
for manual brachytherapy.

A licensee shall use only
brachytherapy sealed sources for
therapeutic medical uses:

- (a) As approved in the
Sealed Source and Device
Registry; or
- (b) In research in
accordance with an effective
Investigational Device Exemption
(IDE) application accepted by the
FDA provided the requirements of
§ 35.49(a) are met.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.404 Release of patients or human research subjects treated with temporary implants.

(a) Immediately after removing the last temporary implant source from a patient or a human research subject, the licensee shall make a radiation survey of the patient or the human research subject with a radiation detection survey instrument to confirm that all sources have been removed. The licensee may not release from confinement for medical care a patient or a human research subject treated by temporary implant until all sources have been removed.

(b) A licensee shall retain a record of patient or human research subject surveys for three years. Each record must include the date of the survey, the name of the patient or the human research subject, the dose rate from the patient or the human research subject expressed as millirem per hour and measured at 1 meter from

Sec. 35.404 Radiation surveys of patients or human research subjects treated with implants.

(a) Immediately after implanting sources in a patient or a human research subject, the licensee shall make a radiation survey of the patient or the human research subject and the adjacent area of use to confirm that no sources have been misplaced.

(b) Immediately after removing the last temporary implant source from a patient or a human research subject, the licensee shall make a radiation survey of the patient or the human research subject with a radiation detection survey instrument to confirm that all sources have been removed.

(c) A licensee shall retain a record of patient or human research subject surveys in accordance with Sec. 35.2404.

§ 35.404 Surveys after source implant and removal.

(a) Immediately after implanting sources in a patient or a human research subject, the licensee shall make a survey to locate and account for all sources that have not been implanted.

(b) Immediately after removing the last temporary implant source from a patient or a human research subject, the licensee shall make a survey of the patient or the human research subject with a radiation detection survey instrument to confirm that all sources have been removed.

(c) A licensee shall retain a record of the surveys required by paragraphs (a) and (b) of this section in accordance with § 35.2404.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

the patient or the human research subject, the survey instrument used, and the initials of the individual who made the survey.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.406 Brachytherapy sources inventory.

Sec. 35.406 Brachytherapy sources inventory.

§ 35.406 Brachytherapy sources accountability.

(a) Promptly after removing them from a patient or a human research subject, a licensee shall return brachytherapy sources to the storage area, and count the number returned to ensure that all sources taken from the storage area have been returned.

(b) A licensee shall make a record of brachytherapy source use which must include:

(1) The names of the individuals permitted to handle the sources;

(2) The number and activity of sources removed from storage, the patient's or the human research subject's name and room number, the time and date they were removed from storage, the number and activity of the sources in storage after the removal, and the initials of the individual who removed the sources from storage;

(3) The number and activity of sources returned to

(a) A licensee shall maintain accountability at all times for all brachytherapy sources in storage or use.

(b) Promptly after removing sources from a patient or a human research subject, a licensee shall return brachytherapy sources to a secure storage area.

(c) A licensee shall maintain a record of the brachytherapy source accountability in accordance with Sec. 35.2406.

(a) A licensee shall maintain accountability at all times for all brachytherapy sources in storage or use.

(b) As soon as possible after removing sources from a patient or a human research subject, a licensee shall return brachytherapy sources to a secure storage area.

(c) A licensee shall maintain a record of the brachytherapy source accountability in accordance with § 35.2406.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

storage, the patient's or the human research subject's name and room number, the time and date they were returned to storage, the number and activity of sources in storage after the return, and the initials of the individual who returned the sources to storage.

(c) Immediately after implanting sources in a patient or a human research subject the licensee shall make a radiation survey of the patient or the human research subject and the area of use to confirm that no sources have been misplaced. The licensee shall make a record of each survey.

(d) A licensee shall retain the records required in paragraphs (b) and (c) of this section for three years.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.410 Safety instruction.

Sec. 35.410 Safety instruction.

§ 35.410 Safety instruction.

(a) The licensee shall provide radiation safety instruction to all personnel caring for the patient or the human research subject undergoing implant therapy.

In addition to the requirements of Sec. 19.12 of this chapter,

(a) The licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that are undergoing implant therapy and cannot be released in accordance with Sec. 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include the --

In addition to the requirements of § 19.12 of this chapter,

(a) The licensee shall provide radiation safety instruction, initially and at least annually, to personnel caring for patients or human research subjects that are undergoing implant therapy and cannot be released in accordance with § 35.75. To satisfy this requirement, the instruction must be commensurate with the duties of the personnel and include the -

To satisfy this requirement, the instruction must describe:

- (1) Size and appearance of the brachytherapy sources;
- (2) Safe handling and shielding instructions in case of a dislodged source;
- (3) Procedures for patient or human research subject control;
- (4) Procedures for visitor control; and

- (1) Size and appearance of the brachytherapy sources;
- (2) Safe handling and shielding instructions;
- (3) Patient or human research subject control;
- (4) Visitor control, including both--
 - (i) Routine visitation of hospitalized individuals in accordance with Sec. 20.1301(a)(1) of this chapter; and
 - (ii) Visitation authorized in accordance with Sec. 20.1301(a)(3); and

- (1) Size and appearance of the brachytherapy sources;
- (2) Safe handling and shielding instructions;
- (3) Patient or human research subject control;
- (4) Visitor control, including both:
 - (i) Routine visitation of hospitalized individuals in accordance with § 20.1301(a)(1) of this chapter; and
 - (ii) Visitation authorized in accordance with § 20.1301(c) of this chapter; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(5) Procedures for notification of the Radiation Safety Officer if the patient or the human research subject dies or has a medical emergency.

(b) A licensee shall retain for three years a record of individuals receiving instruction required by paragraph (a) of this section, a description of the instruction, the date of instruction, and the name of the individual who gave the instruction.

(5) Notification of the authorized user and Radiation Safety Officer, or his or her designee, if the patient or the human research subject dies or has a medical emergency.

(b) A licensee shall retain a record of individuals receiving instruction in accordance with Sec. 35.2310.

(5) Notification of the Radiation Safety Officer, or his or her designee, and an authorized user if the patient or the human research subject has a medical emergency or dies.

(b) A licensee shall retain a record of individuals receiving instruction in accordance with § 35.2310.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.415 Safety precautions.

(a) For each patient or human research subject receiving implant therapy and not released from licensee control pursuant to Section 35.75 of this part, a licensee shall:

(1) Not quarter the patient or the human research subject in the same room with an individual who is not receiving radiation therapy.

(2) Post the patient's or human research subject's door with a "Radioactive Materials" sign and note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or human research subject's room;

(3) Authorize visits by individuals under age 18 only on a case-by-case basis with the approval of the authorized user after consultation with the Radiation Safety Officer; and

(4) Promptly after implanting the material, survey the dose rates in contiguous restricted and unrestricted areas with a radiation measurement

Sec. 35.415 Safety precautions.

(a) For each patient or human research subject receiving brachytherapy and confined pursuant to Sec. 35.75, a licensee shall--

(1) Not quarter the patient or the human research subject in the same room as an individual who is not receiving radiation therapy; and

(2) Visibly post the patient's or human research subject's room with a "Radioactive Materials" sign and note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or human research subject's room.

§ 35.415 Safety precautions.

(a) For each patient or human research subject that is receiving brachytherapy and cannot be released in accordance with § 35.75, a licensee shall --

(1) Not quarter the patient or the human research subject in the same room as an individual who is not receiving brachytherapy;

(2) Visibly post the patient's or human research subject's room with a "Radioactive Materials" sign; and

(3) Note on the door or in the patient's or human research subject's chart where and how long visitors may stay in the patient's or human research subject's room.

(b) A licensee shall have applicable emergency response equipment available near each treatment room to respond to a source --

(1) Dislodged from the patient; and

(2) Lodged within the patient following removal of the source applicators.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

survey instrument to demonstrate compliance with the requirements of Part 20 of this chapter, and retain for three years a record of each survey that includes the time and date of the survey, a plan of the area or list of points surveyed, the measured dose rate at several points expressed in millirem per hour, the instrument used to make the survey, and the initials of the individual who made the survey.

(5) [Removed 62 FR 4120.]

PROPOSED RULE

(b) A licensee shall have available, near each treatment room, emergency response equipment. The emergency response equipment must include, as applicable--

(1) A device to assist in placing the source(s) in the shielded position;

(2) A shielded source/applicator storage container;

(3) Remote handling tools; and

(4) Supplies necessary to surgically remove applicators or sources from a patient or human

FINAL RULE

(c) A licensee shall notify the Radiation Safety Officer, or his or her designee, and an authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) A licensee shall notify the Radiation Safety Officer immediately if the patient or the human research subject dies or has a medical emergency.

research subject treated internally with sealed sources.

(c) A licensee shall notify the authorized user and the Radiation Safety Officer, or his designee, as soon as possible, if the patient or human research subject has a medical emergency and, immediately, if the patient dies.

Section 35.420 Possession of survey instrument.

A licensee authorized to use byproduct material for implant therapy shall have in its possession a portable radiation detection survey instrument capable of detecting dose rates over the range 0.1 millirem per hour to 100 millirem per hour, and a portable radiation measurement survey instrument capable of measuring dose rates over the range 1 millirem per hour to 1000 millirem per hour.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.432 Full calibration measurements of brachytherapy sources.</p>	<p>§ 35.432 Calibration measurements of brachytherapy sealed sources.</p>	
<p>(a) A licensee authorized to use brachytherapy sources for medical use shall perform full calibration measurements on brachytherapy sources before the first medical use of the source or source/appliator configuration.</p>	<p>(a) Before the first medical use of a brachytherapy sealed source on or after [insert date 6 months from publication of the Final Rule], a licensee shall –</p> <p>(1) Determine the source output or activity using a dosimetry system that meets the requirements of § 35.630(a);</p> <p>(2) Determine source positioning accuracy within applicators; and</p> <p>(3) Use published protocols accepted by nationally recognized bodies to meet the requirements of paragraphs (a)(1) and (a)(2) of this section.</p>	
<p>(b) A licensee may use calibration measurements provided by the source manufacturer that are made in accordance with the requirements of this section.</p>	<p>(b) A licensee may use measurements provided by the source manufacturer or by a calibration laboratory accredited by the American Association of Physicists in Medicine that are made in accordance with paragraph (a) of this section.</p>	
<p>(c) To satisfy the requirements of paragraphs (a) and (b) of this section, full calibration measurements</p>	<p>(c) A licensee shall mathematically correct the</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	<p>must include determination of--</p> <p>(1) The output or activity within +/-5 percent;</p> <p>and</p> <p>(2) Source positioning accuracy within applicators.</p> <p>(d) A licensee shall use the dosimetry system described in Sec. 35.630(a) to measure the output or activity of the brachytherapy source.</p> <p>(e) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols by nationally recognized bodies.</p> <p>(f) A licensee shall mathematically correct the outputs or activities determined in paragraph (c) of this section for physical decay at intervals consistent with 1 percent physical decay.</p> <p>(g) A licensee shall retain a record of each calibration in accordance with Sec. 35.2432.</p>	<p>outputs or activities determined in paragraph (a) of this section for physical decay at intervals consistent with 1 percent physical decay.</p> <p>(d) A licensee shall retain a record of each calibration in accordance with § 35.2432.</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.433 Decay of strontium-90 sources for ophthalmic treatments.

(a) Only an authorized medical physicist shall calculate the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments. The decay must be based on the activity determined under § 35.432.

(b) A licensee shall retain a record of the activity of each strontium-90 source in accordance with § 35.2433.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.457 Therapy-related computer systems.

The licensee shall perform acceptance testing on the treatment planning system in accordance with published protocols accepted by nationally recognized bodies. At a minimum, the acceptance testing must include, as applicable, verification of:

- (a) The source-specific input parameters required by the dose calculation algorithm;
- (b) The accuracy of dose, dwell time, and treatment time calculations at representative points;
- (c) The accuracy of isodose plots and graphic displays; and
- (d) The accuracy of the software used to determine radioactive source positions from radiographic images.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.490 Training for use of manual brachytherapy sources.</p>	<p>§ 35.490 Training for use of manual brachytherapy sources.</p>	
<p>Except as provided in Sec. 35.57, the licensee shall require the authorized user of a manual brachytherapy source for the uses authorized under Sec. 35.400 to be a physician who--</p>	<p>Except as provided in § 35.57, the licensee shall require an authorized user of a manual brachytherapy source for the uses authorized under § 35.400 to be a physician who --</p>	
<p>(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission; or</p>	<p>(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or</p>	
<p>(b)(1) Has completed a structured educational program in basic radionuclide handling techniques applicable to the use of manual brachytherapy sources consisting of both--</p>	<p>(b)(1) Has completed a structured educational program in basic radionuclide handling techniques applicable to the use of manual brachytherapy sources that includes --</p>	
<p>(i) 200 hours of didactic training in the following areas –</p>	<p>(i) 200 hours of classroom and laboratory training in the following areas -</p>	
<p>(A) Radiation physics and instrumentation; (B) Radiation protection;</p>	<p>(A) Radiation physics and instrumentation; (B) Radiation protection;</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>(C) Mathematics pertaining to the use and measurement of radioactivity; and</p> <p>(D) Radiation biology;</p> <p>(ii) 500 hours of supervised practical experience, under the supervision of an authorized user at a medical institution, involving—</p> <p>(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;</p> <p>(B) Checking survey meters for proper operation;</p> <p>(C) Preparing, implanting, and removing sealed sources;</p> <p>(D) Maintaining running inventories of material on hand;</p> <p>(E) Using administrative controls to prevent a medical event involving the use of byproduct material;</p> <p>(F) Using emergency procedures to control byproduct material; and</p>	<p>(C) Mathematics pertaining to the use and measurement of radioactivity; and</p> <p>(D) Radiation biology; and</p> <p>(ii) 500 hours of work experience, under the supervision of an authorized user who meets the requirements in § 35.490 or equivalent Agreement State requirements at a medical institution, involving --</p> <p>(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;</p> <p>(B) Checking survey meters for proper operation;</p> <p>(C) Preparing, implanting, and removing sealed sources;</p> <p>(D) Maintaining running inventories of material on hand;</p> <p>(E) Using administrative controls to prevent a medical event involving the use of byproduct material;</p> <p>(F) Using emergency procedures to control byproduct material; and</p>	<p>(C) Mathematics pertaining to the use and measurement of radioactivity; and</p> <p>(D) Radiation biology; and</p> <p>(ii) 500 hours of work experience, under the supervision of an authorized user who meets the requirements in § 35.490 or equivalent Agreement State requirements at a medical institution, involving --</p> <p>(A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;</p> <p>(B) Checking survey meters for proper operation;</p> <p>(C) Preparing, implanting, and removing sealed sources;</p> <p>(D) Maintaining running inventories of material on hand;</p> <p>(E) Using administrative controls to prevent a medical event involving the use of byproduct material;</p> <p>(F) Using emergency procedures to control byproduct material; and</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(2) Three years of supervised clinical experience that includes one year in a formal training program approved by the Residency Review Committee for Radiology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association or equivalent program approved by the NRC, and an additional two years of clinical experience under the supervision of an authorized user; and

(3) Has obtained written certification, signed by a preceptor authorized user, that the requirements in paragraphs (b)(1) and (2) of this section have been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently function as an authorized user of manual

(2) Has obtained three years of supervised clinical experience in radiation oncology, under an authorized user who meets the requirements in § 35.490 or equivalent Agreement State requirements, as part of a formal training program approved by the Residency Review Committee for Radiation Oncology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association. This experience may be obtained concurrently with the supervised work experience required by paragraph (b)(1)(ii) of this section; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.490 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (b)(1) and (b)(2) of this section and has achieved a level of competency

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	<p>brachytherapy sources for the uses listed in Sec. 35.400; and,</p> <p>(4) Following completion of the requirements in paragraph (b)(1) and (2) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part.</p>	<p>sufficient to function independently as an authorized user of manual brachytherapy sources for the medical uses authorized under § 35.400.</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.491 Training for ophthalmic use of strontium-90.

Except as provided in § 35.57, the licensee shall require the authorized user of only strontium-90 for ophthalmic radiotherapy to be a physician who has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of strontium-90 for ophthalmic radiotherapy, and a period of supervised clinical training in ophthalmic radiotherapy as follows --

- (a) 24 hours of classroom and laboratory training that includes --
 - (1) Radiation physics and instrumentation;
 - (2) Radiation protection;
 - (3) Mathematics pertaining to the use and measurement of radioactivity; and
 - (4) Radiation biology;
- (b) Supervised clinical training in ophthalmic radiotherapy under the supervision of an authorized user at a medical institution

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

that includes the use of strontium-90 for the ophthalmic treatment of five individuals that includes --

- (1) Examination of each individual to be treated;
- (2) Calculation of the dose to be administered;
- (3) Administration of the dose; and
- (4) Follow up and review of each individual's case history.

(c) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in 35.490, 35.491, or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in paragraphs (a) and (b) of this section and has achieved a level of competency sufficient to function independently as an authorized user of strontium-90 for ophthalmic use.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart H--Sealed Sources for Diagnosis

Subpart G--Sealed Sources for Diagnosis

Subpart G--Sealed Sources for Diagnosis

Section 35.500 Use of sealed sources for diagnosis.

Sec. 35.500 Use of sealed sources for diagnosis.

§ 35.500 Use of sealed sources for diagnosis.

A licensee shall use the following sealed sources in accordance with the manufacturer's radiation safety and handling instructions:

A licensee shall use only sealed sources for diagnostic medical uses as approved in the Sealed Source and Device Registry.

A licensee shall use only sealed sources for diagnostic medical uses as approved in the Sealed Source and Device Registry.

(a) Iodine-125, americium-241, or gadolinium-153 as a sealed source in a device for bone mineral analysis; and

(b) Iodine-125 as a sealed source in a portable imaging device.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.520 Availability of survey instrument.

A licensee authorized to use byproduct material as a sealed source for diagnostic purposes shall have available for use a portable radiation detection survey instrument capable of detecting dose rates over the range 0.1 millirem per hour to 100 millirem per hour or a portable radiation measurement survey instrument capable of measuring dose rates over the range 1 millirem per hour to 1000 millirem per hour. The instrument must have been calibrated in accordance with Section 35.51 of this part.

Sec. 35.590 Training for use of sealed sources for diagnosis.

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a diagnostic sealed source for the use in a device authorized under Sec. 35.500 to be a physician, dentist, or podiatrist who--

§ 35.590 Training for use of sealed sources for diagnosis.

Except as provided in § 35.57, the licensee shall require the authorized user of a diagnostic sealed source for use in a device authorized under § 35.500 to be a physician, dentist, or podiatrist who --

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(a) Is certified by a speciality board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission; or

(b) Has had 8 hours of classroom and laboratory training in basic radionuclide handling techniques specifically applicable to the use of the device that includes--

- (1) Radiation physics and instrumentation;
 - (2) Radiation protection;
 - (3) Mathematics pertaining to the use and measurement of radioactivity;
 - (4) Radiation biology;
- and
- (5) Training in the use of the device for the uses requested.

(a) Is certified by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or an Agreement State; or

(b) Has had 8 hours of classroom and laboratory training in basic radionuclide handling techniques specifically applicable to the use of the device; the training must include --

- (1) Radiation physics and instrumentation;
 - (2) Radiation protection;
 - (3) Mathematics pertaining to the use and measurement of radioactivity;
 - (4) Radiation biology;
- and
- (5) Training in the use of the device for the uses requested.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart I--Teletherapy

Subpart H--Therapeutic Medical Devices

Subpart H-- Photon Emitting Remote Afterloader Units, Teletherapy Units, and Gamma Stereotactic Radiosurgery Units

Section 35.600 Use of a sealed source in a teletherapy unit.

Sec. 35.600 Use of a sealed source in a device for therapeutic medical uses.

§ 35.600 Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit.

The regulations and provisions of this subpart govern the use of teletherapy units for medical use that contain a sealed source of cobalt-60 or cesium-137.

A licensee shall use sealed sources and devices for therapy as approved in the Sealed Source and Device Registry for medical use.

A licensee shall use sealed sources in photon emitting remote afterloader units, teletherapy units, or gamma stereotactic radiosurgery units for therapeutic medical uses:
 (a) As approved in the Sealed Source and Device Registry; or
 (b) In research in accordance with an effective Investigational Device Exemption (IDE) application accepted by the FDA provided the requirements of § 35.49(a) are met.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.604 Radiation surveys of patients and human research subjects treated with remote afterloaders.</p>	<p>(a) Before releasing a patient or a human research subject from licensee control, a licensee shall make a survey of the patient or the human research subject and the afterloader device with a portable radiation detection survey instrument to confirm that the source(s) has been removed from the patient or human research subject and returned to the safe shielded position.</p> <p>(b) A licensee shall retain a record of patient or human research subject surveys in accordance with Sec. 35.2404.</p>	<p>§ 35.604 Surveys of patients and human research subjects treated with a remote afterloader unit.</p> <p>(a) Before releasing a patient or a human research subject from licensee control, a licensee shall survey the patient or the human research subject and the remote afterloader unit with a portable radiation detection survey instrument to confirm that the source(s) has been removed from the patient or human research subject and returned to the safe shielded position.</p> <p>(b) A licensee shall retain a record of these surveys in accordance with § 35.2404.</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.605 Maintenance and repair restrictions.

Sec. 35.605 Installation, maintenance, and repair.

§ 35.605 Installation, maintenance, adjustment, and repair.

Only a person specifically licensed by the Commission or an Agreement State to perform teletherapy unit maintenance and repair shall:

(a) Install, relocate, or remove a teletherapy sealed source or a teletherapy unit that contains a sealed source; or

(b) Maintain, adjust, or repair the source drawer, the shutter or other mechanism of a teletherapy unit that could expose the source, reduce the shielding around the source, or result in increased radiation levels.

(a) Only a person specifically licensed by the Commission or an Agreement State shall install, maintain, adjust, or repair a device that involves work on the source(s) shielding, the source(s) driving unit, or other electronic or mechanical component that could expose the source, reduce the shielding around the source(s), or compromise the radiation safety of the device or the source(s).

(b) Except for low dose-rate remote afterloader devices, only a person specifically licensed by the Commission or an Agreement State shall install, replace, relocate, or remove a sealed source or source contained in a device,

(c) For a low dose-rate remote afterloader device, only a person specifically

(a) Only a person specifically licensed by the Commission or an Agreement State shall install, maintain, adjust, or repair a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit that involves work on the source(s) shielding, the source(s) driving unit, or other electronic or mechanical component that could expose the source(s), reduce the shielding around the source(s), or compromise the radiation safety of the unit or the source(s).

(b) Except for low dose-rate remote afterloader units, only a person specifically licensed by the Commission or an Agreement State shall install, replace, relocate, or remove a sealed source or source contained in other remote afterloader units, teletherapy units, or gamma stereotactic units.

(c) For a low dose-rate remote afterloader unit, only a person specifically

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

licensed by the Commission or an Agreement State or an authorized medical physicist shall perform the functions listed in paragraph (b) of this section.

(d) A licensee shall retain a record of the installation, maintenance, and repair done on therapeutic medical devices in accordance with Sec. 35.2605.

licensed by the Commission or an Agreement State or an authorized medical physicist shall install, replace, relocate, or remove a sealed source(s) contained in the unit.

(d) A licensee shall retain a record of the installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units in accordance with § 35.2605.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.606 License amendments.

In addition to the changes specified in Section 35.13 of this part, a licensee shall apply for and must receive a license amendment before:

(a) Making any change in the treatment room shielding;

(b) Making any change in the location of the teletherapy unit within the treatment room;

(c) Using the teletherapy unit in a manner that could result in increased radiation levels in areas outside the teletherapy treatment room;

(d) Relocating the teletherapy unit; or

(e) Allowing an individual not listed on the licensee's license to perform the duties of the teletherapy physicist.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.610 Safety instruction.

Sec. 35.610 Safety procedures and instructions for remote afterloaders, teletherapy units, and gamma stereotactic radiosurgery units.

§ 35.610 Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

(a) A licensee shall post instructions at the teletherapy unit console. To satisfy this requirement, these instructions must inform the operator of:

(a) A licensee shall develop, implement, and maintain written procedures for--

(a) A licensee shall --

(1) The procedure to be followed to ensure that only the patient or the human research subject is in the treatment room before turning the primary beam of radiation on to begin a treatment or after a door interlock interruption;

(1) Securing the device, the console, the console keys, and the treatment room when not in use or unattended;

(1) Secure the unit, the console, the console keys, and the treatment room when not in use or unattended;

(2) The procedure to be followed if:

(2) Except for low dose-rate remote afterloaders, ensuring that only the patient or the human research subject is in the treatment room before initiating treatment with the source(s), unless contraindicated, or after a door interlock interruption;

(2) Permit only individuals approved by the authorized user, Radiation Safety Officer, or authorized medical physicist to be present in the treatment room during treatment with the source(s) ;

(i) The operator is unable to turn the primary beam of radiation off with controls outside the treatment room or any other abnormal operation occurs; and

(3) Preventing dual operation of more than one radiation producing device in a treatment room if applicable; and

(3) Prevent dual operation of more than one radiation producing device in a treatment room if applicable;

(ii) The names and telephone numbers of the authorized users and Radiation Safety Officer to be

(4) Responding to an abnormal situation when the operator is unable to place the source(s) in the shielded position, or remove the patient

(4) Develop, implement, and maintain written procedures for responding to an abnormal situation when the operator is

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

immediately contacted if the teletherapy unit or console operates abnormally.

or human research subject from the radiation field with controls from outside the treatment room. This procedure must include--

(i) Instructions for responding to equipment failures and the names of the individuals responsible for implementing corrective actions;

(ii) The process for restricting access to and posting of the treatment area to minimize the risk of inadvertent exposure; and

(iii) The names and telephone numbers of the authorized users, the authorized medical physicist, and the Radiation Safety Officer to be contacted if the device or console operates abnormally.

(b) A copy of the procedures required by Sec. 35.610(a) must be physically located at the unit console.

unable to place the source(s) in the shielded position, or remove the patient or human research subject from the radiation field with controls from outside the treatment room. These procedures must include --

(i) Instructions for responding to equipment failures and the names of the individuals responsible for implementing corrective actions;

(ii) The process for restricting access to and posting of the treatment area to minimize the risk of inadvertent exposure; and

(iii) The names and telephone numbers of the authorized users, the authorized medical physicist, and the Radiation Safety Officer to be contacted if the unit or console operates abnormally.

(b) A copy of the procedures required by paragraph (a)(4) of this section must be physically located at the unit console.

(c) A licensee shall post instructions at the unit

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

console to inform the operator
of --

(1) The location of the
procedures required by
paragraph (a)(4) of this
section; and

(2) The names and
telephone numbers of the
authorized users, the
authorized medical physicist,
and the Radiation Safety
Officer to be contacted if the
unit or console operates
abnormally.

(d) A licensee shall
provide instruction, initially
and at least annually, to all
individuals who operate the
unit, as appropriate to the
individual's assigned duties, in

--

(1) The procedures
identified in paragraph (a)(4)
of this section; and

(2) The operating
procedures for the unit.

(e) A licensee shall
ensure that operators,
authorized medical physicists,
and authorized users
participate in drills of the
emergency procedures,
initially and at least annually.

(f) A licensee shall
retain a record of individuals

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

receiving instruction required by paragraph (d) of this section, in accordance with § 35.2310.

(c) A licensee shall post instructions at the device console to inform the operator of--

(1) The location of the procedures required by Sec. 35.610(a); and

(2) The names and telephone numbers of the authorized users, the authorized medical physicist, and the Radiation Safety Officer to be contacted if the device or console unit or console operates abnormally.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) A licensee shall provide instruction in the topics identified in paragraph (a) of this section to all individuals who operate a teletherapy unit.

(d) A licensee shall provide instruction and practice drills, initially and at least annually, in the procedures identified in paragraph (a) of this section and the operating procedures to all individuals who operate the device, as appropriate to the individual's assigned duties. A licensee shall ensure that operators receive refresher training in the operation of the unit and procedures for periodic spot-checks and full calibrations; and that operators, authorized medical physicists, and authorized users participate in drills of the emergency procedures.

(c) A licensee shall retain for three years a record of individuals receiving instruction required by paragraph (b) of this section, a description of the instruction, the date of instruction, and the name of the individual who gave the instruction.

(e) A licensee shall retain a record of individuals receiving instruction required by paragraph (d) of this section, in accordance with Sec. 35.2310.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.615 Safety precautions.

Sec. 35.615 Safety precautions for remote afterloaders, teletherapy units, and gamma stereotactic radiosurgery units.

§ 35.615 Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

(a) A licensee shall control access to the teletherapy room by a door at each entrance.

(a) A licensee shall control access to the treatment room by a door at each entrance.

(a) A licensee shall control access to the treatment room by a door at each entrance.

(b) A licensee shall equip each entrance to the teletherapy room with an electrical interlock system that will:

(b) A licensee shall equip each entrance to the treatment room with an electrical interlock system that will--

(b) A licensee shall equip each entrance to the treatment room with an electrical interlock system that will --

(1) Prevent the operator from turning the primary beam of radiation on unless each treatment room entrance door is closed;

(1) Prevent the operator from initiating the treatment cycle unless each treatment room entrance door is closed;

(1) Prevent the operator from initiating the treatment cycle unless each treatment room entrance door is closed;

(2) Turn the primary beam of radiation off immediately when an entrance door is opened; and

(2) Cause the sources to be shielded immediately when an entrance door is opened; and

(2) Cause the source(s) to be shielded when an entrance door is opened; and

(3) Prevent the primary beam of radiation from being turned on following an interlock interruption until all treatment room entrance doors are closed and the beam on-off control is reset at the console.

(3) Prevent the primary beam of radiation from being turned on following an interlock interruption until all treatment room entrance doors are closed and the beam on-off control is reset at the console.

(3) Prevent the source(s) from being exposed following an interlock interruption until all treatment room entrance doors are closed and the source(s) on-off control is reset at the console.

(c) A licensee shall equip each entrance to the teletherapy room with a beam condition indicator light.

(c) A licensee shall require any individual entering the treatment room to assure,

(c) A licensee shall require any individual entering the treatment room to assure,

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(d) A licensee shall install in each teletherapy room a permanent radiation monitor capable of continuously monitoring beam status.

(1) A radiation monitor must provide visible notice of a teletherapy unit malfunction that results in an exposed or partially exposed source, and must be observable by an individual entering the teletherapy room.

(2) A radiation monitor must be equipped with a backup power supply separate from the power supply to the teletherapy unit. This backup power supply may be a battery system.

(3) A radiation monitor must be checked with a dedicated check source for proper operation each day before the teletherapy unit is used for treatment of patients or human research subjects.

(4) A licensee shall maintain a record of the check required by paragraph (d)(3) of this section for three years. The record must include the date of the check, notation that the monitor indicates when its detector is and is not exposed,

through the use of appropriate radiation monitors, that radiation levels have returned to ambient levels.

through the use of appropriate radiation monitors, that radiation levels have returned to ambient levels.

(d) Except for low-dose remote afterloader units, a licensee shall construct or equip each treatment room with viewing and intercom systems to permit continuous observation of the patient or the human research subject from the treatment console during irradiation.

(e) For licensed activities where sources are placed within the patient's or human research subject's body, a licensee shall only conduct treatments which allow for expeditious removal of a decoupled or jammed source.

(f) In addition to the requirements specified in paragraphs (a) through (e) of this section, a licensee shall --

(1) For medium dose-rate and pulsed dose-rate remote afterloader units, require --

(i) An authorized medical physicist and either an authorized user or a physician, under the supervision of an

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

and the initials of the individual who performed the check.

(5) If a radiation monitor is inoperable, the licensee shall require any individual entering the teletherapy room to use a survey instrument or audible alarm personal dosimeter to monitor for any malfunction of the source exposure mechanism that may result in an exposed or partially exposed source. The instrument or dosimeter must be checked with a dedicated check source for proper operation at the beginning of each day of use. The licensee shall keep a record as described in paragraph (d)(4) of this section.

(6) A licensee shall promptly repair or replace the radiation monitor if it is inoperable.

authorized user, who has been trained in the operation and emergency response for the unit to be physically present during the initiation of all patient treatments involving the unit; and

(ii) An authorized medical physicist and either an authorized user or an individual, under the supervision of an authorized user, who has been trained to remove the source applicator(s) in the event of an emergency involving the unit, to be immediately available during continuation of all patient treatments involving the unit.

(2) For high dose-rate remote afterloader units, require --

(i) An authorized user and an authorized medical physicist to be physically present during the initiation of all patient treatments involving the unit; and

(ii) An authorized medical physicist and either an authorized user or a physician, under the supervision of an authorized user, who has been trained in the operation and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

emergency response for the unit, to be physically present during continuation of all patient treatments involving the unit.

(3) For gamma stereotactic radiosurgery units, require an authorized user and an authorized medical physicist to be physically present throughout all patient treatments involving the unit.

(4) Notify the Radiation Safety Officer, or his/her designee, and an authorized user as soon as possible if the patient or human research subject has a medical emergency or dies.

(g) A licensee shall have applicable emergency response equipment available near each treatment room to respond to a source --

(1) Remaining in the unshielded position; and

(2) Lodged within the patient following completion of the treatment.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(e) A licensee shall construct or equip each teletherapy room to permit continuous observation of the patient or the human research subject from the teletherapy unit console during irradiation.

PROPOSED RULE

(d) Except for low-dose remote afterloaders, a licensee shall construct or equip each treatment room with viewing and intercom systems to permit continuous observation of the patient or the human research subject from the treatment console during irradiation.

(e) For licensed activities where sources are placed within the patient's or human research subject's body, a licensee shall only conduct treatments which allow for expeditious removal of a decoupled or jammed source.

(f) In addition to the requirements specified in paragraphs (a) through (e) of this section, a licensee shall--

(1) For low dose-rate remote afterloader devices, require--

(i) An authorized user or an authorized medical physicist to be physically present during the initiation of all patient treatments involving the device; and

(ii) An authorized medical physicist and an authorized user or a

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

physician, who has been designated by the authorized user and who is a radiation oncology physician trained in emergency response for the device, to be immediately available during continuation of all patient treatments involving the device.

(2) For high dose-rate remote afterloader devices, require--

(i) An authorized user and an authorized medical physicist to be physically present during the initiation of all patient treatments involving the device; and

(ii) An authorized medical physicist and an authorized user or a physician, who has been designated by the authorized user and who is a radiation oncology physician that has been trained in emergency response for the device, to be physically present during continuation of all patient treatments involving the device.

(3) For pulsed dose-rate remote afterloader devices, require--

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(i) An authorized user and an authorized medical physicist to be physically present during the initiation of all patient treatments involving the device; and

(ii) An authorized medical physicist and an authorized user or a physician, who has been designated by the authorized user and who is a radiation oncology physician that has been trained in emergency response for the device, to be immediately available during continuation of all patient treatments involving the device.

(4) For gamma stereotactic radiosurgery units, require an authorized user and an authorized medical physicist to be physically present throughout all patient treatments involving the unit.

(g) The licensee shall have emergency response equipment available near each treatment room. The emergency response equipment must include, as applicable--

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

- (1) A device to assist in placing the source(s) in the shielded position;
- (2) A shielded source/applicator storage container;
- (3) Remote handling tools; and
- (4) Supplies necessary to surgically remove applicators or sources from a patient or human research subject treated internally with sealed sources.

Section 35.620 Possession of survey instrument.

A licensee authorized to use byproduct material in a teletherapy unit shall have in its possession either a portable radiation detection survey instrument capable of detecting dose rate over the range 0.1 millirem per hour to 100 millirem per hour or a portable radiation measurement survey instrument capable of measuring dose rates over the range 1 millirem per hour to 1,000 millirem per hour.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.630 Dosimetry equipment.

Sec. 35.630 Dosimetry equipment.

§ 35.630 Dosimetry equipment.

(a) A licensee shall have a calibrated dosimetry system available for use. To satisfy this requirement, one of the following two conditions must be met.

(a) A licensee shall have a calibrated dosimetry system available for use. To satisfy this requirement, one of the following two conditions must be met.

(a) Except for low dose-rate remote afterloader sources where the source output or activity is determined by the manufacturer, a licensee shall have a calibrated dosimetry system available for use. To satisfy this requirement, one of the following two conditions must be met.

(1) The system must have been calibrated by the National Institute of Standards and Technology or by a calibration laboratory accredited by the American Association of Physicists in Medicine (AAPM). The calibration must have been performed within the previous two years and after any servicing that may have affected system calibration; or

(1) The system must have been calibrated using a source traceable to the National Institute of Standards and Technology and published protocols approved by nationally recognized bodies or by a calibration laboratory accredited by the American Association of Physicists in Medicine (AAPM). The calibration must have been performed within the previous 2 years and after any servicing that may have affected system calibration; or

(1) The system must have been calibrated using a system or source traceable to the National Institute of Standards and Technology (NIST) and published protocols accepted by nationally recognized bodies; or by a calibration laboratory accredited by the American Association of Physicists in Medicine (AAPM). The calibration must have been performed within the previous 2 years and after any servicing that may have affected system calibration; or

(2) The system must have been calibrated within the previous four years; eighteen to thirty months after that

(2) The system must have been calibrated within the previous 4 years; 18 to 30 months after that calibration,

(2) The system must have been calibrated within the previous 4 years. Eighteen to thirty months after that

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

calibration, the system must have been intercompared at an intercomparison meeting with another dosimetry system that was calibrated within the past twenty-four months by the National Institute of Standards and Technology or by a calibration laboratory accredited by the AAPM. The intercomparison meeting must be sanctioned by a calibration laboratory or radiologic physics center accredited by the AAPM. The results of the intercomparison meeting must have indicated that the calibration factor of the licensee's system had not changed by more than 2 percent. The licensee may not use the intercomparison result to change the calibration factor. When intercomparing dosimetry systems to be used for calibrating cobalt-60 teletherapy units, the licensee shall use a teletherapy unit with a cobalt-60 source. When intercomparing dosimetry systems to be used for calibrating cesium-137 teletherapy units, the licensee shall use a teletherapy unit with a cesium-137 source.

the system must have been intercompared with another dosimetry system that was calibrated within the past 24 months by the National Institute of Standards and Technology or by a calibration laboratory accredited by the AAPM. The results of the intercomparison must have indicated that the calibration factor of the licensee's system had not changed by more than 2 percent. The licensee may not use the intercomparison result to change the calibration factor. When intercomparing dosimetry systems to be used for calibrating sealed sources for therapeutic devices, the licensee shall use a comparable device with beam attenuators or collimators, as applicable, and sources of the same radionuclide as the source used at the licensee's facility.

calibration, the system must have been intercompared with another dosimetry system that was calibrated within the past 24 months by NIST or by a calibration laboratory accredited by the AAPM. The results of the intercomparison must indicate that the calibration factor of the licensee's system had not changed by more than 2 percent. The licensee may not use the intercomparison result to change the calibration factor. When intercomparing dosimetry systems to be used for calibrating sealed sources for therapeutic units, the licensee shall use a comparable unit with beam attenuators or collimators, as applicable, and sources of the same radionuclide as the source used at the licensee's facility.

(b) The licensee shall have a dosimetry system available for use for spot-check output measurements, if applicable. To satisfy this requirement, the system may be compared with a system that has been

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

calibrated in accordance with paragraph (a) of this section. This comparison must have been performed within the previous year and after each servicing that may have affected system calibration. The spot-check system may be the same system used to meet the requirement in paragraph (a) of this section.

(c) The licensee shall retain a record of each calibration, intercomparison, and comparison in accordance with § 35.2630.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) The licensee shall have available for use a dosimetry system for spot-check measurements. To satisfy this requirement, the system may be compared with a system that has been calibrated in accordance with paragraph (a) of this section. This comparison must have been performed within the previous year and after each servicing that may have affected system calibration. The spot-check system may be the same system used to meet the requirement in paragraph (a) of this section.

(c) The licensee shall retain a record of each calibration, intercomparison, and comparison for the duration of the license. For each calibration, intercomparison, or comparison, the record must include the date, the model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by paragraphs (a) and (b) of this section, the correction factor that was determined from the calibration or comparison or

(b) The licensee shall have available for use a dosimetry system for spot-check measurements. To satisfy this requirement, the system may be compared with a system that has been calibrated in accordance with paragraph (a) of this section. This comparison must have been performed within the previous year and after each servicing that may have affected system calibration. The spot-check system may be the same system used to meet the requirement in paragraph (a) of this section.

(c) The licensee shall retain a record of each calibration, intercomparison, and comparison in accordance with Sec. 35.2630.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

the apparent correction factor that was determined from an intercomparison, the names of the individuals who performed the calibration, intercomparison, or comparison, and evidence that the intercomparison meeting was sanctioned by a calibration laboratory or radiologic physics center accredited by AAPM

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.632 Full calibration measurements.

Sec. 35.632 Full calibration measurements on teletherapy units.

§ 35.632 Full calibration measurements on teletherapy units.

(a) A licensee authorized to use a teletherapy unit for medical use shall perform full calibration measurements on each teletherapy unit:

(a) A licensee authorized to use a teletherapy unit for medical use shall perform full calibration measurements on each teletherapy unit--

(a) A licensee authorized to use a teletherapy unit for medical use shall perform full calibration measurements on each teletherapy unit --

(1) Before the first medical use of the unit; and

(1) Before the first medical use of the unit; and

(1) Before the first medical use of the unit; and

(2) Before medical use under the following conditions:

(2) Before medical use under the following conditions:

(2) Before medical use under the following conditions:

(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;

(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;

(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;

(ii) Following replacement of the source or following reinstallation of the teletherapy unit in a new location;

(ii) Following replacement of the source or following reinstallation of the teletherapy unit in a new location;

(ii) Following replacement of the source or following reinstallation of the teletherapy unit in a new location;

(iii) Following any repair of the teletherapy unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and

(iii) Following any repair of the teletherapy unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and

(iii) Following any repair of the teletherapy unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(3) At intervals not exceeding one year.
(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of:
(1) The output within +/-3 percent for the range of field sizes and for the distance or range of distances used for medical use;
(2) The coincidence of the radiation field and the field indicated by the light beam localizing device;
(3) The uniformity of the radiation field and its dependence on the orientation of the useful beam;
(4) Timer constancy and linearity over the range of use;
(5) On-off error; and
(6) The accuracy of all distance measuring and localization devices in medical use.
(c) A licensee shall use the dosimetry system described in Section 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in

(3) At intervals not exceeding 1 year.
(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of--
(1) The output within +/-3 percent for the range of field sizes and for the distance or range of distances used for medical use;
(2) The coincidence of the radiation field and the field indicated by the light beam localizing device;
(3) The uniformity of the radiation field and its dependence on the orientation of the useful beam;
(4) Timer accuracy and linearity over the range of use;
(5) On-off error; and
(6) The accuracy of all distance measuring and localization devices in medical use.
(c) A licensee shall use the dosimetry system described in Sec. 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in

(3) At intervals not exceeding 1 year.
(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of --
(1) The output within +/-3 percent for the range of field sizes and for the distance or range of distances used for medical use;
(2) The coincidence of the radiation field and the field indicated by the light beam localizing device;
(3) The uniformity of the radiation field and its dependence on the orientation of the useful beam;
(4) Timer constancy and linearity over the range of use;
(5) On-off error; and
(6) The accuracy of all distance measuring and localization devices in medical use.
(c) A licensee shall use the dosimetry system described in § 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.

(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with either the procedures recommended by the Scientific Committee on Radiation Dosimetry of the American Association of Physicists in Medicine that are described in Physics in Medicine and Biology Vol. 16, No. 3, 1971, pp. 379-396, or by Task Group 21 of the Radiation Therapy Committee of the American Association of Physicists in Medicine that are described in Medical Physics Vol. 10, No. 6, 1983, pp. 741-771, and Vol. 11, No. 2, 1984, p. 213. (Both of these references have been approved for incorporation by reference by the Director of the Federal Register. Copies of the documents are available for inspection at the NRC Library, 11545 Rockville Pike, Rockville, Maryland 20852-2738. Copies of the documents are also on file at

paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.

(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols approved by nationally recognized bodies.

paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.

(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.

(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section for physical decay for intervals not exceeding 1 month for cobalt-60, 6 months for cesium-137, or at intervals consistent with 1 percent decay for all other nuclides.

(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.

(g) A licensee shall retain a record of each calibration in accordance with § 35.2632.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

the Office of the Federal Register, 800 North Capitol Street NW., suite 700, Washington, DC. A notice of any change in the material will be published in the Federal Register.)

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(e) A licensee shall correct mathematically the outputs determined in paragraph (b)(1) of this section for physical decay for intervals not exceeding one month for cobalt-60 or six months for cesium-137.

(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the licensee's teletherapy physicist.

(g) A licensee shall retain a record of each calibration for the duration of use of the teletherapy unit source. The record must include the date of the calibration, the manufacturer's name, model number, and serial number for both the teletherapy unit and the source, the model numbers and serial numbers of the instruments used to calibrate the teletherapy unit, tables that describe the output of the unit over the range of field sizes and for the range of distances

(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section for physical decay for intervals not exceeding 1 month for cobalt-60, 6 months for cesium-137, or at intervals consistent with 1 percent decay for all other nuclides.

(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.

(g) A licensee shall retain a record of each calibration in accordance with Sec. 35.2632.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

used in radiation therapy, a determination of the coincidence of the radiation field and the field indicated by the light beam localizing device, an assessment of timer linearity and constancy, the calculated on-off error, the estimated accuracy of each distance measuring or localization device, and the signature of the teletherapy physicist.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.633 Full calibration measurements on remote afterloaders.</p> <p>(a) A licensee authorized to use a remote afterloader for medical use shall perform full calibration measurements on each unit—</p> <p>(1) Before the first medical use of the unit;</p> <p>(2) Before medical use under the following conditions:</p> <p>(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;</p> <p>(ii) Following replacement of the source or following reinstallation of the unit in a new location outside the facility; and</p> <p>(iii) Following any repair of the unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and</p> <p>(3) At intervals not exceeding 120 days for high</p>	<p>§ 35.633 Full calibration measurements on remote afterloader units.</p> <p>(a) A licensee authorized to use a remote afterloader unit for medical use shall perform full calibration measurements on each unit --</p> <p>(1) Before the first medical use of the unit;</p> <p>(2) Before medical use under the following conditions:</p> <p>(i) Following replacement of the source or following reinstallation of the unit in a new location outside the facility; and</p> <p>(ii) Following any repair of the unit that includes removal of the source or major repair of the components associated with the source exposure assembly; and</p> <p>(3) At intervals not exceeding 1 quarter for high</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>dose-rate and pulsed dose-rate remote afterloaders; and</p> <p>(4) At intervals not exceeding 1 year for low dose-rate remote afterloaders.</p> <p>(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of:</p> <p>(1) The output within +/-5 percent;</p> <p>(2) Source positioning accuracy to within +/-1 millimeter;</p> <p>(3) Source retraction with backup battery upon power failure; and</p> <p>(4) The operability of the electrically assisted treatment room doors with the high-dose rate remote afterloader unit electrical power turned off.</p>	<p>dose-rate, medium dose-rate, and pulsed dose-rate remote afterloader units with sources whose half-life exceeds 75 days; and</p> <p>(4) At intervals not exceeding 1 year for low dose-rate remote afterloader units.</p> <p>(b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include, as applicable, determination of:</p> <p>(1) The output within +/- 5 percent;</p> <p>(2) Source positioning accuracy to within +/- 1 millimeter;</p> <p>(3) Source retraction with backup battery upon power failure;</p> <p>(4) Length of the source transfer tubes;</p> <p>(5) Timer accuracy and linearity over the typical range of use;</p> <p>(6) Length of the applicators; and</p> <p>(7) Function of the source transfer tubes, applicators, and transfer tube-applicator interfaces.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(c) In addition to the requirements for full calibrations for all remote afterloaders in paragraph (b) of this section, a licensee shall:

(1) For high dose-rate and pulsed dose-rate remote afterloaders, calibrate--

(i) At intervals not exceeding one quarter:

(A) The source guide tubes;

(B) Timer accuracy and linearity over the typical range of use; and

(C) Length of the connectors; and

(ii) Annually, the function of the source tube guides and connectors.

(2) For low dose-rate remote afterloaders, perform an autoradiograph of the source(s) to verify inventory and source(s) arrangement and a spot check of the absolute timer accuracy at intervals not exceeding one quarter.

(d) A licensee shall use the dosimetry system described in Sec. 35.630(a) to measure the output.

(c) A licensee shall use the dosimetry system described in § 35.630(a) to measure the output.

(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.

(e) In addition to the requirements for full calibrations for low dose-rate remote afterloader units in paragraph (b) of this section, a licensee shall perform an autoradiograph of the source(s) to verify inventory and source(s) arrangement at intervals not exceeding 1 quarter.

(f) For low dose-rate remote afterloader units, a licensee may use measurements provided by the source manufacturer that are made in accordance with paragraphs (a) through (e) of this section.

(g) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section for physical decay at intervals

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(e) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols approved by nationally recognized bodies.

(f) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section for physical decay at intervals consistent with 1 percent physical decay.

(g) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (f) of this section must be performed by the authorized medical physicist.

(h) A licensee shall retain a record of each calibration in accordance with Sec. 35.2633.

consistent with 1 percent physical decay.

(h) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (g) of this section must be performed by the authorized medical physicist.

(i) A licensee shall retain a record of each calibration in accordance with § 35.2632.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.635 Full calibration measurements on gamma stereotactic radiosurgery units.</p> <p>(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform full calibration measurements on each unit—</p> <p>(1) Before the first medical use of the unit;</p> <p>(2) Before medical use under the following conditions--</p> <p>(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;</p> <p>(ii) Following replacement of the sources or following reinstallation of the gamma stereotactic radiosurgery unit in a new location; and</p> <p>(iii) Following any repair of the gamma stereotactic radiosurgery unit that includes removal of the</p>	<p>§ 35.635 Full calibration measurements on gamma stereotactic radiosurgery units.</p> <p>(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform full calibration measurements on each unit --</p> <p>(1) Before the first medical use of the unit;</p> <p>(2) Before medical use under the following conditions --</p> <p>(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;</p> <p>(ii) Following replacement of the sources or following reinstallation of the gamma stereotactic radiosurgery unit in a new location; and</p> <p>(iii) Following any repair of the gamma stereotactic radiosurgery unit that includes removal of the</p>	<p>§ 35.635 Full calibration measurements on gamma stereotactic radiosurgery units.</p> <p>(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform full calibration measurements on each unit --</p> <p>(1) Before the first medical use of the unit;</p> <p>(2) Before medical use under the following conditions --</p> <p>(i) Whenever spot-check measurements indicate that the output differs by more than 5 percent from the output obtained at the last full calibration corrected mathematically for radioactive decay;</p> <p>(ii) Following replacement of the sources or following reinstallation of the gamma stereotactic radiosurgery unit in a new location; and</p> <p>(iii) Following any repair of the gamma stereotactic radiosurgery unit that includes removal of the</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>sources or major repair of the components associated with the source assembly; and (3) At intervals not exceeding 1 year. (b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of-- (1) The output within +/-3 percent; (2) Relative helmet factors; (3) Isocenter coincidence; (4) Timer accuracy and linearity over the range of use; (5) On-off error; and (6) Trunnion centricity.</p> <p>(c) A licensee shall use the dosimetry system</p>	<p>sources or major repair of the components associated with the source assembly; and (3) At intervals not exceeding 1 year. (b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of -- (1) The output within +/-3 percent; (2) Relative helmet factors; (3) Isocenter coincidence; (4) Timer accuracy and linearity over the range of use; (5) On-off error; (6) Trunnion centricity; (7) Treatment table retraction mechanism, using backup battery power or hydraulic backups with the unit off; (8) Helmet microswitches; (9) Emergency timing circuits; and (10) Stereotactic frames and localizing devices (trunnions).</p> <p>(c) A licensee shall use the dosimetry system</p>	<p>sources or major repair of the components associated with the source assembly; and (3) At intervals not exceeding 1 year. (b) To satisfy the requirement of paragraph (a) of this section, full calibration measurements must include determination of -- (1) The output within +/-3 percent; (2) Relative helmet factors; (3) Isocenter coincidence; (4) Timer accuracy and linearity over the range of use; (5) On-off error; (6) Trunnion centricity; (7) Treatment table retraction mechanism, using backup battery power or hydraulic backups with the unit off; (8) Helmet microswitches; (9) Emergency timing circuits; and (10) Stereotactic frames and localizing devices (trunnions).</p> <p>(c) A licensee shall use the dosimetry system</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>described in Sec. 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.</p> <p>(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols approved by nationally recognized bodies.</p> <p>(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section at intervals not exceeding 1 month for cobalt-60 and at intervals consistent with 1 percent physical decay for all other radionuclides.</p> <p>(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.</p>	<p>described in § 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.</p> <p>(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.</p> <p>(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section at intervals not exceeding 1 month for cobalt-60 and at intervals consistent with 1 percent physical decay for all other radionuclides.</p> <p>(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.</p>	<p>described in § 35.630(a) to measure the output for one set of exposure conditions. The remaining radiation measurements required in paragraph (b)(1) of this section may be made using a dosimetry system that indicates relative dose rates.</p> <p>(d) A licensee shall make full calibration measurements required by paragraph (a) of this section in accordance with published protocols accepted by nationally recognized bodies.</p> <p>(e) A licensee shall mathematically correct the outputs determined in paragraph (b)(1) of this section at intervals not exceeding 1 month for cobalt-60 and at intervals consistent with 1 percent physical decay for all other radionuclides.</p> <p>(f) Full calibration measurements required by paragraph (a) of this section and physical decay corrections required by paragraph (e) of this section must be performed by the authorized medical physicist.</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(g) A licensee shall retain a record of each calibration in accordance with Sec. 35.2635.

(g) A licensee shall retain a record of each calibration in accordance with § 35.2632.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.634 Periodic spot-checks.

Sec. 35.642 Periodic spot-checks for teletherapy units.

§ 35.642 Periodic spot-checks for teletherapy units.

(a) A licensee authorized to use teletherapy units for medical use shall perform output spot-checks on each teletherapy unit once in each calendar month that include determination of:

(a) A licensee authorized to use teletherapy units for medical use shall perform output spot-checks on each teletherapy unit once in each calendar month that include determination of--

(a) A licensee authorized to use teletherapy units for medical use shall perform output spot-checks on each teletherapy unit once in each calendar month that include determination of --

- (1) Timer constancy, and timer linearity over the range of use;
- (2) On-off error;
- (3) The coincidence of the radiation field and the field indicated by the light beam localizing device;
- (4) The accuracy of all distance measuring and localization devices used for medical use;
- (5) The output for one typical set of operating conditions measured with the dosimetry system described in Section 35.630(b) of this part; and
- (6) The difference between the measurement made in paragraph (b)(5) of this section and the anticipated output, expressed as a

- (1) Timer constancy, and timer linearity over the range of use;
- (2) On-off error;
- (3) The coincidence of the radiation field and the field indicated by the light beam localizing device;
- (4) The accuracy of all distance measuring and localization devices used for medical use;
- (5) The output for one typical set of operating conditions measured with the dosimetry system described in Sec. 35.630(b); and
- (6) The difference between the measurement made in paragraph (b)(5) of this section and the anticipated output, expressed

- (1) Timer constancy, and timer linearity over the range of use;
- (2) On-off error;
- (3) The coincidence of the radiation field and the field indicated by the light beam localizing device;
- (4) The accuracy of all distance measuring and localization devices used for medical use;
- (5) The output for one typical set of operating conditions measured with the dosimetry system described in § 35.630(b); and
- (6) The difference between the measurement made in paragraph (b)(5) of this section and the anticipated output, expressed

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

percentage of the anticipated output (i.e., the value obtained at last full calibration corrected mathematically for physical decay).

(b) A licensee shall perform measurements required by paragraph (a) of this section in accordance with procedures established by the teletherapy physicist. That individual need not actually perform the spotcheck measurements.

(c) A licensee shall have the teletherapy physicist review the results of each spot-check within 15 days. The teletherapy physicist shall promptly notify the licensee in writing of the results of each spot-check. The licensee shall keep a copy of each written notification for three years.

(d) A licensee authorized to use a teletherapy unit for medical use shall perform safety spot-checks of each teletherapy facility once in each calendar month that assure proper operation of:

as a percentage of the anticipated output (i.e., the value obtained at last full calibration corrected mathematically for physical decay).

(b) A licensee shall perform measurements required by paragraph (a) of this section in accordance with written procedures established by the authorized medical physicist. That individual need not actually perform the spot check measurements.

(c) A licensee shall have the authorized medical physicist review the results of each spot-check within 15 days. The authorized medical physicist shall promptly notify the licensee in writing of the results of each spot-check.

(d) A licensee authorized to use a teletherapy unit for medical use shall perform safety spot-checks of each teletherapy facility once in each calendar month and after each source installation to assure proper operation of--

as a percentage of the anticipated output (i.e., the value obtained at last full calibration corrected mathematically for physical decay).

(b) A licensee shall perform measurements required by paragraph (a) of this section in accordance with written procedures established by the authorized medical physicist. That individual need not actually perform the spot-check measurements.

(c) A licensee shall have the authorized medical physicist review the results of each spot-check within 15 days. The authorized medical physicist shall notify the licensee as soon as possible in writing of the results of each spot-check.

(d) A licensee authorized to use a teletherapy unit for medical use shall perform safety spot-checks of each teletherapy facility once in each calendar month and after each source installation to assure proper operation of --

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(1) Electrical interlocks at each teletherapy room entrance;

(2) Electrical or mechanical stops installed for the purpose of limiting use of the primary beam of radiation (restriction of source housing angulation or elevation, carriage or stand travel and operation of the beam on-off mechanism);

(3) Beam condition indicator lights on the teletherapy unit, on the control console, and in the facility;

(4) Viewing systems;

(5) Treatment room doors from inside and outside the treatment room; and

(6) Electrically assisted treatment room doors with the teletherapy unit electrical power turned off.

(e) A licensee shall arrange for prompt repair of any system identified in paragraph (d) of this section that is not operating properly, and shall not use the teletherapy unit following door interlock malfunction until the interlock system has been repaired.

(1) Electrical interlocks at each teletherapy room entrance;

(2) Electrical or mechanical stops installed for the purpose of limiting use of the primary beam of radiation (restriction of source housing angulation or elevation, carriage or stand travel and operation of the beam on-off mechanism);

(3) Source exposure indicator lights on the teletherapy unit, on the control console, and in the facility;

(4) Viewing and intercom systems;

(5) Treatment room doors from inside and outside the treatment room; and

(6) Electrically assisted treatment room doors with the teletherapy unit electrical power turned off.

(e) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(1) Electrical interlocks at each teletherapy room entrance;

(2) Electrical or mechanical stops installed for the purpose of limiting use of the primary beam of radiation (restriction of source housing angulation or elevation, carriage or stand travel and operation of the beam on-off mechanism);

(3) Source exposure indicator lights on the teletherapy unit, on the control console, and in the facility;

(4) Viewing and intercom systems;

(5) Treatment room doors from inside and outside the treatment room; and

(6) Electrically assisted treatment room doors with the teletherapy unit electrical power turned off.

(e) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(f) A licensee shall retain a record of each spot-check required by paragraphs (a) and (d) of this section for three years. The record must include the date of the spot-check, the manufacturer's name, model number, and serial number for both the teletherapy unit and source, the manufacturer's name, model number and serial number of the instrument used to measure the output of the teletherapy unit, an assessment of timer linearity and constancy, the calculated on-off error, a determination of the coincidence of the radiation field and the field indicated by the light beam localizing device, the calculated on-off error, the determined accuracy of each distance measuring or localization device, the difference between the anticipated output and the measured output, notations indicating the operability of each entrance door electrical interlock, each electrical or mechanical stop, each beam condition indicator light, the viewing system and doors, and the signature of the individual

(f) A licensee shall retain a record of each spot-check required by paragraphs (a) and (d), in accordance with Sec. 35.2642

(f) A licensee shall retain a record of each spot-check required by paragraphs (a) and (d) of this section, in accordance with § 35.2642.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

who performed the periodic spot-check.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.636 Safety checks for teletherapy facilities.

(a) A licensee shall promptly check all systems listed in Section 35.634(d) for proper function after each installation of a teletherapy source and after making any change for which an amendment is required by Section 35.606 (a) through (d).

(b) If the results of the checks required in paragraph (a) of this section indicate the malfunction of any system specified in Section 35.634(d), the licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(c) A licensee shall retain for three years a record of the facility checks following installation of a source. The record must include notations indicating the operability of each entrance door interlock, each electrical or mechanical stop, each beam condition indicator light, the viewing system, and doors, and the

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

signature of the Radiation
Safety Officer.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.643 Periodic spot-checks for high dose-rate and pulsed dose-rate remote afterloaders.</p>	<p>§ 35.643 Periodic spot-checks for remote afterloader units.</p>	
<p>(a) A licensee authorized to use high dose-rate or pulsed dose-rate remote afterloaders for medical use shall perform spot-checks on each unit:</p> <ul style="list-style-type: none"> (1) At the beginning of each week of use; (2) At the beginning of each day of use; and (3) After each source installation. <p>(b) The licensee shall have the authorized medical physicist:</p> <ul style="list-style-type: none"> (1) Establish written procedures for performing the spot-checks required in paragraph (a) of this section; and (2) Review the results of each spot-check required by paragraph (a)(1) of this section within 15 days of the check. The authorized medical physicist need not actually perform the spot-check measurements. <p>(c) To satisfy the requirements of paragraphs</p>	<p>(a) A licensee authorized to use a remote afterloader unit for medical use shall perform spot-checks of each remote afterloader facility and on each unit:</p> <ul style="list-style-type: none"> (1) Prior to the first use of a high dose-rate, medium dose-rate, or pulsed dose-rate remote afterloader unit on a given day; (2) Before each patient treatment with a low dose-rate remote afterloader unit; and (3) After each source installation. <p>(b) A licensee shall perform the measurements required by paragraph (a) of this section in accordance with written procedures established by the authorized medical physicist. That individual need not actually perform the spot check measurements.</p> <p>(c) A licensee shall have the authorized medical physicist review the results of each spot-check within 15</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>(a)(1) of this section, spot-checks must, at a minimum--</p> <p>(1) Verify source positioning accuracy;</p> <p>(2) Determine output with the dosimetry system described in Sec. 35.630(b); and</p> <p>(3) Calculate the difference between the measurement made in paragraph (c)(2) of this section and the anticipated output, expressed as a percentage of the anticipated output (i.e., the value obtained at last full calibration mathematically corrected for physical decay).</p> <p>(d) To satisfy the requirements of paragraphs (a)(2) and (a)(3) of this section, spot-checks must, at a minimum, assure proper operation of--</p> <p>(1) Electrical interlocks at each remote afterloader room entrance;</p> <p>(2) Source exposure indicator lights on the remote afterloader unit, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems;</p>	<p>days. The authorized medical physicist shall notify the licensee as soon as possible in writing of the results of each spot-check.</p> <p>(d) To satisfy the requirements of paragraph (a) of this section, spot-checks must, at a minimum, assure proper operation of --</p> <p>(1) Electrical interlocks at each remote afterloader unit room entrance;</p> <p>(2) Source exposure indicator lights on the remote afterloader unit, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems in each high dose-rate, medium dose-rate, and pulsed dose-rate remote afterloader facility;</p> <p>(4) Emergency response equipment;</p> <p>(5) Radiation monitors used to indicate the source position;</p> <p>(6) Timer accuracy;</p> <p>(7) Clock (date and time) in the unit's computer; and</p> <p>(8) Decayed source(s) activity in the unit's computer.</p> <p>(e) If the results of the checks required in paragraph</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(4) Emergency response equipment;
(5) Radiation monitors used to indicate the source position;
(6) Timer constancy;
and
(7) Clock (date and time) in the unit's computer.

(e) In addition to the requirements for spot checks in paragraph (d) of this section, a licensee shall ensure overall proper operation of the unit by conducting a simulated cycle of treatment as part of the spot-checks.

(f) A licensee shall arrange for prompt repair of any system identified in paragraph (c) of this section that is not operating.

(g) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(h) A licensee shall retain a record of each check

(d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(f) A licensee shall retain a record of each check required by paragraph (d) of this section in accordance with § 35.2643.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

required by paragraphs (c)
and (d) of this section in
accordance with Sec.
35.2643.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.644 Periodic spot-checks for low dose-rate remote afterloaders.

(a) A licensee authorized to use low dose-rate remote afterloaders for medical use shall perform spot-checks on each unit prior to each patient treatment and after each source installation that include proper operation of--

- (1) Electrical interlocks at each remote afterloader room entrance;
- (2) Source exposure indicator lights on the remote afterloader unit, on the control console, and in the facility;
- (3) Emergency response equipment;
- (4) Radiation monitors used to indicate the source position;
- (5) Timer constancy; and
- (6) Clock (date and time) in the unit's computer.

(b) In addition to the requirements for spot checks in paragraph (a) of this section, a licensee shall ensure overall proper operation of the unit by conducting a simulated cycle of treatment as part of the spot-checks.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(c) The licensee shall have the authorized medical physicist--

(1) Establish written procedures for performing the spot-checks required in paragraph (a) of this section; and

(2) Review the results of each spot-check required by paragraph (a) of this section within 15 days of the check. The authorized medical physicist need not actually perform the spot-check measurements.

(d) If the results of the checks required in paragraph (a) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(e) A licensee shall retain a record of each check required by paragraph (a) of this section in accordance with Sec. 35.2643.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.645 Periodic spot-checks for gamma stereotactic radiosurgery units.</p> <p>(a) A licensee authorized to use gamma stereotactic radiosurgery units for medical use shall perform spot-checks on each unit—</p> <p>(1) Monthly,</p> <p>(2) At the beginning of each day of use, and</p> <p>(3) After each source installation.</p> <p>(b) The licensee shall have the authorized medical physicist--</p> <p>(1) Establish written procedures for performing the spot-checks required in paragraph (a) of this section; and</p> <p>(2) Review the results of each spot-check required by paragraph (a)(1) of this section within 15 days of the check. The authorized medical physicist need not actually perform the spot-check measurements.</p>	<p>§ 35.645 Periodic spot-checks for gamma stereotactic radiosurgery units.</p> <p>(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform spot-checks of each gamma stereotactic radiosurgery facility and on each unit --</p> <p>(1) Monthly;</p> <p>(2) Prior to the first use of the unit on a given day; and</p> <p>(3) After each source installation.</p> <p>(b) The licensee shall have the authorized medical physicist --</p> <p>(1) Establish written procedures for performing the spot-checks required in paragraph (a) of this section; and</p> <p>(2) Review the results of each spot-check required by paragraph (a)(1) of this section within 15 days of the check. The authorized medical physicist need not actually perform the spot-check measurements.</p>	<p>§ 35.645 Periodic spot-checks for gamma stereotactic radiosurgery units.</p> <p>(a) A licensee authorized to use a gamma stereotactic radiosurgery unit for medical use shall perform spot-checks of each gamma stereotactic radiosurgery facility and on each unit --</p> <p>(1) Monthly;</p> <p>(2) Prior to the first use of the unit on a given day; and</p> <p>(3) After each source installation.</p> <p>(b) The licensee shall have the authorized medical physicist --</p> <p>(1) Establish written procedures for performing the spot-checks required in paragraph (a) of this section; and</p> <p>(2) Review the results of each spot-check required by paragraph (a)(1) of this section within 15 days of the check. The authorized medical physicist need not actually perform the spot-check measurements.</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(c) To satisfy the requirements of paragraph (a)(1) of this section, spot-checks must, at a minimum--

(1) Assure proper operation of--

(i) Treatment table retraction mechanism, using backup battery power or hydraulic/electrical backups with the unit off;

(ii) Helmet microswitches;

(iii) Emergency timing circuits;

(iv) Emergency off buttons; and

(v) Stereotactic frames and localizing devices (trunnions).

(2) Determine--

(i) The output for one typical set of operating conditions measured with the dosimetry system described in Sec. 35.630(b);

(ii) The difference between the measurement made in paragraph (c)(2)(i) of this section and the anticipated output, expressed as a percentage of the anticipated output (i.e., the value obtained at last full

(c) To satisfy the requirements of paragraph (a)(1) of this section, spot-checks must, at a minimum --

(1) Assure proper operation of --

(i) Treatment table retraction mechanism, using backup battery power or hydraulic backups with the unit off;

(ii) Helmet microswitches;

(iii) Emergency timing circuits; and

(iv) Stereotactic frames and localizing devices (trunnions).

(2) Determine --

(i) The output for one typical set of operating conditions measured with the dosimetry system described in § 35.630(b);

(ii) The difference between the measurement made in paragraph (c)(2)(i) of this section and the anticipated output, expressed as a percentage of the anticipated output (i.e., the value obtained at last full

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>calibration corrected mathematically for physical decay);</p> <p>(iii) Source output against computer calculation;</p> <p>(iv) Timer accuracy and linearity over the range of use;</p> <p>(v) On-off error; and</p> <p>(vi) Trunnion centricity.</p> <p>(d) To satisfy the requirements of paragraphs (a)(2) and (a)(3) of this section, spot-checks must assure proper operation of--</p> <p>(1) Electrical interlocks at each gamma stereotactic radiosurgery room entrance;</p> <p>(2) Source exposure indicator lights on the gamma stereotactic radiosurgery unit, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems;</p> <p>(4) Timer termination;</p> <p>(5) Radiation monitors used to indicate room exposures; and</p> <p>(6) Hydraulic cutoff mechanism (if applicable).</p> <p>(e) A licensee shall arrange for prompt repair of any system identified in</p>	<p>calibration corrected mathematically for physical decay);</p> <p>(iii) Source output against computer calculation;</p> <p>(iv) Timer accuracy and linearity over the range of use;</p> <p>(v) On-off error; and</p> <p>(vi) Trunnion centricity.</p> <p>(d) To satisfy the requirements of paragraphs (a)(2) and (a)(3) of this section, spot-checks must assure proper operation of --</p> <p>(1) Electrical interlocks at each gamma stereotactic radiosurgery room entrance;</p> <p>(2) Source exposure indicator lights on the gamma stereotactic radiosurgery unit, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems;</p> <p>(4) Timer termination;</p> <p>(5) Radiation monitors used to indicate room exposures; and</p> <p>(6) Emergency off buttons.</p> <p>(e) A licensee shall arrange, as soon as possible, for repair of any system identified in paragraph (c) of</p>	<p>calibration corrected mathematically for physical decay);</p> <p>(iii) Source output against computer calculation;</p> <p>(iv) Timer accuracy and linearity over the range of use;</p> <p>(v) On-off error; and</p> <p>(vi) Trunnion centricity.</p> <p>(d) To satisfy the requirements of paragraphs (a)(2) and (a)(3) of this section, spot-checks must assure proper operation of --</p> <p>(1) Electrical interlocks at each gamma stereotactic radiosurgery room entrance;</p> <p>(2) Source exposure indicator lights on the gamma stereotactic radiosurgery unit, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems;</p> <p>(4) Timer termination;</p> <p>(5) Radiation monitors used to indicate room exposures; and</p> <p>(6) Emergency off buttons.</p> <p>(e) A licensee shall arrange, as soon as possible, for repair of any system identified in paragraph (c) of</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

paragraph (c) of this section that is not operating properly.

(f) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(g) A licensee shall retain a record of each check required by paragraphs (c) and (d) of this section in accordance with Sec. 35.2645.

this section that is not operating properly.

(f) If the results of the checks required in paragraph (d) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

(g) A licensee shall retain a record of each check required by paragraphs (c) and (d) of this section in accordance with § 35.2645.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.647 Additional technical requirements for mobile remote afterloaders.</p> <p>(a) A licensee providing mobile remote afterloader service shall--</p> <p>(1) Check survey instruments before medical use at each address of use or on each day of use, whichever is more frequent; and</p> <p>(2) Account for all sources before departure from a client's address of use.</p> <p>(b) In addition to the periodic spot-checks required by Sec. 35.643, a licensee authorized to use mobile afterloaders for medical use shall perform checks on each remote afterloader before each address of use. At a minimum, checks must be made to verify the operation of--</p> <p>(1) Electrical interlocks on treatment area access points;</p> <p>(2) Source exposure indicator lights on the remote afterloader, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems;</p>	<p>§ 35.647 Additional technical requirements for mobile remote afterloader units.</p> <p>(a) A licensee providing mobile remote afterloader service shall --</p> <p>(1) Check survey instruments before medical use at each address of use or on each day of use, whichever is more frequent; and</p> <p>(2) Account for all sources before departure from a client's address of use.</p> <p>(b) In addition to the periodic spot-checks required by § 35.643, a licensee authorized to use mobile afterloaders for medical use shall perform checks on each remote afterloader unit before use at each address of use. At a minimum, checks must be made to verify the operation of --</p> <p>(1) Electrical interlocks on treatment area access points;</p> <p>(2) Source exposure indicator lights on the remote afterloader unit, on the control console, and in the facility;</p> <p>(3) Viewing and intercom systems;</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>(4) Applicators and connectors;</p> <p>(5) Radiation monitors used to indicate room exposures;</p> <p>(6) Source positioning (accuracy); and</p> <p>(7) Radiation monitors used to indicate whether the source has returned to a safe shielded position.</p> <p>(c) In addition to the requirements for checks in paragraph (b) of this section, a licensee shall ensure overall proper operation of the remote afterloader unit by conducting a simulated cycle of treatment before use at each address of use.</p> <p>(d) A licensee shall arrange for prompt repair of any system identified in paragraph (b) of this section that is not operating properly.</p> <p>(e) A licensee shall retain a record of each check required by paragraph (b) of</p>	<p>(4) Applicators, source transfer tubes, and transfer tube-applicator interfaces;</p> <p>(5) Radiation monitors used to indicate room exposures;</p> <p>(6) Source positioning (accuracy); and</p> <p>(7) Radiation monitors used to indicate whether the source has returned to a safe shielded position.</p> <p>(c) In addition to the requirements for checks in paragraph (b), a licensee shall ensure overall proper operation of the remote afterloader unit by conducting a simulated cycle of treatment before use at each address of use.</p> <p>(d) If the results of the checks required in paragraph (b) of this section indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.</p> <p>(e) A licensee shall retain a record of each check required by paragraph (b) of</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

this section in accordance with
Sec. 35.2647.

this section in accordance with
§ 35.2647.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.641 Radiation surveys for teletherapy facilities.

Sec. 35.652 Radiation surveys.

§ 35.652 Radiation surveys.

(a) Before medical use, after each installation of a teletherapy source, and after making any change for which an amendment is required by Section 35.606 (a) through (d), the licensee shall perform radiation surveys with a portable radiation measurement survey instrument calibrated in accordance with Section 35.51 of this part to verify that:

(a) In addition to the survey requirement in Sec. 20.1501 of this chapter, a licensee shall make such surveys as defined in the Sealed Source and Device Registry to assure that the maximum radiation levels and average radiation levels from the surface of the main source safe with the source(s) in the shielded position does not exceed the levels stated in the Registry.

(a) In addition to the survey requirement in § 20.1501 of this chapter, a person licensed under this subpart shall make such surveys as defined in the Sealed Source and Device Registry to assure that the maximum radiation levels and average radiation levels from the surface of the main source safe with the source(s) in the shielded position do not exceed the levels stated in the Registry.

(1) The maximum and average dose rates at one meter from the teletherapy source with the source in the off position and the collimators set for a normal treatment field do not exceed 10 millirem per hour and 2 millirem per hour, respectively; and

(b) The licensee shall make the survey required by paragraph (a) of this section at installation of a new source and following repairs to the source(s) shielding, the source(s) driving unit, or other electronic or mechanical component that could expose the source, reduce the shielding around the source(s), or compromise the radiation safety of the device or the source(s).

(b) The licensee shall make the survey required by paragraph (a) of this section at installation of a new source and following repairs to the source(s) shielding, the source(s) driving unit, or other electronic or mechanical component that could expose the source, reduce the shielding around the source(s), or compromise the radiation safety of the unit or the source(s).

(2) With the teletherapy source in the on position with the largest clinically available treatment field and with a scattering phantom in the primary beam of radiation, that:

(c) A licensee shall retain a record of the radiation

(i) Radiation dose rates in restricted areas are not likely

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

to cause any occupationally exposed individual to receive a dose in excess of the limits specified in Section 20.1201 of this chapter; and

(ii) Radiation dose rates in controlled or unrestricted areas are not likely to cause any individual member of the public to receive a dose in excess of the limits specified in Section 20.1301 of this chapter.

(b) If the results of the surveys required in paragraph (a) of this section indicate any radiation dose quantity per unit time in excess of the respective limit specified in that paragraph, the licensee shall lock the control in the off position and not use the unit:

(1) Except as may be necessary to repair, replace, or test the teletherapy unit shielding or the treatment room shielding; or

(2) Until the licensee has received a specific exemption pursuant to Section 20.1301 of this chapter.

surveys required by paragraph (a) of this section in accordance with § 35.2652.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(c) A licensee shall retain a record of the radiation measurements made following installation of a source for the duration of the license. The record must include the date of the measurements, the reason the survey is required, the manufacturer's name, model number and serial number of the teletherapy unit, the source, and the instrument used to measure radiation levels, each dose rate measured around the teletherapy source while in the off position and the average of all measurements, a plan of the areas surrounding the treatment room that were surveyed, the measured dose rate at several points in each area expressed in millirem per hour, the calculated maximum quantity of radiation over a period of one week for each restricted and unrestricted area, and the signature of the Radiation Safety Officer.

PROPOSED RULE

(c) A licensee shall retain a record of the radiation surveys required by paragraph (a) of this section in accordance with Sec. 35.2652.

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.643 Modification of teletherapy unit or room before beginning a treatment program.

(a) If the survey required by Section 35.641 indicates that any individual member of the public is likely to receive a dose in excess of the limits specified in Section 20.1301, of this chapter, the licensee shall, before beginning the treatment program:

(1) Either equip the unit with stops or add additional radiation shielding to ensure compliance with Section 20.1301 of this chapter;

(2) Perform the survey required by Section 35.641 again; and

(3) Include in the report required by Section 35.645 the results of the initial survey, a description of the modification made to comply with paragraph (a)(1) of this section, and the results of the second survey.

(b) As an alternative to the requirements set out in paragraph (a) of this section, a

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

licensee may request a license amendment under Section 20.1301(c) of this chapter that authorizes radiation levels in unrestricted areas greater than those permitted by Section 20.1301(a) of this chapter. A licensee may not begin the treatment program until the license amendment has been issued.

Section 35.645 Reports of teletherapy surveys, checks, tests, and measurements.

A licensee shall mail a copy of the records required in Sections 35.636, 35.641, 35.643, and the output from the teletherapy source expressed as roentgens or rads per hour at one meter from the source and determined during the full calibration required in Section 35.632, to the appropriate Commission Regional Office listed in Section 30.6 of this chapter within thirty days following completion of the action that initiated the record requirement.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.647 Five-year inspection.

Sec. 35.655 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units.

§ 35.655 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units.

(a) A licensee shall have each teletherapy unit fully inspected and serviced during teletherapy source replacement or at intervals not to exceed five years, whichever comes first, to assure proper functioning of the source exposure mechanism.

(a) A licensee shall have each teletherapy unit and gamma stereotactic radiosurgery unit fully inspected and serviced during source replacement or at intervals not to exceed 5 years, whichever comes first, to assure proper functioning of the source exposure mechanism.

(a) A licensee shall have each teletherapy unit and gamma stereotactic radiosurgery unit fully inspected and serviced during source replacement or at intervals not to exceed 5 years, whichever comes first, to assure proper functioning of the source exposure mechanism.

(b) This inspection and servicing may only be performed by persons specifically licensed to do so by the Commission or an Agreement State.

(b) This inspection and servicing may only be performed by persons specifically licensed to do so by the Commission or an Agreement State.

(b) This inspection and servicing may only be performed by persons specifically licensed to do so by the Commission or an Agreement State.

(c) A licensee shall keep a record of the inspection and servicing for the duration of the license. The record must contain the inspector's name, the inspector's license number, the date of inspection, the manufacturer's name and model number and serial number for both the teletherapy unit and source, a list of components inspected, a list of components serviced

(c) A licensee shall keep a record of the inspection and servicing in accordance with Sec. 35.2655.

(c) A licensee shall keep a record of the inspection and servicing in accordance with § 35.2655.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

and the type of service, a list of components replaced, and the signature of the inspector.

Sec. 35.657 Therapy-related computer systems.

The licensee shall:
(a) Verify that the computerized operating system and treatment planning system associated with the therapy device are operating appropriately; and
(b) Perform acceptance testing on the treatment planning system in accordance with published protocols approved by nationally recognized bodies.

§ 35.657 Therapy-related computer systems.

The licensee shall perform acceptance testing on the treatment planning system in accordance with published protocols accepted by nationally recognized bodies. At a minimum, the acceptance testing must include, as applicable, verification of:
(a) The source-specific input parameters required by the dose calculation algorithm;
(b) The accuracy of dose, dwell time, and treatment time calculations at representative points;
(c) The accuracy of isodose plots and graphic displays;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(d) The accuracy of the software used to determine radioactive source positions from radiographic images; and

(e) The accuracy of electronic transfer of the treatment delivery parameters to the treatment delivery unit from the treatment planning system.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.690 Training for use of therapeutic medical devices.</p>	<p>§ 35.690 Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.</p>	<p>Except as provided in § 35.57, the licensee shall require an authorized user of a sealed source for a use authorized under § 35.600 to be a physician who --</p> <p>(a) Is certified by a medical specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been recognized by the Commission or by an Agreement State; or</p> <p>(b)(1) Has completed a structured educational program in basic radionuclide techniques applicable to the use of a sealed source in a therapeutic medical unit that includes --</p> <p>(i) 200 hours of classroom and laboratory training in the following areas -</p> <p>(A) Radiation physics and instrumentation;</p> <p>(B) Radiation protection;</p>
<p>Except as provided in Sec. 35.57, the licensee shall require the authorized user of a sealed source for a use authorized under Sec. 35.600 to be a physician who--</p>	<p>(a) Is certified by a specialty board whose certification process includes all of the requirements in paragraph (b) of this section and whose certification has been approved by the Commission; or</p>	
<p>(b)(1) Has completed a structured educational program in basic radionuclide techniques applicable to the use of a sealed source in a therapeutic medical device consisting of both--</p>	<p>(i) 200 hours of didactic training in the following areas--</p>	
<p>(A) Radiation physics and instrumentation;</p>		

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>(B) Radiation protection;</p> <p>(C) Mathematics pertaining to the use and measurement of radioactivity; and</p> <p>(D) Radiation biology; and</p> <p>(ii) 500 hours of supervised practical experience, under the supervision of an authorized user at a medical institution, involving—</p> <p>(A) Review of the full calibration measurements and periodic spot checks;</p> <p>(B) Preparing treatment plans and calculating treatment doses and times;</p> <p>(C) Using administrative controls to prevent a medical event involving the use of byproduct material;</p> <p>(D) Implementing emergency procedures to be followed in the event of the abnormal operation of the medical device or console;</p> <p>(E) Checking and using survey meters; and</p>	<p>(C) Mathematics pertaining to the use and measurement of radioactivity; and</p> <p>(D) Radiation biology; and</p> <p>(ii) 500 hours of work experience, under the supervision of an authorized user who meets the requirements in § 35.690 or equivalent Agreement State requirements at a medical institution, involving --</p> <p>(A) Reviewing full calibration measurements and periodic spot-checks;</p> <p>(B) Preparing treatment plans and calculating treatment doses and times;</p> <p>(C) Using administrative controls to prevent a medical event involving the use of byproduct material;</p> <p>(D) Implementing emergency procedures to be followed in the event of the abnormal operation of the medical unit or console;</p> <p>(E) Checking and using survey meters; and</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(F) Selecting the proper dose and how it is to be administered; and
(2) Three years of supervised clinical experience that includes one year in a formal training program approved by the Residency Review Committee for Radiology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association or equivalent program approved by the NRC and an additional two years of clinical experience under the supervision of an authorized user; and

(3) Has obtained written certification, signed by a preceptor authorized user, that the required training in paragraphs (b)(1) and (b)(2) of this section has been satisfactorily completed and that the individual has achieved a level of

(F) Selecting the proper dose and how it is to be administered; and
(2) Has completed three years of supervised clinical experience in radiation oncology, under an authorized user who meets the requirements in § 35.690 or equivalent Agreement State requirements, as part of a formal training program approved by the Residency Review Committee for Radiation Oncology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association. This experience may be obtained concurrently with the supervised work experience required by paragraph (b)(1)(ii) of this section; and

(3) Has obtained written certification, signed by a preceptor authorized user who meets the requirements in § 35.690 or equivalent Agreement State requirements, that the individual has satisfactorily completed the requirements in

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	competency sufficient to independently function as an authorized user of the therapeutic medical device for which the individual is requesting authorized user status; and	paragraphs (b)(1) and (b)(2) of this section and has achieved a level of competency sufficient to function independently as an authorized user of the therapeutic medical unit for which the individual is requesting authorized user status.
	(4) Following completion of the requirements in paragraph (b)(1) and (2) of this section, has demonstrated sufficient knowledge in radiation safety commensurate with the use requested by passing an examination given by an organization or entity approved by the Commission in accordance with appendix A of this part.	
Subpart I--Reserved		Subpart I -- [Reserved]

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart J--Training and Experience Requirements

Subpart J--Training and Experience Requirements

Subpart J -- [Reserved]

Section 35.900 Radiation Safety Officer.

Sec. 35.900 Radiation Safety Officer.

Except as provided in Section 35.901, the licensee shall require an individual fulfilling the responsibilities of the Radiation Safety Officer as provided in Section 35.32 to be an individual who:

- (a) Is certified by:
 - (1) American Board of Health Physics in Comprehensive Health Physics;
 - (2) American Board of Radiology;
 - (3) American Board of Nuclear Medicine;
 - (4) American Board of Science in Nuclear Medicine;
 - (5) Board of Pharmaceutical Specialties in Nuclear Pharmacy;
 - (6) American Board of Medical Physics in radiation oncology physics;
 - (7) Royal College of Physicians and Surgeons of Canada in nuclear medicine;
 - (8) American Osteopathic Board of Radiology; or

Except as provided in Sec. 35.57, the licensee shall require an individual fulfilling the responsibilities of the Radiation Safety Officer as provided in Sec. 35.24 to be an individual who--

- (a) Is certified by the--
 - (1) American Board of Health Physics in Comprehensive Health Physics;
 - (2) American Board of Radiology;
 - (3) American Board of Nuclear Medicine;
 - (4) American Board of Science in Nuclear Medicine;
 - (5) Board of Pharmaceutical Specialties in Nuclear Pharmacy;
 - (6) American Board of Medical Physics in radiation oncology physics;
 - (7) Royal College of Physicians and Surgeons of Canada in nuclear medicine;
 - (8) American Osteopathic Board of Radiology; or

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(9) American Osteopathic Board of Nuclear Medicine; or
(b) Has had classroom and laboratory training and experience as follows:
(1) 200 hours of classroom and laboratory training that includes:
(i) Radiation physics and instrumentation;
(ii) Radiation protection;
(iii) Mathematics pertaining to the use and measurement of radioactivity;
(iv) Radiation biology; and
(v) Radiopharmaceutical chemistry; and
(2) One year of full time experience as a radiation safety technologist at a medical institution under the supervision of the individual identified as the Radiation Safety Officer on a Commission or Agreement State license that authorizes the medical use of byproduct material; or
(c) Be an authorized user identified on the licensee's license.

(9) American Osteopathic Board of Nuclear Medicine; or
(b) Has had classroom and laboratory training and experience as follows--
(1) 200 hours of classroom and laboratory training that includes--
(i) Radiation physics and instrumentation;
(ii) Radiation protection;
(iii) Mathematics pertaining to the use and measurement of radioactivity;
(iv) Radiation biology; and
(v) Radiopharmaceutical chemistry; and
(2) One year of full time experience as a radiation safety technologist at a medical institution under the supervision of the individual identified as the Radiation Safety Officer on a Commission or Agreement State license that authorizes the medical use of byproduct material; or
(c) Is an authorized user identified on the licensee's license.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.901 Training for experienced Radiation Safety Officer.

An individual identified as a Radiation Safety Officer on a Commission or Agreement State license before October 1, 1986 need not comply with the training requirements of Section 35.900.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.910 Training for uptake, dilution, and excretion studies.

Except as provided in Sections 35.970 and 35.971, the licensee shall require the authorized user of a radiopharmaceutical in Section 35.100(a) to be a physician who:

- (a) Is certified in:
 - (1) Nuclear medicine by the American Board of Nuclear Medicine;
 - (2) Diagnostic radiology by the American Board of Radiology; or
 - (3) Diagnostic radiology or radiology by the American Osteopathic Board of Radiology;
 - (4) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or
 - (5) American Osteopathic Board of Nuclear Medicine in nuclear medicine; or
- (b) Has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of prepared radiopharmaceuticals, and supervised clinical experience as follows:

Sec. 35.910 Training for uptake, dilution, and excretion studies.

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a radiopharmaceutical in Sec. 35.100(a) to be a physician who--

- (a) Is certified in--
 - (1) Nuclear medicine by the American Board of Nuclear Medicine;
 - (2) Diagnostic radiology by the American Board of Radiology;
 - (3) Diagnostic radiology or radiology by the American Osteopathic Board of Radiology;
 - (4) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or
 - (5) American Osteopathic Board of Nuclear Medicine in nuclear medicine; or
- (b) Has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of prepared radiopharmaceuticals, and supervised clinical experience as follows--
 - (1) 40 hours of classroom and laboratory training that includes--

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(1) 40 hours of classroom and laboratory training that includes:

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics pertaining to the use and measurement of radioactivity;
- (iv) Radiation biology; and
- (v) Radiopharmaceutical chemistry; and

(2) 20 hours of supervised clinical experience under the supervision of an authorized user and that includes:

- (i) Examining patients or human research subjects and reviewing their case histories to determine their suitability for radioisotope diagnosis, limitations, or contraindications;
- (ii) Selecting the suitable radiopharmaceuticals and calculating and measuring the dosages;
- (iii) Administering dosages to patients or human research subjects and using syringe radiation shields;
- (iv) Collaborating with the authorized user in the interpretation of radioisotope test results; and

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics pertaining to the use and measurement of radioactivity;
- (iv) Radiation biology; and
- (v) Radiopharmaceutical chemistry; and

(2) 20 hours of supervised clinical experience under the supervision of an authorized user and that includes--

- (i) Examining patients or human research subjects and reviewing their case histories to determine their suitability for radioisotope diagnosis, limitations, or contraindications;
- (ii) Selecting the suitable radiopharmaceuticals and calculating and measuring the dosages;
- (iii) Administering dosages to patients or human research subjects and using syringe radiation shields;
- (iv) Collaborating with the authorized user in the interpretation of radioisotope test results; and
- (v) Patient or human research subject follow up; or

(c) Has successfully completed a 6-month training

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(v) Patient or human research subject followup; or
(c) Has successfully completed a six-month training program in nuclear medicine as part of a training program that has been approved by the Accreditation Council for Graduate Medical Education and that included classroom and laboratory training, work experience, and supervised clinical experience in all the topics identified in paragraph (b) of this section.

program in nuclear medicine as part of a training program that has been approved by the Accreditation Council for Graduate Medical Education and that included classroom and laboratory training, work experience, and supervised clinical experience in all the topics identified in paragraph (b) of this section.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.920 Training for imaging and localization studies.

Sec. 35.920 Training for imaging and localization studies.

Except as provided in Section 35.970 or 35.971, the licensee shall require the authorized user of a radiopharmaceutical, generator, or reagent kit in Section 35.200(a) to be a physician who:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a radiopharmaceutical, generator, or reagent kit in Sec. 35.200(a) to be a physician who--

(a) Is certified in:

(a) Is certified in--

(1) Nuclear medicine by the American Board of Nuclear Medicine;

(1) Nuclear medicine by the American Board of Nuclear Medicine;

(2) Diagnostic radiology by the American Board of Radiology;

(2) Diagnostic radiology by the American Board of Radiology;

(3) Diagnostic radiology or radiology by the American Osteopathic Board of Radiology;

(3) Diagnostic radiology or radiology by the American Osteopathic Board of Radiology;

(4) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or

(4) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or

(5) American Osteopathic Board of Nuclear Medicine in nuclear medicine; or

(5) American Osteopathic Board of Nuclear Medicine in nuclear medicine; or

(b) Has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of prepared radiopharmaceuticals, generators, and reagent kits, supervised work experience, and supervised clinical experience as follows:

(b) Has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of prepared radiopharmaceuticals, generators, and reagent kits, supervised work experience, and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(1) 200 hours of classroom and laboratory training that includes:

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics pertaining to the use and measurement of radioactivity;
- (iv) Radiopharmaceutical chemistry; and

(v) Radiation biology; and

(2) 500 hours of supervised work experience under the supervision of an authorized user that includes:

- (i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (ii) Calibrating dose calibrators and diagnostic instruments and performing checks for proper operation of survey meters;
- (iii) Calculating and safely preparing patient or human research subject dosages;
- (iv) Using administrative controls to prevent the misadministration of byproduct material;

supervised clinical experience as follows--

(1) 200 hours of classroom and laboratory training that includes--

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics pertaining to the use and measurement of radioactivity;
- (iv) Radiopharmaceutical chemistry; and

(v) Radiation biology; and
(2) 500 hours of supervised work experience under the supervision of an authorized user that includes--

- (i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (ii) Calibrating dose calibrators and diagnostic instruments and performing checks for proper operation of survey meters;
- (iii) Calculating and safely preparing patient or human research subject dosages;
- (iv) Using administrative controls to prevent the medical event of byproduct material;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and
(vi) Eluting technetium-99m from generator systems, measuring and testing the eluate for molybdenum-99 and alumina contamination, and processing the eluate with reagent kits to prepare technetium-99m labeled radiopharmaceuticals; and
(3) 500 hours of supervised clinical experience under the supervision of an authorized user that includes:

(i) Examining patients or human research subjects and reviewing their case histories to determine their suitability for radioisotope diagnosis, limitations, or contraindications;
(ii) Selecting the suitable radiopharmaceuticals and calculating and measuring the dosages;
(iii) Administering dosages to patients or human research subjects and using syringe radiation shields;
(iv) Collaborating with the authorized user in the

(v) Using procedures to contain spilled byproduct material safely and using proper decontamination procedures; and
(vi) Eluting technetium-99m from generator systems, measuring and testing the eluate for molybdenum-99 and alumina contamination, and processing the eluate with reagent kits to prepare technetium-99m labeled radiopharmaceuticals; and
(3) 500 hours of supervised clinical experience under the supervision of an authorized user that includes--

(i) Examining patients or human research subjects and reviewing their case histories to determine their suitability for radioisotope diagnosis, limitations, or contraindications;
(ii) Selecting the suitable radiopharmaceuticals and calculating and measuring the dosages;
(iii) Administering dosages to patients or human research subjects and using syringe radiation shields;
(iv) Collaborating with the authorized user in the

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

interpretation of radioisotope test results; and
(v) Patient or human research subject followup; or
(c) Has successfully completed a six-month training program in nuclear medicine that has been approved by the Accreditation Council for Graduate Medical Education and that included classroom and laboratory training, work experience, and supervised clinical experience in all the topics identified in paragraph (b) of this section.

interpretation of radioisotope test results; and
(v) Patient or human research subject follow up; or
(c) Has successfully completed a 6-month training program in nuclear medicine that has been approved by the Accreditation Council for Graduate Medical Education and that included classroom and laboratory training, work experience, and supervised clinical experience in all the topics identified in paragraph (b) of this section.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.930 Training for therapeutic use of unsealed byproduct material.

Sec. 35.930 Training for therapeutic use of unsealed byproduct material.

Except as provided in Section 35.970, the licensee shall require the authorized user of radiopharmaceuticals in Section 35.300 to be a physician who:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of radiopharmaceuticals in Sec. 35.300 to be a physician who--

(a) Is certified by:

(a) Is certified by--

(1) The American Board of Nuclear Medicine;
(2) The American Board of Radiology in radiology, therapeutic radiology, or radiation oncology;

(1) The American Board of Nuclear Medicine;
(2) The American Board of Radiology in radiology, therapeutic radiology, or radiation oncology;

(3) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or

(3) The Royal College of Physicians and Surgeons of Canada in nuclear medicine; or

(4) The American Osteopathic Board of Radiology after 1984; or

(4) The American Osteopathic Board of Radiology after 1984; or

(b) Has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of therapeutic radiopharmaceuticals, and supervised clinical experience as follows:

(b) Has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of therapeutic radiopharmaceuticals, and supervised clinical experience as follows--

(1) 80 hours of classroom and laboratory training that includes:

(1) 80 hours of classroom and laboratory training that includes--

(i) Radiation physics and instrumentation;

(i) Radiation physics and instrumentation;

(ii) Radiation protection;

(ii) Radiation protection;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(iii) Mathematics pertaining to the use and measurement of radioactivity; and
(iv) Radiation biology; and
(2) Supervised clinical experience under the supervision of an authorized user at a medical institution that includes:
(i) Use of iodine-131 for diagnosis of thyroid function and the treatment of hyperthyroidism or cardiac dysfunction in 10 individuals; and
(ii) Use of iodine-131 for treatment of thyroid carcinoma in 3 individuals.

Section 35.932 Training for treatment of hyperthyroidism.

Except as provided in Section 35.970, the licensee shall require the authorized user of only iodine-131 for the treatment of hyperthyroidism to be a physician with special experience in thyroid disease who has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of iodine-131 for treating hyperthyroidism, and supervised clinical experience as follows:

(iii) Mathematics pertaining to the use and measurement of radioactivity; and
(iv) Radiation biology; and
(2) Supervised clinical experience under the supervision of an authorized user at a medical institution that includes--
(i) Use of iodine-131 for diagnosis of thyroid function and the treatment of hyperthyroidism or cardiac dysfunction in 10 individuals; and
(ii) Use of iodine-131 for treatment of thyroid carcinoma in 3 individuals.

Sec. 35.932 Training for treatment of hyperthyroidism.

Except as provided in Sec. 35.57, the licensee shall require the authorized user of only iodine-131 for the treatment of hyperthyroidism to be a physician with special experience in thyroid disease who has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of iodine-131 for treating hyperthyroidism, and supervised clinical experience as follows--

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(a) 80 hours of classroom and laboratory training that includes:
(1) Radiation physics and instrumentation;
(2) Radiation protection,
(3) Mathematics
pertaining to the use and measurement of radioactivity;
and
(4) Radiation biology; and
(b) Supervised clinical experience under the supervision of an authorized user that includes the use of iodine-131 for diagnosis of thyroid function, and the treatment of hyperthyroidism in 10 individuals.

PROPOSED RULE

(a) 80 hours of classroom and laboratory training that includes--
(1) Radiation physics and instrumentation;
(2) Radiation protection,
(3) Mathematics
pertaining to the use and measurement of radioactivity; and
(4) Radiation biology; and
(b) Supervised clinical experience under the supervision of an authorized user that includes the use of iodine-131 for diagnosis of thyroid function, and the treatment of hyperthyroidism in 10 individuals.

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.934 Training for treatment of thyroid carcinoma.

Sec. 35.934 Training for treatment of thyroid carcinoma.

Except as provided in Section 35.970, the licensee shall require the authorized user of only iodine-131 for the treatment of thyroid carcinoma to be a physician with special experience in thyroid disease who has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of iodine-131 for treating thyroid carcinoma, and supervised clinical experience as follows:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of only iodine-131 for the treatment of thyroid carcinoma to be a physician with special experience in thyroid disease who has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of iodine-131 for treating thyroid carcinoma, and supervised clinical experience as follows--

(a) 80 hours of classroom and laboratory training that includes:

(a) 80 hours of classroom and laboratory training that includes--

(1) Radiation physics and instrumentation;

(1) Radiation physics and instrumentation;

(2) Radiation protection;

(2) Radiation protection;

(3) Mathematics

(3) Mathematics

pertaining to the use and measurement of radioactivity; and

pertaining to the use and measurement of radioactivity; and

(4) Radiation biology; and

(4) Radiation biology; and

(b) Supervised clinical

(b) Supervised clinical

experience under the supervision of an authorized user that includes the use of iodine-131 for the treatment of thyroid carcinoma in 3 individuals.

experience under the supervision of an authorized user that includes the use of iodine-131 for the treatment of thyroid carcinoma in 3 individuals.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.940 Training for use of brachytherapy sources.

Sec. 35.940 Training for use of brachytherapy sources.

Except as provided in Section 35.970, the licensee shall require the authorized user of a brachytherapy source listed in Section 35.400 for therapy to be a physician who:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a brachytherapy source listed in Sec. 35.400 for therapy to be a physician who--

(a) Is certified in:

(a) Is certified in--

(1) Radiology, therapeutic radiology, or radiation oncology by the American Board of Radiology;

(1) Radiology, therapeutic radiology, or radiation oncology by the American Board of Radiology;

(2) Radiation oncology by the American Osteopathic Board of Radiology;

(2) Radiation oncology by the American Osteopathic Board of Radiology;

(3) Radiology, with specialization in radiotherapy, as a British "Fellow of the Faculty of Radiology" or "Fellow of the Royal College of Radiology"; or

(3) Radiology, with specialization in radiotherapy, as a British "Fellow of the Faculty of Radiology" or "Fellow of the Royal College of Radiology"; or

(4) Therapeutic radiology by the Canadian Royal College of Physicians and Surgeons; or

(4) Therapeutic radiology by the Canadian Royal College of Physicians and Surgeons; or

(b) Is in the active practice of therapeutic radiology, has had classroom and laboratory training in radioisotope handling techniques applicable to the therapeutic use of brachytherapy sources, supervised work experience, and supervised clinical experience as follows:

(b) Is in the active practice of therapeutic radiology, has had classroom and laboratory training in radioisotope handling techniques applicable to the therapeutic use of brachytherapy sources, supervised work experience, and supervised clinical experience as follows--

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(1) 200 hours of classroom and laboratory training that includes:

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics

pertaining to the use and measurement of radioactivity; and

- (iv) Radiation biology;

(2) 500 hours of supervised work experience under the supervision of an authorized user at a medical institution that includes:

- (i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (ii) Checking survey meters for proper operation;
- (iii) Preparing, implanting, and removing sealed sources;
- (iv) Maintaining running inventories of material on hand;
- (v) Using administrative controls to prevent the misadministration of byproduct material; and

- (vi) Using emergency procedures to control byproduct material; and

PROPOSED RULE

(1) 200 hours of classroom and laboratory training that includes--

- (i) Radiation physics and instrumentation;
- (ii) Radiation protection;
- (iii) Mathematics

pertaining to the use and measurement of radioactivity; and

- (iv) Radiation biology;

(2) 500 hours of supervised work experience under the supervision of an authorized user at a medical institution that includes--

- (i) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- (ii) Checking survey meters for proper operation;
- (iii) Preparing, implanting, and removing sealed sources;
- (iv) Maintaining running inventories of material on hand;
- (v) Using administrative controls to prevent a medical event involving byproduct material; and
- (vi) Using emergency procedures to control byproduct material; and

(3) Three years of supervised clinical experience that includes one year in a formal

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(3) Three years of supervised clinical experience that includes one year in a formal training program approved by the Residency Review Committee for Radiology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association, and an additional two years of clinical experience in therapeutic radiology under the supervision of an authorized user at a medical institution that includes:

(i) Examining individuals and reviewing their case histories to determine their suitability for brachytherapy treatment, and any limitations or contraindications;

(ii) Selecting the proper brachytherapy sources and dose and method of administration;

(iii) Calculating the dose; and

(iv) Post-administration followup and review of case histories in collaboration with the authorized user.

PROPOSED RULE

training program approved by the Residency Review Committee for Radiology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association, and an additional two years of clinical experience in therapeutic radiology under the supervision of an authorized user at a medical institution that includes—

(i) Examining individuals and reviewing their case histories to determine their suitability for brachytherapy treatment, and any limitations or contraindications;

(ii) Selecting the proper brachytherapy sources and dose and method of administration;

(iii) Calculating the dose; and

(iv) Post-administration follow up and review of case histories in collaboration with the authorized user.

FINAL RULE

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.941 Training for ophthalmic use of strontium-90.

Sec. 35.941 Training for ophthalmic use of strontium-90.

Except as provided in Section 35.970, the licensee shall require the authorized user of only strontium-90 for ophthalmic radiotherapy to be a physician who is in the active practice of therapeutic radiology or ophthalmology, and has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of strontium-90 for ophthalmic radiotherapy, and a period of supervised clinical training in ophthalmic radiotherapy as follows:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of only strontium-90 for ophthalmic radiotherapy to be a physician who is in the active practice of therapeutic radiology or ophthalmology, and has had classroom and laboratory training in basic radioisotope handling techniques applicable to the use of strontium-90 for ophthalmic radiotherapy, and a period of supervised clinical training in ophthalmic radiotherapy as follows--

(a) 24 hours of classroom and laboratory training that includes:

(a) 24 hours of classroom and laboratory training that includes--

- (1) Radiation physics and instrumentation;
- (2) Radiation protection;
- (3) Mathematics pertaining to the use and measurement of radioactivity; and

- (1) Radiation physics and instrumentation;
- (2) Radiation protection;
- (3) Mathematics pertaining to the use and measurement of radioactivity; and

- (4) Radiation biology;
- (b) Supervised clinical training in ophthalmic radiotherapy under the supervision of an authorized user at a medical institution that

- (4) Radiation biology;
- (b) Supervised clinical training in ophthalmic radiotherapy under the supervision of an authorized user at a medical institution that includes the use of strontium-90

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

includes the use of strontium-90 for the ophthalmic treatment of five individuals that includes:
(1) Examination of each individual to be treated;
(2) Calculation of the dose to be administered;
(3) Administration of the dose; and
(4) Followup and review of each individual's case history.

for the ophthalmic treatment of five individuals that includes--
(1) Examination of each individual to be treated;
(2) Calculation of the dose to be administered;
(3) Administration of the dose; and
(4) Follow up and review of each individual's case history.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.950 Training for use of sealed sources for diagnosis.

Sec. 35.950 Training for use of sealed sources for diagnosis.

Except as provided in Section 35.970, the licensee shall require the authorized user of a sealed source in a device listed in Section 35.500 to be a physician, dentist, or podiatrist who:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a sealed source in a device listed in Sec. 35.500 to be a physician, dentist, or podiatrist who--

(a) Is certified in:

(a) Is certified in--

(1) Radiology, diagnostic radiology, therapeutic radiology, or radiation oncology by the American Board of Radiology;

(1) Radiology, diagnostic radiology, therapeutic radiology, or radiation oncology by the American Board of Radiology;

(2) Nuclear medicine by the American Board of Nuclear Medicine;

(2) Nuclear medicine by the American Board of Nuclear Medicine;

(3) Diagnostic radiology or radiology by the American Osteopathic Board of Radiology; or

(3) Diagnostic radiology or radiology by the American Osteopathic Board of Radiology; or

(4) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or

(4) Nuclear medicine by the Royal College of Physicians and Surgeons of Canada; or

(b) Has had 8 hours of classroom and laboratory training in basic radioisotope handling techniques specifically applicable to the use of the device that includes:

(b) Has had 8 hours of classroom and laboratory training in basic radioisotope handling techniques specifically applicable to the use of the device that includes--

(1) Radiation physics, mathematics pertaining to the use and measurement of radioactivity, and instrumentation;

(1) Radiation physics, mathematics pertaining to the use and measurement of radioactivity, and instrumentation;

(2) Radiation biology;

(2) Radiation biology;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(3) Radiation protection;
and
(4) Training in the use of
the device for the uses
requested.

(3) Radiation protection;
and
(4) Training in the use of
the device for the uses
requested.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.960 Training for teletherapy.

Sec. 35.960 Training for use of therapeutic medical devices.

Except as provided in Section 35.970, the licensee shall require the authorized user of a sealed source listed in Section 35.600 in a teletherapy unit to be a physician who:

Except as provided in Sec. 35.57, the licensee shall require the authorized user of a sealed source listed in Sec. 35.600 to be a physician who--

- (a) Is certified in:
 - (1) Radiology, therapeutic radiology, or radiation oncology by the American Board of Radiology;
 - (2) Radiation oncology by the American Osteopathic Board of Radiology;
 - (3) Radiology, with specialization in radiotherapy, as a British "Fellow of the Faculty of Radiology" or "Fellow of the Royal College of Radiology"; or
 - (4) Therapeutic radiology by the Canadian Royal College of Physicians and Surgeons; or
- (b) Is in the active practice of therapeutic radiology, and has had classroom and laboratory training in basic radioisotope techniques applicable to the use of a sealed source in a teletherapy unit, supervised work

- (a) Is certified in--
 - (1) Radiology, therapeutic radiology, or radiation oncology by the American Board of Radiology;
 - (2) Radiation oncology by the American Osteopathic Board of Radiology;
 - (3) Radiology, with specialization in radiotherapy, as a British 'Fellow of the Faculty of Radiology' or 'Fellow of the Royal College of Radiology'; or
 - (4) Therapeutic radiology by the Canadian Royal College of Physicians and Surgeons; or
- (b) Is in the active practice of therapeutic radiology, and has had classroom and laboratory training in basic radioisotope techniques applicable to the use of a sealed source in a therapeutic medical device, supervised work experience, and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

experience, and supervised clinical experience as follows:
 (1) 200 hours of classroom and laboratory training that includes:
 (i) Radiation physics and instrumentation;
 (ii) Radiation protection;
 (iii) Mathematics pertaining to the use and measurement of radioactivity; and
 (iv) Radiation biology;
 (2) 500 hours of supervised work experience under the supervision of an authorized user at a medical institution that includes:
 (i) Review of the full calibration measurements and periodic spot checks;
 (ii) Preparing treatment plans and calculating treatment times;
 (iii) Using administrative controls to prevent misadministrations;
 (iv) Implementing emergency procedures to be followed in the event of the abnormal operation of a teletherapy unit or console; and
 (v) Checking and using survey meters; and

supervised clinical experience as follows--
 (1) 200 hours of classroom and laboratory training that includes--
 (i) Radiation physics and instrumentation;
 (ii) Radiation protection;
 (iii) Mathematics pertaining to the use and measurement of radioactivity; and
 (iv) Radiation biology;
 (2) 500 hours of supervised work experience under the supervision of an authorized user at a medical institution that includes--
 (i) Review of the full calibration measurements and periodic spot-checks;
 (ii) Preparing treatment plans and calculating treatment times;
 (iii) Using administrative controls to prevent medical events;
 (iv) Implementing emergency procedures to be followed in the event of the abnormal operation of the medical device or console; and
 (v) Checking and using survey meters; and
 (3) Three years of supervised clinical experience

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(3) Three years of supervised clinical experience that includes one year in a formal training program approved by the Residency Review Committee for Radiology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association and an additional two years of clinical experience in therapeutic radiology under the supervision of an authorized user at a medical institution that includes:

(i) Examining individuals and reviewing their case histories to determine their suitability for teletherapy treatment, and any limitations or contraindications;

(ii) Selecting the proper dose and how it is to be administered;

(iii) Calculating the teletherapy doses and collaborating with the authorized user in the review of patients' or human research subjects' progress and consideration of the need to modify originally prescribed doses as warranted by patients' or human research

that includes one year in a formal training program approved by the Residency Review Committee for Radiology of the Accreditation Council for Graduate Medical Education or the Committee on Postdoctoral Training of the American Osteopathic Association and an additional two years of clinical experience in therapeutic radiology under the supervision of an authorized user at a medical institution that includes--

(i) Examining individuals and reviewing their case histories to determine their suitability for teletherapy, remote afterloader, or gamma stereotactic radiosurgery treatment, and any limitations or contraindications;

(ii) Selecting the proper dose and how it is to be administered;

(iii) Calculating the doses and collaborating with the authorized user in the review of patients' or human research subjects' progress and consideration of the need to modify originally prescribed doses as warranted by patients' or human research subjects' reaction to radiation; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

subjects' reaction to radiation;
and
(iv) Post-administration
followup and review of case
histories.

(iv) Post-administration
follow up and review of case
histories.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.961 Training for teletherapy physicist.

The licensee shall require the teletherapy physicist to be an individual who:

(a) Is certified by the American Board of Radiology in:

- (1) Therapeutic radiological physics;
- (2) Roentgen ray and gamma ray physics;
- (3) X-ray and radium physics; or

(4) Radiological physics;

or

(b) Is certified by the American Board of Medical Physics in radiation oncology physics; or

(c) Holds a master's or doctor's degree in physics, biophysics, radiological physics, or health physics, and has completed one year of full time training in therapeutic radiological physics and an additional year of full time work experience under the supervision of a teletherapy physicist at a medical institution that includes the tasks listed in Sections 35.59, 35.632, 35.634, and 35.641 of this part.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.970 Training for experienced authorized users.

Physicians, dentists, or podiatrists identified as authorized users for the medical, dental, or podiatric use of byproduct material on a Commission or Agreement State license issued before April 1, 1987 who perform only those methods of use for which they were authorized on that date need not comply with the training requirements of subpart J.

Section 35.971 Physician training in a three month program.

A physician who, before July 1, 1984, began a three month nuclear medicine training program approved by the Accreditation Council for Graduate Medical Education and has successfully completed the program need not comply with the requirements of Sections 35.910 or 35.920.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.972 Recentness of training.

The training and experience specified in this subpart must have been obtained within the 7 years preceding the date of application or the individual must have had related continuing education and experience since the required training and experience was completed.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.980 Training for an authorized nuclear pharmacist.

The licensee shall require the authorized nuclear pharmacist to be a pharmacist who:

(a) Has current board certification as a nuclear pharmacist by the Board of Pharmaceutical Specialties, or

(b)(1) Has completed 700 hours in a structured educational program consisting of both:

(i) Didactic training in the following areas:

(A) Radiation physics and instrumentation;

(B) Radiation protection;

(C) Mathematics pertaining to the use and measurement of radioactivity;

(D) Chemistry of byproduct material for medical use; and

(E) Radiation biology; and

(ii) Supervised experience in a nuclear pharmacy involving the following:

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(A) Shipping, receiving, and performing related radiation surveys;

(B) Using and performing checks for proper operation of dose calibrators, survey meters, and, if appropriate, instruments used to measure alpha- or beta-emitting radionuclides;

(C) Calculating, assaying, and safely preparing dosages for patients or human research subjects;

(D) Using administrative controls to avoid mistakes in the administration of byproduct material;

(E) Using procedures to prevent or minimize contamination and using proper decontamination procedures; and

(2) Has obtained written certification, signed by a preceptor authorized nuclear pharmacist, that the above training has been satisfactorily completed and that the individual has achieved a level of competency sufficient to independently operate a nuclear pharmacy.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.981 Training for experienced nuclear pharmacists.

A licensee may apply for and must receive a license amendment identifying an experienced nuclear pharmacist as an authorized nuclear pharmacist before it allows this individual to work as an authorized nuclear pharmacist. A pharmacist who has completed a structured educational program as specified in Section 35.980(b)(1) before December 2, 1994, and who is working in a nuclear pharmacy would qualify as an experienced nuclear pharmacist. An experienced nuclear pharmacist need not comply with the requirements on preceptor statement (Section 35.980(b)(2)) and recentness of training (Section 35.972) to qualify as an authorized nuclear pharmacist.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
Subpart K--Other Medical Uses of Byproduct Material or Radiation From Byproduct Material	Sec. 35.1000 Other medical uses of byproduct material or radiation from byproduct material.	Subpart K--Other Medical Uses of Byproduct Material or Radiation from Byproduct Material
A licensee may use byproduct material or a radiation source approved for medical use which is not specifically addressed in subparts D through H of this part if--	(a) The applicant or licensee has submitted the information required by Sec. 35.12(d); and (b) The applicant or licensee has received written approval from the Commission in a license and uses the material in accordance with the regulations and specific conditions the Commission considers necessary for the medical use of the material.	§ 35.1000 Other medical uses of byproduct material or radiation from byproduct material.
		A licensee may use byproduct material or a radiation source approved for medical use which is not specifically addressed in subparts D through H of this part if --
		(a) The applicant or licensee has submitted the information required by § 35.12(b) through (d); and (b) The applicant or licensee has received written approval from the Commission in a license or license amendment and uses the material in accordance with the regulations and specific conditions the Commission considers necessary for the medical use of the material.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart L--Records

Subpart L--Records

Sec. 35.2024 Records of authority and responsibilities for radiation protection programs.

§ 35.2024 Records of authority and responsibilities for radiation protection programs.

(a) A licensee shall retain a record of actions taken by the licensee's management in accordance with Sec. 35.24(a) for 5 years. The record must include a summary of the actions taken and a signature of licensee management.

(a) A licensee shall retain a record of actions taken by the licensee's management in accordance with § 35.24(a) for 5 years. The record must include a summary of the actions taken and a signature of licensee management.

(b) The licensee shall retain a current copy of the authorities, duties and responsibilities of the radiation safety officer, as required by Sec. 35.24(d), and a signed copy of the radiation safety officer's willingness to be responsible for implementing the radiation safety program, as required by Sec. 35.24(b). The records must include the signature of the radiation safety officer and licensee management.

(b) The licensee shall retain a current copy of the authority, duties, and responsibilities of the Radiation Safety Officer as required by § 35.24(e), and a signed copy of the Radiation Safety Officer's agreement to be responsible for implementing the radiation safety program, as required by § 35.24(b). The records must include the signature of the Radiation Safety Officer and licensee management.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2026 Records of radiation protection program safety changes.</p> <p>A licensee shall retain a record of each radiation protection program change made in accordance with Sec. 35.26(a) for 5 years. The record must include a copy of the old and new procedures; the effective date of the change; and the signature of the radiation safety officer and the licensee management that reviewed and approved the change.</p>	<p>§ 35.2026 Records of radiation protection program changes.</p> <p>A licensee shall retain a record of each radiation protection program change made in accordance with § 35.26(a) for 5 years. The record must include a copy of the old and new procedures; the effective date of the change; and the signature of the licensee management that reviewed and approved the change.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
Sec. 35.2040 Records of written directives.	§ 35.2040 Records of written directives.	
A licensee shall retain a copy of each written directive as required by Sec. 35.40 for 3 years.	A licensee shall retain a copy of each written directive as required by § 35.40 for 3 years.	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2045 Records of medical events.</p> <p>A licensee shall retain a record of medical events reported pursuant to Sec. 35.3045 for 3 years. The record must contain the licensee's name, names of all the individuals involved, the affected or potentially affected individual's social security number or other identification number if one has been assigned, a brief description of the medical event, why it occurred, the effect on the individual, and the actions taken to prevent recurrence.</p>	<p>§ 35.2045 Records of medical events.</p> <p>(a) A licensee shall retain a record of medical events reported in accordance with § 35.3045 for 3 years.</p> <p>(b) The record must include--</p> <p>(1) The licensee's name;</p> <p>(2) Names of all the individuals involved;</p> <p>(3) The affected or potentially affected individual's social security number or other identification number if one has been assigned;</p> <p>(4) A brief description of the medical event and why it occurred;</p> <p>(5) The effect, if any, on the individual; and</p> <p>(6) The actions, if any, taken or planned to prevent recurrence.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

§ 35.2047 Record of a dose to an embryo/fetus or a nursing child.

(a) A licensee shall retain a record of a dose to an embryo/fetus or a nursing child reported in accordance with § 35.3047 for 3 years.

(b) The record must include--

(1) The licensee's name;

(2) Names of all the individuals involved;

(3) The affected or potentially affected individual's social security number or other identification number if one has been assigned;

(4) A brief description of the event and why it occurred;

(5) The effect, if any, on the embryo/fetus or nursing child; and

(6) The actions, if any, taken or planned to prevent recurrence.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2060 Records of instrument calibrations.</p>	<p>§ 35.2060 Records of calibrations of instruments to measure the activity of unsealed byproduct materials.</p>	
<p>A licensee shall maintain a record of instrument calibrations required by Secs. 35.60 and 35.62 for 3 years. The records must include--</p>	<p>A licensee shall maintain a record of instrument calibrations required by § 35.60 for 3 years. The records must include the model and serial number of the instrument, the date of the calibration, the results of the calibration, and the name of the individual who performed the calibration.</p>	
<p>(a) For constancy, the model and serial number of the instrument, the identity of the radionuclide contained in the check source, the date of the check, and the activity measured, and the name of the individual who performed the check;</p>		
<p>(b) For accuracy, the model and serial number of the instrument, the model and serial number of each source used, the identity of the radionuclide contained in the source and its activity, the date of the test, and the results of the test, and the name of the individual who performed the test;</p>		
<p>(c) For linearity, the model and serial number of the instrument, the calculated activities, the measured</p>		

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

activities, and the date of the test, and the name of the individual who performed the test; and

(d) For geometric dependence, the model and serial number of the instrument, the configuration of the source measured, the activity measured for each volume measured, and the date of the test, and the name of the individual who performed the test.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2061 Records of radiation survey instrument calibrations.</p> <p>A licensee shall maintain a record of radiation survey instrument calibrations required by Sec. 35.61 for 3 years. The record must include--</p> <p>(a) A description of the calibration procedure; and</p> <p>(b) The date of the calibration, a description of the source used and the certified exposure rates from the source, and the rates indicated by the instrument being calibrated, the correction factors deduced from the calibration data, and the name of the individual who performed the calibration.</p>	<p>§ 35.2061 Records of radiation survey instrument calibrations.</p> <p>A licensee shall maintain a record of radiation survey instrument calibrations required by § 35.61 for 3 years. The record must include the model and serial number of the instrument, the date of the calibration, the results of the calibration, and the name of the individual who performed the calibration.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2063 Records of dosages of unsealed byproduct material for medical use.</p> <p>(a) A licensee shall maintain a record of dosage determinations required by Sec. 35.63 for 3 years.</p> <p>(b) To satisfy this requirement, the record must contain the--</p> <p>(1) Radionuclide, generic name, trade name, or abbreviation of the radiopharmaceutical, and its lot number;</p> <p>(2) Patient's or human research subject's name, or identification number if one has been assigned;</p> <p>(3) Prescribed dosage and activity of the dosage at the time of determination, or a notation that the total activity is less than 1.1 MBq (30 <greek-m>Ci);</p> <p>(4) Date and time of the dosage determination; and</p> <p>(5) Name of the individual who determined the dosage.</p>	<p>§ 35.2063 Records of dosages of unsealed byproduct material for medical use.</p> <p>(a) A licensee shall maintain a record of dosage determinations required by § 35.63 for 3 years.</p> <p>(b) The record must contain--</p> <p>(1) The radiopharmaceutical;</p> <p>(2) The patient's or human research subject's name, or identification number if one has been assigned;</p> <p>(3) The prescribed dosage, the determined dosage, or a notation that the total activity is less than 1.1 MBq (30 µCi);</p> <p>(4) The date and time of the dosage determination; and</p> <p>(5) The name of the individual who determined the dosage.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2067 Records of possession of sealed sources and brachytherapy sources.</p> <p>(a) A licensee shall retain records of leak tests required by Sec. 35.67(b) for 3 years. The records must contain the model number, and serial number if one has been assigned, of each source tested, the identity of each source radionuclide and its estimated activity, the measured activity of each test sample, a description of the method used to measure each test sample, the date of the test, and the name of the individual who performed the test.</p> <p>(b) A licensee shall retain records of the semi-annual physical inventory of sealed sources and brachytherapy sources required by Sec. 35.67(g) for 3 years. The inventory records must contain the model number of each source, and serial number if one has been assigned, the identity of each source radionuclide and its nominal activity, the location of each source, and the name of</p>	<p>§ 35.2067 Records of possession of sealed sources and brachytherapy sources.</p> <p>(a) A licensee shall retain records of leak tests required by § 35.67(b) for 3 years. The records must include the model number, and serial number if one has been assigned, of each source tested, the identity of each source radionuclide and its estimated activity, the results of the test, the date of the test, and the name of the individual who performed the test.</p> <p>(b) A licensee shall retain records of the semi-annual physical inventory of sealed sources and brachytherapy sources required by § 35.67(g) for 3 years. The inventory records must contain the model number of each source, and serial number if one has been assigned, the identity of each source radionuclide and its nominal activity, the location of each source, and the name</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

the individual who performed
the inventory.

of the individual who
performed the inventory.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2070 Records of surveys for ambient radiation exposure rate.</p> <p>A licensee shall retain a record of each survey required by Sec. 35.70 for 3 years. The record must include the date of the survey, a plan of each area surveyed, the trigger level established for each area, the detected dose rate at several points in each area expressed in millirem per hour or the removable contamination in each area expressed in disintegrations per minute per 100 square centimeters, the instrument used to make the survey or analyze the samples, and the name of the individual who performed the survey.</p>	<p>§ 35.2070 Records of surveys for ambient radiation exposure rate.</p> <p>A licensee shall retain a record of each survey required by § 35.70 for 3 years. The record must include the date of the survey, the results of the survey, the instrument used to make the survey, and the name of the individual who performed the survey.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2075 Records of the release of individuals containing radiopharmaceuticals or implants.</p>	<p>§ 35.2075 Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material.</p>	
<p>(a) A licensee shall retain records of the release of individuals containing pharmaceuticals or implants in accordance with Sec. 35.75 for 3 years after the date of release</p>	<p>(a) A licensee shall retain a record of the basis for authorizing the release of an individual in accordance with § 35.75, if the total effective dose equivalent is calculated by --</p>	
<p>(b) A licensee shall retain a record in accordance with paragraph (a) of this section that describes the basis for authorizing the release of individuals if the total effective dose equivalent is calculated by--</p>	<p>(1) Using the retained activity rather than the activity administered;</p> <p>(2) Using an occupancy factor less than 0.25 at 1 meter;</p> <p>(3) Using the biological or effective half-life; or</p>	
<p>(1) Using the retained activity rather than the activity administered;</p>	<p>(1) Using the retained activity rather than the activity administered;</p>	
<p>(2) Using an occupancy factor less than 0.25 at 1 meter;</p>	<p>(2) Using an occupancy factor less than 0.25 at 1 meter;</p>	
<p>(3) Using the biological or effective half-life; or</p>	<p>(3) Using the biological or effective half-life; or</p>	
<p>(4) Considering the shielding by tissue.</p>	<p>(4) Considering the shielding by tissue.</p>	
<p>(c) A licensee shall retain a record that the instructions required by Sec.</p>	<p>(b) A licensee shall retain a record that the instructions required by §</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

35.75(b) were provided to a breast-feeding woman if the radiation dose to the infant or child from continued breast-feeding could result in a total effective dose equivalent exceeding 5 mSv (0.5 rem).

35.75(b) were provided to a breast-feeding female if the radiation dose to the infant or child from continued breast-feeding could result in a total effective dose equivalent exceeding 5 mSv (0.5 rem).

(c) The records required by paragraphs (a) and (b) of this section must be retained for 3 years after the date of release of the individual.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2080 Records of administrative and technical requirements that apply to the provision of mobile services.</p> <p>(a) A licensee shall retain a copy of the letter(s) that permits the use of byproduct material at a client's address of use, in accordance with Sec. 35.80(a)(1). This letter must clearly delineate the authority and responsibility of each entity and must be retained for 3 years after the last provision of service.</p> <p>(b) A licensee shall retain the record of each survey required by Sec. 35.80(a)(4) for 3 years. The record must include the date of the survey, a plan of each area that was surveyed, the measured dose rate at several points in each area of use expressed in millirem per hour, the instrument used to make the survey, and the name of the individual who performed the survey.</p>	<p>§ 35.2080 Records of administrative and technical requirements that apply to the provision of mobile medical services.</p> <p>(a) A licensee shall retain a copy of the letter(s) that permits the use of byproduct material at a client's address, as required by § 35.80(a)(1). This letter must clearly delineate the authority and responsibility of the licensee and the client and must be retained for 3 years after the last provision of service.</p> <p>(b) A licensee shall retain the record of each survey required by § 35.80(a)(4) for 3 years. The record must include the date of the survey, the results of the survey, the instrument used to make the survey, and the name of the individual who performed the survey.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2092 Records of waste disposal.</p> <p>A licensee shall maintain records of the disposal of licensed materials made in accordance with Sec. 35.92 for 3 years. The record must include the date of the disposal, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.</p>	<p>§ 35.2092 Records of decay-in-storage.</p> <p>A licensee shall maintain records of the disposal of licensed materials, as required by § 35.92, for 3 years. The record must include the date of the disposal, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2204 Records of molybdenum-99 concentrations.</p> <p>A licensee shall maintain a record of the molybdenum-99 concentration tests required by Sec. 35.204(b) for 3 years. The record must include, for each measured elution of technetium-99m, the ratio of the measures expressed as microcuries of molybdenum per millicurie of technetium, the time and date of the measurement, and the name of the individual who made the measurement.</p>	<p>§ 35.2204 Records of molybdenum-99 concentrations.</p> <p>A licensee shall maintain a record of the molybdenum-99 concentration tests required by § 35.204(b) for 3 years. The record must include, for each measured elution of technetium-99m, the ratio of the measures expressed as kilobecquerel of molybdenum-99 per megabecquerel of technetium-99m (microcuries of molybdenum per millicurie of technetium), the time and date of the measurement, and the name of the individual who made the measurement.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2310 Records of instruction and training.

A licensee shall maintain a record of instructions and training required by Secs. 35.310, 35.410, and 35.610 for 3 years. The record must include a list of the topics covered, the date of the instruction or training, the name(s) of the attendee(s), and the name(s) of the individual(s) who provided the instruction.

§ 35.2310 Records of safety instruction.

A licensee shall maintain a record of safety instructions required by §§ 35.310, 35.410, and 35.610 for 3 years. The record must include a list of the topics covered, the date of the instruction, the name(s) of the attendee(s), and the name(s) of the individual(s) who provided the instruction.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2404 Records of radiation surveys of patients and human research subjects.

A licensee shall maintain a record of the radiation surveys of patients and human research subjects required by Secs. 35.404 and 35.604 for 3 years. Each record must include the date, location, and results of the survey, an identifier for the patient or the human research subject, the survey instrument used, and the name of the individual who made the survey.

§ 35.2404 Records of surveys after source implant and removal.

A licensee shall maintain a record of the surveys required by §§ 35.404 and 35.604 for 3 years. Each record must include the date and results of the survey, the survey instrument used, and the name of the individual who made the survey.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2406 Records of brachytherapy source inventory.</p> <p>(a) A licensee shall maintain a record of brachytherapy source accountability required by Sec. 35.406 for 3 years.</p> <p>(b) For temporary implants, the record must include--</p> <p>(1) The number and activity of sources removed from storage, the time and date they were removed from storage, the name of the individual who removed them from storage, and the location of use; and</p> <p>(2) The number and activity of sources returned to storage, the time and date they were returned to storage, and the name of the individual who returned them from storage.</p> <p>(c) For permanent implants, the record must include--</p> <p>(1) The number and activity of sources removed from storage, the date they were removed from storage, and the name of the individual</p>	<p>§ 35.2406 Records of brachytherapy source accountability.</p> <p>(a) A licensee shall maintain a record of brachytherapy source accountability required by § 35.406 for 3 years.</p> <p>(b) For temporary implants, the record must include --</p> <p>(1) The number and activity of sources removed from storage, the time and date they were removed from storage, the name of the individual who removed them from storage, and the location of use; and</p> <p>(2) The number and activity of sources returned to storage, the time and date they were returned to storage, and the name of the individual who returned them to storage.</p> <p>(c) For permanent implants, the record must include --</p> <p>(1) The number and activity of sources removed from storage, the date they were removed from storage, and the name of the individual</p>	<p>§ 35.2406 Records of brachytherapy source accountability.</p> <p>(a) A licensee shall maintain a record of brachytherapy source accountability required by § 35.406 for 3 years.</p> <p>(b) For temporary implants, the record must include --</p> <p>(1) The number and activity of sources removed from storage, the time and date they were removed from storage, the name of the individual who removed them from storage, and the location of use; and</p> <p>(2) The number and activity of sources returned to storage, the time and date they were returned to storage, and the name of the individual who returned them to storage.</p> <p>(c) For permanent implants, the record must include --</p> <p>(1) The number and activity of sources removed from storage, the date they were removed from storage, and the name of the individual</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

who removed them from storage;
(2) The number and activity of sources returned to storage, the date they were returned to storage, and the name of the individual who returned them to storage; and
(3) The number and activity of sources permanently implanted in the patient or human research subject.

who removed them from storage;
(2) The number and activity of sources not implanted, the date they were returned to storage, and the name of the individual who returned them to storage; and
(3) The number and activity of sources permanently implanted in the patient or human research subject.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2432 Records of full calibrations on brachytherapy sources.

A licensee shall maintain a record of the full calibrations on brachytherapy sources required by Sec. 35.432 for 3 years after the last use of the source. The record must include the date of the calibration; the manufacturer's name, model number, and serial number for the source and instruments used to calibrate the source; the source output; source positioning accuracy within applicators; and the name of the individual or the source manufacturer who performed the calibration.

§ 35.2432 Records of calibration measurements of brachytherapy sealed sources.

(a) A licensee shall maintain a record of the calibrations of brachytherapy sources required by § 35.432 for 3 years after the last use of the source.

(b) The record must include--

(1) The date of the calibration;

(2) The manufacturer's name, model number, and serial number for the source and the instruments used to calibrate the source;

(3) The source output or activity;

(4) The source positioning accuracy within the applicators; and

(5) The signature of the authorized medical physicist.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2605 Records of installation, maintenance, and repair.

A licensee shall retain a record of the installation, maintenance, and repair of therapeutic medical devices as required by Sec. 35.605 for 3 years. For each installation, maintenance, and repair, the record must include the date, description of the service, and name(s) of the individual(s) who performed the work.

§ 35.2605 Records of installation, maintenance, adjustment, and repair.

A licensee shall retain a record of the installation, maintenance, adjustment, and repair of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units as required by § 35.605 for 3 years. For each installation, maintenance, adjustment and repair, the record must include the date, description of the service, and name(s) of the individual(s) who performed the work.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2630 Records of dosimetry equipment.</p> <p>(a) A licensee shall retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with Sec. 35.630 for the duration of the license.</p> <p>(b) For each calibration, intercomparison, or comparison, the record must include--</p> <p>(1) The date;</p> <p>(2) The model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by paragraphs (a) and (b) of Sec. 35.630;</p> <p>(3) The correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and</p> <p>(4) The name(s) of the individual(s) who performed the calibration, intercomparison, or comparison.</p>	<p>§ 35.2630 Records of dosimetry equipment.</p> <p>(a) A licensee shall retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with § 35.630 for the duration of the license.</p> <p>(b) For each calibration, intercomparison, or comparison, the record must include --</p> <p>(1) The date;</p> <p>(2) The manufacturer's name, model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by paragraphs (a) and (b) of § 35.630;</p> <p>(3) The correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and</p> <p>(4) The names of the individuals who performed the calibration, intercomparison, or comparison.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2632 Records of teletherapy full calibrations.</p> <p>(a) A licensee shall maintain a record of the teletherapy full calibrations required by Sec. 35.632 for 3 years.</p> <p>(b) The record must include--</p> <p>(1) The date of the calibration;</p> <p>(2) The manufacturer's name, model number, and serial number for the teletherapy unit, source, and instruments used to calibrate the teletherapy unit;</p> <p>(3) Tables that describe the output of the unit over the range of field sizes and for the range of distances used in radiation therapy;</p> <p>(4) A determination of the coincidence of the radiation field and the field</p>	<p>§ 35.2632 Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations.</p> <p>(a) A licensee shall maintain a record of the teletherapy unit, remote afterloader unit, and gamma stereotactic radiosurgery unit full calibrations required by §§ 35.632, 35.633, and 35.635 for 3 years.</p> <p>(b) The record must include --</p> <p>(1) The date of the calibration;</p> <p>(2) The manufacturer's name, model number, and serial number of the teletherapy, remote afterloader, and gamma stereotactic radiosurgery unit(s), the source, and the instruments used to calibrate the unit(s);</p> <p>(3) The results and an assessment of the full calibrations;</p> <p>(4) The results of the autoradiograph required for low dose-rate remote afterloader units; and</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	indicated by the light beam localizing device; (5) An assessment of timer accuracy and linearity; (6) The calculated on-off error; (7) The estimated accuracy of each distance measuring and localization device; and (8) The signature of the authorized medical physicist who performed the full calibration.	(5) The signature of the authorized medical physicist who performed the full calibration.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2633 Records of remote afterloader full calibrations.

(a) A licensee shall maintain a record of the remote afterloader full calibrations required by Sec. 35.633 for 3 years.

(b) The record must include--

(1) The date of the calibration;

(2) The manufacturer's name, model number, and serial number for the remote afterloader, source, and instruments used to calibrate the unit; the source output;

(3) An assessment of timer accuracy and linearity, source positioning accuracy, source guide tube and connector lengths, and source retraction functionality; and

(4) The signature of the authorized medical physicist who performed the full calibration.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2635 Records of gamma stereotactic radiosurgery unit full calibrations.

(a) A licensee shall maintain a record of the gamma stereotactic radiosurgery full calibrations required by Sec. 35.635 for 3 years.

(b) The record must include--

(1) The date of the calibration;

(2) The manufacturer's name, model number, and serial number for the gamma stereotactic radiosurgery unit, source, and instruments used to calibrate the unit;

(3) The unit output;

(4) An assessment of the relative helmet factors, isocenter coincidence, timer accuracy and linearity, on-off error, and trunnion centricity; and

(5) The signature of the authorized medical physicist who performed the full calibration.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2642 Records of periodic spot-checks for teletherapy units.</p> <p>(a) A licensee shall retain a record of each periodic spot-check for teletherapy units required by Sec. 35.642 for 3 years.</p> <p>(b) The record must include --</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number for the teletherapy unit, source and instrument used to measure the output of the teletherapy unit;</p> <p>(3) An assessment of timer linearity and constancy;</p> <p>(4) The calculated on-off error;</p> <p>(5) A determination of the coincidence of the radiation field and the field indicated by the light beam localizing device;</p> <p>(6) The determined accuracy of each distance measuring and localization device;</p>	<p>§ 35.2642 Records of periodic spot-checks for teletherapy units.</p> <p>(a) A licensee shall retain a record of each periodic spot-check for teletherapy units required by § 35.642 for 3 years.</p> <p>(b) The record must include --</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number of the teletherapy unit, source and instrument used to measure the output of the teletherapy unit;</p> <p>(3) An assessment of timer linearity and constancy;</p> <p>(4) The calculated on-off error;</p> <p>(5) A determination of the coincidence of the radiation field and the field indicated by the light beam localizing device;</p> <p>(6) The determined accuracy of each distance measuring and localization device;</p>	<p>§ 35.2642 Records of periodic spot-checks for teletherapy units.</p> <p>(a) A licensee shall retain a record of each periodic spot-check for teletherapy units required by § 35.642 for 3 years.</p> <p>(b) The record must include --</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number of the teletherapy unit, source and instrument used to measure the output of the teletherapy unit;</p> <p>(3) An assessment of timer linearity and constancy;</p> <p>(4) The calculated on-off error;</p> <p>(5) A determination of the coincidence of the radiation field and the field indicated by the light beam localizing device;</p> <p>(6) The determined accuracy of each distance measuring and localization device;</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(7) The difference between the anticipated output and the measured output;

(8) Notations indicating the operability of each entrance door electrical interlock, each electrical or mechanical stop, each source exposure indicator light, and the viewing and intercom system and doors; and

(9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

(7) The difference between the anticipated output and the measured output;

(8) Notations indicating the operability of each entrance door electrical interlock, each electrical or mechanical stop, each source exposure indicator light, and the viewing and intercom system and doors; and

(9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2643 Records of periodic spot-checks for remote afterloaders.</p> <p>(a) A licensee shall retain a record of each spot-check for remote afterloaders required by Secs. 35.643 and 35.644 for 3 years.</p> <p>(b) The record must include--</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number for the remote afterloader, source, and instrument used to measure the output of the remote afterloader;</p> <p>(3) The difference between the anticipated output and the measured output;</p> <p>(4) Notations indicating the operability of each entrance door electrical interlock, source retraction mechanism, radiation monitors, source exposure indicator lights, viewing and intercom systems if applicable, applicators and connectors, and source positioning accuracy; and</p>	<p>§ 35.2643 Records of periodic spot-checks for remote afterloader units.</p> <p>(a) A licensee shall retain a record of each spot-check for remote afterloader units required by § 35.643 for 3 years.</p> <p>(b) The record must include, as applicable --</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number for the remote afterloader unit and source;</p> <p>(3) An assessment of timer accuracy;</p> <p>(4) Notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom systems, and clock and decayed source activity in the unit's computer; and</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(5) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

(5) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2645 Records of periodic spot-checks for gamma stereotactic radiosurgery units.</p> <p>(a) A licensee shall retain a record of each spot-check for gamma stereotactic radiosurgery units required by Sec. 35.645 for 3 years.</p> <p>(b) The record must include--</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number for the gamma stereotactic radiosurgery unit and the instrument used to measure the output of the unit;</p> <p>(3) The measured source output and source output against computer calculations;</p>	<p>§ 35.2645 Records of periodic spot-checks for gamma stereotactic radiosurgery units.</p> <p>(a) A licensee shall retain a record of each spot-check for gamma stereotactic radiosurgery units required by § 35.645 for 3 years.</p> <p>(b) The record must include --</p> <p>(1) The date of the spot-check;</p> <p>(2) The manufacturer's name, model number, and serial number for the gamma stereotactic radiosurgery unit and the instrument used to measure the output of the unit;</p> <p>(3) An assessment of timer linearity and accuracy;</p> <p>(4) The calculated on-off error;</p> <p>(5) A determination of trunnion centricity;</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(4) Notations indicating the operability of radiation monitors, helmet microswitches, emergency timing circuits, emergency off buttons, electrical interlocks, source exposure indicator lights, viewing and intercom systems, timer termination systems, hydraulic cutoff switch and stereotactic frames and localizing devices (trunnions); and

(5) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Sec. 35.2647 Records of additional technical requirements for mobile remote afterloaders.

(a) A licensee shall retain a record of each check for mobile remote afterloaders required by Sec. 35.647 for 3 years.

(b) The record must include--

(1) The date of the check;

(2) The manufacturer's name, model number, and serial number of the remote afterloader;

(3) Notations accounting for all sources before the licensee departs from a facility;

(4) Notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom system, applicators and connectors, and source positioning accuracy; and

(5) The signature of the individual who performed the check.

§ 35.2647 Records of additional technical requirements for mobile remote afterloader units.

(a) A licensee shall retain a record of each check for mobile remote afterloader units required by § 35.647 for 3 years.

(b) The record must include --

(1) The date of the check;

(2) The manufacturer's name, model number, and serial number of the remote afterloader unit;

(3) Notations accounting for all sources before the licensee departs from a facility;

(4) Notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom system, applicators, source transfer tubes, and transfer tube applicator interfaces, and source positioning accuracy; and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(5) The signature of
the individual who performed
the check.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2652 Records of surveys of therapeutic treatment units.</p> <p>(a) A licensee shall maintain a record of radiation surveys of treatment units made in accordance with Sec. 35.652 for the duration of use of the unit.</p> <p>(b) The record must include--</p> <p>(1) The date of the measurements;</p> <p>(2) The manufacturer's name, model number and serial number of the treatment unit, source, and instrument used to measure radiation levels;</p> <p>(3) Each dose rate measured around the source while the unit is in the off position and the average of all measurements; and</p> <p>(4) The signature of the individual who performed the test.</p>	<p>§ 35.2652 Records of surveys of therapeutic treatment units.</p> <p>(a) A licensee shall maintain a record of radiation surveys of treatment units made in accordance with § 35.652 for the duration of use of the unit.</p> <p>(b) The record must include --</p> <p>(1) The date of the measurements;</p> <p>(2) The manufacturer's name, model number and serial number of the treatment unit, source, and instrument used to measure radiation levels;</p> <p>(3) Each dose rate measured around the source while the unit is in the off position and the average of all measurements; and</p> <p>(4) The signature of the individual who performed the test.</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
<p>Sec. 35.2655 Records of 5-year inspection for teletherapy and gamma stereotactic surgery units.</p> <p>(a) A licensee shall maintain a record of the 5-year inspections for teletherapy and gamma stereotactic radiosurgery units required by Sec. 35.655 for the duration of use of the unit.</p> <p>(b) The record must contain--</p> <p>(1) The inspector's radioactive materials license number;</p> <p>(2) The date of inspection;</p> <p>(3) The manufacturer's name and model number and serial number of both the treatment unit and source;</p> <p>(4) A list of components inspected and serviced, and the type of service; and</p> <p>(5) The signature of the inspector.</p>	<p>§ 35.2655 Records of 5-year inspection for teletherapy and gamma stereotactic radiosurgery units.</p> <p>(a) A licensee shall maintain a record of the 5-year inspections for teletherapy and gamma stereotactic radiosurgery units required by § 35.655 for the duration of use of the unit.</p> <p>(b) The record must contain --</p> <p>(1) The inspector's radioactive materials license number;</p> <p>(2) The date of inspection;</p> <p>(3) The manufacturer's name and model number and serial number of both the treatment unit and source;</p> <p>(4) A list of components inspected and serviced, and the type of service; and</p> <p>(5) The signature of the inspector.</p> <p>(6) The difference between the anticipated output and the measured output;</p>	

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(7) An assessment of source output against computer calculations;

(8) Notations indicating the operability of radiation monitors, helmet microswitches, emergency timing circuits, emergency off buttons, electrical interlocks, source exposure indicator lights, viewing and intercom systems, timer termination, treatment table retraction mechanism, and stereotactic frames and localizing devices (trunnions); and

(9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.

§ 35.2433 Records of decay of strontium-90 sources for ophthalmic treatments.

(a) A licensee shall maintain a record of the activity of a strontium-90 source required by § 35.433 for the life of the source.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) The record must include--
(1) The initial activity of the source and date; and
(2) For each decay calculation, the date and the source activity as determined under § 35.432.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart M--Reports

Subpart M--Reports

Sec. 35.3045 Reports of medical events.

§ 35.3045 Report and notification of a medical event.

(a) A licensee shall report any administration, except for administrations resulting from a direct intervention of a patient or human research subject that could not have been reasonably prevented by the licensee, that results in either--

(a) A licensee shall report any event, except for an event that results from patient intervention, in which the administration of byproduct material or radiation from byproduct material results in --

(1) A dose that differs from the prescribed dose by more than 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin; and

(1) A dose that differs from the prescribed dose by more than 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin; and

(i) The total dose or dosage delivered differs from the prescribed dose or dosage by 20 percent or more; or

(i) The total dose delivered differs from the prescribed dose by 20 percent or more;

(ii) The fractionated dose delivered differs from the prescribed dose, for a single fraction, by 50 percent or more.

(ii) The total dosage delivered differs from the prescribed dosage by 20 percent or more or falls outside the prescribed dosage range; or

(iii) The fractionated dose delivered differs from the prescribed dose, for a single fraction, by 50 percent or more.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(2) A dose that exceeds 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin from any of the following—

(i) An administration of a wrong pharmaceutical;

(ii) An administration of a radiopharmaceutical by the wrong route of administration;

(iii) An administration of a dose or dosage to the wrong individual or human research subject;

(iv) An administration of a dose or dosage delivered by the wrong treatment mode; or

(v) A leaking sealed source.

(3) A dose to the skin or an organ or tissue other than the treatment site that exceeds by 0.5 Sv (50 rem) to an organ or tissue and 20 percent the dose expected from the administration defined in the written directive.

(2) A dose that exceeds 0.05 Sv (5 rem) effective dose equivalent, 0.5 Sv (50 rem) to an organ or tissue, or 0.5 Sv (50 rem) shallow dose equivalent to the skin from any of the following —

(i) An administration of a wrong radioactive drug containing byproduct material;

(ii) An administration of a radioactive drug containing byproduct material by the wrong route of administration;

(iii) An administration of a dose or dosage to the wrong individual or human research subject;

(iv) An administration of a dose or dosage delivered by the wrong mode of treatment; or

(v) A leaking sealed source.

(3) A dose to the skin or an organ or tissue other than the treatment site that exceeds by 0.5 Sv (50 rem) to an organ or tissue and 50 percent or more of the dose expected from the administration defined in the written directive (excluding, for permanent implants, seeds that were implanted in the correct site

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(b) The licensee shall notify by telephone the NRC Operations Center (301-951-0550) no later than the next calendar day after discovery of the medical event.

(c) The licensee shall submit a written report to the appropriate NRC Regional Office listed in Sec. 30.6 of this chapter within 15 days after discovery of the medical event.

- (1) The written report must include--
- (i) The licensee's name;
 - (ii) The name of the prescribing physician;
 - (iii) A brief description of the event;
 - (iv) Why the event occurred;

but migrated outside the treatment site).

(b) A licensee shall report any event resulting from intervention of a patient or human research subject in which the administration of byproduct material or radiation from byproduct material results or will result in an unintended permanent functional damage to an organ or a physiological system, as determined by a physician.

(c) The licensee shall notify by telephone the NRC Operations Center¹ no later than the next calendar day after discovery of the medical event.

(d) The licensee shall submit a written report to the appropriate NRC Regional Office listed in § 30.6 of this chapter within 15 days after discovery of the medical event.

- (1) The written report must include --
- (i) The licensee's name;
 - (ii) The name of the prescribing physician;
 - (iii) A brief description of the event;
 - (iv) Why the event occurred;

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(v) The effect on the individual(s) who received the administration;

(vi) What improvements are needed to prevent recurrence;

(vii) Actions taken to prevent recurrence;

(viii) Whether the licensee notified the individual (or the individual's responsible relative or guardian), and if not, why not; and

(ix) If there was notification, what information was provided.

(2) The report must not contain the individual's name or any other information that could lead to identification of the individual.

(d) The licensee shall notify the referring physician and also notify the individual affected by the medical event no later than 24 hours after its discovery, unless the referring physician personally informs the licensee either that he or she will inform the individual or that, based on medical judgement, telling the individual would be harmful. The

(v) The effect, if any, on the individual(s) who received the administration;

(vi) What actions, if any, have been taken or are planned to prevent recurrence;

(vii) Whether the licensee notified the individual (or the individual's responsible relative or guardian), and if not, why not; and

(viii) If there was notification, what information was provided.

(2) The report may not contain the individual's name or any other information that could lead to identification of the individual.

(e) The licensee shall notify the referring physician and also notify the individual affected by the medical event no later than 24 hours after its discovery, unless the referring physician personally informs the licensee either that he or she will inform the individual or that, based on medical judgement, telling the individual would be harmful. The

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

licensee is not required to notify the individual without first consulting the referring physician. If the referring physician or the affected individual cannot be reached within 24 hours, the licensee shall notify the individual as soon as possible thereafter. The licensee may not delay any appropriate medical care for the individual, including any necessary remedial care as a result of the medical event, because of any delay in notification. To meet the requirements of this section, the notification of the individual receiving the medical event may be made instead to that individual's responsible relative or guardian, when appropriate.

(e) If the individual was notified pursuant to paragraph (d) of this section, the licensee shall also furnish, within 15 days after discovery of the medical event, a written report to the individual by sending either--

(1) A copy of the report that was submitted to the NRC;
or

(2) A brief description of both the event and the consequences as they may affect the individual.

licensee is not required to notify the individual without first consulting the referring physician. If the referring physician or the affected individual cannot be reached within 24 hours, the licensee shall notify the individual as soon as possible thereafter. The licensee may not delay any appropriate medical care for the individual, including any necessary remedial care as a result of the medical event, because of any delay in notification. To meet the requirements of this section, the notification of the individual receiving the medical event may be made instead to that individual's responsible relative or guardian.

(f) If the individual was notified under paragraph (e) of this section, the licensee shall also furnish a written report to the individual within 15 days after discovery of the medical event. A licensee may send either --

(1) A copy of the report that was submitted to the NRC;
or

(2) A brief description of both the event and the

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(f) Aside from the notification requirement, nothing in this section affects any rights or duties of licensees and physicians in relation to each other, to individuals affected by the medical event, or to that individual's responsible relatives or guardians.

consequences as they may affect the individual.

(g) Aside from the notification requirement, nothing in this section affects any rights or duties of licensees and physicians in relation to each other, to individuals affected by the medical event, or to that individual's responsible relatives or guardians.

(h) A licensee shall retain a record of a medical event in accordance with § 35.2045.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE	PROPOSED RULE	FINAL RULE
	<p>Sec. 35.3047 Report of a dose to an embryo/fetus or a nursing child.</p> <p>(a) A licensee shall report any dose to an embryo/fetus that is greater than 5 mSv (500 mrem) absorbed dose that is a result of an administration of byproduct material or radiation from byproduct material to a pregnant individual unless the dose to the embryo/fetus was specifically approved, in advance, by the authorized user.</p> <p>(b) A licensee shall report any dose to a nursing child that is greater than 5 mSv (500 mrem) total effective dose equivalent that is a result of an administration of byproduct material to a breast feeding individual.</p> <p>(c) The licensee shall notify by telephone the NRC Operations Center within 5 days after discovery of a dose to the</p>	<p>§ 35.3047 Report and notification of a dose to an embryo/fetus or a nursing child.</p> <p>(a) A licensee shall report any dose to an embryo/fetus that is greater than 50 mSv (5 rem) dose equivalent that is a result of an administration of byproduct material or radiation from byproduct material to a pregnant individual unless the dose to the embryo/fetus was specifically approved, in advance, by the authorized user.</p> <p>(b) A licensee shall report any dose to a nursing child that is a result of an administration of byproduct material to a breast-feeding individual that --</p> <p>(1) Is greater than 50 mSv (5 rem) total effective dose equivalent; or</p> <p>(2) Has resulted in unintended permanent functional damage to an organ or a physiological system of the child, as determined by a physician.</p> <p>(c) The licensee shall notify by telephone the NRC Operations Center no later than the next calendar day after discovery of a dose to the embryo/fetus or nursing child that</p>

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

embryo/fetus or nursing child that requires a report in paragraphs (a) or (b) in this section.

(d) The licensee shall submit a written report to the appropriate NRC Regional Office listed in Sec. 30.6 no later than 15 days after discovery of a dose to the embryo/fetus or nursing child that requires a report in paragraphs (a) or (b) in this section.

(1) The written report must include--
(i) The licensee's name;
(ii) The name of the prescribing physician;
(iii) A brief description of the event;
(iv) Why the event occurred;
(v) The effect on the embryo/fetus or the nursing child;
(vi) What improvements are needed to prevent recurrence; and
(vii) Actions taken to prevent recurrence.

(2) The report must not contain the individual's or child's name or any other information that could lead to identification of the individual or child.

requires a report in paragraphs (a) or (b) in this section.

(d) The licensee shall submit a written report to the appropriate NRC Regional Office listed in § 30.6 of this chapter no later than 15 days after discovery of a dose to the embryo/fetus or nursing child that requires a report in paragraphs (a) or (b) in this section.

(1) The written report must include --
(i) The licensee's name;
(ii) The name of the prescribing physician;
(iii) A brief description of the event;
(iv) Why the event occurred;
(v) The effect, if any, on the embryo/fetus or the nursing child; and
(vi) What actions, if any, have been taken or are planned to prevent recurrence;

(2) The report must not contain the individual's or child's name or any other information that could lead to identification of the individual or child.

(e) The licensee shall notify the referring physician and also notify the pregnant individual

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

(e) The licensee shall notify the referring physician and also notify the pregnant individual or mother, hereafter referred to as the mother, within 5 days of discovery of an event that would require reporting under paragraph (a) or (b) of this section, unless the referring physician personally informs the licensee either that he or she will inform the mother or that, based on medical judgment, telling the mother would be harmful;

(f) To meet the requirements of this section, the notification of the mother may be made instead to the mother's or child's responsible relative or guardian, when appropriate.

(g) The licensee is not required to notify the mother without first consulting the referring physician. If the referring physician or mother cannot be reached within 5 days, the licensee shall make the appropriate notifications as soon as possible thereafter. The licensee may not delay any appropriate medical care for the embryo/fetus or for the nursing child, including any necessary remedial care as a result of the

or mother, both hereafter referred to as the mother, no later than 24 hours after discovery of an event that would require reporting under paragraph (a) or (b) of this section, unless the referring physician personally informs the licensee either that he or she will inform the mother or that, based on medical judgement, telling the mother would be harmful.

(f) To meet the requirements of this section, the notification may be made to the mother's or child's responsible relative or guardian instead of the mother, when appropriate.

(g) The licensee is not required to notify the mother without first consulting with the referring physician. If the referring physician or mother cannot be reached within 24 hours, the licensee shall make the appropriate notifications as soon as possible thereafter. The licensee may not delay any appropriate medical care for the embryo/fetus or for the nursing child, including any necessary remedial care as a result of the event, because of any delay in notification.

(h) If notification was made under paragraphs (e) and

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

event, because of any delay in notification.

(h) If notification was made pursuant to paragraphs (e) and (f) of this section, the licensee shall also furnish, within 15 days after discovery of the event, a written report to the mother or responsible relative or guardian, by sending either--

(1) A copy of the report that was submitted to the NRC; or

(2) A brief description of both the event and the consequences as they may affect the embryo/fetus or nursing child.

(f) of this section, the licensee shall also furnish a written report to the mother or responsible relative or guardian within 15 days after discovery of the event.

A licensee may send either --

(1) A copy of the report that was submitted to the NRC; or

(2) A brief description of both the event and the consequences as they may affect the embryo/fetus or nursing child.

(i) A licensee shall retain a record of a dose to an embryo/fetus or a nursing child in accordance with § 35.2047.

Sec. 35.3067 Reports of leaking sources.

A licensee shall file a report within 5 days if a leakage test required by Sec. 35.67 reveals the presence of 185 Bq (0.005 μ Ci) or more of removable contamination. The report must be filed with the appropriate NRC Regional Office listed in Sec. 30.6 of this chapter, with a copy to the Director, Office of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001. The written report must include the model number and serial number if assigned, of the leaking source; the radionuclide and its estimated activity; the measured activity of each test sample expressed in microcuries; a description of the method used to measure each test sample; the date of the test; and the action taken.

§ 35.3067 Report of a leaking source.

A licensee shall file a report within 5 days if a leak test required by § 35.67 reveals the presence of 185 Bq (0.005 μ Ci) or more of removable contamination. The report must be filed with the appropriate NRC Regional Office listed in § 30.6 of this chapter, with a copy to the Director, Office of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001. The written report must include the model number and serial number if assigned, of the leaking source; the radionuclide and its estimated activity; the results of the test; the date of the test; and the action taken.

1. The commercial telephone number of the NRC Operations Center is (301) 951-0550.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Subpart K--Enforcement

Subpart N--Enforcement

Subpart N--Enforcement

Section 35.990 Violations.

Sec. 35.4001 Violations.

§ 35.4001 Violations.

(a) The Commission may obtain an injunction or other court order to prevent a violation of the provisions of--

(a) The Commission may obtain an injunction or other court order to prevent a violation of the provisions of--

(a) The Commission may obtain an injunction or other court order to prevent a violation of the provisions of --

(1) The Atomic Energy Act of 1954, as amended;

(1) The Atomic Energy Act of 1954, as amended;

(1) The Atomic Energy Act of 1954, as amended;

(2) Title II of the Energy Reorganization Act of 1974, as amended; or

(2) Title II of the Energy Reorganization Act of 1974, as amended; or

(2) Title II of the Energy Reorganization Act of 1974, as amended; or

(3) A regulation or order issued pursuant to those Acts.

(3) A regulation or order issued pursuant to those Acts.

(3) A regulation or order issued under those Acts.

(b) The Commission may obtain a court order for the payment of a civil penalty imposed under section 234 of the Atomic Energy Act:

(b) The Commission may obtain a court order for the payment of a civil penalty imposed under Section 234 of the Atomic Energy Act:

(b) The Commission may obtain a court order for the payment of a civil penalty imposed under Section 234 of the Atomic Energy Act:

(1) For violations of--

(1) For violations of--

(1) For violations of --

(i) Sections 53, 57, 62, 63, 81, 82, 101, 103, 104, 107, or 109 of the Atomic Energy Act of 1954, as amended;

(i) Sections 53, 57, 62, 63, 81, 82, 101, 103, 104, 107, or 109 of the Atomic Energy Act of 1954, as amended;

(i) Sections 53, 57, 62, 63, 81, 82, 101, 103, 104, 107, or 109 of the Atomic Energy Act of 1954, as amended;

(ii) Section 206 of the Energy Reorganization Act;

(ii) Section 206 of the Energy Reorganization Act;

(ii) Section 206 of the Energy Reorganization Act;

(iii) Any rule, regulation, or order issued pursuant to the sections specified in paragraph (b)(1)(i) of this section;

(iii) Any rule, regulation, or order issued pursuant to the sections specified in paragraph (b)(1)(i) of this section;

(iii) Any rule, regulation, or order issued under the sections specified in paragraph (b)(1)(i) of this section;

(iv) Any term, condition, or limitation of any license issued under the sections specified in paragraph (b)(1)(i) of this section.

(iv) Any term, condition, or limitation of any license issued

(iv) Any term, condition, or limitation of any license issued

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

(2) For any violation for which a license may be revoked under section 186 of the Atomic Energy Act of 1954, as amended.

PROPOSED RULE

under the sections specified in paragraph (b)(1)(i) of this section.
(2) For any violation for which a license may be revoked under Section 186 of the Atomic Energy Act of 1954, as amended.

FINAL RULE

under the sections specified in paragraph (b)(1)(i) of this section.
(2) For any violation for which a license may be revoked under Section 186 of the Atomic Energy Act of 1954, as amended.

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

Section 35.991 Criminal penalties.

(a) Section 223 of the Atomic Energy Act of 1954, as amended, provides for criminal sanctions for willful violation of, attempted violation of, or conspiracy to violate, any regulation issued under sections 161b, 161i, or 161o of the Act. For purposes of section 223, all the regulations in part 35 are issued under one or more of sections 161b, 161i, or 161o, except for the sections listed in paragraph (b) of this section.

(b) The regulations in part 35 that are not issued under subsections 161b, 161i, or 161o for the purposes of section 223 are as follows: Sections 35.1, 35.2, 35.7, 35.8, 35.12, 35.15, 35.18, 35.19, 35.57, 35.100, 35.600, 35.901, 35.970, 35.971, 35.990, 35.991, and 35.999.

Section 35.999 Resolution of conflicting requirements during transition period.

If the rules in this part conflict with the licensee's radiation safety program as identified in its license, and if that

Sec. 35.4002 Criminal penalties.

(a) Section 223 of the Atomic Energy Act of 1954, as amended, provides for criminal sanctions for willful violation of, attempted violation of, or conspiracy to violate, any regulation issued under sections 161b, 161i, or 161o of the Act. For purposes of Section 223, all the regulations in 10 CFR part 35 are issued under one or more of sections 161b, 161i, or 161o, except for the sections listed in paragraph (b) of this section.

(b) The regulations in 10 CFR part 35 that are not issued under subsections 161b, 161i, or 161o for the purposes of Section 223 are as follows: Secs. 35.1, 35.2, 35.7, 35.8, 35.12, 35.15, 35.18, 35.19, 35.65, 35.100, 35.200, 35.300, 35.600, 35.4001, and 35.4002.

§ 35.4002 Criminal penalties.

(a) Section 223 of the Atomic Energy Act of 1954, as amended, provides for criminal sanctions for willful violation of, attempted violation of, or conspiracy to violate, any regulation issued under sections 161b, 161i, or 161o of the Act. For purposes of Section 223, all the regulations in 10 CFR part 35 are issued under one or more of sections 161b, 161i, or 161o, except for the sections listed in paragraph (b) of this section.

(b) The regulations in 10 CFR part 35 that are not issued under subsections 161b, 161i, or 161o for the purposes of Section 223 are as follows: §§ 35.1, 35.2, 35.7, 35.8, 35.12, 35.15, 35.18, 35.19, 35.65, 35.100, 35.200, 35.300, 35.600, 35.4001, and 35.4002 .

10 CFR PART 35, MEDICAL USE OF BYPRODUCT MATERIAL-RULEMAKING MATRIX

CURRENT RULE

PROPOSED RULE

FINAL RULE

license was approved by the Commission before April 1, 1987 and has not been renewed since April 1, 1987, then the requirements in the license will apply. However, if that licensee exercises its privilege to make minor changes in its radiation safety procedures that are not potentially important to safety under Section 35.31 of this chapter, the portion changed must comply with the requirements of this part. At the time of license renewal and thereafter, these amendments to this part shall apply.

ATTACHMENT 9

Summary of NRC Rules with Compatibility or Health and Safety Designations Under the Revision of 10 CFR Part 35 (All other Sections are Category D)

Category A

20.1301 (a)(3) "Dose limits to individual members of the public"

Category B

32.72 (b)(1) & "Manufacture, preparation, or transfer for commercial distribution of radioactive material for medical use under Part 35" (b)(2)(ii) drugs containing byproduct
 32.74 (a) & (a)(3) "Manufacture and distribution of sources or devices"
 35.2 Definitions "Agreement State"
 35.2 Definitions "Sealed source"

Category C

35.2 Definition "Authorized medical physicist"
 35.2 Definition "Authorized nuclear pharmacist"
 35.2 Definition "Authorized user"
 35.2 Definition "Medical use"
 35.2 Definition "Prescribed dosage"
 35.2 Definition "Prescribed dose"
 35.2 Definition "Radiation safety officer"
 35.2 Definition "Treatment site"
 35.6 "Provisions for the protection of human research subjects"
 35.11 "License required"
 35.49 "Suppliers for sealed sources or devices for medical use"
 35.50 "Training for Radiation Safety Officer"
 35.51 "Training for an authorized medical physicist"
 35.55 "Training for an authorized nuclear pharmacist"
 35.57 "Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist"
 35.59 "Recentness of training"
 35.75 "Release of individuals containing unsealed byproduct material or implants containing byproduct material"
 35.190 "Training for uptake, dilution and excretion studies"
 35.290 "Training for imaging and localization studies"

- 35.390 "Training for use of unsealed byproduct material for which a written directive is required"
- 35.392 "Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries)"
- 35.394 "Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries)"
- 35.400 "Use of sealed sources for manual brachytherapy"
- 35.490 "Training for use of manual brachytherapy sources"
- 35.491 "Training for ophthalmic use of strontium-90"
- 35.500 "Use of sealed sources for diagnosis"
- 35.590 "Training for use of sealed sources for diagnosis"
- 35.600 "Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit"
- 35.690 "Training for use of remote afterloader units, teletherapy units, or gamma stereotactic radiosurgery units"
- 35.3045 "Report and notification of a medical event"
- 35.3047 "Report and notification of a dose to an embryo/fetus or a nursing child"
- 35.3067 "Report of a leaking source"

Category D (H&S)

- 35.24 (b) & (f) "Authority and responsibility for radiation protection program"
- 35.27 "Supervision"
- 35.40 "Written directives"
- 35.41 "Procedures for administrations requiring a written directive"
- 35.60 "Possession, use and calibration of instruments to measure activity of unsealed byproduct material"
- 35.61 "Calibration of survey instruments"
- 35.63 "Determination of dosages of unsealed byproduct material for medical use"
- 35.67 "Requirements for possession of sealed sources and brachytherapy sources"
- 35.69 "Labeling of vials and syringes"
- 35.70 "Surveys of ambient radiation exposure rate"
- 35.80 (a)(2), (a)(3), (a)(4) & (b) "Provision of mobile service"
- 35.92 "Decay-in-storage"
- 35.100 "Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required"
- 35.200 "Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required"
- 35.204 "Permissible molybdenum-99 concentration"
- 35.300 "Use of unsealed byproduct material for which a written directive is required"
- 35.310 "Safety instruction"
- 35.315 "Safety precautions"
- 35.404 "Surveys after source implant and removal"
- 35.406 "Brachytherapy sources accountability"
- 35.410 "Safety instruction"
- 35.415 "Safety precautions"

- 35.432 "Calibration measurement of brachytherapy sealed sources"
- 35.433 "Decay of strontium-90 sources for ophthalmic treatment"
- 35.457 "Therapy-related computer systems"
- 35.604 "Surveys of patients and human research subjects treated with a remote afterloader unit"
- 35.605 "Installation, maintenance, adjustment and repair"
- 35.610 "Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units "
- 35.615 "Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units"
- 35.630 "Dosimetry equipment"
- 35.632 "Full calibration measurements on teletherapy units"
- 35.633 "Full calibration measurements on remote afterloader units"
- 35.635 "Full calibration measurements on gamma stereotactic radiosurgery units"
- 35.642 "Periodic spot-checks for teletherapy units"
- 35.643 "Periodic spot-checks for remote afterloader units"
- 35.645 "Periodic spot-checks for gamma stereotactic radiosurgery units"
- 35.647 "Additional technical requirements for mobile remote afterloader units"
- 35.652 "Radiation surveys"
- 35.655 "Five-year inspection for teletherapy and gamma stereotactic radiosurgery units"
- 35.657 "Therapy-related computer systems"

COMPATIBILITY DESIGNATIONS FOR DRAFT FINAL 10 CFR PART 35 - MEDICAL USE OF BYPRODUCT MATERIAL

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.1	Purpose and scope	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2	Definitions		
	Address of use	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Agreement State	[B]	This definition also appears in 10 CFR §150.3(b). For purposes of compatibility, the language of the Part 150 definition should be used and it is assigned to Compatibility Category B.
	Area of use	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	*Authorized medical physicist	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(e), this definition was identified as a Category C because its absence could potentially impair effective communication. This definition impacts effective communication because it is used to delineate that the individual is legally vested with the authority to function as an authorized medical physicist. This definition is also needed because of its impact on the establishment of a nationwide training programs in the medical use area. The essential objective of this definition is to make sure that the medical physicist is a person identified on a license issued by the Commission or by an Agreement State, or meets the minimum requirements in 35.59 and 35.961.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	* Authorized nuclear pharmacist	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(e), this definition was identified as a Category C because its absence could potentially impair effective communication. This definition impacts effective communication because it is used to delineate that the individual is legally vested with the authority to function as an authorized nuclear pharmacist. This definition is also needed because of its impact on the establishment of a nationwide training programs in the medical use area. The essential objective of this definition is to make sure that the nuclear pharmacist is a person identified on a license issued by the Commission or by an Agreement State, or meets the minimum requirements in 35.59 and 35.980.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	* Authorized user	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(e), this definition was identified as a Category C because its absence could potentially impair effective communication. This definition impacts effective communication because it is used to delineate that the individual is legally vested with the authority to function as an authorized user. This definition is also needed because of its impact on the establishment of a nationwide training programs in the medical use area. The essential objective of this definition is to make sure that an authorized user is a person identified on a license issued by the Commission or by an Agreement State, or meets the minimum requirements in this part.
	Brachytherapy	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Brachytherapy source	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Client's address	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Dedicated check source	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	Dentist	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	High dose-rate remote afterloader	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Low dose-rate remote afterloader	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Management	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Manual Brachytherapy	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Medical event	D	Does not meet any of the criteria of Category A, B, C, or D/H&S since the term was not defined in this section, but is defined in 35.3045(a).
	Medical institution	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	* Medical use	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(e), this definition was identified as a Category C. The essential objectives of this definition should be adopted because the lack of it could potentially impair effective communication. The essential objective of this definition is to establish a common understanding regarding the application of radioactive materials and the radiation therefrom to humans as directed by an authorized user.
	Medium dose-rate remote afterloader	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Mobile medical service	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Output	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Patient intervention	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Pharmacist	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	Physician	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Podiatrist	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Preceptor	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	* Prescribed dosage	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs(2)(c) and (2)(e), this definition was identified as a Category C. The essential objectives of this definition should be adopted because the lack of it could potentially preclude an effective review or evaluation by the Commission and Agreement State programs for agreement materials with respect to protection of public health and safety and impair effective communication. The essential objective of this definition is to establish a common understanding regarding the correct quantity of radiopharmaceutical activity prescribed by the authorized user for administration to a patient.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	* Prescribed dose	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(c) and (2)(e), this definition was identified as a Category C. The essential objective of this definition should be adopted because the lack of it could potentially preclude an effective review or evaluation by the Commission and Agreement State programs for agreement materials with respect to protection of public health and safety and impair effective communication. The essential objective of this definition is to establish a common understanding regarding the correct quantity of radiopharmaceutical activity prescribed by the authorized user for administration to a patient.
	Pulsed dose-rate remote afterloader	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	* Radiation safety officer	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(e), this definition was identified as a Category C because its absence could potentially impair effective communication. This definition impacts effective communication because it is used to delineate that the individual is legally vested with the authority to function as radiation safety officer. This definition is also needed because of its impact on the establishment of a nationwide training programs in the medical use area. The essential objective of this definition is to make sure that the RSO is identified on a license issued by the Commission or by an Agreement State, or meets the minimum requirements in 35.59 and 35.900.
	Sealed source	[B]	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (B), paragraph (2)(c), this definition was identified as a Category B because it is a definition of a product that licensees routinely transport in multiple jurisdictions. This definition also appears in 10 CFR §30.4. For purposes of compatibility, the language of the Part 30 definition should be used and it is assigned to Compatibility Category B.
	Sealed source and device registry	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	Stereotactic radiosurgery	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Structured educational program	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Teletherapy	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Temporary jobsite	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Therapeutic dosage	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Therapeutic dose	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
	* Treatment site	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(c) and (2)(e), this definition was identified as a Category C. The essential objective of this definition should be adopted because the lack of it could potentially preclude an effective review or evaluation by the Commission and Agreement State programs for agreement materials with respect to protection of public health and safety and impair effective communication. The essential objective of this definition is to establish a common understanding of the correct anatomical description of the area intended to receive a radiation dose.
	Type of use	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Unit dosage	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
	Written directive	D	Does not meet any of the criteria of Category A, B, C, or D/H&S since the term was not defined in this section, but is defined in 35.40.
§35.5	Maintenance of records	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.6	Provisions for the protection of human research subjects	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (1), this requirement was designated a Category C. The lack of this requirement could create a gap whereby the Federal Policy for the Protection of Human Subjects would not be applied. The essential objective of this requirement is to assure the consistent application of the Federal Policy.
§35.7	FDA, other Federal, and State requirements	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.8	Information collection requirements: OMB Approval	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.10	Implementation	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.11	License required	[C]	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(f), this requirement was designated a Category C because Agreement States should adopt the Part 30 provision as a minimum requirement for their licensees. The general requirement for activities to be licensed appears in 10 CFR § 30.3 which has been designated compatibility category C. If an Agreement State has adopted 10 CFR § 30.3, it is not necessary to adopt this section since the requirements are covered in Part 30.3 and this section would be a duplication of those provisions.
§35.12	Application for license, amendment, or renewal	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.13	License amendments	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.14	Notifications	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.15	Exemptions regarding Type A specific licenses of broad scope	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.18	License issuance	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.19	Specific exemptions	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.24	Authority and responsibilities for the radiation protection program	<p>D - paragraphs (a) (1), (a) (2), (a)(3), (c), (d), (e) and (g)</p> <p>D/H&S- paragraph (b) and (f)</p>	<p>Paragraphs (a)(1), (a)(2), (a)(3), (d) and (g), do not meet any of the criteria of Category A, B, C, or D/H&S. In addition, paragraph (a)(2) was not required because the definitions and training of authorized users, authorized nuclear pharmacists, radiation safety officers and authorized medical physicists are required as a matter of compatibility and would prevent an unqualified individual from working in these positions. Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (E), paragraphs (b) and (f) were designated Category D/H&S because they assist in establishing a minimum level of safety for the medical use of agreement materials since they deal with the implementation of the radiation protection program. The essential objective of paragraph (b) is to assure that the RSO agrees to implement the radiation safety program. The essential objective of paragraph (f) is to require that management provides the RSO with sufficient authority, time and resources to identify radiation problems, initiate corrective actions, stop unsafe operations and verify the implementation of corrective actions.</p> <p>The H&S two or fewer failure test scenario: If unapproved changes are made to the radiation safety program, or if unsafe operations are not stopped, the public and workers could receive radiation exposures in excess of the limits in Parts</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.26	Radiation protection program changes	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.27	Supervision	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S because it assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a worker overexposure and medical event. The essential objectives of this requirement are to assure the instruction of the supervised individuals in radiation safety procedures and policies, and for the supervised individuals to request clarification from authorized users as needed.</p> <p>The H&S two or fewer failure test scenario: If a licensee fails to instruct supervised individuals in radiation safety procedures, regulations and license conditions and radioactive material is mishandled, the public and workers could receive radiation exposures in excess of limits and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.40	Written directives	D/H&S, (a) and (b) except paragraph (c) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S because it assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that correct information on the prescribed dose is communicated to the licensee's staff.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not use written directives for therapeutic medical use and a misinterpretation of the authorized users' orders occurs, a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.41	Procedures for administrations requiring a written directive	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S because it assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure patient identification and dose verification prior to administration to human beings.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not verify patient identity, radioactive material or radiation could be administered to the wrong person, and an exposure in excess of limits could occur. In addition, If the prescribed dose is not verified before administration, the prescribed dose could be exceeded and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.49	Suppliers for sealed sources or devices for medical use	[C]	<p>This section is a bracketed Category C, because it is already required by 10 CFR § 30.32 (g) and § 32.74. Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(f), 10 CFR § 30.32 (g) was designated a Category C. This section requires that licensees, authorized to possess and use sealed sources or devices for medical use, obtain these products from a licensed vendor. In addition, 10 CFR § 32.74, "Manufacture and distribution of sources or devices containing byproduct material for medical use," is designated a Category B. This designation was based upon Handbook 5.9 Part II, "Categorization Criteria," Section (B), paragraph (b), requirements for approval of products that are distributed nationwide. If an Agreement State has adopted 10 CFR § 30.32 (g) and 10 CFR § 32.74, it is not necessary to adopt this section since the requirements are covered by these provisions.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.50	Training for Radiation Safety Officer	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that only persons qualified by training and experience shall be authorized as a Radiation Safety Officer.
§35.51	Training for an authorized medical physicist	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that only persons qualified by training and experience shall be authorized as a medical physicist.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.55	Training for an authorized nuclear pharmacist	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that only persons qualified by training and experience shall be authorized as a nuclear pharmacist.
§35.57	Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user, and nuclear pharmacist	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to exempt experienced authorized users, including Radiation Safety Officers, who are named on a license from the training requirements in §35.50, §35.51 and Subparts C through H.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.59	Recentness of training	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that applicants for authorized user status are current for the proposed use of byproduct material since obtaining the required training and experience.
§35.60	Possession, use, and calibration of instruments to measure the activity of unsealed byproduct material	D/H&S, except paragraph (c) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure the measurement of the dosage with proper instrumentation prior to administration.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not properly measure the dosage with the appropriate instrumentation, the administered dose could differ from the prescribed dose and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.61	Calibration of survey instruments	D/H&S, except paragraphs (a)(3) and (d) are D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials. Without properly calibrated survey instruments over exposures could occur and is needed to demonstrate compliance with Part 20 requirements. The essential objective of this requirement is to assure the possession of calibrated survey instruments.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not calibrate or check survey instruments as required by this rule, and a contamination event occurs, radiation levels in excess of Part 20 limits could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.63	Determination of dosages of unsealed byproduct material for medical use	D/H&S, except paragraph (e) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S because it assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure proper determination of prescribed dosages by proper measurement and/or calculation prior to human use.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not measure a dosage, and a preparation error occurs, a medical event could occur.</p>
§35.65	Authorization for calibration, transmission and reference sources	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.67	Requirements for possession of sealed sources and brachytherapy sources	D/H&S, except paragraph (f) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure. The essential objective of this requirement is to assure the safe handling, periodic leak testing, and inventory of sealed sources.</p> <p>The H&S two or fewer failure test scenario: If the licensee does not follow the manufacturers' instructions, including testing for leakage, and a source is damaged or misplaced, public and worker exposures in excess of limits and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.69	Labeling of vials and syringes	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure. The essential objective of this requirement is to make sure that licensees develop, implement and maintain written procedures for labeling syringes, syringe shields, and vial shields that contain radiopharmaceuticals.</p> <p>The H&S two or fewer failure test scenario: If the syringe, syringe shield, or vial shield is not labeled, then the wrong radiopharmaceutical could be administered, and a medical event could occur. If syringe and vial shields are not used, then a worker could be overexposed.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.70	Surveys of ambient radiation exposure rate	D/H&S, except paragraphs (b) and (c) are D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this requirement was designated a D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and worker overexposure. The essential objective of this requirement is to assure that daily radiation surveys are performed in areas where therapeutic radiopharmaceuticals are used.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not conduct surveys for ambient radiation exposure rates, and an unplanned release of radioactive material occurs, contamination could go undetected and cause public and worker exposure in excess of limits in Part 20.</p>
§35.75	Release of individuals containing unsealed byproduct material or implants containing byproduct material	C, except paragraphs (c) and (d) are D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(a) and (f), this requirement was designated a Category C. This provision assists in establishing a minimum level of safety for the medical use of agreement materials on a nationwide basis by reducing the likelihood of public overexposures in multiple jurisdictions. The essential objective of this requirement is to assure that 0.5 rem TEDE is not exceeded by any individual, and that instructions are provided so that a breast-feeding infant/child does not receive an exposure exceeding 0.1 rem TEDE.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.80	Provision of mobile medical service	<p>D/H&S (a)(2),(a)(3), (a)(4) & (b) for those States which authorize this activity</p> <p>D for other States</p>	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated a D/H&S those Agreement States which authorize this service. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a public exposure. The essential objective of this requirement is to make sure that instruments are possessed, used, calibrated and checked as described in §35.60 and §35.62, that properly operating survey instruments are used and that radiation surveys are performed in areas where therapeutic radiopharmaceuticals are used.</p> <p>The H&S two or fewer failure test scenario: Paragraph (a)(2) and (a)(3) requires that the requirements in §35.60 and §35.62 be applied to mobile services and survey instruments are operating properly. If a mobile service licensee does not measure dosages with a proper operating instrumentation, then the prescribed dose could be exceeded, causing a medical event and the radiation protection limits in Part 20 for workers and the public could be exceeded. Paragraph (a)(4) requires that the mobile service survey all areas of use to assure compliance with Part 20 before leaving the client's address of use. If an exit survey is not conducted and byproduct material is left at the client's address of use, radiation protection limits in Part 20 for</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.92	Decay-in-storage	D/H&S - for those States which authorize this activity D - for other States	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) and (b) are designated as D/H&S for those Agreement States which authorize this activity. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a public overexposure. The essential objective of this requirement is to allow decay in storage for byproduct materials with less than 120 day half-life and to assure that surveys will be performed on the material before disposal.</p> <p>The H&S two or fewer failure test scenario: If byproduct material is allowed to decay-in-storage and is improperly surveyed, it may be disposed of as normal trash. Byproduct material could then be released into the public domain and overexpose workers and members of the public.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.100	Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this provision was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that radiopharmaceuticals are obtained from a licensed vendor or authorized preparer.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not obtain radiopharmaceuticals from a licensed manufacturer or authorized preparer, a preparation error could occur, and a medical event could occur.</p>
§35.190	Training for uptake, dilution and excretion studies	C	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), these requirements were designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians use the radioactive material in this section.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.200	Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, this section was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that radiopharmaceuticals are obtained from a licensed vendor or authorized preparer.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not obtain radiopharmaceuticals from a licensed manufacturer or authorized preparer, and a preparation error occurs, a patient could receive a radiation exposure and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.204	Permissible molybdenum-99 concentration	D/H&S, except c is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) and (b) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that 5.55kBq is not exceeded.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform a Mo-99 measurement and the 5.55kBq limit is exceeded, and this contaminant is administered, a medical event could occur.</p>
§35.290	Training for imaging and localization studies	C	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), these requirements were designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians use the radioactive material in this section.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.300	Use of unsealed byproduct material for which a written directive is required	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated a D/H&S for those Agreement States which authorize this activity. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that radiopharmaceuticals are obtained from a licensed vendor or authorized preparer.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not obtain radiopharmaceuticals from a licensed manufacturer or authorized preparer, a preparation error or medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.310	Safety instruction	D/H&S, except (b) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated a D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of exposure to members of the public. The essential objective of this requirement is to assure that personnel caring for patients receive initial and annual radiation safety instruction.</p> <p>The H&S two or fewer failure test scenario: If the personnel caring for patients are not properly instructed, then personnel could be overexposed and persons visiting the patient could be overexposed. In addition, contaminated material could be released into the public domain, and a public overexposure could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.315	Safety precautions	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated a D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of worker and public exposure. The essential objective of this requirement is to limit unnecessary patient contact with members of the public, and to assure that proper notifications are made if the patient dies.</p> <p>The H&S two or fewer failure test scenario: If the safety precautions are not taken, then hospital personnel could be overexposed and persons visiting the patient could be overexposed. Contaminated material could also be released into the public domain, and public overexposures could occur. In addition, if the proper persons are not notified if a patient dies or, if a medical emergency occurs, other personnel could be overexposed.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.390	Training for use of unsealed byproduct material for which a written directive is required	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians use the radioactive material in this section.
§35.392	Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries)	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians use the radioactive material in this section.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.394	Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries)	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians use the radioactive material in this section.
§35.400	Use of sealed sources for manual brachytherapy	[C]	This section is a bracketed Category C, because it is already required by 10 CFR § 30.32 (g) and § 32.74. Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(f), 10 CFR § 30.32 (g) was designated a Category C. This section requires the use of sealed sources in the SS&D Registry by all specific licensees. In addition, 10 CFR § 32.74, "Manufacture and distribution of sources or devices containing byproduct material for medical use," is designated a Category B. This designation was based upon Handbook 5.9 Part II, "Categorization Criteria," Section (B), paragraph (b), requirements for approval of products that are distributed nationwide. If an Agreement State has adopted 10 CFR § 30.32 (g) and 10 CFR § 32.74, it is not necessary to adopt this section since the requirements are covered by these provisions.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.404	Surveys after source implant and removal	D/H&S, except (c) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) and (b) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event, and overexposures in excess of Part 20 limits. The essential objective of this requirement is to assure that patient and area surveys are performed immediately after removal of sources from the patient or human research subject.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform a patient radiation survey after implanting sources, sources could be misplaced and the public and workers could be exposed to radiation in excess of basic radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.406	Brachytherapy sources accountability	D/H&S, except (c) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) and (b) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a lost source, medical event, public and worker overexposure. The essential objective of this requirement is to assure the accountability of sources after use.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not maintain source accountability a source may be misplaced and the public and workers could be overexposed and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.410	Safety instruction	D/H&S, except (b) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated a D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of exposure to members of the public. The essential objective of this requirement is to assure that personnel caring for a patient receive proper radiation safety instruction.</p> <p>The H&S two or fewer failure test scenario: If the personnel caring for patients are not properly instructed, then the personnel could be overexposed and persons visiting the patient could be overexposed.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.415	Safety precautions	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, these provisions were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of worker and public exposure. The essential objective of these requirements is to limit patient contact with members of the public, to assure the availability of emergency response equipment and to assure that proper notifications are made if the patient dies.</p> <p>The H&S two or fewer failure test scenario: If the safety precautions are not taken, then hospital personnel could be overexposed and persons visiting the patient could be overexposed. In addition, if the proper persons are not notified if a patient dies or, if a medical emergency occurs, other hospital personnel could be overexposed and material could be released into the public domain, and public overexposures could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.432	Calibration measurements of brachytherapy sealed sources	D/H&S, except (d) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, these provisions were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that full calibration measurements on sources are performed before the first use of the source or source/applicator configuration.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform full calibration measurements of brachytherapy sources, a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.433	Decay of strontium-90 sources for ophthalmic treatments	D/H&S (a), except (b) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, these provisions were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that the source is properly decayed to identify the correct source strength.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not properly decay the source and it's strength is not accurately determined, a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.457	Therapy-related computer systems	D/H&S	<p>Based upon Handbook 5.9 Part paragraphs (a) and (b) was designated II, "Categorization Criteria," Section E, this provision was designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood a medical event. The essential objective of this requirement is to assure that therapy related computer systems are functioning properly and testing is done in accordance with national protocols.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not verify that computerized treatment planning systems are operating properly, and a medical event occurs.</p>
§35.490	Training for use of manual brachytherapy sources	C	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(b) and (2)(f), these requirements were designated a Category C by relieving an undue burden on interstate commerce by establishing a minimum level of training for the medical use of agreement materials. The essential objective of this requirement is to assure that properly trained physicians handle sealed sources for human use in this section.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.491	Training for ophthalmic use of strontium-90	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(b) and (2)(f), these requirements were designated a Category C by relieving an undue burden on interstate commerce by establishing a minimum level of training for the medical use of agreement materials. The essential objective of this requirement is to assure that properly trained physicians handle sealed sources for human use in this section.
§35.500	Use of sealed sources for diagnosis	[C]	This section is a bracketed Category C, because it is already required by 10 CFR § 30.32 (g) and § 32.74. Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(f), 10 CFR § 30.32 (g) was designated a Category C. This section requires the use of sealed sources in the SS&D Registry by all specific licensees. In addition, 10 CFR § 32.74, "Manufacture and distribution of sources or devices containing byproduct material for medical use," is designated a Category B. This designation was based upon Handbook 5.9 Part II, "Categorization Criteria," Section (B), paragraph (b), requirements for approval of products that are distributed nationwide. If an Agreement State has adopted 10 CFR § 30.32 (g) and 10 CFR § 32.74, it is not necessary to adopt this section since the requirements are covered by these provisions.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.590	Training for use of sealed sources for diagnosis	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians handle sealed sources for human use in this section.
§35.600	Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit	[C]	This section is a bracketed Category C, because it is already required by 10 CFR § 30.32 (g) and § 32.74. Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraph (2)(f), 10 CFR § 30.32 (g) was designated a Category C. This section requires the use of sealed sources in the SS&D Registry by all specific licensees. In addition, 10 CFR § 32.74, "Manufacture and distribution of sources or devices containing byproduct material for medical use," is designated a Category B. This designation was based upon Handbook 5.9 Part II, "Categorization Criteria," Section (B), paragraph (b), requirements for approval of products that are distributed nationwide. If an Agreement State has adopted 10 CFR § 30.32 (g) and 10 CFR § 32.74, it is not necessary to adopt this section since the requirements are covered by these provisions.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.604	Surveys of patients and human research subjects treated with a remote afterloader unit	D/H&S, except (b) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated D/H&S. This provision assists in establishing a minimum level of safety for the medical use of agreement materials on a nationwide basis by reducing the likelihood of a lost source, medical event, and public/worker overexposure. The essential objective of this requirement is to assure that a survey is conducted on the patient and the device after source(s) removal from the patient or human research subject.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform a patient radiation survey after implanting sources, sources could be misplaced and the public and workers could be exposed to radiation in excess of radiation protection limits in Parts 20 and 35 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.605	Installation, maintenance, adjustment and repair	D/H&S, except (d) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, these provisions were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event and public and worker overexposure. The essential objective of these requirements is to assure that installation, maintenance, and adjustments are performed by a person specifically licensed by the Commission or an Agreement State.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not use a person who is specifically licensed to install and service devices and an equipment failure occurs, the person servicing the device could become overexposed. In addition, the public and workers could receive exposures in excess of the radiation protection limits in Parts 20, if the device is not serviced properly.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.610	Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units	D/H&S except (f) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, these provisions were designated D/H&S. These provisions assist in establishing a minimum level of safety for agreement materials by reducing the likelihood of a medical event and public and worker overexposure. The essential objective of these requirements is to assure the establishment and use of safety procedures and instructions for remote afterloaders, teletherapy units and gamma stereotactic radiosurgery units.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not develop and implement safety procedures and instructions, and radioactive material is mishandled, workers and the public could receive radiation exposures in excess of the radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.615	Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units	D/H&S	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, these provisions were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event, worker and public exposure. The essential objective of these requirements is to assure that controls are implemented, to require the physical presence of the authorized user and/or authorized medical physicist during the use of the device, and to assure the accessibility of emergency equipment.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not control access to therapy equipment and treatment rooms, and an equipment failure occurs, the public and workers could receive exposures in excess of the radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.630	Dosimetry equipment	D/H&S except (c) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) and (b) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure the use of a dosimetry system that has been calibrated in accordance with national standards.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not calibrate and check dosimetry equipment in accordance with industry standards, and an equipment failure occurs, the public and workers could be exposed to radiation in excess of radiation limits in Part 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.632	Full calibration measurements on teletherapy units	D/H&S, except (g) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) through (f) were designated D/H&S. These provisions assist in establishing a minimum level of safety basis by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that the medical device is performing properly.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform full calibration measurements on teletherapy units in accordance with industry standards, and an equipment failure occurs, the public and workers could be exposed to radiation in excess of radiation limits in Part 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.633	Full calibration measurements on remote afterloader units	D/H&S, except (h) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) through (f) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that the medical device is performing properly.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform full calibration measurements on remote afterloaders in accordance with industry standards, and an equipment failure occurs, the public and workers could be exposed to radiation in excess of radiation limits in Part 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.635	Full calibration measurements on gamma stereotactic radiosurgery units	D/H&S, except (g) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) through (f) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that the medical device is performing properly.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform full calibration measurements on gamma stereotactic radiosurgery units in accordance with industry standards, and an equipment failure occurs, the public and workers could be exposed to radiation in excess of radiation limits in Part 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.642	Periodic spot-checks for teletherapy units	D/H&S, except (f) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) through (e) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement material by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that the medical device is performing properly.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform periodic spot checks of teletherapy units, and an equipment failure occurs, the public and workers could receive exposures in excess of radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.643	Periodic spot-checks for remote afterloader units	D/H&S, except (e) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) through (g) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event. The essential objective of this requirement is to assure that the medical device is performing properly.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform periodic spot checks of high dose-rate and pulsed dose-rate remote afterloaders, and an equipment failure occurs, the public and workers could receive exposures in excess of radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.645	Periodic spot-checks for gamma stereotactic radiosurgery units	D/H&S, except (g) is D	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) through (f) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a medical event by checking instrument performance between maintenance and full calibrations. The essential objective of this requirement is to assure that the medical device is performing properly.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform periodic spot checks of gamma stereotactic radiosurgery units, and an equipment failure occurs, then the public and workers could receive exposures in excess of radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.647	Additional technical requirements for mobile remote afterloader units	<p>D/H&S (a) through (d) for those States which authorize this activity.</p> <p>D - for other States.</p>	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraph (a) was designated a D/H&S those Agreement States which authorize this service. This provision assists in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of a public exposure. The essential objective of this requirement is to assure that the appropriate radiation surveys are performed and that the proper administrative controls are utilized.</p> <p>The H&S two or fewer failure test scenario: If these requirements are not adopted, a mobile remote afterloader licensee would not be required to check survey instruments and verify sources after use, then sources could be misplaced and the public and workers could be exposed to radiation in excess of radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.652	Radiation surveys	D/H&S, except (c) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) and (b) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood of public and worker overexposures. The essential objective of this requirement is to assure that exposure to the source(s) in the shielded position does not exceed the levels stated in the SS&D registry.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not perform these radiation surveys, and the radiation safety of the device is compromised, then the public and workers could receive exposures in excess of radiation protection limits in Parts 20.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.655	Five-year inspection for teletherapy and gamma stereotactic radiosurgery units	D/H&S, except (c) is D.	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section E, paragraphs (a) and (b) were designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood a medical event, and public and worker exposure. The essential objective of this requirement is to assure that the inspection and servicing of teletherapy and gamma stereotactic radiosurgery units during source replacements or at intervals not to exceed 5 years and this inspection and servicing is performed by a person specifically licensed by the Commission or an Agreement State.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not have their teletherapy and gamma stereotactic radiosurgery units inspected and serviced at 5-year intervals and an equipment failure occurs, the public and workers could receive radiation exposures in excess of radiation protection limits in Parts 20 and a medical event could occur.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.657	Therapy-related computer systems	D/H&S	<p>Based upon Handbook 5.9 Part paragraphs (a) and (b) was designated II, "Categorization Criteria," Section E, this provision was designated D/H&S. These provisions assist in establishing a minimum level of safety for the medical use of agreement materials by reducing the likelihood a medical event. The essential objective of this requirement is to assure that therapy related computer systems are functioning properly and testing is done in accordance with national protocols.</p> <p>The H&S two or fewer failure test scenario: If a licensee does not verify that computerized treatment planning systems are operating properly, and a medical event occurs.</p>
§35.690	Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units	C	<p>Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2) (b) and (2)(f), this requirement was designated a Category C because they relieve undue burdens on interstate commerce by establishing a minimum level of training for the medical use of agreement materials on a nationwide basis. The essential objective of this requirement is to assure that properly trained physicians use therapeutic medical devices.</p>

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.1000	Other medical uses of byproduct material or radiation from byproduct material	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2024	Records of authority and responsibilities for radiation protection programs	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2026	Records of radiation protection program changes	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2040	Records of written directives	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2045	Records of medical events	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2047	Record of a dose to an embryo/fetus or a nursing child	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2060	Records of calibrations of instruments to measure the activity of unsealed byproduct materials	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.2061	Records of radiation survey instrument calibrations	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2063	Records of dosage of unsealed byproduct material for medical use	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2067	Records of possession of sealed sources and brachytherapy sources	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2070	Records of surveys for ambient radiation exposure rate	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2075	Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2080	Records of administrative and technical requirements that apply to the provision of mobile medical services	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.2092	Records of decay-in-storage	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2204	Records of molybdenum-99 concentrations	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2310	Records of safety instruction	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2404	Records of surveys after source implant and removal	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2406	Records of brachytherapy source accountability	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2432	Records of calibration measurements of brachytherapy sealed sources	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2433	Records of decay of strontium-90 sources for ophthalmic treatments	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.2605	Records of installation, maintenance, adjustment and repair	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2630	Records of dosimetry equipment	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2632	Records of teletherapy, remote afterloader, and gamma stereotactic radiosurgery full calibrations	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2642	Records of periodic spot-checks for teletherapy units	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2643	Records of periodic spot-checks for remote afterloader units	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2645	Records of periodic spot-checks for gamma stereotactic radiosurgery units	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.2647	Records of additional technical requirements for mobile remote afterloader units	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2652	Records of surveys of therapeutic treatment units	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.2655	Records of 5-year inspection of teletherapy and gamma stereotactic radiosurgery units	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.3045	Report and notification of a medical event	C, except paragraphs (b) and (f) are D.	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(c), this requirement was designated a Category C because absent this provision there could be a potential preclusion of an effective review or evaluation by the Commission and Agreement State programs for agreement material with respect to protection of public health and safety. The essential objective of this requirement is to assure that medical events are reported to the radiation control program and in order to assess the effectiveness of the national program for control of Atomic Energy Act materials.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§35.3047	Report and notification of a dose to an embryo/fetus or a nursing child	C, except paragraph (c) is D.	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(c), this requirement was designated a Category C because absent this provision there could be a potential preclusion of an effective review or evaluation by the Commission and Agreement State programs for agreement material with respect to protection of public health and safety. The essential objective of this requirement is to assure that doses to the embryo/fetus or nursing child are reported to the radiation control program.
§35.3067	Report of a leaking source	C	Based upon Handbook 5.9 Part II, "Categorization Criteria," Section (C), paragraphs (2)(c), this requirement was designated a Category C because absent this provision there could be a potential preclusion of an effective review or evaluation by the Commission and Agreement State programs for agreement material with respect to protection of public health and safety. The essential objective of this requirement is to assure that leaking sources are reported to the radiation control program.
§35.4001	Violations	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.
§35.4002	Criminal penalties	D	Does not meet any of the criteria of Category A, B, C, or D/H&S.

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	COMPATIBILITY CATEGORY	COMMENTS
§20.1301 (a)(3)	Dose limits for individual members of the public	A	
§32.72 (b)(1) and (b)(2)(ii)	Manufacture, preparation, or transfer for commercial distribution of radioactive drugs containing byproduct material for medical use under Part 35	B	
§32.74 (a) and (a)(3)	Manufacture and distribution of sources or devices containing byproduct material for medical use	B	

G:\LAB\CHART3.WPD

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

July 7, 1999

CURRENT RULE vs. PROPOSED RULE vs. DRAFT FINAL RULE

CHANGES TO COMPATIBILITY DESIGNATIONS 10 CFR PART 35 - MEDICAL USE OF BYPRODUCT MATERIAL

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.2	Definitions			
	ALARA	[A]	Deleted	Deleted
	Agreement State	[B]	[B]	[B]
	Authorized medical physicist	---	C	C
	Authorized nuclear pharmacist	D	C	C
	Authorized user	C	C	C
	Medical use	C	C	C
	Misadministration	C	Deleted	Deleted
	Prescribed dosage	C	C	C
	Prescribed dose	C	C	C
	Radiation safety officer	D	C	C

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
	Sealed source	[B]	[B]	[B]
	Treatment site	---	C	C
	Written directive	C	D	D
§35.6	Provisions for the protection of human research subjects.	D	C	C
§35.11	License required	[C]	[C]	[C]
§35.20	ALARA program	[D]	Part 20	Part 20
§35.21	Radiation Safety Officer	D/H&S (a) only	Deleted	Deleted
§35.23	Statements of authority and responsibility	D/H&S (a) only	Deleted	Deleted
§35.24	Authority and responsibilities for the radiation protection program	D except (a) is D/H&S	D/H&S (c) & (e) only	D/H&S (b) & (f) only
§35.27	Supervision	D	D/H&S	D/H&S
§35.32	Quality Management program	D/H&S (a)(b) & (c)	Deleted	Deleted

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.33	Notifications, reports, and records of misadministrations	C	Deleted	Deleted
§35.40	Written Directives	---	D/H&S, except (c) is D	D/H&S, (a) & (b) except (c) is D
§35.41	Procedures for administrations requiring a written directive	---	D/H&S	D/H&S
§35.49	Suppliers for sealed sources or devices for medical use	D	[C]	[C]
§35.50	Training for Radiation Safety Officer	D	C	C
§35.51	Training for authorized medical physicist	---	C	C
§35.55	Training for an authorized nuclear pharmacist	D	C	C
§35.57	Training for experienced Radiation Safety Officer, teletherapy or medical physicist, authorized user and nuclear pharmacist	D	C	C

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.59	Recentness of training	D	C	C
§35.60	Possession, use, calibration of instruments to measure the activity of unsealed byproduct material	D	D/H&S except (c) is D	D/H&S except (c) is D
§35.61	Calibration of survey instruments	[D]	D/H&S except (d) is D	D/H&S except (a)(3) & (d) are D
§35.62	Possession, use, calibration, and check of instruments to measure dosages of alpha- or beta-emitting radionuclides	D	D/H&S except (d) is D	Deleted
§35.63	Determination of dosages of unsealed byproduct material for medical use	D/H&S (a) & (b), except (c) is D	D/H&S except (e) is D	D/H&S except (e) is D
§35.67	Requirements for possession of sealed sources & brachytherapy sources	D/H&S	D/H&S except (e) is D	D/H&S except (f) is D
§35.69	Labeling of vials and syringes	D	D/H&S	D/H&S

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.70	Surveys of ambient radiation exposure rate	D/H&S	D/H&S except (c) is D	D/H&S except (b) & (c) are D
§35.75	Release of individuals containing unsealed byproduct material or implants containing byproduct material	C - (a) D/H&S - (b) D - (c) & (d)	C, except (c) & (d) are D	C, except (c) & (d) are D
§35.80	Provision of mobile medical service	D	D except (a) is D/H&S for States that authorize mobile svcs	D - except (a)(2), (a)(3), (a)(4) & (b) are D/H&S for States which authorize mobile services
§35.92	Decay-in-storage	D	D/H&S for those States authorizing & D for all others	D/H&S for States which authorize and D for all others
§35.100	Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required	D/H&S	D/H&S	D/H&S
§35.190	Training for uptake, dilution, and excretion studies	D	C	C

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.200	Use of unsealed byproduct material for imaging and localization for which a written directive is not required	D/H&S	D/H&S	D/H&S
§35.204	Permissible molybdenum-99 concentration	D/H&S	D	D except (c) is D
§35.290	Training for imaging and localization studies	D	C	C
§35.300	Use of unsealed byproduct material for which a written directive is required	D/H&S	D/H&S	D/H&S
§35.310	Safety instruction	D	D/H&S except (b) is D	D/H&S except (b) is D
§35.315	Safety precautions	D	D/H&S	D/H&S
§35.390	Training for use of unsealed byproduct material for which a written directive is required	D	C	C

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.392	Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries)	D	---	C
§35.394	Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries)	D	C	C
§35.400	Use of sources for manual brachytherapy	D	[C]	[C]
§35.404	Surveys after source implant and removal	C	D/H&S except (c) is D	D/H&S except (c) is D
§35.406	Brachytherapy sources accountability	D - except (a) & (c) are D/H&S	D/H&S except (c) is D	D/H&S except (c) is D
§35.410	Safety instruction	D	D/H&S except (b) is D	D/H&S except (b) is D

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.415	Safety precautions	D	D/H&S	D/H&S
§35.432	Calibration measurements of brachytherapy sealed sources	---	D/H&S except (g) is D	D/H&S except (d) is D
§35.433	Decay of strontium-90 sources for ophthalmic treatments	---	---	D/H&S (a), except (b) is D
§35.457	Therapy-related computer systems	---	---	D/H&S
§35.490	Training for use of manual brachytherapy sources	D	C	C
§35.491	Training for ophthalmic use of strontium-90	D	C	C
§35.500	Use of sealed sources for diagnosis	D	[C]	[C]
§35.590	Training for use of sealed sources for diagnosis	D	C	C

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.600	Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit	D	[C]	[C]
§35.604	Surveys of patients and human research subjects treated with a remote afterloader unit	---	D/H&S except (b) is D	D/H&S except (b) is D
§35.605	Installation, maintenance, adjustment and repair	D/H&S	D/H&S except (d) is D	D/H&S except (d) is D
§35.610	Safety procedures and instructions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units	D/H&S	D/H&S except (e) is D	D/H&S except (f) is D
§35.615	Safety precautions for remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units	D/H&S	D/H&S	D/H&S
§35.630	Dosimetry equipment	D/H&S	D/H&S except (c) is D	D/H&S except (c) is D

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.632	Full calibration measurements on teletherapy units	D/H&S	D/H&S except (g) is D	D/H&S except (g) is D
§35.633	Full calibration measurements on remote afterloader units	---	D/H&S except (h) is D	D/H&S except (h) is D
§35.635	Full calibration measurements on gamma stereotactic radiosurgery units	---	D/H&S except (g) is D	D/H&S except (g) is D
§35.642	Periodic spot-check for teletherapy units	D/H&S	D/H&S except (f) is D	D/H&S except (f) is D
§35.643	Periodic spot-check for remote afterloader units	---	D/H&S except (h) is D	D/H&S except (e) is D
§35.645	Periodic spot-checks for gamma stereotactic radiosurgery units	---	D/H&S except (g) is D	D/H&S except (g) is D

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.647	Additional technical requirements for mobile remote afterloaders	---	D except (a-d) are D/H&S for State that authorize	D except (a-d) are D(H/S) for those States which authorize this activity.
§35.652	Radiation surveys	D	D/H&S except (c) is D	D/H&S except (c) is D
§35.655	Five-year inspection for teletherapy & gamma stereotactic radiosurgery units	D/H&S	D/H&S except (c) is D	D/H&S except (c) is D
§35.657	Therapy-related computer systems	---	D/H&S	D/H&S
§35.690	Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units	D	C	C
§35.3045	Reports of medical events	---	C except (b) & (f) are D	C except (b) & (f) are D
§35.3047	Reports of a dose to an embryo/fetus or a nursing child	Part 20	C except (c) is D	C except (c) is D

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

REGULATION SECTION	SECTION TITLE	CURRENT RULE 9/98	PROPOSED RULE 8/98	DRAFT FINAL RULE 5/4/99
§35.3067	Reports of leaking sources	Part 30	C	C
§20.1301 (a)(3)	Dose limits for individual members of the public	A	A	A
§32.72 (b)(1) & (b)(2)(ii)	Manufacture, preparation, or transfer for commercial distribution of radioactive drugs containing byproduct material for medical use under Part 35	B	B	B
§32.74 (a) & (a)(3)	Manufacture & distributor of sources or devices containing byproduct material for medical use	B	B	B

G:\LAB\Delta3.wpd

* This definition provides information that is essential to the common understanding beyond the plain dictionary meaning.

ATTACHMENT 10

SAMPLE LETTER TO BE SENT TO CERTIFICATION BOARDS

American Board of XXX

Dear Dr. _____:

As you know, the Nuclear Regulatory Commission (NRC) is revising its medical use regulations in 10 CFR Part 35, "Medical Use of Byproduct Material." I anticipate the Commission will publish the final rule in the Federal Register in early 2000, with an effective date 6 months after publication. As part of this revision, the regulatory text will no longer incorporate a listing of the specific boards whose diplomates automatically fulfill the training and experience requirements for an authorized medical physicist, authorized nuclear pharmacist, authorized user, or Radiation Safety Officer. Rather, the NRC will recognize certification boards that require individuals to complete the training and experience requirements specified in the regulatory text. Once recognized, the board's name will be placed on the list of recognized boards maintained on the NRC website. This change is being made to eliminate the need for a rulemaking each time a board would be added or deleted.

I am writing to notify you of our intent to initiate the recognition process immediately. Other specialty boards whose diplomates are likely to seek authorization are being similarly notified. If you are interested in having your board recognized by the NRC, please submit a letter to me listing each training and experience section of the rule for which you believe your Board's diplomates should be deemed to have met the requirements. Enclosures 1 and 2 should assist you in preparing your letter. Enclosure 1 lists all requirements where NRC plans to recognize boards. Enclosure 2 is a copy of the draft final regulatory text that lists the training and experience criteria for authorized medical physicists, authorized nuclear pharmacists, authorized users, and Radiation Safety Officers.

Your letter should clearly state that an individual must have completed the training and experience required by a particular section prior to receiving board certification. For example, if your board would like to be recognized under 10 CFR 35.390, "Training for use of unsealed byproduct material for which a written directive is required," the letter should state: The American Board of XXX has reviewed 10 CFR 35.390 and has determined that our certification process requires an individual to meet all the requirements in paragraph (b) of this section prior to being certified by our board."

The letter should be dated and signed by the chief executive of your board. If you have any questions or comments, please contact xxx (301-415-xxxx or E-mail at xxx@nrc.gov) .

Sincerely,

Donald Cool, Ph.D., Director
Division of Industrial and
Medical Nuclear Safety

Enclosures:

1. Areas where NRC plans to recognize boards
2. Applicable training and experience requirements (to be added prior to mailing)

REQUIREMENTS WHERE NRC PLANS TO RECOGNIZE BOARDS

PART 35--MEDICAL USE OF BYPRODUCT MATERIAL

- 35.50 Training for Radiation Safety Officer.
- 35.51 Training for an authorized medical physicist.
- 35.55 Training for an authorized nuclear pharmacist.
- 35.190 Training for uptake, dilution, and excretion studies.
- 35.290 Training for imaging and localization studies.
- 35.390 Training for use of unsealed byproduct material for which a written directive is required.
- 35.392 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 Gigabecquerels (33 millicuries).
- 35.394 Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities greater than 1.22 Gigabecquerels (33 millicuries).
- 35.490 Training for use of manual brachytherapy sources.
- 35.491 Training for ophthalmic use of strontium-90.
- 35.590 Training for use of sealed sources for diagnosis.
- 35.690 Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Enclosure 1