

J McKnight

UNITED STATES
NUCLEAR REGULATORY COMMISSION
OFFICE OF NUCLEAR MATERIAL SAFETY AND SAFEGUARDS
WASHINGTON, D.C. 20555

December 21, 1999

NRC INFORMATION NOTICE 99-33: MANAGEMENT OF WASTES CONTAMINATED WITH
RADIOACTIVE MATERIALS

Addressees: All medical licensees.

Purpose:

The U.S. Nuclear Regulatory Commission (NRC) is issuing this information notice to remind medical use licensees to carefully monitor all waste that may be contaminated with radioactive materials. Waste management facilities not authorized to receive licensable materials are finding wastes contaminated with detectable levels of radioactive materials in waste shipments from hospitals and other medical facilities. It is expected that recipients will review this information for applicability to their facilities and consider actions, as appropriate, to address these issues. However, suggestions contained in this information notice are not new NRC requirements; therefore no specific action nor written response is required.

Description of Circumstances:

In January 1991, NRC issued Information Notice 91-03, entitled "Management of Wastes Contaminated with Radioactive Materials ("Red Bag" Waste and Ordinary Trash)". Since the 1991 information notice, many operators of medical waste incinerators and landfills have been monitoring waste shipments for radioactivity. In several cases, waste shipments from hospitals have contained radioactive materials, with radiation levels that exceeded the waste disposal operators' preset detection levels. In some cases, operators of landfills or medical waste incinerators have rejected the shipments and returned them to the generators. In general, operators of landfills and medical waste incinerators are not authorized to receive or manage radioactive waste.

Case 1: Incident involving a medical waste incinerator

A representative from the Connecticut Department of Environmental Protection (CTDEP) notified the NRC Region I staff that radioactive material had been detected in a dumpster at an incinerator facility. The personnel at the incinerator notified CTDEP that waste in the dumpster, measuring 20 milliroentgen per hour, had tripped their radiation monitors. CTDEP personnel determined that the radioactive waste was localized in a small area within the dumpster and identified the radioactive material as iodine-131. CTDEP personnel tracked the source of the waste to an NRC medical licensee. Because all sources were accounted for during the previous week, the licensee believed that the material was patient generated and not therapeutic. After being contacted by NRC Region I, the licensee segregated the waste from the dumpster and brought the radioactive material back to its facility.

ML993510053

DF03

Case 2: Incident involving a landfill

Browning Ferris Industries (BFI) received a load of solid waste from a hospital in New York City and took it to a transfer station in Brooklyn, NY. The waste was then moved to a landfill in Pennsylvania, where it set off a radiation portal monitor alarm. The waste was rejected by the landfill and returned to the BFI transfer station in Brooklyn. Personnel determined that the waste contained approximately 11.1 gigabecquerel (300 mCi) of technetium-99m and returned it to the hospital.

Case 3: Incident involving housekeeping

The University of Pittsburgh reported the loss of a brachytherapy ribbon containing 0.15 gigabecquerel (4 mCi) of iridium-192 seeds. A Patient Service Technician removed soiled linen from the room of a brachytherapy patient and placed them in the hallway. The linen contained the ribbon with the iridium-192 seeds, which had become displaced from the patient. The technician stated that she was not familiar with the standard precautions of keeping trash and linen in the room until they have been surveyed. For corrective actions, the licensee retrained the nursing staff who have primary care responsibilities for patients receiving brachytherapy in the applicable radiation safety policies, procedures, and precautions. In addition, a written test was administered to each individual to evaluate and assure comprehension of the training information.

Case 4: Incident involving a transfer station

The Virginia Bureau of Radiological Health reported that a radiation monitor at a trash transfer station, located in Washington, D.C., had alarmed, indicating the presence of radioactive material. The waste hauler contractor stated that the waste was from a hospital licensee. The material was identified as a paper cup and a crushed bed pan contaminated with iodine-131. The waste hauler contractor took custody of the radioactive material and held it for decay.

Case 5: Incident involving a transfer station

The licensee reported that a BFI truck in Delray Beach, Florida set off a radiation portal monitor alarm at a transfer station. The container was returned to the licensee, who found a bag containing radioactive medical waste. The radioactive waste was inadvertently placed in the container because of a mix-up during a survey of trash bags. The mix-up occurred when two bags were being transferred to the waste container and the licensee monitor alarm tripped. The first bag was surveyed and found to contain radioactive waste. The employee set both bags aside and went to find a supervisor. While he was gone, another staff member placed the bags into the normal trash container and the container was taken to the transfer station.

Discussion:

Since operators of landfills and medical waste incinerators have installed radiation detection systems, they have become more aware of radioactive materials being shipped to these facilities.

NRC medical licensees are advised that operators of most landfills and medical waste incinerators are not licensed to manage low-level radioactive materials, and may simply reject any shipment that contains detectable levels of radioactivity, regardless of the source.


If a licensee detects radioactivity in its waste, or if an operator of a landfill or medical waste incinerator returns a waste shipment containing detectable levels of byproduct, source or special nuclear material to a licensee, absent an exemption, the licensee must manage the waste as licensed material. The licensee must evaluate the waste in accordance with 10 CFR 20.1501, "Surveys," and manage the storage/disposal of the waste in accordance with the applicable regulations and license conditions. In addition, licensees are reminded that compliance with NRC regulations does not relieve NRC licensees from complying with other local, state, and federal requirements on waste disposal.

Medical use licensees should be aware that radioactive materials may enter their waste-handling process through mechanisms largely beyond their control. Patients undergoing diagnosis and therapy, who are not required to be hospitalized, may discard contaminated items with low, but detectable, levels of radioactivity into waste containers. Therefore, detection of radioactive material in nonradioactive waste streams does not necessarily indicate poor management of radioactive waste or noncompliance with NRC requirements.

Licensees should be aware of areas in which radiation is used to ensure proper disposal of the medical radioactive waste generated. When patients are placed in temporary areas of a medical facility due to unusual circumstances, licensees should have survey systems in place to properly manage the waste. It is suggested that licensees review their programs to assure that radioactive waste is secured, and that workers are adequately trained in safety procedures.

Licensees may want to consider taking further precautions in preventing the improper disposal of radioactive material, since there are still reports of unauthorized recipients receiving nuclear material. Improper transfer of licensed materials to unauthorized recipients is a violation of NRC requirements and will be considered for enforcement action.

This information notice requires no specific action nor written response. If you have any questions about the information in this notice, please contact the technical contact listed below, or the appropriate NRC regional office.



Donald A. Cool, Director
Division of Industrial and
Medical Nuclear Safety
Office of Nuclear Material Safety
and Safeguards

Technical Contact: Anita L. Turner, NMSS
301- 415 - 5508
E-mail: alt@nrc.gov

Attachments:

1. List of Recently Issued NMSS Information Notices
2. List of Recently Issued NRC Information Notices

LIST OF RECENTLY ISSUED
 NMSS INFORMATION NOTICES

Information Notice No.	Subject	Date of Issuance	Issued to
99-32	The Effect of the Year 2000 Issues on Medical Licensees	12/17	All NRC medical licensees
99-31	Operational Controls to Guard Against Inadvertent Nuclear Critically	11/17/99	All NRC licensed fuel cycle conversion, enrichment and fabrication facilities
99-30	Failure of Double Contingency Based on Administrative Controls Involving Laboratory Sampling and Spectroscopic Analysis of Wet Uranium Waste	11/8/99	All fuel cycle licensees and certificates performing laboratory analysis to determine uranium content, in support of administrative criticality safety controls
99-29	Authorized Contents of Spent Fuel Casks	10/28/99	All power reactor licensees and spent fuel storage licensees and applicants
99-28	Recall of Star Brand Fire Protection Sprinkler Heads	9/30/99	All holders of licenses for nuclear power, research and test reactors, and fuel cycle facilities
99-27	Malfunction of Source Retraction Mechanism in Cobalt-60 Teletherapy Treatment Units	9/2/99	All medical licensees authorized to conduct teletherapy treatments
99-26	Safety and Economic Consequences of Misleading Marketing Information	8/24/99	All Distributors and/or Manufacturers of Generally Licensed Products
99-24	Broad-Scope Licensees' Responsibilities for Reviewing and Approving Unregistered Sealed Sources and Devices	7/12/99	All medical licensees' of broad-scope and master materials licensees
99-23	Safety Concerns Related to Related Control Unit Failures of the Nucletron Classic Model High-Dose-Rate Remote Afterloading Brachytherapy Devices	7/6/99	All U. S. NRC medical licensees authorized to use brachytherapy sources in Nucletron Classic Model high-dose-rate remote afterloaders

LIST OF RECENTLY ISSUED
 NRC INFORMATION NOTICES

Information Notice No.	Subject	Date of Issuance	Issued to
99-32	The Effect of the Year 2000 Issues on Medical Licensees	12/17	All NRC medical licensees
99-31	Operational Controls to Guard Against Inadvertent Nuclear Criticality	11/17/99	All NRC licensed fuel cycle conversion, enrichment and fabrication facilities
99-30	Failure of Double Contingency Based on Administrative Controls Involving Laboratory Sampling and Spectroscopic Analysis of Wet Uranium Waste	11/8/99	All fuel cycle licensees and certificates performing laboratory analysis to determine uranium content, in support of administrative criticality safety controls
99-29	Authorized Contents of Spent Fuel Casks	10/28/99	All power reactor licensees and spent fuel storage licensees and applicants
99-01, Rev. 1	Degradation of Prestressing Tendon Systems in Prestressed Concrete Constraints	10/7/99	All holders of operating licensees for nuclear power reactors
99-28	Recall of Star Brand Fire Protection Sprinkler Heads	9/30/99	All holders of licenses for nuclear power, research and test reactors, and fuel cycle facilities
99-27	Malfunction of Source Retraction Mechanism in Cobalt-60 Teletherapy Treatment Units	9/2/99	All medical licensees authorized to conduct teletherapy treatments
99-26	Safety and Economic Consequences of Misleading Marketing Information	8/24/99	All Distributors and/or Manufacturers of Generally Licensed Products
99-25	Year 2000 Contingency Planning Activities	8/10/99	All holders of OLs for nuclear power plants and fuel cycle facilities

NRC medical licensees are advised that operators of most landfills and medical waste incinerators are not licensed to manage low-level radioactive materials, and may simply reject any shipment that contains detectable levels of radioactivity, regardless of the source.

If a licensee detects radioactivity in its waste, or if an operator of a landfill or medical waste incinerator returns a waste shipment containing detectable levels of byproduct, source or special nuclear material to a licensee, absent an exemption, the licensee must manage the waste as licensed material. The licensee must evaluate the waste in accordance with 10 CFR 20.1501, "Surveys," and manage the storage/disposal of the waste in accordance with the applicable regulations and license conditions. In addition, licensees are reminded that compliance with NRC regulations does not relieve NRC licensees from complying with other local, state, and federal requirements on waste disposal.

Medical use licensees should be aware that radioactive materials may enter their waste-handling process through mechanisms largely beyond their control. Patients undergoing diagnosis and therapy, who are not required to be hospitalized, may discard contaminated items with low, but detectable, levels of radioactivity into waste containers. Therefore, detection of radioactive material in nonradioactive waste streams does not necessarily indicate poor management of radioactive waste or noncompliance with NRC requirements.

Licensees should be aware of areas in which radiation is used to ensure proper disposal of the medical radioactive waste generated. When patients are placed in temporary areas of a medical facility due to unusual circumstances, licensees should have survey systems in place to properly manage the waste. It is suggested that licensees review their programs to assure that radioactive waste is secured, and that workers are adequately trained in safety procedures.

Licensees may want to consider taking further precautions in preventing the improper disposal of radioactive material, since there are still reports of unauthorized recipients receiving nuclear material. Improper transfer of licensed materials to unauthorized recipients is a violation of NRC requirements and will be considered for enforcement action.

This information notice requires no specific action nor written response. If you have any questions about the information in this notice, please contact the technical contact listed below, or the appropriate NRC regional office.

Donald A. Cool, Director
 Division of Industrial and
 Medical Nuclear Safety
 Office of Nuclear Material Safety
 and Safeguards

Technical Contact: Anita L. Turner, NMSS
 301- 415 - 5508
 E-mail: alt@nrc.gov

Attachments:

1. List of Recently Issued NMSS Information Notices
2. List of Recently Issued NRC Information Notices

G:\TURNER\99-33.in - OFFICIAL RECORD COPY - *See Previous concurrence

OFC	MSIB*	MSIB	Editor*	MSIB*	IMNS*
NAME	ATurner/Lucy	SMoore	EKraus	JWHickey	DACool
DATE	9/12/99	12/ /99	8/12/99	10/28/99	12/12/99

This information notice requires no specific action nor written response. If you have any questions about the information in this notice, please contact the technical contact listed below, or the appropriate NRC regional office.

(s)
 Donald A. Cool, Director
 Division of Industrial and
 Medical Nuclear Safety
 Office of Nuclear Material Safety
 and Safeguards

Technical Contact: Anita L. Turner, NMSS
 301- 415 - 5508
 E-mail: alt@nrc.gov

Attachments:

1. List of Recently Issued NMSS Information Notices
2. List of Recently Issued NRC Information Notices

G:\Turnerlwaste ...wpd

To receive a copy of this document, indicate in the box: "C"=Copy without attachment/encl. "E"=Copy with attachment/encl. "N"=No copy

12-10-99 CE

OFC	MSIB	MSI B	TechED	MSIB	D/IMNS
	<i>ALT</i> ATurner	<i>JDS</i> SMoore	<i>by FAX</i> EKraus	<i>JWH</i> JWNHickey	<i>DA Cool</i> D/IMNS
DATE	9/2/99	10/2/99	9/1/99	10/28/99	12/12/99

OFFICIAL RECORD COPY