

From: Seung Lee
To: "Sergio_Steinfeld@elgems.com"@GATED.nrcsmtp
Date: Wed, Nov 10, 1999 3:21 PM
Subject: DuPont Label for MG ATC Rod Unit

Good morning, Sergio!

I talked to my staff for the line source, the lead shipping container and the carton box labels (pages 3 and 10 of 12 on 10/21/99). These labels should represent the assayed activities. Or if DuPont wants to use a fixed 450 mCi activity as shown in labels, then DuPont should provide the QA/QC procedures to ensure the line source has exact 450 mCi activity whenever it is assayed. If you have any questions, please let me know.

Thank you very much.

Seung