



GE Medical Systems

P.O. Box 414, W-709
Milwaukee, WI 53201-0414

30 July, 1999

Sueng J. Lee, Mechanical Engineer
Materials Safety and Inspection Branch
Division of Industrial and Medical Nuclear Safety
Office of on Nuclear Material and
Office of Nuclear Material Safety and Safeguards
UNITED STATES NUCLEAR REGULATORY COMMISSION
Washington, D.C. 20555-0001

Attn: Mr. Lee

Enclosed is the reply by ELGEMS to your letter dated July 7, 1999 requesting additional information for our application for MG ATC for Rod Unit p/n ASM 000415.

If you have any questions, please contact me.

Sincerely,

A handwritten signature in black ink that reads "James E. Beebe".


James E. Beebe, Ph.D.
Safety and Regulatory for X-Ray, Nuclear Medicine & PET

Tel: 414-544-3061
Pager: 414-558-7686
Fax: 414-548-5197
Email: James.Beebe@med.ge.com

OVERSIZE PAGE(S)

**NOT CONVERTED
INTO ELECTRONIC
IMAGE FORM.**



**PAPER COPY IS
AVAILABLE IN NRC
FILE CENTER.**



WIRING LIST

FROM	COLOR	TO
P605/1	WHITE	P607
P605/6	WHITE	
P605/3	WHITE	DS605
	WHITE	DS606
P605/4	RED	DS605
	RED	DS606
P605/5 SEE DETAIL A	BLACK	DS605
	GREEN	
	BLACK	DS606
GREEN		
P605/8	BLUE	DS606
P605/9	BLUE	DS605

CHANGE ORD. ELEMENT	DESCRIPTION	APPD.	DATE	REV.	CHANGE TO
R9954003	NEW DOCUMENT		15.03.99	2	



COMPONENT SOURCE INFORMATION

No.	DESCRIPTION	MANUFACTURER	PART NUMBER	QTY.
1	SLOTTED OPTICAL SWITCH	OPTEK	OPB-490	2
2	SOLENOID	DENSITRON	F401-30-287RE	
3	D-TYPE 9 PIN FEMALE	AMP	205203-1	
4	FEMALE PINS AWG 20-24	AMP	66504-2	7
5	S-R NIX	RAYCHEM	CGPT-3/1-0	35 CM
6	CABLE MARKER 3"	PANDUIT	PDL-51	2
7	CABLE TIE D-20 MM	PANDUIT	PJT-1M	
8	SCREWLOCK REV. PAIR	AMP	205817-1	
9	HT. SH. TUB. 3/32" BLK	ALPHA	FIT-221-3/32-BLACK	3 CM

775-3207-1204

ELGEMS P/N: CBL000537 VER. AA

DESIGNER	NAME	DATE	EUROPEAN PROJECT NO.	EQ. EVENT	ASSEMBLY NAME
DESIGNER	SERGIO	07.03.99	21855 218 218 218 218		
REVISION	IDA	15.03.99			
CHECKED	ZILJIC	25.03.99			
APPROVED	SHAJEL	25.03.99			
SCALE			NAME		
1.0:1.0			ATC ROD CABLE		
NON SCALE					
PAGE 1 OF 2			SHEET 1 OF 1		