

From: Seung Lee
To: inet: rflanagan@draximage.com
Date: Thu, Jun 1, 2000 9:32 AM
Subject: Requestion for Additional Information, II

Good morning, Dr. Flanagan!

I received your response yesterday, however, I need more information for leak test method and US distributor:

1. Please provide quantitative, experimental data, or other physical evidence that (1) 10 min. at 130 C and 3 atm is sufficient for the solution to penetrate the containment in those cases where the containment is permeable and likely to leak and (2) the cooling period from 130 C to room temperature is sufficient to reveal any leaks in the containment and indicate how long this period is.
2. The function of a US distributor is, in addition to licensing issues, to assure the quality of the products prior to distribution in the US. Therefore, please delineate how this function for a US distributor will be accomplished.

Please respond within 30 days of the date of this e-mail. If we have not received the requested information within 30 days, we will assume that you have abandoned your application. This assumption is, however, without prejudice and you may resubmit a complete application after that time.

If you have any questions, please let me know.

Thank you very much and have a nice day.

Seung

CC: JPJ2

Mail Envelope Properties

(393665D0.506 : 6 : 20584)

Subject: Requestion for Additional Information, II
Creation Date: Thu, Jun 1, 2000 9:32 AM
From: Seung Lee

Created By: TWFN_DO.TWF4_PO:SJL

Recipients	Action	Date & Time
Post Office inet rflanagan (inet: rflanagan@draximage.com)	Transferred	06/01 9:32 AM
Post Office TWFN_DO.twf4_po	Delivered	06/01 9:32 AM

JPJ2 CC (JPJ2)

Opened

06/01 11:27 AM

Domain.Post Office

inet

TWFN_DO.twf4_po

Delivered

06/01 9:32 AM

Route

inet:draximage.com

TWFN_DO.twf4_po

Files

MESSAGE

Size

2088

Date & Time

Thursday, June 1, 2000 9:32 AM

Options

Auto Delete:

No

Expiration Date:

None

Notify Recipients:

Yes

Priority:

Standard

Reply Requested:

No

Return Notification:

None

Concealed Subject:

No

Security:

Standard

To Be Delivered:

Immediate

Status Tracking:

Delivered & Opened