



November 17, 2011

U.S. Nuclear Regulatory Commission  
Materials Licensing Branch  
Region III  
2443 Warrenville Road  
Suite 210  
Lisle, Illinois 60532-4352

RE: License No. 24-15986-02

Please terminate license number 24-15986-02. We have transferred the following gauges to R.M. Wester and Associates, Inc. (license number 24-20091-01).

<u>Manufacturer</u>	<u>Model</u>	<u>Machine S/N</u>	<u>Cs-137</u>	<u>Am-241:Be</u>
CPN	MC3	M38048219	10mCi SN C8219	40mCi SN A8219
CPN	MC3	M310106027	10mCi SN C6027	50 mCi SN A6027

Please feel free to contact me at 314-492-0100 if you have any questions or need additional information.

Thank you,

Eric Glazier  
Radiation Safety Officer  
URS Corporation  
1001 Highlands Plaza Drive West Suite 300  
St. Louis, MO 63110

URS Corporation  
1001 Highlands Plaza Drive West  
Suite 300  
St. Louis, MO 63110  
Tel: 314.429.0100  
Fax: 314.429.0462

”

RECEIVED FEB 17 2012

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to [Infocollects.Resource@nrc.gov](mailto:Infocollects.Resource@nrc.gov), and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

## CERTIFICATE OF DISPOSITION OF MATERIALS

## LICENSEE NAME AND ADDRESS

URS Corporation  
1001 Highland Plaza Drive West 300  
St. Louis, MO 63110

## LICENSE NUMBER

24-15986-02

## DOCKET NUMBER

030-28557

## LICENSE EXPIRATION DATE

December 31, 2015

## A. LICENSE STATUS (Check the appropriate box)

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

## B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☒ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- ☒ a. Transfer of radioactive materials to the licensee listed below:  
R. M. Wester and Associates, Inc., 215 Indacom Drive, St. Peters, MO 63376 Lic. # 24-20091-01
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:

- c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

## C. SURVEYS PERFORMED AND REPORTED

- ☐ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☐ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☐ 2. A copy of the radiation survey results:
- ☐ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: \_\_\_\_\_ Date \_\_\_\_\_
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☒ a. The results of the latest leak test are attached; and/or ☒ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Eric Glazier	RSO	314.492.0100	

Mail all future correspondence regarding this license to:  
1001 Highlands Plaza Drive West 300, St. Louis, MO 63110

## C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

## PRINTED NAME AND TITLE

Eric Glazier, RSO

## SIGNATURE



## DATE

11/17/2011

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

# R. M. WESTER & ASSOCIATES, INC.

215 INDACOM DRIVE - ST. PETERS, MISSOURI 63376  
(636) 928-9628 - FAX 928-9857

## RADIOACTIVE SEALED SOURCE LEAK TEST REPORT

Test Date: October 4, 2011 Analytical Date: October 4, 2011

### Source Identification:

Manufacturer: CPN Model No.: MC3

Mach. S/N: M38048219

Radionuclide: Cs-137 Activity: 10 mCi Source S/N: C8219

Radionuclide: Am-241:Be Activity: 40 mCi Source S/N: A8219

Sample Submitted by: Kenneth Barnes

Facility: URS Corporation

Address: 1001 Highlands Plaza Dr, West 300

St Louis, MO 63110

The identified sealed source listed above has been tested for leakage of radioactive materials as required by the United States Nuclear Regulatory Commission. The analysis of the wipe material used in testing the sealed source reveals the presence of  $\leq 9.69 \times 10^{-5}$   $\mu\text{Ci}$  of loose contamination for Cs-137 and  $\leq 1.01 \times 10^{-4}$   $\mu\text{Ci}$  of loose contamination for Am-241.

( XX ) This source is acceptable for continued use.

( ) This source has been found to have a level of loose contamination greater than 0.005  $\mu\text{Ci}$  of removable radioactive materials, and should be removed from service immediately.

( N/A ) Operational and performance check of shutter mechanism satisfactory.

Next leak test due by: April 4, 2012

Analysis by: Kenneth Barnes Date: October 4, 2011

Reviewed by: *Kenneth Barnes* Date: October 4, 2011

# R. M. WESTER & ASSOCIATES, INC.

215 INDACOM DRIVE - ST. PETERS, MISSOURI 63376  
(636) 928-9628 - FAX 928-9857

## RADIOACTIVE SEALED SOURCE LEAK TEST REPORT

Test Date: October 4, 2011 Analytical Date: October 4, 2011

### Source Identification:

Manufacturer: CPN Model No.: MC3

Mach. S/N: M310106027

Radionuclide: Cs-137 Activity: 10 mCi Source S/N: C6027

Radionuclide: Am-241:Be Activity: 40 mCi Source S/N: A6027

Sample Submitted by: Kenneth Barnes

Facility: URS Corporation

Address: 1001 Highlands Plaza Dr, West 300

St Louis, MO 63110

The identified sealed source listed above has been tested for leakage of radioactive materials as required by the United States Nuclear Regulatory Commission. The analysis of the wipe material used in testing the sealed source reveals the presence of  $\leq 9.69 \times 10^{-5}$   $\mu\text{Ci}$  of loose contamination for Cs-137 and  $\leq 1.01 \times 10^{-4}$   $\mu\text{Ci}$  of loose contamination for Am-241.

( XX ) This source is acceptable for continued use.

( ) This source has been found to have a level of loose contamination greater than 0.005  $\mu\text{Ci}$  of removable radioactive materials, and should be removed from service immediately.

( N/A ) Operational and performance check of shutter mechanism satisfactory.

Next leak test due by: April 4, 2012

Analysis by: Kenneth Barnes Date: October 4, 2011

Reviewed by: Kenneth Bachmann Date: October 4, 2011



**URS**

URS Corporation  
1001 Highlands Plaza Drive West  
Suite 300  
St. Louis, MO 63110

U.S. Nuclear Regulatory Commission  
Materials Licensing Branch  
Region III  
2443 Warrenville Road  
S210  
Lisle, IL 60532

